

Northumbria Research Link

Citation: Nguyen, T. D., Fu, Richard, Tran, V. T., Gautam, A., Pudasaini, S. and Du, Hejun (2020) Acoustofluidic closed-loop control of microparticles and cells using standing surface acoustic waves. *Sensors and Actuators B: Chemical*, 318. p. 128143. ISSN 0925-4005

Published by: Elsevier

URL: <https://doi.org/10.1016/j.snb.2020.128143>
<<https://doi.org/10.1016/j.snb.2020.128143>>

This version was downloaded from Northumbria Research Link:
<http://nrl.northumbria.ac.uk/id/eprint/42826/>

Northumbria University has developed Northumbria Research Link (NRL) to enable users to access the University's research output. Copyright © and moral rights for items on NRL are retained by the individual author(s) and/or other copyright owners. Single copies of full items can be reproduced, displayed or performed, and given to third parties in any format or medium for personal research or study, educational, or not-for-profit purposes without prior permission or charge, provided the authors, title and full bibliographic details are given, as well as a hyperlink and/or URL to the original metadata page. The content must not be changed in any way. Full items must not be sold commercially in any format or medium without formal permission of the copyright holder. The full policy is available online: <http://nrl.northumbria.ac.uk/policies.html>

This document may differ from the final, published version of the research and has been made available online in accordance with publisher policies. To read and/or cite from the published version of the research, please visit the publisher's website (a subscription may be required.)

Acoustofluidic closed-loop control of microparticles and cells using standing surface acoustic waves

T. D. Nguyen,¹ Y. Q. Fu,² V.T. Tran,¹ A. Gautam¹, S. Pudasaini¹, H. Du,^{1, a)}

¹*School of Mechanical and Aerospace Engineering, Nanyang Technological University, 639798, Singapore*

²*Faculty of Engineering and Environment, Northumbria University, Newcastle upon Tyne, NE1 8ST, UK*

Abstract: Precise, automatic and reliable position control of micro-objects such as single particles, biological cells or bio-organisms is critical for applications in biotechnology and tissue engineering. However, conventional acoustofluidic techniques generally lack reliability and automation capability thus are often incapable of building an efficient and automated system where the biological cells need to be precisely manipulated in three dimensions (3D). To overcome these limitations, we developed an acoustofluidic closed-loop control system which is combined with computer vision techniques and standing surface acoustic waves (SSAWs) to implement selective, automatic and precise position control of an object, such as a single cell or microparticle in a microfluidic chamber. Position of the object is *in situ* extracted from living images that are captured from a video camera. By utilizing the closed-loop control strategy, the object is precisely moved to the desired location in 3D patterns or along designed trajectories by manipulating the phase angle and power signal of the SSAWs. Controlling of breast cancer cells has been conducted to verify the principle and biocompatibility of the control system. This system could be employed to build an automatic system for cell analysis, cell isolation, self-assembling of materials into complex microstructures, or lab-on-chip and organ-on-chip applications.

1. Introduction

Precise, automatic and reliable control of single cells is essential in biotechnology and tissue engineering for tasks of cell sorting and manipulation [1-6], cell analysis and screening [1, 5, 7-9], cell isolation [8, 10], self-assembly of cells and tissues [11-15]. The common technique using optical tweezers [16, 17] could achieve a high precision but usually face the problems of relatively weak forces and highly intense power that causes the damage to the living cells [18]. In the past decades, surface acoustic waves (SAWs) have been extensively studied and considered as a powerful tool to replace the optical tweezer for applications in microfluidics and biomedical engineering due to the noninvasive, biocompatible, label-free, low-power consumption and inexpensive technique [19-21]. SAWs are generated by the propagation of mechanical waves on surface of elastic materials, and these waves have been applied to control the fluid and particles inside the fluid for numerous microfluidic applications [19, 20, 22]. The acoustic tweezers created from the standing SAWs (SSAWs) have been used to precisely trap, transport and manipulate objects within the 2D plane by either modulating the operating frequency or shifting the relative phases of resonant-frequency signals generated from the two opposite inter-digital transducers (IDTs) [18, 23-25]. For manipulating purposes, the phase shifting method has more advantages than frequency sweeping, such as better flexibility and larger working range [26]. In term of single-cell manipulation, SSAWs are able to trap a wide range of the cells with different sizes, shapes, densities and compressibility properties [19].

^{a)} Authors to whom correspondence should be addressed: MHDU@ntu.edu.sg

1 The current manipulation techniques based on SSAWs are mostly using open-loop control
2 strategy which delivers the output manipulation (e.g. shifting the phase or modulating the
3 frequency) independent of the motions of the manipulating objects, (e.g. particle, bubble and cell).
4 These systems have been applied in the fields of microfluidics, especially the manipulation of
5 micro/nanoparticles, microbubbles, cells, molecules to generate multiple functions of focusing [27-
6 29], manipulating [30, 31], separating [32-42], isolating [10, 43], patterning [24, 44-46] and
7 positioning [2, 3, 47]. SSAWs have extensively been used to control rotation and transportation of
8 single cells in the past few years [5, 10, 18, 23, 44, 48-54]. However, the general open-loop control
9 systems face significant challenges, such as unreliability and time-consuming process due to
10 manual control and no detailed information for changes of object's positions. For example, they
11 cannot recognize the unexpected motions of objects caused by noise or external vibrations.
12 Therefore, it is critically demanded for a system to monitor these changes and efficiently adjust
13 object's positions. To solve this problem, a closed-loop control strategy could be implemented to
14 extract the exact position of the objects and precisely adjust their movement. The closed-loop
15 control strategy has been applied in microfluidic system to control the volume of droplet and fluid-
16 height in open reservoirs over recent years [55-58]. However, to the best of our knowledge, there
17 is no any previous report for using close-loop control systems for acoustofluidic devices.

18 In this study, we proposed the integration of a control system into an acoustofluidic device in
19 order to improve the controlling efficiency of cells or microparticles in 3D space using the SSAWs.
20 A computer vision algorithm was employed to detect the current positions of the object from the
21 captured image, which serves as a feedback signal. Based on the feedback, a closed-loop controller
22 then transmits the manipulation signal to adjust position of the particle by shifting the relative phase
23 angles of the frequency signals generated by the opposite IDTs for in-plane motion (x-y
24 coordinates), or adjusting the SAW power for out-of-plane motion (z coordinate). It is worthwhile
25 to note that the integration of closed-loop control strategy technically would not improve the
26 resolution of the SAW device, for example, the resolution of phase-shifting method still depends
27 on the phase division of equipment resolution. The characteristics of the closed-loop control system
28 tested on polystyrene microparticles such as speed, resolution and response time were investigated.
29 Thereby, a microparticle with diameter of 20 μm was separated from the 10 μm microparticles at
30 the speed limit. Automatic control of single microparticle to one location and to follow a 2D
31 arbitrary trajectory was performed. Combination of in-plane and out-of-plane control system was
32 performed to demonstrate the 3D controlling capability. Breast cancer cells (MCF-7) were used to
33 test the principle and compatibility of the control system for biological applications.

34 **2. Method**

35 **2.1. Device fabrication**




Fig. 1. (a) An optical image of the device after fabrication including interdigital transducers (IDTs), polydimethylsiloxane (PDMS) chamber, inlet and outlet. (b) The aligned PDMS chamber filled with DI water.

Fig. 1(a) shows an optical image of the assembled SAW system, and Fig. 1(b) shows the working area inside the chamber. The piezoelectric substrate is 128° Y-cut lithium niobate (LiNbO_3). The IDTs were patterned using the standard photolithography followed by a lift-off process to obtain metal electrodes (Cr/Au, 10 nm/50 nm) on the LiNbO_3 wafer. The IDTs have 60 pairs of fingers with both width and space gap of $75 \mu\text{m}$, corresponding to a SAW wavelength of $300 \mu\text{m}$ and an aperture of 1 cm. A cubic polydimethylsiloxane (PDMS) chamber (Sylgard 184 Silicone Elastomer; Dow Corning, USA) was aligned with the IDT patterns to ensure the forming lines of trapped objects are parallel to the IDT fingers and bonded on top of the piezoelectric substrate after being treated with oxygen plasma. The chamber is 1.5 mm wide (covering five SAW wavelengths), $100 \mu\text{m}$ tall for 2D manipulation or $1000 \mu\text{m}$ for 3D manipulation, and it was prepared using a soft casting method and a mold made using a computer numerical controlled (CNC) machine.

2.2. Cell preparation

Michigan Cancer Foundation-7 (MCF-7) breast cancer cell line, purchased from the American Type Culture Collection, were cultured in Dulbecco's Modified Eagle Medium (Gibco; Thermo Fisher Scientific, USA) supplemented with 10% FBS and 5% antibiotic. The cells were maintained to grow in a humidified environment at 37°C and 5% CO_2 for three to four days until 70 to 80% confluence was reached. After that, they were trypsinized with 0.05% trypsin and collected by centrifuging at $250g$ for 5 mins. The cell pellet was re-suspended in the media and counted by hemocytometer.

2.3. System setup

To record the microparticle or cell images, a charge-coupled device (CCD) camera (DFK 51BG02.H; IVS Imaging, USA) was connected to an optical microscope (M150 Probe Station; Cascade Microtech, USA). The controller program was written in MATLAB R2018b (MathWorks, USA) embedded in the laptop (Ideapad 310; Lenovo, China). The controller was connected to the camera to acquire the live images and the signal generators (MHS-5200P; Ming Wo Electronics, China) to send the manipulating signal (Fig. 2). The signal was then amplified by an RF amplifier (the amplifier characteristics are provided in Fig. S1) before being applied to the IDTs to generate the SSAWs. The resonant frequency was measured using an impedance analyzer (4294A Impedance Analyzer; Agilent, USA). We performed frequency-sweep measurements to scan a wide frequency range, and the resonant frequency was obtained at its highest impedance. The resonant frequencies for the IDT pairs along two orthogonal directions were measured to be 13.193

1 MHz and 12.124 MHz, due to the isotropically piezoelectric property of LiNbO₃ substrate. The
 2 microparticles of two different sizes (polystyrene microspheres with average diameters of 10 μm
 3 and 20 μm; Polysciences, USA) and MCF-7 cells were injected into the chamber using a pipette.

4 **2.4. Working principle**

5
 6 When the SAW device is operated at its resonant frequency, the waves generated from the two
 7 identical but opposite IDTs will propagate on the substrate and interact with the fluid to form the
 8 leaky SAWs and also create the SSAWs [59]. A pressure gradient is generated inside the liquid,
 9 forming pressure nodes (minimum pressure amplitude) and anti-pressure nodes (maximum
 10 pressure amplitude). The suspended objects in fluid are subjected to the combined forces: e.g., (1)
 11 the acoustic radiation force generated from the pressure fluctuations; (2) the Stoke drag force
 12 caused by the acoustic streaming which arises from the nonlinear propagation of a SAW through
 13 an attenuating media [60]; (3) gravity force; and (4) buoyant force. The acoustic radiation force is
 14 given by [61]:

$$F_{\text{rad}} = -\nabla \left\{ V_p \left[\frac{f_1}{2\rho_0 c^2} \langle p_1^2 \rangle - \frac{3\rho_0 f_2}{4} \langle u_1 \cdot u_1 \rangle \right] \right\} \quad (1)$$

where

$$f_1 = 1 - \frac{\rho_0 c^2}{\rho_p c_p^2} \quad (1a)$$

$$f_2 = \frac{2(\rho_p - \rho_0)}{2\rho_p + \rho_0} \quad (1b)$$

15 in which u_1 is the first-order velocity, p_1 is the first-order pressure, V_p is the volume of spherical
 16 object, ρ_p and c_p are density and speed of sound of object, respectively. As we are considering a
 17 quiescent liquid (e.g., water) before the presence of any acoustic wave, the initial pressure and
 18 density are assumed as constants, i.e., p_0 = static pressure, ρ_0 = 998 kg/m³.

19 The Stoke drag force is given by [62]:

$$F_d = 6\pi\mu R (\langle v_2 \rangle - v_p) \quad (2)$$

20 where R is the radius of the object, v_p is the object velocity and v_2 is the second order velocity or
 21 acoustic streaming velocity.

22 From equations (1) and (2), the acoustic radiation force is proportional to the volume (R^3) of the
 23 object while the drag force is proportional to the radius (R) of object. Therefore, the larger size
 24 object are experienced with the larger acoustic radiation force, thus they move much faster to either
 25 the pressure nodes or anti-pressure nodes, depending on the relative density and compressibility of
 26 object to medium [63]. In our experiment, we used the polystyrene beads in the DI water and the
 27 MCF-7 cells in cultured medium, which move towards the pressure nodes when experience the
 28 acoustic radiation force.

29 As the objects are trapped at the acoustic pressure nodes or acoustic tweezers, the in-plane
 30 movement of objects is realized by tuning acoustic tweezer's position. This is achieved by simply
 31 shifting the signal phase-angle of opposite IDTs. If the chamber is high enough (e.g., 1 mm), 3D
 32 aligned patterns with multilayers of 3D acoustic tweezers along the height (or vertical) direction
 33 are generated due to the reflection of acoustic waves from substrate to the top of chamber [45, 49].
 34 The interlayer distances are identical and determined by f/c , where f is the operating frequency and
 35 c is speed of sound in fluid. The object could be trapped into these 3D acoustic tweezers or
 36 levitated/dropped by manipulating the SAW signal power.




Fig. 2 Illustration of closed-loop control system operation. A computer is connected to a camera to track the microparticle's position on feedback line and a SAW device through signal generators and amplifiers (the black box) on forward line. The trapped microparticle can move along x and y directions by shifting the relative phase of signal from opposite IDTs or z direction by turning the power. The computer's monitor shows a tracked microparticle was marked by the red "+" sign and the target position is marked by the black "x" sign

1
 2 The setup and operation of the in-plane closed-loop control system on a microparticle are illustrated
 3 in Fig. 2. It includes a controller (i.e., a computer's algorithm written in MATLAB code), which is
 4 connected to a camera (on the feedback line) and the SAW device through signal generators and
 5 amplifiers (on the forward line). The computer's monitor shows the selected single microparticle
 6 after being trapped into an acoustic tweezer (marked with the red "+" sign) and the target position
 7 (marked with the black "x" sign). The distances between the microparticle and the target position
 8 along both the x and y directions are called the distance errors. The objective of using the closed-
 9 loop controller is to minimize these errors (including x error and y error) to effectively manipulate
 10 the microparticle to the target position. In operation, the controller initially detects the current
 11 position of the microparticle from the optical image, which is *in-situ* captured by microscope's
 12 camera, using a computer vision algorithm on the feedback line. This algorithm relies on the
 13 contrast between black microparticle and the white background, the "regionprops" command in
 14 MATLAB is applied to determine the centroid's position of microparticle. The distance error
 15 between the target position and the microparticle's position is then calculated. As the response, the
 16 controller will change (either increase or decrease) the relative phase angles of the radio-frequency
 17 signal on the forward line to relocate microparticle's position so that the distance error is
 18 minimized. The loop operates in the real time and will be paused when the distance errors in both
 19 x and y directions are smaller than a constant tolerant value, which means that the microparticle
 20 has reached the target position. The purpose of using a constant tolerant value of 5 pixels is to
 21 shorten the matching process. At this time, the controller will load the next target position and
 22 repeat the loop process until the microparticle reaches the targeted positions. The flowchart of
 23 algorithm and the block diagram are presented in Figs. S2 and S4, respectively, in the **SI document**.

24 25 **3. Results and discussion**

26 **3.1. Response time, speed measurement and separation of different-sized microparticles**

27 We firstly evaluated the closed-loop control system by investigating the dynamic response a
 28 single microparticle, as a function of time. Fig. 3(a) and 3(b) show the obtained displacement

1 response curves of a single microparticle with a diameter of 10 μm along x and y directions,
 2 respectively. On each direction, the microparticle took around 30 secs to reach the target position
 3 over the distance of 500 pixels. It should be noted that the distance of 1 μm along x direction
 4 **corresponds to 1.8 pixels, whereas along y direction it corresponds to 2.1 pixels**. Fig. 3(c) shows
 5 the stacked image of the same microparticle before and after applying the closed-loop control,
 6 simultaneously, on both x and y directions to relocate it to a 2D target position (the black plus sign
 7 on the left corner) over the distance of 980 pixels. The response time is strongly dependent on
 8 phase-shifting speed. The linear relationship between the phase change and displacement has been
 9 reported previously in literature [64]:

$$\Delta d = (n-1) \frac{\lambda}{2} + \frac{\lambda}{720} \varphi \quad \varphi \in [0^0, 360^0] \quad (3)$$

10 where Δd is the displacement, $n = 1, 2, 3, \dots$ represents the number of the repetition of changing the
 11 relative phase from 0^0 to 360^0 ; φ is the relative phase opposite IDTs; λ is the wavelength of the
 12 SAW.




Fig. 3. (a) and (b) Response times of a single microparticle with a 10 μm diameter after being controlled to move from 600 to 100 pixels position with the phase-shifting speeds are 8.7 deg/s and 8.79 deg/s along x and y directions, respectively. (c) The stacked images of microparticle's position before and after being controlled simultaneously both x and y directions to move to the target position which is the black "+" sign on the left corner.

13
 14 The in-plane resolution of the SAW device is defined as the minimum distance of object
 15 movement or the smallest increment of displacement that the device can achieve. Since our signal
 16 generator has resolution of 1 degree and the SAW device has the designed wavelength of 300 μm ,
 17 hence, the theoretical minimum displacement for shifting the phase for 1 degree is calculated to be
 18 $\sim 0.42 \mu\text{m}$ (i.e., $300\mu\text{m} \times 1\text{degree}/720\text{degree}$) based on equation (3). In our experiments, it is
 19 challenging to detect this tiny displacement of microparticles by using a standard video camera.
 20 Instead, we recorded the transportation distance of a single microparticle with a diameter of 10 μm
 21 after the phase angle was shifted for 20 degrees. Results showed that the microparticle was moved
 22 ~ 15 pixels in both x and y directions. However, it should be addressed that the pixel's sizes are
 23 slightly different in two directions, which is defined as the pixel aspect ratio of camera sensor. The
 24 dimensions of a single pixel along x and y directions are 2.1 pixels/ μm and 1.8 pixels/ μm ,
 25 respectively. Accordingly, the in-plane resolution of the SAW device was calculated to be ~ 0.36
 26 $\mu\text{m}/\text{deg}$ along the x direction and $\sim 0.42 \mu\text{m}/\text{deg}$ along the y direction which approximately similar
 27 to the theoretical value. It is worthwhile to note that the resolution of the SAW device does not
 28 reflect the resolution of the control system. The resolution of a control system is defined as the

1 smallest movement of microparticle that can be detected by the camera. In our control system, it is
 2 equal to the size of one pixel which is $\sim 0.48 \mu\text{m}/\text{pixel}$ along x direction and $\sim 0.56 \mu\text{m}/\text{pixel}$ along
 3 y direction. Based on this, using a higher resolution camera paves a way to build a higher resolution
 4 control system. With SAW devices operated at a higher frequency or with a smaller wavelength,
 5 their resolution could be further improved. For example, for a SAW device which has a wavelength
 6 of $0.8 \mu\text{m}$ [65], the corresponding resolution is calculated to be $\sim 1.1 \text{ nm}$ (i.e., $0.8/720$). However,
 7 it might then become time-consuming and impractical to control over a large distance at a very
 8 high resolution. This can be resolved by integrating multiple wavelengths into a single IDT [66].
 9 Whereby, the low-resolution manipulation (i.e., using devices with larger wavelengths or smaller
 10 frequencies) can be used for a rough adjustment of movement while the high-resolution
 11 manipulation (i.e., using devices with smaller wavelengths or higher frequencies) for a fine
 12 adjustment of precise movement.




Fig. 4. (a) The relationship between microparticle's velocity and phase-shifting speed. At shifting speed higher than 30 deg/s , the microparticle will be extricated from acoustic tweezer and oscilated back and forth. (b) The response time of $10 \mu\text{m}$ diameter microparticle moved from position 600 pixels to 100 pixels at different phase-shifting speeds along x direction. (c) Controlling the $20 \mu\text{m}$ and $10 \mu\text{m}$ diameter microparticles to the right direction at the speed of 30 degrees/s . Before controlling, all microparticle trapped and aligned along vertical red line. After controlling, $20 \mu\text{m}$ diameter microparticle was separated from $10 \mu\text{m}$ diameter microparticles (scale bar: $50 \mu\text{m}$).

13
 14 Generally, the increase of phase-shifting speed leads to a higher velocity of the microparticle. In
 15 our experiment, we investigated the relation between phase-shifting speed and microparticle's
 16 velocity along x direction. Fig. 4(a) shows that the changes of the microparticle's velocity has a
 17 linear relationship with the phase-shifting speed. With an increase of phase-shifting speed (equal
 18 to or larger than a limit of 30 deg/s), we found that the microparticles could be extricated from the
 19 trapping acoustic tweezer (Fig. 4(a)). This can be explained that while the acoustic radiation force
 20 pushes the microparticle to follow the moving acoustic tweezer, but the drag force caused by the
 21 motion of microparticle in the fluid as described by equation (2) has the opposite direction to pull
 22 the microparticle out of acoustic tweezer. The higher the microparticle's speed, the stronger the
 23 drag force is and thus the easier for the particles to be extricated. However, because the SSAWs
 24 have created an acoustic tweezer pattern with equal intervals [24], after extricated from one

1 acoustic tweezer, the microparticle will be trapped to the next acoustic tweezer. As a result, this
2 phenomenon causes the microparticle to oscillate back and forth (e.g., trapped then released). A
3 video recorded for this phenomenon is shown in the **SI-video 1**.

4 Based on the equation (1), there are three ways which we can increase the speed limit through
5 increasing the amplitude of acoustic radiation force, in the other words, increasing the trapping
6 strength of acoustic tweezer. (i) Firstly, increasing the operating power in order to increase the
7 first-order acoustic pressure. However, applying a higher power will induce more heat inside the
8 chamber [67]. The results of temperature change of fluid inside microchamber for 10 minutes are
9 presented in Figure S3 of **SI document**. Our results suggested that the generated heat can be
10 minimized by integrating a Peltier cooler beneath the SAW device; (ii) Secondly, using larger size
11 microparticles or gathering small microparticles into a large group and (iii) Finally, increasing
12 operating frequency, which will strengthen the acoustic radiation force [44].

13 In fact, the acoustic radiation forces generated from the acoustic pressure field are strongly
14 dependent on the size of microparticles [61, 68]. Various microfluidic devices have relied on this
15 principle to continuously separate or sort the microparticles or cells from different sizes and
16 densities [10, 34, 36-39, 41, 69]. Herein, we took advantage of above extrication phenomenon to
17 separate different-sized microparticles. Two sizes of polystyrene microparticles with diameters of
18 10 μm and 20 μm were used for this investigation. At the initial stage, all of microparticle were
19 trapped into an acoustic tweezer and aligned to the same vertical line (as shown in the red dash in
20 Fig. 4(c)) after the power was turned on to the horizontal pair of IDTs at a power of 28.5 dBm.
21 Then, the relative phase was shifted at a speed of 30 deg/s rightward. The result in Fig. 4(c) shows
22 that the 10 μm microparticles were extricated from the acoustic tweezer whereas the 20 μm
23 microparticles were still trapped and moved. This is because the 20 μm microparticles experienced
24 a larger (~ 8 times based on equation (1)) acoustic radiation force than the 10 μm microparticles,
25 and thus are able to follow the shifting acoustic tweezers while the 10 μm microparticles were
26 extricated. This phenomenon is useful for isolation and separation applications such as separating
27 a specific cell [43].

28 We have further measured the response time with the increase of phase-shifting speed at a given
29 power of 28.5 dBm. The results in Fig. 4(b) show the response time for the microparticle to reach
30 the desired position of 100 pixels is inversely proportional to phase-shifting speed. The shorter the
31 response time, the faster the microparticle moves to the desire location, which generally increases
32 the efficiency of the system.

34 **3.2. Demonstration of 2D, 3D and biological cell manipulation**

35 In this section, we applied the closed-loop system to control a single microparticle in a designed
36 arbitrary trajectory to draw a letter of “NTU”. This is achieved by setting a chain of target positions
37 in the shape of the “NTU” letters written using AutoCAD software as shown in Fig. 5(a). Fig. 5(d)
38 shows how a tracked microparticle approaches and arrives to a target point in the dash line boxes
39 and then continue to move to another target point in the continued line boxes. It should be pointed
40 out that the target position will be changed to the new value when the distance errors are smaller
41 than 5 pixels. Fig. 5(b) shows the experimental trajectory of the microparticle. Each point in this
42 trajectory was the captured position of the microparticle. Fig. 5(c) shows a combined image by
43 stacking multiple images of microparticle at different moments. A recorded video of the whole
44 process is presented in the **SI-video 2**.




Fig. 5. Manipulation of 10 μm diameter single microparticle to write a “NTU” letter. (a) The “NTU” letter which are point series designed using AUTOCAD software. (b) The recorded trajectory of microparticle during the continuous manipulation. (c) The stacked images of microparticle at different moments to show the word of “NTU” (scale bar: 50 μm). (d) The approaching of the microparticle to the first (dash line boxes) and second (continued line boxes) target point.

1
2 In Fig. 5(b), the trajectory of the microparticle are not smooth at some points. This is caused by the
3 external vibrations from ground or equipment and also from the effect of the acoustic streaming
4 that leads to the extrication of microparticle from acoustic tweezer. However, the closed-loop
5 system can regulate the microparticle back to the desired path afterward. To eliminate the external
6 noise, the whole experimental system is recommended to be set on an anti-vibration table.
7 Additionally, an advanced controller such as artificial intelligent algorithm will further improve
8 the performance of the controlling system [70]. The acoustic streaming dominates the motion of
9 microparticles when their sizes are below the frequency-dependent critical crossover size, and vice
10 versa, the acoustic radiation force becomes dominant again [5, 71]. In order to reduce the effect of
11 acoustic streaming, the microparticle’s size is needed to be above their critical size, otherwise, the
12 device needs to be redesigned at a higher operating frequency. Due to the automatic and precise
13 control of microparticle, this method could be employed to couple with the lab-on-chip and organ-
14 on-chip device for precise transportation purposes. This could assist the studies of cell such as cell-
15 cell interaction and communication by controlling the distance of cells or fixing the position of the
16 single cell for analysis or transfection [2].




Fig. 6. (a) The stacked images of 20 μm diameter microparticle at different moments moved along a rectangular shape which observed from the side view. The arrows show the moving direction. (b) Experimental vertical view of the chamber's top surface or PDMS roof and the chamber's bottom surface or substrate. (c) Diagram describe the set-up of target height same as the focal plane and the lifting of microparticle from layer 1 to layer 2 by increasing the power. (d) The experimental result of lifting microparticle with diameter of 10 μm from unfocused to be focused (bottom to top images) at focal plane (scale bar: 50 μm).

1
2 3D assembly using programmable system in combination with acoustic tweezer has seldom been
3 reported. In addition to the manipulation in the x-y coordinates (in-plane), the manipulation along
4 z coordinate (out-of-plane) has been achieved by turning the power applied to SAW device [18,
5 45]. In the 3D chamber, the SSAWs generate multiple and periodic trapping layers along z direction
6 [45]. On each layer, the acoustic radiation force is directed to trapping nodes. After trapping, the
7 microparticles can be either levitated to upper layers by turning on the power of SAW device; or
8 dropped down to lower layers by switching off the SAW power. Basically, turning on the SAW
9 power will increase the difference between upward and downward acoustic radiation forces applied
10 to the microparticle and thus result in a stronger upward net force. This force levitates the
11 microparticle to the higher layer. Afterward, the microparticle can be locked into that new layer by
12 returning the power to the previous value. The dropping of the particle occurs after switching off
13 the power because of the gravity, which is strongly dependent on the weight of microparticle.
14 In terms of controlling range, the manipulation is capable of operating in the whole range of 1 mm
15 height of the chamber. For example, a high power will raise the microparticles to the top surface
16 of the chamber. In term of vertical resolution of the SAW device, the interlayer distance of layers
17 defines the resolution along z direction. In our experiment, this distance is approximated about 60
18 μm and could be reduced by changing to the device with higher operating frequency. In terms of
19 time scale, the raising-up of particle could be quickly achieved but the dropping may be time
20 consuming. Therefore, the heavier microparticles not only accelerate the process but also reduce
21 the effect of the acoustic streaming, thus facilitating the efficient controlling process. The smaller
22 and lighter microparticles become heavier when forming into groups with the aid of SSAWs
23 aggregation [24]. Herein, we combined the out-of-plane control technique with the in-plane control
24 technique to enable the three-dimensional manipulation capability. A demonstration in Fig. 6(a)
25 shows the side-view of 20 μm diameter microparticle at different moments moved along a
26 rectangular shape. The prism was applied to enable the horizontal view as mentioned in our
27 previous study [45]. Initially, the microparticle was fixed at its initial position when turning on the
28 power of 27.1 dBm (point A). After that, it was levitated over a distance of 178 μm by increasing
29 the power to 33.4 dBm and then returning to 27.1 dBm immediately to fix the microparticle to the
30 new position (point B). Phase-shifting was then applied to move the microparticle to the left over

1 a distance of 133 μm (point C). The power was subsequently turned off to drop the microparticle
 2 and turned on again when it reached the same height as the beginning point (point D). Finally, we
 3 shifted the phase in the opposite direction to bring it back to the original position (point A). A
 4 recorded video of this process can be found in the **SI-video 3**. In this recorded video, it is easy to
 5 realize that the microparticle was stopped at different layers when levitating from point A to point
 6 B.

7 In order to integrate the out-of-plane control into the same closed-loop system with the in-plane
 8 control system, we propose to apply the focus function of the microscope to provide a feedback for
 9 the vertical position of the microparticle. Prior to applying the control, the targeted height position
 10 of microparticle relative to the bottom surface of chamber needs to be determined based on the
 11 height of multilayers of 3D pattern where the microparticles are capable to be trapped.
 12 Subsequently, the microscope is focused at the substrate's surface and the exact position of focus
 13 knob is recorded as a reference's value (Fig. 6(b)). Then, we turn the focus knob to the new value
 14 which the subtraction from the reference's value is equal to the target height (layer 2 in Fig. 6(c)).
 15 As long as the focal plane match the target height position, the microparticle will be
 16 levitated/dropped until its image under microscope is sharpest or in-focus. This means that the
 17 vertical position of microparticle has reached to the targeted height position. A demonstration is
 18 shown in Fig. 6(d), in which we turned the power to lift a microparticle with diameter of 10 μm
 19 from the unfocused position to the focused position in a 1 mm-height chamber at the height of 780
 20 μm (about 12th layer from the bottom surface). The block diagram of this out-of-plane control
 21 system has been presented in Figure S5 in the **SI document**. A combination of x-y and z closed-
 22 loop control results in a three-dimensional closed-loop control system. The object could be
 23 controlled in the x-y coordinates and then along the z coordinate, or vice versa. In tissue
 24 engineering, this is a promising technique to build up a micro-bio-system for assembling of 3D
 25 complex constructs of biomaterials.




Fig. 7. The closed-loop control of a single and a group of breast cancer cells. (a) The stacked images of a single MCF-7 after controlling and its rotation direction. (b) The recorded trajectory of the single cell to target position at (100, 100) pixels. The zoomed image on the right shows actual trajectory along the dash line while the solid lines on two sides are noises. (c) Before and after being controlled of a group of MCF-7 cells to the target position at (100, 100) pixels.

26 The application of SAWs for biological cells has been extensively studied in the literature [18,
 27 23, 24, 44, 51-53, 66, 69, 72, 73]. To verify the compatibility of the closed-loop control system
 28 with biological cells, we performed the closed-loop control of single breast cancer cell (MCF-7)
 29 from a random position to the targeted position at the coordinate $(x, y) = (100, 100)$ pixels. The
 30 cell was kept in its original medium with added Trypan Blue stain (0.4%, Life Technologies, USA)
 31

1 for the viability test. Fig. 7(a) shows the stacked images of the tracked cell at different durations.
2 The random shape and rotation of the cell during its moving caused difficulty for the computer
3 vision algorithm to detect the centroid position of the cell. Although there are apparent noises as
4 shown in Fig. 7(b), it was still in a good control pattern because of the high contrast between the
5 cell and background. Moreover, we can extract the real movement of the cell from the recorded
6 trajectory as shown in the zoomed image in Fig. 7(b). A digital filter is recommended to integrate
7 into the computer algorithm, in order to eliminate most of these noises. The cell was still alive after
8 manipulation as it did not take up the dye. A recorded video of this process can be found in the **SI-**
9 **video 4**.

10 In addition, as the cell would rather attach the others and thus form groups, therefore, we have
11 exploited the closed-loop system to control a group of cells. Fig. 7(c) shows the result of a group
12 of MCF-7 cells before and after being relocated to the target position at (100, 100) pixels. In this
13 case, the algorithm was defined to detect many cells at once which could confuse the controller
14 about the current position of the cell. However, our results showed that as long as the distance
15 errors are still existed, the controller kept shifting the relative phase in order to minimize the
16 distance errors. As a result, the group of cell was relocated to the targeted position at (100, 100)
17 pixels. A recorded video of this process can be found in the **SI-video 5**.

18 **4. Conclusion**

19 In conclusion, by integrating a closed-loop control strategy with the acoustic tweezer, our
20 acoustofluidic closed-loop system paves the way for controlling microparticle or cell in an effective
21 manner. A single microparticle is able to follow any desired trajectory automatically in a 3D pattern
22 by the automatic manipulation of relative phases and power of the SAW device. The characteristics
23 of the closed-loop control system have been investigated. Notably, the speed limit of shifting
24 acoustic tweezers was explored to use in order to separate the larger size objects, which is suitable
25 for isolation and separation applications. Controlling of a single and a group of breast cancer cells
26 have been conducted to verify the principle and biocompatibility of the control system. This
27 acoustofluidic closed-loop control system serves as an effective tool to develop an automatic
28 process for cell-cell interaction, cell analysis, cell manipulation as well as self-assembling of
29 materials into complex microstructures. The aid of computer vision in the closed-loop system is
30 also useful to collect the cell images for screening studies.

31
32
33 **Acknowledgment:** The authors gratefully acknowledge the support of (i) Nanyang Technological
34 University and the Ministry of Education of Singapore through a PhD Scholarship; (ii) the UK
35 Engineering and Physical Sciences Research Council (EPSRC) grants EP/P018998/1; (iii) Special
36 Interesting Group of Acoustofluidics funded by UK Fluids Network (EP/N032861/1).

37 **CRedit authorship contribution statement**

38
39 **Tan Dai Nguyen:** Conceptualization, Data curation, Software, Methodology, Formal analysis,
40 Investigation, Visualization, Writing - original draft, Writing - review & editing. **YongQing Fu:**
41 Conceptualization, Funding acquisition, Project administration, Supervision, Writing - review &
42 editing. **Van Thai Tran:** Conceptualization, Investigation, Writing - review & editing. **Sanam**
43 **Pudasaini:** Investigation, Resources. **Archana Gautam:** Investigation, Resources. **Du Hejun:**
44 Conceptualization, Funding acquisition, Resources, Methodology, Project administration,
45 Supervision, Validation.

46 **Competing interests**

1 There is no conflict of interest concerning the work described in this manuscript.

2
3 Appendix A. Supplementary data

4 The MATLAB code is available from the corresponding author upon reasonable request.

5
6 References

- 7 1. Lin, S., D. Chen, and Y. Xie, *Single-Cell Manipulation Technology for Cancer Research*,
8 in *Gastric Cancer Prewarning and Early Diagnosis System*, D. Cui, Editor. 2017, Springer
9 Netherlands: Dordrecht. p. 173-194.
- 10 2. Alam, M.K., et al., *Recent advances in microfluidic technology for manipulation and*
11 *analysis of biological cells (2007–2017)*. *Analytica Chimica Acta*, 2018. **1044**: p. 29-65.
- 12 3. Barani, A., et al., *Microfluidic integrated acoustic waving for manipulation of cells and*
13 *molecules*. *Biosensors and Bioelectronics*, 2016. **85**: p. 714-725.
- 14 4. Pudasaini, S., et al., *Continuous flow microfluidic cell inactivation with use of insulating*
15 *micropillars for multiple electroporation zones*. *ELECTROPHORESIS*. **0**(ja).
- 16 5. Ahmed, D., et al., *Rotational manipulation of single cells and organisms using acoustic*
17 *waves*. *Nature Communications*, 2016. **7**: p. 11085.
- 18 6. Lam, K.H., et al., *Multifunctional single beam acoustic tweezer for non-invasive*
19 *cell/organism manipulation and tissue imaging*. *Scientific Reports*, 2016. **6**: p. 37554.
- 20 7. Heath, J.R., A. Ribas, and P.S. Mischel, *Single-cell analysis tools for drug discovery and*
21 *development*. *Nature Reviews Drug Discovery*, 2015. **15**: p. 204.
- 22 8. Ishii, S., K. Tago, and K. Senoo, *Single-cell analysis and isolation for microbiology and*
23 *biotechnology: methods and applications*. *Applied Microbiology and Biotechnology*, 2010.
24 **86**(5): p. 1281-1292.
- 25 9. Lindström, S. and H. Andersson-Svahn, *Overview of single-cell analyses: microdevices*
26 *and applications*. *Lab on a Chip*, 2010. **10**(24): p. 3363-3372.
- 27 10. Wu, M., et al., *Isolation of exosomes from whole blood by integrating acoustics and*
28 *microfluidics*. *Proceedings of the National Academy of Sciences*, 2017.
- 29 11. Gartner, Z.J. and C.R. Bertozzi, *Programmed assembly of 3-dimensional microtissues with*
30 *defined cellular connectivity*. *Proceedings of the National Academy of Sciences*, 2009.
31 **106**(12): p. 4606-4610.
- 32 12. Whitesides, G.M. and B. Grzybowski, *Self-Assembly at All Scales*. *Science*, 2002.
33 **295**(5564): p. 2418-2421.
- 34 13. Armstrong, P.B., *Cell Sorting Out: The Self-Assembly of Tissues In Vitro*. *Critical Reviews*
35 *in Biochemistry and Molecular Biology*, 1989. **24**(2): p. 119-149.
- 36 14. Kirschner, M. and T. Mitchison, *Beyond self-assembly: From microtubules to*
37 *morphogenesis*. *Cell*, 1986. **45**(3): p. 329-342.
- 38 15. Du, Y., et al., *Directed assembly of cell-laden microgels for fabrication of 3D tissue*
39 *constructs*. *Proceedings of the National Academy of Sciences*, 2008. **105**(28): p. 9522-
40 9527.
- 41 16. Chiou, P.Y., A.T. Ohta, and M.C. Wu, *Massively parallel manipulation of single cells and*
42 *microparticles using optical images*. *Nature*, 2005. **436**(7049): p. 370-372.
- 43 17. Ashkin, A., *Optical trapping and manipulation of neutral particles using lasers*.
44 *Proceedings of the National Academy of Sciences*, 1997. **94**(10): p. 4853-4860.
- 45 18. Guo, F., et al., *Three-dimensional manipulation of single cells using surface acoustic*
46 *waves*. *Proceedings of the National Academy of Sciences*, 2016. **113**(6): p. 1522-1527.

- 1 19. Ding, X., et al., *Surface acoustic wave microfluidics*. Lab on a Chip, 2013. **13**(18): p. 3626-
2 3649.
- 3 20. Lin, S.-C.S., X. Mao, and T.J. Huang, *Surface acoustic wave (SAW) acoustophoresis: now
4 and beyond*. Lab on a Chip, 2012. **12**(16): p. 2766-2770.
- 5 21. Ozcelik, A., et al., *Acoustic tweezers for the life sciences*. Nature Methods, 2018. **15**(12):
6 p. 1021-1028.
- 7 22. Yeo, L.Y. and J.R. Friend, *Surface Acoustic Wave Microfluidics*. Annual Review of Fluid
8 Mechanics, 2014. **46**(1): p. 379-406.
- 9 23. Ding, X., et al., *On-chip manipulation of single microparticles, cells, and organisms using
10 surface acoustic waves*. Proceedings of the National Academy of Sciences, 2012. **109**(28):
11 p. 11105-11109.
- 12 24. Shi, J., et al., *Acoustic tweezers: patterning cells and microparticles using standing surface
13 acoustic waves (SSAW)*. Lab on a Chip, 2009. **9**(20): p. 2890-2895.
- 14 25. Wood, C.D., et al., *Formation and manipulation of two-dimensional arrays of micron-scale
15 particles in microfluidic systems by surface acoustic waves*. Applied Physics Letters, 2009.
16 **94**(5): p. 054101.
- 17 26. Skotis, G.D., et al., *Dynamic acoustic field activated cell separation (DAFACS)*. Lab on a
18 Chip, 2015. **15**(3): p. 802-810.
- 19 27. Antfolk, M., et al., *Focusing of sub-micrometer particles and bacteria enabled by two-
20 dimensional acoustophoresis*. Lab on a Chip, 2014. **14**(15): p. 2791-2799.
- 21 28. Shi, J., et al., *Focusing microparticles in a microfluidic channel with standing surface
22 acoustic waves (SSAW)*. Lab on a Chip, 2008. **8**(2): p. 221-223.
- 23 29. Shi, J., et al., *Three-dimensional continuous particle focusing in a microfluidic channel via
24 standing surface acoustic waves (SSAW)*. Lab on a Chip, 2011. **11**(14): p. 2319-2324.
- 25 30. Zhang, S.P., et al., *Digital acoustofluidics enables contactless and programmable liquid
26 handling*. Nature Communications, 2018. **9**(1): p. 2928.
- 27 31. Jo, M.C. and R. Guldiken, *Particle manipulation by phase-shifting of surface acoustic
28 waves*. Sensors and Actuators A: Physical, 2014. **207**: p. 39-42.
- 29 32. Laurell, T., F. Petersson, and A. Nilsson, *Chip integrated strategies for acoustic separation
30 and manipulation of cells and particles*. Chemical Society Reviews, 2007. **36**(3): p. 492-
31 506.
- 32 33. Shi, J., et al., *Continuous particle separation in a microfluidic channel via standing surface
33 acoustic waves (SSAW)*. Lab on a Chip, 2009. **9**(23): p. 3354-3359.
- 34 34. Ai, Y., C.K. Sanders, and B.L. Marrone, *Separation of Escherichia coli Bacteria from
35 Peripheral Blood Mononuclear Cells Using Standing Surface Acoustic Waves*. Analytical
36 Chemistry, 2013. **85**(19): p. 9126-9134.
- 37 35. Chen, Y., et al., *High-throughput acoustic separation of platelets from whole blood*. Lab
38 on a Chip, 2016. **16**(18): p. 3466-3472.
- 39 36. Destgeer, G., et al., *Continuous separation of particles in a PDMS microfluidic channel via
40 travelling surface acoustic waves (TSAW)*. Lab on a Chip, 2013. **13**(21): p. 4210-4216.
- 41 37. Devendran, C., I. Gralinski, and A. Neild, *Separation of particles using acoustic streaming
42 and radiation forces in an open microfluidic channel*. Microfluidics and Nanofluidics,
43 2014. **17**(5): p. 879-890.
- 44 38. Li, P., et al., *Acoustic separation of circulating tumor cells*. Proceedings of the National
45 Academy of Sciences, 2015. **112**(16): p. 4970-4975.
- 46 39. Nam, J., Y. Lee, and S. Shin, *Size-dependent microparticles separation through standing
47 surface acoustic waves*. Microfluidics and Nanofluidics, 2011. **11**(3): p. 317-326.

- 1 40. Nam, J., et al., *Density-dependent separation of encapsulated cells in a microfluidic*
2 *channel by using a standing surface acoustic wave*. *Biomicrofluidics*, 2012. **6**(2): p.
3 024120.
- 4 41. Nam, J., et al., *Separation of platelets from whole blood using standing surface acoustic*
5 *waves in a microchannel*. *Lab on a Chip*, 2011. **11**(19): p. 3361-3364.
- 6 42. Tsutsui, H. and C.-M. Ho, *Cell separation by non-inertial force fields in microfluidic*
7 *systems*. *Mechanics Research Communications*, 2009. **36**(1): p. 92-103.
- 8 43. Chen, Y., et al., *Rare cell isolation and analysis in microfluidics*. *Lab on a Chip*, 2014.
9 **14**(4): p. 626-645.
- 10 44. Ding, X., et al., *Tunable patterning of microparticles and cells using standing surface*
11 *acoustic waves*. *Lab on a Chip*, 2012. **12**(14): p. 2491-2497.
- 12 45. Nguyen, T.D., et al., *Patterning and manipulating microparticles into a three-dimensional*
13 *matrix using standing surface acoustic waves*. *Applied Physics Letters*, 2018. **112**(21): p.
14 213507.
- 15 46. Zheng, T., et al., *Patterning microparticles into a two-dimensional pattern using one*
16 *column standing surface acoustic waves*. *Sensors and Actuators A: Physical*, 2018. **284**: p.
17 168-171.
- 18 47. Whitesides, G.M., *The origins and the future of microfluidics*. *Nature*, 2006. **442**(7101): p.
19 368-373.
- 20 48. Bernard, I., et al., *Controlled rotation and translation of spherical particles or living cells*
21 *by Surface Acoustic Waves*. *Lab on a Chip*, 2017.
- 22 49. Kang, B., et al., *High-resolution acoustophoretic 3D cell patterning to construct functional*
23 *collateral cylindroids for ischemia therapy*. *Nature Communications*, 2018. **9**(1): p. 5402.
- 24 50. Ding, X., et al., *Standing surface acoustic wave (SSAW) based multichannel cell sorting*.
25 *Lab on a Chip*, 2012. **12**(21): p. 4228-4231.
- 26 51. Lata, J.P., et al., *Surface Acoustic Waves Grant Superior Spatial Control of Cells Embedded*
27 *in Hydrogel Fibers*. *Advanced Materials*, 2016. **28**(39): p. 8632-8638.
- 28 52. Shahid, M.N., et al., *Surface acoustic waves induced micropatterning of cells in gelatin*
29 *methacryloyl (GelMA) hydrogels*. *Biofabrication*, 2017. **9**(1): p. 015020.
- 30 53. Collins, D.J., et al., *Two-dimensional single-cell patterning with one cell per well driven by*
31 *surface acoustic waves*. *Nature Communications*, 2015. **6**: p. 8686.
- 32 54. Tao, X., et al., *3D patterning/manipulating microparticles and yeast cells using ZnO/Si thin*
33 *film surface acoustic waves*. *Sensors and Actuators B: Chemical*, 2019. **299**: p. 126991.
- 34 55. Crawford, D.F., C.A. Smith, and G. Whyte, *Image-based closed-loop feedback for highly*
35 *mono-dispersed microdroplet production*. *Scientific Reports*, 2017. **7**(1): p. 10545.
- 36 56. Miller, E., M. Rotea, and J.P. Rothstein, *Microfluidic device incorporating closed loop*
37 *feedback control for uniform and tunable production of micro-droplets*. *Lab on a Chip*,
38 2010. **10**(10): p. 1293-1301.
- 39 57. Rimsa, R., et al., *A planar surface acoustic wave micropump for closed-loop microfluidics*.
40 *Applied Physics Letters*, 2017. **111**(23): p. 234102.
- 41 58. Soenksen, L.R., et al., *Closed-loop feedback control for microfluidic systems through*
42 *automated capacitive fluid height sensing*. *Lab on a Chip*, 2018. **18**(6): p. 902-914.
- 43 59. Wang, Z. and J. Zhe, *Recent advances in particle and droplet manipulation for lab-on-a-*
44 *chip devices based on surface acoustic waves*. *Lab on a Chip*, 2011. **11**(7): p. 1280-1285.
- 45 60. Sadhal, S.S., *Acoustofluidics 13: Analysis of acoustic streaming by perturbation methods*.
46 *Lab on a Chip*, 2012. **12**(13): p. 2292-2300.

- 1 61. Gor'kov, L.P., *On the Forces Acting on a Small Particle in an Acoustical Field in an Ideal*
2 *Fluid*. Soviet Physics Doklady, 1962. **6**: p. 773-775.
- 3 62. Nama, N., et al., *Numerical study of acoustophoretic motion of particles in a PDMS*
4 *microchannel driven by surface acoustic waves*. Lab on a Chip, 2015. **15**(12): p. 2700-2709.
- 5 63. Hill, M. and N.R. Harris, *Ultrasonic Particle Manipulation*, in *Microfluidic Technologies*
6 *for Miniaturized Analysis Systems*, S. Hardt and F. Schönfeld, Editors. 2007, Springer US:
7 Boston, MA. p. 357-392.
- 8 64. Meng, L., et al., *Precise and programmable manipulation of microbubbles by two-*
9 *dimensional standing surface acoustic waves*. Applied Physics Letters, 2012. **100**(17): p.
10 173701.
- 11 65. Kumar, A.K.S., et al., *High-frequency surface acoustic wave device based on thin-film*
12 *piezoelectric interdigital transducers*. Applied Physics Letters, 2004. **85**(10): p. 1757-1759.
- 13 66. Guo, F., et al., *Controlling cell-cell interactions using surface acoustic waves*. Proceedings
14 of the National Academy of Sciences, 2015. **112**(1): p. 43-48.
- 15 67. Zheng, T., et al., *The role of electric field in microfluidic heating induced by standing*
16 *surface acoustic waves*. Applied Physics Letters, 2018. **112**(23): p. 233702.
- 17 68. Bruus, H., *Acoustofluidics 7: The acoustic radiation force on small particles*. Lab on a
18 Chip, 2012. **12**(6): p. 1014-1021.
- 19 69. Wang, K., et al., *Sorting of tumour cells in a microfluidic device by multi-stage surface*
20 *acoustic waves*. Sensors and Actuators B: Chemical, 2018. **258**: p. 1174-1183.
- 21 70. Gomi, H. and M. Kawato, *Neural network control for a closed-loop System using*
22 *Feedback-error-learning*. Neural Networks, 1993. **6**(7): p. 933-946.
- 23 71. Rogers, P.R., J.R. Friend, and L.Y. Yeo, *Exploitation of surface acoustic waves to drive*
24 *size-dependent microparticle concentration within a droplet*. Lab on a Chip, 2010. **10**(21):
25 p. 2979-2985.
- 26 72. Li, S., et al., *Standing Surface Acoustic Wave Based Cell Coculture*. Analytical Chemistry,
27 2014. **86**(19): p. 9853-9859.
- 28 73. Meng, L., et al., *Transportation of single cell and microbubbles by phase-shift introduced*
29 *to standing leaky surface acoustic waves*. Biomicrofluidics, 2011. **5**(4): p. 044104.
30