

We are IntechOpen, the world's leading publisher of Open Access books Built by scientists, for scientists

4,800

Open access books available

122,000

International authors and editors

135M

Downloads

Our authors are among the

154

Countries delivered to

TOP 1%

most cited scientists

12.2%

Contributors from top 500 universities



WEB OF SCIENCE™

Selection of our books indexed in the Book Citation Index
in Web of Science™ Core Collection (BKCI)

Interested in publishing with us?
Contact book.department@intechopen.com

Numbers displayed above are based on latest data collected.

For more information visit www.intechopen.com



Hardware-Software Partitioning of Digital Signal Processing in Ultrasound Medical Devices a Case Study

Marcin Lewandowski

*Institute of Fundamental Technological Research, Polish Academy of Sciences
Pawinskiego 5B, 02-106 Warsaw,
Poland*

1. Introduction

The development of ultrasound devices and diagnostic methods is closely related with the development of microelectronics and digital signal processing. In most of the state of the art electronic devices, the signal is digitally processed. These devices may be generally categorized based on the number of parallel processing channels. Imaging devices with multi-element linear or phased arrays usually have 16-256 transmission/reception channels. Single channel processing is usually performed in imaging devices with mechanically moved single element sector heads, the so-called “wobbler” and in dedicated Doppler devices. The methods of signal processing are much the same in all device categories.

The paper presents a general overview of ultrasound signal processing and its digital implementation with emphasis on hardware-software partitioning. The available state of the art methods and systems of digital signal processing using both hardware and software are presented as well as the issues pertaining to algorithm implementation methodology. The state of the art system solutions are presented based on the descriptions of representative ultrasound devices, found in literature. The similarities between the presented devices and radio signal processing systems used in telecommunication are also discussed in the paper. Based on device description, the authors present the architecture of processing and communication as well as specific design solutions. The discussed issues and system solutions are analyzed based on two ultrasound medical devices, namely:

- *uScan* - high frequency ultrasonograph with coded transmission (Lewandowski & Nowicki, 2008),
- *digiTDS* - transcranial Doppler system (Lewandowski et al., 2009).

Both devices have been designed and built by the author and co-workers at the Institute of Fundamental Technological Research, Polish Academy of Sciences. They have been designed for commercial purposes, in conformity with medical standards and economic limitations.

The analysis of the presented solution comprises:

- description of the presented systems architecture and of the processing algorithm implementation,
- decisions concerning the design and hardware-software partitioning.

2. Digital signal processing of the ultrasound echoes

Diagnostic medical ultrasound devices utilize ultrasound waves at the frequency of 1-15 MHz. The transmitted ultrasound wave, mainly in the form of a train of short pulses, propagates in the tissue where the structures are reflected and returned to the head as echoes. The signal of echoes is initially amplified and filtered in an analogue chain and next, digitally processed using ADC (Analog-Digital Converter) with 8-14 bit of resolution (Thomenius, 2006). The received high frequency signal (called RF - radio frequency) of the echoes is amplitude- and phase modulated carrier frequency signal. The signal is demodulated in the device to obtain baseband frequency. The demodulated echo signal is further processed, depending on the application (Ali et al., 2008).




Fig. 1. Simplified block diagram of an ultrasound diagnostic device

The chain of ultrasound signal processing (Fig. 1) is much the same as the chain used in digital telecommunication. Therefore, in both cases, similar system solutions and processing methods are applied. For adaptation to quickly changing processing algorithms used in telecommunication, the Software-Defined Radio (SDR) was developed, in which RF signals are processed using software (Reed, 2002).

The generic architecture of SDR systems composed of GPP (General Purpose Processor), DSP (Digital Signal Processor), FPGA (Field Programmable Gate Array) and high frequency front-end blocks was proposed by (Bassam et al., 2009). According to the authors, universal systems of RF signal processing can be implemented based on this architecture.

The most popular example of an SDR platform is the GNU Radio project, and the compatible Universal Software Radio Peripheral (USRP) module, designed by a group of enthusiasts, later used in multiple designs and studies. The USRP module consists of a set of broadband ADC and DAC connected to the FPGA, which in turn communicates with the PC via USB interface. The board is prepared to host analogue transmitter and receiver modules, which tailor the solution to specific applications. The FPGA is responsible for signal modulation and demodulation and stream data to and from PC. The developed digital signal processing in C language with interface to Python script language enables quick and easy prototyping of radio signal processing algorithms. With its universal design and availability of a relatively cheap hardware module, the system is quite popular. The project website (<http://gnuradio.org>) contains information on the application of the project, as well as protocols and telecommunication standards including WiFi, RFID, DVB, DAB, GSM and DECT.

The newer and more advanced version of SDR platform named SORA was developed by Microsoft® Research. It has a significantly larger FPGA (Xilinx Virtex 5) chip and a faster

PCI-e x8 interface. Moreover, a software stack for multicore processors containing drivers and processing libraries, optimized for efficiency and latency minimization was developed (Tan et al., 2009). The authors also presented a demonstrative implementation of SoftWiFi supporting the 802.11a/b/g communication protocol.

The reader will probably notice a very high similarity between the USRP solution and coder-digitizer module of the *uScan* system described later in the paper. It definitely indicates a similarity between the system solutions and the architecture of SDR and ultrasound devices. Interestingly, the author has not found any report on using SDR solutions for ultrasound RF signal processing.

3. Digital signal processing systems

Nowadays, there is a vast number and diversity of digital signal processing systems and methods. Therefore, it is extremely difficult to select an optimal system solution for signal processing of suitable processing power, data throughput and power consumption at a reasonable price. If we limit our choice to the solutions commonly applied in ultrasonography, these will include:

- hardware processing - programmable logic systems (FPGA) and ASIC (Application Specific Integrated Circuits),
- software processing - general purpose processors (GPP), digital signal processors (DSP) and graphic processors (GPU).

Recently there is a tendency to increase software processing because of the development of processor systems and significantly easier implementation process and algorithm debugging as compared to hardware implementations. This tendency has resulted in the development of ultrasound medical devices, requiring a more efficient digital processing and working in a real time regime.

The main advantage of software solutions is easy code modification and simpler testing and debugging methods. Implementations in high level languages (e.g. C language) can be easily simulated and verified on different platforms before being implemented on the target platform. Nowadays, the majority of DSP are programmed in C language and have more advanced tools for development. Thanks to their computing power, easy high-level code development, DSP are the core of processing chain in modern devices. There is an alternative trend in software processing, namely general purpose processors (GPP). Modern PC processors have high computing power and special parallel executive units for multimedia processing. These units are in fact separate vector processors, optimized for digital signal processing. Adequate use of parallel processing and multimedia units can replace several DSP working at the same time. As PCs are frequently part of these devices and are used both as controller and user interface, the implementation of processing algorithms using PC seems natural. Real time application of operational systems, such as Microsoft® Windows is problematic as these are not real time systems and therefore they do not guarantee adequate time precision and the required operations being performed in the required time period.

Recently, graphic processor units (GPU) are used for signal processing; they are traditionally used for 3D graphics rendering in PCs. The quick development of computer graphics resulted in the development of graphic processors with new capabilities. From specialized 3D processors, GPU have been developed into versatile programmable vector processors. Thanks to its massive parallel internal architecture and advanced memory

interfaces, the floating point processing power of GPUs can be even 100 times higher than the power of the fastest PC processors. The new tools for graphic processor programming allow to utilize new computing power, not only for graphic purposes, but also in other, more general applications (Owens et al., 2005).

Hardware implementation of digital signal processing is presently dominated by programmable logic systems (FPGA). Thanks to the availability of cheap FPGA versions, they are used not only in small scale production, but also in products released onto the consumer market. The possibility of reprogramming within the system (also during operation), a very wide range of system density (from thousands to millions of gates) and increasingly advanced design tools (McDonald, 2009) contribute to the growing popularity of FPGA systems. Hardware methods offer the highest potential processing speed (Sirowy & Forin, 2008), however the implementation costs are also high, probably the highest. Therefore, hardware and software methods are frequently applied as complementary methods to achieve the required efficiency and versatility at minimal costs.

The discussion presented above indicates that digital processing algorithms can be implemented in different ways, based on different solutions. The optimal use of various computational tools and system architecture providing development opportunities is a complex problem from the point of view of science and engineering.

4. Ultrasound devices - state of the art

The combined architecture of hardware-software processing devices is currently the main feature of various digital signal processing devices, requiring high processing speed and characterized by a high bandwidth of input data stream. The combined architecture incorporates hardware systems (FPGA/ASIC), responsible for initial processing and aggregation and/or distribution and further software processing systems (DSP, GPP, GPU processors). We can see a clear tendency to extend the scope of software processing and minimize hardware solutions. The popularity of software solutions results not only from easier implementation, testing and debugging, but also from the possibility of adaptation of these devices to quickly changing market requirements and new algorithms.

On the other hand, the state of the art FPGA systems should be now more frequently treated as programmable SOC (System on Chip) systems rather than simple gate arrays. The availability of new tools for algorithm implementation in FPGA (e.g. Xilinx® AccelDSP, Altera® DSP Builder) significantly facilitates the design and transfer of complex computational algorithms without the need to encode them in hardware description languages. The available IP (Intellectual Property) processor blocks (e.g. Xilinx® Microblaze, Altera® NIOS, Actel® Cortex-M) enable the creation of single- and multiprocessor systems directly in FPGA structures. The processor blocks support the implementation of extended instruction sets using hardware solutions built from the surrounding logic (the so called instruction accelerators). Altera® C2H tool is an example of such solution that compiles selected code fragments in C language directly for hardware implementation, warranting 15-73-fold speedup for selected functions (Altera, 2006). Integrated systems based on the reprogrammable FPGA were described by (Garcia et al., 2006).

The model example of combined ultrasound system architecture is a Doppler device developed by (Ricci et al., 2008). The construction of this system is much the same as that of *digiTDS* system presented below. The module designed by Ricci et al. is composed of FPGA (Altera® Stratix) executing hardware processing functions, and DSP (Texas Instruments®

TMS320C67). The module communicates with a PC and is responsible for the control and display functions through the USB interface. The FPGA has implemented a quadrature demodulator, a set of digital filters and a decimator. The initially demodulated and decimated data stream is transferred to the DSP, executing standard Doppler algorithms. Although the reported system is not a commercial solution, its construction is similar to that of the devices offered on the market.

Another device, an ultrasonograph platform ULA-OP, was designed by the same group (Tortoli et al., 2009). Although it is a 64-channel system, it is based on the same general concept. The basic difference is that this system utilizes multiple FPGA chips connected to multichannel ADC and a central hub FPGA, responsible for data aggregation (in total: 5 FPGA Altera® Stratix systems, each having its own 512 MB of DDR memory). The efficient DSP (Texas Instruments® TMS320C6455) is responsible for software processing. The system is connected to PC via USB port and the PC controls the system and displays the results. As we can see, although it has many channels, the general scheme of device design does not differ from that of a dual channel Doppler system.

The next system described in literature is a hand-held 32-channel ultrasonograph, developed by (Lee et al., 2009). Its architecture resembles the ULA-OP system. The main differences are that the ultrasonograph uses a single large FPGA (Xilinx® Virtex 4) and an additional application processor (Intel® XScale PXA270), executing the user interface and display on the integrated LCD panel. The system is completely independent (does not require connection to PC) and optimized for the size and power consumption. The distribution of processing tasks between the hardware and software is in conformity with previous solutions.

1024-channel SARUS (Jensen et al., 2007) system is a unique ultrasound research platform. Unlike the previously described solutions, the system is mainly based on the on-line hardware and off-line software processing. It consists of 64 identical DAUP (Digital Acquisition and Ultrasound Processing) boards, each operating 16 transmitting-receiving channels and an external computer cluster with a disk array. Each of the DAUP boards contains 4 FPGA (Xilinx® Virtex XC4VFX100) with a dedicated 1-4 GB DDR2 memory and one control FPGA (Xilinx® Virtex XC4VFX12). Technical solutions and the potential of SARUS system are overwhelming, however, this is only a research system and its architecture does not allow it to be directly adapted to commercial systems.

The hardware only processing chain for ultrasound system was presented by (Chang et al., 2009). A single-channel acquisition and processing system consists of ADC (Analogue Devices® AD9430) connected to FPGA (Altera® Stratix EP1S60F1020C6), executing the entire processing. Due to the planned application of the above system in studies on mice whose heart rate reaches 13 beats per second, the authors made an attempt to perform imaging with the speed of 400 images per second. The application of simplified and optimized processing algorithms (including envelope detector) allowed to obtain the designed processing speed. The reconstructed images were next transmitted to PC via PCI interface and displayed using a graphics card with the implemented scan-converter. The presented solution indicates how high processing speed can be obtained in systems using hardware processing.

A completely different solution is a C-scan imaging system presented by (Fuller et al., 2009). It has a custom built ultrasonic array transducer, composed of 3600 elements with an integrated transmission-reception electronics. The echoes received from selected depth are then transmitted to a DSP (Analog Devices® Blackfin BF561) which executes the entire

digital signal processing and display on the built-in LCD monitor screen. The obtained display speed of 43 images per second for the C-scan indicates that in some applications, pure software processing is sufficient to obtain a practical outcome.

Interestingly, ready-to-use solutions for commercial ultrasound scanners are offered by Texas Instruments® company, the leading manufacturer of DSP. Texas Instruments® has developed an optimized function library for applications in ultrasonography using the offered TMS320C64 series DSPs (Thomas, 2010). Moreover, (Pailoor & Pradhan, 2008) present specific architectures and solutions for portable medical ultrasonography systems. These are systems with combined hardware-software processing. As in the previous presented systems, the first processing stage (beamforming and demodulation) is executed by hardware, while the subsequent stages are executed by DSP. The TI offers a complete development environment and integrated circuits portfolio to build almost the complete systems (TI doesn't offer FPGA).

The final solution presented in this paper is a cheap ultrasonograph with a single processing channel in the form of an integrated head connected to PC via USB port (Richard et al., 2008). The entire simplified electronics is integrated on a miniature board, placed directly in the ultrasound head. The RF echo signal undergoes analogue demodulation and it is digitally processed and transmitted to PC via USB interface. An analogue demodulation of RF signals is no longer used in the devices; in the presented case its application results from construction simplification. Due to low power consumption, the head is supplied directly from PC via USB. The whole processing and visualization were implemented as software for PC. The quality of an ultrasound image fails to meet contemporary standards, however, on the other hand, its price is unmatched. The concept of system solution, minimization of electronics and total PC processing is similar to that of the *uScan* described further in the paper.

5. Case study

Both systems are based on a mixed model of hardware-software signal processing. The systems of RF echo signal acquisition are based on a fast A/D converter and FPGA, being an interface to the further part of the system. The application of FPGA at the beginning of the digital chains is now a standard solution. Apart from signal acquisition, the FPGA most often executes hardware processing of a digital signal, whose goal is to reduce the data stream for further processing. Next, the processing is executed by DSP, GPP, GPU or combinations of these processors.

5.1 High frequency ultrasound scanner - *uScan*

High frequency ultrasonography is an imaging method using ultrasound waves of more than 20 MHz frequency. Due to the obtained high resolution and low penetration depth of imaging, this method is applied in dermatology, ophthalmology, cosmetology and for examining small animals.

uScan system is a high frequency ultrasound scanner (20-40 MHz) with the implemented coded transmission/excitation function. The coded transmission method involves transmission of long series of impulses and subsequent time compression of the received echoes to restore resolution. With the long bursts, it is possible to obtain a better signal-to-noise ratio, and the resulting image contrast improvement. Unfortunately, the echo time

compression algorithm, which is in fact, matched filter to the transmitted signal, requires a very high computing power.

The device is based on the concept of absolute minimization of electronics and computation based solely on CPU and GPU. The electronic module of coder-digitizer (Fig. 2) executing all the transmission and reception functions is 90 mm × 90 mm in size and is based on a single low-cost FPGA chip (Xilinx® Spartan 3 XC3S200). The transmission involves the use of an arbitrary waveform generator with a 14-bit DAC (Analogue Devices® AD9744) operating at 200 MSPS speed. The acquisition is performed using a 12-bit DAC (Maxim® MAX1214) also working at a speed of 200 MSPS. The digital RF echo signal is transmitted to PC via USB 2.0 interface (Cypress Semiconductor® CY7C68013A). The processing software executes the function of time compression on a CPU processor. The sector geometry conversion function (scan-converter) required to obtain a correct display of an image generated by the sector head is executed by a graphics processor.




Fig. 2. Block diagram of the coder-digitizer module (left), and the board photo of size 90 mm x 90 mm (right)

The module was designed for stream data transmission; therefore it contains only a buffer in FPGA for a single RF signal line. The line data have to leave the buffer before the next acquisition, i.e. they have to be sent to the USB controller system. The USB system contains 4 buffers 512 bytes each, which are used for communication in the ping-pong mode. A typical length of a RF line is 2048 samples, which gives 4096 bytes of data (two bytes per sample). On an ACER TravelMate 4650 notebook (Intel® Pentium M processor with a clock 1.7 GHz, 512 MB of RAM), the USB transmission throughput obtained in the test mode was close to 40 MB/s. However, in real conditions, when signal line acquisition was released periodically by ultrasound head movements, buffer overflow errors were observed. This indicates minor transmission interruption, i.e. changes in instantaneous data reception speed by PC. It was concluded that the “interruptions” occurred at system driver level. Microsoft® Windows does not guarantee the time limit of event and interruptions handling from the input/output devices. The overflow errors result in image loss and entail the necessity of subsequent image acquisition synchronization with the extreme position of the head movement. As the coder-digitizer module featured no larger data buffer, it was necessary to limit the head movement speed (and thus image display) to 10 images per

second, which is still a sufficient speed for still structure imaging. An alternative solution would be to number the transmitted lines and an additional image filling algorithm with the obtained complete lines. Incomplete or deleted lines of the image would be filled by the fragments of previous lines or by interline interpolation.

The digital RF signal processing functions were implemented as a whole in the software domain and distributed between the CPU of the PC and graphic card processor depending on their specificity (Fig. 3). The code compression and detection algorithm is executed on CPU, while the scan-converter and additional post-processing functions are executed on GPU. Such architecture enables execution of more demanding processing systems without the need to employ specialized and expensive multiprocessor systems.




Fig. 3. Data flow of the software RF digital signal processing



Fig. 4. Photo of the complete *uScan* system

Software efficiency tests for processing and visualization were performed on *ACER TravelMate 4650* notebook in configuration: Intel® Pentium M processor with 1.7 GHz clock, 2 MB second level cache, 512 MB RAM, graphic processor Nvidia GeForce Go 6600 with

64 MB of video memory. For an ultrasound head scanning with the speed of 7 images/s, the CPU load was 37%. When extrapolating the processing time, the system was found to be able to run at the speed up to 25 images/s in this configuration.

The obtained solution meets the presented requirements for processing speed and image display in a real time regime (up to 10 frames per second). The applied balanced architecture of digital processing with task distribution between CPU and GPU ensures very high processing efficiency with simultaneous limitation of electronic systems and hardware processing. The limitation of this system is the speed of data transfer via USB interface, as the computing power would allow for significantly faster display. The quick development and increase of CPU/GPU system efficiency ensures the possibility of increasingly complex processing algorithms without the need of hardware replacement.

5.2 Transcranial Doppler System - *digitDS*

The pulse wave Doppler method is applied as a standard for intracranial diagnosing allowing to measure the flow at selected depth. The state of the art diagnostic solutions are mainly focused on complex assessment and monitoring of cerebrovascular flow in different pathological conditions, especially those potentially affecting normal function of the central nervous system.

The *digitDS* device is a multigate transcranial Doppler system. This system is composed of two electronic modules of 130 mm × 82 mm dimensions (Fig. 5), responsible for high frequency signal transmission, acquisition and demodulation, and the PC responsible for Doppler signal processing after demodulation and data presentation. Fig. 6 presents the




Fig. 5. Photos of the *digitDS*'s electronic modules: digital module (left), mixed-signal module (right)




Fig. 6. Simplified block diagram of the *digiTDS* system

system block scheme. On an analogue-digital module board, there are two independent transmission-reception channels (the device supports two heads operation, the so called bilateral operation). A 10-bit DAC (Texas Instruments® DAC5652A) operating at a speed of 64 MSPS is used to generate the transmitted signal. The reception track contains amplifiers with the function of time gain compensation (Analogue Devices® AD8331) and a dual 14-bit ADC (Analogue Devices® AD9640) with sampling frequency of 64 MHz. The digital signals from/to converters are connected to FPGA on the digital module board via intermodule connector. The digital module consists of FPGA (Altera® Cyclone 3 EP3C25), DSP (Analog Devices® Blackfin BF537) as well as USB device controller (PLX® NET2272) and an Ethernet interface.

Unlike the *uScan*, *digiTDS* may run independently of PC, thanks to its own control resources - DSP processor. The system communicates with PC via USB 2.0 interface. The digital processing chain consists of hardware processing in the FPGA and software processing using DSP and PC's CPU. The application of hardware resources during early processing stage allowed to significantly decrease the required data throughput and computing load of PC. It also allowed to use a low-power single-board PC with Intel® ATOM processor and to extend the operating time on batteries. The distribution and implementation of processing tasks were performed on a constructed device prototype. Especially moving the realization of filtration in multiple gates from the software to FPGA enabled significant processor time saving.

Fig. 7 presents the Doppler digital signal processing track on *digiTDS* device with task distribution between the FPGA system and software.




Fig. 7. Doppler digital signal processing chain




Fig. 8. Block diagram of the digital quadrature demodulator with CIC filter implemented in the FPGA

The received RF echo signal, after the amplification and digital conversion, is sent to the quadrature demodulator (DDC - Digital Down Converter) and CIC (Cascaded Integrator Comb) filter, implemented in the FPGA (Fig. 8). The demodulator is composed of two multipliers, memory containing reference signal samples with the frequency of 1-16 MHz and two low-pass filters. The samples of high frequency echo signal are mixed with 14-bit

samples of reference sine and cosine signals. As a result, two signals are formed, namely in-phase (I) and quadrature (Q) signals, which are next subjected to low-pass filtration and decimation so that only baseband frequency components were retained in the signal. The low pass filters and the decimator were implemented in the FPGA as a CIC filter (Harris, 2004). The 4th order CIC filter enables regulation of the Doppler gate size and the degree of filtration and decimation.

The subsequent processing component of the FPGA system is a set of constant echo filters. In a multigate system, the signal in each gate has to be filtered in order to eliminate slow changing signals from still or slowly moving structures. In the FPGA, 200 parallel FIR filters are implemented, each with the length of 64 coefficients; they provide filtration in real time in 100 Doppler gates (independently I & Q signals are filtered in each gate). The cumulative utilization of logical resources of the applied FPGA (Altera® EP3C25) amounted to 5% of logic cells, 39% of internal block RAM and 3% of dedicated 18-bit multipliers. Time closure was obtained for the frequency of 64 MHz, and thus the processing speed matched the speed of input data stream (RF signal sampling). It is of note that the software execution of such a task on Intel® ATOM processor would be impractical.

The processed signal is next sent to PC where it is processed specifically for a given application. Recently, standard Doppler processing methods have been implemented, namely Color and Spectrum. Digital signal processing software algorithms have been implemented using Intel® IPP libraries (Taylor, 2004), which use SSE (Streaming SIMD Extensions) vector extensions of x86 family processors. SSE extensions enable execution of floating point operations on multiple data at the same time, which ensures a very high efficiency of multimedia applications (Martinez-Nieto et al., 2009).




Fig. 9. PC software stack of the *digiTDS* system

The control and user software (Fig. 9) was implemented in Microsoft® Windows Embedded Standard 2009 environment, based on NET 3.5 framework. For communication with the digital module via USB interface, Microsoft® WinUSB drivers were applied. User interface was created based on on .NET WPF (Windows Presentation Foundation). Obtaining short delay of Doppler signal transmission, processing and visualization required optimal data buffering, processing optimization and multithreaded implementation.



Fig. 10. Photo of the complete *digiTDS* system

6. Conclusions

The paper presents technical issues pertaining to the architecture of digital signal processing and task distribution between software and hardware in ultrasound diagnostic devices. The solutions reported in literature were reviewed and the similarities between the presented solutions and SDR systems used in digital telecommunication were discussed.

The two presented examples of combined hardware-software processing architecture in the ultrasound devices represent current state of the art.

The experience gained during designing and implementation of digital processing algorithms on these platforms show their unquestionable advantages, namely:

- possibility of algorithm migration between hardware and software solutions,
- possibility of equalization of load distribution in all system elements and data transfer between sub-systems,
- possibility of optimization of the selected system parameter (time delay, data transfer, power consumption, etc.) by changing the method/place of processing execution.

It is also of note that the problems can be encountered during system development requiring advanced real time processing, namely:

- difficulty in defining the required efficiency and throughput of each system component at the design stage,

- problems with operating the system in real time with minimal time delay of result presentation (image, audio),
- time synchronization of hardware and software subsystems (particularly those controlled by not real-time operating system),
- complicated verification & validation of the whole system, as required for medical device.

The future development of ultrasound medical systems will be interrelated with the development of electronics and digital signal processing systems. The present trend of extending the scope of software processing will be probably continued with the next generation of multicore GPP, DSP and GPU. Rapid development of portable devices can be expected, due to the availability of increasingly fast, integrated and low-power System on Chip integrated circuits.

7. Acknowledgements

Acknowledgements to my co-workers from the Institute of Fundamental Technological Research, Polish Academy of Sciences who cooperated with me in the project development: Andrzej Nowicki, Bogusław Zienkiewicz, Mateusz Walczak, Paweł Karłowicz, Ryszard Tymkiewicz, Piotr Karwat, and Beata Witek.

Funding: The European Regional Development Fund under the Innovative Economy Operational Programme & Ministry of Science and Higher Education, Poland.

8. References


- Ali M., Magee D., Dasgupta U. (2008). Signal Processing Overview of Ultrasound Systems for Medical Imaging, Texas Instruments, White Paper SPRAB12
- Altera (2006). *Automated Generation of Hardware Accelerators With Direct Memory Access from ANSI/ISO Standard C Functions*, ALTERA White Paper, WP-AGHRDWR-1.0
- Bassam S. A., Ebrahimi M. M., Kwan A., Helaoui M., Aflaki M. P., Hammi O., Fattouche M., Ghannouchi F. M. (2009). A Generic architecture for smart multi-standard software defined radio systems, SDR'09 Technical Conference and Product Exposition, 1-4 December 2009, Washington, DC, USA
- Chang J. H., Sun L., Yen J.T., Shung K. K. (2009). Low-cost, high-speed back-end processing system for high-frequency ultrasound B-mode imaging, *IEEE Transactions on Ultrasonics, Ferroelectrics, and Frequency Control*, Vol. 56, No. 7
- Fuller M. I., Owen K., Blalock T. N., Hossack J. A., Walker W. F. (2009). Real Time Imaging with the Sonic Window: A Pocket-Sized, C-Scan, Medical Ultrasound Device, *Proceedings of IEEE International Ultrasonics Symposium*, pp. 196-199, Roma, Italy, 20-23 September 2009, IEEE
- Garcia P., Compton K., Schulte M., Blem E., Fu W. (2006). An Overview of Reconfigurable Hardware in Embedded Systems, *EURASIP Journal on Embedded Systems*, Volume 2006, Article ID 56320
- Harris F. J. (2004). *Multirate Signal Processing for Communication Systems*, Prentice Hall PTR, ISBN 978-0131465114

- Jensen J. A., Hansen M., Tomov B. G., Nikolov S. I., Holten-Lund H. (2007). System Architecture of an Experimental Synthetic Aperture Real-Time Ultrasound System, *Proceedings of IEEE International Ultrasonounds Symposium*, pp. 636-640, New York, USA, 28-31 October 2007, IEEE
- Lee H., Sohn H., Yoon Ch., Mo Yoo Y., Song T. (2009). Software-based Hand-Held Ultrasound Color Doppler Imaging System, *Proceedings of IEEE International Ultrasonounds Symposium*, pp. 1844-1847, Roma, Italy, 20-23 September 2009, IEEE
- Lewandowski M., Nowicki A. (2008). High Frequency Coded Imaging System with RF Software Signal Processing, *IEEE Transactions on Ultrasonounds, Ferroelectrics, and Frequency Control*, 55(8):1878-82
- Lewandowski M., Walczak M., Nowicki A. (2009). Compact Modular Doppler System with Digital RF Processing, *Proceedings of IEEE International Ultrasonounds Symposium*, pp. 1848-1851, Roma, Italy, 20-23 September 2009, IEEE
- Martinez-Nieto D., McDonnell M., Carlston P., Reynolds K., Santos V. (2009). Digital Signal Processing on Intel® Architecture, *Intel® Technology Journal*, Vol. 13, No. 1
- McDonald J., (2009). *Insiders' Guide: FPGAs, Tools, and Boards*, <http://www.eg3.com>
- Owens J. D., Luebke D., Govindaraju N., Harris M., Krüger J., Lefohn A.E., Purcell T.J. (2005). A Survey of General-Purpose Computation on Graphics Hardware, *Eurographics 2005 State of the Art Reports*, August 2005, pp. 21-51.
- Reed J. H. (2002). *Software Radio: A Modern Approach to Radio Engineering*, Prentice Hall, ISBN 978-0130811585
- Ricci S., Dallai A., Boni E., Bassi L., Guidi F., Cellai A., Tortoli P. (2008). Embedded System for Real-Time Digital Processing of Medical Ultrasound Doppler Signals, *EURASIP Journal on Advances in Signal Processing*, Volume 2008, Article ID 418235
- Richard W. D., Zar D. M., Solek R. (2008). A low-cost B-mode USB ultrasound probe, *Ultrason Imaging*, Vol. 30, No. 1, pp. 21-8
- Pailoor R, Pradhan D. (2008). Digital Signal Processor (DSP) for Portable Ultrasound, Texas Instruments , Application Report SPRAB18A
- Sirowy S., Forin A. (2008). Where's the Beef? Why FPGAs Are So Fast, Microsoft Research, Technical Report MSR-TR-2008-130
- Tan T., Zhang J., Fang J., Liu H., Ye Y., Wang S., Zhang Y., Wu H., Wang W., Voelker G. M. (2009). Sora: High Performance Software Radio Using General Purpose Multi-core Processors, *Proceedings of the 6th USENIX Symposium on Networked Systems Design and Implementation*, April 22-24, 2009, Boston, MA, USA
- Taylor S. (2004). *Intel Integrated Performance Primitives*, Intel Press, ISBN 978-0971786134
- Thomas D. (2010). Using TI's Embedded Processor Software Toolkit for Medical Imaging (STK-MED), Texas Instruments, Application Report SPRABB8
- Thomenius K. E. (2006). Instrumentation Design For Ultrasound Imaging, In: *Biomedical Engineering And Design Handbook Vol. 2: Applications*, Myer Kutz, (Ed.), 249-256, McGraw-Hill, ISBN 978-0-07-170474-8

Tortoli P., Bassi L., Boni E., Dallai A., Guidi F., Ricci S. (2009). ULA-OP: An Advanced Open Platform for Ultrasound Research. *IEEE Transactions on Ultrasonics, Ferroelectrics, and Frequency Control*, Vol. 56, No. 10

IntechOpen

IntechOpen



Ultrasound Imaging

Edited by Mr Masayuki Tanabe

ISBN 978-953-307-239-5

Hard cover, 210 pages

Publisher InTech

Published online 11, April, 2011

Published in print edition April, 2011

In this book, we present a dozen state of the art developments for ultrasound imaging, for example, hardware implementation, transducer, beamforming, signal processing, measurement of elasticity and diagnosis. The editors would like to thank all the chapter authors, who focused on the publication of this book.

How to reference

In order to correctly reference this scholarly work, feel free to copy and paste the following:

Marcin Lewandowski (2011). Hardware-Software Partitioning of Digital Signal Processing in Ultrasound Medical Devices a Case Study, Ultrasound Imaging, Mr Masayuki Tanabe (Ed.), ISBN: 978-953-307-239-5, InTech, Available from: <http://www.intechopen.com/books/ultrasound-imaging/hardware-software-partitioning-of-digital-signal-processing-in-ultrasound-medical-devices-a-case-stu>

INTECH

open science | open minds

InTech Europe

University Campus STeP Ri
Slavka Krautzeka 83/A
51000 Rijeka, Croatia
Phone: +385 (51) 770 447
Fax: +385 (51) 686 166
www.intechopen.com

InTech China

Unit 405, Office Block, Hotel Equatorial Shanghai
No.65, Yan An Road (West), Shanghai, 200040, China
中国上海市延安西路65号上海国际贵都大饭店办公楼405单元
Phone: +86-21-62489820
Fax: +86-21-62489821

© 2011 The Author(s). Licensee IntechOpen. This chapter is distributed under the terms of the [Creative Commons Attribution-NonCommercial-ShareAlike-3.0 License](#), which permits use, distribution and reproduction for non-commercial purposes, provided the original is properly cited and derivative works building on this content are distributed under the same license.

IntechOpen

IntechOpen