

We are IntechOpen, the world's leading publisher of Open Access books Built by scientists, for scientists

4,800

Open access books available

122,000

International authors and editors

135M

Downloads

Our authors are among the

154

Countries delivered to

TOP 1%

most cited scientists

12.2%

Contributors from top 500 universities



WEB OF SCIENCE™

Selection of our books indexed in the Book Citation Index
in Web of Science™ Core Collection (BKCI)

Interested in publishing with us?
Contact book.department@intechopen.com

Numbers displayed above are based on latest data collected.
For more information visit www.intechopen.com



Multimode Hyperspectral Imaging for Food Quality and Safety

Fartash Vasefi, Nicholas Booth, Hesam Hafizi and Daniel L. Farkas

Additional information is available at the end of the chapter

<http://dx.doi.org/10.5772/intechopen.76358>

Abstract

Food safety and quality are becoming progressively important, and a failure to implement monitoring processes and identify anomalies in composition, production, and distribution can lead to severe financial and customer health damages. If consumers were uncertain about food safety and quality, the impact could be profound; hence, we need better ways of minimizing such risks. On the data management side, the rise of artificial intelligence, data analytics, the Internet of Things, and blockchain all provide enormous opportunities for supply chain management and liability management, but the impact of any approach starts with the quality of the relevant data. Here, we present state-of-the-art spectroscopic technologies including hyperspectral reflectance, fluorescence imaging as well as Raman spectroscopy, and speckle imaging that are all validated for food safety and quality applications. We believe a multimode approach comprising of a number of these synergetic optical detection modes is needed for the highest performance. We present a plan where our implementations reflect this concept through a multimode tabletop system in the sense that a large, real-time production-level device would be based on more modes than this mid-level one, while a handheld, portable unit may only address fewer challenges, but with a lower cost and size.

Keywords: multimode optical imaging, food contamination, hyperspectral imaging, food quality, multimode data management, machine learning

1. Introduction

There is a great need to assess the composition of food, quantitatively and reproducibly, in order to avoid any unintended scenarios, ranging from a product not being quite what it is

stated to be (e.g., lesser quality fish or olive oil) to intentional adulteration (including by terrorist intent) to random contamination (such as by bacteria that can be lethal). These constitute the application domain of, respectively, food quality, food defense, and food safety. Given the place food occupies in society, and the possible extreme implications of any negative events, there is great interest in bringing the best testing to the task of ensuring the quality and safety of our food supply. Unfortunately, some of the currently prevalent methods (molecular/biochemical/biophysical, such as polymerase chain reaction (PCR), chromatography, mass spectrometry, etc.) are intrinsically too slow to yield results in real time, and also rely on random and very sparse sampling. We believe that the power of light as an investigational tool can be brought to the resulting challenge and focus on this possibility here.

Optical imaging is an approach rapidly growing in popularity and applications due to technological advances that have enabled the production of smaller, less expensive, more efficient, and faster light sources and detectors. These new technologies have facilitated the acquisition of more accurate optical image sets, yielding molecular, structural, and physiological information from targeted samples. There are many different optical measurement techniques used by industry and academic researchers alike, with each technology usually focusing on a specific property of light (intensity, polarization, wavelength, coherence, temporal change, etc.). We believe, however, that no single method can provide the comprehensive analysis of food that is required.

When applied to food samples, the accuracy of optical detection techniques can be limited due to factors such as low penetration depth and lack of contrast, especially for low biomarker concentrations. However, using a strategic combination of multiple optical detection technologies in an optical system that thus becomes *multimode*, the chemical and/or biological detection accuracy can be substantially improved. Each individual detection method can provide a specific and complementary (sometimes even synergetic) piece of information regarding the sample being examined. Thus, by combining a number of these methods, the impact of the individual limitations can be minimized, and their combined strengths may be harnessed to deliver highly specific results.

The advantages of multimode optical imaging include greatly reducing the time required for the initial detection and enumeration of contaminants, with minimal sample preparation, nondestructive evaluation, fast acquisition times, and visualization of the spatial distribution of numerous components simultaneously. These advantages are highly useful in detecting contaminants in food for assessing safety and quality, and the use of multiple modes of detection, properly combined, is essential for effectiveness and performance.

We summarize here optical technologies which are useful in food safety and quality applications, highlighting both successes and limitations, thus underscoring the usefulness of the new, multimode approach we propose.

2. Hyperspectral imaging

Hyperspectral imaging (HSI) is a growing platform technology that functions by integrating conventional imaging and spectroscopy to gain spatial and spectral information from

an object [1]. It is capable of capturing reflectance, transmittance, and fluorescence images in the visible and infrared regions with submillimeter spatial resolution [2] and high spectral resolution (10 nm). While HSI was originally developed for remote sensing [3], it has gained popularity in the field of food safety and analysis with new applications reported in fruits and vegetables [4–20, 34, 37, 42], poultry [21–25], and meat [26–28]. Some advantages HSI has in comparison with other techniques such as RGB imaging, NIR spectroscopy, and multicolor imaging include being able to produce spatial and spectral information, multiconstituent information, and sensitivity to minor components [1].

HSI in the near infrared (NIR) can provide chemical composition of red meat such as prediction of fat, protein, and water content of lamb meat [32]. Moreover, this method enables the detection of certain bacteria in food, such as *E. coli* [33]. Fungal growth on food products is of particular concern due to the potential for detrimental effects on population health ranging from allergic reactions and respiratory problems to the production of mycotoxins. HSI has been deployed to identify fungal species such as *Aspergillus flavus*, *Aspergillus parasiticus*, *Aspergillus niger*, and *Fusarium* spp. which can produce mycotoxins, which are secondary metabolites that are toxic for humans and animals [36, 37].

A common source of contamination for fresh products and other raw materials used to produce food is fecal contamination; hence it would be highly desirable to develop an automatic inspection system for use in the field and on processing lines. Multispectral detection of fecal contamination on apples using HSI imaging was demonstrated by Kim et al. [45]. A HSI system with a range of 450–851 nm was used to examine reflectance images of experimentally contaminated apples. Fecal contamination sites were evaluated using principal component analysis (PCA) with the goal of identifying two to four wavelengths that could be used in an online multispectral imaging system. As shown in **Figure 1**, their results showed that contamination could be identified using either of three wavelengths in the green, red, and NIR regions.

With the use of HSI in the spectral range of 400–1000 nm, *E. coli* loads in grass carp fish have been measured to evaluate microbial spoilage. In 2015, the researchers demonstrated that reflectance HSI in combination with multivariate analysis had the ability to rapidly and noninvasively quantify and visualize the *E. coli* loads in grass carp fish flesh during the spoilage process [35]. Distribution maps, shown in **Figure 2**, were created to allow for visualization of *E. coli* contamination. These distribution maps were vital in that they provided more detailed information of postmortem spoilage development in grass carp flesh. One of the main advantages that HSI has over conventional spectroscopy methods is its ability to visualize distribution maps of the contamination in a pixel-wise manner. By multiplying the regression coefficients of the multiple linear regression model by the spectrum of each pixel in the image, a prediction map was generated for showing the distribution of *E. coli* within the fish flesh. The different *E. coli* loads were represented by different colors from blue to red. As *E. coli* load increased, the color of the images shifted from blue to red, reflecting the growth of bacteria.

In 2013, Feng et al. [36] presented HSI as a nondestructive tool for direct, quantitative determination of Enterobacteriaceae loads on chicken fillets. The authors developed partial least




Figure 1. First and second principal component images obtained using 748–851 nm region of the hyperspectral reflectance image data for (A) fuji, (B) gala, (C) golden delicious, and (D) red delicious apples [45].




Figure 2. These are examples of distribution maps of *E. coli* loads in fish filets. The distribution maps showed how the level of *E. coli* contamination varied from one sample to the next. A shift in color intensity is seen from blue to red, reflecting the increase in *E. coli* contamination [35].

squares regression (PLSR) models and root mean squared errors. After a simplified model was developed, the PLSR model, it was used for predicting Enterobacteriaceae loads in every pixel of the image acquired from HSI, resulting in a new image called a “prediction map.” In this prediction map, a color scale was used to describe the different microbial loads in each spot of the sample. As shown in **Figure 3**, when the microbial loads increase, the images shift from a blue color to a more reddish one, this reflects the growth of bacteria on the chicken fillets.




Figure 3. This is an image of a median-filtered prediction map for validation set using the simplified PLSR model built on three wavelengths (930, 1121, and 1345 nm). Values under each sample predict the Enterobacteriaceae counts (in \log_{10} CFU g^{-1}) [36].

Changes in temperature during cold storage of meat products can lead to undesirable microbial growths, which may affect food safety. A study of the spoilage of beef was reported by Peng et al. [41]; in this work, HSI was exploited to measure biochemical changes within the fresh beef. The research demonstrated that HSI showed potential for real-time and nondestructive detection of bacterial spoilage in beef.

Work performed by Barbin et al. [43] used HSI in the near-infrared range (900–1700 nm) to determine the total viable count and psychotropic plate count in chilled pork during storage. NIR hyperspectral images in the reflectance mode were captured every 48 h from each sample. Assuming that meat spoilage is evident at a microbial load of 107 CFU per gram or cm^2 , the author's defined a cutoff point of 106 CFU/g as an acceptable threshold of freshness. By examining the spectral information that was obtained from the samples, a difference was observed in the wavelength range between 1300 and 1600 nm, where fresh samples had lower absorbance than spoiled samples (see **Figure 4**). This spectral region is commonly assigned to N-H stretch of proteins (amines and amides) and their interactions with water, and it could suggest the occurrence of proteolytic changes, which are recognized as the main indicator for the onset of spoilage in meat products.

In 2016, Everard et al. [51] presented fluorescence HSI coupled with multivariate image analysis techniques utilized for the detection of fecal contaminants on spinach leaves. Violet fluorescence excitation was provided at 405 nm, and light emission was recorded from 464 to 800 nm. Partial least square discriminant analysis (PLSDA) and wavelength ratio methods were compared for detection accuracy for fecal contamination. The PLSDA model had 19% false positives for nonfresh post storage leaves. A wavelength ratio technique using four wavebands (680, 688, 703, and 723 nm) was successful in identifying 100% of fecal contaminants on both fresh and nonfresh leaves.




Figure 4. (a) absorbance spectra for fresh and spoiled samples (after 7 days of storage); (b) second derivative spectra for fresh and spoiled samples showing potentially relevant wavelengths [43].

Detection of fecal contamination on cantaloupes using HS fluorescence imagery was demonstrated by Vargas et al. [46]. HS images of cantaloupes artificially contaminated with a range of diluted bovine feces were acquired from 425 to 774 nm in response to ultraviolet-A (320–400 nm) excitation. Evaluation of images at emission peak wavelengths indicated that 675 nm exhibited the greatest contrast between contaminated and untreated surface areas. Two-band ratios compared with the single-band images enhanced the contrast between the fecal contaminated spots and untreated cantaloupe surfaces.

Yang et al. [47] examined methods to classify fecal contamination on leafy greens. They utilized HS fluorescence imaging system with ultraviolet-A excitation (320–400 nm) for detection of bovine fecal contaminants on the abaxial and adaxial surfaces of romaine lettuce and baby spinach leaves. They applied six spots of fecal contamination to each of the 40 lettuce and spinach leaves. Their results showed that for both lettuce and spinach, the detection of fecal matter was best obtained using the ratio of the signal from 666 nm divided by that from 680 nm, (R values of 0.98 for romaine lettuce and 0.96 for baby spinach).

3. Raman spectroscopy and spectral imaging

Raman spectroscopy is a nondestructive spectroscopic technique, based on the vibrational properties of the constituent molecules, that provides molecular information about the sample under examination. The Raman signal results from molecules being excited by a small

amount of incident light at a specific wavelength. The remitted light has some of its photons shifted to different wavelengths by the addition or subtraction of vibrational energy from some of the tissue intramolecular bonds [44]. Contrast is achieved when the tissue molecular constituents differ enough that the Raman signals from two tissues have different wavelength distributions. Raman spectral imaging (RSI) intertwines Raman spectroscopy and digital imaging to visualize the composition and structure of a target, thereby having great potential for food safety and analysis [29]. Historically Raman imaging systems have only been able to perform Raman measurement at a microscopic level and were unable to evaluate whole surfaces of individual foods. Recent studies have shown a benchtop point-scanning Raman chemical imaging system designed and developed for food safety research [56]. Raman imaging is a highly specific and sensitive technique as it allows for the detection of particular chemicals at low concentrations, such as detecting melamine particles in dry milk. This technique has wide applications, and due to its specificity, it may help detect contaminants in food products of different sizes.

A study aimed at the detection and differentiation of important food and waterborne bacteria (*E. coli*, *Staphylococcus epidermidis*, *Listeria monocytogenes*, and *Enterococcus faecalis*) was performed by Fan et al. [38] using surface-enhanced Raman spectroscopy (SERS) coupled with intracellular nanosilver as SERS substrates. Variations observed in the spectral patterns of bacterial pathogens are due to the different quantity and distribution of cellular components like proteins, phospholipids, nucleic acids, and carbohydrates. SERS coupled with statistical analysis has become very useful in discriminating and detecting bacterial cells, spores, and viruses.

In another study, a portable Raman sensor system was presented with an integrated 671 nm microsystem diode laser as excitation light source for the rapid in situ detection of meat spoilage and bacteria [39]. The system used in this chapter is an example of the reduction in form factor of enabled by recent advances and is comprised of three main components: a handheld measurement head with a laser driver electronics board, the Raman optical bench, and finally, a battery pack. This method was used to rapidly detect meat spoilage in specific pork cuts, *musculus longissimus dorsi* (LD) and *musculus semimembranosus* (SM). The authors were able to determine the total number of mesophilic aerobic microorganisms on the surface of the meat to show possible correlations of the bacterial growth with the measured Raman spectra. In 2007, the food industry faced substantial economic losses following the discovery of melamine, a nitrogen rich chemical, in human and pet foods [48]. In one SERS study which employed SERS-active substrates, the concentration of melamine was measured in wheat gluten, chicken feed, and processed foods such as cake and noodles [49, 50].

4. Speckle imaging

Spoilage and poisoning of food products by microorganisms is a major issue in food safety and human health. As these microorganisms grow and become more active, they cause deterioration of food quality and cause food intoxication. Some of the microorganisms capable of such damage are bacteria, yeast, and mold. As detailed earlier, there have been many different

technologies developed to detect harmful microorganisms in food products such as hyperspectral imaging, Raman spectroscopy, and high-performance liquid chromatography. All these methods have certain intrinsic shortcomings. Factors such as the need for a well-equipped laboratory, high-cost equipment, complicated procedures for sample preparation and long analysis times, and trained professional operators limit their widespread application in the food processing, transportation, marketing, and preservation in various food industries.

A technology that is finding increasing favor by circumventing many of these limitations is laser speckle imaging. Laser speckle imaging has been introduced in this field of application to monitor moving particles in optically inhomogeneous media by analyzing time-varying laser speckle patterns for applications such as measuring meat quality and detecting contaminants. Unlike multiple light scattering in meat which exhibits static and deterministic speckle intensity patterns, light paths associated with the movements of living microorganisms result in time-varying changes in the speckle intensity patterns. Therefore, by detecting the decorrelation in the laser speckle intensity patterns from tissues, the living activities of microorganisms can be detected.

Another advantage of this method is the ability to examine meats sealed with transparent packaging because this method detects time-varying signals in reflected laser beams and transparent plastic does not affect these. Furthermore, the technique can provide rapid assessment as bacterial colonies can be detected within a few seconds [30]. Thus, this method provides an efficient and effective way to detect live bacteria in food products to avoid food toxicity. Speckle imaging systems have been demonstrated to indicate the presence of bacterial colonies and other contaminants in both food and water [31]. Technology such as this may be very effective in the marketplace as food producers or consumers themselves may be able to use them to assess food safety. As mentioned, there are currently several approaches available for detecting low levels of microorganisms in food; however, they require complex equipment, high costs, invasive procedures, and skilled technicians to operate which all act to restrict its widespread adoption and use in the food industry [31].

Work performed by Yoong et al. [53] aimed to detect and quantify various levels of contamination using chicken breast meat samples. The meats contaminated with bacteria had significant decreases in the autocorrelation values over the time lag, whereas the control group (meat treated with a PBS solution) did not show any major changes. The meat treated with a high concentration of bacteria had more significant changes over the time lag compared with the meat treated with a low concentration of bacteria. Moreover, the decrease in the autocorrelation value was proportional to the concentration of the treated bacteria. The measured autocorrelation values were all statistically different from one another ($p < 0.001$), and the decreases in the autocorrelation were proportional to the concentration of bacteria. Thus, the authors were able to show that through various experimental validations, spontaneous bacterial activity caused strong decorrelation in laser speckle dynamics (**Figure 5**).

In 2014, Kim et al. [55] presented a label-free bacterial colony phenotyping technology called bacterial rapid detection using optical scattering technology (BARDOT), which can provide classification for several different types of bacteria. Recent experiments with colonies of *Bacillus* species using speckle imaging show a certain speckle formation that allows for the detection and




Figure 5. This image illustrates the groups attempt at assessing bacterial activity in meat. (A) shows representative autocorrelation amps in meat treated with various concentrations of bacteria at various time lags. (B) Averaged $C(\tau)$ values over the areas in (A) as a function of the time lag. (C) quantification of the autocorrelation values at $\tau = 10s$ [53].

identification of these bacterial species. As the center diameter of the *Bacillus* spp. colony grew from 500 to 900 microns, the average speckle area decreased twofold and the number of small speckles increased sevenfold. As *Bacillus* colonies grow, the average speckle size in the scatter pattern decreases and the number of smaller speckle increases due to the swarming growth characteristics of bacteria within the colony [40]. Singh et al. showed the real-time detection and identification of *Salmonella* colonies grown from inoculated peanut butter, chicken breast, and spinach or from naturally contaminated meat using BARDOT technology (90–100% accuracy) in the presence of background microbiota from naturally contaminated meat [52].

5. Multimode hyperspectral imaging system

Due to the multicomponent nature of foods, their reflectance or fluorescence spectra are complex and chemometric methods using multivariate analysis are needed to extract contaminant-specific information. By varying both the excitation and detection wavelengths and measuring both reflectance and fluorescence emission properties of a food sample, we can fine-tune algorithms for specific foods and contaminants. It has been shown that for biological tissues, dual or multiple excitation fluorescence can increase the specificity and accuracy of classification and quantification of specific sources of fluorescence [54]. Rasch et al. [57] showed the combination of different spectroscopic methods (such as fluorescence and NIR spectroscopy) becomes a promising approach to circumvent such single method inherent limitations and to use optical sensing for in situ mycotoxin detection. Additional chemometric tools are essential to eliminate disturbing factors and to extract the desired biochemical information with respect to contamination with fungi and/or mycotoxins.

An example of a multimode hyperspectral imaging system operates in fluorescence and reflectance modes as well as speckle imaging is shown in **Figure 6** developed by SafetySpect Inc. The system uses spectral band sequential imaging on the detection side. To ensure high signal-to-noise level, camera and spectral selection filter integration time is optimized for each spectral band from visible to the near infrared. The illumination module uses two independent light sources to provide illumination for fluorescence excitation and reflectance measurements using three computer-controlled LED illumination rings. The UVA (375 nm) and blue/violet (420 nm) LED rings provide fluorescence excitation. White LEDs will be used for reflectance illumination. The HSi-440CO hyperspectral imaging system (Gooch & Housego, UK, originally developed by ChromoDynamics, Inc.) incorporated in the proposed system can image and analyze multiple signals in fixed and living cells at video rates. Its tunable filter can switch wavelengths within microseconds. The system acquires multiwavelength, high-spatial and spectral resolution image datasets, and can compute and display quantitative signal-specific images in near real time. The spectrally controllable image capture system can record spectral images of food samples in wavelengths ranging from 450 to 800 nm. The system is configured as a tabletop platform where illumination and detection will operate above the food sample.

In this system, time-varying speckle signals can be quantitatively addressed with the speckle correlation time. A sample containing living microorganisms will have a correlation time way shorter than a static one, and thus contaminated food will be less time-correlated as compared to fresh food due to the spontaneous motility of microorganisms. Correlation time of scattered light from samples, the presence and activity of microorganisms can be quantitatively analyzed.




Figure 6. Configuration of the multimode HSI.

Let us consider $I(x, y, t)$ the image of the sample at time t . The correlation coefficient between two images of the sample at different times is given by the normalized autocorrelation function:

$$C(x, y, t) = \frac{1}{T-\tau} \sum_{t=1}^{T-\tau} I(x, y, t) \cdot I(x, y, t + \tau) \delta t \quad (1)$$

where T is the total acquisition time, δt the time difference, and τ the time lag. In the case of food contamination assessment, the sample is expected to be static and the correlation to be close to the unity. Every decorrelation effect is due, then, to the presence of live microorganisms moving across the sample.

6. Conclusions

There is inherent risk in food (preparing, selling, and consuming it), and we need better ways of minimizing such risk. The number of people who are sickened by problematic food is staggering (it is estimated that 1/6 of the US population is thus affected yearly), and the number of people who die (~3000/year) is unacceptable. If one examines the rather extensive risk management/mitigation literature, it is evident that certain fields of human endeavor (such as air travel) are doing a better job than others in minimizing the undesirable scenarios. A particularly pragmatic take on this field was provided by Dr. J. Reason [58], who developed an approach he termed the Swiss Cheese theory (**Figure 7**). Basically, he posits that we all want to insert countermeasures between us and hazards, to prevent harm, but because we are human and thus imperfect, these countermeasures are like a slice of Swiss cheese. The most logical and direct improvement is to “stack” the slices of cheese, as the holes do not align, and prevention is achieved. Translated to imaging for food safety, this calls for a multimode approach, which is what we propose (see **Figure 8**). The number of modes needed for good performance scales, naturally, with the difficulty of the problem, and we plan to have our implementations reflect this, in the sense that a large, real-time production-level device will be based on more modes than a mid-level (e.g., restaurant) one, while a handheld, portable unit may only address 80% of the challenges, but with ~20% of the cost and size.




Figure 7. Dr. Reason’s Swiss cheese theory of accident causation/prevention.

MULTIMODE OPTICAL IMAGING: A FUNNEL OF METHODS




Figure 8. Multimode imaging as a funnel of methods. The right mix (on the same instrument, in the proper sequence) optimizes performance, speed and cost simultaneously.

Author details

Fartash Vasefi, Nicholas Booth, Hesam Hafizi and Daniel L. Farkas*

*Address all correspondence to: dlfarkas@gmail.com

SafetySpect, Inc., Sherman Oaks, CA, USA

References

- [1] Gowen AA, O'Donnell C, Cullen PJ, Downey G, Frias JM. Hyperspectral imaging—An emerging process analytical tool for food quality and safety control. *Trends in Food Science & Technology*. 2007;**18**(12):590-598
- [2] Kim MS, Chen YR, Mehl PM. Hyperspectral reflectance and fluorescence imaging system for food quality and safety. *Transactions of the ASAE*. 2001;**44**(3):721
- [3] Goetz AF, Vane G, Solomon JE, Rock BN. Imaging spectrometry for earth remote sensing. *Science*. 1985 Jun 7;**228**(4704):1147-1153
- [4] Liu Y, Chen YR, Kim MS, Chan DE, Lefcourt AM. Development of simple algorithms for the detection of fecal contaminants on apples from visible/near infrared hyperspectral reflectance imaging. *Journal of Food Engineering*. 2007 Jul 1;**81**(2):412-418

- [5] Mehl PM, Chen YR, Kim MS, Chan DE. Development of hyperspectral imaging technique for the detection of apple surface defects and contaminations. *Journal of Food Engineering*. 2004 Jan 1;61(1):67-81
- [6] Nicolai BM, Lötze E, Peirs A, Scheerlinck N, Theron KI. Non-destructive measurement of bitter pit in apple fruit using NIR hyperspectral imaging. *Postharvest Biology and Technology*. 2006 Apr 1;40(1):1-6
- [7] Xing J, Bravo C, Jancsó PT, Ramon H, De Baerdemaeker J. Detecting bruises on 'golden delicious' apples using hyperspectral imaging with multiple wavebands. *Biosystems Engineering*. 2005 Jan 1;90(1):27-36
- [8] Xing J, Jancsó P, De Baerdemaeker J. Stem-end/calyx identification on apples using contour analysis in multispectral images. *Biosystems Engineering*. 2007 Feb 1;96(2):231-237
- [9] Xing J, Saeys W, De Baerdemaeker J. Combination of chemometric tools and image processing for bruise detection on apples. *Computers and Electronics in Agriculture*. 2007 Mar 1;56(1):1-3
- [10] Weinstock BA, Janni J, Hagen L, Wright S. Prediction of oil and oleic acid concentrations in individual corn (*Zea mays* L.) kernels using near-infrared reflectance hyperspectral imaging and multivariate analysis. *Applied Spectroscopy*. 2006 Jan 1;60(1):9-16
- [11] Ariana DP, Lu R, Guyer DE. Near-infrared hyperspectral reflectance imaging for detection of bruises on pickling cucumbers. *Computers and Electronics in Agriculture*. 2006 Aug 1;53(1):60-70
- [12] Cheng X, Chen YR, Tao Y, Wang CY, Kim MS, Lefcourt AM. A novel integrated PCA and FLD method on hyperspectral image feature extraction for cucumber chilling damage inspection. *Transactions of the ASAE*. 2004;47(4):1313
- [13] Liu Y, Chen YR, Wang CY, Chan DE, Kim MS. Development of a simple algorithm for the detection of chilling injury in cucumbers from visible/near-infrared hyperspectral imaging. *Applied Spectroscopy*. 2005 Jan;59(1):78-85
- [14] Menesatti P, Urbani G, Lanza G. Spectral imaging Vis-NIR system to forecast the chilling injury onset on citrus fruits. In *V International Postharvest Symposium*. 2004 Jun 6;682:1347-1354
- [15] Monteiro ST, Minekawa Y, Kosugi Y, Akazawa T, Oda K. Prediction of sweetness and amino acid content in soybean crops from hyperspectral imagery. *ISPRS Journal of Photogrammetry and Remote Sensing*. 2007 May 1;62(1):2-12
- [16] Lu R, Peng Y. Hyperspectral scattering for assessing peach fruit firmness. *Biosystems Engineering*. 2006 Feb 1;93(2):161-171
- [17] Qiao J, Wang N, Ngadi MO. Water content and weight estimation for potatoes using hyperspectral imaging. In *2005 ASAE Annual Meeting 2005* (p. 1). American Society of Agricultural and Biological Engineers. p. 053126

- [18] ElMasry G, Wang N, ElSayed A, Ngadi M. Hyperspectral imaging for nondestructive determination of some quality attributes for strawberry. *Journal of Food Engineering*. 2007 Jul 1;81(1):98-107
- [19] Tallada JG, Nagata M, Kobayashi T. Non-destructive estimation of firmness of strawberries (*Fragaria× ananassa* Duch.) using NIR hyperspectral imaging. *Environmental Control in Biology*. 2006;44(4):245-255
- [20] Nagata M, Tallada JG, Kobayashi T. Bruise detection using NIR hyperspectral imaging for strawberry (*Fragaria× ananassa* Duch.). *Environmental Control in Biology*. 2006;44(2):133-142
- [21] Lawrence KC, Windham WR, Park B, Heitschmidt GW, Smith DP, Feldner P. Partial least squares regression of hyperspectral images for contaminant detection on poultry carcasses. *Journal of Near Infrared Spectroscopy*. 2006 Aug 1;14(4):223-230
- [22] Park B, Lawrence KC, Windham WR, Smith DP. Performance of hyperspectral imaging system for poultry surface fecal contaminant detection. *Journal of Food Engineering*. 2006 Aug 1;75(3):340-348
- [23] Park B, Windham WR, Lawrence KC, Smith DP. Contaminant classification of poultry hyperspectral imagery using a spectral angle mapper algorithm. *Biosystems Engineering*. 2007 Mar 1;96(3):323-333
- [24] Kim I, Kim MS, Chen YR, Kong SG. Detection of skin tumors on chicken carcasses using hyperspectral fluorescence imaging. *Transactions of the ASAE*. 2004;47(5):1785
- [25] Jiang H, Yoon SC, Zhuang H, Wang W, Lawrence KC, Yang Y. Tenderness classification of fresh broiler breast fillets using visible and near-infrared hyperspectral imaging. *Meat Science*. 2018 Jan 31;139:82-90
- [26] Qiao J, Ngadi MO, Wang N, Gariépy C, Prasher SO. Pork quality and marbling level assessment using a hyperspectral imaging system. *Journal of Food Engineering*. 2007 Nov 1;83(1):10-16
- [27] Zhu WY, Su WH, Sun DW. Measurement of tenderness of red meats using hyperspectral imaging: A brief review. *Biosystems and Food Engineering Research Review*. 2017 May;22:66
- [28] Xiong Z, Sun DW, Zeng XA, Xie A. Recent developments of hyperspectral imaging systems and their applications in detecting quality attributes of red meats: A review. *Journal of Food Engineering*. 2014 Jul 1;132:1-3
- [29] Qin J, Chao K, Kim MS. Raman chemical imaging system for food safety and quality inspection. *Transactions of the ASABE*. 2010;53(6):1873-1882
- [30] Bianco V, Mandracchia B, Nazzaro F, Marchesano V, Gennari O, Paturzo M, Ferraro P, et. al Food quality inspection by speckle decorrelation properties of bacteria colonies. In: *Optical Methods for Inspection, Characterization, and Imaging of Biomaterials III* (Vol. 10333, p. 103331N). International Society for Optics and Photonics; 2017, June
- [31] Yoon J, Lee K, Park Y. A simple and rapid method for detecting living microorganisms in food using laser speckle decorrelation. *arXiv preprint arXiv:1603.07343*; 2016

- [32] Kamruzzaman M, ElMasry G, Sun DW, Allen P. Non-destructive prediction and visualization of chemical composition in lamb meat using NIR hyperspectral imaging and multivariate regression. *Innovative Food Science & Emerging Technologies*. 2012;**16**:218-226
- [33] Tao F, Peng Y. A method for nondestructive prediction of pork meat quality and safety attributes by hyperspectral imaging technique. *Journal of Food Engineering*. 2014;**126**:98-106
- [34] Pu YY, Feng YZ, Sun DW. Recent progress of hyperspectral imaging on quality and safety inspection of fruits and vegetables: A review. *Comprehensive Reviews in Food Science and Food Safety*. 2015;**14**(2):176-188
- [35] Cheng JH, Sun DW. Rapid quantification analysis and visualization of *Escherichia coli* loads in grass carp fish flesh by hyperspectral imaging method. *Food and Bioprocess Technology*. 2015;**8**(5):951-959
- [36] Feng YZ, ElMasry G, Sun DW, Scannell AG, Walsh D, Morcy N. Near-infrared hyperspectral imaging and partial least squares regression for rapid and reagentless determination of Enterobacteriaceae on chicken fillets. *Food Chemistry*. 2013;**138**(2): 1829-1836
- [37] Del Fiore A, Reverberi M, Ricelli A, Pinzari F, Serranti S, Fabbri AA, Fanelli C, et al. Early detection of toxigenic fungi on maize by hyperspectral imaging analysis. *International Journal of Food Microbiology*. 2010;**144**(1):64-71
- [38] Fan C, Hu Z, Mustapha A, Lin M. Rapid detection of food-and waterborne bacteria using surface-enhanced Raman spectroscopy coupled with silver nanosubstrates. *Applied Microbiology and Biotechnology*. 2011;**92**(5):1053-1061
- [39] Sowoidnich K, Schmidt H, Kronfeldt HD, Schwägele F. A portable 671 nm Raman sensor system for rapid meat spoilage identification. *Vibrational Spectroscopy*. 2012;**62**:70-76
- [40] Kim H, Singh AK, Bhunia AK, Bae E. Laser-induced speckle scatter patterns in *Bacillus* colonies. *Frontiers in Microbiology*. 2014;**5**:537
- [41] Peng Y, Zhang J, Wang W, Li Y, Wu J, Huang H, Jiang W, et al. Potential prediction of the microbial spoilage of beef using spatially resolved hyperspectral scattering profiles. *Journal of Food Engineering*. 2011;**102**(2):163-169
- [42] Gómez-Sanchis J, Moltó E, Gomez-Chova L, Aleixos N, Camps-Valls G, Juste F, Blasco J. Hyperspectral computer vision system for the detection of *Penicillium digitatum* in citrus packing lines. In: 2004 CIGR International Conference, Beijing, China. 2004, October. (pp. 11-14)
- [43] Barbin DF, ElMasry G, Sun DW, Allen P, Morsy N. Non-destructive assessment of microbial contamination in porcine meat using NIR hyperspectral imaging. *Innovative Food Science & Emerging Technologies*. 2013;**17**:180-191
- [44] Lohumi S, Lee S, Lee H, Cho BK. A review of vibrational spectroscopic techniques for the detection of food authenticity and adulteration. *Trends in Food Science & Technology*. 2015 Nov;**46**(1, 1):85-98

- [45] Kim MS, Lefcourt AM, Chao K, Chen YR, Kim I, Chan DE. Multispectral detection of fecal contamination on apples based on hyperspectral imagery: Part I. Application of visible and near-infrared reflectance imaging. *Transactions of the ASAE*. 2002;**45**(6):2027
- [46] Vargas AM, Kim MS, Tao Y, Lefcourt AM, Chen YR, Luo Y, et al. Detection of fecal contamination on cantaloupes using hyperspectral fluorescence imagery. *Journal of Food Science*. 2005;**70**(8):471-476
- [47] Yang CC, Jun W, Kim MS, Chao K, Kang S, Chan DE, Lefcourt A. Classification of fecal contamination on leafy greens by hyperspectral imaging. *Sensing for Agriculture and Food Quality and Safety II*, Proceedings of the SPIE. 2010;**7676**:76760F-767601F
- [48] Domingo E, Tirelli AA, Nunes CA, Guerreiro MC, Pinto SM. Melamine detection in milk using vibrational spectroscopy and chemometrics analysis: A review. *Food Research International*. 2014 Jun 1;**60**:131-139
- [49] Zhou N, Zhou Q, Meng G, Huang Z, Ke Y, Liu J, Wu N. Incorporation of a basil-seed-based surface enhanced Raman scattering sensor with a pipet for detection of melamine. *ACS Sensors*. 2016 Oct 14;**1**(10):1193-1197
- [50] Li X, Feng S, Hu Y, Sheng W, Zhang Y, Yuan S, Zeng H, Wang S, Lu X. Rapid detection of melamine in milk using immunological separation and surface enhanced Raman spectroscopy. *Journal of Food Science*. 2015 Jun 1;**80**(6):C1196-C1201
- [51] Everard CD, Kim MS, Cho H, O'Donnell CP. Hyperspectral fluorescence imaging using violet LEDs as excitation sources for fecal matter contaminate identification on spinach leaves. *Journal of Food Measurement and Characterization*. 2016 Mar;**10**(1, 1):56-63
- [52] Singh AK, Bettasso AM, Bae E, Rajwa B, Dundar MM, Forster MD, Liu L, Barrett B, Lovchik J, Robinson JP, Hirleman ED. Laser optical sensor, a label-free on-plate salmonella enterica colony detection tool. *MBio*. 2014 Feb 28;**5**(1):e01019-e01013
- [53] Yoon J, Lee K, Park Y. A simple and rapid method for detecting living microorganisms in food using laser speckle decorrelation. *arXiv preprint arXiv*. 2016 Mar 18;**1603**:07343
- [54] Sauvage VR, Levene AP, Nguyen HT, Wood TC, Kudo H, Concas D, Thomas HC, Thursz MR, Goldin RD, Anstee QM, Elson DS. Multi-excitation fluorescence spectroscopy for analysis of non-alcoholic fatty liver disease. *Lasers in Surgery and Medicine*. 2011;**43**(5):392-400
- [55] Kim H, Singh AK, Bhunia AK, Bae E. Laser-induced speckle scatter patterns in bacillus colonies. *Frontiers in Microbiology*. 2014 Oct 14;**5**:537
- [56] Qin J, Chao K, Kim MS. Development of a Raman chemical imaging system for food safety inspection. In 2010 Pittsburgh, Pennsylvania, June 20-June 23, 2010 2010 (p. 1). American Society of Agricultural and Biological Engineers
- [57] Rasch C, Kumke M, Löhmannsröben HG. Sensing of mycotoxin producing fungi in the processing of grains. *Food and Bioprocess Technology*. 2010 Dec 1;**3**(6):908-916
- [58] Reason J. Human error: models and management. *BMJ*. 2000;**320**(7237):768-770