

SHORT COMMUNICATION



Does electrical capacitance represent roots in the soil?

Imre Cseresnyés¹ · Eszter Vozáry² · Kálmán Rajkai¹Received: 30 October 2019 / Revised: 6 January 2020 / Accepted: 4 April 2020
© The Author(s) 2020

Abstract

The efficiency of the root capacitance method is still judged contradictorily. It was, therefore, aimed to verify that the electrical capacitance (C_R) and impedance (Z_R) measured between ground electrode (pushed into the soil) and plant electrode (clamped on the stem) really represented the root system present in the soil. An ‘electrode separation experiment’ and a ‘root excision experiment’ were performed on potted maize plants cultivated in arenosol. The partial electrical separation of the ground electrode (restriction of the current flow in the topsoil) caused no significant changes in the capacitance and impedance of either the soil or the plant–soil system. This suggested that the current flows through a multitude of soil solution pathways, not predominantly on the wet soil surface. The progressive excision of the main root branches caused a gradual decrease in C_R (31–39% of the initial value) and an increase in Z_R (4.4- to 5.2-fold that of intact plants). C_R showed a much stronger correlation with root dry mass ($R^2=0.807$) than with stem cross-sectional area ($R^2=0.424$) in intact plants. The closeness of the latter relationship decreased further ($R^2=0.377$) after whole root excision. The results clearly demonstrated that the root system dominated the measured capacitance and impedance, though the role of the stem base was not negligible. In conclusion, the capacitance method is worth considering in future studies as an adequate non-intrusive approach to evaluate root size in the soil.

Keywords Electrical capacitance method · Electrical impedance · Root system size · Root excision · Soil dielectrics

Introduction

The root electrical capacitance (C_R) method is based on the polarization of root membranes, which induces changes in the amplitude and phase of an alternating current (AC) driven through the root system. The C_R value detected between the ground and plant electrodes was significantly correlated with the root dry mass (RDM), root length, and surface area (Chloupek et al. 2010). The in situ technique provided reliable estimates of the whole root system size (RSS) or the absorptive root surface area for both pot- and field-grown herbaceous plants and tree saplings (Kormanek et al. 2016; Postic and Doussan 2016; Cseresnyés et al.

2018). However, as C_R is very sensitive to soil texture and composition, soil water content (SWC), and electrode protocol, data are only comparable for the same species when grown and measured under the same conditions (Chloupek et al. 2010). The great advantage of the method is the possibility of screening numerous plants quickly, and tracking root dynamics by repeated measurements at different phenological stages. The approach is thus ideal for comparison of individuals or genotypes during plant breeding (Svačina et al. 2014), though root morphology cannot be visualized. According to the widely accepted model (Dalton 1995), roots are lossy capacitors (parallel RC circuits), in which root membranes act as dielectrics, separating the conductive root sap and soil solution. The membranes store charges, exhibiting a capacitance proportional to their surface area. Later, Rajkai et al. (2005) proposed a two-dielectric model, which considered the soil capacitance as well.

In contrast, Aulen and Shipley (2012) found the capacitance method unreliable to evaluate RSS in the soil because of the unpredictable effect of external factors on the measured C_R value. Furthermore, Kendall et al. (1982) reported that C_R of field-grown alfalfa plants reduced significantly but

Communicated by O. Ferrarese-Filho.

✉ Imre Cseresnyés
cseresnyes.imre@agrar.mta.hu¹ Centre for Agricultural Research, Institute for Soil Sciences and Agricultural Chemistry, Herman Ottó út 15., Budapest 1022, Hungary² Department of Physics and Control, Szent István University, Somlói út 14–16., Budapest 1118, Hungary

only slightly (by ~10%) by severing most of the root system; however, detailed methodology for root excision was not provided. Dietrich et al. (2012) reached similar conclusions in their root cutting experiment on hydroponically grown barley. The authors stated that the bulk of the root system made a negligible contribution to C_R , which was determined by the capacitance of the tissue between the solution surface and the plant electrode. Furthermore, the C_R detected was directly proportional to the cross-sectional area (CSA) at the solution surface, and the linear C_R –RSS correlations could only be explained as an allometric relationship between RSS and CSA. They also found that, in soils, only the moisture content in the top layer influenced C_R , implying that current preferred to pass between the electrodes through the more conductive soil surface instead of through the roots (Dietrich et al. 2013).

Nevertheless, the revised model suggested by Dietrich et al. (2012, 2013) contradicts several previous and recent observations. For example, the colonization of maize roots with arbuscular mycorrhizal fungi had no influence on RDM and reduced the root length and surface area, whereas it increased C_R due to the enhanced root–soil interface caused by the growth of external hyphae (Cseresnyés et al. 2013). The monitoring of annual crops and young woody species in soil revealed characteristic seasonal changes in C_R , which were related to the dynamics of whole-plant transpiration and root uptake activity (Psarras and Merwin 2000; Vamerli et al. 2009; Cseresnyés et al. 2018). Likely due to similar reasons, root studies in oilseed plants and young trees showed diurnal patterns in spectral polarization signatures (Weigand and Kemna 2019) and in the magnitude of C_R

(Ellis et al. 2013b). Kormanek et al. (2016) found a closer correlation between C_R and root size parameters for beech seedlings immersed in hydroponic solution when the root systems were flattened between two flat-plate electrodes and then when the roots were freely arranged inside a cylindrical electrode.

These observations can hardly be explained by the dominance of the plant stem base in the resulting C_R . A pot trial with soil-grown maize was, therefore, undertaken to verify the hypothesis that (1) the current flows between the ground and plant electrodes not only through the soil but also through the roots; and (2) roots play a key role in determining C_R , confirming the ability of the capacitance method to evaluate RSS in soils.

Materials and methods

Twelve 2.9 L plastic pots (14 cm height, 17 cm upper diameter) were filled with 4.0 kg of air-dried, coarsely sieved arenosol (80.3% sand, 12.7% silt, 7.0% clay) with a $\text{pH}_{\text{H}_2\text{O}}$ of 7.21, 0.23% lime content, 1.31% humus content, and $0.21 \text{ cm}^3 \text{ cm}^{-3}$ SWC at field capacity. An open PVC cylinder (11 cm outer diameter, 8 cm height, and 0.2 cm wall thickness) was installed vertically to a depth of 7 cm on each pot in an eccentric position (Fig. 1). The pots were watered to field capacity on a balance ($\pm 1 \text{ g}$), and then, a single 2-day-old maize seedling (*Zea mays* L., composite cross population) was planted at 2 cm depth in the center of each pot. The plants were cultivated in a growth chamber at 28/20 °C with a 16 h photoperiod and PAR of $500 \mu\text{mol m}^{-2} \text{ s}^{-1}$ with




Fig. 1 Scheme of the electrode separation experiment (Exp. 1). See the text for details

daily irrigation to field capacity. The SWC was checked with a Trime-HD2 TDR meter attached to a Pico32 probe with 11 cm rod length (IMKO GmbH., Ettlingen, Germany).

The parallel electrical capacitance (C_R) and then, after switching the measurement mode, electrical impedance (Z_R) were detected using a GW-8101G LCR bridge (GW Instek Co. Ltd., Taiwan) at 1 kHz and 1 V AC. The ground electrode was a sharpened stainless steel rod (15 cm length and 0.6 cm diameter) pushed vertically into the soil to 12 cm depth 5 cm from the stem. The plant electrode was clamped at the same height (1.5 cm above the soil surface) on each plant through a 0.4 cm wide aluminum strip that bent the stem. Conductivity gel was smeared under the strip.

Electrode separation experiment (Exp. 1): An hour before the electrical measurement, the soil was adjusted to field capacity (checked with TDR meter) both inside and outside the plastic cylinder in each pot. The cylinder edge above the soil surface was wiped dry to ensure electrical insulation. An ‘inner’ and an ‘outer’ ground electrode were positioned inside and outside the cylinder, respectively, both 5 cm away from the stem (Fig. 1). First, C_R and Z_R were measured between the plant electrode and the ‘inner’ ground electrode. The readings were then repeated after relocating the instrument terminal to the ‘outer’ ground electrode, while the plant electrode was left untouched. Three measurements were taken with both electrode combinations and were averaged for each of the plants on five occasions (at plant ages of 7, 14, 22, 30, and 38 days). Before planting, the electrical capacitance and impedance of the moistened soil (C_{Soil} and Z_{Soil}) were detected using two ground electrodes positioned 5 cm apart, first with both of them inside and then with one of them outside the PVC cylinder (three measurements per pot). At the end of the experiment, the cylinders were cautiously lifted out of the soil. Paired t tests were performed to assess statistical differences ($p < 0.05$) in C_{Soil} , Z_{Soil} , C_R , and Z_R between the ‘inner’ and ‘outer’ ground electrode locations for the twelve pots.

Root excision experiment (Exp. 2): All the plants were grown until the 42nd day, when they already have thick crown roots but still no brace roots in contact with the soil. Two perpendicular stem diameters were measured at the soil surface with a digital caliper to determine CSA, after which the plants were tied to wooden sticks to prevent lodging. After checking SWC in the pot, C_R and Z_R were detected for the intact plant. Keeping both electrodes in contact, a crown root was sought by scraping out the wet soil 1–1.5 cm away from the edge of the stem. The root branch was excised below the soil level (~1 cm), the soil was smoothed back carefully, and the LCR meter was read again. By going around the plant, the next crown root was excavated, cut, and buried, followed by the detection of the electrical properties. Root excision and instrumental measurement were continued with the other individual crown roots and then the seminal

roots beneath the stem. After the last coarse-root excision, the roots were washed out of the soil over a 0.2-mm sieve (Oliveira et al. 2000), floated (to minimize the loss of fine roots), oven-dried at 70 °C, and weighed (± 0.001 g). The proximal root segments were cut off the root crown, dried and weighed separately. The total RDM was calculated for each plant. The effect of root excision on the C_R and Z_R values was statistically evaluated with the paired t test. Linear regression was used to relate C_R to CSA or RDM.

Results and discussion

In Exp. 1, C_{Soil} and Z_{Soil} values of 80.8 ± 4.9 nF (mean \pm SD; $n = 12$) and 508 ± 38 Ω , respectively, were measured between the two ground electrodes placed inside the PVC cylinder, while values of 79.9 ± 4.4 nF and 512 ± 40 Ω were obtained when one electrode was positioned outside the cylinder. According to the paired t test, both the decrease in C_{Soil} and the increase in Z_{Soil} caused by electrode separation were non-significant ($p = 0.134, 0.203$). The magnitude of C_R increased and that of Z_R decreased continuously with plant age due to root development. Electrode separation had no significant influence on C_R or Z_R at any measurement time (Table 1). The mean rate of change ranged from -1.09 to 1.28% and from -1.23 to 1.26% for C_R and Z_R , respectively.

In Exp. 2, the 42-day-old intact plants exhibited C_R and Z_R values of 8.44–13.54 nF and 3.83–6.18 k Ω , respectively. A total of 10–14 root branches per plant (5–8 crown roots and 4–7 seminal roots) were excised. The root cutting process resulted in a progressive decrease in C_R to 3.01–4.52 nF (Fig. 2a), which was 31–39% of the initial value, and led to a continuous increase in Z_R to 17.2–29.8 k Ω (Fig. 2b), which was 4.4- to 5.2-fold higher than that of intact plants. The paired t test revealed that both C_R and Z_R changed significantly ($p < 0.001$) even after the first root excision. CSA ranged from 128.6 to 213.1 mm², and showed only weakly significant, linear correlations with the C_R values measured both for the intact plants ($R^2 = 0.424$; $F = 7.35$; $p = 0.022$) and after total root cutting ($R^2 = 0.377$; $F = 6.06$; $p = 0.034$; Fig. 2c). The harvested plants had total RDM values ranging from 2.326 to 4.125 g, 4.2–6.8% of which (0.106–0.256 g) was composed of the proximal root segments excised from the crown terminally. A highly significant linear relationship was found between C_R and the total RDM ($R^2 = 0.807$; $F = 41.7$; $p < 0.001$; Fig. 2d).

The separation of the two ground electrodes by the PVC cylinder caused no significant differences in the C_{Soil} or Z_{Soil} values. This finding directly contradicts the hypothesis that AC flows between the electrodes principally on the wet soil surface. In accordance, the transferal of the ground electrode from the inner to the outer side of the cylinder had no influence on the C_R and Z_R values measured even for young

Table 1 Mean \pm SD ($n=12$) of root electrical capacitance (C_R ; nF) and root electrical impedance (Z_R ; k Ω) detected for maize plants at different ages (DAP: days after planting), including an ‘inner ground

electrode’ (IGE) or an ‘outer ground electrode’ (OGE) in the measurement circuit (see text for details)

DAP	C_R (nF)				Z_R (k Ω)			
	IGE	OGE	RC %	p	IGE	OGE	RC %	p
7	2.10 \pm 0.25	2.13 \pm 0.27	1.28	0.173	40.74 \pm 3.73	41.23 \pm 3.57	1.26	0.135
14	3.68 \pm 0.34	3.70 \pm 0.27	0.61	0.643	20.91 \pm 2.51	20.68 \pm 2.79	-1.23	0.238
22	6.45 \pm 0.68	6.38 \pm 0.64	-0.89	0.224	9.56 \pm 1.24	9.62 \pm 1.23	0.75	0.449
30	8.96 \pm 0.87	8.88 \pm 1.08	-1.09	0.363	6.16 \pm 0.67	6.09 \pm 0.63	-0.88	0.282
38	10.57 \pm 1.34	10.63 \pm 1.35	0.60	0.375	5.16 \pm 0.80	5.19 \pm 0.78	0.66	0.386

RC (%) shows the rate of change in C_R or Z_R when the outer ground electrode was attached instead of the inner one. p represents the significance level obtained using paired t tests




Fig. 2 Changes in **a** root electrical capacitance (C_R in nF; mean \pm SD, $n=12$) and **b** root electrical impedance (Z_R in k Ω ; mean \pm SD, $n=12$) in relation to the number of excised root branches (N_{ERB}) for maize plants. **c** Relationship between C_R and the cross-sectional area of the

stem at the soil surface (CSA; mm 2) for maize plants with intact root systems (filled square; $F=7.35$; $p=0.022$) and after root excision (square; $F=6.06$; $p=0.034$). **d** Relationship between C_R and total root dry mass (RDM; g) for maize plants ($F=41.7$; $p<0.001$)

(7-day-old) plants, whose entire root system probably grew inside the cylinder. These results verify the basic assumption of Dalton's (1995) model, that electric current flows between the electrodes through a multitude of soil solution pathways formed by the structural porosity. A more recent study on canola reported closer relationships between C_R and RSS in dry than in wet soil (Wu et al. 2017), which also contradicts the dominant role of the moistened topsoil in conducting electric current.

The gradual drop in C_R and rise in Z_R with the increasing number of excised root branches clearly demonstrated the significance of the root system in the plant–soil impedance response. The change induced in C_R or Z_R by cutting a root branch was variable. This was presumably due to differences in anatomy, size (surface area), uptake activity, and conductivity not only between the crown and the seminal roots but also between the individual root branches (Ahmed et al. 2018). Pattern of resource allocation was reported to result close allometric association between shoot and root biomass for maize (Gerardo et al. 2013). In our case, the measured C_R showed a much stronger correlation with RDM than with CSA. Additionally, the closeness of the relationship between C_R and CSA decreased further when the whole root system was cut off. The same was also true for Z_R (data not presented).

These results were similar to those obtained by Čermák et al. (2006). Although they reported a significant allometric relationship between the basal area (stem CSA) and absorbing root surface area for several broadleaf and coniferous tree species, the root severing of Norway spruce supported a direct negative correlation between the measured impedance and the active root surface area. Moreover, a 'root cutting experiment' and a 'progressively immersed root system experiment' involving hydroponically grown spinach and tomato provided evidence that only the root part immersed in the medium contributed significantly to the C_R detected (Ozier-Lafontaine and Bajazet 2005).

Opinions still differ on how the various parts of the root system participate in electric conduction and contribute to the measured capacitance. Urban et al. (2011) suggested that, as most of the charge carriers pass through plant tissues near the root neck, the impedance response is predominantly determined by the proximal part of the root–soil interface, and only slightly affected by the distal fine roots. In contrast, Ellis et al. (2013a) showed that most if not the entire length of the root system was electrically connected, so it was the thick proximal roots that had a lesser influence on C_R . These authors found that the stem and soil impedances were principally resistive and, consequently, the roots represented the main capacitance element in the plant–root–soil system (Ellis et al. 2013a, b). In the present case, the increasing Z_R caused by progressive root cutting clearly demonstrated a reduction in the electrically conductive cross-sectional area.

At the end of root excision, the plants exhibited about a third of the initial C_R value. As the root stumps left in the crowns represented only 4–7% of the total RDM (and obviously a negligible part of the total root surface area), the remaining C_R was suspected to be the capacitance of the stem base. This finding partially agrees with the revised model offered by Dietrich et al. (2012, 2013), and demonstrates that the role of the stem in determining the magnitude of C_R is not negligible. The remaining capacitance is thought to be due to the polarization of the epidermal membranes in the root neck and the stem base. The two-dielectric capacitor model assumes that if the soil capacitance is substantially higher than the root capacitance, C_R is basically determined by the root system (Rajkai et al. 2005; Dietrich et al. 2012; Kormanek et al. 2016). In the present study, C_{Soil} for the wet substrate proved to be much higher than C_R , which met the criteria of this model. Importantly, the dielectric behavior of soils may differ considerably from those of hydroponic solutions due to the presence of surface-charged particles such as colloids, and to water binding by the soil matrix (Hilhorst 1998). The pathways of a low-frequency current driven through a solid substrate–root system are presumably determined jointly by the component resistances and capacitances.

The present experiment did not reveal what proportion of the whole root system was responsible for the resultant C_R . However, the results convincingly showed that, in contrast to some previous suggestions, the stem base was an important but not dominant component of C_R under the given experimental conditions. The findings may necessitate developing improved models for the underlying current pathways. The importance of this is based on the inconsistent results obtained for the role of roots in forming the impedance response.

Although it is true that the electrophysical basis of the stem–root–soil circuit is still poorly understood and requires further comprehensive research, the present study provided strong evidence that the electrical capacitance detected between ground and plant electrodes represented the root system growing in the soil environment. Consequently, this measurement technique should be considered as an adequate tool to assess RSS in situ.

Author contribution statement ICs designed and carried out the experimental work, read the relevant literature, and wrote the paper. EV was responsible for giving biophysical explanations of the present findings. KR supervised the work and helped in data interpretation. All the authors read and improved the manuscript critically and approved the submission.

Acknowledgements Open access funding provided by Centre for Agricultural Research. This project was financed by the National Research, Development and Innovation Fund of Hungary (NKFIH K-115714), GINOP-2.3.2-15-2016-00028, and a János Bolyai Research Scholarship from the Hungarian Academy of Sciences. The authors thank Barbara Harasztos for language editing.

Compliance with ethical standards

Conflict of interest The authors declare that they have no conflict of interest.

Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>.

References

- Ahmed MA, Zarebanadkouki M, Meunier F, Javaux M, Kaestner A, Carminati A (2018) Root type matters: measurement of water uptake by seminal, crown, and lateral roots in maize. *J Exp Bot* 69:1199–1206. <https://doi.org/10.1093/jxb/erx439>
- Aulen M, Shipley B (2012) Non-destructive estimation of root mass using electrical capacitance on ten herbaceous species. *Plant Soil* 355:41–49. <https://doi.org/10.1007/s11104-011-1077-3>
- Čermák J, Ulrich R, Staněk Z, Koller J, Aubrecht L (2006) Electrical measurement of tree root absorbing surfaces by the earth impedance method: 2. Verification based on allometric relationships and root severing experiments. *Tree Physiol* 26:1113–1121. <https://doi.org/10.1093/treephys/26.9.1113>
- Chloupek O, Dostál V, Středa T, Psota V, Dvořáčková O (2010) Drought tolerance of barley varieties in relation to their root system size. *Plant Breed* 129:630–636. <https://doi.org/10.1111/j.1439-0523.2010.01801.x>
- Cseresnyés I, Takács T, Végh RK, Anton A, Rajkai K (2013) Electrical impedance and capacitance method: a new approach for detection of functional aspects of arbuscular mycorrhizal colonization in maize. *Eur J Soil Biol* 54:25–31. <https://doi.org/10.1016/j.ejsob.2012.11.001>
- Cseresnyés I, Sztár K, Rajkai K, Füzy A, Mikó P, Kovács R, Takács T (2018) Application of electrical capacitance method for prediction of plant root mass and activity in field-grown crops. *Front Plant Sci* 9:93. <https://doi.org/10.3389/fpls.2018.00093>
- Dalton FN (1995) In-situ root extent measurements by electrical capacitance methods. *Plant Soil* 173:157–165. <https://doi.org/10.1007/BF00155527>
- Dietrich RC, Bengough AG, Jones HG, White PJ (2012) A new physical interpretation of plant root capacitance. *J Exp Bot* 63:6149–6159. <https://doi.org/10.1093/jxb/ers264>
- Dietrich RC, Bengough AG, Jones HG, White PJ (2013) Can root electrical capacitance be used to predict root mass in soil? *Ann Bot* 112:457–464. <https://doi.org/10.1093/aob/mct044>
- Ellis T, Murray W, Kavalieris L (2013a) Electrical capacitance of bean (*Vicia faba*) root systems was related to tissue density—a test for the Dalton model. *Plant Soil* 366:575–584. <https://doi.org/10.1007/s11104-012-1424-z>
- Ellis T, Murray W, Paul K, Kavalieris L, Brophy J, Williams C, Maass M (2013b) Electrical capacitance as a rapid non-invasive indicator of root length. *Tree Physiol* 33:3–17. <https://doi.org/10.1093/treephys/tps115>
- Gerardo R, Gutierrez Boem FH, Fernández MC (2013) Severe phosphorus stress affects sunflower and maize but not soybean root to shoot allometry. *Agron J* 105:1283–1288. <https://doi.org/10.2134/agronj2013.0123>
- Hilhorst MA (1998) Dielectric characterisation of soil. Doctoral Thesis, Wageningen Agricultural University, The Netherlands
- Kendall WA, Pederson GA, Hill RR (1982) Root size estimates of red clover and alfalfa based on electrical capacitance and root diameter measurements. *Grass Forage Sci* 37:253–256. <https://doi.org/10.1111/j.1365-2494.1982.tb01604.x>
- Kormanek M, Głęb T, Klimek-Kopyra A (2016) Modification of the tree root electrical capacitance method under laboratory conditions. *Tree Physiol* 36:121–127. <https://doi.org/10.1093/treephys/tpv088>
- Oliveira MRG, van Noordwijk M, Gaze SR, Brouwer G, Bona S, Mosca G, Hairiah K (2000) Auger sampling, ingrowth cores and pinboard methods. In: Smit AL, Bengough AG, Engels C, van Noordwijk M, Pellerin S, van de Geijn SC (eds) *Root methods: a handbook*. Springer, Berlin, pp 175–210
- Ozier-Lafontaine H, Bajazet T (2005) Analysis of root growth by impedance spectroscopy (EIS). *Plant Soil* 277:299–313. <https://doi.org/10.1007/s11104-005-7531-3>
- Postic F, Doussan C (2016) Benchmarking electrical methods for rapid estimation of root biomass. *Plant Methods* 12:33. <https://doi.org/10.1186/s13007-016-0133-7>
- Psarras G, Merwin IA (2000) Water stress affects rhizosphere respiration rates and root morphology of young ‘Mutsu’ apple trees on M-9 and MM.111 rootstocks. *J Am Soc Hortic Sci* 125:588–595
- Rajkai K, Végh RK, Nacsa T (2005) Electrical capacitance of roots in relation to plant electrodes, measuring frequency and root media. *Acta Agron Hung* 53:197–210. <https://doi.org/10.1556/AAgr.53.2005.2.8>
- Svačina P, Středa T, Chloupek O (2014) Uncommon selection by root system size increases barley yield. *Agron Sustain Dev* 34:545–551. <https://doi.org/10.1007/s13593-013-0160-y>
- Urban J, Bequet R, Mainiero R (2011) Assessing the applicability of the earth impedance method for in situ studies of tree root systems. *J Exp Bot* 62:1857–1869. <https://doi.org/10.1093/jxb/erq370>
- Vamerali T, Bandiera M, Coletto L, Zanetti F, Dickinson NM, Mosca G (2009) Phytoremediation trials on metal- and arsenic-contaminated pyrite wastes (Torvicosa, Italy). *Environ Pollut* 157:887–894. <https://doi.org/10.1016/j.envpol.2008.11.003>
- Weigand M, Kemna A (2019) Imaging and functional characterization of crop root systems using spectroscopic electrical impedance measurements. *Plant Soil* 435:201–224. <https://doi.org/10.1007/s11104-018-3867-3>
- Wu W, Duncan RW, Ma B-L (2017) Quantification of canola root morphological traits under heat and drought stresses with electrical measurements. *Plant Soil* 415:229–244. <https://doi.org/10.1007/s11104-016-3155-z>

Publisher's Note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.