

# International Journal of Research in AYUSH and Pharmaceutical Sciences


## Letter to Editor

### PICRIC ACID; AN ALTERNATIVE SPECTROPHOTOMETRIC REAGENT FOR ESTIMATION OF EDTA SALTS


There are many reagents and techniques by which EDTA can be estimated in water as well as in Raw material. We standardized more quick and easy method for EDTA estimation. It has been reported that PA forms charge transfer complex with compounds. The stability of charge transfer complex depends upon the nature of compound, but mostly these are unstable. PA and EDTA complex show absorbance maximum at 450nm. The aim of our present study was to standardize an alternative method to estimate EDTA in water sample and in pure form by using the complex absorbance property at 450 nm. The 8,12,18,20 and 24 ppm concentration levels were prepared from standard stock solution of EDTA. PA of concentration 0.5gm in 100ml was prepared in chloroform. In each flask 1ml of PA was added and make up the volume of each flask with acetonitrile. The Standard plot was prepared by using EDTA -SALT and recorded Optical density at 450nm (Fig 01). The concentration level of PA was always high then EDTA-SALT in reaction mixture. However excess PA did not interfere at this wave length (450nm).




The EDTA alone Did not show any absorbance in the range of 220 – 400 nm Fig 02 (a, b), but EDTA-PA complex showed an absorbance maximum at 450nm Fig-022(c). We also checked the necessary optimization factors time, temperature and solvent effect, which are directly or indirectly effect the stability and formation of EDTA-PA complex. Stability study of complex with time, and Solvent Effect at ambient temperature are represented respectively in Fig 03, and Fig 04.

*Fig-01;- Standard Curve of EDTA (Salt)*



*Fig 02(a,b);- Spectrogram of PA (346nm) and EDTA respectively*



**Fig-02(c):- Spectrogram of EDTA-PA complex****Fig 03;- Breaking of(Edta -Picric acid) complex with time****Fig-4;- Suitable Solvent for complex formation**

ACN= Acetonitrile  
 MeOH = Methanol  
 DMF = Dimethylformamide

From these results, we conclude that this alternate method is more simple than earlier methods. The study suggests the possibility of estimation of EDTA in raw material and in water samples.

**Acknowledgement:** The Authors are cordially thankful to Ophth Pharma and its management.

**From:**

**Abdul Mannan, Akhterjamal Khan, Mahwash khan, Ghulam Abbas, Zohaib Ahmed, Syed Roohullah**  
 Ophth Pharma (PVT) LTD Plot 241, sect 24, Korangi industrial Area Karachi Pakistan.

**Address for correspondence:**

**Abdul Mannan**

**Email: abdulmunan17@gmail.com**

## REFERENCES

1. Colorimetric Determination of EDTA,, G. W. F .Brady, J. R. Gwilt].Chemical technology and Biotechnology Feb 1962 vol 12(2) page 79-82.
2. Quantitative determination of Disodium edetate in ophthalmic and contact lens care solutions by reversed-phase high-performance liquid chromatography, Lon.Hall, Lioyd Takahashi, J.Pharma.Sce 1988, March vol 77(3) page 247-250.
3. A validated Reverse Phase HPLC Method for the Determination of Disodium EDTA in Meropenem Drug Substance with UV- Detection using precolumn Derivatization Tchnique, bhavil Narola, A.Ssingh, M.Mitra, P.RSanthakumar and T.G.Chandrashekhar]. Analytical Chemistry Insights 2011 Feb(6)7-14.
4. High-performance liquid chromatographic, stability indicating assay for disodium EDTA in ophthalmic preparations, Bauer,J.Heathcote, and D.Kroghs,J.Chromatography 1986, Nov 21 vol 369, page 422-425.
5. Quantitative determination of Edta and its behaviour in radioactive waste solutions using HPLC,, Martina Unger, Elisabeth Mainka, werner. Kong,J.Chromatographicanalysis, Jan 1987, volume 329, (1) pp 50-54.
6. Determination of ethylenediaminetetraacetic acid in injection form by ion-pair chromatography,, C.Pistos, M.Parisssi, M.Poulou,J.Pharmaceutical and Biomedical ANALYSIS JUN 2002, VOL 28(6)PP 1073-1079.
7. Determination of disodium edetate dihydrate in streptomycin by atomic absorption spectrophotometry, July 1974, vol 63(7) page 1131-1138.
8. New EDTA determination method based on ion chromatography with suppressed conduct metric detection, *Journal of Radioanalytical and Nuclear Chemistry*, 302, 1, (161), (2014).
9. Determination of ethylenediaminetetra-acetic acid in aqueous rinses of detergent-washed rubber stoppers of pharmaceutical vials using solid-phase extraction and capillary gas chromatography, Anastasios, Voulgaropoulos, Nicolaos. Tzivanakis J.Pharmaceutical and Biomedical Analysis 1987, Volume 5(7)PP. 687-694.
10. Kinetic-photometric determination of EDTA, zinc and bismuth by interchange reactions of  $>C=N-$  groups, Journal: *Analyst*, 1984 vol,109,pp 1147-1150
11. EDTA; An Alternative spectrophotometric Reagent for Iron Estimation,, J.Analytical Toxicology March/April 1997 vol.21 pp 172-173.