



**University of
Zurich** UZH

**Zurich Open Repository and
Archive**

University of Zurich
Main Library
Strickhofstrasse 39
CH-8057 Zurich
www.zora.uzh.ch

Year: 2014

Prevention of peritendinous adhesions using an electrospun DegraPol polymer tube: a histological, ultrasonographic, and biomechanical study in rabbits

Meier Bürgisser, Gabriella; Calcagni, Maurizio; Müller, Angela; Bonavoglia, Eliana; Fessel, Gion; Snedeker, Jess G; Giovanoli, Pietro; Buschmann, Johanna

Abstract: **PURPOSE** One of the great challenges in surgical tendon rupture repair is to minimize peritendinous adhesions. In order to reduce adhesion formation, a physical barrier was applied to a sutured rabbit Achilles tendon, with two different immobilization protocols used postoperatively. **METHODS** Thirty New Zealand white rabbits received a laceration on the Achilles tendon, sutured with a 4-strand Becker suture, and half of the rabbits got a DegraPol tube at the repair site. While fifteen rabbits had their treated hind leg in a 180° stretched position during 6 weeks (adhesion provoking immobilization), the other fifteen rabbits were recasted with a 150° position after 3 weeks (adhesion inhibiting immobilization). Adhesion extent was analysed macroscopically, via ultrasound and histology. Inflammation was determined histologically. Biomechanical properties were analysed. **RESULTS** Application of a DegraPol tube reduced adhesion formation by approximately 20%—independently of the immobilization protocol. Biomechanical properties of extracted specimen were not affected by the tube application. There was no serious inflammatory reaction towards the implant material. **CONCLUSIONS** Implantation of a DegraPol tube tightly set around a sutured tendon acts as a beneficial physical barrier and prevents adhesion formation significantly—without affecting the tendon healing process.

DOI: 10.1155/2014/656240

Posted at the Zurich Open Repository and Archive, University of Zurich

ZORA URL: <http://doi.org/10.5167/uzh-100053>

Published Version

Originally published at:

Meier Bürgisser, Gabriella; Calcagni, Maurizio; Müller, Angela; Bonavoglia, Eliana; Fessel, Gion; Snedeker, Jess G; Giovanoli, Pietro; Buschmann, Johanna (2014). Prevention of peritendinous adhesions using an electrospun DegraPol polymer tube: a histological, ultrasonographic, and biomechanical study in rabbits. *BioMed Research International*, 2014:656240-656244. DOI: 10.1155/2014/656240

Research Article

Prevention of Peritendinous Adhesions Using an Electrospun DegraPol Polymer Tube: A Histological, Ultrasonographic, and Biomechanical Study in Rabbits

Gabriella Meier Bürgisser,¹ Maurizio Calcagni,¹ Angela Müller,¹ Eliana Bonavoglia,² Gion Fessel,³ Jess G. Snedeker,³ Pietro Giovanoli,¹ and Johanna Buschmann¹

¹ Division of Plastic Surgery and Hand Surgery, University Hospital Zurich, ZKF, Sternwartstrasse 14, 8091 Zurich, Switzerland

² *ab medica*, Via Nerviano 31, 20020 Lainate, Milan, Italy

³ Department of Orthopedics, Uniklinik Balgrist, Forchstrasse 340, 8008 Zurich, Switzerland

Correspondence should be addressed to Johanna Buschmann; johanna.buschmann@usz.ch

Received 4 February 2014; Revised 23 May 2014; Accepted 2 June 2014; Published 2 July 2014

Academic Editor: Ulrich Kneser

Copyright © 2014 Gabriella Meier Bürgisser et al. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Purpose. One of the great challenges in surgical tendon rupture repair is to minimize peritendinous adhesions. In order to reduce adhesion formation, a physical barrier was applied to a sutured rabbit Achilles tendon, with two different immobilization protocols used postoperatively. **Methods.** Thirty New Zealand white rabbits received a laceration on the Achilles tendon, sutured with a 4-strand Becker suture, and half of the rabbits got a DegraPol tube at the repair site. While fifteen rabbits had their treated hind leg in a 180° stretched position during 6 weeks (adhesion provoking immobilization), the other fifteen rabbits were recasted with a 150° position after 3 weeks (adhesion inhibiting immobilization). Adhesion extent was analysed macroscopically, via ultrasound and histology. Inflammation was determined histologically. Biomechanical properties were analysed. **Results.** Application of a DegraPol tube reduced adhesion formation by approximately 20%—independently of the immobilization protocol. Biomechanical properties of extracted specimen were not affected by the tube application. There was no serious inflammatory reaction towards the implant material. **Conclusions.** Implantation of a DegraPol tube tightly set around a sutured tendon acts as a beneficial physical barrier and prevents adhesion formation significantly—without affecting the tendon healing process.

1. Introduction

Tendon repair is a field in surgery where improvements are still welcome. Complications such as scar formation and adhesion in the tendon sheath in the early healing phase up to 6 weeks do occur in 7 to 15% of the cases, which leads to increased work disability and costs [1]. In particular, there is a need to reduce peritendinous adhesion formation because this is reported to be the most common complication following flexor tendon repair in the hand [2]—but also in the Achilles tendons, functional complications caused by adhesion are reported [3]. Hence, different antiadhesion strategies to improve the gliding capability of the regenerating tendon are being developed [4]. First of all, the postoperative treatment has to be optimized with respect to its impact on

the adhesion extent as it is well known that for example in Achilles tendon regeneration the ankle angle plays a crucial role [5]. Secondly, physical antiadhesion barriers consisting of biomaterials such as fibronectin [6], collagen [3], or silk [7] have been reported to show positive effects with respect to reducing adhesion formation. Finally, antiadhesive agents were beneficially administered such as 5-fluorouracil [8], hyaluronic acid, and ADCON-T/N, being a gelatine-polyglycan ester compound [9] or anti-inflammatory drugs like ibuprofen [10].

Hence, concepts of adhesion prevention include postoperative treatment strategies and pharmacological as well as mechanical agents. An effective mechanical barrier should be biocompatible, biodegradable to achieve tissue integration and prevent bulging. Furthermore, it should be easy in

surgical handling during implantation and cause only a low inflammatory response. The polymer DegraPol which is a copolymer of poly-hydroxy-butyrate and ϵ -caprolactone meets all these requirements [11]. It has been shown to be an ideal scaffold material for tenocytes [12]. In recent studies, we checked the cellular response toward classic DegraPol in Achilles tendon repair in a rabbit model (mimicking the biomechanical properties of human flexor tendons of the hand) and found no adversary effects [13]. Subsequently, we synthesized a new more elastic DegraPol polymer in order to facilitate surgical handling [14].

In the study presented here, an electrospun DegraPol tube as an antiadhesive was investigated. Therefore, the adhesion extent of lacerated rabbit Achilles tendons 6 weeks after surgery either with or without the application of a DegraPol tube was determined by macroscopic evaluation, by dynamic ultrasound, and by histology. The rabbit Achilles tendon is surrounded by a leaflet of fascia and it has similar biomechanical properties as the human flexor tendons of the hand [15, 16]. In order to modulate the adhesion in the rabbit Achilles tendon model, we used two different postoperative immobilization protocols: the ankle angle of the cast was changed after 3 weeks from 180° to 150° (adhesion inhibiting immobilization protocol) or the cast had an ankle angle of 180° for the full 6-week period (adhesion provoking immobilization protocol). Biomechanical properties including load until failure, cross sectional area (CSA), stiffness, E-modulus, and failure stress were determined. Additionally, inflammatory reaction was determined based on macrophages and lymphocytes. Hence, the two hypotheses were that (i) the application of a DegraPol tube around the repaired tendon would reduce adhesion formation without an inflammatory reaction in both postoperative treatment models and (ii) the biomechanical properties of the regenerated tendons would not be adversely affected by the implantation of a DegraPol tube.

2. Materials and Methods

2.1. Block Copolymer DegraPol. A biodegradable polyester urethane polymer (trade name DegraPol) with poly-hydroxy-butyrate as a crystalline segment and ϵ -caprolactone as a soft segment was synthesized according to an adapted protocol [14]. Briefly, for the synthesis of the block copolymer, 25 wt% of poly(3-(R)-hydroxybutyrate)-co-(ϵ -caprolactone)-diol ($M_n = 2824 \text{ g mol}^{-1}$) and 75 wt% with $M_n = 1000 \text{ g mol}^{-1}$ poly(ϵ -caprolactone)-diol-co-glycolide (15 mol% glycolide 85 mol% ϵ -caprolactone) were dissolved in 1,4-dioxane and dried by heating and refluxing the solvent over molecular sieves (pore size 0.4 nm). The reaction mixture was cooled to 83°C before the stoichiometric amount, with respect to the two diols, of 2,2,4-trimethylhexane-diisocyanate (TMDI) was added. After about one day of reaction, three portions of dibutyltin dilaurate (20 ppm) were added within 1 day in order to reach molecular weight of 110 kDa. The polymer was precipitated in dry ice cooled hexane isomers and subsequently purified via dissolution in chloroform and filtration over a silicagel 60 (Fluka) column. A second precipitation in cooled ethanol ended the process.

2.2. Scaffold Fabrication and Characteristics. The electrospinning setup was assembled in-house and consisted of a syringe pump (Racel Scientific Instruments Inc., USA), a spinning head consisting of a central stainless steel tube (1 mm inner diameter and 0.3 mm wall thickness, Angst & Pfister AG, Switzerland), a cylindrical rotating aluminum mandrel for fiber collection (length: 100 mm, diameter: 4 mm), and a DC high voltage supply (Glassman High Voltage Inc., USA). A 25 wt% solution of the DegraPol (DP) in chloroform (Fluka, puriss., Switzerland) was prepared by dissolving the polymer under stirring overnight. Electrospinning voltage (15 kV) was applied with a high voltage supply between a needle and the rotating cylindrical collector (20 cm apart from each other). As-spun tubes (with randomly oriented fibres) were removed from the target by slightly swelling them with ethanol (Fluka, puriss., Switzerland) and then dried under vacuum at room temperature. In Figure 1(a), the high elasticity of such a tube is shown. The overall mesh porosity P was calculated according to $P = (1 - \rho/\rho_0) \times 100$ (%), where ρ denotes the density of the electrospun mesh and ρ_0 is the bulk density of the electrospun mesh determined gravimetrically using weights of precisely cut samples of defined area and thickness; P was $75.2 \pm 0.4\%$. The wall thickness was $357.0 \pm 17.7 \mu\text{m}$ as determined by means of scanning electron microscopy (SEM, FEI, Nova NanoSEM 450). The fibres of the electrospun tube were randomly oriented. The tubes had an inner diameter of 4 mm and were cut into pieces of 1 cm in length. Their degradation half-lives were determined to be approximately 3 months.

2.3. Animals. For the *in vivo* study, thirty female New Zealand white rabbits aged 12 to 16 weeks were used (Charles River, Research Models and Services, Germany). They were specific pathogen free (SPF). All animals were housed in pairs in two interconnected cages, each of them with a bottom area of 70 cm \times 70 cm and a height of 62 cm (Indulab, Switzerland). The animals were maintained under controlled conditions: temperature $22 \pm 1^\circ\text{C}$, 45% relative humidity, 15 air changes per hour, and a light/dark rhythm of 12 hours. The rabbits had free access to water (automatic water supply), autoclaved hay, and straw *ad libitum* as well as to standard pellet diet (Kliba Nafag, Nr. 3410, Provimi Kliba AG, Switzerland). Ethical approval was obtained for the experiments from the veterinary office of Zurich, Switzerland (reference numbers 92/2009 and 193/2012). Prior to surgery, all animals were acclimatized to their environment for two weeks.

2.4. Achilles Tendon Repair. Before implantation, the DP tubes were sterilized with ethylene oxide at 38°C. The rabbits received premedication with 65 mg/kg body weight Ketamine and 4 mg/kg Xylazine. A venous catheter was inserted in the marginal ear vein. The rabbits were intubated with Propofol i.v. 0.6 mg/kg–1.3 mg/kg. Anaesthesia was maintained with 1-2% isoflurane. In order to ensure systemic analgesia during the time of operation, 0.2-0.3 mg/kg body weight Butorphanol (Dr. E. Graeb AG, Berne, Switzerland) was applied preoperatively. The hind legs were cleaned with iodine. The Achilles tendon exposure was obtained through




FIGURE 1: (a) Very elastic DegraPol tube used for mechanical blockage of regenerating tendon tissue and surrounding tissue, (b) and (c) implantation of tube in addition to a conventional 4-strand Becker suture, (d) New Zealand white rabbit with a cast having an ankle angle of 180° (fully stretched), (e) extracted tendon specimen with 4-strand suture and (f) extracted tendon specimen with DP tube application in addition to 4-strand suture.

a paratendineal incision of cutis, subcutis, and fascia. The medial and lateral *M. gastrocnemius* of the Achilles tendon complex was then sliced perpendicularly to the length of the tendon 2.0 cm above the *calcaneus* and the two tendon stumps were sutured (4-strand Becker suture) using a USP 4.0 polypropylene thread. In case of DP tube application, one of the two tendon stumps was pulled through the DP tube before suturing (Figures 1(b) and 1(c)). Subsequently the wound was closed with a running suture (using a USP 6.0 polypropylene fiber) of the fascia and interrupted skin. Immediately after surgery, a Duragesic matrix patch (Janssen-Cilag AG, Switzerland) was applied with 4.2 mg Fentanyl per patch in order to provide analgesia for about 72 hours with 25 $\mu\text{g/h}$ Fentanyl. Postoperative treatment included a cast having an angle of 180° at the ankle for a total of 12 rabbits (6 with and 6 without tube; adhesion provoking immobilization) (Figure 1(d)). The cast was well padded and it was changed after 3 weeks. For the other 12 rabbits, the cast was also changed after 3 weeks; however, a cast with a smaller angle of 150° was applied (adhesion inhibiting immobilization). Great attention was paid to make the casts not too tight so that it was tolerated well by the rabbits (they did not bite the cast). Six weeks after surgery, the rabbits were euthanized in deep anaesthesia (100 mg/kg Ketamine and 4 mg/kg Xylazine) with 80 mg/kg Pentobarbital (Esconarkon *ad us. vet.*, Switzerland).

2.5. Treatment Groups. The thirty rabbits were randomly distributed into eight groups with $n = 3$ or 6 (Figure 2). All were operated on one hind leg [17]. The tendons of twenty-four rabbits were analysed histologically (eighteen of them for adhesion scoring (twelve in the 180°/150° and six in the 180°/180° group), six for inflammation analysis, all in the 180°/180° group), with groups classified by the application of a DP tube in addition to a 4-strand Becker suture (application or no application) and by the casting protocol (ankle angle change from 180° to 150° (180°/150° group) or 180° for the full 6-week period with a cast change after 3 weeks (180°/180° group)). The tendons of six rabbits were analysed biomechanically, all receiving a 180°/150° immobilization, including three animals with DP and three without. The counter legs that were not treated (NT) served as control.

2.6. Rating during Tendon Extraction. Achilles tendons including *flexor digitorum superficialis* and surrounding tissue were extracted. During extraction, the extent of adhesion formation was macroscopically evaluated (carefully paying attention to not move the tendon in the sheath) and scored semiquantitatively with 0 = good gliding (no adhesion), 1 = middle gliding (some adhesion) and 2 = no gliding (maximum adhesion). The surgeon was blinded to the treatment.

2.7. Dynamic Ultrasound. The Achilles tendons of all animals were imaged with high-frequency ultrasound before surgery and at the end point of the experiments before explanting the tendons. Ultrasound imaging was performed with an ultrasound unit (iU22 Ultrasound System, Philips Healthcare, Switzerland) with a linear high-frequency hockey-stick probe of 17.5 MHz (L17-5io Broadband Compact Linear Array

Transducer, Philips Healthcare, Switzerland). The examination protocol consisted of a dynamic imaging of the gliding Achilles tendon by moving the *flexor digitorum superficialis* of the rabbit paw. All measurements were done freehand. The films of the gliding tendons were scored with 0 = good gliding (no adhesion), 1 = middle gliding (some adhesion) and 2 = no gliding (maximum adhesion) (see Supplementary Videos 1 and 2 (see Supplementary Material available online at <http://dx.doi.org/10.1155/2014/656240>)). Colour gain adjustment was calibrated on the counter leg that was not treated (healthy side).

2.8. Quantification of Adhesion Extent and Inflammation by Histology. After extraction, the Achilles tendon specimens were immediately frozen at -20°C. After being thawed to room temperature, they were dehydrated, paraffin-embedded, and sectioned into 5 μm thick slices, which were cross sections in the DP tube region (perpendicular to the Achilles tendon). After being deparaffinized with xylene and rehydrating the sections (descending gradient of ethanol), they were differently stained: Picrosirius red and H&E according to commonly established procedures.

Picrosirius red stained sections were used to quantify the adhesion extent at 8x magnification (Leica EZ4D microscope, Switzerland). Here, adhesion formation was quantified in five subsequent cross sections separated by 2.0 mm using a method by Tan et al. [18]. Analysis was done in a blinded fashion. The percentage of adhesion was calculated by the length of the contact region of the tendon under view with the surrounding tissue divided by the total perimeter. The length of the contact region and the whole perimeter were determined using *synedra View 3* software (version 3.1.0.3.).

Inflammation zones based on macrophages and on lymphocytes were evaluated in the H&E-stained sections at 100x and 200x magnification based on their morphology (5 FOV of each object, semiquantitative analysis); scores: 1 = healthy tendon tissue, 2 = some fine films of inflammation, 3 = more fine films of inflammation, 4 = few zones (areas) of inflammation, and 5 = many zones (areas) of inflammation.

2.9. Biomechanical Tests. Before measurement at room temperature (21°C), the tendons were thawed overnight at 4°C. All tendons were harvested from the hind legs including the muscle and the *calcaneus*. On the muscle side, the samples were mounted in serrated clamps after being wrapped in two pieces of cloth to reduce slippage [19], and on the bone side, a device fixing the *calcaneus* in a rectangular position to the tendon was used. All samples were tested in uniaxial tension to failure at 1 mm min⁻¹ speed on a Zwick 1456 tensile testing machine (1 kN load-cell, 0.1 μm extensometer, TestXpert 10, Germany) with preconditioning (10 cycles to 10 N). The samples were sprayed with phosphate buffered saline during measurement in order to prevent drying. Load until failure (N) was determined as the maximum load measured.

The CSA was determined 2.0 cm above the *calcaneus* by a custom designed linear laser scanner adapted by Vergari et al. with $n = 6$ per specimen before tensile testing [20, 21]. The load until failure (N) was divided by the thus-determined




FIGURE 2: Experimental design depicting number of rabbits used for the corresponding surgical and postoperative treatments. Accordingly, the three different analysis approaches are given (adhesion = determination of adhesion extent by macroscopic, dynamic ultrasound and histological analyses; biomechanics = determination of load until failure, cross-sectional area, stiffness, failure stress and elastic modulus; inflammation = determination of macrophage and lymphocyte density). Moreover, links to figure numbers of the respective results are given.

CSA at the repair site (m²) resulting in the failure stress at the repair site (Pa). The elastic modulus (E-Modulus; Pa) was determined in the stress-strain curves. Moreover, the stiffness (N/mm) was determined in the force-elongation curves.

2.10. *Statistical Analysis.* Histomorphometric and biomechanical data were analysed with StatView 5.0.1. One-way analysis of variance (one-way ANOVA) was conducted. Pairwise comparison probabilities (*P*) were calculated using Fisher’s PLSD. *P* values <0.05 were considered significant. Values were expressed as means ± standard deviations.

3. Results

3.1. *Adhesion Provoking and Adhesion Inhibiting Immobilization.* The two different postoperative casting protocols had a significant impact on the adhesion extent. A change of the ankle angle from 180° to 150° after 3 weeks reduced the adhesion extent significantly from 82.7 ± 9.7% to 31.9 ± 9.8% (*P* < 0.05) when the tendon laceration was repaired by a conventional 4-strand Becker suture (Supplementary Table 1). Accordingly, in the tube-treated group, the change of the ankle angle after 3 weeks leads to a reduction in the adhesion extent from 69.3 ± 9.8% to 15.6 ± 10.9% (*P* < 0.05).

3.2. *Impact of DegraPol Tube on Adhesion.* The extent of adhesion formation was independently evaluated by two different persons using a semiquantitative scoring system by macroscopic observation during extraction (MC) and by the percentage of contact regions of the Achilles tendon to its

surrounding tissue calculated from five subsequent histological cross-sections (GMB) (Figure 3). There was a statistically significant difference in adhesion formation between tube-treated specimen and specimen without tube. The implantation of a DP tube reduced the contact area to the surrounding tissue significantly from 31.9 ± 9.8% to 15.6 ± 10.9% in adhesion inhibiting postoperative model (*P* < 0.05), while it reduced the contact area from 82.7 ± 9.7% to 69.3 ± 9.8% in the adhesion provoking postoperative model (*P* < 0.05).

3.3. *Biomechanics.* Figure 4 shows the biomechanical results for rabbits with the 180°/150° casting regime (adhesion inhibiting model). There were no statistical significant differences between the DP treated and the mere 4-strand treated specimen (neither for load until failure (*P* = 0.81) nor for CSA (*P* = 0.13), for stiffness (0.44), for E-Modulus (0.20), and for failure stress (0.72)), indicating no adverse effect of the polymer tube on biomechanical properties of the healing tendon.

3.4. *Correlation of Macroscopic and Histological Evaluation.* Semiquantitative adhesion scores gathered immediately after tendon extraction were correlated to the quantitative analysis of adhesion determined in histological sections (Figure 5(a)). A positive correlation was found (*r*² = 0.86).

3.5. *Correlation of Dynamic Ultrasound and Histological Evaluation.* Semiquantitative adhesion scores from dynamic ultrasound were correlated with the quantitative analysis of




FIGURE 3: Picrosirius red stained 5 μm histological sections of the three groups considered at 8x magnification. *Key:* 4-strand + DP = 4-strand Becker suture and application of DegraPol tube; 4-strand = 4-strand Becker suture; no treatment = control group without any intervention; DP = DegraPol, AT = Achilles tendon; FDS = *Flexor digitorum superficialis*. The black lines in the lower figures denote the whole circumference, while the green lines depict the adhesion formation.

adhesion determined in histological sections (Figure 5(b)). A positive correlation was found ($r^2 = 0.79$).

3.6. Inflammation. Inflammation was semiquantitatively scored and found to be not significantly different whether a DP tube was implanted or not for macrophages (Figure 6(a)); only lymphocyte analysis showed a significant higher cell density for the DP treated specimen ($P = 0.05$) (Figure 6(b)).

4. Discussion

The most common problem arising during tendon healing is the formation of fibrous adhesions to the surrounding tissue [2]. The etiology is based on the destruction of a cell-retentive layer (basement membrane) on the tendon surface. At the site of the tendon rupture, cells exit from the damaged tendons where the basement membrane is damaged [22]. Thus, adhesions are mostly evoked by scarring between two damaged surfaces [23]. Many studies dealing with the application of lubricants have been published [18, 24–26]—the outcome being controversial. For example, while some authors claim a beneficial effect of hyaluronic acid (based on adhesion scoring and cell density) [27], others found adhesion to be unaffected [28]. Physical barriers such as collagen membranes [3], PGA membranes [29], PLGA membranes [10], or spontaneously forming hydrogels [30] positively affected adhesion formation, although their use in the clinical set up is still very scarce due to limited

tissue integration, inflammatory reactions, cell stress caused by lowered pH, or decreasing reduced final strength of the regenerated tendon.

Different rehabilitation protocols are reported to have an impact on the clinical outcome of tendon rupture repairs. Early active motion protocols go along with lower complication rates [31]. In an experimental study by Pihlajamäki et al., the cellular composition at the repair site 6 weeks after surgery of the healing rabbit Achilles tendon was similar whether a cast was applied or not [29]. In contrast, other studies denote the importance of the specific postoperative casting regime [32]. Besides the length of the application period of the cast, the ankle angle is one important parameter [5].

With the intention to drastically modulate the adhesion extent, we used two different postoperative immobilization protocols. The ankle angle of the cast was either changed or not 3 weeks after surgery. Early active or passive motion was not included—in contrast to other studies [33, 34], because a defined and standardized early motion protocol is very difficult to realize in rabbits. Therefore, a cast was applied for a total time period of 6 weeks (end of experiments). While one group received the whole 6 weeks a 180° cast (renewed after 3 weeks, provoking adhesion formation), the other group had a 180° cast for the first 3-week period and then a new 150° cast for the second 3-week period (inhibiting adhesion formation). A totally stretched position was chosen because the Achilles tendon is perfectly relaxed and the rabbit's foot is




FIGURE 4: Load until failure (a), CSA (b), stiffness (c), failure stress (d), and E-modulus (e) determined for the tendons of rabbits with a 180°/150° casting regime. Key: NT = no treatment (contralateral legs); DP = 4-strand Becker suture and a DegraPol tube; 4-strand = 4-strand Becker suture. *P* values <0.05 are marked by asterisk (*). Sample size was *n* = 3 for DP and 4-strand groups, respectively, and *n* = 6 for NT group.

in a position which is usually taken during rest. The adhesion inhibiting protocol including a change in the ankle angle led to a clear reduction of adhesion by over 50%, not only because some initial fibrous adhesion were decreased by the intervention of a positional change among the involved tissue

layers, but also by having the angle a little bit closer to angles occurring during jumping [35].

Mechanical blockage by an electrospun DegraPol polymer tube 6 weeks after surgery was tested in both postoperative treatment models and the adhesion provoking




FIGURE 5: Adhesion scores determined by semiquantitative scoring on observation during tendons extractions (a) and by dynamic ultrasound (b) correlated well with the percentage of adhesions to the surrounding tissue ($r^2 = 0.86$ and 0.79 , resp.). Pairwise comparison probabilities (P) were calculated using Fisher’s PLSD and the Bonferroni post hoc test. All P values are given in Supplementary Table 1. Key: DP = DegraPol tube application; no DP = mere 4-strand suture; NT = no treatment, control; empty circles = immobilization protocol where the ankle angle of the cast was changed from 180° to 150° after 3 weeks; black circles = immobilization protocol where the ankle angle of the cast was 180° during 6 weeks (cast renewed after 3 weeks). Sample size was $n = 3$ for $180^\circ/180^\circ$, $n = 6$ for $180^\circ/150^\circ$ group, and $n = 18$ for NT group.




FIGURE 6: Semiquantitative evaluations of macrophages (a) and lymphocytes (b) in the tendon tissue. Scores: 1 = healthy tendon tissue, 2 = some fine films of inflammation, 3 = more fine films of inflammation, 4 = few zones (areas) of inflammation, 5 = many zones (areas) of inflammation. Sample size was $n = 3$ for all groups; except for NT group ($n = 6$).

as well as the adhesion inhibiting model. We had tested DegraPol before with respect to its cell response *in vivo* 12 weeks after operation [13]. The polymer DegraPol is a biocompatible and totally biodegradable block copolymer. With its hard and soft segments that can be varied user-defined in weight ratios during production, it allows a wide range of degradation half-lives, stabilities, and biomechanical characteristics. Originally, it was developed as a scaffold material for bone reconstruction [36]. However, DegraPol

was also shown to be a good scaffold for tenocytes [37]. Even after having changed its synthesis protocol in order to make the polymer more elastic to facilitate the surgeon’s pulling it over the repair site, no adverse cellular effects 12 weeks after operation were observed upon implantation [14]. The very high elasticity of the electrospun DegraPol mesh denoted by an elongation at break of $544 \pm 68\%$ is a big advantage over other mechanical barrier options—even the best ones with respect to this property—such as electrospun PLGA having

an elongation at break of “only” $324 \pm 78\%$ [38]. Moreover, degrading DegraPol is pH-neutral while other polymers used as implant materials degrade with lowering the pH (like PLGA).

In this study, the quantitative determination of the adhesion extent by histological analysis clearly showed that the implantation of a DegraPol tube around the conventionally sutured tendon reduced the peritendinous adhesion significantly by around 20%—independent of the postoperative treatment mode. Moreover, this antiadhesive effect was clearly supported by our semiquantitative findings based on macroscopic inspection and evaluation and scoring of the gliding capacity by dynamic ultrasound, both correlating positively to the histological findings. Although other studies report much higher reduction of adhesion extent up to 88% by using, for example, an Fe^{3+} crosslinked carboxymethylcellulose barrier, such results are not directly comparable to our findings because the adhesion extent was determined at a different time point—already after 2 weeks [39], when peritendinous adhesion formation is not yet completed [40]. Importantly, the implantation of our electrospun DP tube did not lead to a decrease in failure strength nor in other biomechanical parameters such failure stress, tensile modulus, or stiffness—standing in contrast to other studies using, for example, cortisone as antiadhesive that leads to great problems caused by decreased biomechanical properties after implantation [41].

Though not statistically significantly different, the failure stress and the E-Modulus—two typical material properties [42]—of the DP tube specimen were approximately 30% lower compared to no-tube specimen; this is mainly caused by the approximately 30% larger CSA provoked by the DP tube that is not yet degraded fully after 6 weeks. In addition, inflammatory reaction to the implant material was only found to be marginally increased for lymphocytes, but not for macrophages which was the abundant subpopulation of inflammatory cells, and which might cause problems when materials that are less biocompatible such as cellulose [43] are used as physical barriers.

Although adhesion formation was analyzed by macroscopic evaluation, dynamic ultrasound, and quantitative histology, it might be favorable in future studies to quantify the gliding capability of the rabbit Achilles tendon by range-of-motion analysis [44]. Moreover, inflammatory reactions towards the implant material might be analyzed in more details using immunohistochemistry methods [45].

5. Conclusion

In summary, we could demonstrate that an electrospun DegraPol tube acts as an efficient physical barrier enabling a reduction of peritendinous adhesion without changing biomechanical characteristics and not affecting inflammation—and regardless of the postoperative treatment protocol. In both treatment models, we found a significant adhesion reduction of approximately 20%. The beneficial impact of this electrospun material could be clearly demonstrated for this purpose. Thus, peritendinous adhesion prevention by a DegraPol tube is a promising approach and may find its way

to clinical application after some further modifications will be undertaken; therefore, we plan to load such a tube with tendon growth factors inside in order to accelerate tendon healing, thus generating a device that is physically blocking adhesion formation outside, while simultaneously promoting the healing process inside.

Disclosure

All authors were fully involved in the study and the material within has not been and will not be submitted for publication elsewhere.

Conflict of Interests

As the coauthor Eliana Bonavoglia is an employee of *ab medica*, Italy, the company that holds the license for the polymer DegraPol, there is a potential conflict of interests. For all other coauthors, there is no conflict of interests.

Acknowledgments


The authors are very much indebted for financial support by the *Hartmann-Müller Foundation*, the *Wolferrmann-Nägeli Foundation*, and the *Fonds für Medizinische Forschung der Universität Zürich*. Flora Nicholls and Nikola Cesarovic are highly acknowledged for their help in animal care and in anaesthesia. Vroni Bäuerle is highly acknowledged for casting the rabbits' hind legs. The authors also thank Pia Fuchs for her perfect technical assistance in the preparation of the histological sections and Nico Wick for photography. The authors thank *ab medica*, Lainate, Italy, for providing DegraPol. The work was performed at the University Hospital Zurich, Zürich, Switzerland.

References

- [1] B. T. Jin, “Clinical outcomes associated with flexor tendon repair,” *Hand Clinics*, vol. 21, no. 2, pp. 199–210, 2005.
- [2] M. Jaibaji, “Advances in the biology of zone II flexor tendon healing and adhesion formation,” *Annals of Plastic Surgery*, vol. 45, no. 1, pp. 83–92, 2000.
- [3] H. Zhao, H. G. Guan, J. Gu et al., “Collagen membrane alleviates peritendinous adhesion in the rat Achilles tendon injury model,” *Chinese Medical Journal*, vol. 126, no. 4, pp. 729–733, 2013.
- [4] C. Brochhausen, V. H. Schmitt, T. K. Rajab et al., “Intraperitoneal adhesions—an ongoing challenge between biomedical engineering and the life sciences,” *Journal of Biomedical Materials Research A*, vol. 98, no. 1, pp. 143–156, 2011.
- [5] W. H. Wray, C. Regan, S. Patel, R. May, and S. G. Parekh, “Positional effects of the knee and ankle on the ends of acute Achilles tendon ruptures,” *Foot & Ankle Specialist*, vol. 2, no. 5, pp. 214–218, 2009.
- [6] O. A. Branford, R. A. Brown, D. A. McGrouther, A. O. Grobbelaar, and V. Mudera, “Shear-aggregated fibronectin with anti-adhesive properties,” *Journal of Tissue Engineering and Regenerative Medicine*, vol. 5, no. 1, pp. 20–31, 2011.

- [7] T. Ni, P. Senthil-Kumar, K. Dubbin et al., "A photoactivated nanofiber graft material for augmented Achilles tendon repair," *Lasers in Surgery and Medicine*, vol. 44, no. 8, pp. 645–652, 2012.
- [8] U. Khan, S. Kakar, A. Akali, G. Bentley, and D. A. McGrouther, "Modulation of the formation of adhesions during the healing of injured tendons," *The Journal of Bone and Joint Surgery*, vol. 82, no. 7, pp. 1054–1058, 2000.
- [9] L. McGonagle, M. D. Jones, D. Dowson, and P. S. Theobald, "The bio-tribological properties of anti-adhesive agents commonly used during tendon repair," *Journal of Orthopaedic Research*, vol. 30, no. 5, pp. 775–780, 2012.
- [10] S. Liu, C. M. Hu, F. F. Li, X. J. Li, W. G. Cui, and C. Y. Fan, "Prevention of peritendinous adhesions with electrospun ibuprofen-loaded poly(L-Lactic Acid)-polyethylene glycol fibrous membranes," *Tissue Engineering A*, vol. 19, no. 3-4, pp. 529–537, 2013.
- [11] J. A. Henry, K. Burugapalli, P. Neuenschwander, and A. Pandit, "Structural variants of biodegradable polyesterurethane in vivo evoke a cellular and angiogenic response that is dictated by architecture," *Acta Biomaterialia*, vol. 5, no. 1, pp. 29–42, 2009.
- [12] B. Saad, O. M. Keiser, M. Welti, G. K. Uhlschmid, P. Neuenschwander, and U. W. Suter, "Multiblock copolyesters as biomaterials: in vitro biocompatibility testing," *Journal of Materials Science: Materials in Medicine*, vol. 8, no. 8, pp. 497–505, 1997.
- [13] J. Buschmann, G. Meier-Bürgisser, E. Bonavoglia et al., "Cellular response of healing tissue to DegraPol tube implantation in rabbit Achilles tendon rupture repair: an in vivo histomorphometric study," *Journal of Tissue Engineering and Regenerative Medicine*, vol. 7, no. 5, pp. 413–420, 2013.
- [14] J. Buschmann, M. Calcagni, G. M. Bürgisser et al., "Synthesis, characterization and histomorphometric analysis of cellular response to a new elastic DegraPol polymer for rabbit Achilles tendon rupture repair," *Journal of Tissue Engineering and Regenerative Medicine*, 2012.
- [15] H. J. Goodman and J. Choueka, "Biomechanics of the flexor tendons," *Hand Clinics*, vol. 21, no. 2, pp. 129–149, 2005.
- [16] K. Nagasawa, M. Noguchi, K. Ikoma, and T. Kubo, "Static and dynamic biomechanical properties of the regenerating rabbit Achilles tendon," *Clinical Biomechanics*, vol. 23, no. 6, pp. 832–838, 2008.
- [17] J. Buschmann, A. Müller, F. Nicholls et al., "2D motion analysis of rabbits after Achilles tendon rupture repair and histological analysis of extracted tendons: can the number of animals be reduced by operating both hind legs simultaneously?" *Injury*, vol. 44, no. 10, pp. 1302–1308, 2013.
- [18] V. Tan, A. Nourbakhsh, J. Capo, J. A. Cottrell, M. Meyenhofer, and J. P. O'Connor, "Effects of nonsteroidal anti-inflammatory drugs on flexor tendon adhesion," *Journal of Hand Surgery*, vol. 35, no. 6, pp. 941–947, 2010.
- [19] S. Rigozzi, R. Müller, and J. G. Snedeker, "Local strain measurement reveals a varied regional dependence of tensile tendon mechanics on glycosaminoglycan content," *Journal of Biomechanics*, vol. 42, no. 10, pp. 1547–1552, 2009.
- [20] C. Vergari, P. Pourcelot, L. Holden et al., "A linear laser scanner to measure cross-sectional shape and area of biological specimens during mechanical testing," *Journal of Biomechanical Engineering*, vol. 132, no. 10, Article ID 105001, 7 pages, 2010.
- [21] G. Fessel, J. Cadby, S. Wunderli, Rv. Weeren, and JG. Snedeker, "Dose and time dependent effects of genipin cross-linking on cell viability and tissue mechanics—toward clinical application for tendon repair," *Acta Biomaterialia*, vol. 10, no. 5, pp. 1897–1906, 2014.
- [22] S. H. Taylor, S. Al-Youha, T. van Agtmael et al., "Tendon is covered by a basement membrane epithelium that is required for cell retention and the prevention of adhesion formation," *PLoS ONE*, vol. 6, no. 1, Article ID e16337, 2011.
- [23] J. K. F. Wong, Y. H. Lui, Z. Kapacee, K. E. Kadler, M. W. J. Ferguson, and D. A. McGrouther, "The cellular biology of flexor tendon adhesion formation: an old problem in a new paradigm," *The American Journal of Pathology*, vol. 175, no. 5, pp. 1938–1951, 2009.
- [24] S. J. Bates, E. Morrow, A. Y. Zhang, H. Pham, M. T. Longaker, and J. Chang, "Mannose-6-phosphate, an inhibitor of transforming growth factor- β , improves range of motion after flexor tendon repair," *Journal of Bone and Joint Surgery A*, vol. 88, no. 11, pp. 2465–2472, 2006.
- [25] M. Wiig, K. Olmarker, J. Håkansson, L. Ekström, E. Nilsson, and M. Mahlapuu, "A lactoferrin-derived peptide (PXL01) for the reduction of adhesion formation in flexor tendon surgery: an experimental study in rabbits," *Journal of Hand Surgery: European Volume*, vol. 36, no. 8, pp. 656–662, 2011.
- [26] A. Khanna, M. Friel, N. Gougoulis, U. G. Longo, and N. Maffulli, "Prevention of adhesions in surgery of the flexor tendons of the hand: what is the evidence?" *British Medical Bulletin*, vol. 90, no. 1, pp. 85–109, 2009.
- [27] F. E. Karabekmez and C. Zhao, "Surface treatment of flexor tendon autograft and allograft decreases adhesion without an effect of graft cellularity: a pilot study," *Clinical Orthopaedics and Related Research*, vol. 470, no. 9, pp. 2522–2527, 2012.
- [28] T. de Wit, D. de Putter, W. M. W. Tra et al., "Auto-crosslinked hyaluronic acid gel accelerates healing of rabbit flexor tendons in vivo," *Journal of Orthopaedic Research*, vol. 27, no. 3, pp. 408–415, 2009.
- [29] H. Pihlajamäki, O. Tynninen, P. Karjalainen, and P. Rokkanen, "Enveloping bioabsorbable polyglycolide membrane and immobilization in achilles tendon repair: a comparative experimental study on rabbits," *Journal of Orthopaedic Research*, vol. 26, no. 2, pp. 264–270, 2008.
- [30] N. Ishiyama, T. Moro, K. Ishihara et al., "The prevention of peritendinous adhesions by a phospholipid polymer hydrogel formed in situ by spontaneous intermolecular interactions," *Biomaterials*, vol. 31, no. 14, pp. 4009–4016, 2010.
- [31] K. Hammond, H. Starr, D. Katz, and J. Seiler, "Effect of aftercare regimen with extensor tendon repair: a systematic review of the literature," *Journal of Surgical Orthopaedic Advances*, vol. 21, no. 4, pp. 246–252, 2012.
- [32] N. Maffulli, C. Tallon, J. Wong, K. P. Lim, and R. Bleakney, "Early weightbearing and ankle mobilization after open repair of acute midsubstance tears of the Achilles tendon," *American Journal of Sports Medicine*, vol. 31, no. 5, pp. 692–700, 2003.
- [33] M. Aoki, N. Ogiwara, T. Ohta, and Y. Nabeta, "Early active motion and weightbearing after cross-stitch Achilles tendon repair," *The American Journal of Sports Medicine*, vol. 26, no. 6, pp. 794–800, 1998.
- [34] N. H. M. Mortensen, O. Skov, and P. E. Jensen, "Early motion of the ankle after operative treatment of a rupture of the Achilles tendon—a prospective, randomized clinical and radiographic study," *The Journal of Bone and Joint Surgery A*, vol. 81, no. 7, pp. 983–990, 1999.
- [35] R. S. Simons, "Running, breathing and visceral motion in the domestic rabbit (*Oryctolagus cuniculus*): testing visceral displacement hypotheses," *Journal of Experimental Biology*, vol. 202, no. 5, pp. 563–577, 1999.

- [36] B. Saad, M. Casotti, T. Huber et al., "In vitro evaluation of the biofunctionality of osteoblasts cultured on DegraPol-foam," *Journal of Biomaterials Science, Polymer Edition*, vol. 11, no. 8, pp. 787–800, 2000.
- [37] A. Sukthankar, *DegraPol-btc/btgc: In Vitro Untersuchung der Biokompatibilität für Tenozyten*, University Hospital Zurich, Zurich, Switzerland, 1999.
- [38] Z. X. Meng, Y. S. Wang, C. Ma, W. Zheng, L. Li, and Y. F. Zheng, "Electrospinning of PLGA/gelatin randomly-oriented and aligned nanofibers as potential scaffold in tissue engineering," *Materials Science and Engineering C*, vol. 30, no. 8, pp. 1204–1210, 2010.
- [39] C. L. Zhang, W. M. Tian, and F. Z. Cui, "Evaluation of resorbable barrier for preventing surgical adhesions," *Current Applied Physics*, vol. 5, no. 5, pp. 411–416, 2005.
- [40] Y. F. Wu and J. B. Tang, "Tendon healing, edema, and resistance to flexor tendon gliding. Clinical implications," *Hand Clinics*, vol. 29, no. 2, pp. 167–178, 2013.
- [41] R. N. Wrenn, J. L. Goldner, and J. L. Markee, "An experimental study of the effect of cortisone on the healing process and tensile strength of tendons," *The Journal of Bone & Joint Surgery A*, vol. 36, no. 3, pp. 588–601, 1954.
- [42] N. Juncosa-Melvin, J. T. Shearn, G. P. Boivin et al., "Effects of mechanical stimulation on the biomechanics and histology of stem cell-collagen sponge constructs for rabbit patellar tendon repair," *Tissue Engineering*, vol. 12, no. 8, pp. 2291–2300, 2006.
- [43] F. L. Ashley, R. S. Stone, M. Alonsoartieda et al., "Experimental and clinical studies on the application of monomolecular cellulose filter tubes to create artificial tendon sheaths in digits," *Plastic and Reconstructive Surgery and the Transplantation Bulletin*, vol. 23, no. 5, pp. 526–534, 1959.
- [44] J. Chang, R. Thunder, D. Most, M. T. Longaker, and W. C. Lineaweaver, "Studies in flexor tendon wound healing: neutralizing antibody to TGF- β 1 increases postoperative range of motion," *Plastic and Reconstructive Surgery*, vol. 105, no. 1, pp. 148–155, 2000.
- [45] S. C. Juneja, E. M. Schwarz, R. J. O'Keefe, and H. A. Awad, "Cellular and molecular factors in flexor tendon repair and adhesions: a histological and gene expression analysis," *Connective Tissue Research*, vol. 54, no. 3, pp. 218–226, 2013.



Hindawi

Submit your manuscripts at
<http://www.hindawi.com>

