



THE UNIVERSITY *of* EDINBURGH

Edinburgh Research Explorer

Separation of atomic and molecular ions by ion mobility with an RF carpet

Citation for published version:

Miskun, I, Dickel, T, Andres, SAS, Bergmann, J, Constantin, P, Ebert, J, Geissel, H, Greiner, F, Haettner, E, Hornung, C, Lippert, W, Mardor, I, Moore, I, Plaß, WR, Purushothaman, S, Rink, A-K, Reiter, MP, Scheidenberger, C & Weick, H 2020, 'Separation of atomic and molecular ions by ion mobility with an RF carpet', *International Journal of Mass Spectrometry*. <https://doi.org/10.1016/j.ijms.2020.116450>

Digital Object Identifier (DOI):

[10.1016/j.ijms.2020.116450](https://doi.org/10.1016/j.ijms.2020.116450)

Link:

[Link to publication record in Edinburgh Research Explorer](#)

Document Version:

Peer reviewed version

Published In:

International Journal of Mass Spectrometry

General rights

Copyright for the publications made accessible via the Edinburgh Research Explorer is retained by the author(s) and / or other copyright owners and it is a condition of accessing these publications that users recognise and abide by the legal requirements associated with these rights.

Take down policy

The University of Edinburgh has made every reasonable effort to ensure that Edinburgh Research Explorer content complies with UK legislation. If you believe that the public display of this file breaches copyright please contact openaccess@ed.ac.uk providing details, and we will remove access to the work immediately and investigate your claim.



Separation of atomic and molecular ions by ion mobility with an RF carpet

Ivan Miskun^{a,*}, Timo Dickel^{a,b}, Samuel Ayet San Andrés^{a,b}, Julian Bergmann^a, Paul Constantin^c, Jens Ebert^a, Hans Geissel^{a,b}, Florian Greiner^a, Emma Haettner^b, Christine Hornung^a, Wayne Lippert^a, Israel Mardor^{d,e}, Iain Moore^f, Wolfgang R. Plaß^{a,b}, Sivaji Purushothaman^b, Ann-Kathrin Rink^a, Moritz P. Reiter^{a,g}, Christoph Scheidenberger^{a,b}, Helmut Weick^b

^a*II. Physikalisches Institut, Justus-Liebig-Universität Gießen, 35392 Gießen, Germany*

^b*GSI Helmholtzzentrum für Schwerionenforschung GmbH, 64291 Darmstadt, Germany*

^c*IFIN-HH/ELI-NP, 077125, Măgurele - Bucharest, Romania*

^d*Tel Aviv University, 6997801 Tel Aviv, Israel*

^e*Soreq Nuclear Research Center, 81800 Yavne, Israel*

^f*University of Jyväskylä, 40014 Jyväskylä, Finland*

^g*University of Edinburgh, EH8 9AB Edinburgh, United Kingdom*

Abstract

Gas-filled stopping cells are used at accelerator laboratories for the thermalization of high-energy radioactive ion beams. Common challenges of many stopping cells are a high molecular background of extracted ions and limitations of extraction efficiency due to space-charge effects. At the FRS Ion Catcher at GSI, a new technique for removal of ionized molecules prior to their extraction out of the stopping cell has been developed. This technique utilizes the RF carpet for the separation of atomic ions from molecular contaminant ions through their difference in ion mobility. Results from the successful implementation and test during an experiment with a 600 MeV/u ^{124}Xe primary beam are presented. Suppression of molecular contaminants by three orders of magnitude has been demonstrated. Essentially background-free measurement conditions with less than 1 % of background events within a mass-to-charge range of 25 u/e have been achieved. The technique can also be used to reduce the space-charge effects at the extraction nozzle and in the downstream beamline, thus ensuring high efficiency of ion transport and highly-accurate measurements under space-charge-free conditions.

Keywords: ion mobility, gas cell, beam purification, molecular contamination, low-energy RIB, space charge

1. Introduction

Gas-filled stopping cells are widely used in many in-flight accelerator facilities to slow down high-energy radioactive ion beams (RIBs) and to enable high-accuracy measurements with low-energy ions [1]. Incoming ions are thermalized in a noble gas, extracted and delivered to the measuring device. For efficient thermalization of ion beams produced at relativistic energies, areal densities of the buffer gas of at least a few mg/cm² are required. In order to provide a fast extraction and, therefore, give access to short-lived nuclei, the ions are extracted by a combination of DC and RF electric fields and a gas flow [2, 3, 4, 5, 6]. RF fields are created by so-called RF carpets and funnels, devices with fine electrode structures that repel ions, preventing them from

being lost and guiding them towards an extraction orifice.

An important limitation of the extraction efficiency, encountered by high-density stopping cells, is caused by impurities of the buffer gas. During slowing down and thermalization in the buffer gas, each incoming ion generates about 2.5×10^4 ion-electron pairs per MeV of deposited energy in a helium buffer gas. The ionized buffer gas atoms may undergo charge-exchange reactions with atoms of impurities contained in the buffer gas. Once ionized, the contaminants can form adducts and molecules with mass-to-charge ratios over a very broad range and are extracted out of the stopping cell together with the ion of interest (IOI) [7]. Adducts are loosely-bound and can be fragmented after their extraction with dissociation techniques, such as collision-induced dissociation (CID) [8, 9]. Strongly-bound molecules are harder to remove. They lead to po-

*Corresponding author: ivan.miskun@exp2.physik.uni-giessen.de

tentially harmful space charge in the downstream transport beamline and high-precision experimental devices and, in some cases, might even deteriorate the accuracy of the measurement.

In addition, the extraction efficiency of the stopping cell can also be limited by space-charge effects occurring inside of the cell [10, 11]. Electrons, due to their high mobility, are removed with electric fields out of the stopping volume very quickly. In contrast, the positively-charged ions of the buffer gas and its contaminants have velocities similar to the one of the stopped IOI. At high rates of the incoming beam, this can lead to a build-up of a positive charge inside the stopping volume of the stopping cell. A significant space charge can distort the electric fields and cause additional ion losses in the bulk of the cell [12, 13] or in the extraction orifice [14]. The space-charge effects in the extraction orifice are not universally observed in all gas-filled stopping cells, however, in some cases, they can cause drastic efficiency losses. This limitation is currently one of the main bottlenecks for the production of cooled and re-accelerated RIBs of high intensities.

In this work, a novel separation technique that allows molecular suppression prior to ion extraction out of the stopping cell is presented. This technique utilizes the dependence of the effective repelling field of the RF carpet on ion mobility. It has been used to avoid space charge at the extraction nozzle of a stopping cell and for suppression of molecules, in particular of those that are too strongly-bound to be dissociated by the CID in the RFQ beamline.

2. Experimental Method

Ion mobility spectrometry is commonly used in analytical chemistry and has many applications [15]. There, ion funnels are often employed for radial focusing, accumulation, and bunching of ions [16]. In this work, we propose to use the RF funnel or carpet itself for the separation by ion mobility.

The effective repelling field E_{eff} of the RF carpet depends on different parameters [17, 2] and can be described from the pseudopotential approach by

$$E_{\text{eff}} = 2V_{\text{RF}}^2 \frac{1}{r_0^3} \left(\frac{r}{r_0} \right) \left(K_0 \frac{T}{T_0} \frac{p_0}{p} \right)^2 \frac{m}{q}, \quad (1)$$

where m and q are the mass and the charge of the ions, V_{RF} is the amplitude of the applied RF voltage, r_0 is the half of the electrode pitch size (the distance between centers of neighboring electrodes), K_0 is the reduced ion mobility, T is the temperature ($T_0 = 273.15$ K), and p

is the pressure of the buffer gas ($p_0 = 1.013$ bar). For fixed operating parameters, ions with different mass-to-charge ratios and mobility values experience different effective repelling field. From Eq. (1), it is possible to deduce the dependence of the relative change in the effective repelling field of the RF carpet $\Delta E_{\text{eff}}/E_{\text{eff}}$ on relative changes in the mass-to-charge ratio and the mobility of the ion:

$$\frac{\Delta E_{\text{eff}}}{E_{\text{eff}}} = \frac{\Delta(m/q)}{(m/q)} + \frac{2\Delta K_0}{K_0}. \quad (2)$$

For ions with a value of $K_0^2 \times (m/q)$ lower than a certain threshold value, the effective repelling field is not strong enough to compensate the DC field pushing the ions towards the carpet. These ions are not repelled by the RF carpet and are lost on the surface of the electrodes. This is of great importance as it helps to prevent the extraction of low-mass ionization products out of the stopping cell, in particular the ionized buffer gas. The dependence of E_{eff} on the reduced ion mobility K_0 makes it possible to separate the IOI and molecular contaminants with the same mass-to-charge ratio by their respective ion mobilities with the RF carpet.

The separation by ion mobility at the RF carpet works in the following way. In a buffer gas, molecules and adducts have lower mobility value than atomic ions of the same mass-to-charge ratio, due to their larger size. This means that, according to Eq. (1), for the same operating conditions, the molecular ions have a weaker effective repelling field than the isobaric atomic ions. Thus, with the proper choice of the RF amplitude, it is possible to find a regime whereby the atomic ions are repelled and extracted out of the stopping cell, and the molecular ions are lost on the electrodes of the RF carpet. Furthermore, in helium gas at cryogenic temperatures of 70-80 K, atomic positively-charged ions over a broad mass-to-charge range have similar reduced mobility values within a typical range of 15 to 21 $\text{cm}^2/(\text{Vs})$ [18, 19]. The major charge carrier in helium buffer gas at these temperatures, which is He_3^+ , also has comparable ion mobility of about 18 $\text{cm}^2/(\text{Vs})$ [20]. Therefore, it is possible to find operating conditions that allow the transport of atomic ions in the mass-to-charge region of interest while the molecular contaminants are suppressed. As can be seen from Eq. (2), the effective repelling field is twice as sensitive to the change in the ion mobility than to the change in the mass-to-charge ratio, which allows molecular suppression without significant losses for atomic ions over a broad mass-to-charge range.




Figure 1: Left: schematic figure of the part of the cryogenic stopping cell (CSC) in a sectional view. Right: schematic representation of the pushing and focusing DC fields and the counteracting effective repelling field of the RF carpet.

3. Experimental Setup

The FRS Ion Catcher (FRS-IC) [21, 22] is a test facility for the Low-Energy Branch of the Super-FRS [23, 24] at FAIR. It is installed at the end of the symmetric branch of the fragment separator FRS [25] at GSI and consists of three major parts: a gas-filled cryogenic stopping cell (CSC) [6, 26, 27], an RFQ beamline [26] and a multiple-reflection time-of-flight mass spectrometer (MR-TOF-MS) [28, 29]. The exotic nuclei are produced, separated in-flight and range bunched at relativistic energies of up to 1 GeV/u in the FRS, slowed down in degraders, and thermalized in the CSC. The thermalized ions are guided to the exit side of the stopping cell by a DC electric field. At the exit side, a PCB-based RF carpet with concentric electrodes with a density of 4 electrodes per millimeter is installed. The RF carpet repels and guides the ions towards the extraction nozzle with a combination of RF and radial DC electric fields. This part of the CSC, as well as the electric fields in the vicinity of the RF carpet, are shown schematically in Fig. 1. Once the ions reach the nozzle, they are extracted by the gas flow into the RFQ beamline and transported to the MR-TOF-MS, where high-resolution mass measurements are performed.

For the CSC, many different technological steps were taken to ensure the high cleanliness of the buffer gas. To reduce the amount of residual gas inside the CSC, the CSC is baked at a temperature of about 400 K and constantly pumped by a turbomolecular pump for a few days before the cool-down. This, together with cryogenic operation temperatures of 70-80 K, helps in removing most of the contaminants [27]. The stopping cell itself works as a cryopump, where residual contaminants freeze out on the walls of the device. In addition,

before being supplied into the CSC, the helium buffer gas is purified by an LN₂ cold trap and by a getter (MicroTorr, SAES Pure Gas). In combination, all these steps result in the high cleanliness of the CSC, which is required for ion survival and contaminant suppression.

Nevertheless, some contaminants still enter the CSC in trace amounts together with the helium buffer gas. Mostly, these are noble gases present in helium gas cylinders as residuals. They are very hard to filter out completely. Atoms of noble gases contained in the buffer gas do not harm the ion survival, but can be ionized and extracted when the incoming ion beam is thermalized in the stopping cell. It has been seen that the rates of extracted ions of different noble gases change from one gas cylinder to another.

The separation by ion mobility was investigated in an online experiment at the FRS-IC. The ¹²⁴Xe primary beam was stopped in the CSC and produced ionization of the buffer gas and its contaminants. Extracted ionization products were measured with the MR-TOF-MS in the broadband time-focus shift (TFS) mode [29] at a mass resolving power of about 1000. The CSC was filled with 75 mbar of helium gas at a temperature of 74 K, which corresponds to an areal density of 5.1 mg/cm². A DC push field of 14 V/cm was applied. The ions spent 0 to 170 ms on the carpet surface before their extraction due to a problem with the electronics of the RF carpet, which, nevertheless, did not affect the extraction efficiency [30, 31]. The RF carpet was operated at a frequency of 5.92 MHz. In all measurements, the extraction RFQ was operated as a mass filter, transmitting a mass-to-charge window from 70 u/e to 95 u/e. The isolation-dissociation-isolation technique (IDI) [32] was used as a primary method for the suppression of molecular background. It was implemented by using the RFQ mass-filter as the first isolation step, CID in the RFQ beamline with a voltage step of 50 V, and a low-mass cut-off of the first RFQ of the MR-TOF-MS as the second isolation step. In the mass-to-charge region of the ¹²⁴Xe⁺ ions extracted out of the CSC, all molecular contaminants could be efficiently removed from the measured spectrum by the IDI technique only [32].

4. Results

The rate of the incoming ion beam was about 20,000 ions/s. The stopping efficiency of the CSC in this experiment was about 25 %, so $\sim 5,000$ ¹²⁴Xe ions per second were thermalized in the effective volume of the CSC. ATIMA calculations [33] estimate an average energy deposition by a single incoming ion of 190 MeV, which corresponds to about 9×10^{10} He⁺-electron

pairs per second generated in the CSC. In Figure 2a, the measured mass-to-charge spectrum of lower-mass ionization products is shown. IDI was applied, the RF carpet was operated at $68 V_{\text{peak-peak}}$. Most of the peaks in the spectrum correspond to the ionized stable isotopes of krypton, a residual contaminant of the helium gas from a gas cylinder. Because for several mass lines, the rate of detected ions was larger than 100 ions/s, dead-time effects occurred in the data acquisition system of the MR-TOF-MS for these mass lines. As a result, the measured relative abundances of the krypton isotopes do not precisely correspond to the literature values. The measured counts in each spectrum were normalized to the number of spills and intensity of the incoming ion beam. As can be seen from the spectrum, one isotope of krypton, ^{80}Kr , overlaps with a high-abundant molecular contaminant. The dip in the middle of the peak does not represent a mass separation of two species but is caused by severe dead-time effects. With the IDI method, only about half of this molecular contaminant could be broken up. Besides the molecule at 80 u/e, there is no significant molecular background. For every few projectile ions hitting the CSC, only one Kr ion is produced. The measurement was done with a single helium gas cylinder, so the production rate of the Kr ions can be assumed to be constant. These results demonstrate the good cleanliness of the CSC.

Then, the RF voltage of the RF carpet was lowered from $68 V_{\text{peak-peak}}$ to $48 V_{\text{peak-peak}}$, which did not bring a significant difference to the composition of the spectrum, however, led to the partial suppression of the molecular contaminant, as can be seen from the reduction of dead-time effects in the peak at 80 u/e (Fig. 2b). Afterwards, the RF voltage was further decreased down to $41 V_{\text{peak-peak}}$. The spectrum for this setting is shown in Fig. 2c. The molecular contaminant at the mass line 80 u/e was removed, leaving the ^{80}Kr isotope peak, which fits its natural abundance. This behavior corresponds to the RF carpet operation regime when the effective repelling field is strong enough to repel the atomic ions, but not the molecular ones with lower ion mobility. The transition from the transmission of the molecular contaminant to its significant suppression is rather sharp for a small change in the RF voltage. This indicates that the separation by ion mobility is a very well-defined suppression method.

Further decrease of the repelling RF voltage to $34 V_{\text{peak-peak}}$ led to the suppression of Kr isotopes, as shown in Fig. 2d. This effect occurs due to the mass dependence of the effective repelling field of the RF carpet. The ions of lower masses feel a weaker effective repelling field.




Figure 2: Mass spectra of ionization products extracted out of the CSC measured with the MR-TOF-MS in the TFS mode at a mass resolving power of ~ 1000 . Stable isotopes of krypton and molecular contaminant (MC) are present. IDI is applied. All spectra are normalized to the number of spills and intensity of the incoming ion beam. a) RF carpet is at $68 V_{\text{peak-peak}}$, abundant molecular contaminant overlapping with $^{80}\text{Kr}^+$ is not removed by IDI; b) RF carpet is at $48 V_{\text{peak-peak}}$, no change in spectrum composition despite 30 % lower RF repelling voltage, the molecular contaminant is partly suppressed but not removed; c) RF carpet is at $41 V_{\text{peak-peak}}$, the molecule at 80 u/e is removed, the abundance of the peak at $^{80}\text{Kr}^+$ corresponds to the natural abundance ratio of stable Kr isotopes; d) RF carpet is at $34 V_{\text{peak-peak}}$, Kr isotopes are suppressed.

The relative extraction efficiencies of the measured ions are plotted in Fig. 3. The measured efficiencies for Kr isotopes with masses 78, 82, 83, 84, and 86 u were normalized independently to their highest count rate detected within the RF voltage scan. The measured counts were corrected for the dead-time effects and the fluctuations in the beam intensity. The efficiency of the ^{80}Kr isotope was estimated from the measured efficiencies of ^{82}Kr and ^{83}Kr and their natural abundance ratio. The efficiency of the molecular contaminant at the mass-to-charge line 80 u/e was calculated from the measured efficiency of all ions with 80 u/e and the estimated efficiency of the ^{80}Kr isotope. Plotted error bars reflect the statistical uncertainties and uncertainties introduced by dead-time effects and changes in beam intensity. RF voltages could be measured with an accuracy of $\sim 5\%$. As can be seen from the plot, most stable Kr isotopes follow the same trend in extraction efficiency, which decreases with smaller RF voltage applied to the RF carpet. At the same time, the extraction efficiency of the molecular contaminant drastically drops at voltages below $55 V_{\text{peak-peak}}$. At RF voltages of $37 V_{\text{peak-peak}}$ and $41 V_{\text{peak-peak}}$, the molecular contaminant is suppressed by factors of 220 and 160 stronger compared to the stable Kr isotopes, respectively. The data points for the molecular background reflect the combined efficiency of all other ions measured in the mass-to-charge range from 70 u/e to 95 u/e. The overall background is also strongly suppressed with decreasing RF voltage. Its slope is not as steep as of the molecular contaminant at 80 u/e, because the background is composed of multiple contaminants that all have different ion mobilities. Changes in the steepness of the slope indicate the removal of individual species from the mixture. The different behavior observed for ^{78}Kr suggests that its mass line also overlapped with molecular contaminants, which were not recognized in the measured spectra due to the low statistics and resolving power. For RF voltages lower than $48 V_{\text{peak-peak}}$, the efficiency of ^{78}Kr follows the same trend as for the other Kr isotopes, with an offset due to the removed overlapping contaminants. From the difference in the RF voltage required for 50% extraction efficiency, it is possible to deduce the difference in the reduced ion mobility of ^{83}Kr and the molecular contaminant at 80 u/e using Eq. (1). The mobility ratio of $K_0(^{83}\text{Kr})/K_0(\text{MC80}) = 1.4 \pm 0.2$ is obtained. The relatively large uncertainty of this value is caused by the uncertainties in the measured counts due to the dead-time effects. According to [34], the $^{83}\text{Kr}^+$ ions have a reduced ion mobility in helium gas at 74 K of about $17 \text{ cm}^2/(\text{Vs})$. The estimated reduced mobility value of the molecular contaminant is, therefore,




Figure 3: Relative extraction efficiencies of stable Kr isotopes, molecular contaminant (MC) with $m/q=80$ u/e, and overall molecular background at different RF voltages applied to the RF carpet. Efficiencies of $^{80}\text{Kr}^+$ were estimated from the measured counts of $^{82}\text{Kr}^+$ and $^{83}\text{Kr}^+$ and their natural abundances. Efficiencies of MC were calculated from the total measured counts in 80 u/e peak and estimated $^{80}\text{Kr}^+$ efficiencies. A dramatic efficiency drop for the molecular contaminant is caused by its lower ion mobility value.

$K_0(\text{MC80}) = (12 \pm 2) \text{ cm}^2/(\text{Vs})$. This reduced ion mobility value indicates that the molecule might be a hydrocarbon. However, there is no corresponding isotopic abundance pattern seen in any of the measured spectra. The measured abundances of the peak at 81 u/e, which would correspond to the same molecule with one ^{12}C replaced by ^{13}C , are less than 2% of the abundances of MC80. A high-resolution measurement of this mass-to-charge range with the MR-TOF-MS was not performed during the discussed experiment, so the identification of this molecular contaminant was not done.

In Figure 4, the spectra for $68 V_{\text{peak-peak}}$ and $41 V_{\text{peak-peak}}$ (Fig. 2a and Fig. 2c) are shown on a logarithmic scale. The logarithmic scale reveals low-abundant molecular contaminants, which could not be seen on the linear scale. The spectrum measured with $68 V_{\text{peak-peak}}$ at the RF carpet and no IDI applied is also added to the plot. As can be seen from the first two panels, in this mass-to-charge range, the IDI technique did not bring significant improvement to the composition of the spectrum. Only the contaminant peak at 81 u/e was suppressed by an order of magnitude. At the same time, it can be clearly seen that the separation by ion mobility at the RF carpet works over a very broad mass-to-charge range. The ratio of counts of Kr isotopes to all other counts measured in the range from 70 u/e to 95 u/e in Fig. 4b and Fig. 4c increases by a factor of 125. In Figure 4c, background events amount to




Figure 4: Measured mass spectra of ionization products extracted out of the CSC plotted in logarithmic scale. All spectra are normalized to the number of spills and intensity of the incoming ion beam. a) IDI is not applied, RF carpet is at $68 V_{\text{peak-peak}}$; b) IDI is applied, RF carpet is at $68 V_{\text{peak-peak}}$ (spectrum from Fig. 2a); c) IDI is applied, RF carpet is at $41 V_{\text{peak-peak}}$ (spectrum from Fig. 2c). Separation by ion mobility at the RF carpet strongly suppresses the molecular background over a mass-to-charge range of 25 u/e.

less than 1 % of all measured counts. Both very abundant and low-abundant molecular contaminants are essentially completely removed from the spectrum. These background-free conditions are of a high importance for the measurements of low-yield nuclides as they improve the overall sensitivity of the experiment.

5. Conclusion and Outlook

The separation by ion mobility in gas-filled stopping cells with RF structures was developed and successfully implemented at the FRS-IC. In this work, it was shown for an RF carpet, however, similar results can be also expected for RF funnels. Suppression of molecular contaminants by almost three orders of magnitude for less than 50 % loss of the ions of interest was demonstrated. A background level of less than 1 % for a very broad

mass-to-charge range of 25 u/e was achieved. Neutralization of molecular ions at the RF carpet surface helps to reduce unnecessary space charge effects at the extraction nozzle and in the downstream beamline. In addition, it can be used to remove strongly-bound molecules that are difficult to break up by collision-induced dissociation in the RFQ beamline. In combination with other separation methods, this technique can be a very powerful tool for the suppression of molecular background and significant improvement of the cleanliness of the measured spectrum. Although the ion mobility value of the molecular contaminant obtained in this work ($K_0(\text{MC80}) = (12 \pm 2) \text{ cm}^2/(\text{Vs})$) has a relatively large uncertainty due to the dead-time effects, the developed technique can also be, in principle, used as a complementary method for the measurement of ion mobilities.

It is important to note that the developed technique requires an extended understanding of the interplay of the system's operating conditions and parameters. In addition, the current RF carpet is resonance-driven, so the applied RF voltage and, therefore, the separation by ion mobility may be affected by small temperature changes. This will be avoided in the future by implementing an active RF-amplitude and areal density regulation.

Acknowledgments

This work was supported by the German Federal Ministry for Education and Research (BMBF) under contracts no. 05P16RGFN1 and 05P19RGFN1, by Justus Liebig University Gießen and GSI under the JLU-GSI strategic Helmholtz partnership agreement, by HGS-HIRe, and by the Hessian Ministry for Science and Art (HMWK) through the LOEWE Center HICforFAIR. PC is supported by ELI-NP Phase II (1/07.07.2016, COP, ID 1334).

References

- [1] M. Wada, Genealogy of gas cells for low-energy RI-beam production, *Nucl. Instrum. Methods B* 317 (2013) 450–456.
- [2] M. Wada, Y. Ishida, T. Nakamura, Y. Yamazaki, T. Kambara, H. Ohyama, Y. Kanai, T. M. Kojima, Y. Nakai, N. Ohshima, A. Yoshida, T. Kubo, Y. Matsuo, Y. Fukuyama, K. Okada, T. Sonoda, S. Ohtani, K. Noda, H. Kawakami, I. Katayama, Slow RI-beams from projectile fragment separators, *Nucl. Instrum. Methods B* 204 (2003) 570–581.
- [3] G. Savard, J. Clark, C. Boudreau, F. Buchinger, J. Crawford, H. Geissel, J. Greene, S. Gulick, A. Heinz, J. K. P. Lee, A. Levand, M. Maier, G. Münzenberg, C. Scheidenberger, D. Seweryniak, K. S. Sharma, G. Sprouse, J. Vaz, J. C. Wang, B. J. Zabransky, Z. Zhou, Development and operation of gas catchers to thermalize fusion-evaporation and fragmentation products, *Nucl. Instrum. Methods B* 204 (2003) 582–586.

- [4] L. Weissman, D. J. Morrissey, G. Bollen, D. A. Davies, E. Kwan, P. A. Lofy, P. Schury, S. Schwarz, C. Sumithrarachchi, T. Sun, R. Ringle, Conversion of 92 MeV/u $^{38}\text{Ca}/^{37}\text{K}$ projectile fragments into thermalized ion beams, *Nucl. Instrum. Methods A* 540 (2005) 245–258.
- [5] J. Neumayr, L. Beck, D. Habs, S. Heinz, J. Szerypo, P. Thirolf, V. Varentsov, F. Voit, D. Ackermann, D. Beck, M. Block, Z. Di, S. Eliseev, H. Geissel, F. Herfurth, F. Heßberger, S. Hofmann, H.-J. Kluge, M. Mukherjee, G. Münzenberg, M. Petrick, W. Quint, S. Rahaman, C. Rauth, D. Rodriguez, C. Scheidenberger, G. Sikler, Z. Wang, C. Weber, W. Plaß, M. Breitenfeldt, A. Chaudhuri, G. Marx, L. Schweikhard, A. Dodonov, Y. Novikov, M. Suhonen, The ion-catcher device for SHIP-TRAP, *Nucl. Instrum. Methods B* 244 (2006) 489–500.
- [6] M. Ranjan, S. Purushothaman, T. Dickel, H. Geissel, W. R. Plaß, D. Schäfer, C. Scheidenberger, J. Van de Walle, H. Weick, P. Dendooven, New stopping cell capabilities: RF carpet performance at high gas density and cryogenic operation, *Eur. Phys. Lett.* 96 (2011) 52001.
- [7] Y. Kudryavtsev, B. Bruyneel, M. Huyse, J. Gentens, P. Van den Bergh, P. Van Duppen, L. Vermeeren, A gas cell for thermalizing, storing and transporting radioactive ions and atoms. part i: Off-line studies with a laser ion source, *Nucl. Instrum. Methods B* 179 (2001) 412–435.
- [8] S. A. McLuckey, Principles of collisional activation in analytical mass spectrometry, *J. Am. Soc. Mass Spectrom.* 3 (1992) 599–614.
- [9] P. Schury, G. Bollen, M. Block, D. Morrissey, R. Ringle, A. Prinke, J. Savory, S. Schwarz, T. Sun, Beam purification techniques for low energy rare isotope beams from a gas cell, *Hyperfine Interact.* 173 (2006) 165–170.
- [10] D. J. Morrissey, Extraction of thermalized projectile fragments from gas, *Eur. Phys. J. Special Topics* 150 (2007) 365–366.
- [11] I. D. Moore, New concepts for the ion guide technique, *Nucl. Instrum. Methods B* 266 (2008) 4434–4441.
- [12] A. Takamine, M. Wada, Y. Ishida, T. Nakamura, K. Okada, Y. Yamazaki, T. Kambara, Y. Kanai, T. M. Kojima, Y. Nakai, N. Oshima, A. Yoshida, T. Kubo, S. Ohtani, K. Noda, I. Katayama, P. Hostain, V. Varentsov, H. Wollnik, Space-charge effects in the catcher gas cell of a RF ion guide, *Rev. Sci. Instrum.* 76 (2005) 103503.
- [13] M. P. Reiter, A.-K. Rink, T. Dickel, E. Haettner, F. Heiße, W. R. Plaß, S. Purushothaman, F. Amjad, S. Ayet San Andrés, J. Bergmann, D. Blum, P. Dendooven, M. Diwisch, J. Ebert, H. Geissel, F. Greiner, C. Hornung, C. Jesch, N. Kalantar-Nayestanaki, R. Knöbel, J. Lang, W. Lippert, I. Miskun, I. D. Moore, C. Nociforo, M. Petrick, S. Pietri, M. Pfützner, I. Pohjalainen, A. Prochazka, C. Scheidenberger, M. Takechi, Y. K. Tanaka, H. Weick, J. S. Winfield, X. Xu, Rate capability of a cryogenic stopping cell for uranium projectile fragments produced at 1000 MeV/u, *Nucl. Instrum. Methods B* 376 (2016) 240–245.
- [14] C. Sumithrarachchi, D. Morrissey, S. Schwarz, K. Lund, G. Bollen, R. Ringle, G. Savard, A. Villari, Beam thermalization in a large gas catcher, *Nucl. Instrum. Methods B* 463 (2020) 305–309.
- [15] G. A. Eiceman, Z. Karpas, H. H. Hill Jr, *Ion mobility spectrometry*, CRC press, 2013.
- [16] R. T. Kelly, A. V. Tolmachev, J. S. Page, K. Tang, R. D. Smith, The ion funnel: theory, implementations, and applications, *Mass spectrometry reviews* 29 (2010) 294–312.
- [17] A. V. Tolmachev, I. V. Chernushevich, A. F. Dodonov, K. G. Standing, A collisional focusing ion guide for coupling an atmospheric pressure ion source to a mass spectrometer, *Nucl. Instrum. Methods B* 124 (1997) 112–119.
- [18] L. A. Viehland, Zero-field mobilities in helium: highly accurate values for use in ion mobility spectrometry, *Int. J. Ion Mobil. Spec.* 15 (2012) 21–29.
- [19] G. Visentin, M. Laatiaoui, L. A. Viehland, A. A. Buchachenko, Mobility of the singly-charged lanthanide and actinide cations: Trends and perspectives, *Frontiers in Chemistry* 8 (2020) 438.
- [20] P. L. Patterson, Temperature dependence of helium-ion mobilities, *Phys. Rev. A* 2 (1970) 1154.
- [21] W. R. Plaß, T. Dickel, S. Purushothaman, P. Dendooven, H. Geissel, J. Ebert, E. Haettner, C. Jesch, M. Ranjan, M. P. Reiter, H. Weick, F. Amjad, S. Ayet, M. Diwisch, A. Estrade, F. Farinon, F. Greiner, N. Kalantar-Nayestanaki, R. Knöbel, J. Kurcewicz, J. Lang, I. Moore, I. Mukha, C. Nociforo, M. Petrick, M. Pfuetzner, S. Pietri, A. Prochazka, A.-K. Rink, S. Rinta-Antila, D. Schäfer, C. Scheidenberger, M. Takechi, Y. K. Tanaka, J. S. Winfield, M. I. Yavor, The FRS Ion Catcher - a facility for high-precision experiments with stopped projectile and fission fragments, *Nucl. Instrum. Methods B* 317 (2013) 457–462.
- [22] W. R. Plaß, T. Dickel, I. Mardor, S. Pietri, H. Geissel, C. Scheidenberger, D. Amanbayev, S. A. San Andrés, J. Äystö, D. L. Balabanski, et al., The science case of the FRS Ion Catcher for FAIR Phase-0, *Hyperfine Interactions* 240 (2019) 73.
- [23] H. Geissel, H. Weick, M. Winkler, G. Münzenberg, V. Chichkine, M. Yavor, T. Aumann, K. H. Behr, M. Böhmer, A. Brünle, K. Burkard, J. Benlliure, D. Cortina-Gil, L. Chulkov, A. Dael, J.-E. Ducret, H. Emling, B. Franczak, J. Friese, B. Gastineau, J. Gerl, R. Gernhäuser, M. Hellström, B. Jonsson, J. Kojouharova, R. Kulesa, B. Kindler, N. Kurz, B. Lommel, W. Mittig, G. Moritz, C. Mühle, J. A. Nolen, G. Nymann, P. Roussel-Chomaz, C. Scheidenberger, K.-H. Schmidt, G. Schrieder, B. Sherrill, H. Simon, K. Sümmerer, N. A. Tahir, V. Vysotsky, H. Wollnik, A. F. Zeller, The Super-FRS project at GSI, *Nucl. Instrum. Methods B* 204 (2003) 71–85.
- [24] J. S. Winfield, H. Geissel, J. Gerl, G. Münzenberg, C. Nociforo, W. R. Plaß, C. Scheidenberger, H. Weick, M. Winkler, M. I. Yavor, A versatile high-resolution magnetic spectrometer for energy compression, reaction studies and nuclear spectroscopy, *Nucl. Instrum. Methods A* 704 (2013) 76–83.
- [25] H. Geissel, P. Armbruster, K. H. Behr, A. Brünle, K. Burkard, M. Chen, H. Folger, B. Franczak, H. Keller, O. Klepper, B. Langenbeck, F. Nickel, E. Pfeng, M. Pfützner, E. Roeckl, K. Rykaczewski, I. Schall, D. Schardt, C. Scheidenberger, K. H. Schmidt, A. Schröter, T. Schwab, K. Sümmerer, M. Weber, G. Münzenberg, T. Brohm, H. G. Clerc, M. Fauerbach, J. J. Gaimard, A. Grewe, E. Hanelt, B. Knödler, M. Steiner, B. Voss, J. Weckenmann, C. Ziegler, A. Magel, H. Wollnik, J. P. Dufour, Y. Fujita, D. J. Vieira, B. Sherrill, The GSI projectile fragment separator (FRS): a versatile magnetic system for relativistic heavy ions, *Nucl. Instrum. Methods B* 70 (1992) 286–297.
- [26] M. P. Reiter, Pilot Experiments with Relativistic Uranium Projectile and Fission Fragments Thermalized in a Cryogenic Gas-Filled Stopping Cell, PhD thesis, Justus Liebig University Gießen, 2015.
- [27] M. Ranjan, P. Dendooven, S. Purushothama, T. Dickel, M. P. Reiter, S. Ayet, E. Haettner, I. D. Moore, N. Kalantar-Nayestanaki, H. Geissel, W. R. Plaß, D. Schäfer, C. Scheidenberger, F. Schreuder, H. Timersma, J. Van de Walle, H. Weick, Design, construction and cooling system performance of a prototype cryogenic stopping cell for the Super-FRS at FAIR, *Nucl. Instrum. Methods A* 770 (2015) 87–97.
- [28] W. R. Plaß, T. Dickel, U. Czok, H. Geissel, M. Petrick, K. Reinheimer, C. Scheidenberger, M. I. Yavor, Isobar separation by time-of-flight mass spectrometry for low-energy radioactive ion

- beam facilities, *Nucl. Instrum. Methods B* 266 (2008) 4560–4564.
- [29] T. Dickel, W. R. Plaß, A. Becker, U. Czok, H. Geissel, E. Haettner, C. Jesch, W. Kinsel, M. Petrick, C. Scheidenberger, M. I. Yavor, A high-performance multiple-reflection time-of-flight mass spectrometer and isobar separator for the research with exotic nuclei, *Nucl. Instrum. Methods A* 777 (2015) 172–188.
- [30] I. Miskun, T. Dickel, I. Mardor, C. Hornung, D. Amanbayev, S. Ayet San Andrés, J. Bergmann, J. Ebert, H. Geissel, M. Górska, F. Greiner, E. Haettner, W. R. Plaß, S. Purushothaman, C. Scheidenberger, A.-K. Rink, H. Weick, S. Bagchi, P. Constantin, S. Kaur, W. Lippert, B. Mei, I. Moore, J.-H. Otto, S. Pietri, I. Pohjalainen, A. Prochazka, C. Rappold, M. P. Reiter, Y. K. Tanaka, J. S. Winfield, A novel method for the measurement of half-lives and decay branching ratios of exotic nuclei, *Eur. Phys. J. A* 55 (2019) 148. URL: <https://doi.org/10.1140/epja/i2019-12837-8>. doi:10.1140/epja/i2019-12837-8.
- [31] I. Miskun, A Novel Method for the Measurement of Half-Lives and Decay Branching Ratios of Exotic Nuclei with the FRS Ion Catcher, PhD thesis, Justus Liebig University Gießen, 2019.
- [32] F. Greiner, T. Dickel, S. Ayet San Andrés, J. Bergmann, P. Constantin, J. Ebert, H. Geissel, E. Haettner, C. Hornung, I. Miskun, W. Lippert, I. Mardor, I. Moore, W. R. Plaß, S. Purushothaman, A.-K. Rink, M. P. Reiter, C. Scheidenberger, H. Weick, Removal of molecular contamination in low-energy RIBs by the isolation-dissociation-isolation method, *Nucl. Instrum. Methods B* 463 (2020) 324–326.
- [33] H. Weick, H. Geissel, N. Iwasa, C. Scheidenberger, J. R. Sanchez, Improved accuracy of the code ATIMA for energy loss of heavy ions in matter, *GSI Sci. Rep. 2017* 2018-1 (2018) 130 p.
- [34] L. A. Viehland, E. A. Mason, Transport properties of gaseous ions over a wide energy range, IV, *At. Data Nucl. Data Tables* 60 (1995) 37–95.