



Aalborg Universitet

AALBORG UNIVERSITY
DENMARK

Reply to a letter from Jackson J et al: Effect of pH on thrombin activity measured by calibrated automated thrombinography

Kristensen, Søren Risom; Nybo, Jette; Pedersen, Shona

Published in:
Research and Practice in Thrombosis and Haemostasis

DOI (link to publication from Publisher):
[10.1002/rth2.12371](https://doi.org/10.1002/rth2.12371)

Creative Commons License
CC BY-NC-ND 4.0

Publication date:
2020

Document Version
Publisher's PDF, also known as Version of record

[Link to publication from Aalborg University](#)

Citation for published version (APA):
Kristensen, S. R., Nybo, J., & Pedersen, S. (2020). Reply to a letter from Jackson J et al: Effect of pH on thrombin activity measured by calibrated automated thrombinography. *Research and Practice in Thrombosis and Haemostasis*, 4(6), 1065. <https://doi.org/10.1002/rth2.12371>


General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- ? Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- ? You may not further distribute the material or use it for any profit-making activity or commercial gain
- ? You may freely distribute the URL identifying the publication in the public portal ?

Take down policy

If you believe that this document breaches copyright please contact us at vbn@aub.aau.dk providing details, and we will remove access to the work immediately and investigate your claim.



AUTHOR RESPONSE

Reply to a letter from Jackson J et al: Effect of pH on thrombin activity measured by calibrated automated thrombinography

We wish to thank Joseph Jackson and coworkers for their interest in our paper and for elaborating and commenting on effects of pH on thrombin generation (TG). We agree that it is possible to choose other buffers, and lactic acid has been used for investigation of thrombin-induced fibrin formation.¹ An advantage is that it is a physiological component in blood, although using the highest suggested concentration,¹ it is a much higher concentration than normal levels. We chose 4-(2-hydroxyethyl)-1-piperazineethanesulfonic acid (HEPES) buffer because it is the buffer used in the reagents, and adding this to plasma as described in the paper does not change osmolality.

It is correct that pH increases during the initial procedures before measurement, but we do not think that this fully explains the effect of the HEPES buffer increasing TG since our measurements of pH changes during these procedures indicate that the buffer in the reagents is capable of keeping pH neutral in fresh plasma with an initially normal pH. Nevertheless, we have no other explanation for this effect of HEPES. An ideal solution to avoid escape of CO₂ would be to perform the assay in a chamber with 5% CO₂ atmosphere, but this is probably impractical.

It is an interesting figure (Figure 1 in Jackson et al²) showing the effect of pH on thrombin activity toward the fluorogenic substrate, which is very distinct—data that we did not know. However, we do not think that it is possible to overcome the effect of a high pH in the plasma sample using a calibrator with a matching pH. In the assay, we mix 80 µL of plasma with either 20 µL of trigger reagent or 20 µL of calibrator reagent; that is, in both instances plasma with a high pH is added, and consequently the wells with trigger and calibrator are pH matched (we expect that both of these reagents contain 20 mM of HEPES). We imagine that the activity of several coagulation factors is affected by the abnormal pH as also stated by Jackson et al,² and this will reduce the measured TG.

We also agree that measurements of samples from patients with abnormal pH preferably should be measured at the actual pH in the patient, but this is usually not possible in our assays. Interestingly, however, Mitrophanov et al³ have in a very recent paper shown that the effect of acidosis induced in vitro in plasma samples on TG is minimal, although a potential effect of escape of CO₂ was not controlled in these experiments. We will also stress that the intention with our paper was to describe a preanalytical

problem that can be minimized by keeping plasma samples in closed and filled tubes.

Finally, we want to thank Jackson et al once again for their comments and can only agree with their final remark that there are many considerations when performing the CAT analysis.

RELATIONSHIP DISCLOSURE

The authors declare nothing to report.

AUTHOR CONTRIBUTIONS

SRK wrote the reply. JN and SP read and reviewed the manuscript.

Søren Risom Kristensen MD, DMSc^{1,2}

Jette Nybo BMLS¹

Shona Pedersen MSc, PhD^{1,2}

¹Department of Clinical Biochemistry, Aalborg University Hospital, Aalborg, Denmark

²Department of Clinical Medicine, Aalborg University, Aalborg, Denmark

Email: srk@rn.dk

Handling Editor: Søren Risom Kristensen

ORCID

Søren Risom Kristensen <https://orcid.org/0000-0002-2649-506X>

REFERENCES

1. Sinauridze EI, Volkova RI, Krasotkina YV, Sarbash VI, Ataullakhanov FI. Dynamics of clot growth induced by thrombin diffusing into nonstirred citrate human plasma. *Biochim Biophys Acta*. 1998;1425:607–16.
2. Jackson JW, Surov SS, Liang Y, Parunov L, Ovanesov MV. Effect of pH on thrombin activity measured by calibrated automated thrombinography. *Res Pract Thromb Haemost*. 2020;4.
3. Mitrophanov AY, Szlam F, Sniecinski RM, Levy JH, Reifman J. Controlled multifactorial coagulopathy: effects of dilution, hypothermia and acidosis on thrombin generation in vitro. *Anesth Analg*. 2020;130:1063–76.

This is an open access article under the terms of the Creative Commons Attribution-NonCommercial-NoDerivs License, which permits use and distribution in any medium, provided the original work is properly cited, the use is non-commercial and no modifications or adaptations are made.

© 2020 The Authors. *Research and Practice in Thrombosis and Haemostasis* published by Wiley Periodicals LLC on behalf of International Society on Thrombosis and Haemostasis.