

## Research in Translation

# HIV-1 Viral Load Assays for Resource-Limited Settings

Susan A. Fiscus\*, Ben Cheng, Suzanne M. Crowe, Lisa Demeter, Cheryl Jennings, Veronica Miller, Richard Respass, Wendy Stevens, and the Forum for Collaborative HIV Research Alternative Viral Load Assay Working Group

Tremendous strides have been made in treating HIV-1 infection in industrialized countries. Combination therapy with antiretroviral (ARV) drugs suppresses virus replication, delays disease progression, and reduces mortality. In industrialized settings, plasma viral load assays are used in combination with CD4 cell counts to determine when to initiate therapy and when a regimen is failing. In addition, unlike serologic assays, these assays may be used to diagnose perinatal or acute HIV-1 infection.

Unfortunately, the full benefits of antiretroviral drugs and monitoring tests have not yet reached the majority of HIV-1-infected patients who live in countries with limited resources. In this article we discuss existing data on the performance of alternative viral load assays that might be useful in resource-limited settings. Our search strategy and selection criteria for relevant studies are shown in Box 1.

## Viral Load Assays Currently Used in Developed Countries

Currently there are three HIV-1 RNA assays licensed by the United States Food and Drug Administration: Roche Amplicor HIV-1 Monitor Test, version 1.5, bioMérieux NucliSens HIV-1 QT Assay, and Versant HIV-1 RNA 3.0 Assay (bDNA). These assays have been previously described in detail ([1–7], Table 1) and are validated, undergo quality control by the manufacturers, recognize most HIV-1 subtypes, and are familiar to many clinicians.

However, these kits are expensive (kit cost of \$50–\$100/test in the US), and rely on expensive, often dedicated equipment that can only be used for that assay. Although the manufacturers are decreasing kit costs



DOI: 10.1371/journal.pmed.0030417.g001

**Figure 1.** A Laboratory in Malawi. Many laboratories in low-income settings, such as the laboratory shown here, are poorly resourced. (Photo: Susan Fiscus)

and in some instances equipment costs for resource-limited countries, these assays are still technologically complex and require physical resources, such as uninterrupted electricity, air conditioning, and access to clean water, that may not be available in less-developed countries (see Figure 1) [8]. Newer assays making use of real-time polymerase chain reaction

(Roche TaqMan, Abbott RealTime) or molecular beacon technology (Retina Rainbow, NucliSens EasyQ) are also available, but have not yet been approved by the Food and Drug Administration, and very few papers have been published describing their use.

## Barriers to Using Current Viral Load Assays in Resource-Limited Settings

Resources vary significantly by country. Human resources and infrastructure that might be readily available in major cities, such as trained personnel, clean water, and electricity, may not be found in more rural areas. Reference laboratories that have the personnel, equipment, and infrastructure to perform CD4 cell counts and viral load testing may lack the resources to purchase the kits. Thus, what works for one country, or even for one city within a country, may not apply in all resource-limited settings. Options available to developing countries include performing the

**Funding:** The authors received no specific funding for this article.

**Competing Interests:** No member of the writing team owned stock or was employed by the kit manufacturers mentioned in this article. Some free kits and the loan of equipment were provided by manufacturers for evaluation of the assays and for training purposes: SF received three p24 kits and two Cavid kits as well as the loan of an ELISA reader from PerkinElmer; SC received two p24 kits, six Cavid kits, and the loan of an ELISA reader from Cavid; CJ received two Cavid kits; WS received two Cavid kits and the loan of an ELISA reader from PerkinElmer. In addition, SC received partial support from Cavid to attend the 3rd International AIDS Society Conference in Rio de Janeiro, Brazil, 2005. Some members of the Forum for Collaborative HIV Research Alternative Viral Load Assay Working Group were employed by manufacturers, and they are listed in the Acknowledgments.

**Citation:** Fiscus SA, Cheng B, Crowe SM, Demeter L, Jennings C, et al. (2006) HIV-1 viral load assays for resource-limited settings. *PLoS Med* 3(10): e417. DOI: 10.1371/journal.pmed.0030417

**DOI:** 10.1371/journal.pmed.0030417

This is an open-access article distributed under the terms of the Creative Commons Public Domain

declaration which stipulates that, once placed in the public domain, this work may be freely reproduced, distributed, transmitted, modified, built upon, or otherwise used by anyone for any lawful purpose.

**Abbreviations:** ARV, antiretroviral; ELISA, enzyme-linked immunosorbent assay; PCR, polymerase chain reaction; RT, reverse transcriptase

Susan A. Fiscus is at the University of North Carolina at Chapel Hill, Chapel Hill, North Carolina, United States of America. Ben Cheng and Veronica Miller are at the Forum for Collaborative HIV Research, Washington, D. C., United States of America. Suzanne M. Crowe is at the Macfarlane Burnet Institute for Medical Research and Public Health, Melbourne, Australia. Lisa Demeter is at the University of Rochester School of Medicine and Dentistry, Rochester, New York, United States of America. Cheryl Jennings is at Rush University Medical Center, Chicago, Illinois, United States of America. Richard Respass is at the Centers for Disease Control and Prevention, Atlanta, Georgia, United States of America. Wendy Stevens is at the University of the Witwatersrand and the National Health Laboratory Service, Johannesburg, South Africa.

\* To whom correspondence should be addressed. E-mail: [fiscussa@med.unc.edu](mailto:fiscussa@med.unc.edu)

Research in Translation discusses health interventions in the context of translation from basic to clinical research, or from clinical evidence to practice.

laboratory testing at the local site or transporting the specimens to the reference laboratories for testing. A well-defined infrastructure is needed to ensure proper specimen handling and efficient results reporting.

While viral load kit prices have been decreasing for resource-poor countries, the cost of assay disposables and the establishment of a solid infrastructure still limit the utility of HIV-1 RNA for monitoring ARV therapy in most resource-limited settings. Inexpensive, technologically simpler assays are still needed [9–11]. Table 1 summarizes characteristics of some alternative assays, and a more detailed description, together with a summary of the studies on assay performance, is provided below.

### Ultrasensitive P24 Antigen Assay (PerkinElmer Life and Analytical Sciences)

The ultrasensitive p24 (Ultra p24) antigen assay uses a standard ELISA (enzyme-linked immunosorbent assay) format for the capture and detection of HIV-1 p24 antigens coupled with a specific amplification process to increase the assay sensitivity. Heat denaturation of the plasma prior to binding of p24 antigen in the ELISA step helps dissociate immune complexes and denature the antibodies so they no longer compete for binding to the p24 antigen. The dynamic range of the assay can be increased by using a kinetic read-out with the Quanti-Kin Detection System Software [12]. An early version of the Ultra p24 antigen assay was first described by Jorg Schupbach in 1993 using the PerkinElmer HIV-1 p24 antigen kit [13]. Optimization of the standard p24 antigen assay by Schupbach [14–22] includes an external reagent that improves the antigen detection sensitivity, perhaps by greater dissociation of the immune complexes [21]. A number of investigators around the globe have evaluated the assay, with mixed results.

### Viral Load Monitoring Using the Ultrasensitive P24 Antigen Assay

Schupbach and colleagues have reported that the HIV-1 p24 antigen typically decreases in parallel with HIV-1 RNA in successfully treated patients [17] and is a good prognostic indicator of disease progression

[17,20], at least among those with HIV-1 subtype B infection. P24 antigen detection is a significant inverse correlate of CD4 cell changes in virally suppressed patients [21] as well as in patients studied longitudinally who were either treatment naïve (85% of the population) or treated with dual nucleoside reverse transcriptase inhibitors (15%) [20] or were on a structured treatment interruption study [22].

Some investigators have been able to produce similar results when monitoring infected patients [23–26] while others have had more disappointing results [27–29]. Pascual et al. [23] compared the Roche RNA assay to the Ultra p24 antigen assay in 130 patients from the US and Malawi (subtypes B and C) using the kit buffer. The RNA assay was more sensitive in detecting HIV-1 when compared to the Ultra p24 antigen assay (95.4% vs. 84.6%). The Ultra p24 antigen assay detected 66.7% of specimens with viral loads of less than 10,000 copies/ml, 87% of specimens with viral loads between 10,000 and 100,000 copies/ml, and 97.7% of specimens with viral load greater than 100,000 copies/ml. Similar results were observed in Burkina Faso [24] using the kit buffer where the Ultra p24 antigen assay detected 27%, 80%, and 87% of samples with viral loads <1000, 1000–63,000, and >80,000 copies/ml, respectively. Better results were observed with subtype C clinical specimens where 95%–100% of samples with detectable viral loads

over 400 copies/ml had detectable p24 antigen levels [23,25].

In South Africa, Stevens and colleagues [25] monitored 20 patients (subtype C) who were being treated with highly active antiretroviral therapy. In 19 of 20 patients the Ultra p24 antigen serial results (external buffer) paralleled the RNA results. Ribas and colleagues [26] evaluated 40 patients of different HIV-1 subtypes being treated with ARV therapy. In 33 patients, the changes in p24 antigen (external buffer) correlated well with RNA results, while seven others showed a discrepancy. Since the seven patients with discrepancies during monitoring were infected with seven different subtypes and other patients infected with the same subtypes did not show discrepancies, it is unlikely that the observed differences were related to virus subtype.

In contrast, Prado and colleagues [27] monitored patients in Spain undergoing structured treatment interruption and found little correlation between HIV-1 RNA and the Ultra p24 assay using the kit buffer. In this study, 76% of the treated patients and 49% of the naïve patients showed discordance between HIV-1 RNA detection and p24 detection after viral rebound. Bonard and colleagues [28] tested plasma specimens from 14 patients treated with highly active antiretroviral therapy in Côte d'Ivoire and found weaker changes during treatment with the Ultra p24 antigen assay using the kit buffer compared to the HIV-1 RNA results. Seven of the 14 patients had a dramatic decrease in HIV-1 RNA, but only five of these had similar, though less dramatic, changes in p24 antigen. Of the seven who failed ARV treatment, only three had p24 antigen results that seemed to mirror the RNA results. Use of the external buffer [21] has improved the sensitivity of the assay 2- to 5-fold as well as the correlation with HIV-1 RNA ( $r = 0.74$  vs.  $0.61$ ) [30].

### Infant Diagnosis Using the Ultrasensitive P24 Antigen Assay

Diagnosis of perinatal HIV infection is hindered by the presence of maternal immunoglobulins, which can persist in the infant for as long as 15–18 months. Thus, routine antibody assays cannot be used to diagnose infection until after maternal antibodies have waned.

### Box 1. Search Strategy and Selection Criteria

Members of the Forum for Collaborative HIV Research Alternative Viral Load Assay Working Group include leading academic investigators in this field as well as representatives from the US Centers for Disease Control and Prevention, the National Institutes of Health, charitable organizations that fund validation studies of the assays, and industrial sponsors. We included all peer-reviewed published papers found in PubMed using the search terms: "Alternative HIV-1 viral load assays," "HIV-1 viral load AND resource-limited settings," "real-time HIV-1 PCR," "p24 antigen," "Cavidi," "TaqMan," "EasyQ," "Retina Rainbow," and "Abbott RealTime."

**Table 1.** Summary of Commercially Available Viral Load Assays

System	Key Features (Cutoffs Reflect Manufacturer's Claims and May Not Be Based on the Same Criteria)	Parameters Measured	Advantages	Disadvantages
Roche Molecular Systems (AMPLICOR Monitor Microwell Plate Manual and COBAS Automated)	Cutoff 50 copies/ml (0.5 ml); 400 copies/ml (0.2 ml)	HIV-1 RNA	<ul style="list-style-type: none"> <li>• Equipment can be used for other diseases</li> <li>• Can be used for clades A, B, C, D, E, G</li> <li>• 0.2 to 0.5 ml plasma</li> <li>• High throughput</li> </ul>	<ul style="list-style-type: none"> <li>• Contamination risk</li> <li>• Requires skilled technicians</li> <li>• Cost</li> <li>• Dedicated equipment and space</li> <li>• Need good technical support</li> <li>• Some reports of underestimation of non-B subtypes</li> </ul>
Bayer Diagnostics (VERSANT bDNA 3.0)	Cutoff 75 copies/ml (1.0 ml) in the US; Cutoff 50 copies/ml outside US	HIV-1 RNA	<ul style="list-style-type: none"> <li>• Equipment can be used for other diseases</li> <li>• Can be used for clades A, B, C, D, E, G</li> <li>• High throughput</li> <li>• Less contamination risk compared to PCR</li> </ul>	<ul style="list-style-type: none"> <li>• Requires skilled technicians</li> <li>• Cost</li> <li>• Dedicated equipment and space</li> <li>• Need good technical support</li> <li>• Need 1.0 ml plasma</li> </ul>
bioMérieux (Organon Teknika NucliSens QT)	Cutoff 176 copies/ml (2.0 ml); 400 copies/ml (0.2 ml)	HIV-1 RNA	<ul style="list-style-type: none"> <li>• Equipment can be used for other diseases</li> <li>• Can be used for clades A, B, C, D</li> <li>• Can be used for all biological fluids and dried blood spots (sensitivity issues with blood spots)</li> </ul>	<ul style="list-style-type: none"> <li>• Contamination risk</li> <li>• Requires skilled technicians</li> <li>• Cost</li> <li>• Dedicated equipment and space</li> <li>• Need good technical support</li> <li>• Need at least 1.0 ml plasma to achieve reported sensitivity</li> <li>• Some reports of underestimation of certain non-B subtypes</li> </ul>
PerkinElmer Ultrasensitive p24	Cutoff approximately 30,000 copies/ml with current kit extraction reagent (0.05 ml); cutoff approximately 5,000 copies/ml with external extraction reagent. Cutoff variable due to non-virion associated p24 contribution	p24 antigen	<ul style="list-style-type: none"> <li>• Equipment can be shared with ELISA</li> <li>• Training available at CDC, UNC, Rush and through company</li> <li>• May be used for pediatric diagnosis</li> <li>• Easy training</li> <li>• &lt;1 day turn around time</li> <li>• High throughput</li> <li>• 0.05 ml plasma sample volume</li> </ul>	<ul style="list-style-type: none"> <li>• Dry heat block</li> <li>• Needs more extensive evaluation for use in clinical management</li> <li>• Need to understand the clinical impact of non-virion associated p24</li> <li>• Needs more extensive evaluation in different subtypes</li> <li>• DOS-based software for expanded dynamic range</li> </ul>
Cavidi ExaVir Version 2	Cutoff approximately 500–1000 copies/ml using version 2.0 (1.0 ml)	Reverse transcriptase activity	<ul style="list-style-type: none"> <li>• Easy training</li> <li>• Training available through company and at Macfarlane Burnet Institute for Medical Research and Public Health, Melbourne, Australia</li> <li>• Preliminary data suggest it can be used for all subtypes</li> <li>• Easy to perform assay</li> <li>• Potential use for clinical management</li> <li>• Can be used for NNRTI drug resistance monitoring</li> <li>• Results reported as fg RT/ml as well as RNA equivalent copies/ml equivalents</li> <li>• Simple, inexpensive equipment</li> </ul>	<ul style="list-style-type: none"> <li>• 32° C dedicated incubator and vacuum pump needed</li> <li>• Performance time 3 days</li> <li>• Needs more extensive evaluations for use in clinical management</li> <li>• Needs more extensive evaluations in non-B subtypes</li> <li>• Requires 1 ml plasma, though 0.2 ml may be used with loss of sensitivity</li> <li>• Positive and negative control not supplied</li> <li>• Low throughput</li> </ul>
Homebrew Real-Time PCR	Cutoff varies depending on the system (1 copy/ml [55] to 300 copies/ml [62])	HIV-1 RNA	<ul style="list-style-type: none"> <li>• Low cost reagents</li> <li>• Can be extremely sensitive</li> <li>• Multiplex assays possible</li> </ul>	<ul style="list-style-type: none"> <li>• Expensive equipment costs</li> <li>• No manufacturer's QA for homebrew reagents</li> </ul>
Primagen Retina Rainbow/ NucliSens EasyQ	Cutoff 50 copies/ml (2.0 ml); 500 copies/ml (0.2 ml)	HIV-1 RNA	<ul style="list-style-type: none"> <li>• Compatible with dried fluid spots and plasma, serum, whole blood, mothers milk, etc.</li> <li>• Equipment can be used for other bio-markers</li> <li>• Large dynamic range</li> </ul>	<ul style="list-style-type: none"> <li>• Contamination risk</li> <li>• Skilled technicians</li> <li>• Dedicated equipment and space</li> <li>• Need good technical support</li> <li>• Needs more extensive evaluation in non-B subtypes</li> </ul>
Abbott RealTime	Cutoff ~ 40 copies/ml	HIV-1 RNA	<ul style="list-style-type: none"> <li>• High throughput</li> <li>• Large dynamic range</li> </ul>	<ul style="list-style-type: none"> <li>• Very expensive equipment</li> <li>• Dedicated equipment and space</li> <li>• Need good technical support</li> <li>• Needs more extensive evaluation in non-B subtypes</li> </ul>

CDC, Centers for Disease Control and Prevention; NNRTI, non-nucleoside reverse transcriptase inhibitor; QA, quality assurance; UNC, University of North Carolina  
 DOI: 10.1371/journal.pmed.0030417.t001

In industrialized countries, infants are typically diagnosed within days or weeks of birth by confirmed, repeated detection of HIV-1 DNA or RNA [31–35]. However, given the expense and complexity of nucleic acid testing, the World Health Organization strongly encourages the development of technologically simpler, less expensive assays that can be used to diagnose HIV-1 infection in early infancy [36]. The Ultra p24 assay is gaining support as a tool for detection of HIV-1 infection in infants following mother-to-child transmission. Lyamuya and colleagues [37] showed that even an early version of the heat-denatured p24 antigen assay was 99% sensitive and 100% specific in diagnosing HIV-1 subtypes A and D infection in children in Tanzania. Similar results have been observed with subtype B in Switzerland where the sensitivity was 100% after day 10 of age and specificity was 93.8% (99.2% after neutralization) [38], subtype E in Thailand (sensitivity 100%, specificity 100%) [39], subtype C (sensitivity 97%–98%, specificity 97%–99%) [40,41], and multiple subtypes from the Democratic Republic of Congo using either plasma (sensitivity 92.3%, specificity 100%) or dried plasma spots (sensitivity and specificity 100%) [42]. Others have also modified the assay for use with dried plasma spots [43,44], although only one has been successful in adapting the assay to dried blood spots [45].

### Cavidi ExaVir Load Reverse Transcriptase Assay

The reverse transcriptase (RT) enzyme is extracted from the virus particle in the Cavidi RT assay using a solid phase extraction manifold, and is quantified in a functional assay whereby RT synthesizes BrdU-DNA from a poly-A template bound to a 96-well plate [46]. Synthesized DNA is then quantified using anti-BrdU conjugated to alkaline phosphatase followed by the addition of its substrate. The RT activity in the unknown sample is compared to that of a recombinant RT enzyme standard with a known concentration. The extrapolated result is reported as fg RT/ml of plasma or as HIV-1 RNA equivalents/ml using a conversion factor supplied by the manufacturer. The assay has recently undergone revisions to improve sensitivity [47].

### Five Key Papers on HIV Viral Load in Resource-Limited Settings

Majchrowicz, 2003 [10] One of the first articles written describing the scope of the problem.

Patton et al., 2006 [45] First description of the successful use of dried blood spots with the ultrasensitive p24 antigen assay.

Rouet et al., 2005 [62] First demonstration of real-time PCR technology for monitoring and diagnosis in a resource-limited country.

Schupbach et al., 1996 [15] An early paper comparing the ultrasensitive p24 antigen assay with HIV RNA viral load.

Stevens et al., 2005 [25] First independent comparison of the ultrasensitive p24 antigen assay and the Cavid reverse transcriptase assay in a resource-limited country.

Since the assay measures a virion-associated enzyme, results are usually more comparable to plasma RNA [48]. Additionally, since this is a functional assay for RT and does not rely on specific protein or nucleic acid sequences, it performs well when quantifying any HIV-1 subtype [49–51]. A potential concern is that viruses with heavily mutated RT enzymes may be underestimated when compared to a recombinant wild-type RT enzyme, although preliminary studies suggest this may not be a significant problem [49].

Since this assay is newer than the Ultra p24 antigen assay, fewer data are available and most studies have been conducted either by or in collaboration with the manufacturer. Stevens and colleagues [25] compared the Roche RNA assay with both Ultra p24 antigen (external buffer) and the earlier version (version 1.0) of the Cavidi RT assay and found excellent correlation between RNA and RT results. The concentration of p24 antigen, RT, and RNA decreased in all patients after initiating ARV therapy, except in two individuals who had undetectable p24 and RT at baseline.

Crowe and colleagues have tested both the earlier version (version 1.0) and the more sensitive assay (version 2.0) in patients in Australia [49]. The version 1.0 RT assay had detectable HIV-1 RT present in 98% of samples

( $n = 127$ ) with HIV-1 RNA  $>10,000$  copies/ml. Sensitivity using the version 2.0 assay was higher, with detectable HIV RT in 95% of samples ( $n = 69$ ) with HIV RNA  $>1,000$  copies/ml. A positive association was found between the  $\log_{10}$  HIV RNA copies/ml and  $\log_{10}$  HIV RT fg/ml variables using Pearson correlation ( $r = 0.89$ ,  $p < 0.001$ ;  $n = 189$  for version 1.0;  $r = 0.89$ ;  $n = 85$  for version 2.0). The RT activity closely followed the trend for HIV-1 RNA levels in samples ( $n = 4$ – $10$  per patient) from 10 HIV-1 infected patients with progressive disease [48]. Current ( $n = 40$ ), previous ( $n = 30$ ), or no ( $n = 119$ ) exposure to efavirenz had only a minor effect on the RT assay despite the tight binding of efavirenz to the HIV-1 RT enzyme. There was a  $0.20 \log_{10}$  decrease in viral load in the samples with efavirenz resistance mutations, as measured by RT when compared to RT polymerase chain reaction (PCR), indicating a possible decrease in RT fitness [49].

Lombart and colleagues compared HIV-1 RNA (Roche COBAS Amplicor HIV-1 Monitor Test, version 1.5) with Ultra p24 antigen and the Cavidi RT assay, version 1.0 in samples from Burkina Faso [24]. The RT assay detected 0%, 93%, and 100% of samples with viral loads of  $<10,000$ , 10,000–63,000, and  $>80,000$  copies/ml, respectively. Seyoum and colleagues [50] used the version 1.0 RT assay in a prospective study of 26 untreated patients with HIV in Ethiopia (subtype C), to compare results with the NucliSens QT assay. Although only 0.2 ml of plasma were used in this study instead of the 1.0 ml specified in the package insert, there was significant correlation between the two assays ( $n = 178$ ,  $r = 0.65$ ,  $p < 0.001$ ), and patient RT activity in general mirrored HIV-1 RNA changes. Sivapalasingam and colleagues [51] compared results from the version 2.0 RT assay with the ultrasensitive Roche Amplicor Monitor HIV-1 RNA assay version 1.5 on plasma samples from patients in the United States ( $n = 29$ , subtype B) and Cameroon ( $n = 21$ , primarily CRF02\_AG). They found that RT levels correlated significantly with plasma HIV-1 RNA viral loads using Spearman rank correlation (US:  $r = 0.89$ ;  $p < 0.001$ ; Cameroon:  $r = 0.67$ ,  $p < 0.01$ ). Among 32 samples with viral loads  $>2,000$  copies/ml, 97% had detectable RT activity.

## Real-Time PCR/Molecular Beacon Assays

Real-time PCR such as TaqMan and Abbott RealTime, or molecular beacon assays such as Retina Rainbow or NucliSens EasyQ, might be useful for measuring viral load in resource-limited countries [52–65]. Real-time PCR detects amplicon production in real time with each PCR cycle, and thus does not rely on post-amplification detection of amplicons, which helps reduce the possibility of contamination and improves turnaround time. However, commercially available real-time PCR assays are just as expensive as the more standard nucleic acid viral load tests and also use expensive equipment. The use of in-house versions of these assays can help to reduce kit costs, but laboratories must provide their own reagents such as primers and probes, and optimize their methods. In settings with multiple HIV-1 subtypes, care must be taken in selecting reagents and amplification conditions, such as annealing temperatures. Quality assurance of each batch of such reagents remains problematic.

Although the NucliSens EasyQ assay has been evaluated in South Africa [58] and China [59], and the Abbott RealTime assay in Brazil [61], in general these assays do not provide a simple, less expensive alternative to viral load monitoring in resource-limited settings. Low-level contamination and a relatively high frequency of invalid results requiring repeat testing were some of the problems cited with these studies [58].

Many of these obstacles seem to have been overcome in Abidjan, Côte d'Ivoire, where TaqMan real-time PCR is being routinely used for infant diagnosis and patient monitoring [62]. The limit of quantitation for this assay was 300 copies/ml and RNA results were highly correlated with both the Versant ( $n = 327$ ,  $r = 0.901$ ,  $p < 0.001$ ) and the Monitor ( $n = 101$ ,  $r = 0.856$ ,  $p < 0.001$ ) HIV-1 RNA assay. The assay showed 100% sensitivity ( $n = 57$ ) and specificity ( $n = 210$ ) when used in the early diagnosis of infants compared with the Versant assay. The assay also proved useful for monitoring the response to treatment in 36 children and 46 adults who were initiating antiretroviral therapy.

## Solutions to Specimen Collection and Transport in Rural Areas

Viral load testing is typically performed on plasma, which requires a trained phlebotomist, access to a centrifuge, and a  $-70^{\circ}$  C freezer for storage longer than a few days. Dried blood spots are an ideal alternative for specimen collection in rural areas because the blood is obtained either through a heel-stick or finger-stick and applied to filter paper. Once the blood has completely dried, it can be placed in its own gas impermeable ziplock bag along with desiccant and sent to the testing site at ambient temperature.

Dried blood spots have been used successfully to measure HIV-1 RNA in patients [66–73] and to diagnose perinatal HIV infection [45,73–87]. HIV-1 RNA in dried blood spots is stable for at least one year at room temperature [68,70] and HIV-1 DNA may be stable even longer [85,87], although proteins such as antibody and p24 antigen may be more labile [45,88]. Dried blood spots may also be used for HIV-1 subtyping [89] and genotypic resistance testing (S. Cassol, personal communication). A simplified extraction method used in conjunction with the Roche Amplicor DNA Test version 1.5 [85] or real-time PCR [57], and their use in the Ultra p24 antigen assay [45], makes dried blood spots an attractive way to diagnose infected infants early and get them into clinical care. Further options include the use of a new transport medium that stabilizes HIV-1 RNA in plasma [90] or the use of mobile laboratories in rural districts.

### What Is Needed for Implementation of Simplified Viral Load Testing?

First, each country must determine if the assay will quantify subtypes common in the region and is appropriate for the technical staff and laboratory equipment available. Infrastructure limitations impose significant barriers to implementation. Infrastructure includes both physical resources (water, reliable electricity, air conditioning, refrigeration, other equipment) and human resources (trained technologists). Training tools, trainers, guidelines, and consensus protocols, as well as monitoring and evaluation tools, including proficiency testing programs, all need to be established.

Costs must be taken into consideration—not only the costs of the kits, reagents, and supplies, but also the cost of equipment purchase (Does the assay require expensive equipment?), maintenance (Where is the nearest service technician located and how long does it take someone to repair a broken machine, and at what cost?), kit procurement (How long will it take to get kits from the manufacturer or distributor? Will lengthy delays in shipping or customs render the kits useless?), and the cost of labor (What are the hands-on time labor costs for the assay? Is local training available?). Other questions to consider include: Are there established external quality assurance programs available for the assay? Who will provide these? What will they cost? What approval is required by ministries of health?

Once the assay is deemed acceptable, clinical validation studies in relevant populations are required to prove comparability to the gold standard with respect to sensitivity, specificity, precision, reproducibility, dynamic range, and linearity. It is imperative that the laboratory performing the gold standard assay is participating in national or international quality assurance programs in order for the comparisons to be meaningful.

Lastly, there must be acceptance and understanding by the clinicians who will use the new assay. This will require involvement and education of clinicians with results from the clinical validation studies. Communication at a high level between those implementing laboratory monitoring tests and the clinicians advising government on the use of ARV therapy has been lacking. For example, there is no consensus regarding the level of monitoring that is required. Are ranges of values rather than exact values sufficient to provide inexpensive viral load monitoring? Clinician acceptance must play a role in devising a testing strategy for a given country. Endorsement by an international body such as the World Health Organization would likely increase the acceptability of these assays throughout the world.

### Conclusion

The currently available non-nucleic acid-based assays provide better options for resource-limited settings in terms

of cost and physical requirements. However, they require basic laboratory infrastructure and as such do not meet the need for point-of-care tests for patients without access to central or district hospitals. Much simpler tests, based on dip-stick technology [91] or molecular zipper assays [92], are in development. It is hoped that data describing their applicability will be available soon. ■

## Acknowledgments

Other members of the Forum for Collaborative HIV Research Alternative Viral Load Assay Working Group: Pachamuthu Balakrishnan, YRG Care, Chennai, India; Bernie Branson, Centers for Disease Control and Prevention, Atlanta, Georgia; Deborah Burgess, The Bill and Melinda Gates Foundation, Seattle, Washington; Isabel Cabruja, PerkinElmer Life and Analytical Sciences, Turku, Finland; Polly Clayden, HIV i-Base, London, United Kingdom; Kate Condliffe, Clinton HIV/AIDS Initiative, Boston, Massachusetts; Gary Corrigan, Cavid Tech, Uppsala, Sweden; Michel de Baar, Primagen, Amsterdam, The Netherlands; Robert Downing, Centers for Disease Control and Prevention, Entebbe, Uganda; Ernest Ekong, Nigeria HIV/AIDS Research Network, Lagos, Nigeria, Africa; Larry Fox, LabNow, Inc., Austin, Texas; Gregg Gonsalves, Gay Men's Health Crisis, New York City, New York; Mauro Guarinieri, European AIDS Treatment Group, Bologna, Italy; Martine Guillemin, Médecins Sans Frontières, Paris, France; Nick Hellmann, The Bill and Melinda Gates Foundation, Seattle, Washington; Jon Kaplan, Centers for Disease Control and Prevention, Atlanta, Georgia; Helen Lee, Cambridge University, Cambridge, United Kingdom; Robert Lloyd, Research Think Tank, Inc, Alpharetta, Georgia; Carol Major, Ontario HIV Treatment Network, Ontario, Canada; Lynn Margherio, Clinton HIV/AIDS Initiative, Boston, Massachusetts; Javier Martinez-Picado, Universitat Autònoma de Barcelona, Barcelona, Spain; Daniel Newman, Centers for Disease Control and Prevention, Atlanta, Georgia; Trevor Peter, Clinton Foundation, Botswana; Mark Rayfield, Centers for Disease Control and Prevention, Atlanta, Georgia; Renee Ridzon, The Bill and Melinda Gates Foundation, Seattle, Washington; Tobias Rinke De Wit, PharmAccess, Amsterdam, The Netherlands; Jeff Safrin, Elizabeth Glaser Pediatrics AIDS Foundation, Santa Monica, California; Lisa Spacek, Johns Hopkins University, Baltimore, Maryland; Tom Spira, Centers for Disease Control and Prevention, Atlanta, Georgia; Kenji Tamura, World Health Organization, Geneva, Switzerland; Michael Ussery, National Institutes of Health,

Bethesda, Maryland; Fred Valentine, New York University, New York City, New York; Oliviero Varnier, University of Genoa, Genoa, Italy; Rainer Ziermann, Bayer HealthCare, Berkeley, California.

## References

- Schuurman R, Descamps D, Weverling GJ, Kaye S, Tijnagel J, et al. (1996) Multicenter comparison of three commercial methods for quantification of human immunodeficiency virus type 1 RNA in plasma. *J Clin Microbiol* 34: 3016–3022.
- Holguin A, de Mendoza C, Soriano V (1999) Comparison of three different commercial methods for measuring plasma viraemia in patients infected with non-B HIV-1 subtypes. *Eur J Clin Microbiol Infect Dis* 18: 256–259.
- Nolte FS, Boysza J, Thurmond C, Clark WS, Lennox JL (1998) Clinical comparison of an enhanced-sensitivity branched-DNA assay and reverse transcription-PCR for quantitation of human immunodeficiency virus type 1 RNA in plasma. *J Clin Microbiol* 36: 716–720.
- Nkengasong JN, Kalou M, Maurice C, Bile C, Borget MY, et al. (1998) Comparison of NucliSens and Amplicor monitor assays for quantification of human immunodeficiency virus type 1 (HIV-1) RNA in plasma of persons with HIV-1 subtype A infection in Abidjan, Cote d'Ivoire. *J Clin Microbiol* 36: 2495–2498.
- Dyer JR, Pilcher CD, Shepard R, Schock J, Eron JJ, et al. (1999) Comparison of NucliSens and Roche Monitor assays for quantitation of levels of human immunodeficiency virus type 1 RNA in plasma. *J Clin Microbiol* 37: 447–449.
- Elbeik T, Charlebois E, Nassos P, Kahn J, Hecht FM, et al. (2000) Quantitative and cost comparison of ultrasensitive human immunodeficiency virus type 1 RNA viral load assays: Bayer bDNA quantiplex versions 3.0 and 2.0 and Roche PCR Amplicor monitor version 1.5. *J Clin Microbiol* 38: 1113–1120.
- Murphy DG, Cote L, Fauvel M, Rene P, Vincelette J (2000) Multicenter comparison of Roche COBAS AMPLICOR MONITOR version 1.5, Organon Teknika NucliSens QT with Extractor, and Bayer Quantiplex version 3.0 for quantification of human immunodeficiency virus type 1 RNA in plasma. *J Clin Microbiol* 38: 4034–4041.
- Elbeik T, Delassandro R, Chen Y-MA, Southkov SV, Loftus RA, et al. (2003) Global cost modeling analysis of HIV-1 and HCV viral load assays. *Expert Rev Pharmacoeconomics Outcomes Res* 3: 383–407.
- Crowe S, Turnbull S, Oelrichs R, Dunne A (2003) Monitoring of HIV infection in resource-constrained countries. *Clin Infect Dis* 37 (Suppl 1): S25–S35.
- Majchrowicz M (2003) Beyond antiretroviral access: Low-cost laboratory tests needed for the developing world. *AIDS* 17 (Suppl 4): S12–S15.
- Balakrishnan P, Solomon S, Kumarasamy N, Mayer KH (2005) Low-cost monitoring of HIV infected individuals on highly active antiretroviral therapy (HAART) in developing countries. *Indian J Med Res* 121: 245–355.
- Giacomini M, McDermott JL, Giri AA, Martini I, Lillo FV, et al. (1998) A novel and improved quantitative kinetic software for virologic colorimetric assays. *J Virol Methods* 73: 201–209.
- Schupbach J, Boni J (1993) Quantitative and sensitive detection of immune-complex-dissociated HIV p24 antigen after boiling. *J Virol Methods* 43: 247–256.
- Schupbach J, Boni J, Tomasik Z, Jendis J, Seger R, et al. (1994) Sensitive detection and early prognostic significance of p24 antigen in heat-denatured plasma of human immunodeficiency virus type 1-infected infants. *Swiss Neonatal HIV Study Group. J Infect Dis* 170: 318–324.
- Schupbach J, Flepp M, Pontelli D, Tomasik Z, Luthy R, et al. (1996) Heat-mediated immune complex dissociation and enzyme-linked immunosorbent assay signal amplification render p24 antigen detection in plasma as sensitive as HIV-1 RNA detection by polymerase chain reaction. *AIDS* 10: 1085–1090.
- Boni J, Opravil M, Tomasik Z, Rothen M, Bisset L, et al. (1997) Simple monitoring of antiretroviral therapy with a signal amplification-boosted HIV-1 p24 antigen assay with heat-denatured plasma. *AIDS* 11: F47–F52.
- Ledergerber B, Flepp M, Boni J, Tomasik Z, Come RW, et al. (2000) HIV-1 p24 concentration measured by boosted ELISA of heat-denatured plasma correlates with decline in CD4 cells, progression to AIDS, and survival: Comparison with vira RNA measurement. *J Infect Dis* 181: 1280–1288.
- Burgisser P, Vernazza P, Flepp M, Boni J, Tomasik Z, et al. (2000) Performance of five different assays for the quantification of viral load in persons infected with various subtypes of HIV-1. *Swiss HIV Cohort Study. J Acquir Immune Defic Syndr* 23: 138–144.
- Schupbach J, Boni J, Flepp M, Tomasik Z, Joller H, et al. (2001) Antiretroviral treatment monitoring with an improved HIV-1 p24 antigen test: An inexpensive alternative to tests for viral RNA. *J Med Virol* 65: 225–232.
- Sterling TR, Hoover DR, Astemborski J, Vlahov D, Bartlett JA, et al. (2002) Heat-denatured HIV-1 protein p24 antigen: Prognostic value in adults with early stage disease. *J Infect Dis* 186: 1181–1185.
- Schupbach J, Boni J, Bisset LR, Tomasik Z, Fischer M, et al. (2003) HIV-1 p24 antigen is a significant inverse correlate of CD4 T-cell change in patients with suppressed viremia under long-term antiretroviral therapy. *J Acquir Immune Defic Syndr* 33: 292–299.
- Schupbach J, Gunthard H, Joos B, Fischer M, Boni J, et al. (2005) HIV-1 p24 may persist during long-term highly active antiretroviral therapy, increases little during short treatment breaks, and its rebound after treatment stop correlates with CD4+ T cell loss. *J Acquir Immune Defic Syndr* 40: 250–256.
- Pascual A, Cachafeiro A, Funk ML, Fiscus SA (2002) Comparison of an assay using signal amplification of the heat-dissociated p24 antigen with the Roche Monitor HIV RNA assay. *J Clin Microbiol* 40: 2472–2475.
- Lombart JP, Vray M, Kafando A, Lécène V, Ouédraogo-Taroe R, et al. (2005) Plasma viron reverse transcriptase activity and heat dissociation-boosted p24 antigen assay for HIV load in Burkino Faso, West Africa. *AIDS* 19: 1273–1277.
- Stevens G, Rekhviashvili N, Scott LE, Gonin R, Stevens W (2005) Evaluation of two commercially available, inexpensive alternative HIV-1 viral load assays for assessing viral load in a cohort of HIV-1 infected subtype C patients from South Africa. *J Clin Microbiol* 43: 857–861.
- Ribas SG, Ondo P, Schupbach J, van der Groen G, Fransen K (2003) Performance of a quantitative HIV-1 p24 antigen assay on various HIV-1 subtypes for the follow-up of HIV-1 seropositive individuals. *J Virol Methods* 113: 29–34.
- Prado JG, Shintani A, Boffill M, Clotet B, Ruiz L, et al. (2004) Lack of longitudinal intrapatient correlation between p24 antigenemia and levels of HIV-1 RNA in patients with chronic HIV infection during structured treatment interruptions. *J Clin Microbiol* 42: 1620–1625.
- Bonard D, Rouet F, Toni TA, Minga A, Huet C, et al. (2003) Field evaluation of an improved assay using a heat-dissociated p24 antigen for adults mainly infected with HIV-1 CRF02\_AG strains in Cote d'Ivoire, West Africa. *J Acquir Immune Defic Syndr* 34: 267–273.
- Respass RA, Cachafeiro A, Withum D, Fiscus SA, Newman D, et al. (2005) Evaluation of an

- ultrasensitive p24 antigen assay as a potential alternative to HIV-1 RNA viral load in resource-limited settings. *J Clin Microbiol* 43: 506–508.
30. Jennings C, Danilovic A, Burgess D, Morack R, Fiscus S, et al. A comparison of two non-RNA-based assays for the quantitation of HIV [abstract]. 43rd Interscience Conference on Antimicrobial Agents and Chemotherapy; 2003 September 14–17; Chicago, Illinois, United States of America. American Society for Microbiology. Available: [http://gateway.nlm.nih.gov/robot\\_pages/MeetingAbstracts/102266490.html](http://gateway.nlm.nih.gov/robot_pages/MeetingAbstracts/102266490.html). Accessed 7 September 2006.
  31. Simonds RJ, Brown TM, Thea DM, Orloff SL, Steketee RW, et al. (1998) Sensitivity and specificity of a qualitative RNA detection assay to diagnose HIV infection in young infants. *Perinatal AIDS Collaborative Transmission Study. AIDS* 12: 1545–1549.
  32. Cunningham CK, Charbonneau TT, Song K, Patterson D, Sullivan T, et al. (1999) Comparison of human immunodeficiency virus 1 DNA polymerase chain reaction and qualitative and quantitative RNA polymerase chain reaction in human immunodeficiency virus 1-exposed infants. *Pediatr Infect Dis J* 18: 30–35.
  33. Young NL, Shaffer N, Chaowanachan T, Chotpitayasonondh T, Vanparapar N, et al. (2000) Early diagnosis of HIV-1-infected infants in Thailand using RNA and DNA PCR assays sensitive to non-B subtypes. *J Acquir Immune Defic Syndr* 24: 401–407.
  34. Benjamin DK Jr, Miller WC, Fiscus SA, Benjamin DK, Morse M, et al. (2001) Rational testing of the HIV-exposed infant. *Pediatrics* 108: E3.
  35. Lambert JS, Harris DR, Stiehm ER, Moye JJr, Fowler MG, et al. (2003) Performance characteristics of HIV-1 culture and HIV-1 DNA and RNA amplification assays for early diagnosis of perinatal HIV-1 infection. *J Acquir Immune Defic Syndr* 34: 512–519.
  36. World Health Organization (2004) Scaling up antiretroviral therapy in resource-limited settings: Treatment guidelines for a public health approach. 2003 revision. Available: [http://www.who.int/3by5/publications/documents/arv\\_guidelines/en/](http://www.who.int/3by5/publications/documents/arv_guidelines/en/). Accessed 6 September 2006.
  37. Lyamuya E, Bredberg-Raden U, Massawe A, Urassa E, Kawo G, et al. (1996) Performance of a modified HIV-1 p24 antigen assay for early diagnosis of HIV-1 infection in infants and prediction of mother-to-infant transmission of HIV-1 in Dar es Salaam, Tanzania. *J Acquir Immune Defic Syndr* 12: 421–426.
  38. Nadal D, Boni J, Kind C, Vanirier OE, Steiner F, et al. (1999) Prospective evaluation of amplification-boosted ELISA for heat-denatured p24 antigen for diagnosis and monitoring of pediatric HIV-1 infections. *J Infect Dis* 180: 1089–1095.
  39. Sutthent R, Gaudart N, Chokpaibulkit K, Tanliang N, Kanokinsombath C, et al. (2003) p24 antigen detection assay modified with a booster step for diagnosis and monitoring of HIV-1 infection. *J Clin Microbiol* 41: 1016–1022.
  40. Sherman GG, Stevens G, Stevens W (2004) Affordable diagnosis of HIV infection in infants by p24 antigen detection. *Pediatr Infect Dis J* 23: 173–176.
  41. Zijenah LS, Tobaiwa O, Rusakaniko S, Nathoo KJ, Nhembe M, et al. (2005) Signal-boosted qualitative ultrasensitive p24 antigen assay for diagnosis of subtype C HIV-1 infection in infants under the age of 2 years. *J Acquir Immune Defic Syndr* 39: 391–394.
  42. De Baets AJ, Edidi BS, Kasali MJ, Beelaert G, Schrooten W, et al. (2005) Pediatric human immunodeficiency virus screening in an African district hospital. *Clin Diagn Lab Immunol* 12: 86–92.
  43. Knuchel MC, Tomasik Z, Speck RF, Luthy R, Schubach J (2006) Ultrasensitive quantitative HIV-1 p24 antigen assay adapted to dried plasma spots to improve treatment monitoring in low-resource settings. *J Clin Virol* 36: 64–67.
  44. Li CC, Seidel KD, Coombs RW, Frenkel LM (2005) Detection and quantification of HIV-1 p24 antigen in dried whole blood and plasma on filter paper stored under various conditions. *J Clin Microbiol* 43: 3901–3905.
  45. Patton JC, Sherman GG, Coovadia AH, Stevens WS, Meyers TM (2006) Ultrasensitive HIV-1 p24 antigen assay modified for use on dried whole blood spots as a reliable, affordable test for infant diagnosis. *Clin Vaccine Immunol* 13: 152–155.
  46. Malmsten A, Shao XW, Aperia K, Corrigan GE, Sandstrom E, et al. (2003) HIV-1 viral load determination based on reverse transcriptase activity recovered from human plasma. *J Med Virol* 71: 347–359.
  47. Malmsten A, Shao XW, Sjudahl S, Fredriksson EL, Petterson I, et al. (2005) Improved HIV-1 viral load determination based on reverse transcriptase activity recovered from human plasma. *J Med Virol* 76: 291–296.
  48. Braun J, Plantier JC, Hellot MF, Tuallon E, Gueudin M, et al. (2003) A new quantitative HIV load assay based on plasma viron reverse transcriptase activity for the different types, groups and subtypes. *AIDS* 17: 331–336.
  49. Greengrass VL, Turnbull SP, Hocking J, Dunne AL, Tachedjian G, et al. (2005) Evaluation of a low cost reverse transcription assay for plasma HIV-1 viral load monitoring. *Curr HIV Res* 3: 183–190.
  50. Seyoum E, Wolday D, Girma M, Malmsten A, Meselle T, et al. (2006) Reverse transcriptase activity for quantitation of HIV-1 subtype C in plasma: Relation to RNA copy number and CD4 T-cell count. *J Med Virol* 78: 161–168.
  51. Sivapalasingam S, Essajee S, Nyambi PN, Itri V, Hanna B, et al. (2005) Human immunodeficiency virus (HIV) reverse transcriptase activity correlates with HIV RNA load: Implications for resource-limited settings. *J Clin Microbiol* 43: 3793–3796.
  52. Lewin SR, Vesanan M, Kostriks L, Hurley A, Duran M, et al. (1999) Use of real-time PCR and molecular beacons to detect virus replication in HIV-1 infected individuals on prolonged effective antiretroviral therapy. *J Virol* 73: 6099–6103.
  53. de Baar MP, Timmermans EC, Bakker M, de Rooij E, van Gemen B, et al. (2001) One-tube real-time isothermal amplification assay to identify and distinguish human immunodeficiency virus type 1 subtypes A, B, and C and circulating recombinant forms AE and AG. *J Clin Microbiol* 39: 1895–1902.
  54. de Baar MP, Timmermans EC, Buitelaar M, Westrop M, Dekker JT, et al. (2003) Evaluation of the HIV-1 RNA RetinaTM Rainbow assay on plasma and dried plasma spots: Correlation with the Roche Amplicor HIV-1 Monitor v 1.5 assay. *Antivir Ther* 8 (Suppl 1): S533.
  55. Palmer S, Wiegand AP, Maldarelli F, Bazmi H, Mican JM, et al. (2003) New real-time reverse transcriptase-initiated PCR assay with single-copy sensitivity for human immunodeficiency virus type 1 RNA in plasma. *J Clin Microbiol* 41: 4531–4536.
  56. Ayele W, Pollakis G, Abebe A, Fisseha B, Tegbaru B, et al. (2004) Development of a nucleic acid sequence-based amplification assay that uses gag-based molecular beacons to distinguish between human immunodeficiency virus type 1 subtype C and C' infections in Ethiopia. *J Clin Microbiol* 42: 1534–1541.
  57. Luo W, Yang H, Rathbun K, Pau CP, Ou CY (2005) Detection of HIV-1 DNA in dried blood spots by a duplex real-time PCR assay. *J Clin Microbiol* 43: 1851–1857.
  58. Stevens W, Wiggill T, Horsfield P, Coetzee L, Scott LE (2005) Evaluation of the NucliSens EasyQ assay in HIV-1 infected individuals in South Africa. *J Virol Methods* 124: 105–110.
  59. Yao J, Liu Z, Ko LS, Pan G, Jiang Y (2005) Quantitative detection of HIV-1 RNA using NucliSens EasyQ HIV-1 assay. *J Virol Methods* 129: 40–46.
  60. Rekhviashvili N, Stevens G, Scott L, Stevens W (2006) Fluorogenic LUX primer for quantitation of HIV-1 by real-time RT-PCR. *Mol Biotechnol* 32: 101–110.
  61. Swanson P, Huang S, Holzmayer V, Bodelle P, Yamaguchi J, et al. (2006) Performance of the automated Abbott RealTime HIV-1 assay on a genetically diverse panel of specimens from Brazil. *J Virol Methods* 134: 237–243.
  62. Rouet F, Ekouevi DK, Chaix ML, Burgard N, Inwoley A, et al. (2005) Transfer and evaluation of an automated, low-cost real-time reverse transcription-PCR test for diagnosis and monitoring of HIV-1 infection in a West African resource-limited setting. *J Clin Microbiol* 43: 2709–2717.
  63. Gueudin M, Simon F (2005) Plasma RNA viral load in HIV-1 group O infection by real-time PCR. *Methods Mol Biol* 304: 221–228.
  64. Gueudin M, Plantier JC, Damond F, Roques P, Maucle P, et al. (2003) Plasma viral RNA assay in HIV-1 group O infection by real-time PCR. *J Virol Methods* 113: 43–49.
  65. Candotti D, Temple J, Owusu-Ofori S, Allain JP (2004) Multiplex real-time quantitative real-time PCR assay for hepatitis B virus, hepatitis C virus, and human immunodeficiency virus type 1. *J Virol Methods* 118: 39–47.
  66. Cassol S, Gill MJ, Pilon R, Cormier M, Voigt RF, et al. (1997) Quantification of human immunodeficiency virus type 1 RNA from dried plasma spots collected on filter paper. *J Clin Microbiol* 35: 2795–2801.
  67. Fiscus SA, Brambilla D, Grosso L, Schock J, Cronin M (1998) Quantitation of human immunodeficiency virus type 1 RNA in plasma by using blood dried on filter paper. *J Clin Microbiol* 36: 258–260.
  68. O'Shea S, Mullen J, Corbett K, Christie I, Newell ML, et al. (1999) Use of dried whole blood spots for quantification of HIV-1 RNA. *AIDS* 13: 630–631.
  69. Biggar RJ, Broadhead R, Janes M, Kumwenda N, Taha TE, et al. (2001) Viral levels in newborn African infants undergoing primary HIV-1 infection. *AIDS* 15: 1311–1313.
  70. Brambilla D, Jennings C, Aldrovandi G, Bremer J, Comeau AM, et al. (2003) Multicenter evaluation of use of dried blood and plasma spot specimens in quantitative assays for human immunodeficiency virus RNA: measurement, precision, and RNA stability. *J Clin Microbiol* 41: 1888–1893.
  71. Mwaba P, Cassol S, Nunn A, Pilon R, Chintu C, et al. (2003) Whole blood versus plasma spots for measurement of HIV-1 viral load in HIV-infected African patients. *Lancet* 362: 2067–2068.
  72. Uttayamakul S, Likansonsakul S, Sunthornkachit R, Kuntiranont K, Luoisirirotchankul S, et al. (2005) Usage of dried blood spots for molecular diagnosis and monitoring HIV-1 infection. *J Virol Methods* 128: 128–134.
  73. Rollins NC, Dedicoat M, Danaviah S, Page T, Bishop K, et al. (2002) Prevalence, incidence, and mother-to-child transmission of HIV-1 in rural South Africa. *Lancet* 360: 389.
  74. Cassol S, Salas T, Arella M, Neumann P, Schechter MT, et al. (1991) Use of dried blood spot specimens in the detection of human immunodeficiency virus type 1 by the polymerase chain reaction. *J Clin Microbiol* 29: 667–671.
  75. Cassol SA, Lapointe N, Salas T, Hankins C, Arella M, et al. (1992) Diagnosis of vertical HIV-1 transmission using the polymerase chain reaction and dried blood spot specimens. *J Acquir Immune Defic Syndr* 5: 113–119.

76. Comeau AM, Harris J, McIntosh K, Weiblen BJ, Hoff R, et al. (1992) Polymerase chain reaction in detecting HIV-1 infection among seropositive infants: Relation to clinical status and age and to results of other assays. *J Acquir Immune Defic Syndr* 5: 271–278.
77. Cassol S, Butcher A, Kinard S, Spadaro J, Sy T, et al. (1994) Rapid screening for early detection of mother-to-child transmission of human immunodeficiency virus type 1. *J Clin Microbiol* 32: 2641–2645.
78. Yourno J, Conroy J (1992) A novel polymerase chain reaction method for detection of human immunodeficiency virus in dried blood spots on filter paper. *J Clin Microbiol* 30: 2887–2892.
79. Comeau AM, Hsu HW, Schwerzler M, Mushinsky G, Walter E, et al. (1993) Identifying human immunodeficiency virus infection at birth: Application of polymerase chain reaction to Guthrie cards. *J Pediatr* 123: 252–258.
80. Nyambi PN, Fransen K, De Beenhouwer H, Chomba EN, Temmerman M, et al. (1994) Detection of human immunodeficiency virus type 1 (HIV-1) in heel prick blood on filter paper from children born to HIV-1-seropositive mothers. *J Clin Microbiol* 32: 2858–2860.
81. Biggar RJ, Miley W, Miotti P, Taha TE, Butcher A, et al. (1997) Blood collection on filter paper: A practical approach to sample collection for studies of perinatal HIV transmission. *J Acquir Immune Defic Syndr* 14: 368–373.
82. Panteleeff DD, John G, Nduati R, Mbori-Ngacha D, Richardson B, et al. (1999) Rapid method for screening dried blood samples on filter paper for human immunodeficiency virus type 1 DNA. *J Clin Microbiol* 37: 350–353.
83. Beck IA, Drennan KD, Melvin AJ, Mohan KM, Herz AM, et al. (2001) Simple, sensitive, and specific detection of human immunodeficiency virus type 1 subtype B DNA in dried blood samples for diagnosis in infants in the field. *J Clin Microbiol* 39: 29–33.
84. Fischer A, Lejczak C, Lambert C, Servais J, Makombe N, et al. (2004) Simple DNA extraction method for dried blood spots and comparison of two PCR assays for diagnosis of vertical human immunodeficiency virus type 1 transmission in Rwanda. *J Clin Microbiol* 42: 16–20.
85. Sherman GG, Stevens G, Jones SA, Horsfield P, Stevens WS (2005) Dried blood spots improve access to HIV diagnosis and care for infants in low-resource settings. *J Acquir Immune Defic Syndr* 38: 615–617.
86. Sriwanthana B, Wetprasit N, Chareonsook S, Janejai N, Chareonsiriwatana W (2003) A study to implement early diagnosis of HIV infection in infants born to infected mothers. *Southeast Asian J Trop Med Public Health* 34 (Suppl 3): 221–226.
87. Li CC, Beck IA, Seidel KD, Frenkel LM (2004) Persistence of human immunodeficiency virus type 1 subtype B DNA in dried-blood samples on FTA filter paper. *J Clin Microbiol* 42: 3847–3849.
88. Behets F, Kashamuka M, Pappaioanou M, Green TA, Ryder RW, et al. (1992) Stability of human immunodeficiency virus type 1 antibodies in whole blood dried on filter paper and stored under various tropical conditions in Kinshasa, Zaire. *J Clin Microbiol* 30: 1179–1182.
89. Cassol S, Weniger BG, Babu PG, Salminen MO, Zheng X, et al. (1996) Detection of HIV type 1 env subtypes A, B, C, and E in Asia using dried blood spots: A new surveillance tool for molecular epidemiology. *AIDS Res Hum Retroviruses* 12: 1435–1441.
90. Lee DH, Li L, Andrus L, Prince AM (2002) Stabilized viral nucleic acids in plasma as an alternative shipping method for NAT. *Transfusion* 42: 409–413.
91. Dineva MA, Candotti D, Fletcher-Brown F, Allain JP, Lee H (2005) Simultaneous visual detection of multiple viral amplicons by dipstick assay. *J Clin Microbiol* 43: 4015–4021.
92. Guttikonda S, Wang W, Suresh M (2004) Molecular zipper assays: A simple homosandwich with the sensitivity of PCR. *J Pharm Pharm Sci* 7: 7–16.

