

University of Groningen

A Low-Cost 3-D Printed Smartphone Add-on Spectrometer for Diagnosis of Crop Diseases in Field

Owomugisha, Godliver; Mugagga, Pius K. B.; Melchert, Friedrich; Mwebaze, Ernest; Quinn, John A.; Biehl, Michael

Published in:

Proceedings of the 3rd ACM SIGCAS Conference on Computing and Sustainable Societies

DOI:

[10.1145/3378393.3402252](https://doi.org/10.1145/3378393.3402252)

IMPORTANT NOTE: You are advised to consult the publisher's version (publisher's PDF) if you wish to cite from it. Please check the document version below.

Document Version

Publisher's PDF, also known as Version of record

Publication date:

2020

[Link to publication in University of Groningen/UMCG research database](#)

Citation for published version (APA):

Owomugisha, G., Mugagga, P. K. B., Melchert, F., Mwebaze, E., Quinn, J. A., & Biehl, M. (2020). A Low-Cost 3-D Printed Smartphone Add-on Spectrometer for Diagnosis of Crop Diseases in Field. In *Proceedings of the 3rd ACM SIGCAS Conference on Computing and Sustainable Societies* (pp. 331–332). Association for Computing Machinery. <https://doi.org/10.1145/3378393.3402252>

Copyright

Other than for strictly personal use, it is not permitted to download or to forward/distribute the text or part of it without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license (like Creative Commons).

The publication may also be distributed here under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license. More information can be found on the University of Groningen website: <https://www.rug.nl/library/open-access/self-archiving-pure/taverne-amendment>.

Take-down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Downloaded from the University of Groningen/UMCG research database (Pure): <http://www.rug.nl/research/portal>. For technical reasons the number of authors shown on this cover page is limited to 10 maximum.

A low-cost 3-D printed smartphone add-on spectrometer for diagnosis of crop diseases in field

Godliver Owomugisha^{1,2,3}, Pius K. B. Mugagga², Friedrich Melchert^{1,4}, Ernest Mwebaze², John A. Quinn², Michael Biehl¹

1. University of Groningen, Bernoulli Institute for Mathematics, Computer Science and Artificial Intelligence, P.O. Box 407, 9700 AK Groningen, The Netherlands

2. College of Computing & Information Sciences, Makerere University, P. O. Box 7062, Kampala, Uganda

3. Busitema University, Faculty of Engineering, P. O. Box 236, Tororo, Uganda

4. Fraunhofer Institute for Factory Operation & Automation IFF Sandtorstrasse 22, 39106 Magdeburg, Germany
[g.owomugisha,m.biehl]@rug.nl,friedrich.melchert@iff.fraunhofer.de,piuskavz@gmail.com,[emwebaze,jquinn]@cit.ac.ug

ABSTRACT

We present our initial proof of concept study towards the development of a low-cost 3-D printed smartphone add-on spectrometer. The study aimed at developing a cheap technology (less than 5 USD) to be used for detection of crop diseases in the field using spectrometry. Previously, we experimented with the problem of disease diagnosis using an off-the-shelf and expensive spectrometer (approximately 1000 USD). However, in real world practice, this off-the-shelf device can not be used by typical users (smallholder farmers). Therefore, the study presents a tool that is cheap and user friendly. We present preliminary results and identify requirements for a future version aiming at an accurate diagnostic technology to be used in the field before disease symptoms are visibly seen by the naked eye. Evaluation shows performance of the tool is better than random however below performance of an industry grade spectrometer.

CCS CONCEPTS

- Applied computing → Physical sciences and engineering;
- Computing methodologies → Machine learning.

KEYWORDS

Low-cost, Spectrometry, Crop disease, Diagnosis, 3-D printed, Smartphone

ACM Reference Format:

Godliver Owomugisha^{1,2,3}, Pius K. B. Mugagga², Friedrich Melchert^{1,4}, Ernest Mwebaze², John A. Quinn², Michael Biehl¹. 2020. A low-cost 3-D printed smartphone add-on spectrometer for diagnosis of crop diseases in field. In *ACM SIGCAS Conference on Computing and Sustainable Societies (COMPASS) (COMPASS '20)*, June 15–17, 2020, , Ecuador. ACM, New York, NY, USA, 2 pages. <https://doi.org/10.1145/3378393.3402252>

1 INTRODUCTION

The ability to detect crop disease in-field at a very early stage, especially amongst smallholder farmers, is still a big challenge.

Permission to make digital or hard copies of part or all of this work for personal or classroom use is granted without fee provided that copies are not made or distributed for profit or commercial advantage and that copies bear this notice and the full citation on the first page. Copyrights for third-party components of this work must be honored. For all other uses, contact the owner/author(s).

COMPASS '20, June 15–17, 2020, , Ecuador

© 2020 Copyright held by the owner/author(s).

ACM ISBN 978-1-4503-7129-2/20/06.

<https://doi.org/10.1145/3378393.3402252>

Usually disease identification happens at a stage when it is too late and little can be done to save the situation. This research is an extension of our previous work [5, 6] where diseases in Cassava crop have been identified at an early stage using spectral data acquired by a commercially available spectrometer.

Initially, we analysed spectral data from visibly diseased parts of a leaf as well as parts that appear visibly healthy [5]. Findings of the study indicated that spectral data collected from asymptomatic parts of the leaf has implications for detecting disease in the plants before symptoms are visible.

In [6], we carried out the experiments in a screen house, to rule out the influence of other diseases, pests or severe weather conditions. The results of the study indicate that the presence of the disease can be detected from leaf spectra six weeks earlier before the appearance of visual symptoms.

Although, good results have been attained with this approach, the ability to construct cheap and usable tools to diagnose crop diseases at an early stage is still vital. To tackle this problem, we propose the development of a 3-D printed smartphone add-on spectrometer that can determine the status of a crop from the leaves in field. There have been some initiatives to build a small relatively cheap spectrometer to do these types of analyses e.g. [11], [3]. However, most of the studies on the application of this novel technology are carried out in isolation with no comprehensive information on the most suitable approach.

The key contributions of this research can be seen in two areas. (i) To provide evidence of building a really low-cost technology that can be easily scaled even without a large production line. (ii). To develop an artifact that will actually be able to identify disease early such that smallholder farmers can cause an intervention to their garden. We expect that success in this endeavor will result in a large trial to validate its applicability for use in early diagnosis of disease in the field before it is visibly symptomatic on the plant.

2 MATERIALS AND METHODS

In this section, we describe the setup for the customised spectrometer we constructed in this study. For previous work, we experimented with a commercial device [2].

2.1 Customised spectrometer design

2.1.1 Light source & Diffraction grating. The customised setup in this experiment used a watch battery of capacity 1.5 V and white

light diode as a source of light as a low cost alternative to optical fibre and more powerful light sources. We used a peeled digital versatile disc (DVD) piece as the diffraction grating medium [8].

2.1.2 The digital display. We use a smartphone camera as a receiver. The light split by the DVD falling on the camera is captured as spectrum image. The study also acquired spectral data using the AspectraMini Application [4] defined above. Our hypothesis in this study is that different diseases affect the metabolism of plants differently thus changes in light absorption properties.

2.1.3 Hardware design. We built a 3-D printed smartphone casing to integrate different components and our design aimed at the following: (i). To prevent much light scattering and splitting in order to retain enough information. (ii). To keep external light out of the experiment as to avoid information distortion. We also made the case adjustable to any phone controlling for where the camera is and the size of the phone. Figure 1 shows a full setup.



Figure 1: First prototype

2.2 Methods

We acquired data of two types: (i). Color histograms as a transformation from color RGB spectra. (ii). Spectral data using the already existing software (Aspectra mini Android application) [4].

Plants were grown in a screen house and at week four (4) of growth, they were split into two (2) separate groups. The first one (10 plants) was reserved as a healthy control class (HC) and no disease inoculation was applied to the group. The second group of plants (17 plants) was infected with the cassava brown streak disease (CBSD) virus using grafting inoculation technique [9].

For spectral data we follow the same pre-processing methodology in [6]. We applied computer vision techniques to transform the image spectrum from RGB to HSV color space and attained histograms as image representation. Data from this experiment and the adapter design for the 3D-printed smartphone case and are available at <https://github.com/godliver/3-D-Printouts.git>.

3 RESULTS

We follow the same training and validation strategy as previous [6]. We used standard algorithms (KNN, Extra Trees, Linear SVM) defined under Scikit-learn [7] and also GMLVQ algorithm [10] via the open source GMLVQ toolbox [1]. Result Table 1 shows results for this pilot study.

4 DISCUSSION

We have presented an initial step towards the construction of an innovative low-cost spectrometer that can be used to diagnose

Table 1: Overall accuracy score for customised spectrometer (a, b) vs. Commercial spectrometer (c)

Classifier	Aspectra Mini (a)	Color Histograms (b)	CI-710 device (c)
KNN	0.627	0.521	0.695
Extra Trees	0.671	0.556	0.748
LinearSVM	0.612	0.503	0.812
GMLVQ	0.570	0.522	0.848

disease in plants and infield. Our novel contribution in this area can be seen in the design of the prototype. Previous work [5, 6] showed that spectrometry can gain the smallholder farmer an extra 8 weeks to apply an intervention before disease symptoms become visible. Our experiments in this study aimed at replacing the expensive spectrometer (\$10K) with a cheap version to cost \$5 – \$8. While performance is clearly inferior to the one of the commercial spectrometer, we observe performance above mere guessing. This forms the basis for further improvements. One element to include in a future version is a powerful diffraction grating medium e.g. a prism. We will also experiment with different diodes. The explicit transformation of light emissions to actual spectrograms should facilitate further improvements. Literature [8] also shows efficacy for this type of handcrafted cheap DIY tools. We intend to leverage on that success to provide a diagnostic tool to be used by smallholder farmers in developing countries.

REFERENCES

- [1] Michael Biehl. 2017. A no-nonsense GMLVQ Toolbox, version 2.3. *University of Groningen, The Netherlands*. (2017). <http://www.cs.rug.nl/~biehl/gmlvq>
- [2] CID Bio-Science, Inc. 2010. CI-710 MINIATURE LEAF SPECTROMETER. <http://www.cid-inc.com>
- [3] Joseph Fennell, Charles Veys, Jose Dingle, Joachim Nwezeobi, Sharon van Brunshot, John Colvin, and Bruce Grieve. 2018. A method for real-time classification of insect vectors of mosaic and brown streak disease in cassava plants for future implementation within a low-cost, handheld, in-field multispectral imaging sensor. *Plant Methods* 14 (12 2018). <https://doi.org/10.1186/s13007-018-0350-3>
- [4] Google. 2017. AspectraMini Mobile Application. (2017). <https://play.google.com/store>
- [5] G. Owomugisha, F. Melchert, E. Mwebaze, J. A Quinn, and M. Biehl. 2018. Machine Learning for diagnosis of disease in plants using spectral data. *The 20th International Conference on Artificial Intelligence (2018)*, 9–15.
- [6] Godliver Owomugisha, Ephraim Nuwamanya, John A. Quinn, Michael Biehl, and Ernest Mwebaze. 2020. Early Detection of Plant Diseases Using Spectral Data. In *Proceedings of the 3rd International Conference on Applications of Intelligent Systems (Las Palmas de Gran Canaria, Spain) (APPIS 2020)*. Association for Computing Machinery, New York, NY, USA, Article 26, 6 pages. <https://doi.org/10.1145/3378184.3378222>
- [7] F. Pedregosa, G. Varoquaux, A. Gramfort, V. Michel, B. Thirion, O. Grisel, M. Blondel, P. Prettenhofer, R. Weiss, V. Dubourg, J. Vanderplas, A. Passos, D. Cournapeau, M. Brucher, M. Perrot, and E. Duchesnay. 2011. Scikit-learn: Machine Learning in Python. *Journal of Machine Learning Research* 12 (2011), 2825–2830.
- [8] Public Lab. 2019. Spectrometry. *Public Lab spectrometry* (2019). www.publiclab.org
- [9] GM Rwegasira and Chrissie MER. 2015. Efficiency of non-vector methods of cassava brown streak virus transmission to susceptible cassava plants. *African Journal of Food, Agriculture, Nutrition and Development* (2015).
- [10] Petra Schneider, Michael Biehl, and Barbara Hammer. 2009. Adaptive Relevance Matrices in Learning Vector Quantization. *Neural Comput.* 21, 12 (2009), 3532–3561.
- [11] StellarNet Inc. 2019. Low Cost Spectrometers. *Spectrometers* (2019). www.stellarnet.us