

University of Wollongong

Research Online

---

Faculty of Engineering and Information  
Sciences - Papers: Part A

Faculty of Engineering and Information  
Sciences

---

1-1-2014

## Characterization of MOSkin detector for in vivo skin dose measurement during megavoltage radiotherapy

Weiloong Jong  
*University of Malaya*

Jeannie Hsiu Ding Wong  
*University of Malaya*

Ngiemin Ung  
*University of Malaya*

Kwanhoong Ng  
*University of Malaya*

Gwofuang Ho  
*University of Malaya*

*See next page for additional authors*

Follow this and additional works at: <https://ro.uow.edu.au/eispapers>



Part of the [Engineering Commons](#), and the [Science and Technology Studies Commons](#)

---

### Recommended Citation

Jong, Weiloong; Wong, Jeannie Hsiu Ding; Ung, Ngiemin; Ng, Kwanhoong; Ho, Gwofuang; Cutajar, Dean; and Rosenfeld, Anatoly B., "Characterization of MOSkin detector for in vivo skin dose measurement during megavoltage radiotherapy" (2014). *Faculty of Engineering and Information Sciences - Papers: Part A*. 2888.

<https://ro.uow.edu.au/eispapers/2888>

Research Online is the open access institutional repository for the University of Wollongong. For further information contact the UOW Library: [research-pubs@uow.edu.au](mailto:research-pubs@uow.edu.au)

---

# Characterization of MOSkin detector for in vivo skin dose measurement during megavoltage radiotherapy

## Abstract

In vivo dosimetry is important during radiotherapy to ensure the accuracy of the dose delivered to the treatment volume. A dosimeter should be characterized based on its application before it is used for in vivo dosimetry. In this study, we characterize a new MOSFET-based detector, the MOSkin detector, on surface for in vivo skin dosimetry. The advantages of the MOSkin detector are its water equivalent depth of measurement of 0.07 mm, small physical size with submicron dosimetric volume, and the ability to provide real-time readout. A MOSkin detector was calibrated and the reproducibility, linearity, and response over a large dose range to different threshold voltages were determined. Surface dose on solid water phantom was measured using MOSkin detector and compared with Markus ionization chamber and GAFCHROMIC EBT2 film measurements. Dependence in the response of the MOSkin detector on the surface of solid water phantom was also tested for different (i) source to surface distances (SSDs); (ii) field sizes; (iii) surface dose; (iv) radiation incident angles; and (v) wedges. The MOSkin detector showed excellent reproducibility and linearity for dose range of 50 cGy to 300 cGy. The MOSkin detector showed reliable response to different SSDs, field sizes, surface, radiation incident angles, and wedges. The MOSkin detector is suitable for in vivo skin dosimetry.

## Keywords

detector, during, dose, skin, characterization, radiotherapy, moskin, measurement, vivo, megavoltage

## Disciplines

Engineering | Science and Technology Studies

## Publication Details

Jong, W., Wong, J. H., Ung, N., Ng, K., Ho, G., Cutajar, D. L. & Rozenfeld, A. (2014). Characterization of MOSkin detector for in vivo skin dose measurement during megavoltage radiotherapy. *Journal of Applied Clinical Medical Physics*, 15 (5), 120-132.

## Authors

Weiloong Jong, Jeannie Hsiu Ding Wong, Ngiemin Ung, Kwanhoong Ng, Gwofuang Ho, Dean Cutajar, and Anatoly B. Rosenfeld

## Characterization of MOSkin detector for *in vivo* skin dose measurement during megavoltage radiotherapy

Wei Loong Jong,<sup>1</sup> Jeannie Hsiu Ding Wong,<sup>2</sup> Ngie Min Ung,<sup>1a</sup> Kwan Hoong Ng,<sup>2</sup> Gwo Fuang Ho,<sup>1</sup> Dean L. Cutajar,<sup>3</sup> Anatoly B. Rosenfeld<sup>3</sup>  
*Clinical Oncology Unit,<sup>1</sup> Faculty of Medicine, University of Malaya, Kuala Lumpur, Malaysia; Department of Biomedical Imaging and University of Malaya Research Imaging Centre (UMRIC),<sup>2</sup> Faculty of Medicine, University of Malaya, Kuala Lumpur, Malaysia; Centre for Medical Radiation Physics (CMRP),<sup>3</sup> University of Wollongong, NSW, Australia*  
nmung@ummc.edu.my

Received 28 December, 2013; accepted 15 May, 2014

*In vivo* dosimetry is important during radiotherapy to ensure the accuracy of the dose delivered to the treatment volume. A dosimeter should be characterized based on its application before it is used for *in vivo* dosimetry. In this study, we characterize a new MOSFET-based detector, the MOSkin detector, on surface for *in vivo* skin dosimetry. The advantages of the MOSkin detector are its water equivalent depth of measurement of 0.07 mm, small physical size with submicron dosimetric volume, and the ability to provide real-time readout. A MOSkin detector was calibrated and the reproducibility, linearity, and response over a large dose range to different threshold voltages were determined. Surface dose on solid water phantom was measured using MOSkin detector and compared with Markus ionization chamber and GAFCHROMIC EBT2 film measurements. Dependence in the response of the MOSkin detector on the surface of solid water phantom was also tested for different (i) source to surface distances (SSDs); (ii) field sizes; (iii) surface dose; (iv) radiation incident angles; and (v) wedges. The MOSkin detector showed excellent reproducibility and linearity for dose range of 50 cGy to 300 cGy. The MOSkin detector showed reliable response to different SSDs, field sizes, surface, radiation incident angles, and wedges. The MOSkin detector is suitable for *in vivo* skin dosimetry.

PACS number: 87.55.Qr

Key words: MOSFET, *in vivo* dosimetry, surface dose, skin dose, characterization, quality assurance (QA)

### I. INTRODUCTION

Quality assurance (QA) in radiotherapy is very important in order to ensure the correct functioning of all components in radiotherapy, from treatment planning to the delivery of the treatment.<sup>(1)</sup> Nowadays, advanced radiotherapy techniques, such as intensity-modulated radiotherapy (IMRT), require patient-specific QA to be performed to ensure the accuracy of radiation delivery during radiotherapy. However, these QA and verification procedures may not be sufficient to ensure the accuracy of the entire radiotherapy treatment.

A number of incidents have been reported recently.<sup>(2-4)</sup> Human errors and systematic errors contributed to these incidents. Therefore, towards that end, *in vivo* dosimetry can detect major errors during the delivery of radiotherapy. It also can access clinical relevant differences between planned and delivered dose, record the dose received by the patient, and fulfill legal requirements.<sup>(1)</sup>

<sup>a</sup> Corresponding author: Ngie Min Ung, Clinical Oncology Unit, Faculty of Medicine, University of Malaya, 50603 Kuala Lumpur, Malaysia; phone: (+60)3 7949 2456; fax: (+60)3 7960 3072; email: nmung@ummc.edu.my

In radiotherapy, *in vivo* dosimetry means the measurement of the radiation dose received by a patient during treatment.<sup>(1)</sup> Ideally, a dosimeter should be positioned at the point of interest inside a patient's body. However, in many cases it is not possible to place a dosimeter inside a real patient's body. Hence, the placement of a dosimeter on the surface of the patient's body becomes an alternative.

An ideal *in vivo* dosimeter should possess the following characteristics: (i) tissue equivalent; (ii) small in physical size and has small sensitive volume; (iii) features (e.g., temperature, energy) which are consistent and characterizable; (iv) does not perturb the radiation field; (v) nonhazardous to humans; and (vi) able to provide real-time dosimetric information. Thermoluminescence dosimeter (TLD)<sup>(5-7)</sup> is small in size, but requires a long series of pre- and postirradiation process. Radiochromic film<sup>(8-10)</sup> has excellent dosimetric spatial resolution, is able to provide two-dimensional (2D) dosimetric information<sup>(11)</sup> and is easy to use, but it is not done in real-time and may be affected by improper handling and scanner performance. Semiconductor detectors such as diode<sup>(12-13)</sup> and metal oxide semiconductor field effect transistor (MOSFET)<sup>(14-19)</sup> are able to achieve excellent spatial resolution with their small sensitive volumes. However, the energy, angle, temperature, and dose-rate dependence of semiconductor detectors require rigorous characterization.

The dose deposited on a phantom or patient surface mainly comes from primary photon beam, backscattered radiation from the phantom, as well as radiation contamination from the accelerator. Radiation contamination arises from: (i) treatment head materials and (ii) treatment setup parameters such as source-to-surface distance (SSD), field size, and beam modifier to the surface dose.<sup>(20)</sup> These contaminations will affect the dose in the buildup region. Therefore, it is essential to determine and know the effect of these treatment parameters.

Different terminologies, such as surface dose, skin dose, and entrance dose, have been used to describe the dose measured on the surface of a phantom or a human. The definitions for these terminologies differ according to the point of measurement on the patient or phantom. Surface dose is defined as the dose on the surface of the phantom or human, which is the interface between the air and the surface. Skin dose is defined as the dose at the depth of 0.07 mm.<sup>(21)</sup> Entrance dose is defined as the dose given by the entrance beam at the depth of maximum dose.<sup>(22)</sup>

Characterization of a dosimeter is normally performed at a condition where charged particle equilibrium (CPE) condition exists.<sup>(12-13,19)</sup> However, for *in vivo* skin dosimetry, the dosimeter should be characterized on the surface instead of the depth of maximum dose. This is because the dosimetric condition of skin surface and buildup region is different from the dosimetric condition at the depth of maximum dose. At the interface of two media (air and human tissue), CPE does not exist and there is a steep dose gradient in the buildup region. Therefore, characterization of a dosimeter on surface is needed prior to using it for *in vivo* skin dosimetry.

A MOSFET-based dosimeter, the MOSkin detector was designed and prototyped by the Center for Medical Radiation Physics (CMRP) in the University of Wollongong (UoW). The advantages of the MOSkin detector, such as being small in size with submicron dosimetric volume which provides excellent dosimetry spatial resolution, as well as the ability to provide real-time reading and instant readout, make it suitable for *in vivo* skin dosimetry measurement. The MOSkin detector has been characterized and been used for dose measurement in megavoltage radiotherapy and brachytherapy.<sup>(23-30)</sup>

In this paper, a full characterization of the MOSkin detector on the surface of a phantom simulating the actual condition for *in vivo* skin dosimetry (where non-CPE condition exist) was performed and reported. These include: (i) detector calibration, linearity, reproducibility; (ii) source to surface distance dependence; (iii) field size dependence; (iv) surface dose measurement; (v) angular dependence; and (vi) wedge response. Comparison and verifications were made with previous works with some extension, while benchmarking against different dosimeters that are available commercially and used extensively in radiotherapy centers.

## II. MATERIALS AND METHODS

### A. The MOSkin detector

The MOSkin system is shown in Fig. 1. The MOSkin detector is composed of hermetically sealed MOSFET die with submicron thickness of the sensitive volume into Kapton pigtail strip with thickness of 0.55 mm using “drop-in” packaging technology<sup>(31)</sup> (Fig. 1(a)). The thin reproducible polyamide film acts as an electrical connection and buildup for MOSkin, and gives a water-equivalent depth (WED) of approximately 0.07 mm in tissue, making it a suitable dosimeter for skin dose measurement.<sup>(24)</sup> According to the International Commission on Radiological Protection (ICRP) publication,<sup>(21)</sup> the most radiosensitive layer of epidermis is located at tissue depth of approximately 0.07 mm. A detailed description of the MOSkin dosimetry system can be found in Kwan et al.<sup>(24)</sup> and Qi et al.<sup>(26)</sup> The readout process of MOSkin detector requires measurement of the voltage across the gate of the MOSkin detector under condition of the constant source–drain current that is called the threshold voltage,  $V_{th}$ . The  $V_{th}$  increases with accumulated radiation dose. The readout current corresponds to the thermostable point of the MOSFET to avoid errors associated with thermal instability of the  $V_{th}$ . The sensitivity of the MOSkin detector is defined as the shift of the  $V_{th}$  with the absorption of 1 cGy of radiation dose (Eq. (1)). In this work, the MOSkin measurements were benchmarked against Markus ionization chamber (Markus type 23343 parallel plate ionization chamber; PTW, Freiburg, Germany) and/or GAFCHROMIC EBT2 film (International Specialty Products, Wayne, NJ). All measurements were carried out three times and the mean  $\pm$  1 SD of the readings were reported unless stated otherwise.

$$Sensitivity = \frac{\Delta V_{th}}{1 \text{ cGy}} \quad (1)$$

where  $\Delta V_{th}$  is the change of the threshold voltage in unit Volt (V).




FIG. 1. MOSkin system, MOSkin detector (top right), and the schematic diagram of MOSkin detector in (a) face-up and (b) face-down orientation.

## B. GAFCHROMIC EBT2 film preparation

GAFCHROMIC EBT2 films were cut into sizes of  $1.5 \times 1.5 \text{ cm}^2$ . They were scanned using a flatbed scanner (Epson 1000XL scanner; Epson America, Inc. Long Beach, CA) 24 hours after irradiation to allow for postirradiation color changes.<sup>(32)</sup> The films were scanned in a reflection mode, at a resolution of 96 dots per inch (dpi), 48-bits RGB format, and analyzed using ImageJ 1.46r software (National Institute of Health, Bethesda, MD). Care was taken to scan the films at the center of the scanner to avoid scanner-induced nonuniformity. The films were also scanned in the same orientation to avoid film-induced changes in pixel values.<sup>(33)</sup> Only the red channel was used for analysis. A region of interest (ROI) was selected at the center of the film. A set of standard films was irradiated to establish the calibration curve.

## C. Detector characterization

### C.1 Calibration, linearity, and reproducibility

The MOSkin detectors were calibrated under a Varian Clinac 2100 C/D accelerator (Varian Medical System, Palo Alto, CA) using 6 MV photon beam under standard conditions (1.5 cm depth in  $30 \times 30 \times 15 \text{ cm}^3$  solid water phantom, 100 cm source–surface distance (SSD), and  $10 \times 10 \text{ cm}^2$  field size). Sensitivities of the MOSkin detector have been determined. Linearity measurement of the MOSkin detector was determined for a dose range of 50 cGy to 300 cGy, with an increment of 50 cGy, and the reproducibility was assessed.

In this work, the buildup cap for the Markus ionization chamber was removed in order to position the chamber's effective measurement closer to the surface. Temperature and pressure correction factor, polarity effect correction factor, and ionization recombination correction factor were taken into account for Markus ionization chamber measurements. Parallel plate ionization chambers (Markus ionization chamber) are known to overrespond due to side scatter from the chamber's wall.<sup>(34-36)</sup> In this work, Gerbi and Khan's correction<sup>(36)</sup> (Eqs. (2) and (3)) was only applied in the surface dose measurement (Material & Methods section C.4).

$$P'(d, E) = P(d, E) - \varepsilon(0, E) l e^{-\alpha(d-d_{\max})} \quad (2)$$

$$\varepsilon(0, E) = [-1.66 + (1.982 IR)] [C - 15.8] \quad (3)$$

where,  $P'(d, E)$  is the corrected PDD,  $P(d, E)$  is the measured PDD,  $E$  is the energy,  $l$  is the plate separation (2 mm for Markus PTW 23343),  $\alpha$  is constant (5.5),  $C$  is the sidewall collector distance (0.35 mm for Markus PTW 23343),  $IR$  is the ionization ratio, and  $d$  is the depth of the chamber front window below the surface of phantom surface. The calculated  $\varepsilon(d, E)$  are 10.14 and 6.89 for 6 MV and 10 MV photon beams, respectively.

Except for calibration and dose linearity measurement, all measurements were carried out on the surface of a solid water phantom. Characterization was carried out using a  $30 \times 30 \times 15 \text{ cm}^3$  solid water phantom with a 6 MV photon beam, 100 cm SSD, and  $10 \times 10 \text{ cm}^2$  field size (Fig. 2), unless stated otherwise. This setup is henceforth called the "standard surface setup".

Cheung et al.<sup>(37)</sup> has studied the temperature dependence of this MOSFET-based detector. They reported that this detector shows a variation of 50 mV over the temperature range from  $20^\circ\text{--}40^\circ\text{C}$ . This variation is corresponding to about 10 cGy in dose. However, in order to get an accurate reading, the detector should be placed on phantom or patient approximately 60 s before measurement, to allow thermal equilibrium, and the reading are taken whilst the detector remains on the phantom or patient. The same precaution was also taken throughout this work to reduce the effect of temperature dependence of the detector.




FIG. 2. Schematic diagram of the standard setup of characterization of MOSkin detector on the surface of a solid water phantom.

### C.2 Source-to-surface distance dependence

The MOSkin detector was positioned as per the “standard surface setup”. The response of the MOSkin detector for different distances from the source was measured with SSD, varying from 80 cm to 110 cm with 5 cm increments. One hundred MUs were delivered for each and repeated twice for all measured SSDs. Dose-rate dependence of the MOSkin detector was also evaluated in this section. The dose rate at  $d_{\max}$  was calculated.

### C.3 Field size dependence

The MOSkin detector was set up per “standard surface setup” and irradiated with different field sizes from  $1 \times 1 \text{ cm}^2$  to  $40 \times 40 \text{ cm}^2$  using 6 MV photon beam.

### C.4 Surface dose measurement

The surface dose measured by MOSkin detector for 6 MV and 10 MV photon beams was evaluated. The measured dose by Markus ionization chamber was corrected, based on Gerbi and Khan’s formulae.<sup>(36)</sup>

### C.5 Angular dependence

Conventionally, angular dependence of the MOSFET was carried out at a depth where the CPE condition exists. Kwan et al.<sup>(24)</sup> and Qi et al.<sup>(29)</sup> have measured angular dependence of  $\pm 2\%$  for MOSkin detector in a cylindrical phantom, where the CPE exists, and Qi et al.<sup>(26)</sup> have measured angular dependence of 3.1% on the surface of a solid water phantom. As reported in Scalchi et al.,<sup>(17)</sup> full buildup setup where CPE exists gives better results than a surface setup.

Here we study the angular response of MOSkin detector placed on the surface of a phantom as the detector will be used for *in vivo* skin dose measurement where CPE does not exist. It is important to note that surface dose increases with beam incidence angle and correct measurements of the surface dose or skin dose for different angles of beam incidence are valuable for treatment planning system (TPS) verification. In particular, for the case of tangential beams, the angular response of the MOSkin detector will consist of the increased surface dose due to the beam incident angle and the intrinsic angular response of the MOSkin. The angular response of the MOSkin detector was assessed in face-up and face-down orientation, as shown in Figs. 1(a) and (b), by positioning the MOSkin detector per “standard surface setup”. One hundred

MUs of 6 MV photon were delivered with the accelerator gantry rotated to the angles of  $0^\circ$  to  $75^\circ$ , with  $15^\circ$  increments.

### C.6 Wedge response

In some clinical applications, such as conventional breast radiotherapy, beam modifier devices like physical wedge (PW) or dynamic wedge (DW)<sup>(38-39)</sup> may be required to tilt the dose profile, resulting in an angled isodose curve.<sup>(38)</sup> Beam quality and dose rate of the incident photons may change due to the presence of the beam modifier device.<sup>(40)</sup> For wedge response measurement, the MOSkin detector was irradiated using “standard surface setup”. One hundred MUs were delivered on the detector in an open field and subsequently with the application of PW and DW of  $15^\circ$ ,  $30^\circ$ ,  $45^\circ$ , and  $60^\circ$ .

## III. RESULTS & DISCUSSION

### A. Calibration, linearity and reproducibility

The MOSkin detector showed excellent reproducibility with deviation of less than 1% and excellent linearity ( $R^2 = 0.997$ ) for the dose range of 0 cGy to 300 cGy, as shown in Fig. 3. The reproducibility of the MOSkin detector was determined based on the average standard deviation (1 SD) of three repeated measurements of each dose level in linearity test. The dose linearity verification was carried out at this range because it was deemed to be within the range of a normal fractionated dose in radiotherapy. The average sensitivity of the MOSkin detectors in this study was  $2.53 \pm 0.03$  mV/cGy for 6 MV photon beam. The sensitivity of the MOSkin detector as a function of cumulative dose was found to decrease by  $9 \times 10^{-2}$  mV/cGy for every 10 Gy of delivered dose (Fig. 4). The sensitivity of the MOSkin detectors is expected to decrease as the cumulative dose increases.<sup>(41)</sup> Therefore, it is recommended that periodic recalibration be carried out throughout the detector’s useful lifetime based on the accuracy needed. The readers are referred to Qi et al.<sup>(23)</sup> for details of the MOSkin detector’s lifetime.




Fig. 3. Linearity of MOSkin detector for the dose range of 0 cGy to 300 cGy.






FIG. 4. Change in sensitivity of MOSkin detector as a function of cumulative dose.

### B. Source-to-surface distance dependence

Results of SSD dependence and dose rate dependence measurements for Markus ionization chamber, GAFCHROMIC EBT2 film, and the MOSkin detector are presented in Fig. 5. All measured doses were corrected with inverse square correction factor (ISCF) and normalized to the corrected dose at SSD 100 cm. An ideal SSD independent and dose rate independent dosimeter will have an “Inverse Square Corrected Relative Dose” equal to one.

From Fig. 5, the MOSkin detector, GAFCHROMIC EBT2 film, and Markus ionization chamber showed steady response over 80 cm to 110 cm SSD. The average variation of all SSDs for the MOSkin detector, GAFCHROMIC EBT2 film, and Markus ionization chamber was 0.1%, 1.0%, and 0.5%, respectively.

On the surface of the phantom, the dose deposited is not only due to the primary beam directed from the treatment head, but also from the contaminant electrons which are generated outside of the patient in the air and collimator. This may contribute to the SSD dependence of a skin dosimeter. This contamination is not sufficient to contribute to SSD dependence for large SSD as the electrons produced in the accelerator were of relatively high energy.<sup>(20)</sup> Measurements at shorter SSD exposed the dosimeter to large amounts of low-energy photons scattered by the components in the accelerator and would induce a slight over-response of a dosimeter.




FIG. 5. SSD and dose rate response of Markus ionization chamber, GAFCHROMIC EBT2 film and MOSkin detector corrected with ISCF and normalized to the response at 100 cm SSD. The error bar represents 1 SD of three sets of measurements.

### C. Field size dependence

Qi et al.<sup>(26)</sup> have also studied the field size dependence of the MOSkin detector on phantom surface for the field size of  $5 \times 5 \text{ cm}^2$  to  $30 \times 30 \text{ cm}^2$ . In this study, the field sizes investigated were extended to include small field sizes of as small as  $1 \times 1 \text{ cm}^2$  and larger field size of  $40 \times 40 \text{ cm}^2$ . The field size dependence of surface dose measured by MOSkin detector, as well as EBT2 film and Markus ionization chamber, is shown in Fig. 6. All readings were normalized to  $10 \times 10 \text{ cm}^2$  field size. The three detectors showed an upward trend as field size increases. This is expected and is due to the increase in the backscattered radiation from the phantom and radiation contamination. The radiation contamination was mainly due to the scattered radiation from the flattening filter, with a small portion of scattered radiation as a function of field size.<sup>(42)</sup> The result of MOSkin detector and Attix ionization chamber from Qi et al.<sup>(26)</sup> are also presented in Fig. 6 for comparison. Surface dose increases measured with the MOSkin detectors rose from 0.36 to 2.32 times of the surface dose measurement at  $10 \times 10 \text{ cm}^2$  for the field sizes from  $1 \times 1 \text{ cm}^2$  to  $40 \times 40 \text{ cm}^2$ .

The dose deposited on the surface of the phantom is expected to be only 10%–20% of the maximum dose ( $D_{\text{max}}$ ).<sup>(38)</sup> As field size increases, the surface dose increases and, hence, there is a corresponding reduction in the surface dose gradient.<sup>(17)</sup>

All detectors measured increasing surface dose with increasing field size. For field size  $< 25 \times 25 \text{ cm}^2$ , the MOSkin measurements are in good agreement with EBT2 film. MOSkin detector is advantageous for small field surface dosimetry compared to Markus ionization chamber due to the small sensitive volume of the detector. For field size  $> 25 \times 25 \text{ cm}^2$ , MOSkin detector and Markus ionization chamber measurements showed an enhanced response, compared to the GAFCHROMIC EBT2 film. This may be due to the side scattering effect of Markus ionization chamber and energy dependence of the semiconductor (MOSkin) detector.

Good agreement between the MOSkin detector's response and Attix ionization chamber<sup>(29)</sup> (average difference of 2%) has been observed due to close WED of the detectors used. Earlier reported results of field dependence measured with MOSkin detectors are in agreement (average difference of 3%) with presented results, confirming good reproducibility of WED of the MOSkin detectors.<sup>(26)</sup>




FIG. 6. Field size response of the surface dose measured by MOSkin detector, GAFCHROMIC EBT2 film, and Markus ionization chamber normalized to response at  $10 \times 10 \text{ cm}^2$  radiation field size. The average standard deviation of three sets of measurements for MOSkin detector, GAFCHROMIC EBT2 film, and Markus ionization chamber is 0.034, 0.052, and 0.001, respectively.

#### D. Surface dose measurement

The average sensitivity of the MOSkin detectors was found to be  $2.53 \pm 0.03 \text{ mV/cGy}$  and  $2.50 \pm 0.02 \text{ mV/cGy}$  for 6 MV and 10 MV photon beams, respectively. The MOSkin detectors were 1.01% more sensitive in 6 MV photon beam compared to 10 MV photon beam.

Table 1 shows the measured surface dose of the MOSkin detector, GAFCHROMIC EBT2 film, and the Markus ionization chamber measurement for 6 MV and 10 MV photon beams. The surface doses for 6 MV photon measured by these detectors are higher than 10 MV photon due to the skin sparing effect of 10 MV photon.

Markus ionization chamber measured lowest surface doses for 6 MV and 10 MV photons, followed by the MOSkin detector and GAFCHROMIC EBT2 film. This can be explained by the WED of these detectors. Markus ionization chamber was assumed to have WED of 0 mm after the application of Gerbi and Khan's correction.<sup>(36)</sup> MOSkin detector has a WED of 0.070 mm<sup>(24)</sup> and GAFCHROMIC EBT2 film has a WED of 0.122 mm. The WED of GAFCHROMIC EBT2 film was determined based on the physical depth and density<sup>(43-44)</sup> from the surface to the center of the active layer.

In measuring the surface dose at the depth of 0 cm, the dose difference (%) between the MOSkin detector and Markus ionization chamber was found to be 4.44% and 1.74% for 6 MV and 10 MV photon beams, respectively (Table 1). It is expected because the dose gradient in build-up region for 6 MV photon beam is steeper than that of 10 MV photon beam.

TABLE 1. Comparison of the surface dose (normalized to 100% the dose at  $d_{\text{max}}$ ) with Markus ionization chamber, GAFCHROMIC EBT2 film, and MOSkin detector for 6 MV and 10 MV photons.

Energy (MV)	Markus (%)	EBT2 (%)	MOSkin (%)
6	$15.83 \pm 0.03$	$23.01 \pm 0.07$	$20.27 \pm 0.03$
10	$11.82 \pm 0.00$	$17.89 \pm 0.04$	$13.55 \pm 0.04$

### E. Angular dependence

Figure 7 shows the results of the measured doses which have been normalized to the dose at  $0^\circ$  beam incident angle. The result of MOSkin detector and Attix ionization chamber from Qi et al.<sup>(26)</sup> are also presented in Fig. 7 for comparison. As the beam incident angle increases, the measured surface dose increases because the region of charged particle equilibrium shifts toward the surface. This is in agreement with the results by Scalchi et al.<sup>(17)</sup> and Qi et al.<sup>(26)</sup>

GAFCHROMIC EBT2 film, with its dosimetric properties of tissue equivalence and homogeneous material, is assumed to be of angular independence. Suchowerska et al.<sup>(45)</sup> reported that GAFCHROMIC film shows intrinsic angular dependence of less than 1% when a measurement with the film surface is parallel and perpendicular to the beam direction. MOSkin detector in the face-up orientation showed similar angular response trend to the GAFCHROMIC EBT2 film (Fig. 7). GAFCHROMIC EBT2 film measured higher dose as compared to MOSkin detector in face-up orientation. This may be due to the difference of the WED these detectors (Results & Discussion section D).

When the MOSkin detector was used in the face-down orientation for surface dose measurement, it showed a trend of overresponse when compared with MOSkin detector and GAFCHROMIC EBT2 film. Deviation between MOSkin detector in face-up orientation and face-down orientation is seen to be increasing as the beam incident angle increases. Maximum deviation of 18.5% was found at beam incident angle of  $75^\circ$ . The observed angular dependence between MOSkin detector in the face-up orientation and face-down orientation arises from the effect of the difference in the WED of the detector related to the asymmetric geometry of the detector (inherent anisotropy). For face-down geometry, the WED is approximately 0.9 mm due to a 0.4 mm silicon substrate. Therefore, it is very important to identify the orientation of MOSkin detector when it is used for surface dose measurement. For WED 0.07 mm, MOSkin detector should be used in a face-up mode. Present result was in good agreement with earlier reported result<sup>(26)</sup> with maximum deviation of 2.2 % at angle  $30^\circ$ .




FIG. 7. Relative surface dose measured with GAFCHROMIC EBT2 film and MOSkin detector as the function of beam incident angle and normalized to 1 at  $0^\circ$  beam incident angle with  $10 \times 10 \text{ cm}^2$  field size. The error bar represents 1 SD of three sets of measurements. The average standard deviation of three sets of measurements for MOSkin detector and GAFCHROMIC EBT2 film is 0.033 and 0.036, respectively.

## F. Wedge response

The results of the wedge effect are summarized in Fig. 8. The doses were normalized to the dose measured with open field. The surface dose decreases as the wedge angle increases for both PW and DW. The average variation of MOSkin measurement with Markus ionization measurement for PW and DW was -5.5% and 5.1%, respectively. MOSkin measurement is in close agreement with GAFCHROMIC EBT2 film measurement. GAFCHROMIC EBT2 film responded higher for PW (2.8%) and DW (1.7%), respectively. This is in agreement with other scenarios of surface dose measurements due to higher WED of GAFCHROMIC EBT2 film in comparison with the MOSkin detector.

The presence of any materials between the radiation source and the phantom or patient will alter the dose to the build-up region. In this work, constant MU (100 MUs) was given. The presence of physical wedge under the photon beam will harden the photon beam by absorbing scattered radiation.<sup>(38)</sup> The presence of the wedge will also produce low energy scattered radiation.<sup>(46)</sup>




FIG. 8. Wedges response measurement of Markus ionization chamber, GAFCHROMIC EBT2 film, and MOSkin detector and normalized to 1 at open field. The error bar represents 1 SD of three sets of measurements. The average standard deviation of three sets of measurements for MOSkin detector, GAFCHROMIC EBT2 film, and Markus ionization chamber is 0.033, 0.019, and 0.001, respectively.

## IV. CONCLUSIONS

The surface dose measured with MOSkin detector was investigated on a phantom surface for different SSD, field sizes, surface dose, oblique beams, machine dose rates, and in the presence of the wedge in comparison with EBT2 film and Markus ionization chamber.

MOSkin detector showed deviation of less than 2% over a change of 80 – 110 cm SSD. For field size dependence, they are in agreement with Attix ionization chamber.<sup>(29)</sup> Surface dose measured with the MOSkin detectors increases from 0.36 to 2.32 times of the surface dose measurement at  $10 \times 10 \text{ cm}^2$  for the field sizes from  $1 \times 1 \text{ cm}^2$  to  $40 \times 40 \text{ cm}^2$ . The dose difference between the MOSkin detector and Markus ionization chamber was 4.44% and 1.74% for 6 MV and 10 MV photon beams, respectively. This is due to the different WEDs for both detectors. The dose gradient in the buildup region of 6 MV photon beam is steeper than that of the 10 MV photon beam. When oblique beams are used, surface dose measured with the MOSkin detector increases up to 1.95 times of the normal beam incidence. For angular dependence, MOSkin detector in face-up orientation is in agreement with Attix ionization chamber.<sup>(26)</sup>

Maximum deviation of 18.5% was found with face-down orientation. The orientation (face-up or face-down) of the MOSkin detector must be taken into account when used for skin dose measurement. The MOSkin measurement is in close agreement with GAFCHROMIC EBT2 film for the measurement made under the presence of wedges.

MOSkin detector is suitable detector for *in vivo* skin dosimetry as compared to GAFCHROMIC EBT2 film because of its WED of 0.07 mm. However, due to the difference between the detector's WED and materials, neither can be used as true benchmarked for determining MOSkin detector accuracy as skin dosimeter. Monte Carlo calculation may provide a true benchmark tool for this comparison. However, this is not within the scope of this study.

## ACKNOWLEDGMENTS

We would like to acknowledge HIR Grant UM C/625/1/HIR/MOHE/CHAN/06 (H50001-00-A000020-000001) and UMRG Grant RG507-13HTM for supporting this research. We would also like to thank all the radiographers in the Clinical Oncology Unit of the University Malaya Medical Centre for their assistances in this study.

## REFERENCES

1. Mijnheer B, Beddar S, Izewska J, Refit C. *In vivo* dosimetry in external beam radiotherapy. *Med Phys*. 2013;40(7):1–19.
2. Mayles WP. The Glasgow incident — a physicist's reflections. *Clin Oncol*. 2007;19(1):4–7.
3. Williams MV. Radiotherapy near misses, incidents and errors: radiotherapy incident at Glasgow. *Clin Oncol*. 2007;19(1):1–3.
4. Ortiz Lopez P, Cosset JM, Dunscombe P, et al. ICRP publication 112. A report of preventing accidental exposures from new external beam radiation therapy technologies. *Ann ICRP*. 2009;39(4):1–86.
5. Costa AM, Barbi GL, Bertucci EC, et al. *In vivo* dosimetry with thermoluminescent dosimeters in external photon beam radiotherapy. *Appl Radiat Isot*. 2010;68(4-5):760–62.
6. Rivera T. Thermoluminescence in medical dosimetry. *Appl Radiat Isot*. 2012;71(Suppl):30–34.
7. Kron T. Applications of thermoluminescence dosimetry in medicine. *Radiat Prot Dosim*. 1999;85(1-4):333–40.
8. Butson MJ, Yu PK, Cheung T, Metcalfe P. Radiochromic film for medical radiation dosimetry. *Mater Sci Eng Reports*. 2003;41(3-5):61–120.
9. Devic S. Radiochromic film dosimetry: past, present, and future. *Phys Med*. 2011;27(3):122–34.
10. Niroomand-Rad A, Blackwell CR, Coursey BM, et al. Radiochromic film dosimetry: recommendations of AAPM Radiation Therapy Committee Task Group 55. American Association of Physicists in Medicine. *Med Phys*. 1998;25(11):2093–115.
11. Sim GS, Wong JHD, Ng KH. The use of radiochromic EBT2 film for the quality assurance and dosimetric verification of 3D conformal radiotherapy using Microtek ScanMaker 9800XL flatbed scanner. *J Appl Clin Med Phys*. 2013;14(4):85–95.
12. Meiler RJ, Podgorsak MB. Characterization of the response of commercial diode detectors used for *in vivo* dosimetry. *Med Dosim*. 1997;22(1):31–37.
13. Yorke E, Alecu R, Ding L, et al. Diode *in vivo* dosimetry for patients receiving external beam radiation therapy. Report of Task Group 62 of the Radiation Therapy Committee. Madison, WI: AAPM; 2005.
14. Bharanidharan G, Manigandan D, Devan K, et al. Characterization of responses and comparison of calibration factor for commercial MOSFET detectors. *Med Dosim*. 2006;30(4):213–18.
15. Gopiraj A, Billimagga RS, Ramasubramanian V. Performance characteristics and commissioning of MOSFET as an *in-vivo* dosimeter for high energy photon external beam radiation therapy. *Rep Pract Oncol Radiother*. 2008;13(3):114–25.
16. Ramani R, Russell S, O'Brien P. Clinical dosimetry using MOSFETs. *Int J Radiat Oncol Biol Phys*. 1997;37(4):959–64.
17. Scalchi P, Francescon P, Rajaguru P. Characterization of a new MOSFET detector configuration for *in vivo* skin dosimetry. *Med Phys*. 2005;32(6):1571–78.
18. Halvorsen PH. Dosimetric evaluation of a new design MOSFET *in vivo* dosimeter. *Med Phys*. 2005;32(1):110–17.
19. Scalchi P, Francescon P. Calibration of a MOSFET detection system for 6-MV *in vivo* dosimetry. *Int J Radiat Oncol Biol Phys*. 1998;40(4):987–93.
20. Yadav G, Yadav RS, Kumar A. Skin dose estimation for various beam modifiers and source-to-surface distances for 6MV photons. *J Med Phys*. 2009;34(2):87–92.
21. ICRP. The biological basis for dose limitation in the skin. A report of a Task Group of Committee 1 of the International Commission on Radiological Protection. *Ann ICRP*. 1992;22(2).

22. VanDam J and Marinello G. Methods for in vivo dosimetry in external radiotherapy. Physics for clinical radiotherapy. ESTRO Booklet n.1. Brussels: Garant Publ.; 1994.
23. Qi ZY, Deng XW, Huang SM, et al. Verification of the plan dosimetry for high dose rate brachytherapy using metal-oxide-semiconductor field effect transistor detectors. *Med Phys.* 2007;34(6):2007–13.
24. Kwan IS, Rosenfeld AB, Qi ZY, et al. Skin dosimetry with new MOSFET detectors. *Radiat Meas.* 2008;43(2-6):929–32.
25. Hardcastle N, Soisson E, Metcalfe P, Rosenfeld AB, Tomé WA. Dosimetric verification of helical tomotherapy for total scalp irradiation. *Med Phys.* 2008;35(11):5061–68.
26. Qi ZY, Deng XW, Huang SM, et al. In vivo verification of superficial dose for head and neck treatments using intensity-modulated techniques. *Med Phys.* 2009;36(1):59–70.
27. Kwan IS, Wilkinson D, Cutajar D, et al. The effect of rectal heterogeneity on wall dose in high dose rate brachytherapy. *Med Phys.* 2009;36(1):224–32.
28. Hardcastle N, Cutajar DL, Metcalfe PE, et al. In vivo real-time rectal wall dosimetry for prostate radiotherapy. *Phys Med Biol.* 2010;55(13):3859–71.
29. Qi ZY, Deng XW, Huang SM, et al. Real-time in vivo dosimetry with MOSFET detectors in serial tomotherapy for head and neck cancer patients. *Int J Radiat Oncol Biol Phys.* 2011;80(5):1581–88.
30. Alnawaf H, Butson M, Yu PK. Measurement and effects of MOSKIN detectors on skin dose during high energy radiotherapy treatment. *Australas Phys Eng Sci Med.* 2012;35(3):321–28.
31. Wong JH, Fuduli I, Carolan M, et al. Characterization of a novel two dimensional diode array the “magic plate” as a radiation detector for radiation therapy treatment. *Med Phys.* 2012;39(5):2544–58.
32. Cheung T, Butson MJ, Yu PK. Post-irradiation colouration of Gafchromic EBT radiochromic film. *Phys Med Biol.* 2005;50(20):N281–N285.
33. Butson MJ, Cheung T, Yu PK. Scanning orientation effects on Gafchromic EBT film dosimetry. *Australas Phys Eng Sci Med.* 2006;29(3):281–84.
34. Rawlinson JA, Arlen D, Newcombe D. Design of parallel plate ion chambers for buildup measurements in megavoltage photon beams. *Med Phys.* 1992;19(3):641–48.
35. Bilge H, Cakir A, Okutan M, Acar H. Surface dose measurements with Gafchromic EBT film for 6 MV and 18 MV photon beams. *Phys Med.* 2009;25(2):101–04.
36. Gerbi BJ and Khan FM. Measurement of dose in the buildup region using fixed-separation plane-parallel ionization chambers. *Med Phys.* 1990;17(1):17–26.
37. Cheung T, Butson MJ, Yu PK. Effects of temperature variation on MOSFET dosimetry. *Phys Med Biol.* 2004;49(13):N191–N196.
38. Metcalfe PE, Kron T, Hoban P. The physics of radiotherapy x-rays and electrons. Madison, WI: Medical Physics Publishing; 2007.
39. Shih R, Li XA, Chu JCH. Dynamic wedge versus physical wedge: a Monte Carlo study. *Med Phys.* 2001;28(4):612–19.
40. Lopez Medina A, Teijeiro A, Garcia J, et al. Characterization of electron contamination in megavoltage photon beams. *Med Phys.* 2005;32(5):1281–92.
41. Zilio VO, Joneja OP, Popowski Y, Rosenfeld A, Chawla R. Absolute depth-dose-rate measurements for an 192Ir HDR brachytherapy source in water using MOSFET detectors. *Med Phys.* 2006;33(6):1532–39.
42. Yu MK and Slooboda R. Analytical representation of head scatter factors for shaped photon beams using two-component x-ray source model. *Med Phys.* 1995;22(12):2045–55.
43. ISP. Gafchromic EBT2 Self-developing film for radiotherapy dosimetry. Wayne, NJ: International Specialty Products; 2009.
44. Aland T, Kairn T, Kenny J. Evaluation of a Gafchromic EBT2 film dosimetry system for radiotherapy quality assurance. *Australas Phys Eng Sci Med.* 2011;34(2):251–60.
45. Suchowerska N, Hoban P, Butson M, Davison A, Metcalfe P. Directional dependence in film dosimetry: radiographic and radiochromic film. *Phys Med Biol.* 2001;46(5):1391–97.
46. Öchran TG, Boyer AL, Nyerick CE, Otte VA. Dosimetric characteristics of wedges mounted beyond the blocking tray. *Med Phys.* 1992;19(1):187–94.