The University of Hong Kong The HKU Scholars Hub



Title	Technology-related medication errors – incidence, nature and causes in a tertiary hospital
Author(s)	Samaranayake, NR; Cheung, STD; Chui, WCM; Cheung, BMY
Citation	Hospital Authority Convention 2012, Hong Kong, China, 7-8 May 2012, p. abstract no. SPP-P1.27
Issued Date	2012
URL	http://hdl.handle.net/10722/165046
Rights	Creative Commons: Attribution 3.0 Hong Kong License

errors Methods


Technology-related medication errors – incidence, nature and causes in a tertiary hospital

NR Samaranayake¹, STD Cheung², CMW Chui², BMY Cheung¹

¹Department of Medicine, University of Hong Kong, ²Department of Pharmacy, Queen Mary Hospital, Hong Kong



Analysis of technology-related medication



causes

Definitions

Technology-related errors - Any error that involved technologies used in hospitals to reduce medication


User related errors – Any technology-related error associated with a human failure

Device related errors – Any technology-related error that occurred due to a technical defect of a device

Study process

- 1538 incidents were reported during the period of analysis (2006 - 2010)
- All incidents were reviewed by a pharmacist and technology-related incidents were identified as follows


Pathway for incident analysis




Results

Incidence of technology-related errors

17.3% of all incidents were 'technologyrelated' and most were due to 'user errors'




Types of technology-related errors Most technology-related errors were prescribing errors followed by drug administration and dispensing errors



Technologies involved in technology-related errors





Severity of technology-related errors

12% of technology-related errors reached the patient and 6.4% caused some form of harm to the patient



Most common underlying causes and their % contribution to the occurrence of technologyrelated errors

Incorrect computer entry (49.2%) Similar drug name

appearance (6%) Failure to comply

with policies and procedures (39.1%) **Device related**

problems (1.9%) Lack of supervision (1.1%)

Conclusions

Technologies

involved

Computer

generated

dispensing

Bar-coded

patient ID

Infusion pumps

labels

CMOE

- Technologies have a potential to introduce new errors
- Most technology-related errors are related to user errors than technical defects
- Common underlying causes were incorrect computer entry and staff not complying to policies and procedures
- When using technological interventions, systems need to be improved in a way that errors cannot happen

Staff training and continuous monitoring are also important to minimise technology-related errors