



Certification of mass fractions of aflatoxin B_1 , B_2 , G_1 and G_2 in peanut butter BCR-385R and BCR-401R

G. Buttinger, S. Harbeck, R. Josephs



EUR 23522 EN - 2008

The mission of the IRMM is to promote a common and reliable European measurement system in support of EU policies.

European Commission
Joint Research Centre
Institute for Reference Materials and Measurements

Contact information

Address: Retieseweg 111, 2440 Geel, Belgium
E-mail: jrc-irrm-rm-sales@ec.europa.eu
Tel.: 014/571 705
Fax: 014/590 406

<http://irrm.jrc.ec.europa.eu/>
<http://www.jrc.ec.europa.eu/>

Legal Notice

Neither the European Commission nor any person acting on behalf of the Commission is responsible for the use which might be made of this publication.

***Europe Direct is a service to help you find answers
to your questions about the European Union***

Freephone number (*):

00 800 6 7 8 9 10 11

(*) Certain mobile telephone operators do not allow access to 00 800 numbers or these calls may be billed.

A great deal of additional information on the European Union is available on the Internet. It can be accessed through the Europa server <http://europa.eu/>

EUR 23522 EN
ISBN 978-92-79-09976-2
ISSN 1018-5593
DOI 10.2787/77700

Luxembourg: Office for Official Publications of the European Communities

© European Communities, 2008

Reproduction is authorised provided the source is acknowledged

Printed in Belgium

European Commission

BCR information
REFERENCE MATERIALS

**Certification of mass fractions of aflatoxin
B₁, B₂, G₁ and G₂ in peanut butter**

BCR-385R and BCR-401R

G. Buttinger, S. Harbeck, R. Josephs

European Commission
Joint Research Centre
Institute for Reference Materials and Measurements (IRMM)
Geel, Belgium



Summary

This report describes the preparation of peanut butter (BCR-385R and BCR-401R) matrix reference materials and the certification of their content (mass fraction) of aflatoxins B₁, B₂, G₁ and G₂.

The preparation of the materials, the homogeneity and stability studies and the characterisation are described hereafter and the results are discussed. Uncertainties were calculated in compliance with the Guide to the Expression of Uncertainty in Measurement (GUM) [1], including uncertainties due to possible heterogeneity and instability. The certified values are listed below:

BCR-385R	Certified value ¹⁾	Uncertainty ²⁾	Number of accepted sets of results
Aflatoxin B ₁	1.77 µg/kg	0.30 µg/kg	7
Aflatoxin B ₂	0.48 µg/kg	0.08 µg/kg	6
Aflatoxin G ₁	0.9 µg/kg	0.4 µg/kg	6
Aflatoxin G ₂	0.30 µg/kg	0.12 µg/kg	4
Sum of Aflatoxin B ₁ , B ₂ , G ₁ , G ₂	3.5 µg/kg	0.5 µg/kg ³⁾	

- 1) These values are the mass fractions based on the unweighted mean of accepted results.
- 2) The uncertainties are the expanded uncertainties ($k = 2$) of the values defined in 1).
- 3) The uncertainty for the sum of Aflatoxins B₁, B₂, G₁, G₂ is calculated from the individual

absolute uncertainties as $U_{sum} = 2 \cdot \sqrt{u_{B_1}^2 + u_{B_2}^2 + u_{G_1}^2 + u_{G_2}^2}$.

BCR-401R	Certified value	Number of accepted sets of results
Aflatoxin B ₁	< 0.2 µg/kg ¹⁾	7
Aflatoxin B ₂	< 0.2 µg/kg ²⁾	6
Aflatoxin G ₁	< 0.2 µg/kg ²⁾	6
Aflatoxin G ₂	< 0.2 µg/kg ²⁾	4

- 1) This value is the mass fractions based on the limits of quantification of the methods used and the highest found level of accepted results, with a 80% probability the measurand content is below this level.
- 2) These values are the mass fractions based on the limits of quantification of the methods used and the highest found level of accepted results, with a 95% probability the measurand content is below this level.

Table of contents

Summary	1
Table of contents	3
Glossary	4
1 Introduction	5
2 Participants	6
3 Processing of the peanut materials	8
3.1 BCR-401R peanut butter (very low level)	8
3.2 BCR-385R peanut butter (low level)	8
3.3 Additional characterisation measurements	8
3.3.1 Water content and water activity	8
3.3.2 Particle size measurements	8
4 Homogeneity studies	10
5 Stability studies	12
6 Certification	14
6.1 Design of the study	14
6.2 Results and technical evaluation	15
6.3 Certified values and their uncertainties	19
7 Metrological traceability and commutability	21
8 Instructions for use	22
8.1 Storage conditions	22
8.2 Safety precautions	22
8.3 Use of the material	22
8.4 Use of the certified value	22
9 Acknowledgements	22
10 References	23
Annex A. Homogeneity data	24
Annex B. Stability data	33
Annex C. Certification measurements	41

Glossary

a_w	Water activity
ANOVA	Analysis of variances
ACN	Acetonitrile
CRM	Certified reference material
FAO	Food and Agriculture Organization of the United Nations
GUM	Guide to the Expression of Uncertainty in Measurement
HPLC-FLD	High performance liquid chromatography with fluorescence detection
IAC	Immunoaffinity column
MS_{between}	Mean of squares between groups (ANOVA)
MS_{within}	Mean of squares within groups (ANOVA)
MW	Molecular mass
n	Number of replicates
p	Level of significance
RSD	Relative standard deviation
RSD_r	Relative standard deviation calculated from results under repeatability conditions
RSD_{stab}	Relative standard deviation of all results of the stability study
s	Standard deviation
s_{bb}	Between-bottle (in)homogeneity standard deviation
s_{wb}	Within-bottle standard deviation
SI	International Systems of Units
U	Expanded uncertainty
u	Standard uncertainty
u_{B1}	Uncertainty of the mass fraction of aflatoxin B ₁
u_{B2}	Uncertainty of the mass fraction of aflatoxin B ₂
u_{G1}	Uncertainty of the mass fraction of aflatoxin G ₁
u_{G2}	Uncertainty of the mass fraction of aflatoxin G ₂
u_{Δ}	Combined uncertainty of certified value and measured value
u_{bb}	Relative standard uncertainty due to the inhomogeneity that can be hidden by the method repeatability
u_{bb}	Relative standard uncertainty due to between-bottle (in)homogeneity
u_{cal}	Relative uncertainty of the mass fraction of the calibrants used
u_{char}	Relative uncertainty of the characterisation exercise
u_{CRM}	Combined uncertainty of certified value
U_{CRM}	Expanded, relative uncertainty of certified value
$U_{\text{CRM, abs}}$	Expanded, absolute uncertainty of certified value
u_{its}	Relative uncertainty of long-term stability
u_{meas}	Uncertainty of measurement result
U_{sum}	Expanded uncertainty of the mass fraction of the sum of aflatoxins B ₁ , B ₂ , G ₁ and G ₂
u_{sts}	Relative uncertainty of short-term stability
x	Pre-defined shelf life
\bar{x}	Average of all time points in an isochronous stability study
x_i	Time point i in an isochronous stability study
\bar{y}	Average of all results of a homogeneity study
Δ	Difference between two measurement results
Δ_m	Difference between measured and certified value
v_{MSwithin}	Degrees of freedom of MS_{within}

1 Introduction

Mycotoxins are secondary metabolites of moulds. These toxic metabolites occur as contaminants in a wide range of food and animal feed from plant origin and are therefore a potential risk to human and animal health. Contamination of food and feed can appear at two stages: on the field and/or during storage. Moulds infecting food on the field produce different mycotoxins compared to those moulds infecting food during storage.

The impact of mycotoxins on agricultural production is massive. The Food and Agriculture Organization of the United Nations (FAO) estimates that 25 % of the world-wide production is affected. The 2006 annual report of the rapid alert system for food and feed of the European Union [2] also shows that 30 % of the notifications were due to mycotoxin contaminations. Around 90 % of these notifications concerned aflatoxins and a quarter of these concerned peanuts and peanut products.

Maximum levels for certain mycotoxins have been introduced in the European Union since 1998 [3]. The maximum level for aflatoxin B₁ in peanuts for direct human consumption is 2.0 µg/kg and for the sum of aflatoxins B₁, B₂, G₁ and G₂ 4.0 µg/kg [4].

Aflatoxin B₁ (Figure 1)[2, 3, 6α, 9α-Tetrahydro-4-methoxycyclopenta [c] furo [3', 2':4, 5] furo [2, 3-h] [l] benzopyran-1, 11-dione, CAS Nr. 1162-65-8], aflatoxin B₂ (Figure 1) [2, 3, 6α, 8, 9, 9α-Hexahydro-4-methoxycyclopenta [c] furo [3', 2':4, 5] furo [2, 3-h] [l] benzopyran-1, 11-dione, CAS Nr. 7220-81-7], aflatoxin G₁ (Figure 1) [3, 4, 7α, 10α-Tetrahydro-5-methoxy-1*H*, 12*H* furo [3', 2':4, 5] furo [2, 3-h] pyrano [3, 4-c] [l]-benzopyran-1, 12-dione, CAS Nr. 1165-39-5] and aflatoxin G₂ (Figure 1) [3, 4, 7α, 9, 10, 10α-Hexahydro-5-methoxy-1*H*, 12*H* furo [3', 2':4,5] furo [2, 3-h] pyrano [3, 4-c] [l]-benzopyran-1, 12-dione, CAS Nr. 7241-98-7] are potent liver carcinogens. The aflatoxins are classified as group 1 carcinogens [5].




Figure 1. Molecular structures of aflatoxins

2 Participants

Project management and evaluation:

European Commission, Joint Research Centre, Institute for Reference Materials and Measurements, Reference Materials Unit, BE
(under current scope of ISO Guide 34 accreditation; Belac-268-Test)

Processing:

Wiertz-Eggert-Joerissen, DE
(accredited to ISO 17025 for measurement of aflatoxins in food and feed; DAP-PL-1453.99)

Homogeneity and stability measurements:

Central Science Laboratory, GB
(accredited to ISO 17025 for measurement of aflatoxins in food; UKAS 1642)

Nederlandse Organisatie voor toegepast-natuurwetenschappelijk onderzoek (TNO), NL
(accredited to ISO 17025 for measurement of aflatoxins in food and feed; RvA L027)

University for Natural Resources and Applied Life Sciences, Department of Agrobiotechnology IFA-Tulln, Environmental Biotechnology, AT

Wiertz-Eggert-Joerissen, DE
(accredited to ISO 17025 for measurement of aflatoxins in food and feed; DAP-PL-1453.99)

Certification analysis:

Central Science Laboratory, GB
(accredited to ISO 17025 for measurement of aflatoxins in food; UKAS 1642)

Finnish Customs Laboratory, FI
(accredited to ISO 17025 for measurement of aflatoxins in food and feed; FINAS T006)

Instituto Nacional de Engenharia e Tecnologia (INETI-LIA), PT
(accredited to ISO 17025 for measurement of aflatoxin B₁ in feed; IPAC L0094)

Laboratorio Normativo de Salud Pública de Bilbao, ES
(accredited to ISO 17025 for measurement of aflatoxins in food; ENAC 132/LE326)

LGC Ltd., GB
(accredited to ISO 17025 for measurement of aflatoxins in food and feed; UKAS 0003)

Nederlandse Organisatie voor toegepast-natuurwetenschappelijk onderzoek (TNO), NL
(accredited to ISO 17025 for measurement of aflatoxins in food and feed; RvA L027)

PhytoLab GmbH & Co. KG, DE
(accredited to ISO 17025 for measurement of contaminants by high performance liquid chromatography; SAL-BY-G037-01-05)

Wiertz-Eggert-Joerissen, DE
(accredited to ISO 17025 for measurement of aflatoxins in food and feed; DAP-PL-1453.99)

The timelines for major project tasks are summarized in Table 1.

Table 1. Schedule of the project

Study	Schedule
Processing	January 2002 - September 2002
Homogeneity study	April 2003
Stability studies	March 2003 -March 2007
Certification measurements	March 2006 - August 2006

3 Processing of the peanut materials

3.1 BCR-401R peanut butter (very low level)

250 kg of peanut butter from Lorenz Bahlsen Snack-World (Neuenburg, DE) were homogenised for one hour with 1.25 kg Lecithin EMULFLUID E from Lucas Meyer (Hamburg, DE) in a horizontal kneader from Döpke (Norden, DE).

After assessment of the bulk homogeneity, portions of more than 100 g were filled in aluminium cans. The paste was homogenised once more by using a drill and colour mixer before filling into the filling machine (Hamba, Wuppertal, DE). The aluminium cans were flushed with nitrogen and the lids heat sealed.

3.2 BCR-385R peanut butter (low level)

400 kg of Argentinean roasted peanut kernels were processed to a paste by a coronamill from Probst & Class (Rastatt, DE). The paste was mixed with 2 kg Lecithin EMULFLUID E from Lucas Meyer (Hamburg, DE).

35 kg of contaminated Chinese peanut kernels were roasted at 120 °C in a drying cabinet and processed to a paste by a coronamill from Probst & Class (Rastatt, DE).

170 kg of the Argentinean paste were pre-mixed with 35 kg of the Chinese paste and subsequently homogenised in a horizontal kneader from Döpke (Norden, DE).

After assessment of the bulk homogeneity, portions of more than 100 g were filled in aluminium cans. The paste was homogenised once more by using a drill and colour mixer before filling into the filling machine (Hamba, Wuppertal, DE). The aluminium cans were flushed with nitrogen and the lids heat sealed.

3.3 Additional characterisation measurements

3.3.1 Water content and water activity

The water content and water activity of the two peanut materials were measured by Karl Fischer titration [6] and the dew point technique [7], respectively. 15 bottles were chosen using a random stratified sample picking scheme and analysed in duplicate. The results are summarized in Table 2.

Table 2. Water content and water activity

Material	Water [g/kg]	Water activity (a_w)
BCR-401R	8.0 ± 0.2	0.174 ± 0.004
BCR-385R	16.0 ± 0.9	0.2023 ± 0.0019

3.3.2 Particle size measurements

Three particles size analyses have been carried out for each peanut butter material. The measurements were carried out employing a laser diffraction measuring device from Sympatec (Clausthal, DE). The materials were dispersed in methanol and the particle size distributions of the particles in the resulting suspension, were measured over the range from 0.5 to 875 µm. During the 10 s of measuring time, the samples were stirred with a magnetic stirring bar at 1200 rpm.

Representative graphs of the particle size distributions are shown in Figure 2 and Figure 3 for BCR-385R and BCR-401R, respectively.




Figure 2. Particle size distribution of BCR-385R




Figure 3. Particle size distribution of BCR-401R

4 Homogeneity studies

For the homogeneity study, 31 samples (about 2.0 % of the total batch) of BCR-385R were chosen using a random stratified sample picking scheme and analysed for their aflatoxin content in triplicate. Additionally 31 samples (about 2.0 % of the total batch) of BCR-385R and BCR-401R were chosen using a random stratified sample picking scheme and analysed for total nitrogen content according to Kjeldahl in duplicate.

Samples were measured in a random order to allow distinction between an analytical trend and a trend in the filling sequence. Measurements were performed on five different days. In order to exclude the influence of day to day variance the results were normalized to the mean of the results on this day. The normalized results on these five days (Annex A) were combined and evaluated to detect any trends regarding filling or analysis sequence and to estimate the uncertainty contribution from possible heterogeneity. Therefore the results were evaluated by a one-way analysis of variance (ANOVA). From the results of the ANOVA calculation, the following figures were calculated:

Method repeatability (s_{wb}) expressed as a relative standard deviation is given as follows:

$$s_{wb} = \frac{\sqrt{MS_{within}}}{\bar{y}}$$

MS_{within} : mean square within a bottle from an ANOVA

\bar{y} : average of all results of the homogeneity study

Between-unit variability (s_{bb}) expressed as a relative standard deviation is given by the following equation:

$$s_{bb} = \frac{\sqrt{\frac{MS_{between} - MS_{within}}{n}}}{\bar{y}}$$

$MS_{between}$: mean square among bottles from an ANOVA

n : average number of replicates per bottle

The heterogeneity that can be hidden by method repeatability is defined as follows:

$$u_{bb}^* = \frac{s_{wb}}{\sqrt{n}} \sqrt[4]{\frac{2}{v_{MS_{within}}}}$$

$v_{MS_{within}}$: degrees of freedom of MS_{within}

The larger value of s_{bb} or u_{bb}^* were used as uncertainty contribution for homogeneity, u_{bb} . The distribution of sample averages was checked employing normal probability plots for normal distribution and histograms for unimodal distribution. Data were checked for single and double outliers employing the Grubbs test at a level of confidence of 95 % and 99 %. Outliers were scrutinised but not excluded as no technical reason was found to do so.

Conclusions:

No significant trends regarding filling sequence were detected. Sample averages were not always normal-distributed but followed a unimodal distribution in each case, allowing the use of ANOVA for heterogeneity determination. The following outliers were found for BCR-385R: one single outlier (level of confidence 99%) for aflatoxin B₁, G₁ and G₂. The outlying means were scrutinised but no technical reason for exclusion was found. The individual contributions of heterogeneity to the uncertainty budget are summarized in Table 3 and Table 4.

Table 3. Results of homogeneity study for aflatoxins

BCR-385R	Mean (normalized)	RSD [%]	S _{wb} [%]	S _{bb} [%]	u* _{bb} [%]	u _{bb} ¹ [%]
Aflatoxin						
B ₁	1.00	4.2	7.3	- ²	1.8	1.8
B ₂	1.00	4.1	7.7	- ²	1.9	1.9
G ₁	1.00	16.6	17.5	13.2	4.3	13.2
G ₂	1.00	17.9	21.5	12.8	5.3	12.8

¹ higher value u*_{bb} or S_{bb} taken as contribution of heterogeneity

² cannot be calculated as MS_{within} > MS_{between}

Table 4. Results of homogeneity study for nitrogen

	Mean	RSD [%]	S _{wb} [%]	S _{bb} [%]	u* _{bb} [%]	u _{bb} ¹ [%]
BCR-385R						
Nitrogen	4.68 g/100 g	0.8	1.4	- ²	0.5	0.5
BCR-401R						
Nitrogen	3.90 g/100 g	0.9	1.5	- ²	0.5	0.5

¹ higher value u*_{bb} or S_{bb} taken as contribution of heterogeneity

² cannot be calculated as MS_{within} > MS_{between}

The minimum sample intake is 10 g. The stability and homogeneity studies were performed using 10 g of material proving that the individual samples are homogenous at least to this level.

5 Stability studies

Three stability studies were performed for BCR-385R, one isochronous study over 4 weeks to evaluate stability of the materials during transport and one isochronous study over 18 months and 36 months, respectively, to evaluate stability during storage. The stability of material BCR-401R has not been assessed as its content of aflatoxins is below limits of quantification.

For the short-term study, samples were stored in the dark at 18 °C, 40 °C and for reference at -70 °C. For the 18 months long-term study samples were stored in the dark at 4 °C, -20 °C and for reference at -70 °C and for the 36 months long-term study, samples were stored in the dark at -20 °C and for reference at -70 °C. Two units were stored at each temperature for 0, 1, 2 and 4 weeks (March 2003 - April 2003) for the short-term study and 0, 6, 12 and 18 months (March 2003 – September 2004) and 0, 12, 24 and 36 months (March 2003 – March 2006), respectively, for the long-term studies. After the indicated periods, the samples were transferred to -70 °C for storage until analysis. The results of both long term stability studies were combined by normalizing the individual result to the mean of the results of the zero month samples.

After the final time of each isochronous sample storage scheme the samples were measured together under repeatability conditions in a random order in duplicate. The laboratories employed their in-house HPLC-FLD methods. Documentation of the method employed is kept in the quality system of the respective laboratory. The mycotoxins were quantified using an external calibration and the peak area. The results were not corrected for recovery.

Results (Annex B) were tested for significant trends (degradation, enrichment) due to the storage conditions. Therefore the data points were plotted against time and the regression line calculated.

The uncertainty of stability u_{lts} of the materials was then calculated for the pre-defined shelf life as:

$$u_{lts} = \frac{RSD_{stab}}{\sqrt{\sum (x_i - \bar{x})^2}} \cdot x$$

with RSD_{stab} being the relative standard deviation of all 32 individual results of the relevant stability study, x_i being the time point for each replicate, \bar{x} being the average of all time points and x being the pre-defined shelf life (36 month in this case). Data were checked for single and double outliers employing the Grubbs test at a level of confidence of 95 and 99 %. Outliers were scrutinised but not excluded as no technical reason was found to do so.

Conclusions:

No significant slope at 95 % level of confidence was detected for the material, neither in the short-term study nor in the long-term study.

As no degradation could be observed under any conditions neither, in the short-term nor in the long-term study, it was concluded that no special precautions regarding temperature control during shipment are necessary. The uncertainty of the short-term stability (u_{sts}) is assumed to be negligible since no degradation is expected to happen during this short time.

The materials were stable in both long-term studies. Nevertheless -20 °C was chosen as storage temperature. Using the data from the long-term study, the uncertainty due to possible degradation was calculated for a storage time of 36 months at -20 °C. u_{its} for the materials are summarised in Table 5.

Table 5. Uncertainty contributions due to storage for BCR-385R

Aflatoxin	u_{its} [%]
B ₁	4.8
B ₂	4.5
G ₁	7.9
G ₂	11.5

6 Certification

6.1 Design of the study

The certification exercise was performed in 2006. Eight laboratories were carefully selected to perform the analytical measurements. The laboratories had to prove their measurement capabilities and had to demonstrate experience in the analytical field concerned.

Each laboratory was provided with the following samples:

- 3 units of “Peanut butter (low level aflatoxins)” BCR-385R
- 6 units of “Peanut butter (very low level aflatoxins)” BCR-401R
- 3 ampoules of the common calibrant aflatoxin B₁ in acetonitrile, ERM-AC057
- 3 ampoules of the common calibrant aflatoxin B₂ in acetonitrile, ERM-AC058
- 3 ampoules of the common calibrant aflatoxin G₁ in acetonitrile, ERM-AC059
- 3 ampoules of the common calibrant aflatoxin G₂ in acetonitrile, ERM-AC060

The measurements had to be performed on three different days. On each day one unit of BCR-385R and BCR-401R had to be analysed in duplicate. Additionally individual recovery experiments had to be carried out on each day with the additional sets of BCR-401R.

In order to reveal the recovery rates of the participant’s analytical procedures in the concentration range of the naturally contaminated candidate reference materials, the participants had to spike the “Peanut butter (very low level aflatoxins)”, BCR-401R with a solution of the common calibrants to a mass fraction of each aflatoxin (B₁, B₂, G₁, G₂) of 0.5 µg/kg. No strict procedure was given on how to perform the spiking.

Analyses of the ‘blank’, spiked and, naturally contaminated peanut materials had to be performed together on each day, since the certification results are corrected by the daily recovery factor. External calibrations were based on dilutions of the provided common calibrants. A new calibration had to be performed on each day. The measurement program is visualised in Table 6.

Table 6. Measurement program

Day 1	Day 2	Day 3
Calibration	Calibration	Calibration
1 unit of BCR-385R in duplicate	1 unit of BCR-385R in duplicate	1 unit of BCR-385R in duplicate
1 unit of BCR-401R in duplicate	1 unit of BCR-401R in duplicate	1 unit of BCR-401R in duplicate
1 unit of BCR-401R spiked in triplicate	1 unit of BCR-401R spiked in triplicate	1 unit of BCR-401R spiked in triplicate

6.2 Results and technical evaluation

All laboratories used their in-house methods based on immunoaffinity column (IAC) clean up and reversed phase high performance liquid chromatography with post column derivatisation and fluorescence detection. The individual methods used are summarized in Table 7. Laboratory 6 did not submit any results for technical reasons.

Table 7. Overview of analytical methods used for certification

Lab code	Extraction solvent	Extraction technique	Defatting	IAC producer	Chromatography	Derivatisation
1	Methanol+ water	Blender	Hexane	R-biopharm rhône	Isocratic	PBPB ¹
2	Methanol+ water	Shaker	Hexane	Vicam	Isocratic	PBPB
3	Methanol+ water	Blender	Hexane	Vicam	Isocratic	Kobra cell ²
4	Chloroform+ water	Shaker	Hexane	R-biopharm rhône	Isocratic	Kobra cell
5	Acetonitrile+ water	Blender		R-biopharm rhône	Isocratic	PBPB
7	Acetonitrile+ water	Blender		R-biopharm rhône	Isocratic	Kobra cell
8	Acetonitrile+ water	Ultrasonic bath		Vicam	Gradient	Kobra cell
9	Methanol+ water	Blender	Hexane	R-biopharm rhône	Isocratic	PBPB

¹ bromination with pyridinium hydrobromide perbromide

² electrochemical bromination with potassium bromide

After reception of the data sets, the results were subject to technical evaluation. Results not fulfilling the criteria laid down in Commission Regulation 401/2006 [8] regarding recovery rates were eliminated. Satisfactory recovery rates for the spiking level of 0.5 µg/kg are between 50 % and 120 % [8]. Satisfactory within-laboratory reproducibility is below 50 % based on the Horwitz equation, which all laboratories fulfilled. If the criteria for one measurand were not met on two separate days, the third day results were omitted as well.

This led to the rejection of the following data sets due to not satisfactory recovery rates:

Lab 1: the results of all three days for aflatoxin G₂ in BCR-385R were rejected

Lab 2: the results of one day for aflatoxin B₁ in BCR-385R were rejected

Lab 3: no data was rejected

Lab 4: the results of all three days for aflatoxins G₁ and G₂ in BCR-385R were rejected

Lab 5: no data was rejected

Lab 7: no data was rejected

Lab 8: the results of all three days for aflatoxins G₁ and G₂ in BCR-385R were rejected

Lab 9: the results of one day for aflatoxin B₁ in BCR-385R were rejected

In total 44 values for aflatoxin B₁ from 8 labs, 48 values for aflatoxin B₂ from 8 labs, 36 values of aflatoxins G₁ from 6 labs and 30 values of aflatoxins G₂ for from 5 labs for BCR-385R were accepted after technical scrutiny for further statistical data assessment.

The accepted sets of results were submitted to the following statistical tests:

- Scheffe's multiple t-test to check if the means of two labs are significantly different
- Dixon's test to detect outlying lab means
- Nalimov t-test to detect outlying lab means
- Grubb's test to detect single and double outliers
- Cochran test to check for outlying lab variances
- Bartlett test to check for homogeneity of lab variances
- ANOVA to assess between lab and within lab variances and test their significance employing the SNEDECOR F-test
- Skewness and kurtosis test to assess the normality of the lab means distribution. The later tests are only used if seven or more datasets have been accepted; otherwise normal probability plot have been used.

First, the datasets have been subjected to the Cochran test to check for outlying lab variances. Datasets with outlying variances have been rejected on the basis that all laboratories used a similar method and an outlying variance indicates poor repeatability and therefore a suboptimal control of method performance. The following datasets have been rejected:

- Lab 2 data of aflatoxins B₂ for BCR-385R
- Lab 5 data of aflatoxins G₂ for BCR-385R
- Lab 8 data of aflatoxins B₁ and B₂ for BCR-385R

Table 8. Summary of statistical evaluation for BCR-385R

BCR-385R				
Aflatoxin	B ₁	B ₂	G ₁	G ₂
Number of data sets	7	6	6	4
Number of replicate measurements	38	36	36	24
Mean of means [$\mu\text{g}/\text{kg}$]	1,77	0,48	0,92	0,30
Standard deviation [%]	16,7	14,1	16,1	19,5
Relative standard error [%]	6,3	5,8	6,6	9,8
All data sets compatible two by two? (Scheffe's test)	no	no	no	no
Outlying means? (Dixon test, Nalimov t-test, Grubbs test)	Lab2 Nalimov ($p=0.05$)	no	no	no
Outlying lab variances? (Cochran test)	no	no	no	no
Lab variances homogeneous? (Bartlett test)	yes ($p=0.01$)	yes	yes ($p=0.01$)	yes
Distribution of means normal? (Skewness & kurtosis, normal probability plot)	yes	yes	yes	yes
Variances between labs significantly different? (SNEDECOR)	yes	yes	yes	yes

The accepted individual results after technical and statistical scrutiny are given in Annex C. The results of the statistical tests of the finally considered data for BCR-385R are summarized in Table 8.

The individual results are corrected by the daily recovery. The uncertainty of the daily recovery does not contribute to the overall uncertainty as the relative standard error of the mean of means is used as an estimation of the uncertainty contribution of the characterisation exercise.

The outlying mean of aflatoxins B₁ for BCR-385R lies within two standard deviations of the mean of means and is therefore not significantly different to the mean value.

Only data sets which satisfied the criteria for recovery rate, as mentioned above, were scrutinized and taken into account for the value assessment of BCR-401R (Annex C). The values of laboratory 5 were not taken into account as the limit of detection and limit of quantification were calculated on the basis that BCR-401R is a real blank and are therefore erroneous. The limits of quantification of the methods used were equal to or less than 0.2 µg/kg except for the method used by laboratory 5, where limits of quantification for aflatoxin B₁ and G₁ were higher.

Five values for aflatoxin B₁ were below limit of detection with the highest limit of detection of 0.09 µg/kg (Table 9). Nine values were below the limit of quantification with the highest limit of quantification of 0.2 µg/kg. Two quantified values of 0.19 and 0.16 µg/kg were found as well.

Table 9 Results of characterisation measurements for aflatoxin B₁ of BCR401R

Aflatoxin B ₁ mass fraction in BCR-401R [µg/kg]					
Lab code	Day 1	Day 2	Day 3	LOQ [µg/kg]	LOD [µg/kg]
1	< LOQ	< LOQ	< LOD	0.2	0.05
2	< LOD	< LOD	-	0.13	0.07
3	< LOQ	0.19	< LOQ	0.18	0.07
4	< LOD	< LOD	< LOQ	0.05	0.02
7	< LOQ	< LOQ	< LOQ	0.2	0.02
9	-	0.16	< LOQ	0.14	0.09

Although all values are below 0.2 µg/kg it can not be conclude with a confidence of 95 % that the values is below 0.2 µg/kg. The probability of the value being below 0.2 µg/kg was assessed by calculating the average and its standard deviation. The average was calculated by using the limit of detection for values below the limit of detection and the limit of quantification for values below the limit of quantification but above the limit of detection. This results in an average of 0.10 µg/kg and a standard deviation of 0.08 µg/kg, assuming normal distribution. Therefore it can be concluded that the value is below 0.2 µg/kg with a probability of 80 %.

For comparison the assumption of a rectangular distribution with range from 0 µg/kg to 0.2 µg/kg was made. The uncertainty of the rectangular distribution, expressed as half of the range divided by the square root of three (0.06 µg/kg), would result in a 90 % confidence of the level being below 0.2 µg/kg.

The assumption of a normal distribution was chosen for certification.

Values of four laboratories for aflatoxin B₂ are below limit of detection with the highest limit of detection at 0.03 µg/kg (Table 10). One laboratory reported values of 0.14, 0.10 and one below 0.1 µg/kg (limit of quantification). As all other laboratories report values below 0.03 µg/kg this was taken as an indication that laboratory 3 had a contamination in the analytical process. Nevertheless the two values are below 0.2 µg/kg.

Table 10 Results of characterisation measurements for aflatoxin B₂ of BCR401R

Aflatoxin B₂ mass fraction in BCR-401R [$\mu\text{g}/\text{kg}$]					
Lab code	Day 1	Day 2	Day 3	LOQ [$\mu\text{g}/\text{kg}$]	LOD [$\mu\text{g}/\text{kg}$]
1	< LOD	< LOD	< LOD	0.1	0.03
3	0.10	0.14	< LOQ	0.1	0.03
4	< LOD	< LOD	< LOD	0.05	0.02
7	< LOD	< LOD	< LOD	0.2	0.02
9	< LOD	< LOD	< LOD	0.05	0.03

All values for aflatoxin G₁ except two were below the limits of detection (Table 11). The two values above the limit of detection were nevertheless far below the limit of quantification of 0.2 $\mu\text{g}/\text{kg}$ and below the limit of detection of the other laboratories. The highest limit of detection is 0.15 $\mu\text{g}/\text{kg}$. Although one laboratory has a limit of detection above 0.2 $\mu\text{g}/\text{kg}$ the value is certified as below 0.2 $\mu\text{g}/\text{kg}$ based on the more sensitive methods used by the other laboratories.

Table 11 Results of characterisation measurements for aflatoxin G₁ of BCR401R

Aflatoxin G₁ mass fraction in BCR-401R [$\mu\text{g}/\text{kg}$]					
Lab code	Day 1	Day 2	Day 3	LOQ [$\mu\text{g}/\text{kg}$]	LOD [$\mu\text{g}/\text{kg}$]
1	< LOD	< LOD	< LOD	0.2	0.05
2	< LOD	< LOD	< LOD	0.1	0.07
3	< LOD	< LOD	< LOD	0.18	0.09
7	< LOQ	< LOQ	< LOD	0.2	0.02
9	< LOD	< LOD	< LOD	0.24	0.15

All values for aflatoxin G₂ were below the limits of detection with the highest limit of detection at 0.07 $\mu\text{g}/\text{kg}$ (Table 12).

Table 12 Results of characterisation measurements for aflatoxin G₂ of BCR401R

Aflatoxin G₂ mass fraction in BCR-401R [$\mu\text{g}/\text{kg}$]					
Lab code	Day 1	Day 2	Day 3	LOQ [$\mu\text{g}/\text{kg}$]	LOD [$\mu\text{g}/\text{kg}$]
2	< LOD	< LOD	< LOD	0.07	0.04
3	< LOD	< LOD	< LOD	0.1	0.03
7	< LOD	< LOD	< LOD	0.2	0.02
9	< LOD	< LOD	< LOD	0.12	0.07

6.3 Certified values and their uncertainties

The certified values for BCR-385R are calculated as the mean of means of the accepted datasets. The standard error of the mean of means was used as an estimation of the uncertainty contribution of the characterisation exercise to the mass fractions of the aflatoxins. The standard error is calculated as the standard deviation divided by the square root of accepted data sets.

The combined uncertainty of the certified value includes contributions from the between bottle heterogeneity, long-term storage, the characterisation study and the contribution of the common calibrant. The uncertainty of the mass fraction (aflatoxin in acetonitrile) of the common calibrants propagates in the calibrations and can therefore not be neglected.

The relative combined uncertainty is calculated as the square root of the sum of squares of the relative uncertainties of the individual contributions, according to:

$$u_{CRM} = \sqrt{u_{lts}^2 + u_{bb}^2 + u_{char}^2 + u_{cal}^2}$$

The absolute, expanded uncertainty $U_{CRM, abs}$ is calculated by multiplying the certified value with the relative, expanded uncertainty U_{CRM} .

The values for BCR-401R are certified as to be with a 95 % probability less than 0.2 µg/kg except for aflatoxin B₁, where the probability is 80 %. This is in agreement with the set of accepted results including four results with mass fractions of less than 0.2 µg/kg and the majority of results showing mass fractions below the limits of quantification or even limit of detection of the respective methods.

The certified values are summarised in Table 13 and Table 14.

Table 13. Certified values and their uncertainties for BCR-385R

BCR-385R				
Aflatoxin	B ₁	B ₂	G ₁	G ₂
Certified value [µg/kg]	1,77	0,48	0,9	0,30
u_{lts} [%]	4,8	4,5	7,9	11,5
u_{bb} [%]	1,8	1,9	13,2	12,8
u_{char} [%]	6,3	5,8	6,6	9,8
u_{cal} [%]	1,4	1,0	1,7	1,0
u_{CRM} [%]	8,2	7,6	16,8	19,8
U_{CRM} (k=2) [%]	16,5	15,3	33,7	39,7
U_{CRM} (k=2) [µg/kg]	0,30	0,08	0,4	0,12

Additionally to the values given in Table 13 BCR-385R is certified for the sum of aflatoxins B₁, B₂, G₁ and G₂. The sum of aflatoxins B₁, B₂, G₁ and G₂ is 3.5 ± 0.5 µg/kg. The expanded uncertainty for the sum of aflatoxins B₁, B₂, G₁ and G₂ is calculated from the individual absolute uncertainties according to:

$$U_{sum} = 2 \cdot \sqrt{u_{B_1}^2 + u_{B_2}^2 + u_{G_1}^2 + u_{G_2}^2}$$

Table 14. Certified values for BCR-401R

BCR-401R				
Aflatoxin	B ₁	B ₂	G ₁	G ₂
Certified value [µg/kg]	< 0.2 ¹	< 0.2 ²	< 0.2 ²	< 0.2 ²

¹ with a 80 % probability, the measurand content is below this level

² with a 95 % probability, the measurand content is below this level

7 Metrological traceability and commutability

The aflatoxins mass fractions as stated are defined by the employed reversed phase liquid chromatography methods with post column bromination, fluorescence detection and immunoaffinity clean up. As three different solvents and extraction techniques have been used independence from the extraction method can be assumed.

The certified values for the mass fractions of aflatoxins are traceable via the common calibrants used. The mass fractions of the common calibrants are certified for aflatoxins in acetonitrile. The certified value of the calibrants is traceable to SI due to the gravimetric preparation employed. Therefore the mass fractions of aflatoxins in the CRMs BCR-385R and BCR-401R are traceable to the SI.

BCR-385R is prepared from naturally contaminated material. Therefore there is no reason to assume that BCR-385R would behave differently from natural samples with similar particle size.

8 Instructions for use

8.1 Storage conditions

The materials should be stored at or below -20 °C. However, the European Commission cannot be held responsible for changes that happen during storage of the material at the customer's premises, especially of open samples.

8.2 Safety precautions

The usual laboratory safety precautions apply.

8.3 Use of the material

This material is intended to be used for method performance control and validation purposes. Samples should be allowed to warm to ambient temperature (e.g. overnight) before opening to avoid water condensation. The contents should be thoroughly mixed before sub-samples of at least 10 g are taken. The peanut meal should be weighed out immediately after opening the sachets and the mass fractions of the aflatoxins calculated based on this mass.

8.4 Use of the certified value

For assessing the method performance, the measured values of the CRMs are compared with the certified values following a procedure described by Linsinger [9]. The procedure is described here in brief:

- Calculate the absolute difference between mean measured value and the certified value (Δ_m).
- Combine measurement uncertainty (u_{meas}) with the uncertainty of the certified value (u_{CRM}): $u_{\Delta} = \sqrt{u_{meas}^2 + u_{CRM}^2}$
- Calculate the expanded uncertainty (U_{Δ}) from the combined uncertainty (u_{Δ}) using a coverage factor of two ($k = 2$), corresponding to a confidence interval of approximately 95 %
- If $\Delta_m \leq U_{\Delta}$ then there is no significant difference between the measurement result and the certified value, at a confidence level of about 95 %.

9 Acknowledgements

The authors would like to thank Jacob de Boer (Free University Amsterdam, NL), Petra Gowik (Federal Office of Consumer Protection and Food Safety, DE), Erik Nordkvist (National Veterinary Institute, SVA, SE), Olivier Couteau (IRMM, BE) and Ingrid Zegers (IRMM, BE) for the review of this report. Additionally the authors would like to thank Thomas Linsinger (IRMM, BE) for his technical advise.

10 References

- 1 International Organization for Standardization, Guide to the Expression of Uncertainty in Measurement, (1995) ISO, Geneva, Switzerland
- 2 European Commission, Health & Consumer Protection Directorate General, The Rapid Alert System for Food and Feed (RASFF) Annual Report 2006, http://ec.europa.eu/food/food/rapidalert/report2006_en.pdf
- 3 Commission Regulation EC/1525/98; OJ L 201, 17.7.1998, p. 43
- 4 Commission Regulation EC/1881/2006; OJ L 364, 20.12.2006, p. 5
- 5 World Health Organisation, International Agency for Research on Cancer, IARC Monographs on the Evaluation of Carcinogenic Risks to Humans, Volume 82, (2002), <http://monographs.iarc.fr>
- 6 V. Kestens, P. Conneely, A. Bernreuther, Vaporisation coulometric Karl Fischer titration: A perfect tool for water content determination of difficult matrix reference materials, Food Chemistry, 106 (2008) 1454 - 1459
- 7 S. Rückold, K. H. Grobecker, H. – D. Isengard, Water as a source of errors in reference materials, Fresenius Journal of Analytical Chemistry, 370 (2001) 189 - 193
- 8 Commission Regulation EC/401/2006, OJ L 70, 9.3.2006, p.12
- 9 T. P. J. Linsinger, Comparison of measurement result with the certified value, Application Note 1, July 2005, <http://www.erm-crm.org>

Annex A. Homogeneity data

Table A1. Results of homogeneity study for BCR385R

BCR-385R						
	means (n=3)					mean (n=2)
bottle	B1 [$\mu\text{g}/\text{kg}$]	B2 [$\mu\text{g}/\text{kg}$]	G1 [$\mu\text{g}/\text{kg}$]	G2 [$\mu\text{g}/\text{kg}$]	bottle	nitrogen [g/ 100 g]
30	1.81	0.56	0.74	0.22	20	4.72
82	1.80	0.56	0.78	0.26	57	4.67
143	1.79	0.50	0.78	0.26	124	4.71
176	1.84	0.57	1.13	0.49	157	4.71
238	1.82	0.58	0.75	0.27	215	4.69
289	1.85	0.51	1.27	0.39	261	4.64
340	1.75	0.53	0.81	0.25	318	4.67
397	1.88	0.62	0.71	0.23	353	4.69
432	1.77	0.57	0.88	0.28	408	4.67
478	1.80	0.56	0.69	0.31	468	4.65
529	1.73	0.51	0.70	0.19	518	4.75
597	1.72	0.55	0.86	0.34	564	4.68
646	1.78	0.59	0.84	0.30	618	4.69
678	1.53	0.46	0.66	0.20	664	4.68
741	1.86	0.62	0.83	0.26	709	4.66
790	1.77	0.55	0.90	0.29	776	4.67
843	1.71	0.55	0.74	0.23	819	4.75
894	1.51	0.47	0.62	0.21	859	4.64
945	1.78	0.57	0.73	0.23	922	4.71
981	1.77	0.54	0.79	0.26	967	4.64
1036	1.57	0.47	0.68	0.22	1020	4.68
1095	1.80	0.58	0.94	0.32	1066	4.75
1136	1.90	0.58	0.78	0.35	1102	4.77
1195	1.56	0.46	0.73	0.23	1153	4.63
1242	1.71	0.54	0.82	0.28	1203	4.67
1292	1.80	0.54	0.76	0.35	1254	4.62
1329	1.76	0.57	0.72	0.23	1312	4.67
1390	1.67	0.51	0.69	0.19	1377	4.70
1439	1.66	0.47	0.66	0.18	1409	4.69
1494	1.81	0.57	0.90	0.28	1466	4.70
1524	1.81	0.56	0.99	0.31	1513	4.67

Table A2. Results of homogeneity study normalized by the mean of each measurement day

BCR-385R				
	means (n=3) normalized			
bottle	B1	B2	G1	G2
30	1.002	0.963	0.898	0.836
82	0.996	0.997	0.956	0.999
143	1.070	1.047	1.102	1.185
176	1.014	1.017	1.334	1.329
238	1.026	1.017	0.931	0.944
289	1.140	1.062	1.604	1.608
340	0.999	0.980	1.019	1.042
397	1.041	1.087	0.870	0.864
432	0.982	0.992	1.064	1.089
478	0.994	0.994	0.815	0.847
529	0.987	0.927	0.890	0.792
597	0.951	0.986	1.017	0.934
646	1.006	1.029	1.050	1.051
678	0.943	0.955	0.825	0.820
741	1.031	1.065	1.002	0.967
790	0.980	0.963	1.103	1.094
843	0.976	1.020	0.927	0.935
894	0.928	0.978	0.783	0.853
945	0.986	0.977	0.891	0.888
981	0.983	0.956	0.968	0.975
1036	0.941	0.979	0.966	0.992
1095	0.999	1.003	1.145	1.220
1136	1.048	1.037	0.928	0.944
1195	0.959	0.940	0.926	0.938
1242	0.968	0.954	1.019	1.004
1292	0.992	0.966	0.906	0.946
1329	1.005	1.047	0.912	0.936
1390	1.029	1.065	0.862	0.781
1439	0.990	0.974	0.932	0.823
1494	1.000	0.998	1.104	1.068
1524	1.033	1.027	1.252	1.294

Aflatoxin B1 BCR-385R - Histogram Plot




Figure A1. Histogram of homogeneity results for aflatoxin B₁

Aflatoxin B1 BCR-385R- Graph




Figure A2. Results of homogeneity study for aflatoxin B₁ sorted by the filling sequence

Aflatoxin B2 BCR-385R - Histogram Plot




Figure A3. Histogram of homogeneity results for aflatoxin B₂

Aflatoxin B2 BCR-385R- Graph




Figure A4. Results of homogeneity study for aflatoxin B₂ sorted by the filling sequence

Aflatoxin G1 BCR-385R - Histogram Plot




Figure A5. Histogram of homogeneity results for aflatoxin G₁

Aflatoxin G1 BCR-385R- Graph




Figure A6. Results of homogeneity study for aflatoxin G₁ sorted by the filling sequence

Aflatoxin G2 BCR-385R - Histogram Plot




Figure A7. Histogram of homogeneity results for aflatoxin G₂

Aflatoxin G2 BCR-385R- Graph




Figure A8. Results of homogeneity study for aflatoxin G₂ sorted by the filling sequence

Kjeldahl BCR-385R - Histogram Plot




Figure A9. Histogram of homogeneity results for nitrogen

Kjeldahl BCR-385R- Graph




Figure A10. Results of homogeneity study for nitrogen sorted by the filling sequence

Table A3. Results of homogeneity study for BCR401R

BCR-401R	
	mean (n=2)
bottle	nitrogen [g/ 100 g]
4	3.91
74	3.83
146	3.91
191	3.92
236	3.90
275	3.84
309	3.91
355	3.96
422	3.97
469	3.92
507	3.95
584	3.85
669	3.88
692	3.91
737	3.95
834	3.93
904	3.90
975	3.86
1008	3.92
1062	3.87
1123	3.87
1175	3.89
1249	3.92
1317	3.88
1370	3.95
1432	3.86
1495	3.90
1507	3.87
1549	3.83
1570	3.88
1591	3.91

Kjeldahl BCR401R - Histogram Plot




Figure A11. Histogram of homogeneity results for nitrogen of BCR401R




Figure A12. Results of homogeneity study for nitrogen sorted by the filling sequence of BCR401R

Annex B. Stability data

Table B1. Results of the isochronous studies for aflatoxin B₁

BCR-385R B1 [$\mu\text{g}/\text{kg}$]						
weeks at 18 °C						
samples	0	1	2	4		
1	1.71	1.58	1.72	1.79		
2	1.59	1.55	1.30	1.45		
3	1.59	1.89	1.69	1.56		
4	1.57	1.57	1.37	1.85		
weeks at 40 °C						
samples	0	1	2	4		
1	1.71	1.62	1.79	1.47		
2	1.59	1.61	1.17	1.50		
3	1.59	1.61	1.60	1.37		
4	1.57	1.59	1.46	1.60		
months at -20 °C						
samples	0	6	12	18		
1	1.68	1.53	1.54	1.54		
2	1.54	1.50	1.63	1.58		
3	1.55	1.58	1.15	1.63		
4	1.97	1.49	1.55	1.50		
months at 4 °C						
samples	0	6	12	18		
1		1.33	1.17	0.84		
2	1.54	1.25	1.25	1.46		
3	1.21	1.01	1.03	1.5		
4	1.33	1.39	1.42	1.39		
months at -20 °C						
samples	0	12	24	36		
1	0.89	0.77	0.85	0.85		
2	0.88	0.78	0.77	0.89		
3	0.85	0.81	0.77	0.74		
4	0.77	0.85	0.77	0.74		
months at -20 °C (combined and normalised)						
samples	0	6	12	18	24	36
1	1.00	0.91	0.91	0.91	1.00	1.00
2	0.91	0.89	0.97	0.94	0.91	1.05
3	0.92	0.94	0.68	0.97	0.91	0.87
4	1.17	0.88	0.92	0.89	0.91	0.87
5	1.05		1.00			
6	1.04		0.91			
7	1.00		0.91			
8	0.91		0.91			

BCR-385R Aflatoxin B₁ T=18 °C




Figure B1. Short term stability for aflatoxin B₁ at 18 °C

BCR-385R Aflatoxin B₁ T=40 °C




Figure B2. Short term stability for aflatoxin B₁ at 40 °C

Shelf Life and Associated Ults, T=-20 °C




Figure B3. Long term stability for aflatoxin B₁ at -20 °C with associated u_{lts} for a storage period of 36 month

Table B2. Results of the isochronous studies for aflatoxin B₂

BCR-385R B2 [$\mu\text{g}/\text{kg}$]						
weeks at 18 °C						
samples	0	1	2	4		
1	0.49	0.48	0.53	0.58		
2	0.43	0.41	0.43	0.44		
3	0.47	0.48	0.51	0.46		
4	0.44	0.44	0.45	0.58		
weeks at 40 °C						
samples	0	1	2	4		
1	0.49	0.56	0.53	0.48		
2	0.43	0.45	0.43	0.49		
3	0.47	0.48	0.49	0.43		
4	0.44	0.48	0.47	0.44		
months at -20 °C						
samples	0	6	12	18		
1	0.45	0.48	0.53	0.47		
2	0.45	0.44	0.47	0.49		
3	0.46	0.49	0.42	0.46		
4	0.59	0.46	0.47	0.48		
months at 4 °C						
samples	0	6	12	18		
1		0.24	0.21	0.14		
2	0.26	0.3	0.28	0.28		
3	0.22	0.12	0.16	0.29		
4	0.25	0.24	0.33	0.29		
months at -20 °C						
samples	0	12	24	36		
1	0.23	0.19	0.19	0.2		
2	0.23	0.19	0.19	0.23		
3	0.19	0.19	0.19	0.19		
4	0.19	0.23	0.19	0.19		
months at -20 °C (combined and normalised)						
samples	0	6	12	18	24	36
1	0.92	0.99	1.09	0.96	0.91	0.95
2	0.92	0.90	0.96	1.01	0.91	1.10
3	0.94	1.01	0.86	0.94	0.91	0.91
4	1.21	0.94	0.96	0.99	0.91	0.91
5	1.10		0.91			
6	1.10		0.91			
7	0.91		0.91			
8	0.91		0.91			

BCR-385R Aflatoxin B₂ T=18 °C




Figure B4. Short term stability for aflatoxin B₂ at 18 °C

BCR-385R Aflatoxin B₂ T=40 °C




Figure B5. Short term stability for aflatoxin B₂ at 40 °C

Shelf Life and Associated U_{lts}, T=-20 °C




Figure B6. Long term stability for aflatoxin B₂ at -20 °C with associated u_{lts} for a storage period of 36 month

Table B3. Results of the isochronous studies for aflatoxin G₁

BCR-385R G ₁ [$\mu\text{g}/\text{kg}$]						
weeks at 18 °C						
samples	0	1	2	4		
1	0.68	0.76	0.70	0.83		
2	0.65	0.76	0.59	0.65		
3	0.68	0.66	1.03	0.69		
4	0.95	0.62	0.57	1.86		
weeks at 40 °C						
samples	0	1	2	4		
1	0.68	0.63	0.69	0.70		
2	0.65	0.68	0.50	0.67		
3	0.68	0.64	0.80	0.59		
4	0.95	0.76	0.66	0.65		
months at -20 °C						
samples	0	6	12	18		
1	1.24	0.79	0.79	0.85		
2	0.86	0.91	0.80	0.74		
3	0.75	0.80	0.93	0.84		
4	0.99	0.67	0.87	0.74		
months at 4 °C						
samples	0	6	12	18		
1		0.77	0.5	0.37		
2	0.72	0.55	0.48	0.59		
3	0.64	0.35	0.48	0.88		
4	0.59	0.51	0.78	0.83		
months at -20 °C						
samples	0	12	24	36		
1	0.46	0.5	0.46	0.54		
2	0.5	0.52	0.39	0.54		
3	0.43	0.39	0.39	0.5		
4	0.43	0.43	0.46	0.46		
months at -20 °C (combined and normalised)						
samples	0	6	12	18	24	36
1	1.29	0.82	0.82	0.89	1.01	1.19
2	0.90	0.95	0.83	0.77	0.86	1.19
3	0.78	0.83	0.97	0.88	0.86	1.10
4	1.03	0.70	0.91	0.77	1.01	1.01
5	1.01		1.01			
6	1.10		0.86			
7	0.95		0.86			
8	0.95		1.01			

BCR-385R Aflatoxin G₁ T=18 °C




Figure B7. Short term stability for aflatoxin G₁ at 18 °C

BCR-385R Aflatoxin G₁ T=40 °C




Figure B8. Short term stability for aflatoxin G₁ at 40 °C

Shelf Life and Associated U_{lts}, T=-20 °C




Figure B9. Long term stability for aflatoxin G₁ at -20 °C with associated u_{lts} for a storage period of 36 month

Table B4. Results of the isochronous studies for aflatoxin G₂

BCR-385R G₂ [µg/kg]						
weeks at 18 °C						
samples	0	1	2	4		
1	0.24	0.22	0.19	0.25		
2	0.20	0.23	0.19	0.23		
3	0.21	0.20	0.36	0.19		
4	0.18	0.18	0.22	0.78		
weeks at 40 °C						
samples	0	1	2	4		
1	0.24	0.18	0.20	0.21		
2	0.20	0.22	0.18	0.19		
3	0.21	0.21	0.24	0.17		
4	0.18	0.25	0.20	0.22		
months at -20 °C						
samples	0	6	12	18		
1	0.44	0.22	0.23	0.28		
2	0.26	0.20	0.25	0.20		
3	0.18	0.22	0.31	0.24		
4	0.37	0.22	0.25	0.22		
months at 4 °C						
samples	0	6	12	18		
1		0.17	0.05	0.03		
2	0.17	0.08	0.09	0.12		
3	0.13	0.07	0.09	0.25		
4	0.09	0.07	0.16	0.23		
months at -20 °C						
samples	0	12	24	36		
1	0.15	0.12	0.15	0.15		
2	0.14	0.15	0.12	0.15		
3	0.12	0.12	0.12	0.12		
4	0.15	0.12	0.15	0.15		
months at -20 °C (combined and normalised)						
samples	0	6	12	18	24	36
1	1.41	0.70	0.74	0.90	1.07	1.07
2	0.83	0.64	0.80	0.64	0.86	1.07
3	0.58	0.70	0.99	0.77	0.86	0.86
4	1.18	0.70	0.80	0.70	1.07	1.07
5	1.07		1.07			
6	1.00		0.86			
7	0.86		0.86			
8	1.07		1.07			

BCR-385R Aflatoxin G₂ T=18 °C




Figure B10. Short term stability for aflatoxin G₂ at 18 °C

BCR-385R Aflatoxin G₂ T=40 °C




Figure B11. Short term stability for aflatoxin G₂ at 40 °C

Shelf Life and Associated U_{ITS}, T=-20 °C




Figure B12. Long term stability for aflatoxin G₂ at -20 °C with associated u_{ITS} for a storage period of 36 month

Annex C. Certification measurements

Table C1. Results of characterisation measurements for aflatoxin B₁ of BCR385R

B1 mass fraction in BCR-385R [$\mu\text{g}/\text{kg}$]								
Lab code	Day 1		Day 2		Day 3		Mean	s
1	1.56	1.60	1.30	1.29	1.76	1.39	1.48	0.19
2	2.24	2.89	2.05	2.17	-	-	2.34	0.38
3	1.50	1.51	2.11	2.11	1.41	1.42	1.68	0.34
4	2.10	1.91	1.80	1.95	1.72	1.04	1.75	0.37
5	1.66	1.69	1.85	1.86	1.68	1.95	1.78	0.12
7	1.83	2.23	1.48	1.84	1.90	1.94	1.87	0.24
9	-	-	1.36	1.50	1.49	1.46	1.45	0.06

No Pooling - Lab Means and their StDev for Certification BCR-385R B1




Figure C1 Laboratory means, mean of means and their standard deviation for aflatoxin B₁

Table C2. Results of characterisation measurements for aflatoxin B₂ of BCR385R

B2 mass fraction in BCR-385R [$\mu\text{g}/\text{kg}$]								
Lab code	Day 1		Day 2		Day 3		Mean	s
1	0.54	0.47	0.39	0.35	0.41	0.29	0.41	0.09
3	0.41	0.41	0.49	0.50	0.40	0.40	0.44	0.05
4	0.55	0.49	0.48	0.54	0.53	0.32	0.49	0.09
5	0.57	0.59	0.62	0.61	0.48	0.59	0.58	0.05
7	0.60	0.60	0.48	0.50	0.50	0.55	0.54	0.05
9	0.38	0.42	0.43	0.47	0.44	0.42	0.43	0.03

No Pooling - Lab Means and their StDev for Certification BCR-385R B2




Figure C2 Laboratory means, mean of means and their standard deviation for aflatoxin B₂

Table C3. Results of characterisation measurements for aflatoxin G₁ of BCR385R

G1 mass fraction in BCR-385R [$\mu\text{g}/\text{kg}$]								
Lab code	Day 1		Day 2		Day 3		Mean	s
1	1.02	1.04	0.85	0.74	0.82	0.64	0.85	0.156
2	1.15	1.25	1.21	1.27	0.80	1.19	1.15	0.174
3	0.96	0.97	0.81	0.82	0.41	0.42	0.73	0.254
5	0.75	1.16	1.26	1.11	0.58	1.12	1.00	0.268
7	1.04	1.15	0.75	0.82	1.16	0.93	0.98	0.171
9	0.85	0.87	0.80	0.77	0.85	0.78	0.82	0.042

No Pooling - Lab Means and their StDev for Certification BCR-385R G1




Figure C3. Laboratory means, mean of means and their standard deviation for aflatoxin G₁

Table C4. Results of characterisation measurements for aflatoxin G₂ of BCR385R

G2 mass fraction in BCR-385R [$\mu\text{g}/\text{kg}$]								
Lab code	Day 1		Day 2		Day 3		Mean	s
2	0.42	0.40	0.40	0.41	0.31	0.37	0.39	0.04
3	0.34	0.33	0.30	0.29	0.30	0.30	0.31	0.02
7	0.29	0.28	0.23	0.22	0.28	0.22	0.25	0.03
9	0.25	0.26	0.21	0.25	0.32	0.31	0.27	0.04

No Pooling - Lab Means and their StDev for Certification BCR-385R G2




Figure C4. Laboratory means, mean of means and their standard deviation for aflatoxin G₂

Table C5 Recovery rates for aflatoxin B₁

B1 recovery rate in spiked BCR-401R [%]			
Lab code	Day 1	Day 2	Day 3
1	88	96	70
2	100	113	> 120
3	74	98	83
4	77	79	80
5	111	116	102
7	110	96	95
9	< 50	114	108

Table C6 Recovery rates for aflatoxin B₂

B2 recovery rate in spiked BCR-401R [%]			
Lab code	Day 1	Day 2	Day 3
1	73	88	73
3	88	105	87
4	106	107	109
5	88	98	104
7	109	94	103
9	102	93	92

Table C7 Recovery rates for aflatoxin G₁

G1 recovery rate in spiked BCR-401R [%]			
Lab code	Day 1	Day 2	Day 3
1	67	95	88
2	83	95	103
3	61	87	68
5	82	66	109
7	96	92	91
9	103	94	102

Table C8 Recovery rates for aflatoxin G₂

G2 recovery rate in spiked BCR-401R [%]			
Lab code	Day 1	Day 2	Day 3
2	78	99	107
3	78	95	70
7	102	99	107
9	110	93	50

EUR 23522 EN – Joint Research Centre – Institute for Reference Materials and MeasurementsTitle: Certification of mass fractions in B₁, B₂, G₁, G₂ in peanut butter, BCR-385R and BCR-401R

Author(s): G. Buttinger, S. Harbeck, R. Josephs

Luxembourg: Office for Official Publications of the European Communities

2008 – 43 pp. – 21.0 x 29.7 cm

EUR – Scientific and Technical Research series – ISSN 1018-5593

ISBN 978-92-7909976-2

DOI 10.2787/77700

Abstract

This report describes the preparation of peanut butter (BCR-385R and BCR-401R) matrix reference materials and the certification of their content (mass fraction) of aflatoxins B₁, B₂, G₁ and G₂.

The preparation of the materials, the homogeneity and stability studies and the characterisation are described hereafter and the results are discussed. Uncertainties were calculated in compliance with the Guide to the Expression of Uncertainty in Measurement (GUM) [1], including uncertainties due to possible heterogeneity and instability. The certified values are listed below:

BCR-385R	Certified value ¹⁾	Uncertainty ²⁾	Number of accepted sets of results
Aflatoxin B ₁	1.77 µg/kg	0.30 µg/kg	7
Aflatoxin B ₂	0.48 µg/kg	0.08 µg/kg	6
Aflatoxin G ₁	0.9 µg/kg	0.4 µg/kg	6
Aflatoxin G ₂	0.30 µg/kg	0.12 µg/kg	4
Sum of Aflatoxin B ₁ , B ₂ , G ₁ , G ₂	3.5 µg/kg	0.5 µg/kg ³⁾	

- 1) These values are the mass fractions based on the unweighted mean of accepted results.
- 2) The uncertainties are the expanded uncertainties ($k = 2$) of the values defined in 1).
- 3) The uncertainty for the sum of Aflatoxins B₁, B₂, G₁, G₂ is calculated from the individual absolute uncertainties as $U_{sum} = 2 \cdot \sqrt{u_{B_1}^2 + u_{B_2}^2 + u_{G_1}^2 + u_{G_2}^2}$.

BCR-401R	Certified value	Number of accepted sets of results
Aflatoxin B ₁	< 0.2 µg/kg ¹⁾	7
Aflatoxin B ₂	< 0.2 µg/kg ²⁾	6
Aflatoxin G ₁	< 0.2 µg/kg ²⁾	6
Aflatoxin G ₂	< 0.2 µg/kg ²⁾	4

- 1) This value is the mass fractions based on the limits of quantification of the methods used and the highest found level of accepted results, with a 80% probability the measurand content is below this level.
- 2) These values are the mass fractions based on the limits of quantification of the methods used and the highest found level of accepted results, with a 95% probability the measurand content is below this level.

How to obtain EU publications

Our priced publications are available from EU Bookshop (<http://bookshop.europa.eu>), where you can place an order with the sales agent of your choice.

The Publications Office has a worldwide network of sales agents. You can obtain their contact details by sending a fax to (352) 29 29-42758.

The mission of the JRC is to provide customer-driven scientific and technical support for the conception, development, implementation and monitoring of EU policies. As a service of the European Commission, the JRC functions as a reference centre of science and technology for the Union. Close to the policy-making process, it serves the common interest of the Member States, while being independent of special interests, whether private or national.

