



Certification Report of EQRAIN Plutonium Program N° 10

A. Verbruggen, R. Eykens, F. Kehoe, H. Kühn, S. Richter, Y. Aregbe



The mission of IRMM is to promote a common and reliable European measurement system in support of EU policies.

European Commission

Directorate-General Joint Research Centre
Institute for Reference Materials and Measurements

Contact information

André Verbruggen
European Commission
Directorate-General Joint Research Centre
Institute for Reference Materials and Measurements
Retieseweg 111
B-2440 Geel • Belgium

Email: andre.verbruggen@ec.europa.eu

Tel.: +32 (0)14 571 617

Fax: +32 (0)14 571 863

<http://www.irmm.jrc.be>

<http://www.jrc.ec.europa.eu>

Legal Notice

Neither the European Commission nor any person acting on behalf of the Commission is responsible for the use which might be made of the following information.

A great deal of additional information on the European Union is available on the Internet.

It can be accessed through the Europa server

<http://europa.eu.int>

EUR Report 22787 EN

ISBN 978-92-79-06026-7

ISSN 1018-5593

Luxembourg: Office for Official Publications of the European Communities

© European Communities, 2007

Reproduction is authorised provided the source is acknowledged

Printed in Belgium



EUROPEAN COMMISSION
DIRECTORATE-GENERAL
Joint Research Centre



Certification Report of EQRAIN Plutonium Program N° 10

**A. Verbruggen, R. Eykens, F. Kehoe, H. Kühn,
S. Richter, Y. Aregbe**

Contents

Contents.....	4
Introduction	5
Isotopic amount ratios	5
Measurement of isotopic amount concentrations.....	5
Conclusions.....	8
References	8

Introduction

The EQRAIN Plutonium N°10 (Evaluation de la Qualité des Résultats d'Analyse dans l'Industrie Nucléaire) is an evaluation program set up by CETAMA (la Commission d'Établissement de Méthodes d'Analyses) in order to evaluate the quality of results from nuclear measurement laboratories.

The objective of this programme is to determine the plutonium concentration of a plutonium nitrate solution. Three ampoules were sent to each participant.

At the IM/nuclear group, it was decided to analyse both of the EQRAIN N°10 samples, more specifically sample Pu H53 and Pu H91. The results of this exercise would possibly enable us to identify problems occurring during the handling and measurements by mass-spectrometric Isotope dilution analysis (IDMS) of Pu.

The evaluation of the mass-spectrometry analysis as part of this program could indicate unexplained processes such as the behaviour of the fractionation factor and outlier results not identified during measurements.

Isotopic amount ratios

As the material is originally a MP2 metal, the isotope abundance ratios from the IRMM measurement certificate were used. Recertification of the isotope amount ratios of MP2 was accomplished recently and a certificate was issued valid for 1 January 2007. The values were recalculated for decay to the date of measurement 13 March 2007.

Table 1: Isotopic ratios and abundances

Isotope amount ratio(s)	
$n(^{238}\text{Pu})/n(^{239}\text{Pu})$	0.000 030 78(29)
$n(^{240}\text{Pu})/n(^{239}\text{Pu})$	0.022 432 1(51)
$n(^{241}\text{Pu})/n(^{239}\text{Pu})$	0.000 235 4(31)
$n(^{242}\text{Pu})/n(^{239}\text{Pu})$	0.000 075 70(78)

amount fraction (·100)		mass fraction (·100)	
$n(^{238}\text{Pu})/n(\text{Pu})$	0.003 010(28)	$m(^{238}\text{Pu})/m(\text{Pu})$	0.002 997(28)
$n(^{239}\text{Pu})/n(\text{Pu})$	97.773 31(58)	$m(^{239}\text{Pu})/m(\text{Pu})$	97.764 06(58)
$n(^{240}\text{Pu})/n(\text{Pu})$	2.193 26(49)	$m(^{240}\text{Pu})/m(\text{Pu})$	2.202 24(49)
$n(^{241}\text{Pu})/n(\text{Pu})$	0.023 02(30)	$m(^{241}\text{Pu})/m(\text{Pu})$	0.023 21(30)
$n(^{242}\text{Pu})/n(\text{Pu})$	0.007 402(76)	$m(^{242}\text{Pu})/m(\text{Pu})$	0.007 49(77)

Measurement of isotopic amount concentrations

The isotope concentrations of the Eqrain N°10 sample were measured by IDMS. From both samples, 3 blends were prepared consisting of 0.15 - 0.20 - 0.25 g sample spiked with 5 g IRMM-049c. Prior to mass-spectrometric measurement the samples were chemically conditioned and purified by anion exchange with nitric acid as medium.

The $n(^{239}\text{Pu})/n(^{242}\text{Pu})$ ratios of the blends were measured on the Finnigan MAT262 thermal ionisation mass spectrometer. A turret was loaded with the 6 blends and 5 standards (IRMM-290 A3, E and G3) and stored with following filename: 27116.

The isotopic measurements were carried out in static Faraday multi-collector mode using the total evaporation technique as described in [1, 2] and in working instruction WI-0115 of the IRMM IM-Unit quality system.

The fractionation factor derived from the measurements of the series of IRMM-290 samples was applied in the IDMS GUM workbenches [3, 4] as follows:

$$K_{\text{applied}} = K_{\text{meas}} \cdot K_{\text{base}}$$

where $K_{\text{base}} = 1$ with a relative standard uncertainty = 0.015%

K_{base} is included to eliminate inadvertently arriving at a K factor with negligible uncertainty: always a risk with a small series of measurements.

The overview of measurement results for the samples is shown in Table 2. The K factor, calculated by the ratio of the observed values and certified values for $n(^{242}\text{Pu})/n(^{239}\text{Pu})$ for this part of the campaign is shown in Table 3. Mean values per magazine were calculated as well as the overall value of $K_{\text{meas}} = 0.999\,71 \pm 0.000\,23$ (SE for $n=6$).

For the calculation of the isotopic concentrations of the samples of Eqrain N°10, the values of Table 1, 2 and 3 are the input values of the GUM workbench. In the calculation using GUM workbench, a term δ_i is added to each individual isotope dilution calculation. This term expresses a possible effect in the measurement which contributes to the uncertainty – if present. Each δ has a value of zero and a standard uncertainty that can be adjusted if necessary. Each δ has the same uncertainty, $u(\delta)$.

The value of $u(\delta)$ is assigned after inspection of the absolute difference between individual isotope dilution results and their mean value as given in each case by:

$$\epsilon_i = |C_{\text{Pu}i} - C_{\text{Pu mean}}|$$

In this measurement campaign the uncertainty contribution from each ϵ_i to the final uncertainty of C_{Pu} was 1.7%. This is negligible contribution to the overall uncertainty and shows that the assumption that an extra effect is present in the IDMS measurements is not warranted.

Table 4 shows the result of the IRMM measurements.

Table 5 shows the components of uncertainty for samples H53 and H91 of Eqrain N°10.

All data used can be found in S:\Im UNIT\Secure Data\Project Data\EQRAIN10

All values for plutonium isotope ratios and concentrations are corrected for decay to 13 March 2007.

Table 2: Measurements of blends

	ID	filename	$n(^{242}\text{Pu})/n(^{239}\text{Pu})$
1	FIL 03 H53 015	27308.CSV	0.696 644
2	FIL 05 H53 020	2730806.CSV	0.533 245
3	FIL 07 H53 030	2730808.CSV	0.425 158
4	FIL 09 H91 015	2730810.CSV	0.684 964
5	FIL 11 H91 020	2730812.CSV	0.484 927
6	FIL 13 H91 025	2730814.CSV	0.427 092

Table 3: Calculated fraction factor for $n(^{239}\text{Pu})/n(^{242}\text{Pu})$ ratio

Date	ID	Reference Mat	$K[n(^{242}\text{Pu})/n(^{239}\text{Pu})]$
13/05/04	FIL 04 E3 TE	IRMM-290 E3	0.999 94
13/05/04	FIL 06 G3 TE	IRMM-290 G3	0.999 65
13/05/04	FIL 08 A3 TE	IRMM-290 A3	1.000 04
13/05/04	FIL 10 E3 TE	IRMM-290 E3	0.998 86
13/05/04	FIL 12 G3 TE	IRMM-290 G3	1.000 09
		average	0.999 71
		SE	0.000 23

Table 4: Measured Isotope concentration of Eqrain N°10 samples (Uc $k=2$)

Solution Reference	Pu mass content ($\text{g}\cdot\text{kg}^{-1}$) measured	Pu mass content ($\text{g}\cdot\text{kg}^{-1}$) certified
H53	4.161 2(70)	4.162(4)
H91	4.256 9(64)	4.261(4)

Table 5: Components of uncertainty for IDMS of sample H53

Quantity	Value	Standard Uncertainty	Index
Δt_{MP2}	0.20260 a	$1.56 \cdot 10^{-3}$ a	0.0 %
$R_{238/239}$	$30.833 \cdot 10^{-6}$ mol/mol	$146 \cdot 10^{-9}$ mol/mol	0.0 %
$R_{240/239}$	0.02243240 mol/mol	$2.54 \cdot 10^{-6}$ mol/mol	0.0 %
$R_{241/239}$	$237.77 \cdot 10^{-6}$ mol/mol	$1.55 \cdot 10^{-6}$ mol/mol	0.0 %
$R_{242/239}$	$75.704 \cdot 10^{-6}$ mol/mol	$389 \cdot 10^{-9}$ mol/mol	0.0 %
T_{238}	87.700 a	0.300 a	0.0 %
T_{239}	24110.0 a	30.0 a	0.0 %
T_{240}	6563.00 a	7.00 a	0.0 %
T_{241}	14.2900 a	0.0600 a	0.0 %
T_{242}	$373.50 \cdot 10^3$ a	1100 a	0.0 %
$C_{y,IRMM049c}$	$366.650 \cdot 10^{-9}$ g/mol	$245 \cdot 10^{-12}$ g/mol	62.6 %
$R_{239/242Pu,IRMM049c}$	$2.2120 \cdot 10^{-3}$ mol/mol	$17.0 \cdot 10^{-6}$ mol/mol	0.0 %
Δt_{spike}	10.8695 a	0.0111 a	0.0 %
R_{b1}	1.435450	$144 \cdot 10^{-6}$	0.2 %
R_{b2}	1.875310	$188 \cdot 10^{-6}$	0.2 %
R_{b3}	2.352070	$235 \cdot 10^{-6}$	0.2 %
m_{x1}	0.154320 g	$28.9 \cdot 10^{-6}$ g	0.5 %
m_{x2}	0.201620 g	$28.9 \cdot 10^{-6}$ g	0.3 %
m_{x3}	0.254390 g	$28.9 \cdot 10^{-6}$ g	0.2 %
m_{y1}	4.996700 g	$289 \cdot 10^{-6}$ g	0.0 %
m_{y2}	4.999000 g	$289 \cdot 10^{-6}$ g	0.0 %
m_{y3}	5.016400 g	$289 \cdot 10^{-6}$ g	0.0 %
K_{meas}	1.000285	$228 \cdot 10^{-6}$	7.3 %
K_{base}	1.000000	$150 \cdot 10^{-6}$	3.2 %
δ_1	0.0	$10.0 \cdot 10^{-9}$	8.4 %
δ_2	0.0	$10.0 \cdot 10^{-9}$	8.4 %
δ_3	0.0	$10.0 \cdot 10^{-9}$	8.4 %
γ_{Pu}	$4.16119 \cdot 10^{-3}$ g/g	$3.51 \cdot 10^{-6}$ g/g	

Table 6: Components of uncertainty for IDMS of sample H91

Quantity	Value	Standard Uncertainty	Index
Δt_{MP2}	0.20260 a	$1.56 \cdot 10^{-3}$ a	0.0 %
$R_{238/239}$	$30.833 \cdot 10^{-6}$ mol/mol	$146 \cdot 10^{-9}$ mol/mol	0.0 %
$R_{240/239}$	0.02243240 mol/mol	$2.54 \cdot 10^{-6}$ mol/mol	0.0 %
$R_{241/239}$	$237.77 \cdot 10^{-6}$ mol/mol	$1.55 \cdot 10^{-6}$ mol/mol	0.0 %
$R_{242/239}$	$75.704 \cdot 10^{-6}$ mol/mol	$389 \cdot 10^{-9}$ mol/mol	0.0 %
T_{238}	87.700 a	0.300 a	0.0 %
T_{239}	24110.0 a	30.0 a	0.0 %

Quantity	Value	Standard Uncertainty	Index
T ₂₄₀	6563.00 a	7.00 a	0.0 %
T ₂₄₁	14.2900 a	0.0600 a	0.0 %
T ₂₄₂	373.50·10 ³ a	1100 a	0.0 %
C _{y.IRMM049c}	366.650·10 ⁻⁹ g/mol	245·10 ⁻¹² g/mol	78.0 %
R _{239/242Pu.IRMM049c}	2.2120·10 ⁻³ mol/mol	17.0·10 ⁻⁶ mol/mol	0.0 %
Δt _{spike}	10.8695 a	0.0111 a	0.0 %
R _{b1}	1.459930	146·10 ⁻⁶	0.2 %
R _{b2}	2.062160	206·10 ⁻⁶	0.2 %
R _{b3}	2.341420	234·10 ⁻⁶	0.2 %
m _{x1}	0.153500 g	28.9·10 ⁻⁶ g	0.7 %
m _{x2}	0.217300 g	28.9·10 ⁻⁶ g	0.3 %
m _{x3}	0.245800 g	28.9·10 ⁻⁶ g	0.3 %
m _{y1}	4.995200 g	289·10 ⁻⁶ g	0.0 %
m _{y2}	5.008400 g	289·10 ⁻⁶ g	0.0 %
m _{y3}	4.988100 g	289·10 ⁻⁶ g	0.0 %
K _{meas}	1.000285	228·10 ⁻⁶	9.1 %
K _{base}	1.000000	150·10 ⁻⁶	3.9 %
δ ₁	0.0	6.00·10 ⁻⁹	2.3 %
δ ₂	0.0	6.00·10 ⁻⁹	2.3 %
δ ₃	0.0	6.00·10 ⁻⁹	2.3 %
	γ _{Pu}	4.25693·10 ⁻³ g/g	3.22·10 ⁻⁶ g/g

Conclusions

The measurement of the samples provided through the Eqrain N°10 program has proven to be very useful. A number of conclusions could be drawn:

Analysis of the components of uncertainty on the ratio of measured values of IDMS on the MAT262 gave very satisfactory result. The major contributor to the total uncertainty is the uncertainty of the spike isotopic reference material used (62.6% and 78.0%); the influence of the fractionation is ca 10-15%. The variation within the IDMS sub-samples can contribute up to 25 % of the total uncertainty.

The results of the independent measurements of the samples show that we can have high confidence in our capacity to provide reliable results. We have found acceptable agreement with the certified value demonstrating the reliability of our procedures, methodology and competence of the people.

References

- [1] R. Fiedler, Int. J. of Mass Spectr., 146,147 (1995) 91-97.
- [2] E. L. Callis and R. M. Abernathy, Int. J. of Mass Spectr., 103, (1991) 93-105.
- [3] International Organisation for Standardisation, Guide to the Expression of Uncertainty in Measurements, ©ISO, ISBN 92-67-10188-9, Geneva, Switzerland 1993
- [4] GUM Workbench, Metrodata GmbH, www.metrodata.de

European Commission

**EUR 22787 EN – DG Joint Research Centre, Institute for Reference Materials and Measurements –
Certification Report of EQRAIN Plutonium Program N° 10**

Authors: A. Verbruggen, R. Eykens, F. Kehoe, H. Kühn , S. Richter, Y. Aregbe

Luxembourg: Office for Official Publications of the European Communities

2007 – 9 pp. – 21.0 x 29.7 cm

EUR - Scientific and Technical Research series; ISSN 1018-5593

ISBN 978-92-79-06026-7

Abstract

The EQRAIN Plutonium N°10 (Evaluation de la Qualité des Résultats d'Analyse dans l'Industrie Nucléaire) is an evaluation program set up by CETAMA (la **C**ommission d'**E**tablishement de **M**ethodes d'**A**nalyses) in order to evaluate the quality of results from nuclear measurement laboratories.

The objective of this programme is to determine the plutonium concentration of a plutonium nitrate solution. Three ampoules were sent to each participant. At the IM/nuclear group, it was decided to analyse both of the EQRAIN N°10 samples, more specifically sample Pu H53 and Pu H91. The results of this exercise would possibly enable us to identify problems occurring during the handling and measurements by mass-spectrometric Isotope dilution analysis (IDMS) of Pu.

The evaluation of the mass-spectrometry analysis as part of this program could indicate unexplained processes such as the behaviour of the fractionation factor and outlier results not identified during measurements.



EUROPEAN COMMISSION
DIRECTORATE-GENERAL
Joint Research Centre

The mission of the Joint Research Centre is to provide customer-driven scientific and technical support for the conception, development, implementation and monitoring of European Union policies. As a service of the European Commission, the JRC functions as a reference centre of science and technology for the Community. Close to the policy-making process, it serves the common interest of the Member States, while being independent of special interests, whether private or national.

LA-NA-22787-EN-C

ISBN 978-92-79-06026-7



9 789279 060267



Publications Office
Publications.eu.int