

---

# JRC Scientific and Technical Reports

---

EC/WHO Harmonization Programme for Air Quality Measurements:

## **The evaluation of the Interlaboratory Comparison Exercise for SO<sub>2</sub>, O<sub>3</sub>, NO and NO<sub>2</sub> Langen 20<sup>th</sup>-25<sup>th</sup> September 2009**

**Claudio A. Belis, Friedrich Lagler, Maurizio Barbieri,  
Hans-Guido Mücke, Klaus Wirtz and Volker Stummer**

EUR 24376 EN - 2010

The mission of the JRC-IES is to provide scientific-technical support to the European Union's policies for the protection and sustainable development of the European and global environment.

European Commission  
Joint Research Centre  
Institute for Environment and Sustainability

**Contact information**

Address: via Fermi, 2749 T.P. 441, 21027 Ispra (VA), Italy  
E-mail: [claudio.belis@jrc.ec.europa.eu](mailto:claudio.belis@jrc.ec.europa.eu)  
Tel.: +39 0332 786644  
Fax: +39 0332 785236

<http://ies.jrc.ec.europa.eu/>  
<http://www.jrc.ec.europa.eu/>

**Legal Notice**

Neither the European Commission nor any person acting on behalf of the Commission is responsible for the use which might be made of this publication.

***Europe Direct is a service to help you find answers  
to your questions about the European Union***

**Freephone number (\*):**

**00 800 6 7 8 9 10 11**

(\*) Certain mobile telephone operators do not allow access to 00 800 numbers or these calls may be billed.

A great deal of additional information on the European Union is available on the Internet. It can be accessed through the Europa server <http://europa.eu/>

JRC 58578

EUR 24376 EN  
ISBN 978-92-79-15853-7  
ISSN 1018-5593  
doi:10.2788/94507

Luxembourg: Publications Office of the European Union

© European Union, 2010

Reproduction is authorised provided the source is acknowledged

*Printed in Italy*

In collaboration with:

Gerard G.; Maetz P.; Posset R.; Grozdanovski L.; Atanasov I.; Pavli P.; Gjerek M.; Parvanova M.; Panayotov N.; Molis J.; Gaizutis T.; Alebić-Juretić A.; Zubak V.; Soldatenko S.; Mikhina L.



WHO COLLABORATING CENTRE FOR AIR QUALITY  
MANAGEMENT AND AIR POLLUTION CONTROL

at the

FEDERAL ENVIRONMENTAL AGENCY



## Executive Summary

From the 20<sup>th</sup> to the 25<sup>th</sup> of September 2009 in Langen (DE), 4 National Reference Laboratories (NRL) of AQUILA network and 3 laboratories of the World Health Organisation (WHO) Euro-Region met for an interlaboratory comparison exercise (IE) to evaluate their proficiency in the analysis of inorganic gaseous pollutants covered by European Air Quality Directives (SO<sub>2</sub>, NO, NO<sub>2</sub> and O<sub>3</sub>).

Most of the laboratories participating in the IE used automated CEN reference methods, which are mandatory in the EU, while some laboratories of the WHO Euro-Region performed analysis using manual methods.

In this report proficiency evaluation was made at different degrees for each laboratory taking into account the differences in the methodologies and the completeness of the information provided by participants. For the laboratories who expressed their uncertainty, performance was evaluated using two criteria, providing information on their proficiency to the European Commission and supporting the national quality control systems.

In terms of criteria imposed by the European Commission (that are not mandatory for WHO laboratories), 71% of the results reported by National Reference Laboratories (AQUILA network) were good both in terms of measured values and reported uncertainties. Another 23% of the results had good measured values, but the reported uncertainties were either too high (19%) or too small (4%). There were neither questionable nor unacceptable values.

AQUILA laboratories presented good comparability among participants for NO<sub>2</sub>, O<sub>3</sub>, and SO<sub>2</sub>. The relative reproducibility limit for NO was above the objective deriving from the standard deviation for proficiency assessment.

For WHO laboratories using automated techniques, the results are satisfactory for SO<sub>2</sub>, NO<sub>2</sub> and NO measurement methods, while one laboratory needs further investigation of their O<sub>3</sub> measurements. The laboratory using manual methods presented results comparable to those of the automated methods for NO and O<sub>3</sub> but there were questionable results for NO<sub>2</sub> and SO<sub>2</sub> and unsatisfactory results for NO<sub>2</sub>.

## Contents

1. Introduction .....	1
2. Communication and time schedule .....	3
3. Participants .....	3
4. The Analytical methods.....	4
5. The preparation of test mixtures.....	4
6. The evaluation of laboratory's measurement proficiency .....	6
5.1 z' - score .....	6
5.2 E <sub>n</sub> - number.....	10
7. Discussion .....	15
8. Conclusions .....	17
9. References .....	18
Annex A. Assigned values .....	20
Annex B. The results of the IE .....	22
Reported values for NO.....	22
Reported values for NO <sub>2</sub> .....	24
Reported values for SO <sub>2</sub> .....	27
Reported values for O <sub>3</sub> .....	30
Annex C. The precision of standardized measurement methods.....	33

## List of tables

<b>Table 1:</b>	<b>The list of participating organisations.....</b>	<b>3</b>
<b>Table 2:</b>	<b>The sequence program of generated test gases.....</b>	<b>5</b>
<b>Table 3:</b>	<b>The standard deviation for proficiency assessment <math>\sigma_p</math> as a linear function of concentration .....</b>	<b>6</b>
<b>Table 4:</b>	<b>The general assessment of proficiency results. ....</b>	<b>16</b>
<b>Table 5:</b>	<b>The validation of assigned values (X).....</b>	<b>21</b>
<b>Table 6:</b>	<b>Reported values for NO test gas 1. ....</b>	<b>22</b>
<b>Table 7:</b>	<b>Reported values for NO test gas 2. ....</b>	<b>23</b>
<b>Table 8:</b>	<b>Reported values for NO test gas 3. ....</b>	<b>23</b>
<b>Table 9:</b>	<b>Reported values for NO<sub>2</sub> test gas 4.....</b>	<b>24</b>
<b>Table 10:</b>	<b>Reported values for NO<sub>2</sub> test gas 5.....</b>	<b>24</b>
<b>Table 11:</b>	<b>Reported values for NO<sub>2</sub> test gas 6.....</b>	<b>25</b>
<b>Table 12:</b>	<b>Reported values for NO<sub>2</sub> test gas 7.....</b>	<b>25</b>
<b>Table 13:</b>	<b>Reported values for NO<sub>2</sub> test gas 8.....</b>	<b>26</b>
<b>Table 14:</b>	<b>Reported values for SO<sub>2</sub> test gas 9.....</b>	<b>27</b>
<b>Table 15:</b>	<b>Reported values for SO<sub>2</sub> test gas 10.....</b>	<b>27</b>
<b>Table 16:</b>	<b>Reported values for SO<sub>2</sub> test gas 11.....</b>	<b>28</b>
<b>Table 17:</b>	<b>Reported values for SO<sub>2</sub> test gas 12.....</b>	<b>28</b>
<b>Table 18:</b>	<b>Reported values for SO<sub>2</sub> test gas 13.....</b>	<b>29</b>
<b>Table 19:</b>	<b>Reported values for O<sub>3</sub> test gas 14.....</b>	<b>30</b>
<b>Table 20:</b>	<b>Reported values for O<sub>3</sub> test gas 15.....</b>	<b>30</b>
<b>Table 21:</b>	<b>Reported values for O<sub>3</sub> test gas 16.....</b>	<b>31</b>
<b>Table 22:</b>	<b>Reported values for O<sub>3</sub> test gas 17.....</b>	<b>31</b>
<b>Table 23:</b>	<b>Reported values for O<sub>3</sub> test gas 18.....</b>	<b>32</b>
<b>Table 24:</b>	<b>The R and r of NO standard measurement method.....</b>	<b>34</b>
<b>Table 25:</b>	<b>The R and r of NO<sub>2</sub> standard measurement method*.....</b>	<b>35</b>
<b>Table 26:</b>	<b>The R and r of SO<sub>2</sub> standard measurement method.....</b>	<b>36</b>
<b>Table 27:</b>	<b>The R and r of O<sub>3</sub> standard measurement method.....</b>	<b>37</b>

## List of figures

Figure 1:	The z'-score evaluations of NO measurements.....	7
Figure 2:	The z'-score evaluations of NO <sub>2</sub> measurements.....	8
Figure 3:	The z'-score evaluations of SO <sub>2</sub> measurements.....	8
Figure 4:	The z'-score evaluations of O <sub>3</sub> measurements.....	9
Figure 5:	Bias of participant's NO measurement results.....	11
Figure 6:	Bias of participant's NO <sub>2</sub> measurement results.....	12
Figure 7:	Bias of participant's SO <sub>2</sub> measurement results.....	13
Figure 8:	Bias of participant's O <sub>3</sub> measurement results.....	14
Figure 9:	The decision diagram for general assessment of proficiency results.....	15
Figure 10:	Reported values for NO test gas 1.....	22
Figure 11:	Reported values for NO test gas 2.....	23
Figure 12:	Reported values for NO test gas 3.....	23
Figure 13:	Reported values for NO <sub>2</sub> test gas 4.....	24
Figure 14:	Reported values for NO <sub>2</sub> test gas 5.....	24
Figure 15:	Reported values for NO <sub>2</sub> test gas 6.....	25
Figure 16:	Reported values for NO <sub>2</sub> test gas 7.....	25
Figure 17:	Reported values for NO <sub>2</sub> test gas 8.....	26
Figure 18:	Reported values for SO <sub>2</sub> test gas 9.....	27
Figure 19:	Reported values for SO <sub>2</sub> test gas 10.....	27
Figure 20:	Reported values for SO <sub>2</sub> test gas 11.....	28
Figure 21:	Reported values for SO <sub>2</sub> test gas 12.....	28
Figure 22:	Reported values for SO <sub>2</sub> test gas 13.....	29
Figure 23:	Reported values for O <sub>3</sub> test gas 14.....	30
Figure 24:	Reported values for O <sub>3</sub> test gas 15.....	30
Figure 25:	Reported values for O <sub>3</sub> test gas 16.....	31
Figure 26:	Reported values for O <sub>3</sub> test gas 17.....	31
Figure 27:	Reported values for O <sub>3</sub> test gas 18.....	32
Figure 28:	The R and r of NO standard measurement method as a function of concentration.....	34
Figure 29:	The R and r of NO <sub>2</sub> standard measurement method as a function of concentration.....	35
Figure 30:	The R and r of SO <sub>2</sub> standard measurement method as a function of concentration.....	36
Figure 31:	The R and r of O <sub>3</sub> standard measurement method as a function of concentration.....	37

## Abbreviations:

AQUILA	Network of National Reference Laboratories for Air Quality
CO	Carbon monoxide
DQO	Data Quality Objective
ERLAP	European Reference Laboratory of Air Pollution
EC	European Commission
GPT	Gas phase titration
IE	Intercomparison Exercise – Interlaboratory comparison [14], [23]
IES	Institute for Environment and Sustainability
ISO	International Organization for Standardization
JRC	Joint Research Centre
NO	Nitrogen monoxide
NO <sub>2</sub>	Nitrogen dioxide
NO <sub>x</sub>	the oxides of nitrogen, the sum of NO and NO <sub>2</sub>
NRL	National Reference Laboratory
O <sub>3</sub>	Ozone
SO <sub>2</sub>	Sulphur dioxide
UBA	Umweltbundesamt (Germany)
WHO	World Health Organization Collaborating Centre for Air Quality
CC-EURO	Management and Air Pollution Control, Berlin

## Mathematical Symbols:

<i>symbol</i>	<i>explanation</i>
$E_n$	$E_n$ – number statistic (ISO 13528; [14])
$r$	repeatability limit (ISO 5725; [15])
$R$	reproducibility limit (ISO 5725; [15])
$\sigma_p$	the standard deviation for proficiency assessment (ISO 13528; [14])
$x^*$	robust average (Annex C ISO 13528; [14])
$s^*$	robust standard deviation (Annex C ISO 13528; [14])
$s_r$	repeatability standard deviation (ISO 5725; [15])
$s_R$	reproducibility standard deviation (ISO 5725; [15])
$U_X$	The expanded uncertainty of the assigned/reference value (ISO 13528; [14])
$U_{xi}$	The expanded uncertainty of the participant's value
$u_X$	The standard uncertainty of the assigned/reference value (ISO 13528; [14])
$X$	Assigned/reference value (ISO 13528; [14])
$x_i$	the average of three values reported by the participant $i$ (for particular parameter and concentration level) (ISO 5725; [15])
$x_{ij}$	$j$ -th reported value of participant $i$ (for particular parameter and concentration level) (ISO 5725; [15])
$z'$	$z'$ -score statistic (ISO 13528; [14])



## **1. Introduction**

Directive 2008/50/EC [1] on ambient air quality and cleaner air for Europe sets a framework for a harmonized air quality assessment in Europe. One important objective of the Directive is that the ambient air quality shall be assessed on the basis of common methods and criteria. It deals with the air pollutants sulphur dioxide (SO<sub>2</sub>), nitrogen dioxide (NO<sub>2</sub>) and monoxide (NO), particulate matter, lead, benzene, carbon monoxide (CO) and ozone (O<sub>3</sub>). Among others it specifies the reference methods for measurements and Data Quality Objectives (DQO) for the accuracy of measurements.

The European Commission (EC) has supported the development and publication of reference measurement methods [2], [3] and [4] as European standards. Appropriate calibration methods [5], [8] and [9] have been standardised by the International Organization for Standardization (ISO).

As foreseen in the Directive, the European Reference Laboratory of Air Pollution (ERLAP) of the Institute for Environment and Sustainability (IES) at the Joint Research Centre (JRC) organizes interlaboratory comparison exercises (IE) to assess and improve the status of comparability of measurements of National Reference Laboratories (NRL) of each Member State of the European Union.

The World Health Organization Collaborating Centre for Air Quality Management and Air Pollution Control, Berlin (WHO CC) is carrying out similar activities since 1994 [10] [11], but with a view to obtaining harmonized air quality data for health related studies. Their program integrates within the WHO EURO region, which includes public health institutes and other national institutes - especially from the Central Eastern Europe, Caucasus and countries from Central Asia.

Starting in 2004, it has been decided to bring together the efforts of both the JRC-ERLAP and WHO CC and to coordinate activities as far as possible, with a view to optimize resources and have better international harmonization. The following report deals with the IE that took place from the 20<sup>th</sup> to the 25<sup>th</sup> of September 2009 in the UBA Pilotstation in Langen (DE) in joint cooperation of EC/JRC/IES/ERLAP and WHO CC.

ERLAP has been organizing IEs since 1990 aiming at evaluating the comparability of measurements carried out by NRLs and promoting information exchange among the expert laboratories. Nowadays the main objective, in accordance with the Network of National Reference Laboratories for Air Quality (AQUILA), comprises a more systematic approach that offers alert mechanism for the purposes of the EC and is also useful to NRLs in quality assurance of their implemented quality systems. The methodology for the organisation of IEs was developed by ERLAP and is described in a position paper on the organization of interlaboratory comparison exercises for gaseous air pollutants [13].

This evaluation scheme was adopted in December 2008 and is applied to all IEs since then. It contains common criteria to alert the EC on possible performance failures which do not rely solely on the uncertainty claimed by participants. The evaluation scheme implements the z'-score method [14] with the uncertainty requirements for calibration gases stated in the European standards [2], [3] and [4], which are consistent with the DQOs of European Directives.

According to AQUILA's view, NRLs with an overall unsatisfactory performance in the z'-score evaluation (one unsatisfactory or two questionable results per parameter) need to repeat their participation in the following IE in order to demonstrate remediation measures [13]. In addition, considering that the evaluation scheme should be useful to participants for accreditation according to ISO 17025, they are requested to include their measurement's uncertainty [14]. Hence, participants

measurement results (measurement values and uncertainties) are compared to the assigned values applying the E<sub>n</sub> – number method [14].

Beside the proficiency of participating laboratories, the repeatability and reproducibility of standardized measurement methods [15], [16] and [17] are evaluated as well. These group evaluations are useful indicators of trends in measurement quality over different IEs.

## 2. Communication and time schedule

The IE was announced in May 2009 to the members of the AQUILA network and the WHO CC representative. A registration letter was sent by WHO to interested parties and the registration was closed with the list of 7 participating laboratories. The participants were required to bring their own measurement instruments, data acquisition equipment and travelling standards (to be used for calibrations or checks during the IE).

The participants were invited to arrive on Sunday 20<sup>th</sup> September for the installation of their equipment (Table 2). The calibration and generation of NO analysers was carried out on Monday 21<sup>st</sup>, the calibration and generation of NO<sub>2</sub> analysers was carried out on Tuesday 22<sup>nd</sup>, the calibration and generation of SO<sub>2</sub> analysers was carried out on Wednesday 23<sup>rd</sup> and the calibration and generation of O<sub>3</sub> analysers was carried out on Friday 24<sup>th</sup>. The test gases generation finished on Thursday 24<sup>th</sup> at 16:45.

## 3. Participants

All participants were organizations dealing with the routine ambient air monitoring or health related studies. The national representatives came from EU member states Belgium, Bulgaria, Lithuania, Slovenia, and from non EU members Croatia, Ukraine, and Macedonia.

Country	Name of Organization	IE code	Network	method
Germany	Federal Environment Agency	A		automatic
Croatia	Teaching Institute of Public Health	B	WHO	auto/man
Belgium	Institut Scientifique de Service Public (ISSeP) Cellule Interrégionale de l'Environnement (CELINE) Belgium Environmental Agency	C	AQUILA	automatic
Ukraine	State Institution „O.M. Marzeyev Institute for Hygiene and Medical Ecology, Academy of Medical Sciences of Ukraine”, The Laboratory for Ambient Air Hygiene and Risk Assessment	D	WHO	manual
Lithuania	Environmental Protection Agency	E	AQUILA	automatic
Bulgaria	Executive Environment Agency	F	AQUILA	automatic
Slovenia	The Environmental Agency of the Republic of Slovenia	G	AQUILA	automatic
Macedonia	Central environmental laboratory - Calibration Laboratory from Ministry of Environment and Physical Planning	H	WHO	automatic

Table 1: The list of participating organisations.

#### **4. The Analytical methods**

NRLs of EU member states are required to implement CEN standard methods while non EU member states of the WHO Euro Region may use either CEN methods or methods officially accepted in their countries.

In the present interlaboratory comparison were used the following CEN automatic methods:

NO	Chemiluminescence [3]
NO <sub>2</sub>	Chemiluminescence [3]
SO <sub>2</sub>	Ultraviolet fluorescence [2]
O <sub>3</sub>	Ultraviolet photometry [4]

In addition, the following manual methods were used:

Teaching Institute of Public Health (Croatia) - IE Code B:

NO <sub>2</sub>	Spectrophotometry, modified Griess-Saltzman [5]
-----------------	---

This laboratory provided NO<sub>2</sub> measurements using two different techniques: CEN standard method and the manual method described above. The complete evaluation was made only on the results obtained with the standard method.

Institute for Hygiene and Medical Ecology “Marzeyev” , Academy of Medical Sciences of Ukraine – Laboratory for Ambient Air Hygiene and Risk Assessment (Ukraine) - IE Code D.

NO	Photocolorimetric method, modified Griess-Saltzman [6]
NO <sub>2</sub>	Photocolorimetric method, modified Griess-Saltzman [6]
SO <sub>2</sub>	Photocolorimetric hydrogen peroxide absorption method [6]
O <sub>3</sub>	Photocolorimetric Potassium iodide method, [6]

This was the only laboratory providing data measured exclusively with manual methods.

#### **5. The preparation of test mixtures**

The UBA Pilotstation facility is described in [10]. During this IE, gas mixtures were prepared for SO<sub>2</sub>, O<sub>3</sub>, NO and NO<sub>2</sub> at concentration levels around European Air Quality limit values, critical levels and assessment thresholds.

The test mixtures were prepared by the dilution of gases from cylinders containing high concentration of NO, NO<sub>2</sub> or SO<sub>2</sub> using thermal mass flow controllers [9]. O<sub>3</sub> was added using an ozone generator.

The participants were required to report three half-hour-mean measurements for each concentration level in order to evaluate the repeatability of standardized measurement methods. Zero concentration

levels were generated for one hour and one half-hour-mean measurement was reported. The sequence program of generated test gases is given in Table 2:

day	start time	duration □	operation	NO (nmol/mol)	NO <sub>2</sub> (nmol/mol)	SO <sub>2</sub> (nmol/mol)	O <sub>3</sub> (nmol/mol)
20-Sep	15:00	3	installation				
21-Sep	8:00	1	calibration				
21-Sep	9:00	2.30	NO test gas 1	0			
21-Sep	11:45	1.30	NO test gas 2	200			
21-Sep	13:30	1.30	NO test gas 3	20			
22-Sep	8:45	1	NO <sub>2</sub> test gas 4		0		
22-Sep	10:00	1.30	NO <sub>2</sub> test gas 5		200		
22-Sep	11:45	1.30	NO <sub>2</sub> test gas 6		100		
22-Sep	13:30	1.30	NO <sub>2</sub> test gas 7		60		
22-Sep	15:15	1.30	NO <sub>2</sub> test gas 8		20		
23-Sep	8:45	1	SO <sub>2</sub> test gas 9			0	
23-Sep	10:00	1.30	SO <sub>2</sub> test gas 10			130	
23-Sep	11:45	1.30	SO <sub>2</sub> test gas 11			45	
23-Sep	13:30	1.30	SO <sub>2</sub> test gas 12			20	
23-Sep	15:15	1.30	SO <sub>2</sub> test gas 13			5	
24-Sep	8:45	1	O <sub>3</sub> test gas 14				0
24-Sep	10:00	1.30	O <sub>3</sub> test gas 15				300
24-Sep	11:45	1.30	O <sub>3</sub> test gas 16				100
24-Sep	13:30	1.30	O <sub>3</sub> test gas 17				60
24-Sep	15:15	1.30	O <sub>3</sub> test gas 18				20

**Table 2: The sequence program of generated test gases.**

## 6. The evaluation of laboratory's measurement proficiency

To evaluate the participants measurement proficiency was applied the methodology described in ISO 13528 [14]. It has been agreed among the AQUILA members to take the measurement results of UBA as the assigned/reference values for the whole IE [13]. The traceability of UBA's measurement results and the method applied to validate them are presented in Annex A.

All data reported by participating laboratories are presented in Annex B.

As it is described in the Position Paper [13], the proficiency of the participants was assessed by calculating two performance indicators. The first performance indicator (z'-score) tests if the difference between the participants measured value and the assigned/reference value remains within the limits of a common criterion, while the second performance indicator (E<sub>n</sub>-number) tests if the difference between the participants measured values and assigned/reference value remains within the limits of a criterion, that is calculated individually for each participant, taking into account the uncertainty of the participants measurement and the uncertainty of the assigned/reference value.

### 5.1 z'- score

The z'- score statistic is calculated according to ISO 13528 [14] as:

$$z' = \frac{x_i - X}{\sqrt{\sigma_p^2 + u_x^2}} = \frac{x_i - X}{\sqrt{(a \cdot X + b)^2 + u_x^2}} \quad (1)$$

where 'x<sub>i</sub>' is a participant's run average value, 'X' is the assigned/reference value, 'σ<sub>p</sub>' is the 'standard deviation for proficiency assessment' and 'u<sub>x</sub>' is the standard uncertainty of assigned value. For 'a' and 'b' see Table 3:.

In the European standards [2], [3] and [4] the uncertainties for calibration gases used in ongoing quality control are prescribed. In fact, it is stated that the maximum permitted expanded uncertainty for calibration gases is 5% and that 'zero gas' shall not give instrument reading higher than the detection limit. As one of the tasks of NRLs is to supply calibration gas mixtures, the 'standard deviation for proficiency assessment' (σ<sub>p</sub>) [14] is calculated in *fitness-for-purpose* manner from requirements given in European standards.

Over the whole measurement range σ<sub>p</sub> is calculated by linear interpolation between 2.5% at the calibration point (75% of calibration range) and the limit of detection at zero concentration level. On November 2008 AQUILA members agreed to set σ<sub>p</sub> equal to 1 ppb at zero concentration of SO<sub>2</sub>, O<sub>3</sub>, NO, NO<sub>2</sub>.

The limits of detection of studied measurement methods were evaluated from the data of previous IEs. The linear function parameters of σ<sub>p</sub> are given in Table 3:


Gas	σ <sub>p</sub> =a·c+b	
	a	b
		nmol/mol
SO <sub>2</sub>	0.022	1
O <sub>3</sub>	0.020	1
NO	0.024	1
NO <sub>2</sub>	0.020	1

**Table 3: The standard deviation for proficiency assessment σ<sub>p</sub> as a linear function of concentration (c) with linear function parameters: slope (a) and intercept (b).**


The z'-score evaluation allows the following criteria to be used for the assessment of results:

- $|z'| \leq 2$  are designated satisfactory.
- $2 < |z'| \leq 3$  are designated questionable.
- $|z'| > 3$  are designated unsatisfactory. Scores falling in this range are very unusual and are taken to indicate that the cause of the event should be investigated and remedied.


The results of z'-score evaluation are presented in bar plots (Figure 1: to Figure 5:) in which the z'-scores of each participant are grouped together, and assessment criteria are presented as  $z'=\pm 2$  and  $z'=\pm 3$  lines.



**Figure 1:**The z'-score evaluations of NO measurements are given for each participant and each tested test gas. The evaluations according to the Test Gas step with nominal concentration are: 1 (0 nmol/mol), 2 (200 nmol/mol), 3 (20 nmol/mol), The assessment criteria are presented as  $z'=\pm 2$  and  $z'=\pm 3$  lines. They represent the limits for the questionable and unsatisfactory results.




**Figure 2: The z'-score evaluations of NO<sub>2</sub> measurements** are given for each participant and each tested gas. The evaluations according to the Test Gas step with nominal concentration are: 4 (0 nmol/mol), 5 (200 nmol/mol), 6 (100 nmol/mol), 7 (60 nmol/mol), 8 (20 nmol/mol). The assessment criteria are presented as  $z'=\pm 2$  and  $z'=\pm 3$  lines. They represent the limits for the questionable and unsatisfactory results.



**Figure 3: The z'-score evaluations of SO<sub>2</sub> measurements** are given for each participant and each tested concentration level. The evaluations according to the Test Gas step with nominal concentration are: 9 (0 nmol/mol), 10 (130 nmol/mol), 11 (45 nmol/mol), 12 (20 nmol/mol), 13 (5 nmol/mol). The assessment criteria are presented as  $z'=\pm 2$  and  $z'=\pm 3$  lines. They represent the limits for the questionable and unsatisfactory results.





**Figure 4:**The z'-score evaluations of O<sub>3</sub> measurements are given for each participant and each tested test gas. The evaluations are in the order of increasing concentrations (run number order (with nominal concentration) is: 14 (0 nmol/mol), 15 (300 nmol/mol), 16 (100 nmol/mol), 17 (60 nmol/mol), 18 (20 nmol/mol). The assessment criteria are presented as z'= $\pm$ 2 and z'= $\pm$ 3 lines. They represent the limits for the questionable and unsatisfactory results.


## 5.2 E<sub>n</sub>- number

The normalised deviations [14] (E<sub>n</sub>) were calculated according to:

$$E_n = \frac{x_i - X}{\sqrt{U_{x_i}^2 + U_X^2}} \quad (2)$$


where ‘X’ is the assigned/reference value with an expanded uncertainty ‘U<sub>X</sub>’ and ‘x<sub>i</sub>’ is the participant’s average value with an expanded uncertainty ‘U<sub>Xi</sub>’. Satisfactory results are the ones for which  $|E_n| \leq 1$ .

In Figure 5: to Figure 8: the bias of each participant ( $x_i - X$ ) are plotted and error bars are used to denote the value of denominator of equation 2 ( $\sqrt{U_{x_i}^2 + U_X^2}$ ). These plots represent also the E<sub>n</sub>-number evaluations where, considering the E<sub>n</sub> criteria ( $|E_n| \leq 1$ ), all results with error bars touching or crossing x-axis are satisfactory. Reported standard uncertainties (Annex B) that are bigger than “standard deviation for proficiency assessments” (σ<sub>p</sub>, Table 3:) are considered not fit-for-purpose and are denoted with “\*” in the x-axis of each figure.




**Figure 5: Bias of participant's NO measurement results**

together with the expanded uncertainty of bias presented with error bar are given for each tested gas (Table 2:). Results with error bars touching or crossing the x-axis are satisfactory. For each evaluation the test gas step together with the participants rounded run average (nmol/mol) is given. The '\*' mark indicates reported standard uncertainties bigger than  $\sigma_p$ .




**Figure 6: Bias of participant's NO<sub>2</sub> measurement results**  
 together with the expanded uncertainty of bias presented with error bar are given for each tested gas (see Table 2:). Results with error bars touching or crossing the x-axis are satisfactory. For each evaluation the test gas step together with the participants rounded run average (nmol/mol) is given. The '\*' mark indicates reported standard uncertainties bigger than  $\sigma_p$ .



**Figure 7: Bias of participant's SO<sub>2</sub> measurement results**

together with the expanded uncertainty of bias presented with error bar are given for each tested gas (Table 2:). The results with error bars touching or crossing the x-axis are satisfactory. For each evaluation the test gas step together with the participants rounded run average (nmol/mol) is given. The "\*" mark indicates reported standard uncertainties bigger than  $\sigma_p$ .




**Figure 8: Bias of participant's O<sub>3</sub> measurement results**

together with the expanded uncertainty of bias presented with error bar are given for each tested gas (see Table 2:). Results with error bars touching or crossing the x-axis are satisfactory. For each evaluation the test gas step together with the participants rounded run average (nmol/mol) is given. The '\*' mark indicates reported standard uncertainties bigger than  $\sigma_p$ .

## 7. Discussion

For a general assessment of the quality of each result a decision diagram was developed (Figure 9:) that categorises results in seven categories (a1 to a7). The general comments for each category are:

- a1: measurement result is completely satisfactory
- a2: measurement result is satisfactory (z'-score satisfactory and En-number ok) but the reported uncertainty is too high
- a3: measured value is satisfactory (z'-score satisfactory) but the reported uncertainty is underestimated (En-number not ok)
- a4: measurement result is questionable (z'-score questionable) but due to a high reported uncertainty can be considered valid (En-number ok)
- a5: measurement result is questionable (z'-score questionable and En-number not ok)
- a6: measurement result is unsatisfactory (z'-score unsatisfactory) but due to a high reported uncertainty can be considered valid (En-number ok)
- a7: measurement result is unsatisfactory (z'-score unsatisfactory and En-number not ok)



**Figure 9:**The decision diagram for general assessment of proficiency results.

The results of the IE were assigned to categories according to the diagram given in Figure 9: and are presented in Table 4:.

	run number	conc. level	IE code						
			B	C	D	E	F	G	H
NO (nmol/mol)	1	0.40	NU	a1	NV	NV	a3	NU	a1
	2	202.47	NU	a1	a2	NV	a2	a2	a2
	3	20.43	NU	a1	a2	NV	a1	a1	a1
NO <sub>2</sub> (nmol/mol)	4	0.00	NU	a3	NU	a1	a1	NU	a2
	5	203.40	NU	a1	a4	a2	a2	a2	a2
	6	101.72	NU	a1	a2	a1	a1	a2	a2
	7	60.46	NU	a1	a4	a1	a1	a2	a2
	8	19.67	NU	a1	a6	a1	a1	a1	a1
SO <sub>2</sub> (nmol/mol)	9	0.00	NV	a3	NV	a1	a1	NU	a1
	10	128.27	NV	a1	a2	a1	a1	a2	a1
	11	44.00	NV	a1	a4	a1	a1	a2	a1
	12	19.57	NV	a1	a4	a1	a1	a1	a1
	13	4.80	NV	a1	a2	a1	a1	a1	a1
O <sub>3</sub> (nmol/mol)	14	0.40	NV	a1	NV	a1	a1	NU	a1
	15	297.80	NV	a1	a2	a1	a2	a2	a5
	16	99.60	NV	a1	a2	a1	a1	a2	a2
	17	59.80	NV	a1	a2	a1	a1	a2	a1
	18	20.03	NV	a1	a2	a1	a1	a1	a2

**Table 4: The general assessment of proficiency results.**  
 “NV”: no values reported; “NU”: no uncertainty values reported



## 8. Conclusions

The proficiency evaluation scheme has provided an assessment of the participants measured values and their evaluated uncertainties. Some of the measured values were provided without an uncertainty estimation (11%), in these cases it was only possible to estimate  $z'$  scores. In terms of the criteria imposed by the European Commission ( $\sigma_p$ ), 71% of the results reported by NRLs fall into 'a1' category and are good both in terms of measured values and evaluated uncertainties. Most of the remnant measured values are good but the evaluated uncertainties are either too high (19%), category 'a2', or too small, category 'a3' (4%).

The relative number of 'a2' cases, where participant's evaluated uncertainty is higher than the common IE criterion, was lower than in previous IEs but still relevant. The common IE criterion is confirmed to be realistic by comparison to the reproducibility standard deviation obtained at this (Annex C) and other IEs [19]. The mentioned criterion is derived from the European standards' uncertainty requirements, which are explicit at high concentrations. Since the uncertainty requirements at zero concentration are not quantitatively stated in the European standards, the IE criteria at zero concentration were set by decision of AQUILA members in November 2008. The slight improvement in the compliance of participant's uncertainty at low concentrations with respect to previous IEs is likely ascribable to the application of these new criteria.

No questionable or unsatisfactory results were observed among NRLs.

Concerning WHO laboratories using automated techniques, the results are satisfactory for SO<sub>2</sub>, NO<sub>2</sub> and NO measurement methods. The laboratory using manual methods presents results comparable to those of the automated methods for NO and O<sub>3</sub> but their uncertainty is too high.

Considering all WHO laboratories as a whole, there are 12% questionable results, categories 'a4' and 'a5', and one unsatisfactory result ('a6'). Laboratory D, which uses only manual methods, presents two "a4" and one "a6" in its NO<sub>2</sub> measurements while two SO<sub>2</sub> measurements fall in the "a4" category. Laboratory H has one "a5" result in the highest O<sub>3</sub> level. In the first case, the differences are clearly attributable to the use of methods which are not equivalent to CEN reference methods [2], [3]. Laboratory H should investigate the causes of the discrepancy but no action is needed at this stage.

The better comparability among AQUILA-NRL's results is the one observed in SO<sub>2</sub> measurements while NO measurements present the worst performance in terms of reproducibility. The relative reproducibility limits derive from criteria imposed by the European Commission ( $\sigma_p$ ). Levels are below the objective at the highest studied concentration: 8.6% for NO<sub>2</sub>, 5.8% for O<sub>3</sub>, and 4.7% for SO<sub>2</sub>. Comparability is considered unsatisfactory for NO where the relative reproducibility limit is 14.2% while the objective is 13.0%. On the other hand, the evaluation of the comparability for this compound has to be interpreted with caution due to the reduced number of levels (only three) and labs (only four). The repeatability and reproducibility of NO<sub>2</sub> depend on the concentration of both NO<sub>2</sub> and NO. Therefore, contemporary acquisition of NO and NO<sub>2</sub> would be necessary to evaluate the combined effect of these two gases.

## 9. References

- [1] Directive 2008/50/EC of the European Parliament and of the Council of 21 May 2008 on ambient air quality and cleaner air for Europe, L 152, 11.06.2008
- [2] EN 14212:2005, Ambient air quality - Standard method for the measurement of the concentration of sulphur dioxide by ultraviolet fluorescence
- [3] EN 14211:2005, Ambient air quality - Standard method for the measurement of the concentration of nitrogen dioxide and nitrogen monoxide by chemiluminescence
- [4] EN 14625:2005, Ambient air quality - Standard method for the measurement of the concentration of ozone by ultraviolet photometry
- [5] Perry R. & R.J. Young Eds: Handbook of Air Pollution Analysis, Chapman& Hall, London, 1977, p. 268-270.
- [6] Ukrainian State Sanitary Rules for the Protection of Ambient Air of the Settlements (against chemical and biological contaminations): Order of the Ministry of Public Health, July 9, 1997 (ДСП № 201 ДСП – 201 – 97)
- [7] ISO 6143:2001, Gas analysis - Comparison methods for determining and checking the composition of calibration gas mixtures
- [8] ISO 6144:2003, Gas analysis - Preparation of calibration gas mixtures - Static volumetric method
- [9] ISO 6145-7:2001, Gas analysis - Preparation of calibration gas mixtures using dynamic volumetric methods - Part 7: Thermal mass-flow controllers
- [10] Mücke H.-G., (2008), Air quality management in the WHO European Region – Results of a quality assurance and control programme on air quality monitoring (1994-2004), Environment International, EI-01718
- [11] Mücke H.-G., et al. (2000), European Intercomparison workshop on air quality monitoring vol.4 – Measuring NO, NO<sub>2</sub>, O<sub>3</sub> and SO<sub>2</sub> – Air Hygiene Report 13, WHO Collaboration Centre for Air Quality Management and Air Pollution Control, ISSN 0938 - 9822
- [12] <http://ies.jrc.ec.europa.eu/aquila-homepage.html>
- [13] AQUILA POSITION PAPER N. 37. (2008) Protocol for intercomparison exercise. Organisation of intercomparison exercises for gaseous air pollution for EU national air quality reference laboratories and laboratories of the WHO EURO region [http://ies.jrc.ec.europa.eu/uploads/fileadmin/H04/Air\\_Quality/N%2037%20final%20version%20IE%20organisation%20and%20evaluation.pdf](http://ies.jrc.ec.europa.eu/uploads/fileadmin/H04/Air_Quality/N%2037%20final%20version%20IE%20organisation%20and%20evaluation.pdf)
- [14] ISO 13528:2005, Statistical methods for use in proficiency testing by interlaboratory comparisons

- [15] ISO 5725-1:1994, Accuracy (trueness and precision) of measurement methods and results – Part 1: General principles and definitions
- [16] ISO 5725-2:1994, Accuracy (trueness and precision) of measurement methods and results – Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method
- [17] ISO 5725-6:1994, Accuracy (trueness and precision) of measurement methods and results - Part 6: Use in practice of accuracy values
- [18] Mücke H-G, et al. (1996). European Intercomparison Workshops on Air Quality Monitoring. Vol. 2 – Measuring of CO, NO, NO<sub>2</sub> and O<sub>3</sub> – Air Hygiene Report 9. Berlin, Germany: WHO Collaborating Centre for Air Quality Management and Air Pollution Control; ISSN 0938-9822.
- [19] The evaluation of the Intercomparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> carried out in October 2007 in Essen.
- [20] GUM Workbench, The Tool for Expression of Uncertainty of Measurements
- [21] [http://www.qmbalance.com/pages/level1/procontrol\\_features\\_en.html](http://www.qmbalance.com/pages/level1/procontrol_features_en.html)
- [22] ISO Guide 43-1:1997, Proficiency testing by interlaboratory comparisons - Part 1: Development and operation of proficiency testing schemes.
- [23] Viallon, J., et al. (2006), International comparison CCQM-P28: Ozone at ambient level, Metrologia, 43, Tech. Suppl., 08010, doi:10.1088/0026-1394/43/1A/08010

## ***Annex A. Assigned values***

The assigned values of tested concentration levels were derived from UBA measurements which are calibrated against the certified reference values of CRMs and are traceable to international standards. In this perspective the assigned values are reference values as defined in the ISO 13528 [14].

UBA SO<sub>2</sub>, NO<sub>2</sub> and NO analysers were calibrated using primary calibration gas mixtures prepared according to the methodology described in the ISO 6144 [5]. The procedure and the device for generating primary calibration gases is described elsewhere [18]. Gas mixtures for the calibration experiment were produced from the reference mixtures by dynamic dilution method using mass flow controllers [9].

SO<sub>2</sub>, NO<sub>2</sub> and NO gas mixtures manufactured by Air Liquide and certified by UBA ( $U \leq 2\%$ ) were used as internal standards.

For the reference gas mixture composition evaluation and for the calibration experiment evaluation two computer applications were used, the “GUM WORKBENCH” [20] and “ProControl®” [22].

For O<sub>3</sub> measurements, the primary standard NIST photometer SRP 29 was used [24].

UBA’s measurement results were validated by comparison to the group statistics ( $x^*$  and  $s^*$ ) for every parameter and concentration level of the IE. These statistics are calculated from participants, applying the robust method described in the Annex C of the ISO 13528 [14]. The validation is taking into account UBA’s measurement result ( $X$ ) and its standard uncertainty ( $u_X$ ) as given in expression 3 [14]:

$$\frac{|x^* - X|}{\sqrt{\frac{(1,25 \cdot s^*)^2}{p} + u_X^2}} < 2 \quad (3)$$

Where ‘ $x^*$ ’ and ‘ $s^*$ ’ represent robust average and robust standard deviation respectively and ‘ $p$ ’ is the number of participants.

In Table 5: all inputs for expression 6 are given and all UBA’s measurement results are confirmed to be valid.

run	unit	X	uX'	$\bar{x}^*$	$s^*$	valid.
NO_1	nmol/mol	0.40	0.10	0.25	0.15	OK
NO_2	nmol/mol	202.47	2.10	205.06	3.89	OK
NO_3	nmol/mol	20.43	0.31	20.59	0.22	OK
NO2_4	nmol/mol	0.00	0.10	0.15	0.24	OK
NO2_5	nmol/mol	203.40	2.11	200.72	5.55	OK
NO2_6	nmol/mol	101.72	1.07	99.94	0.86	OK
NO2_7	nmol/mol	60.46	0.67	59.14	1.43	OK
NO2_8	nmol/mol	19.67	0.31	19.12	0.56	OK
SO2_9	nmol/mol	0.00	0.10	-0.02	0.39	OK
SO2_10	nmol/mol	128.27	1.51	128.41	2.21	OK
SO2_11	nmol/mol	44.00	0.83	44.12	0.88	OK
SO2_12	nmol/mol	19.57	0.71	19.59	0.18	OK
SO2_13	nmol/mol	4.80	0.68	4.90	0.09	OK
O3_14	nmol/mol	0.40	0.30	0.30	0.20	OK
O3_15	nmol/mol	297.80	1.54	297.57	5.12	OK
O3_16	nmol/mol	99.60	0.66	99.31	2.24	OK
O3_17	nmol/mol	59.80	0.52	59.80	1.52	OK
O3_18	nmol/mol	20.03	0.43	19.79	0.49	OK

**Table 5: The validation of assigned values (X) by comparison to the robust averages ( $\bar{x}^*$ ) with taking into the account the standard uncertainties of assigned values (uX'), and robust standard deviations ( $s^*$ ) as denoted by expression 6.**

Due to the reduced length of the bench (only 8 m), compared to those used in other IEs, the lack of homogeneity between the beginning and the end is assumed to be negligible. The evaluation of the uncertainty of the bench confirmed that differences along the distribution line are below the limit of detection.

## Annex B. The results of the IE

The reported values, presented also in graphs, are given in this annex. The participants were asked to report results ( $x_{ij}$ ,  $u(x_i)$  and  $U(x_i)$ ) expressed in mol/mol units. For all the runs except concentration levels 0, also average ( $\bar{x}_i$ ) and standard deviation ( $s_i$ ) of each participant are presented. As a group evaluation robust average ( $x^*$ ) and robust standard deviation ( $s^*$ ) were calculated (applying the procedure described in Annex C of ISO 13528) for each run, and are presented in the following tables. The assigned value is indicated on the graphs with the red line and the individual laboratories expanded uncertainties ( $U(x_i)$ ) are indicated with error bars.

### Reported values for NO

parameter: NO				all units are nmol/mol			
test gas 1				$x^*: 0.3$		$s^*: 0.2$	
	A	B	C	D	F	G	H
$x_{i,1}$	0.40	0.32	0.30		0.15	0.00	-0.01
$u(x_i)$	0.10		0.02		0.01		0.76
$U(x_i)$	0.20		0.05		0.02		1.51

Table 6: Reported values for NO test gas 1.




Figure 10: Reported values for NO test gas 1.

parameter: NO				all units are nmol/mol			
test gas 2				205.1		s*: 3.9	
	A	B	C	D	F	G	H
xi,1	202.50	205.54	205.20	192.00	192.00	207.10	210.67
xi,2	202.40	205.07	205.02	208.00	193.00	206.70	210.19
xi,3	202.50	205.58	205.03	176.00	193.00	206.90	210.07
xi	202.47	205.40	205.08	192.00	192.67	206.90	210.31
si	0.06	0.28	0.10	16.00	0.58	0.20	0.32
u(xi)	2.01		4.46	16.17	5.95	7.80	6.35
U(xi)	4.03		8.92	51.44	11.90	15.50	12.69

Table 7: Reported values for NO test gas 2.




Figure 11: Reported values for NO test gas 2.

parameter: NO				all units are nmol/mol			
test gas 3				x*: 20.6		s*: 0.2	
	A	B	C	D	F	G	H
xi,1	20.50	21.24	20.64	20.80	19.51	20.60	20.56
xi,2	20.40	21.21	20.82	16.00	19.52	20.70	20.52
xi,3	20.40	21.23	20.58	22.40	19.60	20.60	20.69
xi	20.43	21.23	20.68	19.73	19.54	20.63	20.59
si	0.06	0.02	0.13	3.33	0.05	0.06	0.09
u(xi)	0.30		0.50	3.49	0.60	0.80	0.85
U(xi)	0.61		1.00	11.09	1.20	1.50	1.69

Table 8: Reported values for NO test gas 3.




Figure 12: Reported values for NO test gas 3.

## Reported values for NO<sub>2</sub>

parameter: NO <sub>2</sub>				all units are nmol/mol				
test gas 4				x*: 0.1		s*: 0.2		
	A	B	C	D	E	F	G	H
xi,1	0.00	0.39	0.44	7.33	-0.07	0.15	0.20	-0.09
u(xi)	0.10		0.02		0.22	0.01		1.00
U(xi)	0.20		0.05		0.45	0.02		2.00

Table 9: Reported values for NO<sub>2</sub> test gas 4.




Figure 13: Reported values for NO<sub>2</sub> test gas 4.

parameter: NO <sub>2</sub>				all units are nmol/mol				
test gas 5				x*: 200.7		s*: 5.6		
	A	B	C	D	E	F	G	H
xi,1	202.16	194.21	200.58	200.87	198.20	194.00	204.00	206.77
xi,2	203.77	195.31	200.85	191.67	198.49	195.00	206.00	207.11
xi,3	204.28	196.64	200.71	169.32	197.65	195.00	207.10	207.24
xi	203.40	195.39	200.71	187.29	198.11	194.67	205.70	207.04
si	1.11	1.22	0.14	16.23	0.43	0.58	1.57	0.24
u(xi)	2.01		4.46	16.38	5.25	6.21	7.70	7.13
U(xi)	4.03		8.92	52.00	10.50	12.42	15.40	14.25

Table 10: Reported values for NO<sub>2</sub> test gas 5.




Figure 14: Reported values for NO<sub>2</sub> test gas 5.



parameter: NO <sub>2</sub>					all units are nmol/mol			
test gas 6					x*: 99.9		s*: 0.9	
	A	B	C	D	E	F	G	H
xi,1	101.89	100.03	99.89	95.25	98.85	97.00	103.10	99.95
xi,2	101.68	100.55	99.88	86.96	99.54	97.50	102.90	99.47
xi,3	101.58	100.32	99.78	105.90	99.85	97.50	102.60	99.00
xi	101.72	100.30	99.85	96.04	99.41	97.33	102.87	99.47
si	0.16	0.26	0.06	9.49	0.51	0.29	0.25	0.48
u(xi)	1.03		2.24	8.28	2.63	2.95	3.90	3.94
U(xi)	2.05		4.48	26.35	5.26	5.90	7.70	7.87

Table 11: Reported values for NO<sub>2</sub> test gas 6.




Figure 15: Reported values for NO<sub>2</sub> test gas 6.

parameter: NO <sub>2</sub>					all units are nmol/mol			
test gas 7					x*: 59.1		s*: 1.4	
	A	B	C	D	E	F	G	H
xi,1	60.53	59.72	59.25	74.08	59.30	57.50	60.90	57.67
xi,2	60.22	59.40	59.37	63.50	59.02	57.60	60.60	57.43
xi,3	60.63	59.47	58.89	63.50	59.00	57.60	60.60	57.31
xi	60.46	59.53	59.17	67.03	59.11	57.57	60.70	57.47
si	0.21	0.17	0.25	6.11	0.17	0.06	0.17	0.18
u(xi)	0.64		1.39	6.25	1.56	1.71	2.30	2.58
U(xi)	1.29		2.79	19.85	3.13	3.42	4.60	5.15

Table 12: Reported values for NO<sub>2</sub> test gas 7.




Figure 16: Reported values for NO<sub>2</sub> test gas 7.

parameter: NO <sub>2</sub>					all units are nmol/mol			
test gas 8					x*: 19.1		s*: 0.6	
	A	B	C	D	E	F	G	H
<b>xi,1</b>	19.67	18.86	19.43	21.16	19.47	19.05	19.40	17.94
<b>xi,2</b>	19.47	18.67	19.13	26.45	19.83	19.10	19.30	17.94
<b>xi,3</b>	19.87	18.62	19.02	31.30	19.60	19.10	19.10	17.86
<b>xi</b>	19.67	18.72	19.19	26.30	19.63	19.08	19.27	17.91
<b>si</b>	0.20	0.13	0.21	5.07	0.18	0.03	0.15	0.05
<b>u(xi)</b>	0.30		0.53	5.24	0.52	0.57	0.70	0.92
<b>U(xi)</b>	0.61		1.06	16.66	1.04	1.14	1.40	1.83

Table 13: Reported values for NO<sub>2</sub> test gas 8.




Figure 17: Reported values for NO<sub>2</sub> test gas 8.

## Reported values for SO<sub>2</sub>

parameter: SO <sub>2</sub>				all units are nmol/mol			
test gas 9				x*: 0.0		s*: 0.4	
	A	C	D	E	F	G	H
xi,1	0.00	0.48		-0.10	0.15	-0.50	-0.07
u(xi)	0.10	0.01		0.25	0.01		0.69
U(xi)	0.20	0.02		0.50	0.02		1.37

Table 14: Reported values for SO<sub>2</sub> test gas 9.




Figure 18: Reported values for SO<sub>2</sub> test gas 9.

parameter: SO <sub>2</sub>				all units are nmol/mol			
test gas 10				x*: 128.4		s*: 2.2	
	A	C	D	E	F	G	H
xi,1	128.20	126.06	121.25	126.85	127.20	129.80	132.30
xi,2	128.20	126.14	131.67	126.98	127.80	130.00	132.17
xi,3	128.40	125.99	142.21	127.17	128.20	130.50	132.33
xi	128.27	126.06	131.71	127.00	127.73	130.10	132.27
si	0.12	0.08	10.48	0.16	0.50	0.36	0.09
u(xi)	1.47	2.03	10.65	2.54	3.72	6.00	2.96
U(xi)	2.93	4.05	33.87	5.08	7.42	11.90	5.91

Table 15: Reported values for SO<sub>2</sub> test gas 10.




Figure 19: Reported values for SO<sub>2</sub> test gas 10.

parameter: SO <sub>2</sub>				all units are nmol/mol			
test gas 11				x*: 44.1		s*: 0.9	
	A	C	D	E	F	G	H
xi,1	44.20	43.59	47.25	44.95	42.90	44.60	45.50
xi,2	43.90	43.42	47.25	44.76	43.20	44.10	45.02
xi,3	43.90	43.28	52.67	44.60	43.30	43.90	44.99
xi	44.00	43.43	49.06	44.77	43.13	44.20	45.17
si	0.17	0.16	3.13	0.18	0.21	0.36	0.29
u(xi)	0.82	0.72	3.40	0.89	1.30	2.00	0.85
U(xi)	1.63	1.44	10.84	1.79	2.60	4.10	1.69

Table 16: Reported values for SO<sub>2</sub> test gas 11.




Figure 20: Reported values for SO<sub>2</sub> test gas 11.

parameter: SO <sub>2</sub>				all units are nmol/mol			
test gas 12				x*: 19.6		s*: 0.2	
	A	C	D	E	F	G	H
xi,1	19.60	19.40	21.06	19.69	19.50	19.70	20.07
xi,2	19.60	19.38	26.33	19.72	19.40	19.50	20.09
xi,3	19.50	19.42	21.06	19.73	19.50	19.50	20.00
xi	19.57	19.40	22.82	19.71	19.47	19.57	20.05
si	0.06	0.02	3.04	0.02	0.06	0.12	0.05
u(xi)	0.71	0.32	3.18	0.39	0.62	0.90	1.10
U(xi)	1.42	0.64	10.13	0.78	1.24	1.80	2.19

Table 17: Reported values for SO<sub>2</sub> test gas 12.




Figure 21: Reported values for SO<sub>2</sub> test gas 12.

parameter: SO <sub>2</sub>				all units are nmol/mol			
test gas 13				x*: 4.9		s*: 0.1	
	A	C	D	E	F	G	H
xi,1	4.80	4.99	5.27	5.02	4.80	5.10	4.87
xi,2	4.80	4.94	5.27	4.99	4.83	4.80	4.91
xi,3	4.80	4.91	7.35	4.96	4.83	5.00	4.89
xi	4.80	4.95	5.96	4.99	4.82	4.97	4.89
si	0.00	0.04	1.20	0.03	0.02	0.15	0.02
u(xi)	0.68	0.09	1.34	0.09	0.14	0.20	0.94
U(xi)	1.36	0.18	4.26	0.19	0.28	0.50	1.87

Table 18: Reported values for SO<sub>2</sub> test gas 13.




Figure 22: Reported values for SO<sub>2</sub> test gas 13.

### Reported values for O<sub>3</sub>

parameter: O <sub>3</sub>				all units are nmol/mol			
test gas 14				x*: 0.3		s*: 0.2	
	A	C	D	E	F	G	H
xi,1	0.40	0.30		0.55	0.20	0.00	-0.22
u(xi)	0.30	0.37		0.20	0.01		1.00
U(xi)	0.60	0.74		0.40	0.02		2.00

Table 19: Reported values for O<sub>3</sub> test gas 14




Figure 23: Reported values for O<sub>3</sub> test gas 14.

parameter: O <sub>3</sub>				all units are nmol/mol			
test gas 15				x*: 297.6		s*: 5.1	
	A	C	D	E	F	G	H
xi,1	296.60	292.96	282.50	293.43	291.80	301.70	310.76
xi,2	298.10	294.43	308.50	294.85	292.80	303.60	313.22
xi,3	298.70	295.29	291.67	296.44	293.00	304.70	314.87
xi	297.80	294.23	294.22	294.91	292.53	303.33	312.95
si	1.08	1.18	13.19	1.51	0.64	1.52	2.07
u(xi)	1.26	3.33	13.35	4.97	9.05	18.20	4.40
U(xi)	2.51	6.65	42.48	9.58	18.10	36.40	8.79

Table 20: Reported values for O<sub>3</sub> test gas 15.




Figure 24: Reported values for O<sub>3</sub> test gas 15.

parameter: O <sub>3</sub>				all units are nmol/mol			
test gas 16				x*: 99.3		s*: 2.2	
	A	C	D	E	F	G	H
xi,1	100.20	98.59	95.83	98.49	96.60	101.80	104.51
xi,2	99.40	97.97	108.33	98.07	97.00	101.10	103.92
xi,3	99.20	97.78	105.00	97.81	97.00	100.90	103.82
xi	99.60	98.11	103.05	98.12	96.87	101.27	104.08
si	0.53	0.42	6.47	0.34	0.23	0.47	0.37
u(xi)	0.58	1.27	6.63	1.59	2.95	6.10	3.49
U(xi)	1.17	2.54	21.10	3.18	5.90	12.20	6.98

Table 21: Reported values for O<sub>3</sub> test gas 16.




Figure 25: Reported values for O<sub>3</sub> test gas 16.

parameter: O <sub>3</sub>				all units are nmol/mol			
test gas 17				x*: 59.8		s*: 1.5	
	A	C	D	E	F	G	H
xi,1	60.00	59.04	58.33	59.48	58.20	61.00	62.31
xi,2	59.70	58.91	63.30	59.32	58.00	60.90	62.05
xi,3	59.70	58.89	56.67	59.22	58.00	60.80	61.99
xi	59.80	58.95	59.43	59.34	58.07	60.90	62.12
si	0.17	0.08	3.45	0.13	0.12	0.10	0.17
u(xi)	0.49	0.85	3.61	0.96	1.75	3.70	1.99
U(xi)	0.98	1.70	11.48	1.92	3.50	7.30	3.97

Table 22: Reported values for O<sub>3</sub> test gas 17.




Figure 26: Reported values for O<sub>3</sub> test gas 17.

parameter: O <sub>3</sub>				all units are nmol/mol			
test gas 18				x*: 19.8		s*: 0.5	
	A	C	D	E	F	G	H
xi,1	20.00	19.46	19.17	19.69	19.10	20.20	20.23
xi,2	20.00	19.41	22.50	19.74	19.20	20.20	20.18
xi,3	20.10	19.43	20.83	19.73	19.20	20.20	20.19
xi	20.03	19.43	20.83	19.72	19.17	20.20	20.20
si	0.06	0.03	1.67	0.03	0.06	0.00	0.03
u(xi)	0.43	0.46	1.82	0.64	0.59	1.20	1.59
U(xi)	0.86	0.93	5.81	1.28	1.18	2.40	3.17

Table 23: Reported values for O<sub>3</sub> test gas 18.




Figure 27: Reported values for O<sub>3</sub> test gas 18.



### ***Annex C. The precision of standardized measurement methods***

For the main purpose of monitoring trends between different IEs undertaken in the framework of the EC/WHO Harmonization Programme for Air Quality Measurements, the precision of standardized SO<sub>2</sub>, O<sub>3</sub> and NO<sub>x</sub> measurement methods [2], [3] and [4] as implemented by NRLs was evaluated. Applied methodology is described in ISO 5725-1, -2 and -6 [15], [16] and [17]. The precision experiment has involved four laboratories, for NO<sub>x</sub>, SO<sub>2</sub> and O<sub>3</sub> measurement methods. Five concentration levels were tested, for O<sub>3</sub>, SO<sub>2</sub> and NO<sub>2</sub>, and three for NO. The data consistency and outlier tests were uncertain due to the small number of participants. No outliers were detected.

The repeatability standard deviation ( $s_r$ ) was calculated in accordance with ISO 5725-2 as the square root of average within laboratory variance. The repeatability limit ( $r$ ) is calculated using equation 4 [17]. It represents the biggest difference between two test results found on an identical test gas by one laboratory using the same apparatus within the shortest feasible time interval, that should not be exceeded on average more than once in 20 cases in the normal and correct operation of method.

$$r = t_{95\%,8} \cdot \sqrt{2} \cdot s_r \quad (4)$$

The reproducibility standard deviation ( $s_R$ ) was calculated in accordance with ISO 5725-2 as the square root of sum of repeatability and between laboratory variance. The reproducibility limit ( $R$ ) is calculated using equation 5 [17]. It represents the biggest difference between two measurements on an identical test gas reported by two laboratories, which should not occur on average more than once in 20 cases in the normal and correct operation of method.

$$R = t_{95\%,3} \cdot \sqrt{2} \cdot s_R \quad (5)$$

The repeatability standard deviation was evaluated with 8 (4·(3-1)) degrees of freedom ( $\nu$ ) and reproducibility standard deviation with 3 (4-1) degrees of freedom. The critical range student factors ( $t_{\alpha,\nu}$ ) are 2,31 and 3,18 respectively.

In this annex are presented the repeatability and reproducibility limits of measurement methods ( $r$ ,  $R$ ), compared to the reproducibility from common criteria ( $R(\text{from } \sigma_p)$ ) calculated by substituting  $s_R$  in equation 5 with the 'standard deviation for proficiency assessment' (Table 3:). Comparison between  $R$  and  $R(\text{from } \sigma_p)$  serves to indicate that  $\sigma_p$  is realistic ([14] 6.3.1) and that the general methodology implemented by NRLs fulfil the criteria set in the standard for limiting uncertainty ( $\sigma_p$ ).

NO data (nmol/mol)			
AQUILA LABS			
group average	repeatability limit : r	reproducibility limit : R	reproducibility limit (relative)
0.2		0.6	14.2%
20.3	0.3	2.4	
201.8	1.0	28.6	

Table 24: The R and r of NO standard measurement method.




Figure 28: The R and r of NO standard measurement method as a function of concentration.

NO <sub>2</sub> data (nmol/mol)			
AQUILA LABS			
group average	repeatability limit : r	reproducibility limit : R	reproducibility limit (relative)
0.1		1.0	
19.4	0.5	1.2	
59.4	0.6	5.0	
100.2	0.9	8.5	
200.5	2.9	17.3	8.6%

Table 25: The R and r of NO<sub>2</sub> standard measurement method\*.




Figure 29: The R and r of NO<sub>2</sub> standard measurement method as a function of concentration\*.

\*No simultaneous NO, NO<sub>2</sub> data are available for these runs.

SO <sub>2</sub> data (nmol/mol)			
AQUILA LABS			
group average	repeatability limit : r	reproducibility limit : R	reproducibility limit (relative)
0.1		1.1	
4.9	0.2	0.4	
19.5	0.2	0.5	
43.9	0.7	2.6	
127.8	0.9	6.0	4.7%

Table 26: The R and r of SO<sub>2</sub> standard measurement method.




Figure 30: The R and r of SO<sub>2</sub> standard measurement method as a function of concentration.

O <sub>3</sub> data (nmol/mol)			
AQUILA LABS			
group average	repeatability limit : r	reproducibility limit : R	reproducibility limit (relative)
0.3		0.8	
19.7	0.1	1.7	
59.4	0.4	4.1	
98.8	1.3	6.8	
296.6	3.9	17.1	5.8%

Table 27: The R and r of O<sub>3</sub> standard measurement method.




Figure 31: The R and r of O<sub>3</sub> standard measurement method as a function of concentration.

European Commission

**EUR 24376 EN – Joint Research Centre – Institute for Environment and Sustainability**

Title: The evaluation of the Interlaboratory Comparison Exercise for SO<sub>2</sub>, O<sub>3</sub>, NO and NO<sub>2</sub> Langen 20th-25th September 2009

Authors: Claudio A. Belis, Friedrich Lagler, Maurizio Barbieri, Hans-Guido Mücke, Klaus Wirtz and Volker Stummer

Luxembourg: Publications Office of the European Union

2010 – 48 pp. – 29.7 x 21 cm

EUR – Scientific and Technical Research series – ISSN 1018-5593

ISBN 978-92-79-15853-7

doi:10.2788/94507

**Abstract**

From the 20<sup>th</sup> to the 25<sup>th</sup> of September 2009 in Langen (DE), 4 national reference laboratories (NRL) of AQUILA network and 3 laboratories of the World Health Organisation (WHO) Euro-Region met for an interlaboratory comparison exercise (IE) to evaluate their proficiency in the analysis of inorganic gaseous pollutants covered by European Air Quality Directives (SO<sub>2</sub>, NO, NO<sub>2</sub> and O<sub>3</sub>).

Most of the laboratories participating in the IE used automated CEN reference methods, which are mandatory in the EU, while some laboratories of the WHO Euro-Region performed analysis using manual methods.

In this report proficiency evaluation was made at different degrees for each laboratory taking into account the differences in the methodologies and the completeness of the information provided by participants. For the laboratories who expressed their uncertainty, performance was evaluated using two criteria, providing information on their proficiency to the European Commission and supporting the national quality control systems.

In terms of criteria imposed by the European Commission (that are not mandatory for WHO laboratories), 71% of the results reported by National Reference Laboratories (AQUILA network) were good both in terms of measured values and reported uncertainties. Another 23% of the results had good measured values, but the reported uncertainties were either too high (19%) or too small (4%). There were neither questionable nor unacceptable values.

AQUILA laboratories presented good comparability among participants for NO<sub>2</sub>, O<sub>3</sub>, and SO<sub>2</sub>. The relative reproducibility limit for NO was above the objective deriving from the standard deviation for proficiency assessment.

For WHO laboratories using automated techniques, the results are satisfactory for SO<sub>2</sub>, NO<sub>2</sub> and NO measurement methods, while one laboratory needs further investigation of their O<sub>3</sub> measurements.

The laboratory using manual methods presented results comparable to those of the automated methods for NO and O<sub>3</sub> but there were questionable results for NO<sub>2</sub> and SO<sub>2</sub> and unsatisfactory results for NO<sub>2</sub>.

### **How to obtain EU publications**

Our priced publications are available from EU Bookshop (<http://bookshop.europa.eu>), where you can place an order with the sales agent of your choice.

The Publications Office has a worldwide network of sales agents. You can obtain their contact details by sending a fax to (352) 29 29-42758.

The mission of the JRC is to provide customer-driven scientific and technical support for the conception, development, implementation and monitoring of EU policies. As a service of the European Commission, the JRC functions as a reference centre of science and technology for the Union. Close to the policy-making process, it serves the common interest of the Member States, while being independent of special interests, whether private or national.

