

# Evaluation of the Laboratory Comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> Langen 23<sup>rd</sup> - 28<sup>th</sup> October 2011

EC Harmonization Program for Air Quality Measurements



Maurizio Barbieri, Volker Stummer, Friedrich Lagler, Hans-Guido Mücke  
2012

Institute for Environment and Sustainability

Contact information

Friedrich Lagler

Address: Joint Research Centre, Via Enrico Fermi 2749, TP 442, 21027 Ispra (VA), Italy

E-mail: [friedrich.lagler@jrc.ec.europa.eu](mailto:friedrich.lagler@jrc.ec.europa.eu)

Tel.: +39 0332 789990

Fax: +39 0332 789931

<http://www.jrc.ec.europa.eu/>

This publication is a Reference Report by the Joint Research Centre of the European Commission.

Legal Notice

Neither the European Commission nor any person acting on behalf of the Commission is responsible for the use which might be made of this publication.

Europe Direct is a service to help you find answers to your questions about the European Union Free phone number (\*): 00 800 6 7 8 9 10 11

(\*) Certain mobile telephone operators do not allow access to 00 800 numbers or these calls may be billed.

A great deal of additional information on the European Union is available on the Internet. It can be accessed through the Europa server <http://europa.eu/>.

JRC72562

EUR 25387 EN

ISBN 978-92-79-25367-6

ISSN 1831-9424

doi:10.2788/33334

Luxembourg: Publications Office of the European Union, 2012

© European Union, 2012

Reproduction is authorised provided the source is acknowledged.

Printed in Italy

In collaboration with:

Stummer, V.; Schuster, A.; Meyer-Arndt, R.; Wirtz, K.; Beslic, I.; Davila, S.; Sega, K.; Grozdanovski, L.; Atanasov, I.; Molis, J.; Gaizutis, T.; Kislova, O.; Karev, A. ; Adasnik-Spasic, L.; Sekulic, Z.; Sostaric, A. ; Mykhina, L.; Petruk, L.



WHO  
Collaborating Centre for Air Quality Management  
and Air Pollution Control  
at the Federal Environment Agency



	<b>NAME</b>	<b>VERSION</b>	<b>DATE</b>
<b>AUTHOR</b>	M. BARBIERE	DRAFT 1	15/05/2012
<b>REVIEW</b>	F. LAGLER	DRAFT 2	16/05/2012
<b>REVIEW</b>	N. JENSEN	DRAFT 3	29/05/2012
<b>REVIEW</b>	H.-G. MÜCKE	DRAFT 4	13/06/2012
<b>REVIEW</b>	V. STUMMER	DRAFT 5	20/06/2012
<b>APPROVAL</b>	N. JENSEN	1.0	29/06/2012

## Executive Summary

From the 23<sup>rd</sup> to the 28<sup>th</sup> of October 2011 seven Laboratories of the World Health Organization (WHO) European-Region met for another joint JRC-ERLAP/WHO inter-laboratory comparison exercise (IE) at the National Air Quality Reference laboratory at the German Federal Environment Agency in Langen Germany to evaluate their proficiency in the analysis of inorganic gaseous pollutants (SO<sub>2</sub>, CO, NO, NO<sub>2</sub> and O<sub>3</sub>) covered by the European Air Quality Directive 2008/50/EC.

Most of the laboratories participating in the IE used automated instruments while one laboratory performed analysis using manual methods.

The proficiency evaluation, where each participant's bias was compared to two criteria, provides information on compliance with Data Quality Objectives and measurement capabilities of the National Air Quality Laboratories to the European Commission and can be used by participants in their laboratory's quality system.

In terms of criteria imposed by the European Commission (that are not mandatory for WHO laboratories), 59.4% of the results reported by National Reference Laboratories (AQUILA network) were good both in terms of measured values and reported uncertainties. Another 39.9% of the results had good measured values, but the reported uncertainties were either too high. Only one reported value (0.7%) has been evaluated as questionable.

The comparability of results among AQUILA participants at the highest generated concentration levels, excluding outliers, is acceptable for CO and NO measurements while SO<sub>2</sub>, O<sub>3</sub> and NO<sub>2</sub> measurements showed less satisfactory results.

## Contents

1. Introduction.....	9
1.1 Communication and time schedule .....	11
1.2 Participants .....	11
1.3 The preparation of test mixtures .....	14
2. The evaluation of laboratory's measurement proficiency.....	15
2.1 z' - score .....	15
2.2 E <sub>n</sub> - number .....	19
3. Discussion .....	25
4. Conclusions .....	27
5. References .....	29
Annex A. Assigned values .....	31
Annex B. The results of the IE .....	33
Reported values for SO <sub>2</sub> .....	33
Reported values for CO .....	36
Reported values for O <sub>3</sub> .....	39
Reported values for NO .....	42
Reported values for NO <sub>2</sub> .....	44
Annex C. The precision of standardized measurement methods .....	47
Annex D. The scrutiny of results for consistency and outlier test .....	53

## List of tables

Table 1: The list of participating institutions.	11
Table 2: The list of instruments used by participants.	12
Table 3: The sequence program of generated test gases	14
Table 4: The standard deviation for proficiency assessment ( $\sigma_p$ ).	15
Table 5: The general assessment of proficiency results. "nd" is referring to values not reported.	26
Table 6: history of the results in the last IE	27
Table 7: z'-score summary	28
Table 8: The validation of assigned values (X)	32
Table 9: Reported values for SO <sub>2</sub> run 0.	33
Table 10: Reported values for SO <sub>2</sub> run 1.	34
Table 11: Reported values for SO <sub>2</sub> run 2.	34
Table 12: Reported values for SO <sub>2</sub> run 3.	35
Table 13: Reported values for SO <sub>2</sub> run 4.	35
Table 14: Reported values for CO run 0.	36
Table 15: Reported values for CO run 1.	36
Table 16: Reported values for CO run 2.	37
Table 17: Reported values for CO run 3.	37
Table 18: Reported values for CO run 4.	38
Table 19: Reported values for CO run 5.	38
Table 20: Reported values for O <sub>3</sub> run 0.	39
Table 21: Reported values for O <sub>3</sub> run 1	39
Table 22: Reported values for O <sub>3</sub> run 2.	40
Table 23: Reported values for O <sub>3</sub> run 3.	40
Table 24: Reported values for O <sub>3</sub> run 4.	41
Table 25: Reported values for NO run 0.	42
Table 26: Reported values for NO run 1.	42
Table 27: Reported values for NO run 2.	43
Table 28: Reported values for NO <sub>2</sub> run 0.	44
Table 29: Reported values for NO <sub>2</sub> run 1.	44
Table 30: Reported values for NO <sub>2</sub> run 2.	45
Table 31: Reported values for NO <sub>2</sub> run 3.	45
Table 32: Reported values for NO <sub>2</sub> run 4.	46
Table 33: Critical values of t used in the repeatability (r) and reproducibility (R) evaluation.	47
Table 34: The R and r of SO <sub>2</sub> standard measurement method.	48
Table 35: The R and r of CO standard measurement method.	49
Table 36: The R and r of O <sub>3</sub> standard measurement method.	50
Table 37: The R and r of NO standard measurement method.	51
Table 38: The R and r of NO <sub>2</sub> standard measurement method.	52
Table 39: "Genuine" statistical outliers according to Grubb's one outlying observation test.	53

## List of figures

Figure 1: The z'-score evaluations of SO <sub>2</sub> measurements	16
Figure 2: The z'-score evaluations of CO measurements	16
Figure 3: The z'-score evaluations of O <sub>3</sub> measurements	17
Figure 4: The z'-score evaluations of NO measurements	17
Figure 5: The z'-score evaluations of NO <sub>2</sub> measurements	18
Figure 6: Bias of participant's SO <sub>2</sub> measurement results	20
Figure 7: Bias of participant's CO measurement results	21
Figure 8: Bias of participant's O <sub>3</sub> measurement results	22
Figure 9: Bias of participant's NO measurement results	23
Figure 10: Bias of participant's NO <sub>2</sub> measurement results	24
Figure 11: The decision diagram for general assessment of proficiency results.	25
Figure 12: Reported values for SO <sub>2</sub> run 0.	33
Figure 13: Reported values for SO <sub>2</sub> run 1.	34
Figure 14: Reported values for SO <sub>2</sub> run 2.	34
Figure 15: Reported values for SO <sub>2</sub> run 3.	35
Figure 16: Reported values for SO <sub>2</sub> run 4.	35
Figure 17: Reported values for CO run 0	36
Figure 18: Reported values for CO run 1.	36
Figure 19: Reported values for CO run 2.	37
Figure 20: Reported values for CO run 3.	37
Figure 21: Reported values for CO run 4.	38
Figure 22: Reported values for CO run 5.	38
Figure 23: Reported values for O <sub>3</sub> run 0.	39
Figure 24: Reported values for O <sub>3</sub> run 1.	39
Figure 25: Reported values for O <sub>3</sub> run 2.	40
Figure 26: Reported values for O <sub>3</sub> run 3.	40
Figure 27: Reported values for O <sub>3</sub> run 4.	41
Figure 28: Reported values for NO run 0.	42
Figure 29: Reported values for NO run 1.	42
Figure 30: Reported values for NO run 2.	43
Figure 31: Reported values for NO <sub>2</sub> run 0.	44
Figure 32: Reported values for NO <sub>2</sub> run 1.	44
Figure 33: Reported values for NO <sub>2</sub> run 2.	45
Figure 34: Reported values for NO <sub>2</sub> run 3.	46
Figure 35: Reported values for NO <sub>2</sub> run 4.	46
Figure 36: The R and r of SO <sub>2</sub> standard measurement method as a function of concentration.	48
Figure 37: The R and r of CO standard measurement method as a function of concentration.	49
Figure 38: The R and r of O <sub>3</sub> standard measurement method as a function of concentration.	50
Figure 39: The R and r of NO standard measurement method as a function of concentration.	51
Figure 40: The R and r of NO <sub>2</sub> standard measurement method as a function of concentration.	52

## Abbreviations

AQUILA	Network of National Reference Laboratories for Air Quality <a href="http://ies.jrc.ec.europa.eu/aquila-homepage.html">http://ies.jrc.ec.europa.eu/aquila-homepage.html</a>
CO	Carbon monoxide
DQO	Data Quality Objective
ERLAP	European Reference Laboratory of Air Pollution <a href="http://ies.jrc.ec.europa.eu/our-activities/support-for-member-states/erlap.html">http://ies.jrc.ec.europa.eu/our-activities/support-for-member-states/erlap.html</a>
EC	European Commission
GPT	Gas Phase Titration
IE	Inter-laboratory Comparison Exercise
IES	Institute for Environment and Sustainability
ISO	International Organization for Standardization
JRC	Joint Research Centre
NO	Nitrogen monoxide
NO <sub>2</sub>	Nitrogen dioxide
NO <sub>x</sub>	the oxides of nitrogen, the sum of NO and NO <sub>2</sub>
NRL	National Reference Laboratory
O <sub>3</sub>	Ozone
SO <sub>2</sub>	Sulphur dioxide
WHO-CC	World Health Organization Collaborating Centre for Air Quality Management and Air Pollution Control, Berlin

## Mathematical Symbols

*symbol explanation*

$\alpha$	converter efficiency (EN 14211; [4])
$E_n$	$E_n$ – number statistic (ISO 13528; [13])
$r$	repeatability limit (ISO 5725; [14])
$R$	reproducibility limit (ISO 5725; [14])
$\sigma_p$	standard deviation for proficiency assessment (ISO 13528; [13])
$x^*$	robust average (Annex C ISO 13528; [13])
$s^*$	robust standard deviation (Annex C ISO 13528; [13])
$s_r$	repeatability standard deviation (ISO 5725; [14])
$s_R$	reproducibility standard deviation (ISO 5725; [14])
$U_{X'}$	expanded uncertainty of the assigned/reference value (ISO 13528; [13])
$U_{x_i}$	expanded uncertainty of the participant's value
$u_{X'}$	standard uncertainty of the assigned/reference value (ISO 13528; [13])
$X$	assigned/reference value (ISO 13528; [13])
$x_i$	average of three values reported by the participant $i$ (for particular parameter and concentration level) (ISO 5725; [14])
$x_{i,j}$	$j$ -the reported value of participant $i$ (for particular parameter and concentration level) (ISO 5725; [14])
$z'$	$z'$ -score statistic (ISO 13528; [13])



## 1. Introduction

Directive 2008/50/EC [1] on ambient air quality and cleaner air for Europe sets a framework for a harmonized air quality assessment in Europe. One important objective of the Directive is that the ambient air quality shall be assessed on the basis of common methods and criteria. It deals with the air pollutants sulphur dioxide (SO<sub>2</sub>), nitrogen dioxide (NO<sub>2</sub>) and monoxide (NO), particulate matter, lead, benzene, carbon monoxide (CO) and ozone (O<sub>3</sub>). Among others it specifies the reference methods for measurements and Data Quality Objectives (DQO) for the accuracy of measurements.

The European Commission (EC) has supported the development and publication of reference measurement methods for CO [2], SO<sub>2</sub> [3], NO-NO<sub>2</sub> [4] and O<sub>3</sub> [5] as European standards. Appropriate calibration methods [6], [7] and [8] have been standardized by the International Organization for Standardization (ISO).

As foreseen in the Air Quality Directive, the European Reference Laboratory of Air Pollution (ERLAP) of the Institute for Environment and Sustainability (IES) at the Joint Research Centre (JRC) organizes inter-laboratory comparison exercises (IE) to assess and improve the status of comparability of measurements of National Reference Laboratories (NRL) of each Member State of the European Union.

The World Health Organization Collaborating Centre for Air Quality Management and Air Pollution Control, Berlin (WHO-CC) is carrying out similar activities since 1994 [9] [10] [31] [33], but with a view to obtaining harmonized air quality data for health related studies. Their program integrates within the WHO European Region, which includes public health and other environmental institutes - especially from countries of Central and Eastern Europe, Caucasus and Central Asia.

Starting in 2004, it has been decided to bring together the efforts of both the JRC-ERLAP and WHO-CC and to coordinate activities as far as possible, with a view to optimize resources and have better international harmonization.

The following report deals with the IE that took place from the 23<sup>rd</sup> to the 28<sup>th</sup> of October 2011 at the National Reference laboratory for Air Pollution, German Federal Environment Agency (UBA) in Langen, Germany in joint cooperation of EC/ JRC/IES/ERLAP and WHO-CC.

Since few decades in Europe IE are organized aiming at evaluating the comparability of measurements carried out by NRLs and promoting information exchange among the expert laboratories.

Currently, a more systematic approach has been adopted, in accordance with the Network of National Reference Laboratories for Air Quality (AQUILA) [11], aiming both at providing an alert mechanism for the purposes of the EC legislation and at supporting the implementation of quality schemes by NRLs. The methodology for the organization of IE was developed by ERLAP in collaboration with AQUILA and is described in a paper on the organization of laboratory comparison exercises for gaseous air pollutants [12].

This evaluation scheme was adopted in December 2008 and is applied to all IE since then. It contains common criteria to alert the EC on possible performance failures which do not rely solely on the uncertainty claimed by participants. The evaluation scheme implements the z'-score method [13] with the uncertainty requirements for calibration gases stated in the European standards [2], [3], [4] and [5], which are consistent with the DQOs of European Directives.

According to the said document, NRLs with an overall unsatisfactory performance in the z'-score evaluation (one unsatisfactory or two questionable results per parameter) ought to repeat their participation in the following IE in order to demonstrate remediation measures [12]. In addition, considering that the evaluation scheme should be useful to participants for accreditation according to ISO 17025, they are requested to include their measurement

uncertainty. Hence, participants' results (measurement values and uncertainties) are compared to the assigned values applying the E<sub>n</sub> - number method [13].  
Beside the proficiency of participating laboratories, the repeatability and reproducibility of standardized measurement methods [14], [15] and [16] are evaluated as well. These group evaluations are useful indicators of trends in measurement quality over different IE.

### 1.1 Communication and time schedule

The IE was announced in March 2011 to the members of the AQUILA network and the WHO-CC representative. Registration was opened on March 2011. A registration letter was sent by WHO-CC to interested parties and the registration was closed with the list of seven participating laboratories.

The participants were required to bring their own measurement instruments, data acquisition equipment and travelling standards (to be used for calibrations or checks during the IE).

The participants were invited to arrive on Sunday, 23<sup>rd</sup> October 2011, for the installation of their equipment. On Monday (24/10/2011) morning the generation of NO gas mixtures started at 9:00. On Tuesday morning at 8:45 the zero air analysis for NO<sub>2</sub> measurement started. SO<sub>2</sub> and CO measurement was carried out on Wednesday 8:45. O<sub>3</sub> was measured on Thursday from 8:45 am till 16:45 when the IE ended.

### 1.2 Participants

All participating laboratories belonged to institutions dealing with routine ambient air quality monitoring or to institutions involved in public health protection. The representatives came from following countries: Croatia, Macedonia, Lithuania, Russia, Serbia, Ukraine and Germany.

Country	Laboratory	Code	Network	Method
Croatia	Institute for Medical Research and Occupational Health (IMI)	B	WHO	automatic
Macedonia	Ministry of Environment and Physical Planning (MOEPP)	C	WHO	automatic
Lithuania	Environmental Protection Agency (AAA)	D	AQUILA	automatic
Russian Federation	State Environmental Institution 'Mosecommonitoring' (MOSECOM)	E	WHO	automatic
Serbia	Institute of Public Health (IPH_S)	F	AQUILA	automatic
Ukraine	State Institution 'O.M. Marzeev Institute of Hygiene and Medical Ecology, Academy of Medical Sciences of Ukraine' (IHME)	G	WHO	auto/manual
Germany	Federal Environment Agency (UBA)	H	AQUILA	automatic

**Table 1: The list of participating institutions.**

Table 2 reports the manufacturer and model of the instrumentation used by every participant during the inter-laboratory comparison exercise included those used in the calculation of the assigned values.

As a whole, the instrumentation belongs to five different manufacturers with the exception of SO<sub>2</sub> where four brands are present.

The list contains the information reported by participants and by no means can be considered as an implicit or explicit endorsement of the organizers to any specific type of instrumentation. All participants have used automatic analyzer beside Ukraine laboratory that used a semi-automatic method.

Gas	Lab Code	Instrument
CO	B	APMA-370, 2010
	C	Thermo Environment, TEI 48C
	D	Horiba Ltd., 2011, NDIR, APMA 370
	E	OPTEC, model K-100 (№58-1-04)
	F	HORIBA, 2008, APMA 370
	G	-
	H	HORIBA, 2009, APMA 370
	NOX	B
C		Thermo Environment, TEI 42C
D		Horiba Ltd, 2011, chemiluminescence, APNA 370,
E		Monitor Europe, ME-9841B (№ 09-1638)
F		Horiba, 2008, APNA370
G		-
H		HORIBA, 2004, APNA 360
O <sub>3</sub>		B
	C	Thermo Environment, TEI 49C.
	D	Horiba Ltd, 2011, NDUV, APOA 370,
	E	Monitor Europe, ME-9810B (№ M1692-M343)
	F	HORRIBA, 2008, APOA 370
	G	-
	H	Thermo Scientific, 2009, 49i
	SO <sub>2</sub>	B
C		Thermo Environment, TEI 43C
D		Horiba Ltd., 2011, UV fluorescence, APSA 370
E		Monitor Europe, ME-9850B (№M1704-M654)
F		HORIBA, 2009, APSA 370
G		-
H		HORIBA , 2005. APSA 360

Table 2: The list of instruments used by participants.

Semi-automatic method adopted by laboratory G:

- NO<sub>2</sub> method is based on the interaction of nitrogen dioxide and sulfanilic acid with a formation of diazo compound which sets off an azo dye in reaction with  $\alpha$ -naphthylamin. Diazo compound colors the solution from light rose to red-violet. Amount of nitrogen dioxide is determined by color intensity (manual, photocolorimetric method, wave length of 540 nm). Range of measurements and error: 0.02 to 0.64 mg/m<sup>3</sup>; d= + 25 %
- NO method is based on the oxidation of nitrogen oxide of chromic acid till dioxide and on the catching of the dioxide with the help of potassium iodine. The diazo compound is formed during the interaction of nitrogen dioxide with sulfanilic acid. This diazo compound is colored from light rose to red-violet while reacting with  $\alpha$ -naphthylamin. Amount of nitrogen dioxide is determined by color intensity (manual, photocolorimetric method, wave length of 540 nm). Range of measurements and error: 0.013 to 0.28 mg/m<sup>3</sup>; d= + 25 %
- O<sub>3</sub> method is based on the displacement of iodine with ozone while ozone is adsorbed by potassium iodine with a buffer based on boric acid. Extracted iodine is determined with a spectrometric measurement, wave length of 325 nm (manual, photo-colorimetric method). Range of measurements and error: 0.01 to 1.0 mg/m<sup>3</sup>; d= + 25 %

- SO<sub>2</sub> method is based on the oxidation of sulphurous gas in the process of its catching from the air with the solution of potassium chlorate or hydrogen peroxide with a further turbidimetric determination of forming sulphat-ion with barium chloride (manual, photocolometric method, wave length of 400 nm). Range of measurements and error: 0.01 – 0.8 mg/m<sup>3</sup>; d= + 25 %.

### 1.3 The preparation of test mixtures

The facility of the UBA National Reference Laboratory is described in [9]. During this IE, gas mixtures were prepared for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> at concentration levels around limit values, critical levels and assessment thresholds set by European Air Quality Directive [1].

The test mixtures were prepared by the dilution of gases from cylinders containing high concentration of NO, NO<sub>2</sub>, SO<sub>2</sub> or CO using thermal mass flow controllers [8]. O<sub>3</sub> was added using an ozone generator.

The participants were required to report three half-hour-mean measurements for each concentration level (run) in order to evaluate the repeatability of standardized measurement methods. Zero concentration levels were generated for one hour and one half-hour-mean measurement was reported. The sequence program of generated test gases is given in Table 3.

day	start time	duration	parameter	installation	calibration	Zero Air	NO	NO2	O3	CO	SO2
		h				nmol/mol	nmol/mol	nmol/mol	nmol/mol	μmol/mol	nmol/mol
23-Oct	13:00	3.5	/	X							
24-Oct	8:45	0.15	/		X						
24-Oct	9:00	2.5	NO			0					
24-Oct	11:45	1.5	NO				200				
24-Oct	13:30	1.5	NO				20				
25-Oct	8:45	0.30	NO2			0					
25-Oct	10:00	1.5	NO2					200			
25-Oct	11:45	1.5	NO2					100			
25-Oct	13:30	1.5	NO2					60			
25-Oct	15:15	1.5	NO2					20			
26-Oct	8:45	1	SO2			0					
26-Oct	10:00	1.5	SO2								130
26-Oct	11:45	1.5	SO2								45
26-Oct	13:30	1.5	SO2								20
26-Oct	15:15	1.5	SO2								5
26-Oct	17:00	1	CO			0					
26-Oct	18:00	2	CO							8	
26-Oct	20:00	2	CO							6	
26-Oct	22:00	2	CO							3	
27-Oct	0:00	2	CO							1	
27-Oct	2:00	2	CO							4.5	
27-Oct	8:45	1	O3			0					
27-Oct	10:00	1.5	O3						300		
27-Oct	11:45	1.5	O3						100		
27-Oct	13:30	1.5	O3						60		
27-Oct	15:15	1.5	O3						20		
28-Oct	8:45	0.15	evaluation								
28-Oct	9:00	3	dismantling								

Table 3: The sequence program of generated test gases

## 2. The evaluation of laboratory's measurement proficiency

To evaluate the participants measurement proficiency the methodology described in ISO 13528 [13] was applied. It has been agreed among the AQUILA members to take the measurement results of UBA as the assigned/reference values for the whole IE [12]. The traceability of UBA's measurement results and the method applied to validate them are presented in Annex A.

All data reported by participating laboratories are presented in Annex C.

As it is described in the said position paper [12], the proficiency of the participants was assessed by calculating two performance indicators. The first performance indicator (z'-score) tests whether the difference between the participants measured value and the assigned/reference value remains within the limits of a common criterion. The second performance indicator (E<sub>n</sub>-number) tests if the difference between the participants measured values and assigned/reference value remains within the limits of a criterion, that is calculated individually for each participant, from the uncertainty of the participants measurement result and the uncertainty of the assigned/reference value.

### 2.1 z' - score

The z'-score statistic is calculated according to ISO 13528 [13] as:

$$z' = \frac{x_i - X}{\sqrt{\sigma_p^2 + u_x^2}} = \frac{x_i - X}{\sqrt{(a \cdot X + b)^2 + u_x^2}} \quad \text{Equation 1}$$

where 'x<sub>i</sub>' is a participant's run average value, 'X' is the assigned/reference value, 'σ<sub>p</sub>' is the 'standard deviation for proficiency assessment' and 'u<sub>x</sub>' is the standard uncertainty of assigned value. For 'a' and 'b' see Table 4.

In the European standards [2], [3], [4] and [5] the uncertainties for calibration gases used in ongoing quality control are prescribed. In fact, it is stated that the maximum permitted expanded uncertainty for calibration gases is 5% and that 'zero gas' shall not give instrument reading higher than the detection limit. As one of the tasks of NRLs is to supply calibration gas mixtures, the 'standard deviation for proficiency assessment' (σ<sub>p</sub>) [13] is calculated in fitness-for-purpose manner from requirements given in European standards.

Over the whole measurement range σ<sub>p</sub> is calculated by linear interpolation between 2.5% at the calibration point (75% of calibration range) and the limit of detection at zero concentration level. The limits of detection of studied measurement methods were evaluated from the data of previous IE. The linear function parameters of σ<sub>p</sub> are given in Table 4:

Gas	σ <sub>p</sub> =a·c+b	
	a	b
	nmol/mol	
SO <sub>2</sub>	0.022	1
CO	0.024	100
O <sub>3</sub>	0.020	1
NO	0.024	1
NO <sub>2</sub>	0.020	1

**Table 4: The standard deviation for proficiency assessment (σ<sub>p</sub>).**

σ<sub>p</sub> is a linear function of concentration (c) with parameters: slope (a) and intercept (b).

The assessment of results in the z'-score evaluation is made according to the following criteria:


- $|z'| \leq 2$  are considered satisfactory.
- $2 < |z'| \leq 3$  are considered questionable.
- $|z'| > 3$  are considered unsatisfactory. Scores falling in this range are very unusual and are taken as evidence that an anomaly has occurred that should be investigated and corrected.

The results of z'-score evaluation are presented in bar plots (Figure 1 to Figure 5) in which the z'-scores of each participant are grouped together, and assessment criteria are presented as  $z' = \pm 2$  and  $z' = \pm 3$  lines.



**Figure 1: The z'-score evaluations of SO<sub>2</sub> measurements**


Scores are given for each participant and each tested concentration level (run). Run number order (with nominal concentration) is: 0 (0 nmol/mol), 1 (130 nmol/mol), 2 (45 nmol/mol), 3 (20 nmol/mol), 4 (5 nmol/mol). The assessment criteria are presented as  $z' = \pm 2$  (blue line) and  $z' = \pm 3$  (red line). They represent the limits for the questionable and unsatisfactory results.



**Figure 2: The z'-score evaluations of CO measurements**


Scores are given for each participant and each tested concentration level (run). Run number order (with nominal concentration) is: 0 (0 µmol/mol), 1 (8 µmol/mol), 2 (6 µmol/mol), 3 (3 µmol/mol), 4 (1 µmol/mol), 5 (4.5 µmol/mol). The assessment criteria are presented as  $z' = \pm 2$  (blue line) and  $z' = \pm 3$  (red line). They represent the limits for the questionable and unsatisfactory results.





**Figure 3: The z'-score evaluations of O<sub>3</sub> measurements**

Scores are given for each participant and each concentration level (run). Run number order (with nominal concentration) is: 0 (0 nmol/mol), 1 (300 nmol/mol), 2 (100 nmol/mol), 3 (60 nmol/mol), 4 (20 nmol/mol). The assessment criteria are presented as  $z'=\pm 2$  (blue line) and  $z'=\pm 3$  (red line). They represent the limits for the questionable and unsatisfactory results.



**Figure 4: The z'-score evaluations of NO measurements**

Scores are given for each participant and each tested concentration level (run). Run number order (with nominal concentration) is: 0 (0 nmol/mol), 1 (200 nmol/mol), 2 (20 nmol/mol). The assessment criteria are presented as  $z'=\pm 2$  (blue line) and  $z'=\pm 3$  (red line). They represent the limits for the questionable and unsatisfactory results.



**Figure 5: The z'-score evaluations of NO<sub>2</sub> measurements**

Scores are given for each participant and each concentration level (run). Run number order (with nominal concentration) is: 0 (0 nmol/mol), 1 (200 nmol/mol), 2 (100 nmol/mol), 3 (60 nmol/mol), 4 (20 nmol/mol). The assessment criteria are presented as  $z'=\pm 2$  (blue line) and  $z'=\pm 3$  (red line). They represent the limits for the questionable and unsatisfactory results.


## 2.2 E<sub>n</sub> - number

The normalized deviations [13] (E<sub>n</sub>) were calculated according to:

$$E_n = \frac{x_i - X}{\sqrt{U_{x_i}^2 + U_X^2}} \quad \text{Equation 2}$$


where 'X' is the assigned/reference value with an expanded uncertainty 'U<sub>X</sub>' and 'x<sub>i</sub>' is the participant's average value with an expanded uncertainty 'U<sub>x<sub>i</sub></sub>'. Satisfactory results are the ones for which  $|E_n| \leq 1$ .

In Figure 6 to Figure 10 the bias of each participant (x<sub>i</sub>-X) are plotted and error bars are used to show the value of denominator of Equation 2 ( $\sqrt{U_{x_i}^2 + U_X^2}$ ). These plots represent also the E<sub>n</sub>-number evaluations where, considering the E<sub>n</sub> criteria ( $|E_n| \leq 1$ ), all results with error bars touching or crossing x-axis are satisfactory. Reported standard uncertainties (Annex B) that are bigger than "standard deviation for proficiency assessments" (σ<sub>p</sub>, Table 4) are considered not fit-for-purpose and are denoted with "\*" in the x-axis of each figure.



**Figure 6: Bias of participant's SO<sub>2</sub> measurement results**


Expanded uncertainty of bias for each run is presented as error bar. The results with error bars touching or crossing the x-axis are satisfactory. For each evaluation the run number (numbers 0 to 4) together with the participants rounded run average (nmol/mol) is given. The '\*' mark indicates reported standard uncertainties bigger than  $\sigma_p$ .



**Figure 7: Bias of participant's CO measurement results**


Expanded uncertainty of bias for each run is presented as error bar. Results with error bars touching or crossing the x-axis are satisfactory. For each evaluation the run number (numbers 0 to 5) together with the participants rounded run average ( $\mu\text{mol/mol}$ ) is given. The '\*' mark indicates reported standard uncertainties bigger than  $\sigma_p$ .

EC harmonization program for Air Quality Measurement  
 Evaluation of the Laboratory Comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub>, Langen 23<sup>rd</sup>-28<sup>th</sup> October 2011




**Figure 8: Bias of participant's O<sub>3</sub> measurement results**

Expanded uncertainty of bias for each run is presented as error bar. Results with error bars touching or crossing the x-axis are satisfactory. For each evaluation the run number (numbers 0 to 4) together with the participants rounded run average (nmol/mol) is given. The '\*' mark indicates reported standard uncertainties bigger than  $\sigma_p$ .



**Figure 9: Bias of participant's NO measurement results**

Expanded uncertainty of bias for each run is presented as error bar. Results with error bars touching or crossing the x-axis are satisfactory. For each evaluation the run number (numbers 0 to 2) together with the participants rounded run average (nmol/mol) is given. The '\*' mark indicates reported standard uncertainties bigger than  $\sigma_p$ .



**Figure 10: Bias of participant's NO<sub>2</sub> measurement results**

Expanded uncertainty of bias is presented as error bar for NO<sub>2</sub> run numbers 0, 1, 2, 3 and 4. Results with error bars touching or crossing the x-axis are satisfactory. For each evaluation the run number together with the participants rounded run average (nmol/mol) is given. The '\*' mark indicates reported standard uncertainties bigger than  $\sigma_p$ .



### 3. Discussion

For a general assessment of the quality of each result a decision diagram was developed (Figure 11) that results in seven categories (1 to 7). The general comments for each category are:

- **1:** measurement result is completely satisfactory
- **2:** measurement result is satisfactory (z'-score satisfactory and En-number ok) but the reported uncertainty is too high
- **3:** measured value is satisfactory (z'-score satisfactory) but the reported uncertainty is underestimated (En-number not ok)
- **4:** measurement result is questionable (z'-score questionable) but due to a high reported uncertainty can be considered valid (En-number ok)
- **5:** measurement result is questionable (z'-score questionable and En-number not ok)
- **6:** measurement result is unsatisfactory (z'-score unsatisfactory) but due to a high reported uncertainty can be considered valid (En-number ok)
- **7:** measurement result is unsatisfactory (z'-score unsatisfactory and En-number not ok)




Figure 11: The decision diagram for general assessment of proficiency results.

The results of the IE were assigned to categories according to the diagram given in Figure 11 and are presented in Table 5.

	run number	Ref. conc. level	IE code						
			B	C	D	E	F	G	
CO (µmol/mol)	0	0.003	2	1	1	2	1	nd	
	1	7.864	1	1	2	2	2	nd	
	2	5.901	1	1	2	2	2	nd	
	3	2.951	1	1	2	2	2	nd	
	4	0.976	1	1	2	1	1	nd	
	5	4.428	1	1	1	2	2	nd	
NO (nmol/mol)	0	0.00	1	1	1	1	1	1	
	1	200.03	1	1	1	2	2	2	
	2	19.21	1	1	1	2	1	2	
NO <sub>2</sub> (nmol/mol)	0	-0.47	2	2	1	1	1	1	
	1	197.16	1	2	2	2	2	2	
	2	100.32	1	1	1	2	2	2	
	3	60.33	1	1	1	2	2	2	
	4	20.68	1	1	1	2	1	2	
O <sub>3</sub> (nmol/mol)	0	0.04	1	1	1	1	1	1	
	1	299.67	1	1	2	2	2	2	
	2	100.23	1	1	2	2	2	4	
	3	59.79	1	1	2	2	2	2	
	4	20.02	1	1	1	1	1	2	
SO <sub>2</sub> (nmol/mol)	0	0.20	1	1	1	1	1	1	
	1	129.99	1	1	2	2	2	2	
	2	45.05	1	1	1	2	2	2	
	3	20.14	1	1	1	1	1	2	
	4	5.17	1	1	1	1	1	2	

Table 5: The general assessment of proficiency results. "nd" is referring to values not reported.

## 4. Conclusions

The proficiency evaluation scheme has provided an assessment of the participants measured values and their evaluated uncertainties.

In terms of the criteria imposed by the European Directive ( $\sigma_p$ ) 59.4% of the results reported by WHO/AQUILA laboratories fall into category '1' and are good both in terms of measured values and evaluated uncertainties. Among the remaining results the 39.9% presented good measured values but the evaluated uncertainties were too high (category '2') and 0.7% of results (category '4') were questionable compared to z-score and OK for the En-number.

As in previous IE, the adopted criteria for high concentrations were the standard deviations for proficiency assessment, deriving from the European Standards' uncertainty requirements. The reproducibility standard deviations obtained at this IE (Annex C) and previous IE [20], [21], [22], [23], [24], [25], [33] are comparable to the mentioned criteria. On the other hand, the uncertainty criteria for zero levels were those set in AQUILA's position paper [12].

In the present IE compared to the past (see Table 6) it has been found a low share of results in category '1'.

A relative high percentage of results falling in category '2' was found and it could be useful to investigate in detail the procedure to calculate the uncertainty used by the participants.

IE	Categories %						
	1	2	3	4	5	6	7
Apr-08	68.4	18.1	7.3	1.0	1.0	2.6	1.6
Oct-08 (I)	37.9	40.8	14.2	0.6	3.6	1.0	1.9
Oct-08 (II)	34.3	38.9	23.7	1.0	2.0	0.0	0.0
Sep-09	60.8	29.9	3.1	4.1	1.0	1.0	0.0
Oct-09	85.0	5.7	7.5	0.4	1.4	0.0	0.0
Jun-10	84.6	8.1	4.4	0.7	2.3	0.0	0.0
Sep-11	86.0	7.9	5.4	0.0	0.3	0.0	0.3
Oct-11 (I)	78.5	12.5	7.6	0.0	1.3	0.0	0.0
Oct-11 (II)	59.4	39.9	0.0	0.7	0.0	0.0	0.0

Table 6: history of the results in the last IE

Comparability of results among participants at the highest concentration level (from Figure 36 to Figure 40), excluding outliers, is acceptable for NO and CO measurements while NO<sub>2</sub> and O<sub>3</sub> and SO<sub>2</sub> one showed less satisfactory results.

The relative reproducibility limits, at the highest studied concentration levels, are 11.5% for SO<sub>2</sub>, 11.0% for CO, 8.5% for O<sub>3</sub>, 9.9% for NO and 10.3% for NO<sub>2</sub>. Only NO and CO are within the objective derived from criteria imposed by the European Commission ( $\sigma_p$ ).

During this IE the performance of all participants has been quite positive. Only two outliers have been identified at zero level for NO and CO (Annex D) and 1 straggler for NO.

In this exercise there were no unsatisfactory results in the z'-score evaluations. Laboratory G obtained one questionable result for O<sub>3</sub>. The good performance of this IE is above the average of the last years as shown in Table 7.

<i>IE</i>	<i>Site</i>	<i>Questionable</i>	<i>Unsatisfactory</i>	<i>Satisfactory</i>
June-05	Ispra (IT)	2.3%	2.3%	95.5%
June-07	Ispra (IT)	1.9%	0.3%	97.8%
October-07	Essen (DE)	4.6%	2.2%	93.2%
April-08	Ispra (IT)	2.1%	4.1%	93.8%
October 2008_1	Ispra (IT)	4.2%	2.9%	92.9%
October 2008_2	Ispra (IT)	3.0%	0.0%	97.0%
September-09	Langen (DE)	4.7%	0.9%	94.3%
October-09	Ispra (IT)	1.8%	0.0%	98.2%
June-10	Ispra (IT)	3.0%	0.0%	97.0%
September-11	Ispra (IT)	0.3%	0.3%	99.4%
October-11	Ispra (IT)	1.3%	0.0%	98.7%
October-11	Langen (DE)	0.7%	0.0%	99.3%

**Table 7: z'-score summary**

## 5. References

- [1] Directive 2008/50/EC of the European Parliament and of the Council of 21 May 2008 on ambient air quality and cleaner air for Europe, L 152, 11.06.2008
- [2] EN 14626:2005, Ambient air quality - Standard method for the measurement of the concentration of carbon monoxide by non-dispersive infrared spectroscopy
- [3] EN 14212:2005, Ambient air quality - Standard method for the measurement of the concentration of sulphur dioxide by ultraviolet fluorescence
- [4] EN 14211:2005, Ambient air quality - Standard method for the measurement of the concentration of nitrogen dioxide and nitrogen monoxide by chemiluminescence
- [5] EN 14625:2005, Ambient air quality - Standard method for the measurement of the concentration of ozone by ultraviolet photometry
- [6] ISO 6143:2001, Gas analysis - Comparison methods for determining and checking the composition of calibration gas mixtures
- [7] ISO 6144:2003, Gas analysis - Preparation of calibration gas mixtures - Static volumetric method
- [8] ISO 6145-7:2001, Gas analysis - Preparation of calibration gas mixtures using dynamic volumetric methods - Part 7: Thermal mass-flow controllers
- [9] Mücke H.-G., (2008), Air quality management in the WHO European Region – Results of a quality assurance and control programme on air quality monitoring (1994-2004), Environment International, EI-01718
- [10] Mücke H.-G., et al. (2000), European Intercomparison workshop on air quality monitoring vol.4 – Measuring NO, NO<sub>2</sub>, O<sub>3</sub> and SO<sub>2</sub> – Air Hygiene Report 13, WHO Collaboration Centre for Air Quality Management and Air Pollution Control, ISSN 0938 - 9822
- [11] <http://ies.jrc.ec.europa.eu/aquila-project/aquila-homepage.html>
- [12] AQUILA POSITION PAPER N. 37, (2008) Protocol for intercomparison exercise. Organisation of intercomparison exercises for gaseous air pollution for EU national air quality reference laboratories and laboratories of the WHO EURO region [http://ies.jrc.ec.europa.eu/uploads/fileadmin/H04/Air\\_Quality/N%2037%20final%20version%20IE%20organisation%20and%20evaluation.pdf](http://ies.jrc.ec.europa.eu/uploads/fileadmin/H04/Air_Quality/N%2037%20final%20version%20IE%20organisation%20and%20evaluation.pdf)
- [13] ISO 13528:2005, Statistical methods for use in proficiency testing by interlaboratory comparisons
- [14] ISO 5725-1:1994, Accuracy (trueness and precision) of measurement methods and results – Part 1: General principles and definitions
- [15] ISO 5725-2:1994, Accuracy (trueness and precision) of measurement methods and results – Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method
- [16] ISO 5725-6:1994, Accuracy (trueness and precision) of measurement methods and results - Part 6: Use in practice of accuracy values
- [17] Mücke H.-G., (2008), Air quality management in the WHO European Region – Results of a quality assurance and control programme on air quality monitoring (1994-2004), Environment International, EI-01718
- [18] De Saeger E. et al., (1997) European comparison of Nitrogen Dioxide calibration methods, EUR 17661
- [19] ISO 15337:2009, Ambient air - Gas phase titration - Calibration of analysers for ozone
- [20] Kapus M. et al. (2009) The evaluation of the Intercomparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> carried out in June 2007 in Ispra . JRC scientific and technical reports. EUR 23804.
- [21] Kapus M. et al. (2009) The evaluation of the Intercomparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> - April 2008. JRC scientific and technical reports. EUR 23805.

- [22] Kapus M. et al. (2009) The evaluation of the Intercomparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> 6-9 October 2008. JRC scientific and technical reports. EUR 23806.
- [23] Kapus M. et al. (2009) The evaluation of the Intercomparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> 13-16 October 2008.. JRC scientific and technical reports. EUR 23807.
- [24] Belis C. A. et al. (2010) The evaluation of the Interlaboratory comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> Langen 20-25 September 2009.
- [25] Belis C. A. et al. (2010) The evaluation of the Interlaboratory comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub> 19-22 October 2009.
- [26] Viallon J. et al 2009 Metrologia 46 08017. Final report, on-going key comparison BIPM.QM-K1: Ozone at ambient level, comparison with JRC, 2008. doi: 10.1088/0026-1394/46/1A/08017
- [27] Viallon, J., et al. (2006), International comparison CCQM-P28: Ozone at ambient level, Metrologia, 43, Tech. Suppl., 08010, doi:10.1088/0026-1394/43/1A/08010
- [28] Tanimoto, H., et al. (2006), Intercomparison of ultraviolet photometry and gas-phase titration techniques for ozone reference standards at ambient levels, Journal of Geophysical Research, vol. 111, D16313, doi:10.1029/2005JD006983
- [29] GUM Workbench, The Tool for Expression of Uncertainty of Measurements
- [30] VDI 2449 Part3: 2001, Measurement methods test criteria- General method for the determination of the uncertainty of calibratable measurement methods.
- [31] Mücke H-G, et al. (1996). European Intercomparison Workshops on Air Quality Monitoring. Vol. 2 – Measuring of CO, NO, NO<sub>2</sub> and O<sub>3</sub> – Air Hygiene Report 9. Berlin, Germany: WHO Collaborating Centre for Air Quality Management and Air Pollution Control; ISSN 0938-9822.
- [32] ISO 17043:2010, Conformity assessment -- General requirements for proficiency testing
- [33] C. A. Belis, F. Lagler, M. Barbieri, H.G. Mücke, K. Wirtz and V. Stummer (2009) The evaluation of the Interlaboratory Comparison Exercise for SO<sub>2</sub>, O<sub>3</sub>, NO and NO<sub>2</sub> Langen 20th-25th September 2009.
- [34] ISO 6144:2003, Gas analysis - Preparation of calibration gas mixtures - Static volumetric method

### **Annex A. Assigned values**

The assigned values of tested concentration levels (run) were derived from UBA measurements which are calibrated against the certified reference values of CRMs and are traceable to international standards. In this perspective the assigned values are reference values as defined in the ISO 13528 [13].

UBA's SO<sub>2</sub>, CO and NO analysers were calibrated according to the methodology described in the ISO 6143 [6]. The procedure and the device for generating primary calibration gases is described elsewhere [31]. Gas mixtures for the calibration experiment were produced from the reference mixtures by static volumetric dilution method ISO 6144 [34].

SO<sub>2</sub>, CO and NO gas mixtures manufactured by Air Liquide and certified by UBA (U ≤ 2%) were used as internal standards.

For the reference gas mixture composition evaluation and for the calibration experiment evaluation two computer applications were used, the "GUM WORKBENCH" [20] and "ProControl®" [31].

For O<sub>3</sub> measurements, the primary standard NIST photometer SRP 29 was used.

UBA's measurement results were validated by comparison to the group statistics (x\* and s\*) for every parameter and concentration level of the IE. These statistics are calculated from participants, applying the robust method described in the Annex C of the ISO 13528 [13]. The validation is taking into account UBA's measurement result (X) and its standard uncertainty (u<sub>X</sub>) as given in Equation 3[13]:

$$\frac{|x^* - X|}{\sqrt{\frac{(1,25 \cdot s^*)^2}{p} + u_X^2}} < 2 \quad \text{Equation 3}$$

Where 'x\*' and 's\*' represent robust average and robust standard deviation respectively and 'p' is the number of participants.

In Table 8 all inputs for expression **Error! Reference source not found.** are given and all UBA's measurement results are confirmed to be valid.

As a group evaluation robust average (x\*) and robust standard deviation (s\*) were calculated (applying the procedure described in Annex C of ISO 13528 for each run and are presented in the following Table 8.

run	unit	X	uX'	x*	s*	p	val.
NO2_0	nmol/mol	-0.47	0.71	-0.061	0.179	7	OK
NO2_1	nmol/mol	197.163	2.26	196.038	5.284	7	OK
NO2_2	nmol/mol	100.32	1.3	98.606	3.457	7	OK
NO2_3	nmol/mol	60.33	0.95	58.483	2.377	7	OK
NO2_4	nmol/mol	20.677	0.74	20.033	0.971	7	OK
O3_0	nmol/mol	0.04	0.05	0.013	0.074	7	OK
O3_1	nmol/mol	299.667	3.52	299.912	5.91	7	OK
O3_2	nmol/mol	100.23	1.29	100.256	2.852	7	OK
O3_3	nmol/mol	59.79	0.92	60.444	1.659	7	OK
O3_4	nmol/mol	20.017	0.6	19.758	0.681	7	OK
SO2_0	nmol/mol	0.2	0.43	0.007	0.343	7	OK
SO2_1	nmol/mol	129.987	1.42	130.118	2.728	7	OK
SO2_2	nmol/mol	45.047	0.64	45.119	1.699	7	OK
SO2_3	nmol/mol	20.143	0.48	20.111	1.003	7	OK
SO2_4	nmol/mol	5.167	0.44	5.016	0.412	7	OK
CO_0	µmol/mol	0.003	0.022	0.002	0.003	6	OK
CO_1	µmol/mol	7.8637	0.086	7.964	0.271	6	OK
CO_2	µmol/mol	5.9013	0.066	5.908	0.092	6	OK
CO_3	µmol/mol	2.9513	0.038	2.98	0.112	6	OK
CO_4	µmol/mol	0.976	0.024	1.019	0.061	6	OK
CO_5	µmol/mol	4.428	0.057	4.449	0.129	6	OK

**Table 8: The validation of assigned values (X)**

by comparison to the robust averages (x\*) with taking into account the standard uncertainties of assigned values (uX'), and robust standard deviations (s\*) as denoted by Equation 3.

The homogeneity of test gas was evaluated from measurements at the beginning and end of the distribution line. From the relative differences between beginning and end measurements, average and standard deviation were calculated, and the uncertainty of test gas due to lack of homogeneity was calculated as the sum of squares of these average and standard deviation. The upper and lower limits of bias due to homogeneity was evaluated to be smaller than 0.5% which constitutes the relative standard uncertainty of 0,3% of each concentration level. The standard uncertainties of assigned/reference values (u<sub>X'</sub>) were calculated with Equation 4 and used in the proficiency evaluations of chapter 2.

$$u_{X'}^2 = u_X^2 + \left( X \cdot u_{\text{homogeneity}} \right)^2 \quad \text{Equation 4}$$



## Annex B. The results of the IE

In this annex are reported participant's results, presented both in tables and graphs. For each run, participants were asked to report 3 results representing 30 minutes measurement each ( $x_{ij}$ ). In this annex are presented the reported data and their uncertainty  $u(x_i)$  and  $U(x_i)$  expressed in mol/mol units.

For all the runs except concentration levels 0, also average ( $\bar{x}_i$ ) and standard deviation ( $s_i$ ) of each participant are presented.

The assigned value is indicated on the graphs with the red line and the individual laboratories expanded uncertainties ( $Ux_i$ ) are indicated with error bars.

### Reported values for SO<sub>2</sub>

values	laboratories						
	B	C	D	E	F	G	H
$x_{i,1}$ (nmol/mol)	0.15	-0.30	0.00	-0.50	0.95	0.00	0.20
$u(x_i)$ (nmol/mol)	0.51	0.68	0.15	0.98	0.60	0.00	0.43
$U(x_i)$ (nmol/mol)	1.01	1.37	0.30	1.96	1.20	0.00	0.86

Table 9: Reported values for SO<sub>2</sub> run 0.




Figure 12: Reported values for SO<sub>2</sub> run 0.

values	laboratories						
	B	C	D	E	F	G	H
xi,1 (nmol/mol)	129.56	131.67	124.96	129.00	137.35	124.36	129.96
xi,2 (nmol/mol)	129.34	131.76	125.29	128.00	137.40	131.06	130.11
xi,3 (nmol/mol)	129.57	131.80	125.47	128.00	137.45	138.47	129.89
Xi (nmol/mol)	129.49	131.74	125.24	128.33	137.40	131.29	129.98
Si (nmol/mol)	0.13	0.06	0.25	0.57	0.05	7.05	0.11
u(xi) (nmol/mol)	3.00	2.95	4.09	8.02	8.86	7.22	1.37
U(xi) (nmol/mol)	6.00	5.90	8.18	16.04	17.72	22.98	2.74

Table 10: Reported values for SO<sub>2</sub> run 1.




Figure 13: Reported values for SO<sub>2</sub> run 1.

values	laboratories						
	B	C	D	E	F	G	H
xi,1 (nmol/mol)	44.89	45.32	42.88	44.00	48.13	49.33	45.11
xi,2 (nmol/mol)	44.81	45.38	42.79	44.00	48.24	42.09	44.98
xi,3 (nmol/mol)	44.75	45.38	42.79	44.00	48.22	47.83	45.05
Xi (nmol/mol)	44.81	45.36	42.82	44.00	48.19	46.41	45.04
Si (nmol/mol)	0.07	0.03	0.05	0.00	0.05	3.82	0.06
u(xi) (nmol/mol)	1.04	1.35	1.40	2.75	3.11	3.97	0.62
U(xi) (nmol/mol)	2.08	2.70	2.80	5.50	6.22	12.64	1.25

Table 11: Reported values for SO<sub>2</sub> run 2.




Figure 14: Reported values for SO<sub>2</sub> run 2.

values	laboratories						
	B	C	D	E	F	G	H
x <sub>i,1</sub> (nmol/mol)	19.88	20.07	18.93	19.10	21.79	23.92	20.10
x <sub>i,2</sub> (nmol/mol)	19.88	20.02	18.95	19.50	21.85	21.52	20.18
x <sub>i,3</sub> (nmol/mol)	19.82	19.88	18.87	19.50	21.78	17.94	20.15
X <sub>i</sub> (nmol/mol)	19.86	19.99	18.91	19.36	21.80	21.12	20.14
S <sub>i</sub> (nmol/mol)	0.03	0.09	0.04	0.23	0.03	3.00	0.04
u(x <sub>i</sub> ) (nmol/mol)	0.61	1.09	0.62	1.01	1.41	3.17	0.47
U(x <sub>i</sub> ) (nmol/mol)	1.22	2.18	1.24	2.03	2.81	10.09	0.95

Table 12: Reported values for SO<sub>2</sub> run 3.




Figure 15: Reported values for SO<sub>2</sub> run 3.

values	laboratories						
	B	C	D	E	F	G	H
x <sub>i,1</sub> (nmol/mol)	4.94	4.73	4.73	4.70	6.05	6.70	5.19
x <sub>i,2</sub> (nmol/mol)	4.87	4.65	4.70	4.70	6.03	5.74	5.16
x <sub>i,3</sub> (nmol/mol)	4.92	4.54	4.76	4.70	5.99	5.38	5.15
X <sub>i</sub> (nmol/mol)	4.91	4.64	4.73	4.70	6.02	5.94	5.16
S <sub>i</sub> (nmol/mol)	0.03	0.09	0.03	0.00	0.03	0.68	0.02
u(x <sub>i</sub> ) (nmol/mol)	0.32	0.93	0.15	0.98	0.39	0.83	0.43
U(x <sub>i</sub> ) (nmol/mol)	0.64	1.87	0.31	1.96	0.78	2.63	0.87

Table 13: Reported values for SO<sub>2</sub> run 4.




Figure 16: Reported values for SO<sub>2</sub> run 4.

## Reported values for CO

values	laboratories					
	B	C	D	E	F	H
$x_{i,1}$ ( $\mu\text{mol/mol}$ )	0.000	0.056	0.000	0.000	0.010	0.003
$u(x_i)$ ( $\mu\text{mol/mol}$ )	0.500	0.054	0.010	0.170	0.018	0.022
$U(x_i)$ ( $\mu\text{mol/mol}$ )	1.000	0.109	0.020	0.340	0.035	0.044

Table 14: Reported values for CO run 0.




Figure 17 Reported values for CO run 0

values	laboratories					
	B	C	D	E	F	H
$x_{i,1}$ ( $\mu\text{mol/mol}$ )	7.676	7.762	7.994	8.320	8.120	7.859
$x_{i,2}$ ( $\mu\text{mol/mol}$ )	7.680	7.773	7.999	8.260	8.191	7.864
$x_{i,3}$ ( $\mu\text{mol/mol}$ )	7.679	7.782	7.995	8.300	8.237	7.868
$\bar{X}_i$ ( $\mu\text{mol/mol}$ )	7.678	7.772	7.996	8.293	8.183	7.864
$S_i$ ( $\mu\text{mol/mol}$ )	0.002	0.010	0.003	0.031	0.059	0.005
$u(x_i)$ ( $\mu\text{mol/mol}$ )	0.212	0.233	0.360	0.550	0.543	0.083
$U(x_i)$ ( $\mu\text{mol/mol}$ )	0.424	0.466	0.720	1.100	1.086	0.166

Table 15: Reported values for CO run 1.




Figure 18: Reported values for CO run 1.

values	laboratories					
	B	C	D	E	F	H
$x_{i,1}$ ( $\mu\text{mol/mol}$ )	5.758	5.878	5.995	5.860	5.991	5.901
$x_{i,2}$ ( $\mu\text{mol/mol}$ )	5.756	5.888	5.990	5.920	5.999	5.901
$x_{i,3}$ ( $\mu\text{mol/mol}$ )	5.751	5.890	5.989	5.900	6.011	5.902
$\bar{X}_i$ ( $\mu\text{mol/mol}$ )	5.755	5.885	5.991	5.893	6.000	5.901
$S_i$ ( $\mu\text{mol/mol}$ )	0.004	0.006	0.003	0.031	0.010	0.001
$u(x_i)$ ( $\mu\text{mol/mol}$ )	0.163	0.193	0.270	0.400	0.397	0.064
$U(x_i)$ ( $\mu\text{mol/mol}$ )	0.326	0.386	0.540	0.800	0.794	0.128

Table 16: Reported values for CO run 2.




Figure 19: Reported values for CO run 2.

values	laboratories					
	B	C	D	E	F	H
$x_{i,1}$ ( $\mu\text{mol/mol}$ )	2.884	3.029	3.000	2.880	3.166	2.952
$x_{i,2}$ ( $\mu\text{mol/mol}$ )	2.881	3.025	2.995	2.880	3.168	2.951
$x_{i,3}$ ( $\mu\text{mol/mol}$ )	2.881	3.035	2.998	2.880	3.172	2.951
$\bar{X}_i$ ( $\mu\text{mol/mol}$ )	2.882	3.030	2.998	2.880	3.169	2.951
$S_i$ ( $\mu\text{mol/mol}$ )	0.002	0.005	0.003	0.000	0.003	0.001
$u(x_i)$ ( $\mu\text{mol/mol}$ )	0.091	0.106	0.200	0.200	0.209	0.037
$U(x_i)$ ( $\mu\text{mol/mol}$ )	0.182	0.211	0.400	0.390	0.419	0.074

Table 17: Reported values for CO run 3.




Figure 20: Reported values for CO run 3.

values	laboratories					
	B	C	D	E	F	H
$x_{i,1}$ ( $\mu\text{mol/mol}$ )	0.956	1.091	1.003	0.980	1.069	0.976
$x_{i,2}$ ( $\mu\text{mol/mol}$ )	0.952	1.093	1.000	1.040	1.072	0.976
$x_{i,3}$ ( $\mu\text{mol/mol}$ )	0.953	1.094	1.000	1.040	1.073	0.976
$X_i$ ( $\mu\text{mol/mol}$ )	0.954	1.093	1.001	1.020	1.071	0.976
$S_i$ ( $\mu\text{mol/mol}$ )	0.002	0.002	0.002	0.035	0.002	0.000
$u(x_i)$ ( $\mu\text{mol/mol}$ )	0.042	0.063	0.135	0.070	0.071	0.024
$U(x_i)$ ( $\mu\text{mol/mol}$ )	0.084	0.126	0.270	0.140	0.142	0.048

Table 18: Reported values for CO run 4.




Figure 21: Reported values for CO run 4.

values	laboratories					
	B	C	D	E	F	H
$x_{i,1}$ ( $\mu\text{mol/mol}$ )	4.310	4.496	4.478	4.800	4.340	4.427
$x_{i,2}$ ( $\mu\text{mol/mol}$ )	4.309	4.505	4.481	4.880	4.352	4.428
$x_{i,3}$ ( $\mu\text{mol/mol}$ )	4.309	4.509	4.480	4.860	4.359	4.429
$X_i$ ( $\mu\text{mol/mol}$ )	4.309	4.503	4.480	4.847	4.350	4.428
$S_i$ ( $\mu\text{mol/mol}$ )	0.001	0.007	0.002	0.042	0.010	0.001
$u(x_i)$ ( $\mu\text{mol/mol}$ )	0.122	0.142	0.050	0.320	0.288	0.055
$U(x_i)$ ( $\mu\text{mol/mol}$ )	0.244	0.284	0.100	0.640	0.575	0.110

Table 19: Reported values for CO run 5.




Figure 22: Reported values for CO run 5.

## Reported values for O<sub>3</sub>

values	laboratories						
	B	C	D	E	F	G	H
$x_{i,1}$ (nmol/mol)	-0.09	-0.04	0.37	0.00	0.26	0.00	0.04
$u(x_i)$ (nmol/mol)	0.54	1.00	0.25	0.93	1.00	0.00	0.55
$U(x_i)$ (nmol/mol)	1.09	2.00	0.50	1.86	2.00	0.00	0.10

Table 20: Reported values for O<sub>3</sub> run 0.




Figure 23: Reported values for O<sub>3</sub> run 0.

values	laboratories						
	B	C	D	E	F	G	H
$x_{i,1}$ (nmol/mol)	291.70	302.18	293.96	297.00	289.68	288.53	297.57
$x_{i,2}$ (nmol/mol)	296.40	309.75	298.38	304.00	294.50	299.16	300.11
$x_{i,3}$ (nmol/mol)	298.38	313.04	300.53	313.00	295.53	313.43	301.32
$\bar{X}_i$ (nmol/mol)	295.49	308.32	297.62	304.66	293.23	300.37	299.66
$S_i$ (nmol/mol)	3.43	5.56	3.35	8.02	3.12	12.49	1.91
$u(x_i)$ (nmol/mol)	2.95	4.39	11.77	22.24	13.34	12.66	3.40
$U(x_i)$ (nmol/mol)	5.80	8.79	23.54	44.47	26.68	40.29	6.80

Table 21: Reported values for O<sub>3</sub> run 1




Figure 24: Reported values for O<sub>3</sub> run 1.

values	laboratories						
	B	C	D	E	F	G	H
xi,1 (nmol/mol)	99.51	104.85	100.67	104.00	99.14	99.89	100.67
xi,2 (nmol/mol)	98.78	104.37	100.04	103.00	98.45	87.36	100.07
xi,3 (nmol/mol)	98.74	104.19	99.88	103.00	98.30	88.54	99.95
$\bar{X}_i$ (nmol/mol)	99.01	104.47	100.19	103.33	98.63	91.93	100.23
Si (nmol/mol)	0.43	0.34	0.41	0.57	0.44	6.91	0.38
u(xi) (nmol/mol)	1.00	2.49	3.97	7.53	4.49	7.07	1.25
U(xi) (nmol/mol)	2.00	4.98	7.94	15.06	8.98	22.50	2.50

Table 22: Reported values for O<sub>3</sub> run 2.




Figure 25: Reported values for O<sub>3</sub> run 2.

values	laboratories						
	B	C	D	E	F	G	H
xi,1 (nmol/mol)	59.15	62.58	60.09	61.00	59.07	68.66	59.97
xi,2 (nmol/mol)	58.90	62.55	60.05	61.00	58.99	57.22	59.73
xi,3 (nmol/mol)	58.62	62.64	59.93	60.00	58.82	61.35	59.67
$\bar{X}_i$ (nmol/mol)	58.89	62.59	60.02	60.66	58.96	62.41	59.79
Si (nmol/mol)	0.26	0.04	0.08	0.57	0.12	5.79	0.15
u(xi) (nmol/mol)	0.80	1.98	2.38	4.42	2.68	5.96	0.90
U(xi) (nmol/mol)	1.60	3.96	4.76	8.85	5.37	18.95	1.80

Table 23: Reported values for O<sub>3</sub> run 3.




Figure 26: Reported values for O<sub>3</sub> run 3.



values	laboratories						
	B	C	D	E	F	G	H
x <sub>i,1</sub> (nmol/mol)	19.30	20.85	20.13	19.00	19.68	21.53	19.99
x <sub>i,2</sub> (nmol/mol)	19.56	20.90	20.20	19.00	19.76	19.32	20.03
x <sub>i,3</sub> (nmol/mol)	19.58	20.74	20.20	19.00	19.75	16.68	20.03
X <sub>i</sub> (nmol/mol)	19.48	20.83	20.17	19.00	19.73	19.17	20.01
S <sub>i</sub> (nmol/mol)	0.15	0.08	0.04	0.00	0.04	2.42	0.02
u(x <sub>i</sub> ) (nmol/mol)	0.60	1.30	0.80	0.93	0.90	2.59	0.60
U(x <sub>i</sub> ) (nmol/mol)	1.20	2.60	1.60	1.86	1.80	8.23	1.20

Table 24: Reported values for O<sub>3</sub> run 4.




Figure 27: Reported values for O<sub>3</sub> run 4.

## Reported values for NO

values	laboratories						
	B	C	D	E	F	G	H
xi,1 (nmol/mol)	0.05	0.01	0.11	0.00	0.56	0.00	0.00
u(xi) (nmol/mol)	0.58	0.75	0.25	0.10	0.60	0.00	0.90
U(xi) (nmol/mol)	1.16	1.50	0.50	0.21	1.20	0.00	1.80

Table 25: Reported values for NO run 0.




Figure 28: Reported values for NO run 0.

values	laboratories						
	B	C	D	E	F	G	H
xi,1 (nmol/mol)	197.42	199.88	208.22	202.00	198.58	216.77	199.91
xi,2 (nmol/mol)	197.32	199.78	208.70	201.00	198.65	187.69	200.06
xi,3 (nmol/mol)	197.36	200.07	208.81	202.00	198.64	201.29	200.11
Xi (nmol/mol)	197.36	199.91	208.57	201.66	198.62	201.91	200.02
Si (nmol/mol)	0.05	0.14	0.31	0.57	0.03	14.55	0.10
u(xi) (nmol/mol)	5.00	4.21	5.17	14.78	8.39	14.72	2.19
U(xi) (nmol/mol)	10.00	8.42	10.35	29.56	16.77	46.83	4.39

Table 26: Reported values for NO run 1.




Figure 29: Reported values for NO run 1.

values	laboratories						
	B	C	D	E	F	G	H
$x_{i,1}$ (nmol/mol)	18.65	18.23	20.90	21.00	19.52	19.37	19.26
$x_{i,2}$ (nmol/mol)	18.53	18.12	20.75	21.00	19.51	17.83	19.21
$x_{i,3}$ (nmol/mol)	18.57	18.04	20.49	22.00	19.57	22.13	19.16
$\bar{X}_i$ (nmol/mol)	18.58	18.13	20.71	21.33	19.53	19.77	19.21
$S_i$ (nmol/mol)	0.06	0.09	0.20	0.57	0.03	2.17	0.05
$u(x_i)$ (nmol/mol)	0.70	0.84	0.53	1.57	0.75	2.34	0.92
$U(x_i)$ (nmol/mol)	1.40	1.69	1.06	3.13	1.65	7.45	1.85

Table 27: Reported values for NO run 2.




Figure 30: Reported values for NO run 2.

## Reported values for NO<sub>2</sub>

values	laboratories						
	B	C	D	E	F	G	H
$x_{i,1}$ (nmol/mol)	-0.18	-0.07	-0.08	0.50	0.04	0.00	-0.47
$u(x_i)$ (nmol/mol)	1.12	1.00	0.25	0.17	0.60	0.00	0.71
$U(x_i)$ (nmol/mol)	2.24	2.00	0.50	0.35	1.20	0.00	1.42

Table 28: Reported values for NO<sub>2</sub> run 0.




Figure 31: Reported values for NO<sub>2</sub> run 0.

values	laboratories						
	B	C	D	E	F	G	H
$x_{i,1}$ (nmol/mol)	190.84	197.37	202.04	189.00	195.94	189.13	196.85
$x_{i,2}$ (nmol/mol)	189.91	196.70	202.19	190.00	195.46	213.61	196.87
$x_{i,3}$ (nmol/mol)	189.48	196.71	202.76	191.00	195.61	197.56	197.77
$\bar{X}_i$ (nmol/mol)	190.07	196.92	202.33	190.00	195.67	200.10	197.16
$S_i$ (nmol/mol)	0.69	0.38	0.38	1.00	0.24	12.43	0.52
$u(x_i)$ (nmol/mol)	4.18	5.17	5.02	13.93	11.46	12.60	2.18
$U(x_i)$ (nmol/mol)	8.36	10.34	10.04	27.85	22.91	40.09	4.36

Table 29: Reported values for NO<sub>2</sub> run 1.




Figure 32: Reported values for NO<sub>2</sub> run 1.

values	laboratories						
	B	C	D	E	F	G	H
xi,1 (nmol/mol)	95.51	96.22	102.11	96.00	98.39	111.98	100.28
xi,2 (nmol/mol)	95.29	95.68	102.13	96.00	98.12	100.98	100.90
xi,3 (nmol/mol)	95.18	95.53	102.06	96.00	98.08	94.51	99.78
Xi (nmol/mol)	95.32	95.81	102.10	96.00	98.19	102.49	100.32
Si (nmol/mol)	0.16	0.36	0.03	0.00	0.16	8.83	0.56
u(xi) (nmol/mol)	2.70	2.50	2.53	7.04	5.75	9.00	1.26
U(xi) (nmol/mol)	5.40	5.00	5.07	14.07	11.50	28.63	2.52

Table 30: Reported values for NO<sub>2</sub> run 2.




Figure 33: Reported values for NO<sub>2</sub> run 2.

values	laboratories						
	B	C	D	E	F	G	H
xi,1 (nmol/mol)	57.55	56.65	61.74	59.00	59.10	61.75	60.53
xi,2 (nmol/mol)	57.50	56.51	61.69	59.00	59.14	51.22	60.31
xi,3 (nmol/mol)	57.57	56.28	61.36	58.00	59.07	54.03	60.15
Xi (nmol/mol)	57.54	56.48	61.59	58.66	59.10	55.66	60.33
Si (nmol/mol)	0.03	0.18	0.20	0.57	0.03	5.45	0.19
u(xi) (nmol/mol)	1.80	1.63	1.53	4.30	3.46	5.61	0.93
U(xi) (nmol/mol)	3.60	3.26	3.07	8.60	6.92	17.85	1.86

Table 31: Reported values for NO<sub>2</sub> run 3.




Figure 34: Reported values for NO<sub>2</sub> run 3.

values	laboratories							
	B	C	D	E	F	G	H	
$x_{i,1}$ (nmol/mol)	19.68	18.57	21.31	21.00	20.20	21.63	20.97	
$x_{i,2}$ (nmol/mol)	19.68	18.57	21.14	21.00	19.93	18.99	20.66	
$x_{i,3}$ (nmol/mol)	19.63	18.56	20.92	20.00	20.06	17.41	20.40	
$\bar{X}_i$ (nmol/mol)	19.66	18.56	21.12	20.66	20.06	19.34	20.67	
$S_i$ (nmol/mol)	0.02	0.00	0.19	0.57	0.13	2.13	0.28	
$u(x_i)$ (nmol/mol)	0.44	1.21	0.53	1.52	1.17	2.29	0.74	
$U(x_i)$ (nmol/mol)	0.88	2.42	1.07	3.03	2.35	7.29	1.48	

Table 32: Reported values for NO<sub>2</sub> run 4.




Figure 35: Reported values for NO<sub>2</sub> run 4.

### **Annex C. The precision of standardized measurement methods**

For the main purpose of monitoring trends between different IE the precision of standardized SO<sub>2</sub>, CO, O<sub>3</sub> and NO<sub>x</sub> measurement methods [2], [3], [4] and [5] as implemented by NRLs was evaluated. Applied methodology is described in ISO 5725-1, -2 and -6 [14], [15] and [16].

The precision experiment has involved a total of 7 laboratories the actual number of labs (p<sub>j</sub>) varying from run to run (Table 33). Laboratory G didn't reported results for CO. For run 0 was requested only one value so repeatability cannot be evaluated. Five concentration levels were tested for CO, four levels for O<sub>3</sub>, SO<sub>2</sub> and NO<sub>2</sub>, and two for NO. Outlier tests were performed and results are reported in Annex D.

The repeatability standard deviation (s<sub>r</sub>) was calculated in accordance with ISO 5725-2 as the square root of average within laboratory variance. The repeatability limit (r) is calculated using Equation 5 [16]. It represents the biggest difference between two test results found on an identical test gas by one laboratory using the same apparatus within the shortest feasible time interval, that should not be exceeded on average more than once in 20 cases in the normal and correct operation of method.

$$r = t_{95\%,v} \cdot \sqrt{2} \cdot s_r \quad \text{Equation 5}$$

The reproducibility standard deviation (s<sub>R</sub>) was calculated in accordance with ISO 5725-2 as the square root of sum of repeatability and between laboratory variance. The reproducibility limit (R) is calculated using Equation 6 [16]. It represents the biggest difference between two measurements on an identical test gas reported by two laboratories, which should not occur on average more than once in 20 cases in the normal and correct operation of method.

$$R = t_{95\%,v} \cdot \sqrt{2} \cdot s_R \quad \text{Equation 6}$$

The repeatability standard deviation was evaluated with (p<sub>j</sub>\*(3-1)) degrees of freedom (v) and reproducibility standard deviation with (p<sub>j</sub>-1) degrees of freedom. The critical range student factors (t<sub>α,v</sub>) are reported in Table 33.

parameter	run	p <sub>j</sub>	t critical value 95% for r	t critical value 95% for R
CO	1,2,3,4,5	6	2.179	2.571
NO	1,2	7	2.145	2.447
NO <sub>2</sub>	1,2,3,4	7	2.145	2.447
O <sub>3</sub>	1,2,3,4	7	2.145	2.447
SO <sub>2</sub>	1,2,3,4	7	2.145	2.447

**Table 33: Critical values of t used in the repeatability (r) and reproducibility (R) evaluation.**

The repeatability (r) and reproducibility (R) limits of measurement methods are presented from Table 34 to Table 38 and from Figure 36 to Figure 40. It is also reported the 'reproducibility from common criteria (R (from  $\sigma_p$ ))' calculated by substituting  $s_R$  in Equation 6 with a 'standard deviation for proficiency assessment' (Table 4). Comparison between R and R (from  $\sigma_p$ ) serves to indicate that  $\sigma_p$  is realistic ([13] par. 6.3.1) or from the other point of view, that the general methodology implemented by NRLs is appropriate for  $\sigma_p$ .

SO <sub>2</sub> data (nmol/mol) without outliers			
group average	repeatability limit : r	reproducibility limit : R	reproducibility limit (relative)
0.1		1.6	11.5%
5.2	0.8	2.2	
20.2	3.5	4.7	
45.2	4.4	7.2	
130.5	8.1	15.0	

Table 34: The R and r of SO<sub>2</sub> standard measurement method.




Figure 36: The R and r of SO<sub>2</sub> standard measurement method as a function of concentration.



CO data (µmol/mol) without outliers			
group average	repeatability limit : r	reproducibility limit : R	reproducibility limit (relative)
0.012		0.081	
1.019	0.044	0.202	
2.985	0.009	0.395	
4.486	0.055	0.699	
5.904	0.042	0.326	
7.964	0.085	0.873	11.0%

Table 35: The R and r of CO standard measurement method.




Figure 37: The R and r of CO standard measurement method as a function of concentration.

O <sub>3</sub> data (nmol/mol) without outliers			
group average	repeatability limit : r	reproducibility limit : R	reproducibility limit (relative)
0.1		0.6	8.5%
19.8	2.8	3.4	
60.5	6.7	8.1	
99.7	8.0	15.9	
299.9	19.5	25.6	

Table 36: The R and r of O<sub>3</sub> standard measurement method.




Figure 38: The R and r of O<sub>3</sub> standard measurement method as a function of concentration.

NO data (nmol/mol) without outliers			
group average	repeatability limit : r	reproducibility limit : R	reproducibility limit (relative)
0.1		0.7	9.9%
19.6	2.6	4.6	
201.2	16.7	20.0	

Table 37: The R and r of NO standard measurement method.




Figure 39: The R and r of NO standard measurement method as a function of concentration.

NO <sub>2</sub>			
group average	repeatability limit : r	reproducibility limit : R	reproducibility limit (relative)
0.02		0.94	10.3%
13.41	0.22	2.13	
20.22	0.16	2.51	
58.96	0.29	6.36	
99.78	0.61	11.60	
119.43	0.87	12.25	

Table 38: The R and r of NO<sub>2</sub> standard measurement method.




Figure 40: The R and r of NO<sub>2</sub> standard measurement method as a function of concentration.

### ***Annex D. The scrutiny of results for consistency and outlier test***

The precision evaluation (Annex C) focuses on data that are as much as possible the reflection of every day work of NRLs and thus represents the comparability of participant's standard operating procedures.

For that reason a procedure for the detection of exceptional errors (error during typing, slip in performing the measurement or the calculation, wrong averaging interval, malfunction of instrumentation, etc.) was applied.

In this procedure were carried out tests for data consistency and statistical outliers as described in ISO 5725-2.

Laboratories showing some form of statistical inconsistency were requested to investigate the cause of discrepancies. Laboratories were allowed to correct their results in case of identification of exceptional errors. Subsequently, data were considered definitive and "Grubb's one outlying observation test" was performed. If detected outliers were removed and "Grubb's one outlying observation test" was repeated until no more outliers were observed.

During this IE the statistical outliers presented in the table below are related only to zero levels:

<b>parameter</b>	<b>run</b>	<b>laboratory</b>	<b>measured value</b>	<b>failing test</b>	<b>confidence level</b>
NO	0	F	0.56	G1 maximum	1%, 5%
CO	0	C	0.056	G1 maximum	1%, 5%

**Table 39: "Genuine" statistical outliers according to Grubb's one outlying observation test.**

The precision of standardized measurement methods reported in Annex C are calculated using the database without outliers.

European Commission  
EUR 25387 – Joint Research Centre – Institute for Environment and Sustainability

Title: Evaluation of the Laboratory Comparison Exercise for SO<sub>2</sub>, CO, O<sub>3</sub>, NO and NO<sub>2</sub>, Langen  
23<sup>rd</sup>-28<sup>th</sup> October 2011

Author(s): Maurizio Barbieri, Volker Stummer, Friedrich Lagler, Hans-Guido Muecke

Luxembourg: Publications Office of the European Union

2012 – 56 pp. – 21.0 x 29.7 cm

EUR – Scientific and Technical Research series - ISSN 1831-9424

ISBN 978-92-79-25367-6

doi:10.2788/33334

### **Abstract**

From the 23<sup>rd</sup> to the 28<sup>th</sup> of October 2011 in Langen (DE), 7 Laboratories of WHO/AQUILA (Network of European Air Quality Reference Laboratories) met at an laboratory comparison exercise to evaluate their proficiency in the analysis of inorganic gaseous pollutants covered by European Directive about air quality (SO<sub>2</sub>, CO, NO, NO<sub>2</sub> and O<sub>3</sub>).

The proficiency evaluation, where each participant's bias was compared to two criteria, provides information on the current situation and capabilities to the European Commission and can be used by participants in their quality control system.

On the basis of criteria imposed by the European Commission, 59.4% of the results reported by the laboratories were good both in terms of measured values and reported uncertainties. Another 39.9% of the results had good measured values, but the reported uncertainties were too high and only 0.7% delivered questionable results.

Comparability of results among participants at the highest concentration level, excluding outliers, is acceptable for CO and NO measurements while SO<sub>2</sub>, O<sub>3</sub> and NO<sub>2</sub> measurements showed less satisfactory results.

As the Commission's in-house science service, the Joint Research Centre's mission is to provide EU policies with independent, evidence-based scientific and technical support throughout the whole policy cycle.

Working in close cooperation with policy Directorates-General, the JRC addresses key societal challenges while stimulating innovation through developing new standards, methods and tools, and sharing and transferring its know-how to the Member States and international community.

Key policy areas include: environment and climate change; energy and transport; agriculture and food security; health and consumer protection; information society and digital agenda; safety and security including nuclear; all supported through a cross-cutting and multi-disciplinary approach.



Publications Office

ISBN 978-92-79-25367-6

