

Automated Electrowetting-Based Digital Microfluidic Chips for Next-Generation Sequencing and Sample Preparation

SELECT**BIO**

Faruk Azam, Shubha Jain, Dr. H. N. Unni Department of Biomedical, IIT Hyderabad

<u>Introduction</u>

Next-generation sequencing (NGS) technology is a prominent tool for Identifying and characterizing unknown pathogens, but its utility in high throughput biodefense and public health applications is currently limited by the lack of fast, efficient, and reliable automated DNA sample preparation methods like automated multiplexing techniques which are used for switching the voltage of electrodes, deciding the threshold voltage, for clinical diagnostic to protect the cell from damage and avoiding the cross talk between electrodes because of electrostatic effect to maintain droplet on proper track. We would like to address those limitations to enhance the system performance for digital microfluidic (DMF) platform to function as a fluid distribution hub.

Computational Methods

The Comsol Multiphysics platform has been used to achieve effective transport of droplets on DMF platform using the Electrowetting on dielectric (EWOD) principle. In particular we use Laminar Two-Phase Flow, Moving Mesh (tpfmm) for simulation of Lippmann-Young equation i.e change in contact angle and Laminar Two Phase Flow, Level Set (tpf) for simulation of 2D droplet transport.


Lippmann-Young equation

$$cos\theta(V) = cos\theta(0) + \frac{\varepsilon_0 \varepsilon}{2\gamma_{LG} d} V^2$$


Brochard's model Eq

$$U = \frac{\varepsilon_0 \varepsilon (1 - \cos \theta(V))}{6 \mu dl \sin \theta(V)} V^2$$


Geometry and Principle



Three phase interphase line of solid, liquid and gas.



Electrowetting Principle




Diagrammatic view of electrodes switching


Time t1 t2 t3 t4 t5


Electrodes


Result

We demonstrated droplet transport with angle modification on four electrodes single row platform only by 15V DC.









Conclusion

Using the information gathered from the above simulation, by integrating the concept of microfluidic mixer with magnetic beads, make feasible to study the next-generation sequencing applications like genomic DNA sequencing of microbial species from single cell and protein RNA interaction.

References

- 1. H. Kim et.al, *PLoS ONE*, vol. 8, no. 7, Article ID e68988, 2013.
- 2. Kim H et.al, *J Lab Autom*. 2011; 16: 405–414.
- 3. K. Chakrabarty, IEEE International conference on computer design, San Josh, CA, Oct. 1-4, 2007, 93-100.
- 4. M. Schneider et.al, *ChemPhysChem*, vol. 9, no. 4, pp. 641–645, Mar. 2008.
- 5. W. H. Huang et.al, Anal. Technol. Biomed. Life Sci., vol. 866, no. 1/2, pp. 104–122, Apr. 2008.

Advances in Next Generation Sequencing