

Technical University of Denmark



The Fge Magnetometer and the Intermagnet 1 Second Standard

Pedersen, Lars William; Merenyi, Laszlo

Publication date:
2012

Document Version
Peer reviewed version

[Link back to DTU Orbit](#)

Citation (APA):

Pedersen, L. W., & Merenyi, L. (2012). The Fge Magnetometer and the Intermagnet 1 Second Standard. Abstract from XVIth IAGA Workshop on Geomagnetic Observatory Instruments Data Acquisition and Processing , Hyderabad, India.

DTU Library Technical Information Center of Denmark

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

THE FGE MAGNETOMETER AND THE INTERMAGNET 1 SECOND STANDARD

PEDERSEN, Lars William⁽¹⁾, Merenyi, Laszlo⁽²⁾

⁽¹⁾ DTU Space, Denmark, lawp@space.dtu.dk

⁽²⁾ MFGI, Hungary, merenyi.laszlo@mfgi.hu

During the last 2 years, the Danish Space Center, DTU Space, has tested and made new features to the well known FGE 3-axis Fluxgate Magnetometer. Due to the new INTERMAGNET standard for 1 second data, we have tested the magnetometer to prove that it fulfills the demands. We will show that the instrument can have a very good linear frequency response up to 20 Hz, depending on the used lowpass filter and gain. Also the delay in the signal is well defined and can be taken into account in the datalogger system, not only in new FGE systems but in most FGE systems placed in observatories round the world.

The noise of the used fluxgate sensors seems to be the only point that can not recently fulfill the new standard, but we will show how close the noise is to the specified limit.

During the tests we have used a suitable digital filter that fulfills the new demands for the frequency response: the filter is linear from DC to 0.2 Hz and has at least of 60 dB attenuation at 0.5 Hz and above. With the MagRec-4B and ObsDaq ADC data acquisition system, operating with 128 Hz sampling rate and GPS time synchronization, the time stamp accuracy can be better than 0.01 sec.

We have also developed a new electronic board for differential output of the three components X, Y and Z, and with this board it is possible to upgrade older FGE systems to use fast seismic dataloggers or similar instruments (like Magrec-4B/ObsDaq datalogger system) in parallel with old datalogger systems. So after such an upgrade the observatory's original FGE magnetic data supporting very good baseline stability can still be acquired together with new data, supporting better noise characteristics, frequency response and timing accuracy.

We will also show how the INTERMAGNET Square Wave System can be used to test the frequency response and to find the delay in the whole system, from sensor to data acquisition system.

⁽¹⁾ Danish National Space Center, DTU Space
Elektrovej, Building 327,
DK-2800, Kgs. Lyngby, Denmark

⁽²⁾ Geological and Geophysical Institute of Hungary
Stefania ut 14.,
H-1143, Budapest, Hungary

Session: Observatory Instruments and Techniques, Oral

Formatted: Danish

Formatted: Danish

Formatted: Danish

Field Code Changed

Formatted: Danish

Field Code Changed

Formatted: Danish

Formatted: Danish