

## Integration of agglutination assay for protein detection in microfluidic disc using Blu-ray optical pickup unit and optical fluid scanning - DTU Orbit (08/11/2017)

### Integration of agglutination assay for protein detection in microfluidic disc using Blu-ray optical pickup unit and optical fluid scanning

We present a novel strategy for thrombin detection by combining a magnetic bead based agglutination assay and low-cost microfluidic disc. The detection method is based on an optomagnetic readout system implemented using a Blu-ray optical pickup unit (OPU) as main optoelectronic component. The assay, from sample to answer, is fully integrated on a microfluidic disc which embeds on-disc mixing ensuring full automation of the assay along with less sample-to-answer time compared to similar methods. Moreover, we compare the optomagnetic readout to the cluster size distribution determined using a commercial optical scanning imaging instrument.

#### General information

State: Published

Organisations: Department of Micro- and Nanotechnology, Nanoprobes, Magnetic Systems

Authors: Uddin, R. (Intern), Burger, R. (Intern), Donolato, M. (Intern), Fock, J. (Intern), Creagh, M. (Intern), Hansen, M. F. (Intern), Boisen, A. (Intern)

Number of pages: 3

Pages: 1807-1809

Publication date: 2015

#### Host publication information

Title of host publication: Proceedings of MicroTAS 2015

Main Research Area: Technical/natural sciences

Conference: MicroTAS 2015, Gyeongju, Korea, Republic of, 25/10/2015 - 25/10/2015

Agglutination assay, Aptamer, Magnetic beads, Centrifugal Microfluidics, Thrombin, Microfluidic disc

Electronic versions:

Integration\_of\_agglutination\_assay.pdf

Source: PublicationPreSubmission

Source-ID: 118520731

Publication: Research - peer-review › Article in proceedings – Annual report year: 2015