

# Scanning SQUID microscopy in a cryogen-free dilution refrigerator

Cite as: Rev. Sci. Instrum. **92**, 083704 (2021); <https://doi.org/10.1063/5.0047652>

Submitted: 15 February 2021 . Accepted: 09 August 2021 . Published Online: 24 August 2021

 D. Low, G. M. Ferguson, Alexander Jarjour, et al.



View Online



Export Citation



CrossMark

## ARTICLES YOU MAY BE INTERESTED IN

[New design yields robust large-area framing camera](#)






Review of Scientific Instruments **92**, 083103 (2021); <https://doi.org/10.1063/5.0049110>

[An ultra-stable 1.5#T permanent magnet assembly for qubit experiments at cryogenic temperatures](#)

Review of Scientific Instruments **92**, 085106 (2021); <https://doi.org/10.1063/5.0055318>

[On the stability of thoriated tungsten cathodes in strong magnetic fields](#)

Review of Scientific Instruments **92**, 083510 (2021); <https://doi.org/10.1063/5.0049430>

	<p>Nanopositioning Systems</p> 	<p>Modular Motion Control</p> 	<p>AFM and NSOM Instruments</p> 	<p>Single Molecule Microscopes</p> 
---	--	--	---	--

# Scanning SQUID microscopy in a cryogen-free dilution refrigerator

Cite as: Rev. Sci. Instrum. 92, 083704 (2021); doi: 10.1063/5.0047652

Submitted: 15 February 2021 • Accepted: 9 August 2021 •

Published Online: 24 August 2021



View Online



Export Citation



CrossMark

D. Low,<sup>1,a)</sup>  G. M. Ferguson,<sup>1</sup> Alexander Jarjour,<sup>1</sup> Brian T. Schaefer,<sup>1</sup>  Maja D. Bachmann,<sup>2,3</sup>   
Philip J. W. Moll,<sup>4</sup> and Katja C. Nowack<sup>1,5,a)</sup> 

## AFFILIATIONS

<sup>1</sup>Laboratory of Atomic and Solid-State Physics, Cornell University, Ithaca, New York 14853, USA

<sup>2</sup>Max Planck Institute for Chemical Physics of Solids, D-01187 Dresden, Germany

<sup>3</sup>School of Physics and Astronomy, University of St. Andrews, St. Andrews KY16 9SS, United Kingdom

<sup>4</sup>Laboratory of Quantum Materials (QMAT), Institute of Materials, École Polytechnique Fédéral de Lausanne (EPFL), 1015 Lausanne, Switzerland

<sup>5</sup>Kavli Institute at Cornell for Nanoscale Science, Cornell University, Ithaca, New York 14853, USA

<sup>a)</sup>Authors to whom correspondence should be addressed: [dhl88@cornell.edu](mailto:dhl88@cornell.edu) and [kcn34@cornell.edu](mailto:kcn34@cornell.edu)

## ABSTRACT

We report a scanning superconducting quantum interference device (SQUID) microscope in a cryogen-free dilution refrigerator with a base temperature at the sample stage of at least 30 mK. The microscope is rigidly mounted to the mixing chamber plate to optimize thermal anchoring of the sample. The microscope housing fits into the bore of a superconducting vector magnet, and our design accommodates a large number of wires connecting the sample and sensor. Through a combination of vibration isolation in the cryostat and a rigid microscope housing, we achieve relative vibrations between the SQUID and the sample that allow us to image with micrometer resolution over a 150  $\mu\text{m}$  range while the sample stage temperature remains at base temperature. To demonstrate the capabilities of our system, we show images acquired simultaneously of the static magnetic field, magnetic susceptibility, and magnetic fields produced by a current above a superconducting micrometer-scale device.

© 2021 Author(s). All article content, except where otherwise noted, is licensed under a Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>). <https://doi.org/10.1063/5.0047652>

## I. INTRODUCTION

Superconducting quantum interference devices (SQUIDs) are among the most sensitive magnetic sensors available and have been widely used for magnetic imaging at cryogenic temperatures. Scanning SQUID microscopy can be used to image static stray magnetic fields above a sample, to measure the local magnetic response, and to image the magnetic field produced by applied and spontaneous currents in a device. This technique has been applied to a wide variety of quantum materials and mesoscopic devices, including unconventional superconductors,<sup>1</sup> topological insulators,<sup>2,3</sup> complex oxides,<sup>4–8</sup> superconducting and normal metal rings,<sup>9–11</sup> unconventional Josephson junctions,<sup>12,13</sup> and a variety of graphene structures.<sup>14–17</sup> Some physical phenomena require cooling to sub-Kelvin temperatures, which can be achieved in a dilution refrigerator (DR). In a traditional DR, an inner vacuum chamber is immersed in a

liquid <sup>4</sup>He bath. Within the vacuum chamber, a smaller bath of liquid <sup>4</sup>He is pumped to lower its temperature to ~1.5 K. This bath pre-cools a circulating <sup>4</sup>He/<sup>3</sup>He mixture that ultimately provides continuous cooling to a few millikelvins. In a cryogen-free DR, all liquid <sup>4</sup>He baths are replaced by a cryocooler, often a pulse tube, which provides cooling to ~3 K. Closed-cycle circulation of a <sup>4</sup>He/<sup>3</sup>He mixture is still used in these systems to achieve temperatures of a few millikelvins at the mixing chamber.

A cryogen-free DR is attractive due to rising prices and uncertain supply of liquid helium. In addition, no <sup>4</sup>He transfers are necessary, which can interrupt measurements and prevent running experiments remotely over extended periods of time. Cryogen-free DRs are often built with a large sample volume as they do not require a <sup>4</sup>He bath to envelop the inner vacuum chamber. However, vibrations from the cryocooler put cryogen-free DRs at a major disadvantage. During the pulse tube cooling cycle, high-pressure helium gas is


pushed in and out of the pulse tube at a frequency of  $\sim 1$  Hz, rendering scanning probe microscopy in a cryogen-free DR challenging. Several scanning probe microscopes operating at few Kelvin temperatures have been reported, including a tuning fork atomic force microscope,<sup>18</sup> a nitrogen vacancy center microscope,<sup>19</sup> and scanning SQUID microscopes.<sup>20,21</sup> To date, several few Kelvin cryogen-free platforms with low vibrations are commercially available that are compatible with scanning probe microscopy (e.g., from Advanced Research Systems, attocube, RHK Technology, Montana Instruments). However, only a few scanning probe microscopes have been reported that operate in cryogen-free DRs, including a scanning gate microscope<sup>22</sup> and a scanning tunneling microscope.<sup>23</sup> In both cases, custom spring stages were used to mechanically isolate the microscope from the cryostat.

Specifically, scanning SQUID microscopy has been implemented in two different types of cryogen-free cryostats with base temperatures of  $\sim 3$ – $4$  K, but not in a cryostat reaching lower temperatures. In Ref. 21, the authors used a cryostat with built-in mechanical isolation between the cryocooler and the cold plate on which the microscope is mounted. In Ref. 20, the microscope was placed in a pulse tube-based cryostat similar to our own but without the dilution unit and associated cold plates. A custom spring stage was used to significantly reduce vibrations in the microscope. However, such mechanical isolation often comes at the cost of reduced thermal anchoring: a less rigid mechanical connection implies less thermal conductivity between the microscope and the cold plate. A weak thermal connection can prevent rapid measurements over a large scan window due to the heat piezoelectric elements generate while moving. This is particularly challenging when operating near the base temperature in a DR due to the limited cooling power. In addition, a spring stage is difficult to combine with mechanically rigid low-loss coaxial connections to either the SQUID or the sample. Such connections are necessary to implement a dispersive SQUID readout<sup>24–26</sup> or to deliver gigahertz and fast rise time excitations to a sample. Combining a spring stage with a superconducting magnet is challenging as well, since a strong magnetic field may exert force on the microscope and cause undesirable motion.<sup>22</sup> In principle, scanning probe microscopes are only affected by relative motion of the probe and the sample. This suggests that a rigid microscope, in which the probe and the sample move together in response to the pulse-tube-induced vibrations, offers an alternative to using spring stages.

Here, we report the implementation of a scanning SQUID microscope in a cryogen-free DR with a base temperature of  $\sim 10$  mK. The microscope is operated in the bore of a superconducting vector magnet. We avoid the use of a spring stage in order to optimize for low sample temperatures. To achieve an acceptable level of vibrations, we designed the microscope prioritizing rigidity while still maintaining a scan range of  $150 \times 150 \times 100 \mu\text{m}^3$  and  $6 \times 6 \times 6 \text{ mm}^3$  coarse positioning range.

## II. DESCRIPTION OF THE SCANNING SQUID MICROSCOPE

Figure 1 shows drawings of the scanning SQUID microscope mounted in the DR. Computer-aided design drawings of the microscope as shown in panels (b) and (c) are available in Ref. 27. We use a Bluefors BF-LD 400 DR with a base temperature of 10 mK and



**FIG. 1.** Drawings of the experimental setup and the scanning probe microscope. (a) Cross-sectional view of a dilution refrigerator with a scanning probe microscope. The room temperature vacuum can is shown. Light tight cans at 60, 4, and 1 K and various cross beams on both frames are omitted for clarity. (b) Cold finger and microscope housing that attach to the mixing chamber plate. (c) Microscope housing cutaway showing coarse positioners, the piezoelectric scanner, and the sample mount. The piezoelectric benders are shown in dark gray. The sample (shown in purple) is mounted face-down on the sample mount. The SQUID (shown in blue) is mounted on a brass cantilever glued to a printed circuit board shown in panel (e) and attached (screws not shown) to the horizontal piezoelectric bender providing vertical motion. (d) Close up view of the SQUID and the sample.

400  $\mu\text{W}$  cooling power at 100 mK. Cooling to  $\sim 3$  K is provided by a Cryomech PT-415 pulse tube cooler implemented with a remote motor option. A superconducting vector magnet from American Magnetics, Inc. can apply 6 T perpendicular to and 1 T in any in-plane direction of the sample. A few modifications designed and implemented by Bluefors reduce the vibrations of the mixing chamber plate. The DR rests on two nested aluminum frames to reduce the mechanical coupling between the pulse tube cooler and the interior of the cryostat. All cold plates, vacuum cans, and the superconducting magnet are supported by the inner frame, while the pulse tube and related components are mounted onto the outer frame. The two nested frames are only connected via an edge-welded stainless

steel bellows to maintain vacuum and via copper braids inside the vacuum chamber to thermally link the cold plates and the pulse tube. The top plate of the cryostat is bolted to heavy metal plates ( $\sim 400$  kg). These metal plates rest on air springs mounted to the inner frame. The remote motor head of the pulse tube, which is a source of vibrations, is housed in a sound isolation box that is mounted on the outer frame and connected to the pulse tube by high- and low-pressure flexible lines. We implemented a linear motor driver for the cold head motor, as suggested in Ref. 22. The high-pressure lines connecting the remote motor head and the compressor are run through thick fiberglass sleeving to reduce the acoustic noise in the lab. The still pump line is passed through a concrete block between the cryostat and the gas handling cabinet. The gas handling cabinet and the compressor are in a utility space that is well separated from the lab space.

The microscope consists of a cold finger and the microscope housing. Each is machined from a single block of copper and bolted together on machined flat, keyed surfaces using four brass screws. The cold finger is mounted to the mixing chamber plate using six brass screws. The outer diameter of the housing is set by the 68 mm diameter bore of the superconducting magnet. The length of the cold finger is chosen such that the sample is positioned at the center of the magnet. Windows in the microscope housing [see Fig. 1(b)] can be opened by removing copper plates to get visual access when aligning the SQUID to the sample at room temperature.

All wiring runs along the cold finger and enters the microscope housing through a box at the bottom of the cold finger. This box has two removable plates, which allows us to include and change desired wiring feedthroughs. A light tight gold-plated copper can (not shown) is bolted below the wiring feedthrough box on the cold finger. This design allows us to, in principle, route fully filtered and shielded wiring to the microscope. All machined parts are gold-plated.

The SQUID is mounted on a home-built piezoelectric scanner similar to the ones described in Refs. 20 and 21, which in turn is mounted on a three-axis stack of attocube coarse positioners [see Fig. 1(c)]. We chose bearing-based attocube ANPx311 positioners with 6 mm coarse range. Two machined brackets adapt a high-load ANPx311 positioner to act as a  $z$  positioner. The attocube positioners are made of grade 2 (99% pure) titanium. The scanner is assembled from piezoelectric bimorphs and machined Macor parts.<sup>28,29</sup> Two pairs of “s-benders” move the SQUID in a horizontal plane, while an additional cantilever piezo allows for vertical motion. Imaging range and stiffness compete in the scanner design, since both depend on the dimensions of the benders. Here, we chose  $x, y$  “s-benders” with lateral dimensions  $38.1 \times 6.4$  mm<sup>2</sup> and 0.13 mm thickness. The  $z$  bender has  $25.4 \times 6.35$  mm<sup>2</sup> lateral dimensions and is 0.19 mm thick. The benders are joined by Macor blocks using EPO-TEK H70e epoxy. The scanner is attached to a top Macor plate as shown in Fig. 1. This plate is mounted with spring-loaded screws to a bottom Macor plate, which in turn is rigidly mounted on the coarse positioners. This arrangement allows us to adjust the alignment angles between the SQUID and the sample through adjusting the spring-loaded screws. Our scan range at cryogenic temperatures is  $\sim 150$   $\mu\text{m}$  in the  $x, y$  directions and 110  $\mu\text{m}$  in the  $z$ -direction. Typical scan rates during imaging are between 1 and 20  $\mu\text{m}/\text{s}$ . The scanner and attocube stack are mounted on a copper puck that slides out at the bottom of the microscope housing, allowing for easy


SQUID replacement. The copper puck is threaded and bolted on two adjacent sides to the microscope housing with four brass screws on each side. To detect when the SQUID touches down on the sample, the SQUID is mounted on a flexible brass cantilever. We monitor the capacitance between this cantilever and a ground plane on the printed circuit board. At touchdown, a sharp increase in the capacitance is detected. These cantilevers are typically 5 mm long, 2 mm wide, and 50  $\mu\text{m}$  thick.

The sample is mounted upside-down on a puck that slides into the top of the microscope housing to allow for easy sample replacement and is fastened using four brass screws. A ruthenium oxide thermometer is mounted on the sample mount. This thermometer is factory calibrated to 50 mK and we extrapolate its calibration curve to obtain approximate temperature values below 50 mK. The mixing chamber plate reaches a base temperature of  $\sim 10$  mK. The sample mount and mixing chamber plate temperature track each other down to 30 mK. Below 30 mK, the resistance of the sample mount thermometer saturates. This is likely due to self-heating of the thermometer rather than saturation of the sample mount temperature.<sup>30</sup> This is consistent with the observation of changes in the magnetic response in some samples below the saturation point of the thermometer as the mixing chamber cools further. A more detailed and quantitative analysis will be reported elsewhere. Finally, we note that the electron temperature of mesoscale devices, including electrical contacts for transport measurements, can differ from the temperature of the sample mount, since for such samples, heating and cooling is dependent on the filtering and thermal anchoring of the electrical connections to the sample. Importantly, no appreciable heating is observed during scanning at a rate of  $\sim 20$   $\mu\text{m}/\text{s}$  even at base temperature.

Three wire bundles with 12 twisted pairs are connected to the microscope [two bundles with 36 American Wire Gauge (AWG) phosphor bronze, and one bundle with 36 AWG Cu and NbTi/CuN wires for low resistance connections]. Two of these are used to operate the piezoelectric scanner, the attocube coarse positioners, and the SQUID. One bundle is available for sample connections. In addition, the cryostat includes rigid and semi-rigid coaxial wiring to the mixing chamber plate. The combination of 24 wires available to connect to the sample and large coarse positioning range allows us to image devices with several contacts and electrostatic gates, perform full transport characterization of these devices, and image multiple devices in a single cooldown.

Figure 2 shows an example of measurements on a CeIrIn<sub>5</sub> microstructure studied in Ref. 31 taken with the scanning SQUID microscope described here. We used a SQUID with a  $\sim 1.5$   $\mu\text{m}$  sensitive area (pickup loop) and a  $\sim 6$   $\mu\text{m}$  on-chip field coil, which enables local magnetic susceptibility measurements. Applying a current to the field coil applies a small magnetic field to the sample. The SQUID has a gradiometric design such that the current in the field coil only couples a minimal amount of flux directly into the SQUID.<sup>32</sup> This amount can be calibrated with the SQUID retracted from the sample. A finite magnetic response by the sample to the field applied with the field coil modifies the flux in the SQUID by an amount that is proportional to magnetic susceptibility of the sample. The device was fabricated from a bulk single crystal of CeIrIn<sub>5</sub> using a focused ion beam (FIB). A lamella is cut out from the crystal [outline of the lamella is shown in Fig. 2(a) as white dashed line], placed on a substrate, and contacted with a gold layer. The gold is removed





**FIG. 2.** Scanning SQUID images of a superconducting  $\text{CeIrIn}_5$  microstructure fabricated using a focused ion beam. (a) Scanning electron microscope (SEM) image of the device with contacts 1–4. (b) Image of the stray magnetic field above the device. The image shows a combination of Meissner screening (red) of a small background field and vortices (blue) in the device. (c) Image of the magnetic field produced by a total current of  $\sim 56 \mu\text{A}$  flowing from contact 1 to 3. (d) Image of the magnetic susceptibility. Bright regions are strongly diamagnetic and therefore superconducting, whereas dark regions are non-superconducting. Images were acquired simultaneously at 225 mK. All scale bars are  $25 \mu\text{m}$ . Part of the data was published in Ref. 31.


in the active area of the device, and four electrical contacts are separated by cutting trenches (black in the image) using a FIB again (see Ref. 31 for details). The images were taken at 225 mK at which all parts of the  $\text{CeIrIn}_5$  structure were superconducting. The images shown in Figs. 2(b)–2(d) showcase different imaging modes: imaging static stray magnetic fields [Fig. 2(b)], imaging the magnetic field produced by an AC current applied to the device [Fig. 2(c)], and imaging the local magnetic susceptibility [Fig. 2(d)]. The static magnetic image [Fig. 2(b)] shows a combination of Meissner screening of a small background field and vortices in the structure. The current was applied from contact 1 to 3. The local magnetic susceptibility shows a strong diamagnetic response above the superconducting structure. All images shown in Figs. 2(b)–2(d) were acquired simultaneously with the sample mount temperature at 225 mK. The magnetic properties of the microstructure do not significantly change below this temperature.

### III. CHARACTERIZATION OF THE RELATIVE MOTION BETWEEN THE SQUID AND THE SAMPLE


The pulse tube causes vibrations in the cryostat. These only affect the imaging if they cause a relative motion between the SQUID and the sample. We characterize this relative motion by following a method reported in Ref. 33 based on analyzing the excess flux noise caused by vibrations in areas with large magnetic field gradients.

Figure 3(a) shows a vortex in a microstructure fabricated from  $\text{Sr}_2\text{RuO}_4$ . The image of the vortex has a tear-drop shape, which is due to the shape of the SQUID's sensitive area.<sup>32,34</sup> The sensitive area is formed by a circular loop that connects with two leads to the main body of the SQUID. Magnetic flux couples into a small unshielded opening between the leads which causes a deviation from a perfectly circular shape of the sensitive area. We obtain the magnetic flux gradients in the  $(x, y)$  direction [Figs. 3(b) and 3(c)] by numerically differentiating Fig. 3(a). To obtain the flux gradient in the  $z$  direction [Fig. 3(d)], we took a second measurement at  $\Delta z = 0.6 \mu\text{m}$  higher than Fig. 3(a). Relative motion between the SQUID and the sample causes noise in the measured flux signal that depends on the strength and direction of the gradient and the magnitude and direction of the motion. The intrinsic flux noise of the SQUID has a white noise floor on the order of  $1 \mu\Phi_0/\sqrt{\text{Hz}}$  and a  $1/f$  tail below 100 Hz. Given the magnitude of the flux gradients of  $\sim 10 \mu\Phi_0/\mu\text{m}$ , the flux noise of  $1 \mu\Phi_0/\sqrt{\text{Hz}}$  corresponds to an approximate sensitivity of  $0.1 \text{ nm}/\sqrt{\text{Hz}}$  for detecting relative motion between the SQUID and the sample. Each pixel in Fig. 3(a) is the average of a four-second time trace. The Fourier transforms of each time trace provide a position-dependent flux noise spectral density. Examples of spatial maps of the noise spectral density at a few frequencies are shown in the left panels of Fig. 4.

To find the power spectral density of the vibrations, we model the flux power spectral density at each pixel  $(i, j)$  and frequency  $f$  as a sum of vibration-induced noise and intrinsic SQUID and electrical



**FIG. 3.** (a) Image of a superconducting vortex in a FIB defined microstructure. (b)–(d) Flux gradients in the (b)  $x$ , (c)  $y$ , and (d)  $z$  directions. The  $x$  and  $y$  gradients are obtained from numerically differentiating (a), whereas the  $z$  gradient is obtained from measurements taken at two heights separated by  $\Delta z = 0.6 \mu\text{m}$ . The tear-drop shape is due to the shape of the SQUID's sensitive area.



**FIG. 4.** Spatial maps of the flux noise spectral density (left panel) at (a) 8.5 Hz, (b) 241 Hz, (c) 316 Hz, and (d) 480 Hz and corresponding fits to the model in Eq. (1) (right panel). Vibrations along the  $(x, y, z)$  directions are extracted from the fits.

noise,<sup>20</sup>

$$[\hat{\Phi}(f)]_{ij}^2 = \left( \left[ \frac{\partial \Phi}{\partial x} \right]_{ij} \hat{X}(f) + \left[ \frac{\partial \Phi}{\partial y} \right]_{ij} \hat{Y}(f) + \left[ \frac{\partial \Phi}{\partial z} \right]_{ij} \hat{Z}(f) \right)^2 + (\hat{N}(f))^2. \quad (1)$$


Here,  $\hat{\Phi}(f)$  denotes the flux noise amplitude spectral density,  $\partial\Phi/\partial(x, y, z)$  are the flux gradients shown in Fig. 3,  $\hat{X}(f)$ ,  $\hat{Y}(f)$ ,  $\hat{Z}(f)$  are the vibrations along the  $(x, y, z)$  direction, and  $\hat{N}(f)$  models intrinsic SQUID noise and is strictly positive. In applying this model, we assume that the vibrations are independent of the motion of the benders. This is reasonable because the typical deflection of the benders is small compared to their size. The mechanical properties are therefore not expected to change significantly within the scan range. This assumption is also supported by the absence of significant scan distortions even at the extremes of our scan range. We fit the spatial map of the noise to Eq. (1) to obtain the spatial

vibrations  $\hat{X}$ ,  $\hat{Y}$ ,  $\hat{Z}$  and electrical noise  $\hat{N}$  at each frequency. We use statistical bootstrapping with 200 trials to avoid local optima and determine the 95% confidence intervals of the fits.<sup>35</sup>

In Figs. 4(a)–4(d), we show examples of maps of flux noise spectral density (left panels) and the respective fits to the model in Eq. (1) (right panels). Figures 4(b)–4(d) closely resemble the flux gradients in the  $(y, z, x)$  directions, respectively, as shown in Fig. 3. This suggests that the vibrations at these frequencies are mostly in the  $(y, z, x)$  directions, respectively. Figure 4(a) shows an example of motion along both the  $x$  and  $y$  direction. We find good agreement between the flux noise spectral density maps and the fits. These fits allow us to estimate the relative motion in the  $(x, y, z)$  directions of the SQUID with respect to the sample as a function of frequency.

Figure 5 shows the vibration spectral densities for in-plane and out-of-plane directions obtained by fitting at each frequency. Error bars estimated from the 95% confidence intervals of the fits are less than  $1 \text{ nm}/\sqrt{\text{Hz}}$  at all frequencies and omitted to improve readability. Systematic uncertainty comes from uncertainty in the conversion of voltage applied to the piezoelectric scanners to the induced displacement in  $\mu\text{m}$ . This conversion is only used in the final step of the analysis, since all measurements, fits, and computations are carried out in terms of voltages applied to the piezoelectric scanners. We determine the conversion factors for motion in  $x$  and  $y$  as  $175 \pm 10 \text{ nm/V}$  from imaging lithographically defined samples with known dimensions and features. To obtain the conversion factor for vertical motion, we probe the magnetic field produced by a current carrying loop. Using the known height dependence of the magnetic field above the center of the loop, we estimate the conversion to be  $149 \pm 6 \text{ nm/V}$ .

Pressurizing the room-temperature air springs that float the top plate of the cryostat has the most notable impact on the sample-to-SQUID vibrations. The air springs significantly suppress vibrations below 200 Hz. However, we find a surprising amount of motion in a frequency band between 450 and 490 Hz. The amplitude of the vibrations in this range vary significantly in amplitude over long time scales. However, we could not correlate their behavior with any changes in the cryostat (such as the temperature) or the lab environment. From the structure of the noise spectral density maps, we can determine that the most pronounced peak at  $\sim 480 \text{ Hz}$  corresponds to motion along the  $x$ -direction [see, for example, Fig. 4(d)]. Figure 5 is based on a dataset in which these vibrations are particularly pronounced. The integrated in-plane (out-of-plane) vibrations are  $31 \text{ nm}$  ( $27 \text{ nm}$ ) from 0.25 to 450 Hz compared to  $106 \text{ nm}$  ( $39 \text{ nm}$ )



**FIG. 5.** Vibration spectral densities for in-plane (blue) and out-of-plane (orange) relative motion of the SQUID and the sample. Measurements were taken with the air spring pressurized and the mixing chamber temperature at 80 mK with the  $^3\text{He}/^4\text{He}$  mixture circulating. The vertical axis ranges to  $25 \text{ nm}/\sqrt{\text{Hz}}$ , cutting off the highest peak that reaches  $63 \text{ nm}/\sqrt{\text{Hz}}$  at  $\sim 480 \text{ Hz}$  in the in-plane vibration spectrum.

from 0.25 to 1000 Hz. With the air springs turned off, we observed integrated in-plane (out-of-plane) vibrations of 90 nm (32 nm) from 0.25 to 450 Hz.

We have not observed a noticeable difference between the vibrations with and without the circulation of the mixture. We have recorded the motion of the mixing chamber plate along the vertical direction using a geophone, since we cannot measure the sample-to-SQUID vibrations with the pulse tube turned off. With the vacuum cans closed and the fridge at room temperature, no noticeable motion is detected with the pulse tube turned off; however, with the pulse tube turned on, some motion is clearly present at higher frequencies. However, the amplitudes and exact positions of peaks are not strongly correlated with the peaks shown in Fig. 5. We speculate that the motion we observe above 200 Hz has contributions from the pulse tube and that resonances in the microscope shape the spectrum we observe, but that an additional source of vibrations is intermittently present in the lab.

#### IV. CONCLUSIONS

We demonstrate the operation of a scanning SQUID microscope with several imaging modes in a cryogen-free DR. We avoid the use of a spring stage and designed a rigid microscope housing to optimize for thermal anchoring of the sample. Our microscope allows for a large number of wires and coaxial connections and fits in the bore of a superconducting magnet. The temperature of the sample stage reaches at least 30 mK and does not increase significantly during scanning. In Ref. 20, the authors reported the relative sensor-to-sample vibrations in a cryogen-free Bluefors cryostat with a 2.8 K base temperature with and without a spring stage. Our performance falls in between these two benchmarks. We believe that this is achieved due to a combination of a more rigid construction of the microscope and lower vibrations in the cryostat itself.

#### ACKNOWLEDGMENTS

We thank Eric Smith, Jihoon Kim, and Kevin Nangoi for help with the construction of the microscope; Bluefors for technical support and discussions; and Eric D. Bauer, Filip Ronning, Naoki Kikugawa, and Andrew P. Mackenzie for growing and providing the crystals for the microstructures. This work was supported by the U.S. Department of Energy, Office of Basic Energy Sciences, Division of Materials Sciences and Engineering, under Award No. DE-SC0015947 (scanning SQUID imaging and implementation of millikelvin microscopes) and the Cornell Center of Materials Research with funding from the NSF MRSEC program under Award No. DMR-1719875 (SQUID and microscope design). Fabrication of the microstructures was supported by the Max Planck Society and by the Deutsche Forschungsgemeinschaft (DFG, German Research Foundation) (Grant No. MO 3077/1-1) and the European Research Council (ERC) under the European Union's Horizon 2020 research and innovation program (Grant Agreement No. 715730).

#### DATA AVAILABILITY

The data that support the findings of this study are available from the corresponding authors upon reasonable request. The

drawings of the microscope are available at <https://doi.org/10.5281/zenodo.4752118>.<sup>27</sup>

#### REFERENCES

- 1 J. R. Kirtley, "Fundamental studies of superconductors using scanning magnetic imaging," *Rep. Prog. Phys.* **73**, 126501 (2010).
- 2 K. C. Nowack, E. M. Spanton, M. Baenninger, M. König, J. R. Kirtley, B. Kalisky, C. Ames, P. Leubner, C. Brüne, H. Buhmann, L. W. Molenkamp, D. Goldhaber-Gordon, and K. A. Moler, "Imaging currents in HgTe quantum wells in the quantum spin Hall regime," *Nat. Mater.* **12**, 787–791 (2013).
- 3 E. O. Lachman, A. F. Young, A. Richardella, J. Cuppens, H. R. Naren, Y. Anahory, A. Y. Meltzer, A. Kandala, S. Kempinger, Y. Myasoedov, M. E. Huber, N. Samarth, and E. Zeldov, "Visualization of superparamagnetic dynamics in magnetic topological insulators," *Sci. Adv.* **1**, e1500740 (2015).
- 4 J. A. Bert, B. Kalisky, C. Bell, M. Kim, Y. Hikita, H. Y. Hwang, and K. A. Moler, "Direct imaging of the coexistence of ferromagnetism and superconductivity at the LaAlO<sub>3</sub>/SrTiO<sub>3</sub> interface," *Nat. Phys.* **7**, 767–771 (2011).
- 5 B. Kalisky, E. M. Spanton, H. Noad, J. R. Kirtley, K. C. Nowack, C. Bell, H. K. Sato, M. Hosoda, Y. Xie, Y. Hikita, C. Woltmann, G. Pfanzelt, R. Jany, C. Richter, H. Y. Hwang, J. Mannhart, and K. A. Moler, "Locally enhanced conductivity due to the tetragonal domain structure in LaAlO<sub>3</sub>/SrTiO<sub>3</sub> heterointerfaces," *Nat. Mater.* **12**, 1091–1095 (2013).
- 6 X. R. Wang, C. J. Li, W. M. Lu, T. R. Paudel, D. P. Leusink, M. Hoek, N. Poccia, A. Vailionis, T. Venkatesan, J. M. D. Coey, E. Y. Tsymlal, Ariando, and H. Hilgenkamp, "Imaging and control of ferromagnetism in LaMnO<sub>3</sub>/SrTiO<sub>3</sub> heterostructures," *Science* **349**, 716–719 (2015).
- 7 Y. Frenkel, N. Haham, Y. Shperber, C. Bell, Y. Xie, Z. Chen, Y. Hikita, H. Y. Hwang, E. K. H. Salje, and B. Kalisky, "Imaging and tuning polarity at SrTiO<sub>3</sub> domain walls," *Nat. Mater.* **16**, 1203–1208 (2017).
- 8 D. V. Christensen, Y. Frenkel, Y. Z. Chen, Y. W. Xie, Z. Y. Chen, Y. Hikita, A. Smith, L. Klein, H. Y. Hwang, N. Pryds, and B. Kalisky, "Strain-tunable magnetism at oxide domain walls," *Nat. Phys.* **15**, 269–274 (2019).
- 9 J. R. Kirtley, C. C. Tsuei, Ariando, C. J. M. Verwijs, S. Harkema, and H. Hilgenkamp, "Angle-resolved phase-sensitive determination of the in-plane gap symmetry in YBa<sub>2</sub>Cu<sub>3</sub>O<sub>7- $\delta$</sub> ," *Nat. Phys.* **2**, 190–194 (2006).
- 10 N. C. Koshnick, H. Bluhm, M. E. Huber, and K. A. Moler, "Fluctuation superconductivity in mesoscopic aluminum rings," *Science* **318**, 1440–1443 (2007).
- 11 H. Bluhm, N. C. Koshnick, J. A. Bert, M. E. Huber, and K. A. Moler, "Persistent currents in normal metal rings," *Phys. Rev. Lett.* **102**, 136802 (2009).
- 12 S. M. Frolov, M. J. A. Stoumire, T. A. Crane, D. J. Van Harlingen, V. A. Oboznov, V. V. Ryazanov, A. Ruosi, C. Granata, and M. Russo, "Imaging spontaneous currents in superconducting arrays of  $\pi$ -junctions," *Nat. Phys.* **4**, 32–36 (2008).
- 13 E. M. Spanton, M. Deng, S. Vaitiekėnas, P. Krogstrup, J. Nygård, C. M. Marcus, and K. A. Moler, "Current-phase relations of few-mode InAs nanowire Josephson junctions," *Nat. Phys.* **13**, 1177–1181 (2017).
- 14 D. Halbertal, M. Ben Shalom, A. Uri, K. Bagani, A. Y. Meltzer, I. Marcus, Y. Myasoedov, J. Birkbeck, L. S. Levitov, A. K. Geim, and E. Zeldov, "Imaging resonant dissipation from individual atomic defects in graphene," *Science* **358**, 1303–1306 (2017).
- 15 A. Marguerite, J. Birkbeck, A. Aharon-Steinberg, D. Halbertal, K. Bagani, I. Marcus, Y. Myasoedov, A. K. Geim, D. J. Perello, and E. Zeldov, "Imaging work and dissipation in the quantum Hall state in graphene," *Nature* **575**, 628–633 (2019).
- 16 A. Uri, S. Grover, Y. Cao, J. A. Crosse, K. Bagani, D. Rodan-Legrain, Y. Myasoedov, K. Watanabe, T. Taniguchi, P. Moon, M. Koshino, P. Jarillo-Herrero, and E. Zeldov, *Nature* **581**, 47–52 (2020).
- 17 C. L. Tschirhart, M. Serlin, H. Polshyn, A. Shragai, Z. Xia, J. Zhu, Y. Zhang, K. Watanabe, T. Taniguchi, M. E. Huber, and A. F. Young, "Imaging orbital ferromagnetism in a moiré Chern insulator," [arXiv:2006.08053](https://arxiv.org/abs/2006.08053) *Science* **372**(6548), 1323–1327 (2020).
- 18 F. P. Quacquarelli, J. Puebla, T. Scheler, D. Andres, C. Bödefeld, B. Sips, C. Dal Savio, A. Bauer, C. Pfleiderer, A. Erb, and K. Karrai, "Scanning probe microscopy

- in an ultra-low vibration closed-cycle cryostat: Skyrmion lattice detection and tuning fork implementation,” *Microsc. Today* **23**, 12–17 (2015).
- <sup>19</sup>M. Pelliccione, A. Jenkins, P. Oavartchaiyapong, C. Reetz, E. Emmanouilidou, N. Ni, and A. C. Bleszynski Jayich, “Scanned probe imaging of nanoscale magnetism at cryogenic temperatures with a single-spin quantum sensor,” *Nat. Nanotechnol.* **11**, 700–705 (2016).
- <sup>20</sup>L. Bishop-Van Horn, Z. Cui, J. R. Kirtley, and K. A. Moler, “Cryogen-free variable temperature scanning SQUID microscope,” *Rev. Sci. Instrum.* **90**, 063705 (2019).
- <sup>21</sup>Y. Shperber, N. Vardi, E. Persky, S. Wissberg, M. E. Huber, and B. Kalisky, “Scanning SQUID microscopy in a cryogen-free cooler,” *Rev. Sci. Instrum.* **90**, 053702 (2019).
- <sup>22</sup>M. Pelliccione, A. Sciambi, J. Bartel, A. J. Keller, and D. Goldhaber-Gordon, “Design of a scanning gate microscope for mesoscopic electron systems in a cryogen-free dilution refrigerator,” *Rev. Sci. Instrum.* **84**, 033703 (2013).
- <sup>23</sup>A. M. J. den Haan, G. H. C. J. Wijts, F. Galli, O. Usenko, G. J. C. van Baarle, D. J. van der Zalm, and T. H. Oosterkamp, “Atomic resolution scanning tunneling microscopy in a cryogen free dilution refrigerator at 15 mK,” *Rev. Sci. Instrum.* **85**, 035112 (2014).
- <sup>24</sup>M. Hatridge, R. Vijay, D. H. Slichter, J. Clarke, and I. Siddiqi, “Dispersive magnetometry with a quantum limited SQUID parametric amplifier,” *Phys. Rev. B* **83**, 134501 (2011).
- <sup>25</sup>F. Foroughi, J.-M. Mol, T. Müller, J. R. Kirtley, K. A. Moler, and H. Bluhm, “A micro-SQUID with dispersive readout for magnetic scanning microscopy,” *Appl. Phys. Lett.* **112**, 252601 (2018).
- <sup>26</sup>E. M. Levenson-Falk, N. Antler, and I. Siddiqi, “Dispersive nanoSQUID magnetometry,” *Supercond. Sci. Technol.* **29**, 113003 (2016).
- <sup>27</sup>D. Low, G. M. Ferguson, A. Jarjour, B. T. Schaefer, and K. C. Nowack (2021). “Design files for a scanning superconducting quantum interference device operated in a cryogen-free dilution refrigerator,” Zenodo. <https://doi.org/10.5281/zenodo.4752118>.
- <sup>28</sup>J. Siegel, J. Witt, N. Venturi, and S. Field, “Compact large-range cryogenic scanner,” *Rev. Sci. Instrum.* **66**, 2520–2523 (1995).
- <sup>29</sup>P. G. Björnsson, B. W. Gardner, J. R. Kirtley, and K. A. Moler, “Scanning superconducting quantum interference device microscope in a dilution refrigerator,” *Rev. Sci. Instrum.* **72**, 4153–4158 (2001).
- <sup>30</sup>Y. J. Song, A. F. Otte, V. Shvarts, Z. Zhao, Y. Kuk, S. R. Blankenship, A. Band, F. M. Hess, and J. A. Stroscio, “Invited review article: A 10 mK scanning probe microscopy facility,” *Rev. Sci. Instrum.* **81**, 121101 (2010).
- <sup>31</sup>M. D. Bachmann, G. M. Ferguson, F. Theuss, T. Meng, C. Putzke, T. Helm, K. R. Shirer, Y.-S. Li, K. A. Modic, M. Nicklas, M. König, D. Low, S. Ghosh, A. P. Mackenzie, F. Arnold, E. Hassinger, R. D. McDonald, L. E. Winter, E. D. Bauer, F. Ronning, B. J. Ramshaw, K. C. Nowack, and P. J. W. Moll, “Spatial control of heavy-fermion superconductivity in CeIrIn<sub>3</sub>,” *Science* **366**, 221–226 (2019).
- <sup>32</sup>M. E. Huber, N. C. Koshnick, H. Bluhm, L. J. Archuleta, T. Azua, P. G. Björnsson, B. W. Gardner, S. T. Halloran, E. A. Lucero, and K. A. Moler, “Gradiometric micro-SQUID susceptometer for scanning measurements of mesoscopic samples,” *Rev. Sci. Instrum.* **79**, 053704 (2008).
- <sup>33</sup>D. Schiessl, J. R. Kirtley, L. Paulius, A. J. Rosenberg, J. C. Palmstrom, R. R. Ullah, C. M. Holland, Y.-K.-K. Fung, M. B. Ketchen, G. W. Gibson, and K. A. Moler, “Determining the vibrations between sensor and sample in SQUID microscopy,” *Appl. Phys. Lett.* **109**, 232601 (2016).
- <sup>34</sup>J. R. Kirtley, L. Paulius, A. J. Rosenberg, J. C. Palmstrom, C. M. Holland, E. M. Spanton, D. Schiessl, C. L. Jermain, J. Gibbons, Y.-K.-K. Fung, M. E. Huber, D. C. Ralph, M. B. Ketchen, G. W. Gibson, and K. A. Moler, “Scanning SQUID susceptometers with sub-micron spatial resolution,” *Rev. Sci. Instrum.* **87**, 093702 (2016).
- <sup>35</sup>B. Efron and R. Tibshirani, “Bootstrap methods for standard errors, confidence intervals, and other measures of statistical accuracy,” *Stat. Sci.* **1**, 77 (1986).