

***PROTOTYPE IMPINGER GELEMBUNG GANDA (IGG) Uji Kinerja Prototype Impinger
Gelembung Ganda pada Pengukuran Gas Buang Boiler PS II PT. Pertamina (Persero) RU III
Plaju-Sungai Gerong***

***PROTOTYPE DOUBLE BUBBLE IMPINGER (IGG) Performance at Boiler Flue Gas Measurement
PS II PT. Pertamina (Persero) RU III Plaju-Sungai Gerong Measurement***

Hilwatullisan

Jurusan Teknik Kimia Politeknik Negeri Sriwijaya, Jl. Sriwijaya Negara Bukit Besar
Palembang 30139; Telp 0711 353414 e-mail Polsri ac.id

Abstract

The conscious of society with air pollution because flue gas from souch of pollutant in big city is increasing. From any do not move souch although from move souch. So, it need the alternative equipment to monitoring and measuring the hazardous gas. One of them is using double bubble impinger system, it can measure 4 paramaters like NO_x, SO_x, H₂S and NH₃. The method is using in analyze absorp solution in IGG based on Indonesia National standart in 2005. the test of IGG in Boiler PS II Pertamina RU III Plaju. The result of measurement there was different value between emission gas concentration from Pertamina and IGG. Concentration of gas emission derived from IGG were 15,54 ppm, 4,74 ppm, 0,11 ppm, 0,053 ppm whereas Pertamina were 3 ppm, 1 ppm, 0 ppm dan 0 ppm. Based on measuring the performance both IGG and AUER *Pump Detector & Tube* could not be compare because the differences of measuring showed the significant value.

Keyword : Impinger, Flue Gas, Emision