



THE UNIVERSITY *of* EDINBURGH

Edinburgh Research Explorer

Aortic valve imaging using ^{18}F -sodium fluoride: impact of triple motion correction

Citation for published version:

Lassen, ML, Tzolos, E, Massera, D, Cadet, S, Bing, R, Kwiecinski, J, Dey, D, Berman, DS, Dweck, MR, Newby, DE & Slomka, PJ 2022, 'Aortic valve imaging using ^{18}F -sodium fluoride: impact of triple motion correction', *EJNMMI physics*, vol. 9, no. 1. <https://doi.org/10.1186/s40658-022-00433-7>

Digital Object Identifier (DOI):

[10.1186/s40658-022-00433-7](https://doi.org/10.1186/s40658-022-00433-7)

Link:

[Link to publication record in Edinburgh Research Explorer](#)

Document Version:

Publisher's PDF, also known as Version of record

Published In:

EJNMMI physics

General rights

Copyright for the publications made accessible via the Edinburgh Research Explorer is retained by the author(s) and / or other copyright owners and it is a condition of accessing these publications that users recognise and abide by the legal requirements associated with these rights.

Take down policy

The University of Edinburgh has made every reasonable effort to ensure that Edinburgh Research Explorer content complies with UK legislation. If you believe that the public display of this file breaches copyright please contact openaccess@ed.ac.uk providing details, and we will remove access to the work immediately and investigate your claim.




ORIGINAL RESEARCH

Open Access



Aortic valve imaging using ^{18}F -sodium fluoride: impact of triple motion correction

Martin Lyngby Lassen^{1,2}, Evangelos Tzolos^{3,4}, Daniele Massera⁵, Sebastien Cadet³, Rong Bing⁴, Jacek Kwiecinski^{3,6}, Damini Dey⁷, Daniel S. Berman³, Marc R. Dweck⁴, David E. Newby⁴ and Piotr J. Slomka^{1*} 

*Correspondence:

piotr.slomka@cshs.org

¹ Department of Medicine
(Division of Artificial
Intelligence in Medicine),
Cedars-Sinai Medical Center,
8700 Beverly Blvd Ste. Metro
203, Los Angeles, CA 90048,
USA

Full list of author information
is available at the end of the
article

Abstract

Background: Current ^{18}F -NaF assessments of aortic valve microcalcification using ^{18}F -NaF PET/CT are based on evaluations of end-diastolic or cardiac motion-corrected (ECG-MC) images, which are affected by both patient and respiratory motion. We aimed to test the impact of employing a triple motion correction technique ($3 \times \text{MC}$), including cardiorespiratory and gross patient motion, on quantitative and qualitative measurements.

Materials and methods: Fourteen patients with aortic stenosis underwent two repeat 30-min PET aortic valve scans within (29 ± 24) days. We considered three different image reconstruction protocols; an end-diastolic reconstruction protocol (standard) utilizing 25% of the acquired data, an ECG-gated (four ECG gates) reconstruction (ECG-MC), and a triple motion-corrected ($3 \times \text{MC}$) dataset which corrects for both cardiorespiratory and patient motion. All datasets were compared to aortic valve calcification scores (AVCS), using the Agatston method, obtained from CT scans using correlation plots. We report SUV_{max} values measured in the aortic valve and maximum target-to-background ratios (TBR_{max}) values after correcting for blood pool activity.

Results: Compared to standard and ECG-MC reconstructions, increases in both SUV_{max} and TBR_{max} were observed following $3 \times \text{MC}$ (SUV_{max} : Standard = 2.8 ± 0.7 , ECG-MC = 2.6 ± 0.6 , and $3 \times \text{MC}$ = 3.3 ± 0.9 ; TBR_{max} : Standard = 2.7 ± 0.7 , ECG-MC = 2.5 ± 0.6 , and $3 \times \text{MC}$ = 3.3 ± 1.2 , all p values ≤ 0.05). $3 \times \text{MC}$ had improved correlations (R^2 value) to the AVCS when compared to the standard methods (SUV_{max} : Standard = 0.10, ECG-MC = 0.10, and $3 \times \text{MC}$ = 0.20; TBR_{max} : Standard = 0.20, ECG-MC = 0.28, and $3 \times \text{MC}$ = 0.46).

Conclusion: $3 \times \text{MC}$ improves the correlation between the AVCS and SUV_{max} and TBR_{max} and should be considered in PET studies of aortic valves using ^{18}F -NaF.

Keywords: Motion correction, PET/CT, Cardiac PET, ^{18}F -sodium fluoride, Aortic valve imaging

Background

Positron emission tomography (PET) utilizing ^{18}F -sodium fluoride (^{18}F -NaF) combined with computed tomography angiography (CTA) permits identification of microcalcification activity and progression of disease in aortic valves [1–6]. Moreover, ^{18}F -NaF PET is

currently being used as an endpoint in several ongoing clinical trials assessing the efficacy of novel therapies in aortic stenosis [7–9].

Cardiac ^{18}F -NaF PET imaging is affected by motion (cardiac and respiratory [cardiorespiratory] and patient) and by technical challenges associated with the current acquisition protocols for ^{18}F -NaF PET, which can last for up to 30 min. The long imaging protocols were initially designed to ensure sufficient count statistics to obtain images of diagnostic quality [10]. Unfortunately, patient motion during imaging protocols with long acquisition times degrades the image quality of the scans [10] and consequently reduces the quantitative accuracy and the test–retest repeatability [11, 12]. In addition, variations in the tracer injection-to-scan delays affect the quantitative assessment of the lesions with ^{18}F -NaF PET [13–15]. While the impact of these factors is of keen interest in studies of aortic valve microcalcification, to date, it has been only evaluated in studies of coronary plaques [10–15]. Best possible image quality and spatial resolution are of critical importance in studies of microcalcification in native valves to understand the pathophysiology of aortic stenosis, and in studies of bioprosthetic valves where accurate localization of ^{18}F -NaF uptake is essential.

To this end, we aimed to test the hypothesis that correcting for cardiorespiratory and patient motion ($3 \times \text{MC}$) will lead to improved image quality (signal-to-noise ratio [SNR]) [10], and increased semi-quantitative measurements (maximum standardized uptake values [SUV_{max}] and maximum target-to-background ratios [TBR_{max}]) [11] in studies of microcalcification activity in the aortic valve.

Methods

Study population

Fourteen patients with aortic stenosis underwent repeated ^{18}F -NaF aortic valve PET/CT examinations with the two occasions 4 weeks apart as a part of the ongoing study investigating the effect of drugs used to treat osteoporosis on the progression of calcific aortic stenosis (SALTIRE II) [7]. Of note, only the two baseline scans (scan 1 and scan 2) of the SALTIRE II study were evaluated in this study; thus, no disease progression was expected between the two scans. These scans were obtained originally to check the test–retest variability of the entire imaging procedure [10, 16].

This study was approved by the Scottish Research Ethics Committee and the Medicines and Healthcare Products Regulatory Authority of the United Kingdom, and the study was performed in accordance with the declaration of Helsinki. All participants signed written informed consent.

Imaging protocol

PET/CT All patients underwent 30-min listmode PET-emission scans one hour following ^{18}F -NaF injection (target injection 125 MBq). All scans were acquired on a 128-slice Biograph mCT system (Siemens Healthineers, Knoxville, USA). All patients had a low-dose CT scan for attenuation correction purposes prior to the PET acquisitions (120 kV, 50 mAs, 3-mm slice thickness). Three ECG leads were used for the detection of cardiac motion (ECG gating). No additional external markers for tracking patient or respiratory motion were employed.

CT angiography For anatomical identification of the aortic valves, all patients had a cardiac CT angiogram (CTA) for each scan session immediately after emission scanning. All CTA images were acquired using prospective gating, a 330-ms rotation time, and a body-mass index (BMI)-dependent voltage (BMI < 25 kg/m², 100 kV, BMI ≥ 25 kg/m², 120 kV). All patients had beta-blockers administered, either orally or intravenously to achieve a target heart rate of < 60 /min. Iodinated contrast was administered in bolus form (400 mg/mL) with an injection rate of 5–6 mL/s after determining the appropriate trigger delay, defined as a test-bolus of 20 mL of contrast material.

Calcium scoring Aortic valve CT calcium scoring was acquired for all patients to estimate the calcium scores for the repeat PET/CT scans and used for primary outcome of the study. A non-contrast ECG-gated CT was performed at each visit using the same scanner, electrocardiogram gating, and a standardized protocol (120 kV CARE Dose4D [Siemens], 3-mm slice thickness, spiral acquisition, 70% R-R interval, inspiratory breath-hold). CT calcium scoring was performed by an experienced operator using dedicated software (Vitrea Advanced; Toshiba Systems) on axial views, with care taken to exclude calcium originating from the ascending aorta, left ventricular outflow tract, and coronary arteries. The calcium score was recorded in Agatston units.

Motion detection and image reconstruction

Three different image reconstructions were evaluated in this study; (1) an end-diastolic image reconstruction employing 25% of the counts (Standard) [17] a cardiac motion-corrected dataset (ECG-MC) reconstruction employing four cardiac phases [10], and a triple motion-corrected (3 × MC) dataset employing 16 cardiorespiratory phases per each patient position phase as determined from the listmode. Thus, gross patient motion had variable number of phases. The 3 × MC reconstructions were obtained using a previously described gating protocol (Fig. 1) [12]. In brief, information on cardiac contraction was obtained using ECG gating. Respiratory and patient motion was detected retrospectively using data-driven gating. In brief, the data-driven gating employed a center-of-mass-based analysis of single-slice-rebinned sinograms (3D) obtained for every 200 ms [11, 18]. Information on the respiratory motion was extracted from the diaphragm region only using segmentation on the diaphragm area from the co-registered low-dose CT attenuation correction map, whereas gross body motion was detected by evaluating the whole PET field-of-view (longitudinal direction) as in work by Büther et al. [19] for respiratory motion detection. The motion detection was obtained using four time-of-flight bins covering the central part of the PET system (transaxial direction). For automatic extraction of patient repositioning events, the center-of-mass signal was filtered with a band-stop filter (0.2–0.5 Hz, equivocal to 12–30 respiratory cycles per minute) and a moving-average filter (3 s) to remove stochastic noise in the signal. Patient repositioning events were defined as changes in the center-of-mass baseline (> 0.5 mm over 3 s (sudden repositioning) or > 0.3 mm over 15 s (gradual repositioning events) [11]. The respiratory signal was obtained using a band-pass filter of the extracted center-of-mass signal (0.2–0.5 Hz), followed by a moving-average filter (3 s) to remove stochastic noise in the signal.




Image reconstruction

All PET images were reconstructed using vendor-provided software (JS-Recon12, Siemens, Knoxville, USA), using corrections for both point-spread and time-of-flight using two iterations and 21 subsets. All reconstructions were performed using a voxel-grid of $256 \times 256 \times 109$ indices ($2.73 \times 2.73 \times 2.037$ mm) and filtered with a 5-mm Gaussian post-filter.

Image registration

All PET datasets were co-registered to the CTA images, using anatomical landmarks as described in a previous study [10]. In brief, the images were co-registered based on matching of the blood pool of the two imaging modalities by overlaying the four cardiac chambers for the PET and CTA images. Following PET to CTA co-registration, the PET-PET image registration (PET motion correction) was performed within a sphere (radius = 20 mm) enclosing the aortic valve, thereby, creating ECG-MC and 3 × MC image-sets. The motion correction was performed using nonlinear registrations (daemons) [20]. All motion correction was performed in FusionQuant, thus, employing a post-reconstruction motion correction technique registering the 3D reconstructed frames (FusionQuant, Cedars-Sinai Medical Center) [10].


Image analysis

PET Quantification Using the co-registered PET and CTA images, the valvular uptake was measured using a three-dimensional polyhedron (6-mm thickness) encompassing the valve (Fig. 2). In addition, background blood pool uptake was measured in the center of the right atrium (cylindric volume: radius = 8 mm, height = 9 mm) [12, 16]. Of note, the background blood pool activity was extracted from the right atrium as it offers reduced coefficient of variation in the measurements compared to measurements obtained in the vena cava and brachiocephalic vein, although at the cost of slightly increased blood pool activities [16]. The same volume of interests (VOIs) were used for all three reconstruction protocols. All scans were read in a blinded fashion, with readings of the images 6 weeks apart to reduce the risk of bias in the assessment [10]. All the images were read by a cardiologist certified in echocardiography and computed tomography angiography.

To compensate for variations in the background blood pool activity introduced by variations in the injection-to-scan delays [14], the activity observed in the background at time t (in minutes) after the injection was normalized to a standardized injection-to-scan delay of 60 min using a previously described correction factor [12]:

$$\text{SUV}_{\text{Background corrected}} = \text{SUV}_{\text{background}} \times e^{-0.004 \times (60-t)} \quad (1)$$

TBR_{max} measurements were obtained by normalizing the SUV_{max} measured in the segmented valve to the corrected mean standardized uptake value obtained in the right atrium ($\text{SUV}_{\text{Corrected Background}}$) (Eq. 2).



$$TBR_{\max} = \frac{SUV_{\max}}{SUV_{\text{Corrected Background}}} \quad (2)$$

To assess the image quality of the resulting reconstructions (Standard, ECG-MC, and $3 \times MC$), we evaluated the signal -to-noise ratio (SNR) obtained in the background. The SNR was defined as the SUV_{\max} obtained in the valve normalized to the standard deviation obtained in the background blood pool (right atrium) [10]. The magnitude of the patient repositioning was measured from the PET-to-PET co-registration using the inverse motion vector fields.

Statistical analysis

The data were tested for normality using the Shapiro–Wilk test, using corrections for multiple comparisons (Bonferroni corrections). Continuous variables were presented as mean \pm standard deviation. Test–retest repeatability of the TBR_{\max} assessments was compared using the repeatability coefficient (RC; $RC = 1.96 \times \text{standard deviation}$ with the standard deviation being expressed in % difference from the initial scans. Of note, the standard deviation was calculated from the 14 patients included in this study) and using the Kendall's Tau measure [12]. Tests for variations in TBR_{\max} assessments were evaluated using Pitman–Morgan analyses with p values < 0.05 were considered statistically significant.

Results

All patients underwent two ^{18}F -NaF PET/CTA scans within one month. The patients had an average AVCS of $1259 \pm 908\text{AU}$. Of the 14 patients, 11 were acquired with the arms placed above the head, while the remaining three patients were unable to have the scans performed with the arms elevated. Patient demographics are shown in Table 1.

Acquisition and motion estimation

The average dose was 124 ± 6 MBq of ^{18}F -NaF. On average, the patients had injection-to-scan delays of 68 ± 11 min (range 59–99 min). During the 30-min acquisitions, patients were found to reposition themselves 3.6 ± 1.3 times (range 1–6). The arm position did not seem to affect the number of repositioning events (Repositioning events: arms up = 4 ± 1 , arms down = 4 ± 1). Acquisition individual maximum translations in contrast to the reference frame was of magnitude of 14 mm (3D motion) (3D motion: 14.3 ± 2.3 mm, range = [11.1; 21.8 mm]) (Additional file 1: Supplementary figure 1).

SUV_{\max} and TBR_{\max} assessments


The $3 \times MC$ reconstruction protocol resulted in increased SUV_{\max} and TBR_{\max} values when compared to both the standard and ECG-MC protocols (SUV_{\max} : Standard = 2.8 ± 0.7 , ECG-MC = 2.6 ± 0.6 , and $3 \times MC = 3.3 \pm 0.9$; TBR_{\max} : Standard = 2.7 ± 0.7 , ECG-MC = 2.5 ± 0.6 and $3 \times MC = 3.4 \pm 1.2$, all p values < 0.05) (Fig. 3). Of note, the median variation in the injection-to-scan delay (intra-patient delay) was 6.5 min (inter quartile range = [1.5; 10.5 min]). Correlations plots of the aortic valve calcium score (AVCS) and the SUV_{\max} and TBR_{\max} (Fig. 4) were improved for the $3 \times MC$ technique when compared to the Standard and ECG-MC



Table 1 Patient demographics

Demographics	Value
Age (years)	73 ± 7
Gender (Males)	10 (67.7)
Body mass index (BMI)	27.2 ± 4.3
<i>Cardiovascular risk factors</i>	
Diabetes mellitus	4 (26.7)
Current smoker	6 (40.0)
Hypertension	11 (73.3)
Hyperlipidemia	10 (67.7)
TIA/CVA	2 (13.3)
Renal disease	0 (0)
CABG	2 (13.3)
PCI	4 (23.7)
<i>Medications</i>	
ACE inhibitor	6 (40.0)
ARB	3 (20.0)
Beta blocker	7 (56.6)
Statin	9 (60.0)
<i>Aortic stenosis grade</i>	
Mild	7 (47.6)
Moderate	4 (26.7)
Severe	4 (26.7)
Aortic valve calcium score (Agatston)	1269 [246–5774]

Continuous variables reported as mean ± SD or median [range]; categorical variables reported as n (%)

TIA transient ischemic attack, CVA cerebrovascular accident





techniques (R^2 ; SUV_{max} : Standard = 0.10, ECG-MC = 0.10, and $3 \times MC$ = 0.20; TBR_{max} : Standard = 0.20, ECG-MC = 0.28, and $3 \times MC$ = 0.46).

Image quality

Lower SNR were obtained for the standard (end-diastolic) image approach ($SNR = 25.7 \pm 10.1$) when compared to the motion-corrected datasets (SNR : ECG-MC = 37.3 ± 12.9 , $3 \times MC$ = 45.7 ± 16.9) (both $p < 0.0001$). Noteworthy, the SNR obtained for $3 \times MC$ has improved when compared to the ECG-MC ($p < 0.001$) (Fig. 5).


Test–retest repeatability measurements


Similar test–retest repeatability coefficient were obtained for all three reconstruction protocols (SUV_{max} : Standard = 25.6%, ECG-MC = 20.7%, $3 \times MC$ = 20.3%; TBR_{max} : Standard = 14.4%, ECG-MC = 14.8%, $3 \times MC$ = 13.9%, all $p \geq 0.79$) (Fig. 6). Similarly, the Kendall's Tau measures showed improved reliability of the $3 \times MC$ reconstruction protocols, although not being statistically significant (SUV_{max} : Standard = 0.74, ECG-MC = 0.69, $3 \times MC$ = 0.74; TBR_{max} : Standard = 0.80, ECG-MC = 0.78%, $3 \times MC$ = 0.85, all $p \geq 0.79$). Figures 7 and 8 show two cases of test–retest repeatability of the tracer-uptake before and after motion correction.

Discussion

We evaluated the use of a novel $3 \times MC$ PET reconstruction technique tested for the first time in PET imaging of aortic valves, including cardiorespiratory and gross patient motion correction. The impact of the $3 \times MC$ protocol was evaluated on five criteria assessing the quantitative and qualitative assessments of the images: SUV_{max} , TBR_{max} , SNR, the test–retest repeatability, and the correlation between the quantitative measures and AVCS. The $3 \times MC$ was superior to the standard (end-diastolic) and ECG-MC imaging protocols without affecting the test–retest repeatability. The improved localized uptake observed in the aortic valves for the $3 \times MC$ reconstructions correlated better with the AVCS than the standard and ECG-MC protocols, which in combination with the increased SNR might aid the understanding of the pathophysiology in native and bioprosthetic valve diseases.


The assessment of aortic valve microcalcification with ^{18}F -NaF can be used to predict aortic valve stenosis progression [1, 4, 16, 21]. Studies to date have established the association between increased TBR_{max} and aortic stenosis progression using either standard (end-diastolic) or ECG-MC image series [1, 10, 15]. However, it is uncertain how much patient and respiratory motion can affect quantitative assessments of ^{18}F -NaF





uptake (SUV_{max} and TBR_{max}). It is therefore of great interest to evaluate potential image improvements offered by the $3 \times MC$ protocol in the context of aortic valve microcalcification. Specifically, the improved co-localization of PET activity with areas of aortic valve calcification is of high interest in studies investigating the various causes of leaflet calcification, i.e., mechanical stress at the leaflet coaptation points versus commissures. Moreover, it might help differentiate between activities originating in the aortic valve from activity from nearby structures (left main stem, mitral valve, left atrium, etc.) by reducing the spillover effect of those structures. This is of importance in cases of bioprosthetic valves, and transcatheter aortic valve implantation (TAVI), where localizing the source of PET activity can help differentiate between bioprosthetic valve leaflet degeneration and remote activity originating from the valve struts or crushed native leaflets in case of TAVI [22]. In this context, partial volume effects strongly affect the activity observed in the aortic valves and their surrounding tissues. It is anticipated that the $3 \times MC$ reconstruction protocol may ameliorate the impact of the apparent partial volume effects because of the co-registration of the gated images with significantly reduced intra-gate motion compared to the standard and ECG-MC reconstruction protocols where several motion patterns blur the resulting images. Therefore, it is believed that the $3 \times MC$ reduces the impact of the partial volume effects while also aiding toward better test–retest repeatability.

In the current study, all patients demonstrated an increase in both SUV_{max} and TBR_{max} when the data were corrected using $3 \times MC$ (cohort-based increase: $SUV_{max} = 26\%$,



$TBR_{max} = 33\%$) (Fig. 3), indicating loss of signal when not applying these correction techniques. The improved SUV_{max} and TBR_{max} assessments for the $3 \times MC$ both had preserved repeatability measures and led to a twofold increase in the correlations to the aortic valve calcium score when compared to the Standard and ECG-MC reported results; thus, suggesting that $3 \times MC$ might improve predictions on disease progression (Figs. 3, 4, 6, 7, 8) [1]. Based on these findings, we can conclude that the motion patterns across different patients vary widely, depending on the respiratory translations and patient motion patterns during the acquisitions, which in some cases can introduce variations in the quantitative values exceeding 40% (average increases of the four scans with substantial changes in SUV_{max} and TBR_{max} values: $SUV_{max} = 42\%$, $TBR_{max} = 72\%$). The increase in the correlation scores observed for the TBR_{max} and aortic valve calcium score may translate into lower number of patients required in studies investigating the effects of interventions on ^{18}F -NaF PET uptake (as a marker of calcification activity) because any true effect will not be covered by noise within the region of interest.

Motion during the scans has been shown to have a detrimental impact on the image quality, measured as SNR [10, 13]. In concordance with a previous study of coronary plaques, the SNR was significantly reduced for the standard imaging protocol when compared to the motion-corrected protocols [13]. The low SNR observed for the standard protocols is introduced by two arms, low count statistics (25% of the acquired data used for the analyses), and respiratory and patient motion blur embedded in the ECG-gated

images [12]. Introducing motion correction (ECG-MC and $3 \times$ MC), the SNR improved as all the data were used in the analyses partly owing to the increased count statistics in the images (100% of the acquired data). The further improvement observed for the $3 \times$ MC was introduced by the corrections for both respiratory and patient motion, which reduced the residual blur introduced to the ECG-gated reconstructions, as shown in Fig. 4. While SNR was significantly reduced using ECG-MC alone compared to $3 \times$ MC, ECG-MC provides greatly improved SNR when compared to the standard assessment and therefore should be considered when $3 \times$ MC is not possible. Of note, the gated reconstructions took 3 min to reconstruct each gate using a 3 year-old PC, introducing a reconstruction time of 48 min per repositioning event.

While some of the discrepancies, in general, may be attributed to changes in the pathological disease, only minor changes in the SUV_{\max}/TBR_{\max} test–retest variability is expected to be introduced by disease progression owing to the slow microcalcification processes which may change the calcification burden by 24% per year (and thus, only 1–2% change in the SUV_{\max}/TBR_{\max} burden should be expected) [16]. Therefore, the discrepancy in the impact of the ECG-MC and $3 \times$ MC observed for SNR, TBR_{\max} , and SUV_{\max} indicate that the different motion patterns (cardiac contraction, respiratory and gross patient motion) affect image quality to varying extent. The improved quantitative and qualitative findings observed for the $3 \times$ MC protocol suggest that cardiac contraction is of less importance to correct for in studies of aortic valves when compared to the patient repositioning events and the respiratory translations. Such results were expected as the cardiac contraction affects mainly the coronary arteries, where the right coronary artery has been reported to shift up to 26 mm [23] which is in contrast to the average displacement of the aortic valves (~ 12 mm) [24]. In contrast, respiratory motion has been reported to displace the heart by up to 21 mm [25] and patient repositioning events might shift the heart 5–15 mm [11]. In this context, patient motion (repositioning events) usually happens with low frequency (approximately 3.5 times per scan) and although the translations are few, they introduce non-periodical shifts of the respiratory baseline and, thus, affect the cardiorespiratory motion correction.

Another important finding in our study is the robustness of the data-driven motion detection techniques, as applied to valve imaging studies. A previous study has shown that motion detection techniques employing these assessments are affected by reductions in the count rates [26]. The count rates are mainly affected by two variables: the injected activity and injection-to-scan delay. While the previous study utilizing a $3 \times$ MC protocol (focusing on coronary plaques) had similar injection-to-scan delays as the current valve study (59–99 min) [12], the injection doses, in general, are halved for the studies of aortic valve microcalcification (to 125 MBq) [7, 10, 17], as compared to studies of coronary plaques (250 MBq) [12, 14, 27]).

Given the large dose-reduction in the current study, it was important to test the robustness of the extracted motion patterns (respiratory and patient). In this context, we evaluated the robustness of the technique by evaluating the impact of the data-driven motion detection techniques through assessments of the TBR_{\max} , SNR, and test–retest repeatability of the TBR_{\max} assessments. First, the number of repositioning events reported in this study was in concordance with previous studies [11, 12], suggesting that the robustness of the motion detection is preserved. Motion correction of the

detected repositioning and respiratory events introduced an increase in both SUV_{max} , TBR_{max} , and SNR while preserving a high test–retest repeatability which is in concordance with already established imaging protocols [28]. These findings strongly indicate that the data-driven motion detection technique developed for patient and respiratory motion detection algorithm provides robust and reliable results in valve imaging even when using a low-dose imaging protocol.

Limitations

In this study, the number of patients was limited to 14 who underwent two PET/CTA scans within a month—this number is limited by the difficulty in obtaining such repeated PET/CT scans with short time interval. Nevertheless, we were able to report substantial improvements for SUV_{max} and TBR_{max} following $3 \times MC$, which correlated better to the calcium scores. The attenuation correction was not motion-corrected prior to image reconstruction, which might pose another limitation. However, in a previous study from our center, we showed that a respiratory averaged CT attenuation correction did not change the quantitative assessment. Therefore, we do consider this a limitation of this study [29]. Another limitation was the use of only four cardiac and four respiratory gates for each patient repositioning event during the scans; however, double gating imposes count limitations especially in low-dose studies as studied. Further improvement in TBR_{max} values is possible with an increasing number of cardiorespiratory gates or introducing motion correction during the reconstruction, whereby noise in the images will be suppressed.

Conclusion

$3 \times MC$ employed in ^{18}F -NaF aortic valve imaging substantially improves the correlation between the obtained aortic valve calcium score and SUV_{max} , TBR_{max} and should be considered in PET studies of aortic valves using ^{18}F -NaF.

Abbreviations

PET: Positron emission tomography; CTA: Computed tomography angiography; ECG- MC: Cardiac contraction motion-corrected; $3 \times MC$: Triple motion-corrected; TBR_{max} : Maximum target-to-background ratio; SUV: Standardized uptake value; VOI: Volume of interest; AVCS: Aortic valve calcification score.

Supplementary Information

The online version contains supplementary material available at <https://doi.org/10.1186/s40658-022-00433-7>.

Additional file 1: Supplementary Figure 1. Maximum translations (3D) of the aortic valve obtained for the repeat scans. On average, the aortic valve was translated 14.3 mm during the scans, with a maximum translation of 21.8 mm.

Authors' contributions

MLL interpreted the data, wrote the gating software used for the analysis, and drafted the manuscript. DM drafted the work and assisted in the data analysis. ET interpreted the data and drafted the manuscript. RB, JK, DD, and DSB helped to draft the manuscript. MRD and DEN designed the work and helped to revise the manuscript. PJS helped to interpret the data and draft the manuscript. All authors have approved the manuscript submission and agreed to the submission guidelines.

Funding

This research was supported in part by Grant R01HL135557 from the National Heart, Lung, and Blood Institute/National Institute of Health (NHLBI/NIH). The content is solely the responsibility of the authors and does not necessarily represent the official views of the National Institutes of Health. In addition, the study was supported by Siemens Medical

Systems. The study was also supported by a Grant ("Cardiac Imaging Research Initiative") from the Miriam & Sheldon G. Adelson Medical Research Foundation. DEN is supported by the British Heart Foundation (CH/09/002, RG/16/10/32375, RE/18/5/34216) and is the recipient of a Wellcome Trust Senior Investigator Award (WT103782AIA). MRD is supported by the Sir Jules Thorn Biomedical Research Award (JTA/15) and the British Heart Foundation (FS/14/78/31020).

Availability of data and materials

The data is not publicly available because this study is still ongoing.

Declarations

Ethical approval and consent to participate

All procedures performed in studies involving human participants were in accordance with the ethical standards of the institutional research committee (Edinburgh, UK) and with the 1964 Helsinki declaration and its later amendments or comparable ethical standards. Informed consent was obtained from all individual participants included in the study.

Consent for publication

All co-authors have read and approved the submission of the publication.

Competing interests

No relevant conflict of interests relevant to this manuscript exist. This study was in part supported by Siemens.

Author details

¹Department of Medicine (Division of Artificial Intelligence in Medicine), Cedars-Sinai Medical Center, 8700 Beverly Blvd Ste. Metro 203, Los Angeles, CA 90048, USA. ²Department of Clinical Physiology, Nuclear Medicine and PET and Cluster for Molecular Imaging, Department of Biomedical Sciences, Rigshospitalet and University of Copenhagen, Copenhagen, Denmark. ³Department of Imaging, Cedars-Sinai Medical Center, 8700 Beverly Blvd Ste. Metro 203, Los Angeles, CA 90048, USA. ⁴British Heart Foundation Centre for Cardiovascular Science, Clinical Research Imaging Centre, Edinburgh Heart Centre, University of Edinburgh, Edinburgh, UK. ⁵Leon H. Charney Division of Cardiology, New York University School of Medicine, New York, NY, USA. ⁶Department of Interventional Cardiology and Angiology, Institute of Cardiology, Warsaw, Poland. ⁷Department of Biomedical Sciences, Cedars-Sinai Medical Center, 8700 Beverly Blvd Ste. Metro 203, Los Angeles, CA 90048, USA.

Received: 20 September 2021 Accepted: 12 January 2022

Published online: 29 January 2022

References

1. Dweck MR, Jenkins WSA, Vesey AT, Pringle MAH, Chin CWL, Malley TS, et al. 18F-sodium fluoride uptake is a marker of active calcification and disease progression in patients with aortic stenosis. *Circ Cardiovasc Imaging*. 2014;7:371–8.
2. Bing R, Dweck MR. Aortic valve and coronary 18F-sodium fluoride activity: a common cause? *J Nucl Cardiol*. 2019. <https://doi.org/10.1007/s12350-019-01901-x>.
3. Nakamoto Y, Kitagawa T, Sasaki K, Tatsugami F, Awai K, Hirokawa Y, et al. Clinical implications of 18F-sodium fluoride uptake in subclinical aortic valve calcification: its relation to coronary atherosclerosis and its predictive value. *J Nucl Cardiol*. 2019. <https://doi.org/10.1007/s12350-019-01879-6>.
4. Jenkins WSA, Vesey AT, Shah ASV, Pawade TA, Chin CWL, White AC, et al. Valvular ¹⁸F-fluoride and ¹⁸F-fluorodeoxyglucose uptake predict disease progression and clinical outcome in patients with aortic stenosis. *J Am Coll Cardiol*. 2015;66:1200–1. <https://doi.org/10.1016/j.jacc.2015.06.1325>.
5. Doris MK, Everett RJ, Shun-Shin M, Clavel MA, Dweck MR. The role of imaging in measuring disease progression and assessing novel therapies in aortic stenosis. *JACC Cardiovasc Imaging*. 2019;12:185–97.
6. Irkle A, Vesey AT, Lewis DY, Skepper JN, Bird JLE, Dweck MR, et al. Identifying active vascular microcalcification by 18F-sodium fluoride positron emission tomography. *Nat Commun*. 2015;6:1–11.
7. Trials C. Study Investigating the Effect of Drugs Used to Treat Osteoporosis on the Progression of Calcific Aortic Stenosis. (SALTIRE II). Study Investig. Eff. Drugs Used to Treat Osteoporos. *Progress. Calcif. Aortic Stenosis*. (SALTIRE II). 2014. <https://clinicaltrials.gov/ct2/show/NCT02132026?co>.
8. Peeters FECM, Van Mourik MJW, Meex SJR, Bucerius J, Schalla SM, Gerretsen SC, et al. Bicuspid aortic valve stenosis and the effect of vitamin K2 on calcification using 18F-sodium fluoride positron emission tomography/magnetic resonance: the BASIK2 rationale and trial design. *Nutrients*. 2018;10.
9. Kim H-S. PCSK9 Inhibitors in the Progression of Aortic Stenosis. 2017. <https://clinicaltrials.gov/ct2/show/NCT03051360>.
10. Massera D, Doris MK, Cadet S, Kwiecinski J, Pawade TA, Peeters FECM, et al. Analytical quantification of aortic valve 18F-sodium fluoride PET uptake. *J Nucl Cardiol*. 2020;27:962–72. <https://doi.org/10.1007/s12350-018-01542-6>.
11. Lassen ML, Kwiecinski J, Cadet S, Dey D, Wang C, Dweck MR, et al. Data-driven gross patient motion detection and compensation: Implications for coronary 18F-NaF PET imaging. *J Nucl Med Soc Nucl Med*. 2019;60:830–6.
12. Lassen ML, Kwiecinski J, Dey D, Cadet S, Germano G, Berman DS, et al. Triple-gated motion and blood pool clearance corrections improve reproducibility of coronary 18F-NaF PET. *Eur J Nucl Med Mol Imaging*. 2019;46:2610–20.
13. Lassen ML, Kwiecinski J, Dey D, Cadet S, Germano G, Berman D, et al. Triple motion correction including cardiorespiratory and gross patient motion: application in coronary plaque imaging using PET. *J Nucl Med Soc Nucl Med*. 2019;60:104.

14. Kwiecinski J, Berman DS, Lee S-E, Dey D, Cadet S, Lassen ML, et al. Three-hour delayed imaging improves assessment of coronary 18 F-sodium fluoride PET. *J Nucl Med*. 2019;60:530–5. <https://doi.org/10.2967/jnumed.118.217885>.
15. Andrews JPM, MacNaught G, Moss AJ, Doris MK, Pawade T, Adamson PD, et al. Cardiovascular 18F-fluoride positron emission tomography-magnetic resonance imaging: a comparison study. *J Nucl Cardiol*. 2019. <https://doi.org/10.1007/s12350-019-01962-y>.
16. Pawade TA, Carlidge TRG, Jenkins WSA, Adamson PD, Robson P, Lucatelli C, et al. Optimization and reproducibility of aortic valve 18F-fluoride positron emission tomography in patients with aortic stenosis. *Circ Cardiovasc Imaging*. 2016;9:1–11.
17. Doris MK, Rubeaux M, Pawade T, Otaki Y, Xie Y, Li D, et al. Motion-Corrected imaging of the aortic valve with 18 F-NaF PET/CT and PET/MRI: a feasibility study. *J Nucl Med*. 2017;58:1811–4. <https://doi.org/10.2967/jnumed.117.194597>.
18. Lassen ML, Beyer T, Berger A, Beitzke D, Rasul S, Büther F, et al. Data-driven, projection-based respiratory motion compensation of PET data for cardiac PET/CT and PET/MR imaging. *J Nucl Cardiol*. 2020;27:2216–30.
19. Büther F, Ernst I, Dawood M, Kraxner P, Schäfers M, Schober O, et al. Detection of respiratory tumour motion using intrinsic list mode-driven gating in positron emission tomography. *Eur J Nucl Med Mol Imaging*. 2010;37:2315–27.
20. Rubeaux M, Joshi N, Dweck MR, Fletcher A, Motwani M, Thomson LE, et al. Demons versus level-set motion registration for coronary 18 F-sodium fluoride PET. *Proc SPIE Int Soc Opt Eng*. 2016; 97843Y.
21. Tzolos E, Dweck MR. Aortic valve stenosis—multimodality assessment with PET/CT and PET/MRI. *Br J Radiol*. 2019;93:20190688.
22. Carlidge TRG, Doris MK, Sellers SL, Pawade TA, White AC, Pessotto R, et al. Detection and prediction of bioprosthetic aortic valve degeneration. *J Am Coll Cardiol*. 2019;73:1107–19.
23. Shechter G, Resar JR, McVeigh ER. Displacement and velocity of the coronary arteries: cardiac and respiratory motion. *IEEE Trans Med Imaging*. 2006;25:369–75.
24. Plonek T, Berezowski M, Kurcz J, Podgorski P, Sasiadek M, Rylski B, et al. The evaluation of the aortic annulus displacement during cardiac cycle using magnetic resonance imaging. *BMC Cardiovasc Disord*. 2018;18:1–6.
25. Dawood M, Büther F, Stegger L, Jiang X, Schober O, Schäfers M, et al. Optimal number of respiratory gates in positron emission tomography: a cardiac patient study. *Med Phys*. 2009;36:1775–84.
26. Büther F, Ernst I, Frohwein LJ, Pouw J, Schäfers KP, Stegger L. Data-driven gating in PET: influence of respiratory signal noise on motion resolution. *Med Phys*. 2018;45:3205–13.
27. Bellinge JW, Francis RJ, Majeed K, Watts GF, Schultz CJ. In search of the vulnerable patient or the vulnerable plaque: 18 F-sodium fluoride positron emission tomography for cardiovascular risk stratification. *J Nucl Cardiol*. 2018;25:1774–83. <https://doi.org/10.1007/s12350-018-1360-2>.
28. Rasmussen JH, Fischer BM, Aznar MC, Hansen AE, Vogelius IR, Ofgren JL, et al. Reproducibility of ¹⁸F-FDG PET uptake measurements in head and neck squamous cell carcinoma on both PET/CT and PET/MR. *Br J Radiol*. 2015;88:15–21.
29. Tzolos E, Lassen ML, Pan T, Kwiecinski J, Cadet S, Dey D, et al. Respiration-averaged CT versus standard CT attenuation map for correction of 18F-sodium fluoride uptake in coronary atherosclerotic lesions on hybrid PET/CT. *J Nucl Cardiol*. 2020. <https://doi.org/10.1007/s12350-020-02245-7>.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Submit your manuscript to a SpringerOpen[®] journal and benefit from:

- Convenient online submission
- Rigorous peer review
- Open access: articles freely available online
- High visibility within the field
- Retaining the copyright to your article

Submit your next manuscript at ► [springeropen.com](https://www.springeropen.com)
