

Prospective Multicenter Trial of Cervical Arthroplasty With the ROTAIO[®] Cervical Disc Prosthesis

Steffen Fleck (✉ sfleck@uni-greifswald.de)

University Medicine Greifswald <https://orcid.org/0000-0002-2465-2893>

Anna Maria Lang

Medical University Innsbruck

Jens Lehmborg

Klinikum Bogenhausen

Julia Fee Landscheidt

Klinikum Bogenhausen

Rüdiger Gerlach

University Medicine Erfurt

Julian Rathert

University Medicine Erfurt

Ralph T. Schär

Bern University Hospital, University of Bern

Christian Ulrich

Lindenhofspital Bern

Jan-Uwe Müller

University Medicine Greifswald

Sebastian Hartmann

Medical University Innsbruck

Claudius Thomé

Medical University Innsbruck

Research Article

Keywords: cervical disc prosthesis, cervical arthroplasty, outcome, complications, adjacent segment disease

Posted Date: April 23rd, 2021

DOI: <https://doi.org/10.21203/rs.3.rs-450035/v1>

License:  This work is licensed under a Creative Commons Attribution 4.0 International License. [Read Full License](#)

Abstract

Objective

Anterior Cervical Discectomy and Arthroplasty (ACDA) is an established treatment for degenerative cervical disc disease and seems to be an alternative to fusion in minimizing the risk of Adjacent Segment Disease (ASD). The ROTAIO® cervical disc prosthesis is a novel unconstrained implant with a variable center of rotation aiming at physiological motion. The objective of this multicenter prospective trial was to evaluate clinical outcome and complications within 2 years.

Material and Methods

120 patients (72 females and 48 males with a median age of 43.0 years; range: 23 to 60 years) underwent ACDA (ROTAIO®, SIGNUS Medical, Alzenau, Germany) and were prospectively followed for 24 months. Preoperative complaints were mainly associated with radiculopathy (n=104) or myelopathy (n=16). There were 108 monosegmental and 12 bisegmental procedures including 6 hybrid constructs.

Clinical outcome was evaluated at 3, 12 and 24 months by the Visual Analogue Scale (VAS) for head, neck and arm pain, the Neck Disability Index (NDI), the Work Limitation Questionnaire (WL-26), the Patient Satisfaction Index (PSI) and a Quality of Life Questionnaire (SF-36). The Nurick Score, the Modified Japanese Orthopaedic Association Score overall satisfaction and the amount of analgesics were assessed.

Results

Highly significant clinical improvements were observed according to NDI and VAS (p<0.0001 (arm); p<0.001 (neck); p=0.002 (head)) at all postoperative time points. Analgetic medication could be reduced after 3 months in 91.3%, after 12 months in 87.1% and after 24 months in 95.2% of patients. Doctor's visits for cervical spine problems have been reduced in 93.8% after 24 months.

Patient's overall satisfaction was high after 3, 12 and 24 months with 83.5%, 78.4% and 79.1% of patients, while 4.1%, 6.8% and 7.0% respectively were not satisfied. The composite success rate was 77.5% after 12 months and 76.9% after 24 months. There were no major complications in this series. Slight subsidence of the prosthesis was observed in 2 patients and 3 patients demonstrated fusion after 24 months. 2 patients developed symptomatic foraminal stenosis, so that implant removal and fusion was performed.

Conclusion

The ROTAIO® cervical disc prosthesis is a safe and efficient treatment option for symptomatic degenerative disc disease demonstrating excellent clinical results at 2 years. Outcome proves to be stable over time with very low revision rates.

Purpose

Since its introduction in the 1950s (Smith & Robinson, 1958) Anterior Cervical Discectomy and Fusion (ACDF) has become a standard surgical procedure for the treatment of cervical disc disease in patients with radiculopathy or myelopathy. ACDF is performed to achieve neural decompression, segmental stabilization and to maintain cervical lordosis. ACDF yields good clinical outcome and high fusion rates (Fraser & Hartl, 2007) (Shriver et al., 2015). However, fusion sacrifices the mobility of the operated segment leading to increased biomechanical forces at the level of the non-fused adjacent segments and this may accelerate Adjacent Segment Disease (ASD) (Anderson et al., 2012a; Reitman et al., 2004). Anterior Cervical Discectomy and Arthroplasty (ACDA) has been introduced as an alternative that preserves segmental mobility of the operated segment aiming at decreasing the risk of ASD. Although recent randomized clinical trials (Coric et al., 2011; Heller et al., 2009; Mummaneni et al., 2007; Murrey et al., 2009; Sasso & Best, 2008) have compared both techniques, this issue is still controversially debated (Anderson et al., 2012a; Fallah et al., 2012; Shim et al., 2006; Song et al., 2011) (Bartels et al., 2010) and mid- to long-term results are sparse. Currently, there are many disc prostheses with different biomechanical properties commercially available, which claim to imitate physiologic motion. Standard ball and socket designs and (semi-) constrained devices, however, don't allow uncoupled translation and are thus thought to force the facet joints into non-physiologic movements. As this may interfere with successful outcome over time, our group focused on an unconstrained disc prosthesis with uncoupled translation. (Kang et al., 2010; Rousseau et al., 2008).

The aim of this multicenter, multinational prospective observational study was therefore to evaluate clinical outcome and complications with the newly developed artificial cervical disc prosthesis ROTAIO® (SIGNUS Medizintechnik GmbH, Alzenau, Germany) within a follow-up period of 24 months.

Materials And Methods

The study complied with the Declaration of Helsinki and was approved by all local ethics committees of the involved centers (initial approval: Ethics Commission, Medical University Innsbruck, Austria). All patients gave written informed consent.

Patient population

Consecutive patients meeting the inclusion criteria were prospectively included from July 2014 to January 2019.

Inclusion criteria:

Loading [MathJax]/jax/output/CommonHTML/jax.js

- (2) Degenerative Disc Disease (DDD) at one or two levels between C3/4 to C6/7 and a disc height of at least 50% in comparison to other segments,
- (3) no improvement of clinical symptoms after at least 6 weeks of conservative treatment or significant or progressive neurological deficits at the time of presentation,
- (4) a minimum NDI of 15 points (30%).

Exclusion criteria:

1. cervical instability demonstrated on flexion / extension radiographs
2. kyphotic index segment
3. non-mobile index level and/or increased osteochondrosis in index disc / facet joints
4. previous surgery of the cervical spine

Device Design

The ROTAIO® cervical disc prosthesis (Fig.1) is used to replace the disc after anterior cervical discectomy. The aim of disc replacement is to restore disc height and to maintain physiological motion of the index segment. The ROTAIO® prosthesis is a new unconstrained implant with a variable center of rotation (COR) and uncoupled translation to enable physiological facet joint-guided motion. The prosthesis consists of a superior and an inferior end plate (Titanium alloy to ISO 5832-3), on which the sliding elements (Cobalt-Chrome alloy to ISO 5832-12) are anchored and secured by means of a fixation pin. Primary stabilization is achieved via toothed surfaces of the end plates, which are additionally blasted to increase surface area and allow rapid bony integration.

Surgical procedure

Surgery was performed by board-certified neurosurgeons who had attended manufacturer *s* \in *struction* \leq *ctures*. A standard anterior or cervical approach with discectomy \rightarrow mywasused \in *clud* \in *g* microsurgical resection of the posterior or lateral *s* instructions during slight distraction of the intervertebral space.

Clinical evaluation

All patients underwent clinical examination on the day before surgery, postoperatively prior to discharge, and 3, 12 and 24 months after surgery. In addition to a clinical examination including neurological status validated self-assessment outcome measures were used: Visual Analogue Scale (VAS) (range 0-10) for head, neck and arm pain separately, Patient's Satisfaction Index (PSI), Neck Disability Index (NDI) (range 0-50), Work Limitation Questionnaire (WL-26), and Quality of Life Questionnaire (SF-36) (MacDermid et al., 2009) (Arumugam et al., 2013). The Nurick-Score for classification of gait disturbance and the Modified Japanese Orthopaedic Association Score (mJOA) were recorded. Complications related to the implant, fusion of the index level and surgical procedures at the index level (revision surgery) or at an adjacent level (for ASD) were recorded. A Composite Success Rate was defined as the combination of (1) improvement in NDI ($\geq 15\%$), (2) stable or improved neurological status, (3) no secondary operation, (4) no implant-associated complication.

Imaging protocol

Plain anteroposterior and lateral radiographs in neutral position and lateral radiographs in flexion and extension were obtained in each patient. Furthermore, all patients underwent magnetic resonance imaging before surgery. Computed tomography was additionally applied at the discretion of the surgeon.

Statistical Analysis

Data are presented as the mean value \pm standard deviation (SD) of the mean.

The trial was designed to detect an absolute difference between pre- and postoperative data. Comparisons were performed with the use of an unpaired t-test, or in case of nonparametric values with the Wilcoxon-Mann-Whitney-U test. Furthermore, χ^2 was used for differentiation of categories.

Differences were considered statistically significant for two-tailed p-values < 0.05 (power of 80%, maximum dropout rate of 20%). Data analysis was performed with IBM SPSS Statistics Vers. 20.0.

Results

Demographic Data

120 patients (60% female, 40% male) with a median age of 43.0 years (range: 23 to 60 years) were included. The majority of patients underwent ACDA at the C5/6 (58.8%) and the C6/7 (49.6%) levels. Preoperative complaints were mainly associated with radiculopathy (n=104; 86.7%) or myelopathy (n=16; 13.3%). There were 108 monosegmental and 12 bisegmental procedures including 6 hybrid constructs comprised of ACDF with cage fusion and ACDA with the ROTAIO® disc prosthesis. Minimum follow-up of 12 months was available in 115 patients and 92 patients had completed 2-year follow-up.

Clinical evaluation

The median duration of symptoms amounted to 5 months. Demographics and preoperative symptoms are listed in table 1. There were some differences

Loading [MathJax]/jax/output/CommonHTML/jax.js ic (MP) patients.

Highly significant clinical improvements were observed for **VAS** arm ($p < 0.0001$), neck ($p < 0.001$) and head ($p = 0.0022$) at all time points. Significant **functional improvements** were observed with a $\geq 15\%$ decrease in **NDI** in $\geq 90\%$ at 12 and 24 months after surgery. Differences of RP and MP are noticed in table 3.

The neurological status according to muscle **strength** and **mJOA** remained stable or improved in 92.5% and 96.9% respectively. The **composite success rate** was 77.5% after 12 months and 76.9% after 24 months.

Analgesic medication could be reduced after 3 months in 91.3%, after 12 months in 87.1% and after 24 months in 95.2% of patients, but increased in 8.8%, 12.9% and 4.8%, respectively. Preoperatively, the analgesics consumption was significantly higher in the RP group than in the MP group (see table 1). As expected, improvement and success rate tended to be more pronounced in the RP group.

The **WL-26** clearly demonstrated a reduction of work limitations ($p < 0.0001$ at 3, 12 and 24 months). Health-related Quality of Life (**SF-36**) revealed a highly significant improvement ($p < 0.0001$) for the following items within 3, 12 and 24 months: body function, social function, psychologic well-being, physical pain, vitality, and overall health perception.

Patient's overall satisfaction was high after 3, 12 and 24 months with 83.5%, 78.4% and 79.1% of patients, while 4.1%, 6.8% and 7.0% respectively were not satisfied. **Doctor's visits** for cervical spine problems have been reduced in 93.8% after 24 months and increased in 6.2%.

Complications

There were no major complications in this series. Temporary morbidity related to the anterior cervical approach but not the implant per se, like recurrent nerve palsy and significant dysphagia, occurred in 2 patients. Initially, 2 patients experienced intermittent and transient cracking noises. Slight subsidence of the prosthesis was observed in 2 patients and 3 patients demonstrated fusion after 24 months. 2 patients developed clinical problems associated with foraminal stenosis after 3 and 9 months, respectively, so that implant removal and fusion was performed. Revision rate thus amounted to 2% at the index level and no procedures at the adjacent levels within 2 years.

Discussion

Anterior Cervical Discectomy and Fusion (ACDF) is a standard procedure for the treatment of degenerative cervical disc disease. Based on the notion that preserving motion reduces the risk of Adjacent Segment Disease (ASD) (Chang et al., 2007; Robertson et al., 2005), Anterior Cervical Discectomy with Arthroplasty (ACDA) has been introduced as an alternative to fusion in the 1990's. Although clinical outcome is well documented for both techniques (Anderson et al., 2012a; Fallah et al., 2012), some patients will experience persistent or increasing symptoms over time.

Adjacent Segment Disease

Despite its well documented benefits, ACDF may cause ASD in mid- and long-term follow-up (Goffin et al., 2003; Ishihara et al., 2004; Katsuura et al., 2001; Kim et al., 2009; Song et al., 2011; Yao et al., 2015). Biomechanical studies have shown increased intradiscal stress and motion compensation in the levels adjacent to the fusion site (Eck et al., 2002; Schwab et al., 2006) with a change of the center of rotation in adjacent levels postoperatively (Dvorak et al., 1993) (Eck et al., 2002) (Anderson et al., 2012b). Although this is considered by some authors to be the underlying cause for ASD, it is still controversial if this is attributable to the biomechanical effects of fusion or to the natural history of cervical degeneration (Anderson et al., 2012a; Bartels et al., 2010; Seo & Choi, 2008).

Reoperation due to ASD has been documented at a rate of 2.9% annually after ACDF (Hilibrand et al., 1999). ACDA is considered as an alternative to ACDF preserving normal cervical kinetics and biomechanics (Goffin et al., 2002). Thus, the rate of additional surgeries may be reduced with less stress on adjacent levels using ACDA (Upadhyaya et al., 2012) (Traynelis, 2006). The pooled surgery rate for ASD after disc prosthesis (ACDA) was 3.8% (0.9 -7.6%) within a follow-up of up to 84 months summarizing 13 randomized controlled trials (RCT) (Kang et al., 2015). Although clinical short-term results are satisfactory (Goffin et al., 2003; Heller et al., 2009; Murrey et al., 2009), there are only a few studies reporting mid- to long-term results (Ding et al., 2012; Garrido et al., 2010; Goffin et al., 2010; Walraevens et al., 2010). Garrido et al (Garrido et al., 2010) reported improved functional outcome for ACDF and ACDA on 24 and 48 months follow-up with no degradation of the outcome measures between 2 and 4 years after surgery. This is in concordance with the results of Goffin et al. (Goffin et al., 2010), who reported consistent if not improved clinical results at 4- and 6-years follow-up compared to the 1- and 2-years postoperative results. Our study supports these data that patients improved significantly after surgery and the clinical results remained stable on mid-term follow-up.

The protective effect of ACDA on the adjacent discs remains controversial. In the single level arm of their prospective cohort study, Kim et al. (Kim et al., 2009) observed ASD in 13% of all patients treated with ACDA compared to 23% in the ACDF group at a median follow-up of 19 month. Walraevens et al. reported ASD in the adjacent upper and lower segment to the operated site for up to 8 years after ACDA (Walraevens et al., 2010). Similar observations were made by Ding et al. (Ding et al., 2012). They observed mild ASD in the adjacent levels in approx. 23% of all patients. The degeneration mainly manifested as new formation or enlargement of an anterior osteophyte. However, no degeneration in clinical outcome occurred due to the lack of a direct relation between radiographic and clinical ASD (Ding et al., 2012). In our cohort with the new ROTAIO® prosthesis, no patient required adjacent level surgery within 2 years.

Quality of motion

After ACDA emphasis is often placed on presence and magnitude of motion as assessed by ROM, while quality of motion by parameters like instantaneous COR, COR, and instantaneous axis of rotation has just recently been identified as important for evaluating changes in the cervical motion pattern (Anderst et al., 2013; Guo et al., 2019; Jonas et al., 2018; Penning, 1988). Anderst et al. demonstrated that the instantaneous COR was generally fixed in the longitudinal direction, but it translated in the anterior-posterior direction during flexion-extension (Anderst et al., 2013). If translation is not adequately possible, non-

physiologic stress on the facet joints at the index level ensues, which may cause facet joint syndrome, as it has commonly been seen in lumbar disc arthroplasty. Liu et al. evaluated the instantaneous COR located at the superior half of the lower vertebral body height and the posterior half of its width, and changing with age (Liu et al., 2014). It has been postulated that these further findings should be considered in clinical practice and when designing disc prostheses (Guo et al., 2019).

Although the overall effectiveness of ACDA has already been demonstrated, the kinematic properties of the various designs differ substantially (Galbusera, Bellini, Brayda-Bruno, et al., 2008; Kang et al., 2010). The Bryan disc prosthesis with its almost unconstrained design retained kinematic motion adequately (Pickett et al., 2005; Powell et al., 2010; Ryu et al., 2013), (Fleck et al., 2017) (Kowalczyk et al., 2011), yielding a near-physiological rotation at the index level (Galbusera, Bellini, Raimondi, et al., 2008). Ball-and socket designs like the Prestige LP (semiconstrained design) and the Prodisc-C (semiconstrained with fixed axis of rotation), however, did not fully restore normal mobility in view of ROM and COR, which may cause secondary problems over time (Rousseau et al., 2008). Particularly neck pain can be an ongoing problem after ACDA as a result of abnormal forces and load on the facet joints.

As standard ball and socket designs and (semi-) constrained devices do not allow uncoupled translation and are thus thought to force the facet joints into non-physiologic movements as mentioned above. As this may interfere with successful outcome over time, our group focused on an unconstrained disc prosthesis with uncoupled translation. The low revision rate and the stable clinical results over time in this series seem to support these considerations.

Clinical Outcome

In 2016, preliminary clinical and radiographic results with the ROTAIO® cervical prosthesis demonstrated excellent results (Oberbauer et al., 2016). Our present results with more than 100 patients in a multicenter prospective trial confirm these findings with excellent clinical outcome. Pain relief, reduction of analgesics consumption, functional improvement, reduction of disability, patient satisfaction and quality of life were found to be very high and at least comparable to previous IDE trials. Revision rate was very low and no implant failure was observed. No surgical procedure due to ASD was performed within 2 years.

Nevertheless, longer follow-up is necessary to prove durability and functionality of the prosthesis. In view of our current data, however, the ROTAIO® prosthesis is a suitable alternative to ACDF and other available prostheses. The particular biomechanical characteristics with uncoupled translation and a variable center of rotation may allow physiological cervical spine motion with low fusion and ASD rates.

Limitations

This prospective observational multicenter study of consecutive patients has received research support by the manufacturer, although clinical data was assessed and analyzed by the investigators. The study was not intended to compare the ROTAIO® results to ACDF or other prostheses. Follow-up is currently limited to 2 years, so that long-term sequelae cannot yet be adequately monitored.

Conclusion

The ROTAIO® cervical disc prosthesis with its unconstrained design with uncoupled translation and variable center of rotation is a safe and effective treatment option for symptomatic degenerative disc disease. Good to excellent clinical results and very low secondary surgery rates after a follow-up of 24 months could be demonstrated in this prospective, observational study.

Abbreviations

ACDA	anterior cervical discectomy and arthroplasty
ACDF	anterior cervical decompression and fusion
ASD	adjacent segment disease
COR	center of rotation
DDD	degenerative disc disease
FSU	functional spinal unit
MP	myelopathy
NDI	neck disability index
n.s.	not significant
NSAID	non-steroidal anti-inflammatory drugs
PSI	patient`s satisfaction index
RCT	randomized clinical trial
ROI	region of interest

Loading [MathJax]/jax/output/CommonHTML/jax.js

RP	radiculopathy
SF-36	short form -36
VAS	visual analogue scale
WL-26	work limitation
QoL	quality of life

Declarations

Acknowledgement:

We thank for statistical analysis Gerd Schöneberg and Jörg Schnitker, Institute of Applied Statistics, Bielefeld, Germany.

Disclosure

The authors' institutions have received funding to conduct this trial. CT, CU, RG and JUM are consultants to SIGNUS Medizintechnik GmbH.

Publishing: The manuscript has not been submitted to another journal for simultaneous consideration.

Funding: The authors' institutions have received funding to conduct this trial from SIGNUS Medizintechnik, Alzenau, Germany.

Disclosure: C.Thomé, C. Ulrich, R. Gerlach and JU. Müller are consultants to SIGNUS Medizintechnik GmbH.

Ethics: The study complied with the Declaration of Helsinki and was approved by all local ethics committees of the involved centers (initial approval: Ethics Commission, Medical University Innsbruck, Austria). All patients gave written informed consent.

Data transparency: There is availability of data and material on demand.

Code availability: Software application or custom code: not applicable.

Author's contribution: All authors meet all 3 of the following conditions in accordance with the "Consensus Statement on Surgery Journals Authorship—2005":

1) Authors make substantial contributions to conception and design, and/or acquisition of data, and/or analysis and interpretation of data;

2) Authors participate in drafting the article or revising it critically for important intellectual content; and

3) Authors give final approval of the version to be submitted and any revised version. **Consent for publication:** The authors transfer to Springer (respective to owner if other than Springer and for U.S. government employees: to the extent transferable) the non-exclusive publication rights and he warrants that his/her contribution is original and that he/she has full power to make this grant. The author signs for and accepts responsibility for releasing this material on behalf of any and all co-authors. This transfer of publication rights covers the non-exclusive right to reproduce and distribute the article, including reprints, translations, photographic reproductions, microform, electronic form (offline, online) or any other reproductions of similar nature. The author may self-archive an author-created version of his article on his own website and his institution's repository, including his final version; however he may not use Springer's PDF version which is posted on www.springerlink.com. Furthermore, the author may only post his version provided acknowledgement is given to the Journal and Springer as one of the original places of publication and a link is inserted to the published article on Springer's website.

For the authors

Steffen Fleck

References

1. Anderson, P. A., Sasso, R. C., Hipp, J., Norvell, D., Raich, A., & Hashimoto, R. (2012a). kinematics of the Cervical Adjacent Segments Following Disc Arthroplasty and Anterior Discectomy and Fusion: A Systematic Review And Meta-Analysis. *Spine (Phila Pa 1976)*. <https://doi.org/10.1097/BRS.0b013e31826d6628>
2. Anderson, P. A., Sasso, R. C., Hipp, J., Norvell, D. C., Raich, A., & Hashimoto, R. (2012b). Kinematics of the cervical adjacent segments after disc arthroplasty compared with anterior discectomy and fusion: a systematic review and meta-analysis. *Spine (Phila Pa 1976)*, 37(22 Suppl), S85-95. <https://doi.org/10.1097/BRS.0b013e31826d6628>
3. Anderst, W., Baillargeon, E., Donaldson, W., Lee, J., & Kang, J. (2013). Motion path of the instant center of rotation in the cervical spine during in vivo dynamic flexion-extension: implications for artificial disc design and evaluation of motion quality after arthrodesis. *Spine (Phila Pa 1976)*, 38(10), E594-601. <https://doi.org/10.1097/BRS.0b013e31828ca5c7>
4. Arumugam, V., Macdermid, J. C., & Grewal, R. (2013). Content analysis of work limitation, stanford presenteeism, and work instability questionnaires using international classification of functioning, disability, and health and item perspective framework. *Rehabil Res Pract*, 2013, 614825. <https://doi.org/10.1155/2013/614825>

5. Bartels, R. H., Donk, R., & Verbeek, A. L. (2010). No justification for cervical disk prostheses in clinical practice: a meta-analysis of randomized controlled trials. *Neurosurgery*, 66(6), 1153-1160; discussion 1160. <https://doi.org/10.1227/01.neu.0000369189.09182.5f>
6. Chang, U. K., Kim, D. H., Lee, M. C., Willenberg, R., Kim, S. H., & Lim, J. (2007). Changes in adjacent-level disc pressure and facet joint force after cervical arthroplasty compared with cervical discectomy and fusion. *J Neurosurg Spine*, 7(1), 33-39. http://www.ncbi.nlm.nih.gov/entrez/query.fcgi?cmd=Retrieve&db=PubMed&dopt=Citation&list_uids=17633485
7. Coric, D., Nunley, P. D., Guyer, R. D., Musante, D., Carmody, C. N., Gordon, C. R., Laurysen, C., Ohnmeiss, D. D., & Boltes, M. O. (2011). Prospective, randomized, multicenter study of cervical arthroplasty: 269 patients from the Kineflex|C artificial disc investigational device exemption study with a minimum 2-year follow-up: clinical article. *J Neurosurg Spine*, 15(4), 348-358. <https://doi.org/10.3171/2011.5.SPINE10769>
8. Ding, C., Hong, Y., Liu, H., Shi, R., Hu, T., & Li, T. (2012). Intermediate clinical outcome of Bryan Cervical Disc replacement for degenerative disk disease and its effect on adjacent segment disks. *Orthopedics*, 35(6), e909-916. <https://doi.org/10.3928/01477447-20120525-33>
9. Dvorak, J., Panjabi, M. M., Grob, D., Novotny, J. E., & Antinnes, J. A. (1993). Clinical validation of functional flexion/extension radiographs of the cervical spine. *Spine (Phila Pa 1976)*, 18(1), 120-127. <https://doi.org/10.1097/00007632-199301000-00018>
10. Eck, J. C., Humphreys, S. C., Lim, T. H., Jeong, S. T., Kim, J. G., Hodges, S. D., & An, H. S. (2002). Biomechanical study on the effect of cervical spine fusion on adjacent-level intradiscal pressure and segmental motion. *Spine (Phila Pa 1976)*, 27(22), 2431-2434. <https://doi.org/10.1097/00007632-200211150-00003>
11. Fallah, A., Akl, E. A., Ebrahim, S., Ibrahim, G. M., Mansouri, A., Foote, C. J., Zhang, Y., & Fehlings, M. G. (2012). Anterior cervical discectomy with arthroplasty versus arthrodesis for single-level cervical spondylosis: a systematic review and meta-analysis. *PLoS One*, 7(8), e43407. <https://doi.org/10.1371/journal.pone.0043407>
12. Fleck, S. K., Langner, S., Rosenstengel, C., Kessler, R., Matthes, M., Muller, J. U., Langner, I., Marx, S., & Schroeder, H. W. (2017). 3-Tesla Kinematic MRI of the Cervical Spine for Evaluation of Adjacent Level Disease After Monosegmental Anterior Cervical Discectomy and Arthroplasty: Results of 2-Year Follow-Up. *Spine (Phila Pa 1976)*, 42(4), 224-231. <https://doi.org/10.1097/BRS.0000000000001705>
13. Fraser, J. F., & Hartl, R. (2007). Anterior approaches to fusion of the cervical spine: a metaanalysis of fusion rates. *J Neurosurg Spine*, 6(4), 298-303. <https://doi.org/10.3171/spi.2007.6.4.2>
14. Galbusera, F., Bellini, C. M., Brayda-Bruno, M., & Fornari, M. (2008). Biomechanical studies on cervical total disc arthroplasty: a literature review. *Clin Biomech (Bristol, Avon)*, 23(9), 1095-1104. <https://doi.org/10.1016/j.clinbiomech.2008.06.002>
15. Galbusera, F., Bellini, C. M., Raimondi, M. T., Fornari, M., & Assietti, R. (2008). Cervical spine biomechanics following implantation of a disc prosthesis. *Med Eng Phys*, 30(9), 1127-1133. <https://doi.org/10.1016/j.medengphy.2008.02.002>
16. Garrido, B. J., Taha, T. A., & Sasso, R. C. (2010). Clinical outcomes of Bryan cervical disc arthroplasty a prospective, randomized, controlled, single site trial with 48-month follow-up. *J Spinal Disord Tech*, 23(6), 367-371. <https://doi.org/10.1097/BSD.0b013e3181bb8568>
17. Goffin, J., Casey, A., Kehr, P., Liebig, K., Lind, B., Logroscino, C., Pointillart, V., Van Calenbergh, F., & van Loon, J. (2002). Preliminary clinical experience with the Bryan Cervical Disc Prosthesis. *Neurosurgery*, 51(3), 840-845; discussion 845-847.
18. Goffin, J., Van Calenbergh, F., van Loon, J., Casey, A., Kehr, P., Liebig, K., Lind, B., Logroscino, C., Sgrambiglia, R., & Pointillart, V. (2003). Intermediate follow-up after treatment of degenerative disc disease with the Bryan Cervical Disc Prosthesis: single-level and bi-level. *Spine (Phila Pa 1976)*, 28(24), 2673-2678. http://www.ncbi.nlm.nih.gov/entrez/query.fcgi?cmd=Retrieve&db=PubMed&dopt=Citation&list_uids=14673368
19. Goffin, J., van Loon, J., Van Calenbergh, F., & Lipscomb, B. (2010). A clinical analysis of 4- and 6-year follow-up results after cervical disc replacement surgery using the Bryan Cervical Disc Prosthesis. *J Neurosurg Spine*, 12(3), 261-269. <https://doi.org/10.3171/2009.9.spine09129>
20. Guo, Z., Cui, W., Sang, D. C., Sang, H. P., & Liu, B. G. (2019). Clinical Relevance of Cervical Kinematic Quality Parameters in Planar Movement. *Orthop Surg*, 11(2), 167-175. <https://doi.org/10.1111/os.12435>
21. Heller, J. G., Sasso, R. C., Papadopoulos, S. M., Anderson, P. A., Fessler, R. G., Hacker, R. J., Coric, D., Cauthen, J. C., & Riew, D. K. (2009). Comparison of BRYAN cervical disc arthroplasty with anterior cervical decompression and fusion: clinical and radiographic results of a randomized, controlled, clinical trial. *Spine (Phila Pa 1976)*, 34(2), 101-107. <https://doi.org/10.1097/BRS.0b013e31818ee263>
22. Hilibrand, A. S., Carlson, G. D., Palumbo, M. A., Jones, P. K., & Bohlman, H. H. (1999). Radiculopathy and myelopathy at segments adjacent to the site of a previous anterior cervical arthrodesis. *J Bone Joint Surg Am*, 81(4), 519-528. <https://doi.org/10.2106/00004623-199904000-00009>
23. Ishihara, H., Kanamori, M., Kawaguchi, Y., Nakamura, H., & Kimura, T. (2004). Adjacent segment disease after anterior cervical interbody fusion. *Spine J*, 4(6), 624-628. http://www.ncbi.nlm.nih.gov/entrez/query.fcgi?cmd=Retrieve&db=PubMed&dopt=Citation&list_uids=15541693
24. Jonas, R., Demmelmaier, R., Hacker, S. P., & Wilke, H. J. (2018). Comparison of three-dimensional helical axes of the cervical spine between in vitro and in vivo testing. *Spine J*, 18(3), 515-524. <https://doi.org/10.1016/j.spinee.2017.10.065>
25. Kang, H., Park, P., La Marca, F., Hollister, S. J., & Lin, C. Y. (2010). Analysis of load sharing on uncovertebral and facet joints at the C5-6 level with implantation of the Bryan, Prestige LP, or ProDisc-C cervical disc prosthesis: an in vivo image-based finite element study. *Neurosurg Focus*, 28(6), E9. <https://doi.org/10.3171/2010.3.FOCUS1046>
26. Kang, J., Shi, C., Gu, Y., Yang, C., & Gao, R. (2015). Factors that may affect outcome in cervical artificial disc replacement: a systematic review. *Eur Spine J*, 24(9), 2023-2032. <https://doi.org/10.1007/s00586-015-4096-6>
27. Katsuura, A., Hukuda, S., Saruhashi, Y., & Mori, K. (2001). Kyphotic malalignment after anterior cervical fusion is one of the factors promoting the degenerative process in adjacent intervertebral levels. *Eur Spine J*, 10(4), 320-324. http://www.ncbi.nlm.nih.gov/entrez/query.fcgi?cmd=Retrieve&db=PubMed&dopt=Citation&list_uids=11563618

28. Kim, S. W., Limson, M. A., Kim, S. B., Arbatin, J. J., Chang, K. Y., Park, M. S., Shin, J. H., & Ju, Y. S. (2009). Comparison of radiographic changes after ACDF versus Bryan disc arthroplasty in single and bi-level cases. *Eur Spine J*, *18*(2), 218-231. http://www.ncbi.nlm.nih.gov/entrez/query.fcgi?cmd=Retrieve&db=PubMed&dopt=Citation&list_uids=19127374
29. Kowalczyk, I., Lazaro, B. C., Fink, M., Rabin, D., & Duggal, N. (2011). Analysis of in vivo kinematics of 3 different cervical devices: Bryan disc, ProDisc-C, and Prestige LP disc. *J Neurosurg Spine*, *15*(6), 630-635. <https://doi.org/10.3171/2011.8.SPINE11273>
30. Liu, B., Liu, Z., VanHoof, T., Kalala, J., Zeng, Z., & Lin, X. (2014). Kinematic study of the relation between the instantaneous center of rotation and degenerative changes in the cervical intervertebral disc. *Eur Spine J*, *23*(11), 2307-2313. <https://doi.org/10.1007/s00586-014-3431-7>
31. MacDermid, J. C., Walton, D. M., Avery, S., Blanchard, A., Etruw, E., McAlpine, C., & Goldsmith, C. H. (2009). Measurement properties of the neck disability index: a systematic review. *J Orthop Sports Phys Ther*, *39*(5), 400-417. <https://doi.org/10.2519/jospt.2009.2930>
32. Mummaneni, P. V., Burkus, J. K., Haid, R. W., Traynelis, V. C., & Zdeblick, T. A. (2007). Clinical and radiographic analysis of cervical disc arthroplasty compared with allograft fusion: a randomized controlled clinical trial. *J Neurosurg Spine*, *6*(3), 198-209. <https://doi.org/10.3171/spi.2007.6.3.198>
33. Murrey, D., Janssen, M., Delamarter, R., Goldstein, J., Zigler, J., Tay, B., & Darden, B. (2009). Results of the prospective, randomized, controlled multicenter Food and Drug Administration investigational device exemption study of the ProDisc-C total disc replacement versus anterior discectomy and fusion for the treatment of 1-level symptomatic cervical disc disease. *Spine J*, *9*(4), 275-286. <https://doi.org/10.1016/j.spinee.2008.05.006>
34. Obernauer, J., Landscheidt, J., Hartmann, S., Schubert, G. A., Thome, C., & Lumenta, C. (2016). Cervical arthroplasty with ROTAIO(R) cervical disc prosthesis: first clinical and radiographic outcome analysis in a multicenter prospective trial. *BMC Musculoskelet Disord*, *17*, 11. <https://doi.org/10.1186/s12891-016-0880-7>
35. Penning, L. (1988). Differences in anatomy, motion, development and aging of the upper and lower cervical disk segments. *Clin Biomech (Bristol, Avon)*, *3*(1), 37-47. [https://doi.org/10.1016/0268-0033\(88\)90124-6](https://doi.org/10.1016/0268-0033(88)90124-6)
36. Pickett, G. E., Rouleau, J. P., & Duggal, N. (2005). Kinematic analysis of the cervical spine following implantation of an artificial cervical disc. *Spine (Phila Pa 1976)*, *30*(17), 1949-1954. <https://doi.org/10.1097/01.brs.0000176320.82079.ce>
37. Powell, J. W., Sasso, R. C., Metcalf, N. H., Anderson, P. A., & Hipp, J. A. (2010). Quality of spinal motion with cervical disk arthroplasty: computer-aided radiographic analysis. *J Spinal Disord Tech*, *23*(2), 89-95. <https://doi.org/10.1097/BSD.0b013e3181991413>
38. Reitman, C. A., Hipp, J. A., Nguyen, L., & Esses, S. I. (2004). Changes in segmental intervertebral motion adjacent to cervical arthrodesis: a prospective study. *Spine (Phila Pa 1976)*, *29*(11), E221-226.
39. Robertson, J. T., Papadopoulos, S. M., & Traynelis, V. C. (2005). Assessment of adjacent-segment disease in patients treated with cervical fusion or arthroplasty: a prospective 2-year study. *J Neurosurg Spine*, *3*(6), 417-423. http://www.ncbi.nlm.nih.gov/entrez/query.fcgi?cmd=Retrieve&db=PubMed&dopt=Citation&list_uids=16381202
40. Rousseau, M. A., Cottin, P., Levante, S., Nogier, A., Lazennec, J. Y., & Skalli, W. (2008). In vivo kinematics of two types of ball-and-socket cervical disc replacements in the sagittal plane: cranial versus caudal geometric center. *Spine (Phila Pa 1976)*, *33*(1), E6-9. <https://doi.org/10.1097/BRS.0b013e31815e5dce>
41. Ryu, W. H., Kowalczyk, I., & Duggal, N. (2013). Long-term kinematic analysis of cervical spine after single-level implantation of Bryan cervical disc prosthesis. *Spine J*, *13*(6), 628-634. <https://doi.org/10.1016/j.spinee.2013.02.046>
42. Sasso, R. C., & Best, N. M. (2008). Cervical kinematics after fusion and bryan disc arthroplasty. *J Spinal Disord Tech*, *21*(1), 19-22. http://www.ncbi.nlm.nih.gov/entrez/query.fcgi?cmd=Retrieve&db=PubMed&dopt=Citation&list_uids=18418131
43. Schwab, J. S., Diangelo, D. J., & Foley, K. T. (2006). Motion compensation associated with single-level cervical fusion: where does the lost motion go? *Spine (Phila Pa 1976)*, *31*(21), 2439-2448. <https://doi.org/10.1097/01.brs.0000239125.54761.23>
44. Seo, M., & Choi, D. (2008). Adjacent segment disease after fusion for cervical spondylosis; myth or reality? *Br J Neurosurg*, *22*(2), 195-199. http://www.ncbi.nlm.nih.gov/entrez/query.fcgi?cmd=Retrieve&db=PubMed&dopt=Citation&list_uids=18348013
45. Shim, C. S., Lee, S. H., Park, H. J., Kang, H. S., & Hwang, J. H. (2006). Early clinical and radiologic outcomes of cervical arthroplasty with Bryan Cervical Disc prosthesis. *J Spinal Disord Tech*, *19*(7), 465-470. http://www.ncbi.nlm.nih.gov/entrez/query.fcgi?cmd=Retrieve&db=PubMed&dopt=Citation&list_uids=17021408
46. Shriver, M. F., Lewis, D. J., Kshetry, V. R., Rosenbaum, B. P., Benzel, E. C., & Mroz, T. E. (2015). Pseudoarthrosis rates in anterior cervical discectomy and fusion: a meta-analysis. *Spine J*, *15*(9), 2016-2027. <https://doi.org/10.1016/j.spinee.2015.05.010>
47. Smith, G. W., & Robinson, R. A. (1958). The treatment of certain cervical-spine disorders by anterior removal of the intervertebral disc and interbody fusion. *J Bone Joint Surg Am*, *40-A*(3), 607-624.
48. Song, K. J., Choi, B. W., Jeon, T. S., Lee, K. B., & Chang, H. (2011). Adjacent segment degenerative disease: is it due to disease progression or a fusion-associated phenomenon? Comparison between segments adjacent to the fused and non-fused segments. *Eur Spine J*, *20*(11), 1940-1945. <https://doi.org/10.1007/s00586-011-1864-9>
49. Traynelis, V. C. (2006). Cervical arthroplasty. *Clin Neurosurg*, *53*, 203-207. <https://www.ncbi.nlm.nih.gov/pubmed/17380753>
50. Upadhyaya, C. D., Wu, J. C., Trost, G., Haid, R. W., Traynelis, V. C., Tay, B., Coric, D., & Mummaneni, P. V. (2012). Analysis of the three United States Food and Drug Administration investigational device exemption cervical arthroplasty trials. *J Neurosurg Spine*, *16*(3), 216-228. <https://doi.org/10.3171/2011.6.SPINE10623>
51. Walraevens, J., Demaerel, P., Suetens, P., Van Calenbergh, F., van Loon, J., Vander Sloten, J., & Goffin, J. (2010). Longitudinal prospective long-term radiographic follow-up after treatment of single-level cervical disk disease with the Bryan Cervical Disc. *Neurosurgery*, *67*(3), 679-687; discussion 687. <https://doi.org/10.1227/01.neu.0000377039.89725.f3>

Tables

Table 1: demographics and preoperative symptoms

	%	χ^2 (p)	Radiculopathy (RP)	Myelopathy (MP)	overall
age (median) (yrs)		0.1273	42.7 ± 8.4	39.2 ± 7.5	43 (23-60)
female	60.0				
male	40.0				
leading symptoms (RP vs. MP)			87%	13%	
ASA (II)	80.7				
ASA (III)	19.3				
VAS overall		0.1120	5.9 ± 2.4	4.7 ± 2.0	
daily analgetics	84	0.0479	87%	67%	
doctor`s visits (≥ monthly)	51				
physiotherapy		0.0577	68%	30%	
alcohol		0.0225	51%	11%	
smoking	42				

Table 2: clinical results (focusing on visual analogue scale (VAS) and patient`s satisfaction index (PSI))

	preoperative	3 months	12 months	24 months
VAS overall (0-10)	6.5 ± 2.2	-4.5 ± 2.8	-4.2 ± 3.1	-4.4 ± 2.7
VAS head (0-10)	3.0 ± 2.8	-1.6 ± 2.6	-1.3 ± 2.9	-1.0 ± 3.1
VAS neck (0-10)	5.7 ± 2.4	-3.8 ± 2.5	-3.7 ± 2.9	-3.4 ± 2.7
VAS arm (0-10)	6.1 ± 2.6	-4.4 ± 3.1	-4.1 ± 3.0	-4.3 ± 3.0
PSI (satisfied vs. nonsatisfied)		83.5% vs. 4.1%	78% vs. 7%	79% vs. 7%
	-			

Table 3: functional outcome

	Radiculopathy (RP) (%)	Myelopathy (MP) (%)	RP and MP (%)	Significance (p) (χ^2)
<i>12 months</i>				
NDI (decrease \geq 15%)	92.6	70	90.0	0.02
strength and mJOA (stable or improved)	91.4	100	92.5	0.34
complication	6.3	9.1	6.6	0.75
composite success rate	80	60	77.5	0.16
<i>24 months</i>				
NDI (decrease \geq 15%)	93.0	75.0	90.8	0.10
strength and mJOA (stable or improved)	96.5	100.0	96.9	0.59
complication	8.9	7.7	8.7	0.99
composite success rate	78.9	62.5	76.9	0.30

Figures



Figure 1

The ROTAIO® cervical disc prosthesis allows uncoupled anterior translation upon flexion and posterior translation upon extension mimicking natural disc motion.