



OPEN ACCESS

Open Access
Scan to access more
free content

ORIGINAL ARTICLE

A blinded international study on the reliability of genetic testing for GGGGCC-repeat expansions in *C9orf72* reveals marked differences in results among 14 laboratories

Chizuru Akimoto,¹ Alexander E Volk,² Marka van Blitterswijk,³ Marleen Van den Broeck,^{4,5} Claire S Leblond,⁶ Serge Lumbroso,⁷ William Camu,⁸ Birgit Neitzel,⁹ Osamu Onodera,¹⁰ Wouter van Rheenen,¹¹ Susana Pinto,¹² Markus Weber,¹³ Bradley Smith,¹⁴ Melanie Proven,¹⁵ Kevin Talbot,¹⁶ Pamela Keagle,¹⁷ Alessandra Chesi,¹⁸ Antonia Ratti,^{19,20} Julie van der Zee,^{4,5} Helena Alstermark,¹ Anna Birve,¹ Daniela Calini,^{19,20} Angelica Nordin,¹ Daniela C Tradowsky,² Walter Just,² Hussein Daoud,⁶ Sabrina Angerbauer,⁹ Mariely DeJesus-Hernandez,³ Takuya Konno,¹⁰ Anjali Lloyd-Jani,¹⁵ Mamede de Carvalho,¹² Kevin Mouzat,⁷ John E Landers,¹⁷ Jan H Veldink,¹¹ Vincenzo Silani,^{19,20} Aaron D Gitler,¹⁸ Christopher E Shaw,¹⁴ Guy A Rouleau,⁶ Leonard H van den Berg,¹¹ Christine Van Broeckhoven,^{4,5} Rosa Rademakers, Peter M Andersen,^{1,21} Christian Kubisch²

► Additional material is published online only. To view please visit the journal online (<http://dx.doi.org/10.1136/jmedgenet-2014-102360>).

For numbered affiliations see end of article.

Correspondence to

Dr Peter M Andersen, Department of Pharmacology and Clinical Neuroscience, Umeå University, Umeå SE-90185, Sweden; peter.andersen@neuro.umu.se

CA and AEV wish it to be known that, in their opinion, the first two authors should be regarded as joint first authors. PMA and CK wish it to be known that, in their opinion, the last two authors should be regarded as joint last authors.

Received 13 February 2014
Accepted 10 March 2014
Published Online First
4 April 2014



CrossMark

To cite: Akimoto C, Volk AE, van Blitterswijk M, et al. *J Med Genet* 2014;**51**:419–424.

ABSTRACT

Background The GGGGCC-repeat expansion in *C9orf72* is the most frequent mutation found in patients with amyotrophic lateral sclerosis (ALS) and frontotemporal dementia (FTD). Most of the studies on *C9orf72* have relied on repeat-primed PCR (RP-PCR) methods for detection of the expansions. To investigate the inherent limitations of this technique, we compared methods and results of 14 laboratories.

Methods The 14 laboratories genotyped DNA from 78 individuals (diagnosed with ALS or FTD) in a blinded fashion. Eleven laboratories used a combination of amplicon-length analysis and RP-PCR, whereas three laboratories used RP-PCR alone; Southern blotting techniques were used as a reference.

Results Using PCR-based techniques, 5 of the 14 laboratories got results in full accordance with the Southern blotting results. Only 50 of the 78 DNA samples got the same genotype result in all 14 laboratories. There was a high degree of false positive and false negative results, and at least one sample could not be genotyped at all in 9 of the 14 laboratories. The mean sensitivity of a combination of amplicon-length analysis and RP-PCR was 95.0% (73.9–100%), and the mean specificity was 98.0% (87.5–100%). Overall, a sensitivity and specificity of more than 95% was observed in only seven laboratories.

Conclusions Because of the wide range seen in genotyping results, we recommend using a combination of amplicon-length analysis and RP-PCR as a minimum in a research setting. We propose that Southern blotting techniques should be the gold standard, and be made obligatory in a clinical diagnostic setting.

INTRODUCTION

In 2011, an expansion of a GGGGCC-repeat in the gene ‘Chromosome 9 open reading frame 72’ (*C9orf72*) was identified as a cause of amyotrophic lateral sclerosis (ALS, OMIM614260) and frontotemporal dementia (FTD, OMIM105550).^{1,2} The following 3-years series of publications reported that a large proportion of ALS (1–30%) and FTD (6–30%) Caucasian patients carry a *C9orf72* repeat expansion,^{3–5} making this mutation the most common known genetic cause of ALS and FTD, and one of the most frequent genetic alterations causing neurodegenerative diseases overall. In one of the initial reports, a combination of amplicon-length analysis, repeat-primed PCR (RP-PCR) assays, and Southern blot (SB) was used for detection and calculation of the repeat numbers.¹ SB is regarded as the gold standard for detecting large polynucleotide repeat expansions,⁶ but it is relatively expensive, cumbersome and time consuming, and up to 10 µg of high-quality DNA is needed for a single analysis. It is not surprising, therefore, that in nearly all studies published during 2011–2013, the much simpler, cheaper and faster-to-perform PCR-based screening methods were used.² By using amplification primers flanking the repeat motif, the amplicon-length analysis allows determination of the exact repeat numbers of alleles with up to 30 repeats, and thus, is able to exclude a pathological repeat expansion if two different alleles in the wild-type range are detected. In RP-PCR, at least two primers are used: one primer that hybridises outside the repeat motif, and one primer that binds to the repeat motif itself. In most protocols a third

primer is applied that hybridises to an oligonucleotide tail of the repeat motif binding primer, also leading to the term triplet-primed PCR.⁷ In this RP-PCR, a large GGGGCC-repeat expansion typically gives rise to a ‘saw-tooth’ or ‘stutter’ pattern, which has been taken as evidence for the presence of a disease-associated expansion.^{1–2} Since the first reports of this mutation, a large number of studies have reported on the epidemiology, clinical, psychological and imaging features, and postmortem neuropathology of *C9orf72* repeat expansion carriers with ALS, FTD, Alzheimer disease, Parkinson’s disease, and other neurodegenerative diseases, as well as in healthy individuals.⁸ In many of these studies, the *C9orf72*-genotype was assessed using RP-PCR alone without a confirmatory SB analysis. Only recently, large-scale studies using SB analysis have been published.^{9–10} PCR-based *C9orf72* screening is also used in clinical diagnostic testing of affected individuals, as well as in predictive testing of healthy individuals at-risk of ALS and FTD. The aim of this study is to determine the sensitivity and specificity of different *C9orf72* genotyping methods, and to establish recommendations for molecular testing of the GGGGCC-repeat expansion in *C9orf72*. Fourteen experienced genetic laboratories participated in this study, and all laboratories were requested to perform RP-PCR with or without amplicon-length analyses on the same 78 samples.

MATERIALS AND METHODS

Study design

At the initiative of the ALS research laboratory at Umeå University, Sweden, 20 laboratories with extensive experience in performing DNA analysis for neurodegenerative diseases, including ALS and FTD, were invited to participate in this study. Eleven research laboratories and two diagnostic laboratories agreed to participate, and including Umeå University a total of 14 laboratories participated. Umeå University sent 400 ng DNA from 78 familial ALS or FTD cases to 13 laboratories: eight in Europe, four in North America, and one in Asia. All participating research laboratories had already established the methodologies for identifying the GGGGCC-repeat expansion in *C9orf72* and published at least one manuscript regarding *C9orf72*. Each laboratory was asked to analyse the 78 DNA samples according to their own procedures and classify the results. Umeå University collected all results independently and analysed them in a blinded fashion. Eleven laboratories used RP-PCR and amplicon-length analysis, and three laboratories used RP-PCR alone. Independently to RP-PCR and amplicon-length analysis, SB was performed in three laboratories.

Patients and DNA extraction

Blood samples of 78 familial ALS or FTD patients from 32 families living in the Nordic countries, Switzerland and Portugal were collected. The sex ratio was 1.75 males per female. The ALS patients were diagnosed according to the EFNS consensus diagnostic criteria,¹¹ and the FTD patients according to the Neary criteria.¹² Autopsies were performed in 13 patients confirming the diagnosis of ALS or FTD neuropathologically. Whole venous blood was drawn into EDTA-containing vacuum tubes and following centrifugation, the buffy coat was isolated. The samples were collected during the time period 1993–2012 and stored as buffy coat in -80°C freezers until DNA extraction. The DNA was extracted according to the manufacturer’s protocol with the DNA extraction kit NUCLEON BACC2 (GE Healthcare, Piscataway, New Jersey, USA) and DNA from the same extraction batch was sent to all laboratories. Ethical review boards in Sweden (The Regional Medical Review Board for

Northern Sweden), Switzerland (Ethikkommission des Kantons St Gallen), and Portugal (Hospital de Santa Maria Ethics Committee, Lisbon) approved this study, and all participants gave informed written consent.

RP-PCR, amplicon-length analyses, and SB

The RP-PCR and amplicon-length analyses were done according to each laboratory’s own method, and these are listed in online supplementary table S3. The SB protocols are listed in online supplementary table S5.

RESULTS

Genotyping results: comparison of the results in 14 laboratories

The overall results of 14 laboratories (laboratory A–N) are presented in table 1. When combining amplicon-length analysis and RP-PCR, sensitivity and specificity above 95% were found in seven laboratories (A–E, G and L; 50%). The mean sensitivity of the combined results was 95.0% (73.9–100%), and the mean specificity was 98.0% (87.5–100%). Using RP-PCR analysis alone, a sensitivity and specificity of more than 95% were found in six laboratories (A–D, G and L; 42.9%). The mean sensitivity of RP-PCR alone was 94.3% (71.7–100%) and the mean specificity was 97.3% (87.5–100%). The classification determined by RP-PCR alone, therefore, changed eight genotype calls in four laboratories (E, F, J and K). Or put differently, 6 of 14 laboratories (42.8%) failed to correctly classify some samples when they performed RP-PCR analysis only. By combining RP-PCR with fragment-length analysis, four laboratories still reported false negative or false positive samples. A significant number of samples (1–10) were unclassifiable/difficult to classify in eight of the laboratories even when RP-PCR and fragment-length analysis were combined.

Genotyping results: comparison of genotyped samples

The genotyping results of all 78 samples are summarised in online supplementary tables S1 and S2 available online. For 50 samples, the results among all laboratories were consistent. In the RP-PCR-only analysis, there were three individuals (6.5, 26.3 and 32.3) who were either misclassified or failed to classify in RP-PCR; these three samples had between 23 and 32 repeats on the wild-type allele, in addition to a large expansion on the other allele. In the amplicon-length analysis, two samples (samples 8.1 and 8.2) from the same family (number 8) were hard to classify: six laboratories identified two amplicons, four laboratories identified one amplicon, and one laboratory classified the samples as ‘undecided’. We sequenced these samples and revealed that one allele was a wild-type allele with two repeats, and the other allele had six repeats with a complex of 15 bp deletion/17 bp insertion mutation just after the repeat motif (online supplementary figure S1A,B). We found one more sample (sample 32.1) that carried the same nucleotide variant on one allele, and this sample also carried a large repeat expansion on the other allele (online supplementary figure S1C,D). This mutation is located in the low-complexity sequence region of *C9orf72*,¹³ and consequently, may interfere with the PCR-based genotyping method.

As a reference, SB was performed on all samples in a blinded fashion by three laboratories. There was complete concordance among the three laboratories: 46 samples carried a repeat expansion and 32 samples lacked the expansion (online supplementary figure S2).

Table 1 The results of 14 laboratories on 78 samples

	Laboratories													
	A	B	C	D	E	F	G	H	I	J	K	L	M	N
RP-PCR results														
GGGGCC-repeat expansion														
Yes	46	46	46	46	42	48	45	43	43	45	34	44	41	43
No	32	32	32	32	32	28	33	32	32	30	33	32	33	28
Q	0	0	0	0	4	2	0	3	3	3	11	2	4	7
False positive	0	0	0	0	0	2	0	0	0	2	1	0	0	0
False negative	0	0	0	0	1	0	1	0	0	0	2	0	1	0
Sensitivity (%)	100	100	100	100	91.3	100	97.8	93.5	93.5	93.5	71.7	95.7	89.1	93.5
Specificity (%)	100	100	100	100	96.8	87.5	100	100	100	93.8	96.9	100	100	87.5
Unclassified (%)	0	0	0	0	5.1	2.6	0	3.8	3.8	3.8	14.1	2.6	5.1	9.0
Amplicon-length analysis results														
Numbers of amplicon														
0	0	0	1	0	0	0	0	0	0	0	0	N.D.	N.D.	N.D.
1	54	54	55	56	56	54	54	51	56	54	51	N.D.	N.D.	N.D.
2	24	24	22	22	22	24	24	24	22	24	24	N.D.	N.D.	N.D.
Q	0	0	0	0	0	0	0	3	0	0	3	N.D.	N.D.	N.D.
RP-PCR+amplicon-length analysis results														
GGGGCC-repeat expansion														
Yes	46	46	46	46	46	48	45	43	43	43	34	44	41	43
No	32	32	32	32	32	29	33	32	32	30	34	32	33	28
Q	0	0	0	0	0	1	0	3	3	5	10	2	4	7
False positive	0	0	0	0	0	2	0	0	0	0	0	0	0	0
False negative	0	0	0	0	0	0	1	0	0	0	2	0	1	0
Sensitivity (%)	100	100	100	100	100	100	97.8	93.5	93.5	93.5	73.9	95.7	89.1	93.5
Specificity (%)	100	100	100	100	100	90.6	100	100	100	93.8	100	100	100	87.5
Unclassified (%)	0	0	0	0	0	1.3	0	3.8	3.8	6.4	12.8	2.6	5.1	9.0

No, sample numbers without large GGGGCC-repeat expansion; N.D., not determined; Q, sample numbers of undecided or unclassified samples Yes: sample numbers with large GGGGCC-repeat expansion.

DISCUSSION

Our blinded multicenter study demonstrates limitations of PCR-based techniques used to assess *C9orf72* GGGGCC-repeats, emphasising the need of detailed technical consensus guidelines for diagnostic and research settings.

Comparison of the methods of RP-PCR and amplicon-length analysis

Based on RP-PCR results alone, 50 samples (64.1%) showed congruent results among the participating laboratories. There are many variable RP-PCR protocols (see online supplementary table S3) and a comprehensive comparison is difficult. In summary, the RP-PCR protocols of the four laboratories (A–D) with 100% sensitivity and specificity, we found that laboratories A and C used almost identical methods. Laboratory B used primers that were a modification of the ones previously published,¹ and they deleted the unspecific linker region between the fluorescence tag and *C9orf72*-specific sequence. In laboratory D, the RP-PCR was performed with only two primers. The primers sets P1, 2, 3 and P4, 5, 6 were the most commonly used with five laboratories using each set, and two of the laboratories that obtained 100% sensitivity and specificity used primers set P1, 2, 3. Among the laboratories that failed to classify samples in concordance with the SB results (having false positive and/or false negative), we found that one used a very short PCR elongation time, the PCR products were diluted extensively before capillary electrophoresis, and the

concentration of deaza-dGTP was very low, or deaza-dGTP was not used at all.

In the amplicon-length analysis alone, 72 samples (92.3%) had concordant results, highlighting the reliability of this technique, regardless of the differences in PCR reactions and PCR protocols. The set of primers were the same in all laboratories except one, but the PCR reactions and PCR protocols were all different between the laboratories (see online supplementary table S3). Thus, it is not possible to identify specific parameters that could explain the incorrect results for the six samples that were not concordant between the laboratories.

Analyses of the results

All five laboratories that obtained full concordance of the PCR-based and SB results used RP-PCR in combination with amplicon-length analysis. None of the laboratories that performed RP-PCR alone reported the correct genotype in all samples. The sensitivity and specificity increased, and the percentage of unclassified samples decreased in three laboratories (E, F and K) when they performed RP-PCR and amplicon-length analysis. Accordingly, a combination of amplicon-length analysis and RP-PCR methodology is recommended to obtain the highest level of sensitivity and specificity, but it should be emphasised that a high risk of misclassification as either false positive or false negative (6 samples in four laboratories) still exists.

The RP-PCR results alone seemed difficult to interpret if one allele with a relatively large number (20–32) of

GGGGCC-repeats was present in combination with a large repeat expansion on the other allele. This was demonstrated by three individuals who could not be genotyped using RP-PCR alone in five laboratories, possibly because the saw-tooth pattern curve of the large expanded alleles were hidden behind the peaks of the allele with 20–32 repeats (figure 1). Similarly, this intermediate 20–32 repeat allele could easily mimic an expanded allele and become a cause of false positive results.

Based on our data, there is no common definition of what a ‘GGGGCC-repeat expansion’ in *C9orf72* is, and laboratories

classify it in different ways. For example, some laboratories used clear cut-offs, for example, of more than 24 or 30 repeats, while other laboratories used the definition that a saw-tooth pattern in RP-PCR corresponds to a GGGGCC-repeat expansion (see online supplementary table S4). In this study, there was no false positive result based on different classifications of what is an ‘expansion’. This study was designed to compare the genotyping results of *C9orf72* among laboratories using the same DNA. An identical amount of DNA was sent to all laboratories, but the concentration and quality of DNA may have changed





Figure 1 Results from RP-PCR and amplicon-length analysis of sample 6.5. (A) A-1 is the full shape of RP-PCR and A-2 is a 16-times scale-up (Y-axis) shape of A-1. It is possible to see the clear saw-tooth pattern after scale-up. A-3 is the result of amplicon-length analysis and there is a number of small peaks before the highest last peak. It may be hard to identify the amplicon numbers because of these small peaks. (B) The RP-PCR figures and classifications of the same sample in different laboratories. The scale of the Y-axis is the same as in A-2.

Figure 2 Flow chart for *C9orf72* genotyping in a scientific setting.



during shipping and handling. Another possible limitation of the present study is that most reactions were only performed once in each laboratory, and the accuracy may be improved if reactions were repeated in case of doubt, for example, with more DNA (however, only two laboratories requested more DNA).

Additionally, all laboratories that participated in this study used their own protocols, and hence, this study was not designed to thoroughly assess every single step in the protocol under similar circumstances within laboratories.

Proposed methods for GGGGCC-repeat expansion genotyping in *C9orf72*

There is an urgent need for broad consensus on analysing GGGGCC-repeat expansions in *C9orf72*, which is particularly important in a clinical setting (for diagnosing ALS or FTD, or when performing predictive testing of at-risk individuals), but also for research purposes. A possible algorithm for *C9orf72* genotyping is presented in figure 2. In conclusion, we recommend, that as an absolute minimum, a combination of amplicon-length analysis and RP-PCR should be performed. We recommend using good quality and quantity of DNA and primers, an appropriate concentration of deaza-dGTP, and a minimum elongation time of 3 min. When results are questionable, we suggest (1) expand the analysed scale in the analysis software, (2) use a higher concentration of PCR products in the capillary electrophoresis and (3) repeat RP-PCR with a higher amount of DNA. Though five of the 14 laboratories got full concordance with SB using PCR-based techniques only, the high risk for misgenotyping using only PCR-based techniques as performed here in nine laboratories, and the devastating consequences misgenotyping may have in clinical practise, make us conclude that SB should always be employed in a diagnostic setting, and should be the preferred method in a research setting of smaller number of samples (eg, analysis of the expansion in autopsy tissue specimens). RP-PCR plus amplicon-length analysis should be used in a research setting and when many samples are to be analysed, for example, in an epidemiological study. Optimally, also is such studies should samples with a saw-tooth pattern be confirmed to have an expansion by SB (figure 2).

Author affiliations

- ¹Department of Pharmacology and Clinical Neuroscience, Umeå University, Umeå, Sweden
- ²Institute of Human Genetics, Ulm University, Ulm, Germany
- ³Department of Neuroscience, Mayo Clinic, Jacksonville, Florida, USA
- ⁴Neurodegenerative Brain Diseases Group, Department of Molecular Genetics, VIB, University of Antwerp—CDE, Antwerp, Belgium
- ⁵Diagnostic Service Facility, Laboratory of neurogenetics, Institute Born-Bunge, University of Antwerp, Antwerp, Belgium
- ⁶Department of Neurology and Neurosurgery, Montreal Neurological Institute and Hospital, McGill University, Montreal, Quebec, Canada
- ⁷Department of Biochemistry, Nimes University Hospital, Nimes Cedex 9, France
- ⁸Center SLA, Montpellier University Hospital, Hôpital Gui-de-Chauliac, Montpellier Cedex 5, France
- ⁹Medizinisch Genetisches Zentrum, München, Germany
- ¹⁰Department of Neurology, Brain Research Institute, Niigata University, Niigata, Japan
- ¹¹Department of Neurology, Brain Center Rudolf Magnus, University Medical Center Utrecht, Utrecht, The Netherlands
- ¹²Faculty of Medicine-University of Lisbon, Instituto de Medicina Molecular, Hospital de Santa Maria, University of Lisbon, Alameda Universidade, Lisbon, Portugal
- ¹³Department of neurology, Kantonsspital St. Gallen and University Hospital, St. Gallen, Switzerland
- ¹⁴Institute of Psychiatry, King's College London and King's Health Partners, London, UK
- ¹⁵Oxford Medical Genetics Laboratories, Churchill Hospital, Oxford, England
- ¹⁶Nuffield Department of Clinical Neurosciences, University of Oxford, John Radcliffe hospital, Oxford, UK
- ¹⁷Department of Neurology, University of Massachusetts Medical School, Worcester, Massachusetts, USA
- ¹⁸Department of Genetics, Stanford University School of Medicine, Stanford, California, USA
- ¹⁹Department of Pathophysiology and Transplantation, "Dino Ferrari" Center, Università degli Studi di Milano, Milan, Italy
- ²⁰Department of Neurology and Laboratory of Neuroscience, IRCCS Istituto Auxologico Italiano, Milan, Italy
- ²¹Department of Neurology, University of Ulm, Ulm, Germany

Acknowledgements We thank the patients and their relatives for participating in this study. We also wish to thank Ann-Charloth Nilsson (Umeå University), Birgit Schmoll (Ulm University), Peter Sapp (University of Massachusetts Medical School), Jovana Kantar and Gonzague Sacaze (Nimes university hospital) for excellent technical assistance.

Contributors CA, AV, PMA and CK designed the study, and with MvB and RR wrote the first drafts of the manuscript. MvB, CA and AV also performed lab analysis and interpreted the results, as did MvdB, CL, SL, WC, BN, OO, WvR, BS, MP, KT, PK, AC, AR, JvdZ, HA, AB, DC, AN, DT, WJ, HD, SA, MD-H, TK, AL-J, KM, JL, JV, VS, AG, CS, GR, LvdB, CvB, SP, McC, MW and PMA accrued the patient cohort. PMA and CK were the PIs on the project.

Funding This project was funded by the Swedish Science Council, the Brain Research Foundation, Mr B Hällsten's Brain Research Foundation, The Ulla-Carin Lindquist's Foundation for ALS Research, the Knut and Alice Wallenberg Foundation, Swedish Brain Power, the European Community's Health Seventh Framework Programme (FP7/2007–2013) (grant agreement no. 259867), The Belgian Science Policy Office Interuniversity Attraction Poles (IAP) programme, the Flemish Government supported Europe Initiative on Centers of Excellence in Neurodegeneration (CoEN), the Flemish Government initiated Methusalem excellence research programme, Alzheimer Research Foundation, the Medical Foundation Queen Elisabeth, the Research Foundation Flanders (FWO) and the FWO provided a postdoctoral scientist fellowship to JvdZ, University of Antwerp Research Fund, the Swiss ALS Foundation, the Italian Ministry of Health (RF-2009-1473856), Grant-in-Aid for the Research Committee of CNS Degenerative Diseases and Comprehensive Research on Disability Health and Welfare from the Ministry of Health, Labour and Welfare in Japan and Dr Van Blitterswijk is supported by the Milton Safenowitz Post-Doctoral Fellowship for ALS research from the ALS Association.

Competing interests None.

Ethical approval The Medical Ethical Review Boards in Sweden, Switzerland and Portugal.

Provenance and peer review Not commissioned; externally peer reviewed.

Open Access This is an Open Access article distributed in accordance with the Creative Commons Attribution Non Commercial (CC BY-NC 3.0) license, which permits others to distribute, remix, adapt, build upon this work non-commercially, and license their derivative works on different terms, provided the original work is properly cited and the use is non-commercial. See: <http://creativecommons.org/licenses/by-nc/3.0/>

REFERENCES

- DeJesus-Hernandez M, Mackenzie IR, Boeve BF, Boxer AL, Baker M, Rutherford NJ, Nicholson AM, Finch NA, Flynn H, Adamson J, Kouru N, Wojtas A, Sengdy P, Hsiung GY, Karydas A, Sealey WW, Josephs KA, Coppola G, Geschwind DH, Wszolek ZK, Feldman H, Knopman DS, Petersen RC, Miller BL, Dickson DW, Boylan KB, Graff-Radford NR, Rademakers R. Expanded GGGGCC hexanucleotide repeat in noncoding region of *C9orf72* causes chromosome 9p-linked FTD and ALS. *Neuron* 2011;72:245–56.
- Renton AE, Majounie E, Waite A, Simón-Sánchez J, Rollinson S, Gibbs JR, Schymick JC, Laaksovirta H, van Swieten JC, Myllykangas L, Kalimo H, Paetau A, Abramzon Y, Remes AM, Kaganovich A, Scholz SW, Duckworth J, Ding J, Harmer DW, Hernandez DG, Johnson JO, Mok K, Ryten M, Trabuni D, Guerreiro RJ, Orrell RW, Neal J, Murray A, Pearson J, Jansen IE, Sondervan D, Seelaar H, Blake D, Young K, Halliwell N, Callister JB, Toulson G, Richardson A, Gerhard A, Snowden J, Mann D, Neary D, Nalls MA, Peuralinna T, Jansson L, Isoviita VM, Kaivorinne AL, Hölttä-Vuori M, Ikonen E, Sulkava R, Benatar M, Wu J, Chiò A, Restagno G, Borghera G, Sabatelli M, ITALSGEN Consortium Heckerman D, Rogava E, Zinman L, Rothstein JD, Sendtner M, Drepper C, Eichler EE, Alkan C, Abdullaev Z, Pack SD, Dutra A, Pak E, Hardy J, Singleton A, Williams NM, Heutink P, Pickering-Brown S, Morris HR, Tienari PJ, Traynor BJ. A hexanucleotide repeat expansion in *C9orf72* is the cause of chromosome 9p21-linked ALS-FTD. *Neuron* 2011;72:257–68.
- Gijssels I, Van Langenhove T, van der Zee J, Sleegers K, Philtjens S, Kleinberger G, Janssens J, Bettens K, van Cauwenbergh C, Pereson S, Engelborghs S, Sieben A, De Jonghe P, Vandenbergh R, Santens P, De Bleecker J, Maes G, Bäumer V, Dillen L, Joris G, Couijt I, Corsmit E, Elinck E, Van Dongen J, Vermeulen S, Van den Broeck M, Vaerenberg C, Mattheijssens M, Peeters K, Robberecht W, Cras P, Martin JJ, De Deyn PP, Cruts M, Van Broeckhoven C. A *C9orf72* promoter repeat expansion in a Flanders-Belgian cohort with disorders of the frontotemporal lobar degeneration-amyotrophic lateral sclerosis spectrum: a gene identification study. *Lancet Neurol* 2012;11:54–65.
- Smith BN, Newhouse S, Shatunov A, Vance C, Topp S, Johnson L, Miller J, Lee Y, Troakes C, Scott KM, Jones A, Gray I, Wright J, Hortobágyi T, Al-Sarraj S, Rogeli B, Powell J, Lupton M, Lovestone S, Sapp PC, Weber M, Nestor PJ, Schelhaas HJ, Asbroek AA, Silani V, Gellera C, Taroni F, Ticonni N, Van Den Berg L, Veldink J, Van Damme P, Robberecht W, Shaw PJ, Kirby J, Pall H, Morrison KE, Morris A, de Belleruche J, Vianney de Jong JM, Baas F, Andersen PM, Landers J, Brown RH Jr, Weale ME, Al-Chalabi A, Shaw CE. The *C9orf72* expansion mutation is a common cause of ALS+/-FTD in Europe and has a single founder. *Eur J Hum Genet* 2013;21:102–8.
- Majounie E, Renton AE, Mok K, Dopper EG, Waite A, Rollinson S, Chio A, Restagno G, Nicolaou N, Simon-Sanchez J, van Swieten JC, Abramzon Y, Johnson JO, Sendtner M, Pamplett R, Orrell RW, Mead S, Sidle KC, Houlden H, Rohrer JD, Morrison KE, Pall H, Talbot K, Ansorge O, Chromosome 9-ALS/FTD Consortium; French research network on FTL/FTLD/ALS; ITALSGEN Consortium Hernandez DG, Arepalli S, Sabatelli M, Mora G, Corbo M, Giannini F, Calvo A, Englund E, Borghero G, Floris GL, Remes AM, Laaksovirta H, McCluskey L, Trojanowski JQ, Van Deerlin VM, Schellenberg GD, Nalls MA, Drory VE, Lu CS, Yeh TH, Ishiura H, Takahashi Y, Tsuji S, Le Ber I, Brice A, Drepper C, Williams N, Kirby J, Shaw P, Hardy J, Tienari PJ, Heutink P, Morris HR, Pickering-Brown S, Traynor BJ. Frequency of the *C9orf72* hexanucleotide repeat expansion in patients with amyotrophic lateral sclerosis and frontotemporal dementia: a cross-sectional study. *Lancet Neurol* 2012;11:323–30.
- Curtis-Cioffi KM, Rodrigues DA, Rodrigues VC, Cicarelli RM, Scarel-Caminaga RM. Comparison between the polymerase chain reaction-based screening and the Southern blot methods for identification of fragile X syndrome. *Genet Test Mol Biomarkers* 2012;16:1303–8.
- Hantash FM, Goos DG, Tsao D, Quan F, Buller-Burckle A, Peng M, Jarvis M, Sun W, Strom CM. Qualitative assessment of FMR1 (CGG)_n triplet repeat status in normal, intermediate, premutation, full mutation, and mosaic carriers in both sexes: implications for fragile X syndrome carrier and newborn screening. *Genet Med* 2010;12:162–73.
- Beck J, Poulter M, Hensman D, Rohrer JD, Mahoney CJ, Adamson G, Campbell T, Uphill J, Borg A, Fratta P, Orrell RW, Malaspina A, Rowe J, Brown J, Hodges J, Sidle K, Polke JM, Houlden H, Schott JM, Fox NC, Rossor MN, Tabrizi SJ, Isaacs AM, Hardy J, Warren JD, Collinge J, Mead S. Large *C9orf72* hexanucleotide repeat expansions are seen in multiple neurodegenerative syndromes and are more frequent than expected in the UK population. *Am J Hum Genet* 2013;92:345–53.
- Dobson-Stone C, Hallupp M, Loy CT, Thompson EM, Haan E, Sue CM, Panegyres PK, Razquin C, Seijo-Martinez M, Rene R, Gascon J, Campdelacreu J, Schmolz B, Volk AE, Brooks WS, Schofield PR, Pastor P, Kwok JB. *C9orf72* repeat expansion in Australian and Spanish frontotemporal dementia patients. *PLoS ONE* 2013;8:e56899.
- van Blitterswijk M, DeJesus-Hernandez M, Niemantsverdriet E, Murray ME, Heckman MG, Diehl NN, Brown PH, Baker MC, Finch NA, Bauer PO, Serrano G, Beach TG, Josephs KA, Knopman DS, Petersen RC, Boeve BF, Graff-Radford NR, Boylan KB, Petrucelli L, Dickson DW, Rademakers R. Association between repeat sizes and clinical and pathological characteristics in carriers of *C9orf72* repeat expansions (Xpansize-72): a cross-sectional cohort study. *Lancet Neurol* 2013;12:978–88.
- Andersen PM, Abrahams S, Borasio GD, de Carvalho M, Chio A, Van Damme P, Hardiman O, Kollewe K, Morrison KE, Petri S, Pradat PF, Silani V, Tomik B, Wasner M, Weber M. EFNS Task Force on Management of Amyotrophic Lateral Sclerosis. EFNS guidelines on the Clinical Management of Amyotrophic Lateral Sclerosis (MALS)—revised report of an EFNS task force. *Eur J Neurol* 2012;19:360–75.
- Neary D, Snowden J, Mann D. Frontotemporal dementia. *Lancet Neurol* 2005;4:771–80.
- van der Zee J, Gijssels I, Dillen L, Van Langenhove T, Theuns J, Engelborghs S, Philtjens S, Vandenbulcke M, Sleegers K, Sieben A, Bäumer V, Maes G, Corsmit E, Borroni B, Padovani A, Archetti S, Pernecky R, Diehl-Schmid J, de Mendonça A, Miltenberger-Miltenyi G, Pereira S, Pimentel J, Nacmias B, Bagnoli S, Sorbi S, Graff C, Chiang HH, Westerlund M, Sanchez-Valle R, Llado A, Gelpi E, Santana I, Almeida MR, Santiago B, Frisoni G, Zanetti O, Bonvicini C, Synofzik M, Maetzler W, Vom Hagen JM, Schöls L, Heneka MT, Jessen F, Matej R, Parobkova E, Kovacs GG, Ströbel T, Sarafov S, Tournev I, Jordanova A, Danek A, Arzberger T, Fabrizio GM, Testi S, Salmon E, Santens P, Martin JJ, Cras P, Vandenbergh R, De Deyn PP, Cruts M, Van Broeckhoven C, van der Zee J, Gijssels I, Dillen L, Van Langenhove T, Theuns J, Philtjens S, Sleegers K, Bäumer V, Maes G, Corsmit E, Engelborghs S, De Deyn PP, Cras P, Engelborghs S, De Deyn PP, Vandenbulcke M, Borroni B, Padovani A, Archetti S, Pernecky R, Diehl-Schmid J, Synofzik M, Maetzler W, Müller Vom Hagen J, Schöls L, Synofzik M, Maetzler W, Müller Vom Hagen J, Schöls L, Heneka MT, Jessen F, Ramirez A, Kurzweil D, Sachtleben C, Mairer W, de Mendonça A, Miltenberger-Miltenyi G, Pereira S, Firmo C, Pimentel J, Sanchez-Valle R, Llado A, Antonell A, Molinuevo J, Gelpi E, Graff C, Chiang HH, Westerlund M, Graff C, Kinhult Ståhlbom A, Thonberg H, Nennesmo I, Börjesson-Hanson A, Nacmias B, Bagnoli S, Sorbi S, Bessi V, Piaceri I, Santana I, Santiago B, Santana I, Helena Ribeiro M, Rosário Almeida M, Oliveira C, Massano J, Garret C, Pires P, Frisoni G, Zanetti O, Bonvicini C, Sarafov S, Tournev I, Jordanova A, Tournev I, Kovacs GG, Ströbel T, Heneka MT, Jessen F, Ramirez A, Kurzweil D, Sachtleben C, Mairer W, Jessen F, Matej R, Parobkova E, Danel A, Arzberger T, Maria Fabrizio G, Testi S, Ferrari S, Cavallaro T, Salmon E, Santens P, Cras P; European Early-Onset Dementia Consortium. A pan-European study of the *C9orf72* repeat associated with FTD: geographic prevalence, genomic instability, and intermediate repeats. *Hum Mutat* 2013;34:363–73.

SUPPLEMENTARY MATERIAL

**A blinded international study on the reliability of genetic testing
for GGGGCC-repeat expansions in *C9orf72* reveals marked
differences in results among 14 laboratories**

Chizuru Akimoto, Alexander E Volk *et al.*

Supplementary table 1. The results of 78 samples.

RP-PCR results		
All laboratories same results	50 samples (64.1%)	
1 laboratory different result	20 samples (25.6%)	13 samples (16.7%): 1 laboratory Q result 5 samples (6.41%): 1 laboratory FP result 2 samples (2.56%): 1 laboratory FN result
2 laboratories different results	5 samples (6.41%)	4 samples (5.12%): 2 laboratories Q results 1 sample (1.28%): 1 laboratory FN and 1 laboratory Q results
6 laboratories different results	2 samples (2.56%)	1 sample (1.28%): 1 laboratory FN and 5 laboratories Q results 1 sample (1.28%): 6 laboratories Q results
7 laboratories different results	1 sample (1.28%)	1 sample (1.28%): 1 laboratory FN and 6 laboratories Q results
Amplicon-length analysis results		
All laboratories same results	72 samples (92.3%)	
1 laboratory different result	2 samples (2.56%)	1 sample (1.28%): 1 laboratory different peak number 1 sample (1.28%): 1 laboratory Q result
2 laboratories different results	1 sample (1.28%)	1 sample (1.28%): 1 laboratory different peak number and 1 laboratory Q result
3 laboratories different results	1 sample (1.28%)	1 sample (1.28%): 1 laboratory different peak number and 2 laboratories Q results
5 laboratories different results	2 samples (2.56%)	2 samples (2.56%): 4 laboratories different peak number and 1 laboratory Q result
RP-PCR + amplicon-length analysis results		
All laboratories same results	53 samples (67.9%)	
1 laboratory different result	19 samples (24.4%)	16 samples (20.5%): 1 laboratory Q result 2 samples (2.56%): 1 laboratory FP result 1 sample (1.28%): 1 laboratory FN result
2 laboratories different results	3 samples (3.85%)	2 samples (2.56%): 2 laboratories Q results 1 sample (1.28%): 1 laboratory FP and 1 laboratory Q results
5 laboratories Different results	2 samples (2.56%)	1 sample (1.28%): 1 laboratory FN and 4 laboratories Q results 1 sample (1.28%): 5 laboratories Q results
6 laboratories Different results	1 sample (1.28%)	1 sample (1.28%): 1 laboratory FN and 5 laboratories Q results

FP: False positive, FN: False negative, Q: Undecided or unclassified sample.

Supplementary table 2. The GGGGCC-repeat numbers on 78 samples from 32 families.

Sample	Allele 1	Allele 2	Sample	Allele 1	Allele 2
1.1	7	Expanded	16.1	2	Expanded
1.2	6	Expanded	16.3	2	Expanded
2.1	2	Expanded	17.1	2	6
2.2	2	Expanded	17.2	2	5
3.1	5	12	18.1	5	Expanded
3.2	2	Expanded	18.2	2	Expanded
4.1	4	Expanded	19.1	2	2
4.2	4	Expanded	19.2	2	5
4.3	8	Expanded	20.1	5	13
4.4	4	Expanded	20.2	5	6
5.1	2	2	21.1	5	Expanded
5.2	2	2	21.2	5	Expanded
6.1	5	5	22.1	11	Expanded
6.2	2	Expanded	22.2	2	Expanded
6.3	2	Expanded	23.1	8	Expanded
6.4	2	Expanded	23.2	8	Expanded
6.5	23-24	Expanded	24.1	2	Expanded
6.6	8	Expanded	24.2	2	Expanded
6.7	8	Expanded	25.1	2	2
6.8	2	2	25.2	2	8
7.1	2	6	26.1	2	Expanded
7.2	2	5	26.2	2	Expanded
8.1	2	6*	26.3	23-25	Expanded
8.2	2	6*	27.1	2	Expanded
9.1	2	10	27.2	8	Expanded
9.2	2	8	27.3	5	Expanded
10.1	5	Expanded	27.4	5	Expanded
10.2	2	Expanded	28.1	2	7
11.1	2	Expanded	28.2	5	Expanded
11.2	2	Expanded	29.1	2	5
12.1	2	5	29.2	2	6
12.2	2	8	30.1	2	6
12.3	2	5	30.2	2	5
13.1	2	Expanded	31.1	7	Expanded
13.2	2	Expanded	31.2	2	2
14.1	2	2	31.3	2	Expanded
14.2	2	8	32.1	6*	Expanded
15.1	2	6	32.2	2	Expanded
15.2	6	10	32.3	28-32	Expanded

*: mutant allele.

Supplementary table 3A. Total methods of 14 laboratories.

Laboratory	A	B	C	D	E	F	G	H	I	J	K	L	M	N
RP-PCR method														
Primers*	P1, 2, 3	P7, 8, 9	P1, 2, 3	P4, 13	P4, 5, 6	P10, 11, 12	P4, 5, 6	P1, 2, 3	P4, 5, 6	P4, 5, 6	No	P1, 2, 3	P4, 5, 6	P1, 2, 3
PCR reaction*	PR1	PR2	PR1	PR3	PR4	PR5	PR6	PR1	PR7	PR8	No	PR1	PR6	PR9
Primer 1 dose (µM)	1.4	0.2	1.4	0.2	0.4	0.33	1.0	1.4	1.4	2.5	No	1.4	1.0	1.4
Primer 2 dose (µM)	0.7	0.2	0.7	0.2	0.2	0.33	1.0	0.7	0.7	2.5	No	0.7	1.0	0.7
Primer 3 dose (µM)	1.4	0.2	1.4	-	0.4	0.33	1.0	1.4	1.4	0.25	No	1.4	1.0	1.4
Deaza-GTP (mM)	0.18	0.2	0.18	0.2	0.0375	0	0.25	0.18	0.18	0	No	0.18	0.25	0.16
DNA dose (ng/µl)	7.1	5.0	2.14	2.0	20	10.0	12.5	7.1	12.0	3.3	No	7.1	5.0	3.3
PCR protocol*	PP1	PP2	PP3	PP4	PP5	PP6	PP7	PP8	PP7	PP9	No	PP1	PP7	PP1
Extension time (min)	3	8-18	3	3	40 sec	2-10	8-16	3	8-16	8-16	No	3	8-16	3
Total cycle	40	45	40	32	65	35	35	43	35	34	No	40	35	40
Analysis machine*	M1	M2	M2	M1	M1	M1	M1	M1	M3	M1	No	M1	M1	M2
PCR product conc.	10%	10%	25%	20%	0.47%	16.2%	3.3%	0.94%	8.8%	10%	No	21%	4.8%	13.3%
Analysis software*	A1	A2	A1	A1	A1	A5	A2	A2	A2	A2	No	A1	A1	A3
Sensitivity	100%	100%	100%	100%	91.3%	100%	97.8%	93.5%	93.5%	93.5%	71.7%	95.7%	89.1%	93.5%
Specificity	100%	100%	100%	100%	96.8%	87.5%	100%	100%	100%	93.8%	96.9%	100%	100%	87.5%
Unclassified	0%	0%	0%	0%	5.1%	2.6%	0%	3.8%	3.8%	3.8%	14.1%	2.6%	5.1%	9.0%
Amplicon length analysis method														
Primers*	P14, 16	P14, 16	P14, 16	P15, 16	P14, 16	P17, 18	P14, 16	P14, 16	P15, 16	P14, 16	No	N.D.	N.D.	N.D.
PCR reaction*	PR10	PR11	PR12	PR13	PR10	PR14	PR15	PR16	PR13	PR17	No	N.D.	N.D.	N.D.
DNA dose (ng/µl)	2.2	5.0	2.4	4.2	2.0	3.3	3.35	7.1	2.0	3.3	No			
PCR protocol*	PP10	PP11	PP12	PP13	PP14	PP15	PP16	PP17	PP18	PP19	No	N.D.	N.D.	N.D.
Extension time (min)	1.5	1	2	2	1.5	1	1	3	2	8-16	No			
Total cycle	34	38	30	35	35	33	36	40	35	34	No			
Analysis machine*	M1	M2	M2	4% gel	M1	M1	M1	M1	3% gel	M1	No	N.D.	N.D.	N.D.
PCR product conc.	10%	0.5%	25%		0.47%	16.2%	0.17%	0.47%		10%	No			
Analysis software*	A1	A2	A1		A1	A5	A2	A2		A2	No			

*: Listed in supplementary table 3B and 3C, N.D.: not done. No: no method disclosed.

Supplementary table 3B. List of products and equipment.

Primers for RP-PCR	
P1	FAM-AGTCGCTAGAGGCGAAAGC
P2	tacgcatcccagtttgagacgGGGGCCGGGGCCGGGGCCGGGG
P3	tacgcatcccagtttgagacg
P4	FAM-tgtaaaacgacggccagtCAAGGAGGGAAACAACCGCAGCC
P5	caggaaacgctatgaccGGGCCCGCCCCGACCACGCCCGGCCCGGCCCGG
P6	caggaaacgctatgacc
P7	FAM-CAAGGAGGGAAACAACCGCAGCC
P8	ggataacaattcacacaggGGGCCCGCCCCGACCACGCCCGGCCCGGCCCGG
P9	ggataacaattcacacagg
P10	FAM-AGTACTCGCTGAGGGTGAAC
P11	cgtacgcatcccagtttgagaGCCCGGCCCGGCCCGG
P12	cgtacgcatcccagtttgaga
P13	cacgacgttgtaaacgaCCCCGGGCCCGGCCCGG

Primers for amplicon-length analysis	
P14	FAM-CAAGGAGGGAAACAACCGCAGCC
P15	CAAGGAGGGAAACAACCGCAGCC
P16	GCAGGCACCGCAACCGCAG
P17	FAM-CAGGTGTGGGTTTAGGAGGT
P18	CCAGCTTCGGTCAGAGAAAT
Size standard	
S1	GeneScan 500XL ROX and 500ROX Size Standard (Applied Biosystems, USA)
S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
S3	GeneScan 400HD Rox Size Standard (Applied Biosystems, USA)
S4	GeneScan 600LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	
M1	3730 and 3730xl DNA Analyzer (Applied Biosystems, USA)
M2	3130 and 3130xl Genetic Analyzer (Applied Biosystems, USA)
M3	3500 Genetic Analyzer (Applied Biosystems, USA)
Analysis software	

A1	Peak Scanner Software v1.0 or v2.0 (Applied Biosystems, USA)
A2	GeneMapper Software v4.0 or v4.1 (Applied Biosystems, USA)
A3	Geotyper v4 (Applied Biosystems, USA)
A4	Agarose gel
A5	In-house developed Tracl genotyping software

Supplementary table 3C-1. Concentration of ingredients in RP-PCR.

PCR reaction number	PR1	PR2	PR3	PR4	PR5	PR6	PR7	PR8	PR9
Laboratory	ACHL	B	D	E	F	GM	I	J	N
Qiagen buffer *1	-	-	-	1x	-	1x	-	-	-
FastStart PCR Master *2	1x	-	-	-	-	-	-	-	-
Taq DNA polymerase (units)	-	-	1.8	0.75	-	2.5	1.0	-	-
Other buffer and polymerase	-	*4	*5	-	*6	-	*7	*8	*9
Primer 1 (μ M)	1.4	0.2	0.2	0.4	0.33	1.0	1.4	2.5	1.4
Primer 2 (GGGGCC repeat) (μ M)	0.7	0.2	0.2	0.2	0.33	1.0	0.7	2.5	0.7
Primer 3 (μ M)	1.4	0.2	-	0.4	0.03	1.0	1.4	0.25	1.4
dCTP, dATP, dTTP (each) (mM)	-	0.2	0.2	0.05	-	0.25	0.18	-	-
7-deaza-2-deoxy GTP (mM)	0.18	0.2	0.2	0.0375	-	0.25	0.18	-	0.16
dGTP (mM)	-	-	-	0.0125	-	-	-	-	-
dNTPs (each) (mM)	-	-	-	-	0.5	-	-	-	-
DMSO	7%	5%	5%	-	-	5%	-	-	7%
Betaine (M)	-	-	1.0	-	2.0	1.0	-	1.75	0.83
Q solution *3	1x	-	-	1x	-	-	-	-	-
MgCl ₂ (mM)	0.89	2.0	-	-	-	-	-	-	0.83

*1: *Taq* DNA polymerase (QIAGEN, Germany)

*2: FastStart PCR Master mix (Roche, Switzerland)

*3: Q solution (QIAGEN, Germany)

*4: *Taq* Robust (Kapa Biosystems, USA), Kapa Enhancer 1x.

*5: RG25 buffer (160mM NH₄SO₄, 670mM Tris-HCl (pH 8.8), 1% Tween, 25mM MgCl₂)

*6: Expand Long Template PCR system (Roche, Switzerland), Buffer 2 1x, Titanium *Taq* polymerase 2x.

*7: AccuPrime GC-rich DNA polymerase (Invitrogen, USA), Buffer A 1x.

*8: Extensor Mastermix (Thermo Scientific, USA)

*9: True Allele PCR Premix (Applied Biosystems, USA), PCR master mix 1x.

Supplementary table 3C-2. Concentration of ingredients in amplicon-length analysis PCR.

PCR reaction number	PR10	PR11	PR12	PR13	PR14	PR15	PR16	PR17
Used laboratory	AE	B	C	DI	F	G	H	J
Primer 4 (μM)	0.4	0.2	0.4	0.2	0.3	1.0	0.5	0.1
Primer 5 (μM)	0.4	0.2	0.4	0.2	0.3	1.0	0.5	0.1
dCTP, dATP, dTTP (each) (mM)	-	0.2	-	-	-	0.25	-	-
7-deaza-2-deoxy GTP (mM)	-	0.2	0.2	-	-	0.25	0.18	-
dNTP (each) (mM)	-	-	0.25	-	0.3	-	-	-
DMSO (%)	-	5.0	4.0	-	-	5.0	7.1	1.0
Betaine (M)	-	-	1.0	-	1.0	1.0	-	-
Q solution *1	-	-	-	-	-	-	1x	-
MgCl ₂ (mM)	-	2.0	-	-	-	-	0.89	-
Kit buffer	1x	1x	1x	1x	1x	1x	1x	1x
Kit polymerase (U)	-	0.2	1.25	0.4	0.3	2.5	-	-
Kit and polymerase	*2	*3	*4	*5	*6	*7	*8	*9

*1: Q solution (QIAGEN, Germany)

*2: One *Taq* 2X Master Mix with GC Buffer (New England BioLabs, UK)

*3: KAPA Enhncer and *Taq* Robust (Kapa Biosystems, USA)

*4: AmpliTaq and GeneAmp PCR buffer (Roche, Switzerland)

*5: AccuPrime GC-Rich DNA Polymerase (Invitrogen, USA)

*6: Kapa HiFi Hotstart PCR kit, GC buffer (Kapa Biosystems, USA)

*7: *Taq* DNA Polymerase (QIAGEN, Germany)

*8: FastStart PCR Master mix (Roche, Switzerland)

*9: Megamix Blue (Microzone, UK)

Supplementary table 3C-3. PCR protocols for RP-PCR (PP1-9).

PP1, Laboratory A, L, N			PP2, Laboratory B			PP3, Laboratory C		
Temp	Time	Cycle	Temp	Time	Cycle	Temp	Time	Cycle
95°C	15 min	1	98°C	10 min	1	95°C	8 min	1
94°C	1 min	2	98°C	35 sec	13	95°C	1 min	8
70°C	1 min		65-58°C	2 min	Every cycle	70-56°C	1 min	Every cycle
72°C	3 min		72°C	8 min	-1°C	72°C	3 min	-2°C
94°C	1 min	3	97°C	35 sec	32	95°C	1 min	32
68°C	1 min		58°C	2 min	Every cycle	56°C	1 min	
72°C	3 min		72°C	8-18 min	+20 sec	72°C	3 min	
94°C	1 min	4	72°C	10 min	1	72°C	10 min	1
66°C	1 min		4°C	hold	1	4°C	hold	1
72°C	3 min							
94°C	1 min	5						
64°C	1 min							
72°C	3 min							
94°C	1 min	6	PP4, Laboratory D			PP5, Laboratory E		
62°C	1 min		Temp	Time	Cycle	Temp	Time	Cycle
72°C	3 min		95°C	5 min	1	95°C	5 min	1
94°C	1 min	7	95°C	30 sec	32	95°C	30 sec	49
60°C	1 min		68°C	3 min		70-53.2°C	30 sec	Every cycle
72°C	3 min		72°C	7 min		68°C	40 sec	-0.35°C
94°C	1 min	8	12°C	hold	1	95°C	30 sec	16
58°C	1 min					58°C	30 sec	
72°C	3 min					68°C	40 sec	
94°C	1 min	5				68°C	10 min	1
56°C	1 min					12°C	hold	1
72°C	3 min							
72°C	10 min	1						
4°C	hold	1						
PP6, Laboratory F			PP7, Laboratory G, I, M			PP8, Laboratory H		
Temp	Time	Cycle	Temp	Time	Cycle	Temp	Time	Cycle
98°C	10 min	1	98°C	10 min	1	95°C	5 min	1
97°C	35 sec	10	97°C	35 sec	10	95°C	30 sec	8
53°C	2 min		64°C	2 min		70-56°C	30 sec	Every cycle
68°C	2 min		68°C	8 min		72°C	3 min	-2°C
97°C	35 sec	25	97°C	35 sec	25	95°C	30 sec	35
53°C	2 min	Every cycle	64°C	2 min	Every cycle	56°C	30 sec	
68°C	2-10 min	+20 sec	68°C	8-16 min	+20 sec	72°C	3 min	
68°C	10 min	1	16°C	hold	1	72°C	20 min	1
4°C	hold	1				4°C	hold	1

PCR protocols for RP-PCR (continued).

PP9, Laboratory J

Temp	Time	Cycle
95°C	5 min	1
95°C	30 sec	8
70-56°C	30 sec	Every cycle
72°C	8 min	-2°C
95°C	30 sec	26
56°C	30 sec	Every cycle
72°C	8-16 min	+20 sec
72°C	7 min	1
4°C	hold	1

Supplementary table 3C-4. PCR protocols for amplicon-length analysis (PP10-19).

PP10, Laboratory A			PP11, Laboratory B			PP12, Laboratory C		
Temp	Time	Cycle	Temp	Time	Cycle	Temp	Time	Cycle
98°C	5 min	1	97°C	5 min	1	95°C	5 min	1
97°C	30 sec	10	97°C	30 sec		95°C	1 min	
65-56°C	30 sec	Every cycle	60°C	30 sec	10	56°C	1 min	30
68°C	1.5 min	-1°C	72°C	1 min		72°C	2 min	
97°C	30 sec		97°C	10 sec		72°C	10 min	1
55°C	30 sec	24	60°C	15 sec	28	4°C	hold	1
68°C	1.5 min		72°C	30 sec				
68°C	10 min	1	72°C	10 min	1			
12°C	hold	1	4°C	hold	1			

PCR protocols for Amplicon length analysis (continued).

PP13, Laboratory D

Temp	Time	Cycle
95°C	3 min	1
95°C	30 sec	35
68°C	30 sec	
72°C	2 min	
72°C	10 min	1
12°C	hold	1

PP14, Laboratory E

Temp	Time	Cycle
98°C	5 min	1
97°C	30 sec	11
65-55°C	30 sec	Every cycle -1°C
68°C	1.5 min	
97°C	30 sec	24
55°C	30 sec	
68°C	1.5 min	
68°C	10 min	1
12°C	hold	1

PP15, Laboratory F

Temp	Time	Cycle
95°C	5 min	1
98°C	20 sec	33
59°C	15 sec	
72°C	1 min	
72°C	5 min	1
4°C	hold	1

PP16, Laboratory G

Temp	Time	Cycle
94°C	4 min	1
94°C	30 sec	36
62°C	30 sec	
72°C	1 min	
72°C	10 min	1

PP17, Laboratory H

Temp	Time	Cycle
95°C	4 min	1
95°C	30 sec	14
70-57°C	30 sec	Every cycle -1°C
72°C	3 min	
95°C	30 sec	26
56°C	30 sec	
72°C	3 min	
72°C	10 min	1
4°C	hold	1

PP18, Laboratory I

Temp	Time	Cycle
95°C	3 min	1
95°C	30 sec	35
63°C	30 sec	
72°C	2 min	
72°C	10 min	1
16°C	hold	1

PP19, Laboratory J

Temp	Time	Cycle
95°C	5 min	1
95°C	30 sec	8
70-56°C	30 sec	Every cycle -2°C
72°C	8 min	
95°C	30 sec	26
56°C	30 sec	Every cycle +20 sec
72°C	8-16 min	
72°C	7 min	1
4°C	hold	1

Supplementary table 3D. Methods of each laboratory.

Laboratory A**RP-PCR**

List

Primer 1	P1	FAM-AGTCGCTAGAGGCGAAAGC
Primer 2 (with GGGGCC repeat)	P2	tacgcatcccagtttgagacgGGGGCCGGGGCCGGGGCCGGGG
Primer 3	P3	tacgcatcccagtttgagacg
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730x/DNA Analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR1)

Products	Final conc.
FastStart PCR Master	1x
Primer 1	1.4 µM
Primer 2	0.7 µM
Primer 3	1.4 µM
7-deaza-2-deoxy GTP	0.18 mM
DMSO	7 %
Q solution	1x
MgCl ₂	0.89 mM
DNA	7.1 ng/µl
Total	14.0 µl

FastStart PCR Master (Roche, Switzerland)
Q solution (QIAGEN, Germany).

PCR protocol (PP1)

Temp	Time	Cycle
95°C	15 min	1
94°C	1 min	
70-58°C	1 min	2-8 *
72°C	3 min	
94°C	1 min	
56°C	1 min	5
72°C	3 min	
72°C	10 min	1
4°C	hold	1

* Every 2°C decrease add 1 cycle.

Capillary electrophoresis

PCR product	1.0 µl
Size standard	0.5 µl
Formamide	8.5 µl
Total	10.0 µl

95°C 5min and on ice.
Analyse with DNA analyzer.

Amplicon length analysis

List

Primer 4	P14	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730x/ DNA Analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR10)

Products	Final conc.
Primer 4	0.4 µM
Primer 5	0.4 µM
Kit buffer	1x
DNA	2.2 ng/µl
Total	15 µl

Kit buffer: One *Taq* 2x Master Mix with GC Buffer (New England BioLabs, UK).

PCR protocol (PP10)

Temp	Time	Cycle
98°C	5 min	1
97°C	30 sec	10
65-56°C	30 sec	Every cycle -1°C
68°C	1.5 min	
97°C	30 sec	
55°C	30 sec	24
68°C	1.5 min	
68°C	10 min	1
12°C	hold	1

Capillary electrophoresis

PCR product	1.0 µl
Size standard	0.5 µl
Formamide	8.5 µl
Total	10 µl

95°C 5min and on ice.
Analyse with DNA analyzer.

Supple table 3D. Methods of each laboratory

Laboratory B**RP-PCR**

List

Primer 1	P7	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 2 (with GGGGCC repeat)	P8	ggataacaatttcacacaggGGGCCCCGCCCGACCACGCCCCCGGC CCCGGCCCGG
Primer 3	P9	ggataacaatttcacacagg
Size standard	S2	GeneScan 500 LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M2	3130xl Genetic Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper Software v4.0 (Applied Biosystems, USA)

PCR reaction (PR2)		PCR protocol (PP2)			Capillary electrophoresis
Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H ₂ O=1:1
KAPA Enhancer	1x	98°C	10 min	1	PCR product 2.0 µl Size standard 0.1 µl Formamide 7.9 µl
Primer 1	0.2 µM	98°C	35 sec	13	
Primer 2	0.2 µM	65-58°C	2 min	Every cycle	Total 10.0 µl
Primer 3	0.2 µM	72°C	8 min	-1°C	
dCTP, dATP, dTTP (each)	0.2 mM	97°C	35 sec	32	95°C 3min and analyse with genetic analyzer.
7-deaza-2-deoxy GTP	0.2 mM	58°C	2 min	Every cycle	
DMSO	5%	72°C	8-18 min	+20 sec	
MgCl ₂	2.0 mM	72°C	10 min	1	
Taq Robust	0.2 U	72°C	10 min	1	
DNA	5 ng/µl	72°C	10 min	1	
Total	20 µl	4°C	hold	1	

KAPA Enhancer and *Taq* Robust (Kapa Biosystems, USA). Ramping up and down: 0.5°C/sec

Amplicon length analysis

List

Primer 4	P14	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M2	3130xl Genetic Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper Software v4.0 (Applied Biosystems, USA)

PCR reaction (PR11)		PCR protocol (PP11)			Capillary electrophoresis
Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H ₂ O=1:39
KAPA Enhancer	1x	97°C	5 min	1	Diluted PCR product 2.0 µl Size standard 0.1 µl Formamide 7.9 µl
Primer 4	0.2 µM	97°C	30 sec	10	
Primer 5	0.2 µM	60°C	30 sec	10	Total 10.0 µl
dCTP, dATP, dTTP (each)	0.2 mM	72°C	1 min	28	
7-deaza-2-deoxy GTP	0.2 mM	97°C	10 sec	28	95°C 3min and analyse with genetic analyzer.
DMSO	5%	60°C	15 sec	28	
MgCl ₂	2.0 mM	72°C	30 sec	1	
Taq Robust	0.2 U	72°C	10 min	1	
DNA	5 ng/µl	72°C	10 min	1	
Total	20 µl				

KAPA Enhancer and *Taq* Robust (Kapa Biosystems, USA)

Laboratory C**RP-PCR**

List

Primer 1	P1	FAM-AGTCGCTAGAGGCGAAAGC
Primer 2 (with GGGGCC repeat)	P2	tacgcatcccagtttgagacgGGGGCCGGGGCCGGGGCCGGGG
Primer 3	P3	tacgcatcccagtttgagacg
Size standard	S2	GeneScan 500 LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M2	3130xl genetic analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR1)

Products	Final conc.
FastStart PCR Master	1x
Primer 1	1.4 μ M
Primer 2	0.7 μ M
Primer 3	1.4 μ M
7-deaza-2-deoxy GTP	0.18 mM
DMSO	7 %
Q solution	1x
MgCl ₂	0.89 mM
DNA	2.14 ng/ μ l
Total	28 μl

FastStart PCR Master (Roche, Switzerland)
Q solution (QIAGEN, Germany)

PCR protocol (PP3)

Temp	Time	Cycle
95°C	8 min	1
95°C	1 min	8
70-56°C	1 min	Every cycle
72°C	3 min	-2°C
95°C	1 min	
56°C	1 min	32
72°C	3 min	
72°C	10 min	1
4°C	hold	1

Capillary electrophoresis

PCR product	2.5 μ l
Size standard	0.25 μ l
Formamide	7.25 μ l
Total	10.0 μl
95°C 2min and 4°C. Analyse with genetic analyzer.	

Amplicon length analysis

List

Primer 4	P14	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M2	3130xl genetic analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR12)

Products	Final conc.
Primer 4	0.4 μ M
Primer 5	0.4 μ M
7-deaza-2-deoxy GTP	0.2 mM
dNTPs (each)	0.25 mM
DMSO	4 %
Betaine	1.0 M
GeneAmp PCR buffer	1x
AmpliTaQ DNA polymerase	1.25 U
DNA	2.4 ng/ μ l
Total	25.0 μl

AmpliTaQ and GeneAmp PCR buffer (Roche, Switzerland)

PCR protocol (PP12)

Temp	Time	Cycle
95°C	5 min	1
95°C	1 min	
56°C	1 min	30
72°C	2 min	
72°C	10 min	1
4°C	hold	1

Capillary electrophoresis

PCR product	2.5 μ l
Size standard	0.25 μ l
Formamide	7.25 μ l
Total	10.0 μl
95°C 2min and 4°C. Analyse with genetic analyzer.	

Laboratory D**RP-PCR**

List

Primer 1	P4	FAM-tgtaaacgacggccagtCAAGGAGGGAAACAACCGCAGCC
Primer 2 (with GGGGCC repeat)	P13	cacgacgttgtaaacgaCCCCGGCCCCGGCCCCGG
Primer 3	-	none
Size standard	S2	GeneScan 500LIZ Size Standard (Life Technologies, USA)
Analysis machine	M1	3730xL DNA Analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v2.0 (Applied Biosystems, USA)

PCR reaction (PR3)

Products	Final conc.
RG25 buffer	1x
Primer 1	0.2 µM
Primer 2	0.2 µM
dCTP, dATP, dTTP (each)	0.2 mM
7-deaza-2-deoxy GTP	0.2 mM
DMSO	5%
Betaine	1.0 M
Taq DNA polymerase	1.8 U
DNA	2.0 ng/µl
Total	20.0 µl

RG25 buffer (160mM NH₄SO₄, 670mM Tris-HCl (pH 8.8), 1% Tween, 25mM MgCl₂)

PCR protocol (PP4)

Temp	Time	Cycle
95°C	5 min	1
95°C	30 sec	32
68°C	3 min	
72°C	7 min	1
12°C	hold	1

Capillary electrophoresis

PCR product	2.0 µl
Size standard	0.5 µl
Formamide	7.5 µl
Total	10.0 µl
95°C 5min and analyse with DNA analyzer.	

Amplicon length analysis

List

Primer 4	P15	CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Analysis system	A4	4 % agarose gel

PCR reaction (PR13)

Products	Final conc.
Primer 4	0.2 µM
Primer 5	0.2 µM
AccuPrime buffer A	1x
AccuPrime Taq	0.4 U
DNA	4.2 ng/µl
Total	30.0 µl

AccuPrime GC-Rich DNA Polymerase (Invitrogen, USA)

PCR protocol (PP13)

Temp	Time	Cycle
95°C	3 min	1
95°C	30 sec	
68°C	30 sec	35
72°C	2 min	
72°C	10 min	1
12°C	hold	1

Laboratory E**RP-PCR**

List

Primer 1	P4	FAM-tgtaaaacgacggccagtCAAGGAGGGAAACAACCGCAGCC
Primer 2 (with GGGGCC repeat)	P5	caggaacagctatgaccGGGCCCCGCCCCGACCACGCCCCGCCCCG GCCCCG
Primer 3	P6	caggaacagctatgacc
Size standard	S1	GeneScan 500ROX Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR4)**PCR protocol (PP5)****Capillary electrophoresis**

Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H ₂ O=1.5:100
Qiagen buffer	1x	95°C	5 min	1	Diluted PCR product 5.0 µl Size standard 0.07 µl H ₂ O 10.0 µl
Primer 1	0.4 µM	95°C	30 sec	49	
Primer 2	0.2 µM	70-53.2°C	30 sec	Every cycle -0.35°C	Total 15.7 µl
Primer 3	0.4 µM	68°C	40 sec		
ATP, CTP, TTP (each)	0.05 mM				95°C 5min and analyse with DNA analyzer.
GTP	0.0125 mM	95°C	30 sec		
7-deaza-2-deoxy GTP	0.0375 mM	58°C	30 sec	16	
Q solution	1x	68°C	40 sec		
Taq DNA polymerase	0.75 U				
DNA	20 ng/µl	68°C	10 min	1	
Total	25.0 µl	12°C	hold	1	

Taq DNA polymerase (QIAGEN,
Germany)

Amplicon length analysis

List

Primer 4	P14	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Size standard	S1	GeneScan 500ROX Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR10)**PCR protocol (PP14)****Capillary electrophoresis**

Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H ₂ O=1.5:100
Primer 4	0.4 µM	98°C	5 min	1	Diluted PCR product 5.0 µl Size standard 0.07 µl H ₂ O 10.0 µl
Primer 5	0.4 µM	97°C	30 sec	11	
Kit buffer	1x	65-55°C	30 sec	Every cycle -1°C	Total 15.7 µl
DNA	2.0 ng/µl	68°C	1.5 min		
Total	25.0 µl	97°C	30 sec		95°C 5min and analyse with DNA analyzer.
Kit buffer: One Taq 2x Master Mix with GC Buffer (New England BioLabs, UK)		55°C	30 sec	24	
		68°C	1.5 min		
		68°C	10 min	1	
		12°C	hold	1	

Supple table 3D. Methods of each laboratory

Laboratory F**RP-PCR**

List

Primer 1	P10	FAM-AGTACTCGCTGAGGGTGAAC
Primer 2 (with GGGGCC repeat)	P11	cgtacgcatcccagtttgagaGCCCCGCCCCGCCCCGG
Primer 3	P12	cgtacgcatcccagtttgaga
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A5	In-house developed Tracl genotyping software

PCR reaction (PR5)

Products	Final conc.
Buffer 2 (kit)	1x
Primer 1	0.33 μ M
Primer 2	0.33 μ M
Primer 3	0.03 μ M
dNTP (each)	0.5 mM
Betaine	2.0 M
Polymeras (TiTaq 50x)	2x
DNA	10 ng/ μ l
Total	20.0 μl

Expand Long Template PCR system
(Roche, Switzerland)**PCR protocol (PP6)**

Temp	Time	Cycle
98°C	10 min	1
97°C	35 sec	10
53°C	2 min	
68°C	2 min	
97°C	35 sec	25
53°C	2 min	Every cycle
68°C	2-10 min	+20 sec
68°C	10 min	1
4°C	hold	1

Capillary electrophoresis

PCR product	2.0 μ l
Size standard	0.35 μ l
Formamide	10.0 μ l
Total	12.35 μl
95°C 1.5min and analyse with DNA analyzer.	

Amplicon length analysis

List

Primer 4	P17	FAM-CAGGTGTGGGTTTAGGAGGT
Primer 5	P18	CCAGCTTCGGTCAGAGAAAT
Size standard	S4	GeneScan 600LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A5	In-house developed Tracl genotyping software

PCR reaction (PR14)

Product	Final conc.
Primer 4	0.3 μ M
Primer 5	0.3 μ M
dNTPs (each)	0.3 mM
Betaine	1.0 M
Kapa HiFi GC buffer	1x
Polymerase (kit)	0.3U
DNA	3.3 ng/ μ l
Total	15.0 μl

Kapa HiFi Hotstart PCR kit (Kapa Biosystems, USA)

PCR protocol (PP15)

Temp	Time	Cycle
95°C	5 min	1
98°C	20 sec	33
59°C	15 sec	
72°C	1 min	
72°C	5 min	1
4°C	hold	1

Ramping rate: 50%

Capillary electrophoresis

PCR product	2.0 μ l
Size standard	0.35 μ l
Formamide	10.0 μ l
Total	12.35 μl
95°C 1.5min and analyse with DNA analyzer.	

Laboratory G**RP-PCR**

List

Primer 1	P4	FAM-tgtaaacgacggccagtCAAGGAGGGAAACAACCGCAGCC
Primer 2 (with GGGGCC repeat)	P5	caggaaacagctatgaccGGGCCC GCCCGACCGCCCCGCCCCG GCCCCG
Primer 3	P6	caggaaacagctatgacc
Size standard	S3	GeneScan 400HD Rox Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper v4.0 (Applied Biosystems, USA)

PCR reaction (PR6)		PCR protocol (PP7)			Capillary electrophoresis	
Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H2O=1:5	
Qiagen buffer	1x	98°C	10 min	1		
Primer 1	1.0 µM	97°C	35 sec	10	Diluted PCR product	2.0 µl
Primer 2	1.0 µM	64°C	2 min		Size standard	0.12 µl
Primer 3	1.0 µM	68°C	8 min		Formamide	7.88 µl
dCTP, dATP, dTTP (each)	0.25 mM				Total	10.0 µl
7-deaza-2-deoxy GTP	0.25 mM	97°C	35 sec	25	95°C 2min and analyse with DNA analyzer.	
DMSO	5%	64°C	2 min	Every cycle		
Betaine	1.0 M	68°C	8-16 min	+20 sec		
Taq DNA polymerase	2.5 U					
DNA *1	12.5 ng/µl	Reduce ramping rate up and down to 0.5°C/sec.				
Total	20.0 µl					

Taq DNA polymerase, buffer (QIAGEN, Germany)

*1: Initial DNA concentration ~250 ng/µl.

Amplicon length analysis

List

Primer 4	P14	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Size standard	S3	GeneScan 400HD Rox Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper v4.0 (Applied Biosystems, USA)

PCR reaction (PR15)		PCR protocol (PP16)			Capillary electrophoresis	
Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H2O=1:180	
Primer 4	1.0 µM	94°C	4 min	1		
Primer 5	1.0 µM	94°C	30 sec	36	Diluted PCR product	3.0 µl
dCTP, dATP, dTTP (each)	0.25 mM	62°C	30 sec		Size standard	0.10 µl
7-deaza-2-deoxy GTP	0.25 mM	72°C	1 min		HiDi Formamide	6.90 µl
DMSO	5%			1	Total	10.0 µl
Betaine	1.0 M	72°C	10 min		95°C 2min and analyse with DNA analyzer.	
Qiagen buffer	1x					
Taq DNA polymerase	2.5 U					
DNA *2	3.35 ng/µl					
Total	20.0 µl					

Taq DNA polymerase, buffer (QIAGEN, Germany)

*2: Initial DNA concentration ~67 ng/µl.

Supple table 3D. Methods of each laboratory

Laboratory H**RP-PCR**

List

Primer 1	P1	FAM-AGTCGCTAGAGGCGAAAGC
Primer 2 (with GGGGCC repeat)	P2	tacgcatcccagtttgagacgGGGGCCGGGGCCGGGGCCGGGG
Primer 3	P3	tacgcatcccagtttgagacg
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper v4.0 (Applied Biosystems, USA)

PCR reaction (PR1)**PCR protocol (PP8)****Capillary electrophoresis**

Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H ₂ O=1:19
FastStart PCR Master	1x	95°C	5 min	1	
Primer 1	1.4 µM	95°C	30 sec	8	Diluted PCR product 2.0 µl
Primer 2	0.7 µM	70-56°C	30 sec	Every cycle	Formamide mix 8.6 µl
Primer 3	1.4 µM	72°C	3 min	-2°C	Total 10.6 µl
7-deaza-2-deoxy GTP	0.18 mM				
DMSO	7 %	95°C	30 sec		Formamide mix=8µl Size standard + 850µl Formamide
Q solution	1x	56°C	30 sec	35	
MgCl ₂	0.89	72°C	3 min		
DNA	7.1 ng/µl				No denaturation. Analyze with DNA analyzer.
Total	28.1 µl	72°C	20 min	1	
		4°C	hold	1	

FastStart PCR Master (Roche,
Switzerland)

Q solution (QIAGEN, Germany)

Amplicon length analysis

List

Primer 4	P14	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper v4.0 (Applied Biosystems, USA)

PCR reaction (PR16)**PCR protocol (PP17)****Capillary electrophoresis**

Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H ₂ O=1:39
Primer 4	0.5 µM	95°C	4 min	1	
Primer 5	0.5 µM	95°C	30 sec	14	Diluted PCR product 2.0 µl
7-deaza-2-deoxy GTP	0.18 mM	70-57°C	30 sec	Every cycle	Formamide mix 8.6 µl
DMSO	7.1 %	72°C	3 min	-1°C	Total 10.6 µl
Q solution	1x				
MgCl ₂	0.89 mM	95°C	30 sec		Formamide mix=8µl Size standard + 850µl Formamide
FastStart PCR master	1x	56°C	30 sec	26	
DNA	7.1 ng/µl	72°C	3 min		
Total	28.0 µl	72°C	10 min	1	No denaturation. Analyze with DNA analyzer.
		4°C	hold	1	

FastStart PCR Master (Roche,
Switzerland)

Q solution (QIAGEN, Germany)

Laboratory I**RP-PCR**

List

Primer 1	P4	FAM-tgtaaacacgacggccagtCAAGGAGGGAAACAACCGCAGCC
Primer 2 (with GGGGCC repeat)	P5	caggaaacagctatgaccGGGCCCCGCCCGACACGCCCCGGCCCCG GCCCCGG
Primer 3	P6	caggaaacagctatgacc
Size standard	S1	GeneScan 500ROX Size Standard (Applied Biosystems, USA)
Analysis machine	M3	3500 Genetic Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper v4.0 (Applied Biosystems, USA)

PCR reaction (PR7)**PCR protocol (PP7)****Capillary electrophoresis**

Products	Final conc.	Temp	Time	Cycle	PCR product	1.5 µl	
AccuPrime buffer A	1x	98°C	10 min	1	Size standard	0.5 µl	
Primer 1	1.4 µM	97°C	35 sec	10	Formamide	15.0 µl	
Primer 2	0.7 µM	64°C	2 min		25	Total	17.0 µl
Primer 3	1.4 µM	68°C	8 min			Every cycle +20 sec	95°C 5min and analyse with genetic analyzer.
dCTP, dATP, dTTP (each)	0.18 mM	97°C	35 sec	1			
7-deaza-2-deoxy GTP	0.18 mM	64°C	2 min				
Taq DNA polymerase	1.0 U	68°C	8-16 min				
DNA	12 ng/µl						
Total	25.0 µl	16°C	hold				

AccuPrime GC-rich DNA polymerase,
buffer A (Invitrogen, USA)

Ramp rate adjusted to 64% for the
denaturation and extension steps.
Ramp rate adjusted to 25% for the
annealing step.

Amplicon length analysis

List

Primer 4	P15	CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Analysis system	A4	3% agarose gel

PCR reaction (PR13)**PCR protocol (PP18)**

Products	Final conc.	Temp	Time	Cycle
Primer 4	0.2 µM	95°C	3 min	1
Primer 5	0.2 µM	95°C	30 sec	35
Kit buffer	1x	63°C	30 sec	
Kit polymerase	0.4 U	72°C	2 min	
DNA	2.0 ng/µl			1
Total	25.0 µl	72°C	10 min	
Kit: AccuPrime GC-rich DNA polymerase (Invitrogen, USA)		16°C	hold	1

Laboratory J**RP-PCR**

List

Primer 1	P4	FAM-tgtaaacgacggccagtCAAGGAGGGAAACAACCGCAGCC
Primer 2 (with GGGGCC repeat)	P5	caggaaacagctatgaccGGGCCCCGCCCGACCACGCCCGGCCCGGCCCCGG
Primer 3	P6	caggaaacagctatgacc
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730x/DNA Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper v4.1 (Applied Biosystems, USA)

PCR reaction (PR8)		PCR protocol (PP9)			Capillary electrophoresis	
Products	Final conc.	Temp	Time	Cycle	PCR product	1.0 µl
Extensor Mastermix	1x	95°C	5 min	1	Size standard	0.5 µl
Primer 1	2.5 µM	95°C	30 sec	8	Formamide	8.5 µl
Primer 2	2.5 µM	70-56°C	30 sec	Every cycle	Total	10.0 µl
Primer 3	0.25 µM	72°C	8 min	-2°C	95°C 2min and on ice.	
Betaine	1.75 M	95°C	30 sec	26	Analyse with DNA analyzer.	
DNA	3.3 ng/µl	56°C	30 sec	Every cycle		
Total	15 µl	72°C	8-16 min	+20 sec		
Extensor Mastermix (Thermo Scientific, USA)		72°C	7 min	1		

Amplicon length analysis

List

Primer 4	P14	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730x/ DNA Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper v4.1 (Applied Biosystems, USA)

PCR reaction (PR17)		PCR protocol (PP19)			Capillary electrophoresis	
Products	Final conc.	Temp	Time	Cycle	PCR product	1.0 µl
Primer 4	0.1 µM	95°C	5 min	1	Size standard	0.5 µl
Primer 5	0.1 µM	95°C	30 sec	8	Formamide	8.5 µl
DMSO	1%	70-56°C	30 sec	Every cycle	Total	10.0 µl
Megamix Blue	1x	72°C	8 min	-2°C	95°C 2min and on ice.	
DNA	3.3 ng/µl	95°C	30 sec	26	Analyse with DNA analyzer.	
Total	15 µl	56°C	30 sec	Every cycle		
Megamix Blue (Microzone, UK)		72°C	8-16 min	+20 sec		
		72°C	7 min	1		
		4°C	hold	1		

Laboratory L**RP-PCR**

List

Primer 1	P1	FAM-AGTCGCTAGAGGCGAAAGC
Primer 2 (with GGGGCC repeat)	P2	tacgcatcccagtttgagacgGGGGCCGGGGCCGGGGCCGGGG
Primer 3	P3	tacgcatcccagtttgagacg
Size standard	S1	GeneScan 500XL ROX Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730x/DNA Analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR1)

Products	Final conc.
FastStart PCR Master	1x
Primer 1	1.4 µM
Primer 2	0.7 µM
Primer 3	1.4 µM
7-deaza-2-deoxy GTP	0.18 mM
DMSO	7 %
Q solution	1x
MgCl ₂	0.89 mM
DNA	100 ng
Total	28.0 µl

FastStart PCR Master (Roche, Switzerland)
Q solution (QIAGEN, Germany)

PCR protocol (PP1)

Temp	Time	Cycle
95°C	15 min	1
94°C	1 min	2-8 *
70-58°C	1 min	
72°C	3 min	
94°C	1 min	5
56°C	1 min	
72°C	3 min	
72°C	10 min	1

* Every 2°C decrease add 1 cycle.

Capillary electrophoresis

PCR product	2.0 µl
Size standard	0.5 µl
HiDi Formamide	7.0 µl
Total	9.5 µl

95°C 3min and analyse with DNA analyzer.

Amplicon length analysis...not performed.

Laboratory M**RP-PCR**

List

Primer 1	P4	FAM-tgtaaaacgacggccagtCAAGGAGGGAAACAACCGCAGCC
Primer 2 (with GGGGCC repeat)	P5	caggaaacagctatgaccGGGCCCCGCCCCGACCACGCCCCGGCCCCG GCCCCGG
Primer 3	P6	caggaaacagctatgacc
Size standard	S2	GeneScan 500 LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730x/DNA Analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR6)		PCR protocol (PP7)			Capillary electrophoresis	
Products	Final conc.	Temp	Time	Cycle	PCR product	
Qiagen buffer	1x	98°C	10 min	1	PCR product	1.0 µl
Primer 1	1.0 µM	97°C	35 sec	10	Size standard	0.1 µl
Primer 2	1.0 µM	64°C	2 min		Formamide	19.9 µl
Primer 3	1.0 µM	68°C	8 min		Total	21.0 µl
dCTP, dATP, dTTP (each)	0.25 mM	97°C	35 sec	25	95°C 5min and analyse with DNA analyzer.	
7-deaza-2-deoxy GTP	0.25 mM	64°C	2 min	Every cycle		
DMSO	5%	68°C	8-16 min	+20 sec		
Betaine	1.0 M	Ramping up and down: 0.5°C/sec				
Taq DNA polymerase	2.5 U					
DNA	5.0 ng/µl					
Total	20.0 µl					

Taq DNA polymerase, buffer (QIAGEN, Germany)

Amplicon length analysis...not performed.

Supple table 3D. Methods of each laboratory

Laboratory N**RP-PCR**

List

Primer 1	P1	FAM-AGTCGCTAGAGGCGAAAGC
Primer 2 (with GGGGCC repeat)	P2	tacgcatcccagtttgagacgGGGGCCGGGGCCGGGGCCGGGG
Primer 3	P3	tacgcatcccagtttgagacg
Size standard	S1	GeneScan 500ROX Size Standard (Applied Biosystems, USA)
Analysis machine	M2	3130 Genetic Analyzer (Applied Biosystems, USA)
Analysis software	A3	Genotyper v4 (Applied Biosystems, USA)

PCR reaction (PR9)

Products	Final conc.
True Allele PCR premix	1x
Primer 1	1.4 µM
Primer 2	0.7 µM
Primer 3	1.4 µM
7-deaza-2-deoxy GTP	0.16 mM
DMSO	7%
Betaine	0.83 M
MgCl ₂	0.83 mM
DNA	3.3ng/µl
Total	30.0 µl

True Allele PCR premix (Applied Biosystems, USA)

PCR protocol (PP1)

Temp	Time	Cycle
95°C	15 min	1
94°C	1 min	2-8 *
70-58°C	1 min	
72°C	3 min	
94°C	1 min	5
56°C	1 min	
72°C	3 min	
72°C	10 min	1
4°C	hold	1

* Every 2°C decrease add 1 cycle.
Ramp speed: 1°C/sec**Capillary electrophoresis**

PCR product	2.0 µl
Size standard	0.5 µl
Formamide	12.5 µl
Total	15.0 µl

95°C 2min and analyse with DNA analyzer.

Amplicon length analysis...not performed.

**Supplementary table 4. The definitions of an expansion of the GGGGCC-repeat in
C9orf72 as used in 13 laboratories.**

GGGGCC-repeat number (4 laboratories)

- > 24 repeats
- ≥ 30 repeats
- > 30 repeats (2 laboratories)

RP-PCR shape (2 laboratories)

Clear saw-tooth sloping pattern

GGGGCC-repeat number + RP-PCR shape (1 laboratory)

>30 repeats + Stutter profile in RP-PCR

GGGGCC-repeat number + RP-PCR shape + Amplicon-length analysis result (1 laboratory)

Saw-tooth pattern over 460 bp (30 repeats) in RP-PCR + One amplicon in amplicon-length analysis

RP-PCR shape + Amplicon-length analysis result (1 laboratory)

Saw-tooth pattern in RP-PCR + One amplicon in amplicon-length analysis

RP-PCR shape + Amplicon-length analysis result + Southern blot result (4 laboratories)


- Saw-tooth pattern with concave shape in RP-PCR + One amplicon in amplicon-length analysis + Expanded band in Southern blot
 - One amplicon in amplicon-length analysis + Saw-tooth profile in RP-PCR + Expanded band in Southern blot
 - One amplicon in amplicon-length analysis + Long stutter in RP-PCR + Expanded band in Southern blot (2 laboratories)
-

Supplementary table 5. Summary of the Southern blot methods of three laboratories.


	Laboratory I	Laboratory II	Laboratory III
Labelling	Radioisotope	Radioisotope	DIG system
Dose of gDNA	10 µg	10 µg	6-10 µg
Probe primer-F	TGACACACCAAGCGTCA TCT	TATTAAGGTTCGCACAC GCTA	AGAACAGGACAAGTTGCC
Probe primer-R	CACTGTGGGAGCAGTGT CAT	GACGGCTGACACACCAA G	AACACACACCTCCTAAAC C
Probe length	1000bp	210 bp	241 bp
Restriction enzyme	Xba I and Hind III	Xba I and Hind III	Xba I

Supplementary Figure 1a,b,c,d

A



Amplicon-length analysis



129bp: Peak of wild type allele with 2 repeats.
131bp: Peak of mutant allele with 6 repeats.

B


Sequence

WT CGCGCTAGGGGCGGGGCGGGGCGTGGTCGGGGCGGGCCCGGGGCGGGCCCGGGG

Mut. CGCGCTAGGGGCGGGGCGGGGCGGGGCGGGGCGGGGCGGGGCGGGGCGGGCCCGGGG


1 2 deletion 15bp

1 2 3 4 5 6 insertion 17bp




Supplementary figure 1a.b.c.d (cont.)

C



D

Sequence



Supplementary figure 1. Detailed analyses of the results of the three samples with mutations.








(A) The Rp-PCR and amplicon length analysis of sample 8.1. In RP-PCR there are four peaks before the wild-type allele peak at 290 bp. In amplicon-length analysis there are two peaks at 129bp (wild-type allele) and 131bp (the mutant allele), respectively and the two peaks are close together.

As a consequence, four laboratories interpreted this sample as having one amplicon.

- (B) The sequence of sample 8.1. The wild-type allele has two GGGGCC-repeats and the mutant allele has six GGGGCC-repeats with a complex 15 deletion/17bp insertion mutation. Sample 8.1 and Sample 8.2 both carries this mutation.
- (C) The RP-PCR and amplicon length analysis results of sample 32.1. In RP-PCR there are four peaks before the 290bp peak (mutant allele) and a saw-tooth pattern peaks (expanded allele). In amplicon-length analysis there is one peak at 131bp (mutant allele) with 6 repeats.
- (D) The sequence of 32.1 illustrating the mutant allele which has six repeats with a complex 15bp deletion/17bp insertion mutation.




Supplementary Figure 2

Laboratory I




Supplementary Figure 2 (cont.)

Laboratory II



Supplementary Figure II (cont.)

Laboratory III



Supplementary Figure 2. The Southern blot results in three laboratories.

Laboratory III ran out of DNA from one individual (sample 31.1). This laboratory therefore only

performed SB on 77 samples.

PC: Positive control, NC: negative control, +: sample with large repeat expansion, - :sample without large repeat expansion.