

University of Groningen

GRACE Project: model-based control of a nonlinear actuator

Fonseca Aguilar, Oscar; Muñoz Arias, Mauricio; Stavenga, Doekele

Published in:
 Benelux Meeting of Systems and Control

IMPORTANT NOTE: You are advised to consult the publisher's version (publisher's PDF) if you wish to cite from it. Please check the document version below.

Document Version
 Final author's version (accepted by publisher, after peer review)

Publication date:
 2019

[Link to publication in University of Groningen/UMCG research database](#)

Citation for published version (APA):
 Fonseca Aguilar, O., Muñoz Arias, M., & Stavenga, D. (2019). GRACE Project: model-based control of a nonlinear actuator. In *Benelux Meeting of Systems and Control* (2019 ed.). KU Leuven.

Copyright

Other than for strictly personal use, it is not permitted to download or to forward/distribute the text or part of it without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license (like Creative Commons).

The publication may also be distributed here under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license. More information can be found on the University of Groningen website: <https://www.rug.nl/library/open-access/self-archiving-pure/taverne-amendment>.

Take-down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Downloaded from the University of Groningen/UMCG research database (Pure): <http://www.rug.nl/research/portal>. For technical reasons the number of authors shown on this cover page is limited to 10 maximum.

GRACE Project: model-based control of a nonlinear actuator

Óscar Fonseca-Aguilar

Dept. of Mechatronics Engineering
Costa Rica Institute of Technology
P.O. Box 159-7050,
Cartago, Costa Rica

Email: osfonseca@estudiantec.cr

Mauricio Muñoz-Arias and Doekele Stavenga

Faculty of Science and Engineering
University of Groningen
Nijenborgh 4, 9747 AG
Groningen, The Netherlands

Email: m.munoz.arias@rug.nl, d.g.stavenga@rug.nl

1 Introduction

Natural phenomena have proven to be a great source of inspiration for developing specialized devices. Insect compound eyes offer promising insights for the development of new artificial navigation and vision systems [5]. In recent years, a Goniometric Research Apparatus for Compound Eyes (GRACE) has been developed. The aim of the apparatus is to obtain a rapid mapping of compound eyes. Initially a non-automatized system was used [4], [5]. The apparatus has been automatized into a specialized mechatronical device with 6 degrees of freedom, with an upper and lower optical stage and a digital camera (Figure 1). In the last upgrade an auto-focusing algorithm sets the position of the studied insect using the sharpness of the image by controlling the position of Motor Z. Motors X and Y adjust the lateral position, depending on the image details. To control the position of the camera system, which is attached to a Zeiss microscope in the upper stage, a highly nonlinear actuator for a Elero DC motor has been developed. The actuator holds the microscope, which, due to a lead screw, is severely affected by gravity and a high friction coefficient. An energy-based model is here applied using the port-Hamiltonian framework (pH), taking the mass of the system in the screw as an external force, and based on the self-locking property [1]. An optical rotational encoder is the sensor used in the actuator. The proposed controller is a dynamical extension based on the results of [2]. Experimental results can be obtained with a longitudinal accuracy of the microscope position of 5 μ m. The developed GRACE system allows the massive collection of detailed images of insect eyes, i.e. of flies and butterflies.

2 Experimental setup

The GRACE system is shown in Figure 1. First, the M axis is in repose when the DC motor is not powered. The aforementioned is due to the motion transformation mechanism, a lead screw presents a nonlinear behavior, dependant of the direction of motion and friction. The above is considered as an external force entering the rotational system.

References

[1] Budynas, R.G. and Nisbett, J.K., 2008. Shigley's mechanical engineering design (Vol. 8). New York: McGraw-




Figure 1: Schematics of the GRACE system at University of Groningen [3].

Hill.

[2] Dirksz, D.A., Scherpen, J.M. and Ortega, R., 2008, Interconnection and damping assignment passivity-based control for port-Hamiltonian mechanical systems with only position measurements. 47th IEEE CDC, 4957–4962.

[3] Sahu, K.K., 2017, Automated image feature identification for motorized mapping of insect eyes, Master Thesis, University of Groningen.

[4] Stavenga, D.G., 2002. Reflections on colourful ommatidia of butterfly eyes. Journal of Experimental Biology, 205, p.1077–1085.

[5] Stavenga, D.G., 2005, Modern optical tools for studying insect eyes. Methods in insect sensory neuro-science. Edited by TA Christensen. CRC Press, Florida,USA, pp 159-184.

[6] van der Schaft, A.J., 2000, L_2 -gain analysis of nonlinear systems and nonlinear state-feedback h_∞ control. IEEE Trans. on Aut. Contr., 37(6), 770–784.