

University of Groningen

Surgical Applications of Compliant Mechanisms

Thomas, Theodosia Lourdes; Venkiteswaran, Venkatasubramanian Kalpathy;
Ananthasuresh, G. K.; Misra, Sarthak

Published in:
Journal of mechanisms and robotics-Transactions of the asme

DOI:
[10.1115/1.4049491](https://doi.org/10.1115/1.4049491)

IMPORTANT NOTE: You are advised to consult the publisher's version (publisher's PDF) if you wish to cite from it. Please check the document version below.

Document Version
Final author's version (accepted by publisher, after peer review)

Publication date:
2021

[Link to publication in University of Groningen/UMCG research database](#)

Citation for published version (APA):

Thomas, T. L., Venkiteswaran, V. K., Ananthasuresh, G. K., & Misra, S. (2021). Surgical Applications of Compliant Mechanisms: A Review. *Journal of mechanisms and robotics-Transactions of the asme*, 13(2), [020801]. <https://doi.org/10.1115/1.4049491>

Copyright

Other than for strictly personal use, it is not permitted to download or to forward/distribute the text or part of it without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license (like Creative Commons).

The publication may also be distributed here under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license. More information can be found on the University of Groningen website: <https://www.rug.nl/library/open-access/self-archiving-pure/taverne-amendment>.

Take-down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Downloaded from the University of Groningen/UMCG research database (Pure): <http://www.rug.nl/research/portal>. For technical reasons the number of authors shown on this cover page is limited to 10 maximum.

Surgical Applications of Compliant Mechanisms - A Review

Theodosia Lourdes Thomas

Surgical Robotics Laboratory
 Department of Biomechanical Engineering
 University of Twente
 7500 AE Enschede, The Netherlands
 Email: t.l.thomas@utwente.nl

Venkatasubramanian Kalpathy Venkiteswaran

Surgical Robotics Laboratory
 Department of Biomechanical Engineering
 University of Twente
 7500 AE Enschede, The Netherlands
 Email: v.kalpathyvenkiteswaran@utwente.nl

G. K. Ananthasuresh

Multidisciplinary and Multiscale Device and Design Lab
 Department of Mechanical Engineering
 Indian Institute of Science
 Bengaluru, Karnataka 560012, India
 Email: suresh@iisc.ac.in

Sarthak Misra

Surgical Robotics Laboratory
 Department of Biomechanical Engineering
 University of Twente
 7500 AE Enschede, The Netherlands

 Surgical Robotics Laboratory
 Department of Biomedical Engineering
 University of Groningen
 University Medical Centre Groningen
 9713 GZ Groningen, The Netherlands
 Email: s.misra@utwente.nl

Current surgical devices are mostly rigid and are made of stiff materials, even though their predominant use is on soft and wet tissues. With the emergence of compliant mechanisms (CMs), surgical tools can be designed to be flexible and made using soft materials. CMs offer many advantages like monolithic fabrication, high precision, no wear, no friction and no need for lubrication. It is therefore beneficial to consolidate the developments in this field and point to challenges ahead. With this objective, in this paper, we review the application of CMs to surgical interventions. The scope of the review covers five aspects that are important in the development of surgical devices: (i) conceptual design and synthesis, (ii) analysis, (iii) materials, (iv) manufacturing, and (v) actuation. Furthermore, the surgical applications of CMs are assessed by classification into five major groups, namely, (i) grasping and cutting, (ii) reachability and steerability, (iii) transmission, (iv) sensing, (v) implants and deployable devices. The scope and prospects of surgical devices using CMs are also discussed.

BC in the form of bows, which were the primary hunting tools [1]. While reviewing the history of urethral catheterization, Bloom *et al.* [2] noted that ancient Chinese medical texts used lacquer-coated compliant tubular leaves of allium fistulosum (bunched onion) as catheters. They also mention that Sushruta, the author of ancient Indian surgical text, described tubes of gold and silver coated with ghee (clarified butter) used for catheterization. Ancient Greek and Roman surgeons too are known to have used flexible silver tubes in surgery. Over the years, CMs have seen several applications in surgical procedures. Furthermore, the applications of CMs have been extended to aerospace and automotive industries, microelectromechanical systems (MEMS), actuators and sensors, high precision instruments, and robots [3,4].

CMs have gained significant attention in the last few decades as they offer many advantages over traditional rigid-body mechanisms. A CM has monolithic structure, which reduces the number of assembly steps, thus simplifying the fabrication process and requiring reduced maintenance [5]. High precision is attained and the need for lubrication is eliminated due to absence of contact among members that causes wear, friction, backlash, and noise [6].

The merits of CMs have led to a proliferation of studies that implement CM, especially in the medical field [7]. Many variants of CMs have been designed as surgical devices to perform various functions. The structural compli-

1 Introduction

Compliant mechanisms (CMs) are designed to achieve transfer or transformation of motion, force, or energy through elastic deformation of flexible elements. Devices that implement CMs can be traced back to as early as 8000

ance integrated in the main body of a device is exploited to perform object manipulation tasks such as grasping, cutting, retracting and suturing for surgical procedures in the form of ablation, laparoscopy, endoscopy, and biopsy, to mention a few. Additionally, easy miniaturization of CMs enables the device to reach remote difficult-to-access surgical sites as seen in the design of several continuum manipulators [8]. CMs also serve a secondary function in the device to transmit force/motion, as observed in some surgical robots [9, 10]. Applications of CMs are found in microactuators, MEMS and micro-scale surgical devices as well [11–14]. Force sensing using CMs to monitor tool-tissue interaction has also been demonstrated, which serves as a feedback for safe operation of the device inside the human body [15, 16]. The potential of CMs made using biocompatible materials has been realized in the development of biomedical implants, stents and deployable devices [17–19].

There is a growing body of literature that provides a useful account of the design process of CMs [6, 20]. However, there is no detailed investigation into the different aspects to be considered while designing surgical devices using CMs. It poses a problem for those with little to no experience in the medical field on what approach to follow, to go from initial concept to final prototype. This paper aims to provide an overview of this process which involves five major aspects: (i) CM conceptual design and synthesis, (ii) analysis, (iii) material selection, (iv) fabrication methods, and (v) actuation methods. Furthermore, this paper also reviews the existing literature on surgical devices that use CMs by classification into five major groups: (i) grasping and cutting, (ii) reachability and steerability, (iii) transmission, (iv) sensing, (v) implants and deployable devices. We conclude this paper by addressing the associated challenges and provide an outlook on future scope.

2 Design Aspects

This section presents the various methods used during the design process of surgical devices that use CMs. The process begins with synthesis of the CM, followed by optimization to satisfy the intended functional requirements and constraints are identified. Various methods of generating or synthesizing CMs have been explored by researchers. Howell [4] describes four techniques used in the synthesis of CMs: Freedom and Constraint Topologies (FACT); Building Blocks; Topology Optimization; and Rigid-Body-Replacement. Hegde and Ananthasuresh [21, 22] introduced a Selection Maps method for conceptual design and synthesis of CMs. The five aforementioned synthesis methods are explained briefly in Table 1. However, many compliant surgical devices are designed without explicit use of these conventional synthesis methods. This may be because the synthesis methods developed for CMs mostly apply to input-output transmission characteristics rather than guiding and maneuvering. The scope of the expected functions of surgical devices, described later in the paper, offers a huge opportunity for designers. Therefore, the synthesis methods and subsequent classification of devices is not discussed in detail in

this review.

During synthesis of CM, selection of suitable material is crucial to ensure failure prevention. It is generally desirable to have large deformation of a CM, while ensuring the strain is small and the stress stays within limits. This depends on the Young's modulus and the failure strength of the material. From a clinical standpoint, other criteria that need to be considered are the biocompatibility, chemical resistance, elasticity, transparency, strength, temperature resistance, and most importantly, sterilizability of the chosen material [23]. Table 2 describes the materials and different fabrication methods that are suitable for making surgical devices. The four commonly used 3D printing technologies for rapid prototyping compliant surgical devices are also described in Table 2. While punching and blanking technique is used in meso-scale compliant grippers, electrical discharge machining (EDM) is most widely used for micro-scale fabrication of flexure-based continuum manipulators and grippers. Pop-up book MEMS fabrication is an emerging multi-material technique of fabricating MEMS and micro-scale surgical devices. Milling and laser cutting are conventional subtractive manufacturing methods used for surgical manipulators and their constituent parts like wrist and end effector. Although injection molding was not typically used in the making of surgical devices reviewed in this paper, it is an economical way of mass manufacturing implants and medical plastics.

The method of actuation is an important aspect to be considered in the design of a CM. Based on the specific function that the CM serves in the design, various actuation methods have been demonstrated in literature. Table 3 presents commonly-used actuation methods of CMs which are suited to surgical applications, along with their advantages and limitations. Cable-driven actuation is the most widely used method among continuum manipulators and steerable instruments. Shape memory alloys (SMAs) and piezoelectric materials are seen more in high precision devices and for micro/nano manipulation. While fluidic actuation is used in few flexible surgical instruments, there is a gradual increase towards the use of magnetic actuation in designing surgical devices for precise contactless control.

3 Surgical Applications

This section presents a review of the different surgical applications of CMs. The applications of CMs in surgical devices can be broadly classified into five major groups: (i) grasping and cutting, (ii) reachability and steerability, (iii) transmission, (iv) sensing, and (v) implants and deployable devices. Fig. 2 is an overview of this classification showing examples of surgical devices designed for each of these groups of applications, while Fig. 1 depicts the distribution of the number of surgical devices in each group. These are explored in detail in the remainder of this section.

3.1 Grasping and Cutting

CMs have been used to develop forceps, scissors, graspers, and needle holders for performing different sur-

Table 1. Description of synthesis methods for Compliant Mechanisms (CMs), stating their applications and limitations. References of work relevant to each method are provided in square brackets. Numbers in **bold** refer to the papers that describe the general approach of the method and numbers in *italics* refer to the surgical devices reviewed in this paper which are designed using the particular method.

Synthesis Method	Description	Applications and Limitations
Freedom and Constraint Topologies (FACT) [169–171] <i>[none]</i>	Provides topological solution for known freedom space and constraint space based on screw theory, in which twists and wrenches are used to represent constraints and degree-of-freedom of compliant elements.	<ul style="list-style-type: none">• Synthesizing CMs with small to intermediate deflections.• Research on large deformation analysis and representation of elastomechanics, dynamics characteristics and parasitic errors is limited.
Building Blocks [172, 173] <i>[114, 118]</i>	Two main approaches based on: (i) instant centers and compliance ellipsoids, and (ii) flexible building blocks and optimization.	<ul style="list-style-type: none">• Synthesizing CMs with intermediate to large deflections.• Infeasible geometry may result depending on the chosen basic building block.
Topology Optimization [174–176] <i>[118, 41, 112, 134, 146]</i>	Uses optimization algorithms to search for best CM topology to realize the design objective, subject to desired requirements and constraints generally through finite element methods.	<ul style="list-style-type: none">• Most widely-used CM synthesizing method within surgical devices, with its ability to generate solutions from a wide design space.• Difficult to account for localized stresses and buckling. Resulting topologies are sometimes difficult to manufacture, warranting 3D printing or post-processing for manufacturing.
Rigid-Body-Replacement [1, 177–179] <i>[61, 113, 118]</i>	Utilizes the pseudo-rigid-body model to replace compliant members and joints with equivalent rigid links and movable joints, with springs for capturing elastic deformation energy.	<ul style="list-style-type: none">• Reduced-order method that relies on established rigid-body kinematics methods, providing more intuitive analysis.• Accuracy of analysis suffers with increase in the complexity of CM.
Selection Maps [21, 22, 180–182] <i>[none]</i>	Uses a catalog of CMs whose inherent stiffness and inertia characteristics are captured in two-port spring-lever and spring-mass-lever models for matching the user-specifications for the purpose of selection.	<ul style="list-style-type: none">• Can incorporate practical considerations of material selection, manufacturability, strength, and scaling.• Limited to single-input-single-output CMs at present.

Table 2. Description of fabrication methods and materials suitable for compliant mechanisms (CMs) in surgical applications. The pros (+) and cons (–) of each method are described, along with examples of surgical devices made using the given method. Numbers of references are given in square brackets.

Fabrication Method		Materials	Surgical Devices		Pros and Cons
Rapid Prototyping	PolyJet [47, 91, 92, 130, 135, 183, 184]	Biocompatible materials like MED625FLX, MED610 and MED620.	Flexible surgical manipulators, tooltips and catheters.		+ Suitable for small parts with intricate details and printed with high precision.
	Stereolithography (SL) [105, 106, 185]	Photopolymer resins.	Flexible surgical instruments and surgical robot joints.		– Vulnerable to heat and light degradation. – Provides limited mechanical strength.
	Selective Laser Sintering (SLS) [9, 41, 59, 186, 187]	Biocompatible polymers such as polyetheretherketone (PEEK), poly(vinyl alcohol) (PVA), polycaprolactone (PCL) and poly(L-lactic acid) (PLLA).	Surgical robot joints.		+ Uses a wide-variety of materials that provide good mechanical performance.
	Selective Laser Melting (SLM) [60, 188, 189]	Biocompatible metals like steel, titanium alloys and cobalt-chrome.	Surgical continuum manipulators with flexure hinges and bone implants.		– Expensive. – Produces rough surface finish.
Subtractive Manufacturing	Milling [31, 68, 120, 125, 127, 189, 190]	Metals like stainless steel, aluminium and titanium. Plastics like nylon, acrylonitrile butadiene styrene (ABS), polyether ether ketone (PEEK), polyvinyl chloride (PVC).	Surgical manipulators and associated supports like wrist, fixtures and end-effector.		+ Accurate, precise and repeatable machining applicable on a wide variety of materials. – High initial machinery and tooling costs. – Difficult to model complex 3D parts.
	Laser Cutting [49, 126, 191, 192]	Plastics like acrylic, ABS and delrin. And metals like stainless steel, aluminium and titanium.	Endoscopic manipulators and surgical tooltips with intricate patterned cuts.		+ Contactless cutting with accuracy and speed. – Not suitable for cutting parts with very wide thickness. – Releases toxic fumes that needs good ventilation provision.

Fabrication Method		Materials	Surgical Devices	Pros and Cons
Micro-scale Fabrication	Electrical Discharge Machining (EDM) [30, 31, 44, 53, 55, 61, 77, 80, 85, 99, 127, 132, 193, 194]	Conductive materials like titanium, inconel and kovar.	Miniature components like coronary stents, implants, grippers and micro-scale flexures for compliant manipulators.	<ul style="list-style-type: none">+ Suitable to fabricate biocompatible surfaces as it can create an oxide layer on the surface to enhance biological attachment.– Expensive.– Fabricating parts with complex shapes require specially designed fixtures and takes more time.
	Pop-up Book MEMS Fabrication [13, 14, 195]	3D multi-material fabrication using a flexible polyimide layer (Kapton®), by DuPont de Nemours, Inc.) and structural layers (304 Stainless Steel), with adhesive (Dupont FR1500 acrylic adhesive).	MEMS and micro-surgical devices.	<ul style="list-style-type: none">+ Monolithic meso- and micro-structures made can be inserted through small incisions and ‘pop-up’ to perform their function.+ Soft fluidic micro-actuators can also be integrated in the fabrication process.– Risk of peel failure.– Castellated hinge failure due to stress concentrations.
	Punching and Blanking [196]	Sheet form of metals like steel, aluminium and plastics. like PEEK, nylon and delrin.	Meso-scale compliant grippers.	<ul style="list-style-type: none">+ Low cost and fast process.– Cutting complex geometry is difficult.– Negatively affects the quality of edges of cut out part.
Mass Manufacturing	Injection Molding [23, 155, 197, 198]	Plastics like polyvinyl chloride (PVC), styrene acrylonitrile copolymer (SAN), polycarbonate (PC), and polyester. Metals like titanium alloys.	Surgical implants and medical plastics.	<ul style="list-style-type: none">+ Efficient and economic manufacturing method that is automated to produce high output in one step.– High initial tooling costs and long lead times.– For high fatigue resistance and increased lifetime, mold designs for CMs should orient the polymer chain in specific directions.

Table 3. Description of actuation methods for surgical devices, stating their advantages (+) and limitations (−) in surgical applications and integration with CMs. Numbers of references are given in square brackets.

Actuation Method	Surgical Applications	Integration with Compliant Mechanisms
Cable-driven actuation [10, 42, 53, 55, 59, 199–202]	<p>Surgical robotic systems and flexible surgical instruments.</p> <ul style="list-style-type: none"> + Uses lightweight and flexible cables for deformation of the structure. – Miniaturization is challenging due to the associated cables and moment arms. 	<ul style="list-style-type: none"> + Ability to transmit force/motion to remote joints and application points enables convenient location of the actuation unit away from the workspace of the device. – High pretension in cable is necessary to reduce backlash and hysteresis.
Shape memory alloys (SMAs) [98, 150, 151, 154, 203–205]	<p>Internal actuators for instruments like biopsy forceps, hingeless graspers, endoscopic and laparoscopic instruments, among others. Also used in stents, stent grafts and in orthopaedics as correction rods and fracture fixators.</p> <ul style="list-style-type: none"> + Similar hysteresis behaviour with bone and tendons and low sensitivity to MRI. + Shape memory effect provides a collapsible form during insertion and expands after deployment. – Limited by rise in temperature caused by heating. 	<ul style="list-style-type: none"> + Reliable control on actuating CM by training the SMA to fine-tune the performance. + Offers high power-to-weight ratio. + Easy to embed in complex structures. – Generally activated by Joule heating while deactivation takes place via convection heat transfer, which leads to a slow response time.
Piezoelectric materials [43, 44, 47, 143, 206, 207]	<p>Actuators for micro/nano manipulation.</p> <ul style="list-style-type: none"> + Delivers sub-nanometer positioning accuracy and is compact in size. – Expensive to fabricate. 	<ul style="list-style-type: none"> + Offers high response speed. + Large force-to-weight ratio. – Limited by low strain range. – Transmission of forces to remote location is challenging.
Magnetic actuation [38, 51, 61, 92, 104, 208]	<p>Endoscopic devices and surgical instruments with inherent compliance.</p> <ul style="list-style-type: none"> + Precise positioning and control. 	<ul style="list-style-type: none"> + Enables contactless actuation of CM. – Adversely affected upon scaling to large surgical workspace.
Flexible fluidic actuators [14, 31, 209]	<p>Flexible surgical instruments.</p> <ul style="list-style-type: none"> + Safe to operate under radiation and magnetic fields. + Ability of the inflatable membranes to lose and regain their shape facilitates the insertion of instrument inside a patient's body. 	<ul style="list-style-type: none"> + Causes no relative motion between parts, no wear and there is no need for lubrication. – Risk of leakages, and controlling pressure is more complex when compared to electrical signals used in motors and other conventional actuators.




Fig. 1. Contribution of different surgical applications of compliant mechanisms, showing the distribution of the number of surgical devices reviewed in this paper in each application group.

gical tasks such as grasping, cutting, suturing, and holding tissue. For instance, Frecker *et al.* [24, 25] designed a multifunctional compliant instrument with forceps and scissors using topology optimization and fabricated a 5.0 mm diameter stainless steel prototype. Subsequently, a miniaturized prototype was developed by applying size and shape optimization [26, 27]. Recently, a compliant forceps with serpentine flexures was designed to overcome the problem of parallel motion found in traditional forceps with ‘U’ shaped flexure [28]. Cronin *et al.* [29] demonstrated an endoscopic suturing instrument by optimizing a compliant design that provides sufficient puncture force with maximum distal opening of the suture arms.

Several forms of grasping tools have been investigated which utilize the flexibility and stiffness that a CM can offer with different geometry, materials and fabrication techniques. For example, an underactuated compliant gripper made of five phalanges was designed to have large shape-adaptation capability and the deformation was shared by many joints so as to increase the lifetime of the device [30]. A polymer-based MIS shaft instrument was developed using a hybrid effector mechanism combining compliant joints and conventional pin joints [31]. A three-fingered laparoscopic grasper for finger articulation was designed using flexures, leading to distribution of the grasping force, and thereby minimizing tissue perforation [32]. A multi-material design was utilized for a compliant narrow-gauge surgical forceps for laparoscopic and endoscopic procedures [33]. Large grasping forces were realized through a hybrid design approach by having some regions with high stiffness and other regions with greater flexibility to provide larger jaw openings. In subsequent work, a design optimization routine was carried out to maximize the tool performance, validating the grasping potential of a meso-scale contact-aided compliant forceps [34, 35]. Recently, the grasping performance of a compliant surgical grasper was enhanced by functional grading which introduces material with elastic nonlinearity at certain segments of the grasper, while reducing the maximum overall stress [36].

The introduction of robot-assisted surgery has led to many designs of CM-based grasping end effectors, in order to deliver efficient manipulation with high dexterity. Piccin *et al.* [37] showed that a flexible needle grasping device for medical robots has a higher threshold force and stiffness before slipping, compared to a rigid-body needle grasping device. In another work by Forbrigger *et al.* [38], the distal dexterity of a brain tissue resection robot was enhanced by a magnetically-driven forceps made with flexible beams and eliminating the need for an external mechanical or electrical transmission to actuate the end effector.

The monolithic nature of CMs makes them easier to fabricate when compared to the pivoted jaw configurations of current grasping tools [39]. Hence, CM was used in developing a disposable compliant forceps for HIV patients in which, the Q-joints methods was employed to replace a conventional pin-joint [40]. Later, Sun *et al.* [41] synthesized the shape of a disposable compliant forceps for traditional open surgical applications using topology optimization. Subsequently, an adaptive grasping function of the forceps to overcome damaging sensitive organs during both open surgery and robot-assisted minimally invasive surgery (MIS) was devised using topology optimization [42].

At micrometer scale, CM-based microgrippers and micromanipulators have been developed based on flexure hinges and cantilever beam structures. A microgripper made up of piezoelectric bending unimorphs was demonstrated by Haddab *et al.* [43]. Accurate manipulation of a hybrid compliant gripper was achieved using a combination of flexure hinges and a bias spring [44]. Ease in grasping and accurate tool positioning of a micro-forceps was provided by optimizing the jaw design to minimize actuation force, internal stresses, and size [45]. Yang *et al.* [46] demonstrated the opening and closing of the jaws of a compliant micrograsper and microcutter for ophthalmic surgery, by using a cylindrical package tube pulled through the device. While the use of CMs contributes to the elimination of Coulomb friction and backlash, they have some inherent drawbacks. As noted in the design of a low cost flexure-based handheld mechanism for micromanipulation, a drift in the major axis is caused by the imperfect rotation of most compliant joints [47]. Flexure hinges have limited range of angular motion depending on the geometry and material properties of the hinge, and cantilever structures fail to produce perfect parallel motion [44, 48]. However, topology optimization aided by intuition has been used to design CM grippers with parallel-jaw motion.

3.2 Reachability and Steerability

This section describes applications of CMs to increase range of motion and enhance steerability of the surgical instruments to reach difficult to reach surgical sites inside the body. Single-port laparoscopic and endoscopic procedures are adversely affected by limited maneuverability of surgical instruments through confined spaces and narrow visual view inside the human body. Therefore, a steerable endoscopic instrument was developed using three coaxial tubes

Compliant Mechanisms in Surgical Interventions



Fig. 2. An overview of different surgical applications of compliant mechanisms (CMs). Images: Grasper (© 2018 IEEE. Reprinted, with permission, from [36], originally from [210]); Forceps (© 2019 IEEE. Reprinted, with permission, from [38]); Needle holder (Reproduced with permission of ASME, from [37]; permission conveyed through Copyright Clearance Center, Inc.); Suture (Reproduced with permission of ASME, from [29]; permission conveyed through Copyright Clearance Center, Inc.); Cutting (Reproduced with permission of ASME, from [27]; permission conveyed through Copyright Clearance Center, Inc., originally from [25]); Range of reach (© 2019 Simi *et al.*, Reproduced from [51], Licensed under CC BY 4.0); Continuum Manipulator (© 2020 Thomas *et al.*, Reproduced from [61]); Articulated CM (© 2019 IEEE. Reprinted, with permission, from [9]); Tremor compensation (© 2005 IEEE. Reprinted, with permission, from [106]); Dexterous control (Reproduced with permission of ASME, from [107]; permission conveyed through Copyright Clearance Center, Inc.); Pop-up book MEMS (© 2013 IEEE. Reprinted, with permission, from [195]); Statically balanced CM (Reproduced with permission of ASME, from [210]; permission conveyed through Copyright Clearance Center, Inc.); Displacement-amplifying CM (© 2013 IEEE. Reprinted, with permission, from [131]); Force decoupling (© 2012 IEEE. Reprinted, with permission, from [16]); Displacement-amplifying CM (© 2013 IEEE. Reprinted, with permission, from [139]); Orthopedic implant (Image courtesy of Halverson *et al.* (Brigham Young University, USA) from [17]; Intraocular implant (Reprinted from [142], © 2012, with permission from Elsevier); Stent (Reprinted from [151], © 2006, with permission from Elsevier); Heart valve (© 2009 Herrmann *et al.*, Reproduced from [145]).

that slide together concentrically to form a single tube [49]. The design offers additional flexibility due to narrow cuts in the tube and more room in the lumen as the steering mechanism resides in the tubular wall. A review on the different joint types used in the steerable tips of MIS instruments is described by Jelínek *et al.* [50]. To maximize the span of an endoscopic camera, Simi *et al.* [51] modelled a compliant joint in a magnetic levitation system and potential to reduce instrument collision inside the body was shown. Similarly, a flexure-based foldable and steerable CM was reported for providing stereo vision capture in laparoscopic surgery with a pair of miniature cameras [52].

Continuum manipulators are devices that can be precisely steered inside the body to reach difficult-to-access surgical sites. CMs have been used to design flexible miniaturized continuum manipulators for robot-assisted surgery. For example, a 2 degree-of-freedom (DoF) flexible distal tip for enhanced dexterity of endoscopic robot surgery was constructed with a flexible tube cut into a structure consisting of a series of rings connected by thin elastic joints [53, 54]. A similar design was used in a flexible micro manipulator for neurosurgery [55, 56]. A two-section tendon-driven continuum robot with a backbone cut into flexures from a pipe was designed to enhance tip positioning and offer large viewing angles in endoscopic surgery [57, 58]. A multi-arm snake-like robot for MIS was developed using flexible overtube structure as a spine which guides endoscope and other instruments, and two manipulator arms at its tip made of three separate flexure hinge sections [59]. Since beam flexure structures suffer from stress concentrations in the corners, as well as fatigue, a snake-like surgical robot composed of flexible joints based on helical spring was designed [60]. Furthermore, to prevent axial compression, circular rolling contacts were introduced at each turn of the helix. Recently, a contactless mode of actuation and steering of a monolithic metallic compliant continuum manipulator with flexures using magnetic fields was demonstrated [61].

Notched-tube compliant joint mechanisms are variants of aforementioned continuum manipulators, where different shapes, sizes and patterns of notches made on tubes can enable different DoFs and range of motion [62]. For instance, a flexible manipulator arm for single port access abdominal surgery was made from a superelastic nitinol tube with triangular notches [63–66]. A needle-sized wrist made from a nitinol tube with rectangular cutouts was developed to increase the DoF and dexterity of needle laparoscopic surgery (needlescopy) surgical tools [67, 68]. Eastwood *et al.* [69] designed asymmetric notch joints for surgical robots and noted that decreasing the joint's tube diameter and increasing notch depth favours compact bending of the manipulator, but leads to significant reduction in stiffness. Hence, a contact-aided compliant notched-tube joint for surgical manipulation was introduced to improve the stiffness and bending compactness, while operating in confined workspaces [62]. In another work, a cable-driven dexterous continuum manipulator (DCM) comprising two nested superelastic nitinol tubes with notches was designed for removing osteolytic lesions with enhanced volumetric exploration [70–78]. In subse-

quent work, a flexible ring curette made of thin and long pre-curved ring nitinol strips was designed to pass through the open lumen of the DCM [79]. The integration of DCM to a da Vinci actuation box (Intuitive Surgical, Inc., USA) as a hand-held actuator was also shown [80, 81]. In related work, a flexible cutter and an actuation unit to control the DCM were designed to study its buckling behavior during the cutting procedure [82]. The designs of a debriding tool that passes through the lumen of DCM and a steerable drill following a curved-drilling approach to remove lesions were also investigated [83, 84]. Subsequently, by using the curved drilling technique, a bendable medical screw made of two arrays of orthogonal notches along its shaft was devised for internal fixation of bone fractures [85, 86].

Concentric tube robots (CTRs) are a special type of continuum manipulators that are made of multiple precurved elastic tubes that are concentrically nested within one another [87]. CTRs have been deployed for “follow-the-leader” insertion and their steering is not affected by the tissue interaction forces [88]. Thus, they have found several applications as steerable needles and miniaturized surgical manipulators [89].

Some surgical manipulators rely on CMs to enhance articulation. For instance, a compliant articulation structure for surgical tool tips using nitinol was designed to increase the functional workspace and deliver a large blocked force [90]. Other work studied the use of corner- filleted flexure hinge-based compliant joints in a compliant grasper integrated to a 2-DoF surgical tooltip, and circular guide members were added to strengthen the load carrying capacity of the slender compliant joints [91]. Later, a 3-DoF surgical tooltip with modified serpentine flexures and magnetic coupling was developed [92]. Arata *et al.* [93] designed a prototype of 2-DoF articulated laparoscopic surgical instrument using a CM to move two spring blades at the tip. Thereafter, a 4-DoF compliant manipulator was proposed consisting of springs designed to deform locally, reducing the bending radius [94]. A subsequent study on the variation of range of motion and rigidity of elastic moments revealed that to achieve a higher range of motion, there will be a trade-off with the lower values of output force and the precision, and vice versa [95].

The flexibility provided by CMs can be extended to positively affect some specific surgical applications. For instance, a compliant endoscopic ablation probe composed of an array of compliant tines was designed to generate target spherical heating zones, and improve the distribution of heat in the ablation zone [96, 97]. A 3 DoF microrobotic wrist for needlescopy was fabricated using MEMS technology [98, 99]. It was based on a CM derived from a reference parallel kinematics mechanism architecture with three legs, which offered increased instantaneous mobility. A compliant instrument for preparing the subtalar joint for arthroscopic joint fusion was developed, having a shaft design which was compliant in only one direction and stiff in the other two directions to resist and transmit machining forces [100]. In subsequent work, a sideways-steerable instrument joint was designed for meniscectomy that increases range of motion and reachability within the knee joint while operating

through small portal of the body [101, 102]. It consisted of a compliant rolling-contact element (CORE) which was rotated by flexural steering beams configured in a parallel-ogram mechanism. Steerability of kinked bevel-tip needles was improved through the use of a flexure-based needle tip design while minimizing tissue damage, as the flexure keeps the needle in place during insertion [103].

3.3 Transmission

Transmission refers to the use of CMs in augmenting an actuator in the transfer of force, displacement or energy. In some surgical devices, CMs made for force or displacement transmission serve as an input or feedback for the principal function of the device. For example, the translation motion of a medical robot for ENT (ear, nose and throat) surgery was provided using compliant linear joints fabricated by 3D printing [9]. Yim *et al.* [104] showed passive deformation and recovery of a magnetically actuated compliant capsule endoscopic robot, by having its structure based on a Sarrus linkage and circular flexure hinges.

The traditional compliant rolling-contact element (CORE) joint involves joining two half cylinders with flexures. Derived from CORE, the Split CORE was integrated to a wrist design provided by Intuitive Surgical Inc. to create a 3 DoF gripping mechanism [105]. Lan *et al.* [10] developed an adjustable constant-force forceps for robot-assisted surgical manipulation to aid in grasping soft tissues. It employs a compliant constant-torque mechanism made using flexible arms to transmit the required force to forceps tips. The motion of a flexure-based parallel manipulator for an active handheld microsurgical instrument was tracked in order to cancel the hand tremors using piezo-actuators [106]. Awtar *et al.* [107] developed FlexDexTM, a minimally invasive surgical tool frame that is attached to the surgeon's forearm to enhance dexterity and provide intuitive control. The design projects a two-DoF virtual center of rotation for the tool handle at the surgeon's wrist using transmission strips, making it stiff about one axis and compliant in the orthogonal axis.

In microsurgery applications, the concept of pop-up book MEMS has found a few applications. For example, pop-up components made of flexible hinges were designed to realize an articulating microsurgical gripper and a flexural return spring to passively open the gripper [13]. A multi-articulated robotic arm was fabricated by introducing soft elastomeric materials into the pop-up book MEMS process, and mounted on top of an endoscope model demonstrating potential surgical applications such as tissue retraction [14].

A drawback of CMs is that energy efficiency is challenged due to energy storage in the flexible members of the mechanism [108]. Herder and Van Den Berg [109] introduced the principle of a statically balanced compliant mechanism (SBCM) to circumvent this problem for a partially compliant statically balanced laparoscopic grasper (SBLG), in which a negative stiffness mechanism negates the elastic forces of the CM. Drent and Herder [110] developed a numerical optimization model for total range of motion of a

SBLG with normal springs (with non-zero free length) and a constant force transfer function. Powell and Frecker [111] designed a compensation mechanism of a compliant forceps for ophthalmic surgery using a rigid link slider-crank mechanism with a nonlinear spring, which balances the potential energy of the CM. de Lange *et al.* [112] used topology optimization for a SBCM, which resulted in reduced actuation force of a SBLG. Tolou and Herder [113] modelled a partially compliant SBLG using pairs of pre-stressed initially-curved pinned-pinned beams made of linear elastic material that resulted in reduced Von Mises stress and balancing error. Hoetmer *et al.* [114] investigated a building block approach in designing SBCM, since the pseudo-rigid-body method and the topology optimization did not consider an optimization process and the stress constraints, respectively. Subsequently, the first physical demonstration of SBCM with fully compliant elements was shown by taking into account stiffness, range of motion, and stress [115]. Lassooij *et al.* [116] used pre-curved straight-guided beams that are preloaded collinear with the direction of actuation of a fully compliant SBLG with a near zero stiffness, also demonstrating its bi-stable behaviour. Earlier, Stapel and Herder [117] had carried out a feasibility study of a fully compliant SBLG using the pseudo-rigid-body method. In subsequent work, Lamers *et al.* [118] developed a fully compliant SBLG with zero stiffness and zero operation force.

3.4 Sensing

Sensing application refers to the use of CMs in detecting or measuring physical quantities. Several kinds of sensors rely on the change in deflection or stiffness of CMs in conjunction with other transducers like optical sensors and strain gauges to measure physical parameters. Alternatively, vision-based force sensing integrated with miniature grippers was reported by Reddy *et al.* [119]. Subsequently, a compliant end-effector to passively limit the force in tele-operated tissue-cutting using the vision-based force sensing for haptic feedback was demonstrated [120].

Force sensing forms an integral part of different surgical applications that involve tissue palpation, pulling and pushing of tissue during biopsy, to name a few. A miniature microsurgical instrument tip force sensor during robot-assisted manipulation was developed using a double-cross flexure beam configuration [121]. It can provide uniform force sensitivity in all directions at the instrument tip by altering the vertical separation between the beam crosses. A force-torque transducer based on flexural-jointed Stewart platform was integrated to an MIS instrument's tip to enable 6-axis force sensing capability [122].

Magnetic resonance imaging (MRI)-compatible force sensors, in particular, benefit from a CM-based design as the metallic and electric elements can be placed outside MRI. The force sensing element typically consists of an elastic body which deforms under the influence of an applied force, which in turn is measured by a transducer like optical fiber. For example, high accuracy and high sensitivity to displacement was demonstrated using optical micrometry by sup-

porting the force detector with thin annular plates which convert applied force into minute displacement [123]. Later, a parallel plate structure was chosen to design a uniaxial force sensor due to its directionality and simplicity, offering better accuracy including hysteresis characteristics and axial interference than the previous design [124].

Different types of flexible elements can be adapted in the design of force sensors. Analysing the mechanical design of sensing elements, a polymer torsion beam guided in rotation by a ball bearing and supported by compliant linkages was proposed in the development of an MRI-compatible torque sensor [125]. The sensor design was further improved for a 2-DoF haptic interface by using a sensing body made of two blades fixed between the optical head and the reflective target [126]. The blade causes a displacement of the optical head upon application of force by the subject and prevents deformation in other directions, thereby minimizing cross-sensitivity. Later, an ultrasonic motor torque sensor using flexible hinges was also developed [127]. A 3-axis optical fiber force sensor for MRI applications was designed using a 3-DoF compliant platform made of 3 identical cantilever beams with their supports, offering flexibility in response to axial forces and bending moments and high stiffness to withstand axial torque [128]. A 3-axis optical force sensor made of two parallelogram-like segments of helical circular engravings that can provide intrinsic axial/ lateral overload protection during prostate needle placement was developed [129]. Similarly, a triaxial catheter tip force sensor having flexures and integrated reflector was developed for cardiac procedures [130]. The flexures are designed so that the axial and lateral forces cause different deformation of the flexures which leads to different amounts of light getting reflected and detected by the photo detectors.

A challenge with multi-axial force sensors lies in the decoupling of forces along the axes as observed in the study by Gao *et al.* [131]. Linear decoupling methods proved to be inaccurate since local deformation of flexures affects the strains measured. A method to decouple pulling and grasping forces of a 2-DOF compliant forceps was derived using the serial connections of two torsional springs which was realized by optimizing the shape of two circular-type flexure hinges [16]. However, rotational perturbation of forceps, sideways forces acting at the forceps, and fabrication errors introduced disturbances in the force measurement. Gonenc *et al.* [45] demonstrated axial-transverse force decoupling in their flexure design of micro-forceps for robot-assisted vitreoretinal surgery. Peirs *et al.* [132] decoupled the deformations caused by axial and radial forces of a micro optical force sensor for minimally invasive robotic surgery, using four identical parallelograms placed in an axisymmetric arrangement. Fifanski *et al.* [133] developed a flexure-based in-vivo force sensor that can measure forces in 3D using individual optical fibers. As flexure-based force sensors cause undesirable transverse moments, twists and lateral deflections, making it difficult to measure forces along the different axes, Tan *et al.* [134] presented a potential solution of decoupling the force measurements using topology optimization to design the elastic frame structure.

Other factors to be considered while designing force sensors include thermal sensitivity, hysteresis, plastic deformation and friction due to contact between internal components that can alter the elastic behaviour of flexures [135]. Kumar *et al.* [136] developed a force sensor using a compliant version of the Sarrus mechanism and strain gauges. Their elastic model could not address the hysteresis, viscoelastic effects, and non-linearities in the prototype caused by fabrication process. To increase the sensitivity of force sensors, Krishnan and Ananthasuresh [137] evaluated several displacement-amplifying compliant mechanisms (DaCMs) and proposed a general design methodology using application-specific topology optimization. Furthermore, a study by Turkseven and Ueda [138, 139] showed that a DaCM-based force sensor with lower sensitivity can enhance the performance of the sensor by reducing hysteresis and improving signal-to-noise ratio. CMs can also be used to passively sense force and respond in surgical situations. An instance of this was discussed in the context of endoscopy simulation [140], which could also be used in virtual surgical trials. In this work, a CM was designed to convert radial force experienced by the inner rim of a ring into circumferential motion of the ring that can be measured using an encoder.

3.5 Implants and Deployable Devices

Implants are medical devices embedded inside the body via surgery to replace or enhance damaged biological tissue. Within this review, different applications of implants designed using CMs are discussed. FlexSuReTM, a spinal implant based on the geometry of Lamina Emergent Torsional (LET) joint was developed to restore normal motion to the degenerate spine [141]. The LET joint is made from a lamina, and torsion of beams results in flexibility in multiple directions similar to the intervertebral disc. An intraocular implant with CM-based silicon linkages was designed to amplify the displacement of a piezoelectric bender and provide an almost tilt-free translational displacement of the lens for optical imaging quality [142]. Krucinski *et al.* [143] showed that the flexural stresses of bioprosthetic heart valves can be reduced by incorporating a flexible or expansile supporting stent into the valve design.

Within the context of this paper, deployable devices refer to CMs designed to change in shape and size that facilitate insertion of the surgical device in a compact form to reduce invasiveness of the procedure. For example, Chen *et al.* [144] designed an intra-cardiac magnetic resonance imaging (ICMRI) catheter consisting of folded imaging coil during vascular navigation (4.5 mm in diameter). Upon deployment, it forms a circular loop (40 mm in diameter) to image a 40mm field of view. Herrmann *et al.* [145] developed a bistable heart valve prosthesis that can be folded inside a catheter and percutaneously inserted for delivery to the patient's heart for implantation. In designing cardiovascular stents, topology optimization was used to generate optimal geometry of stent cells and maximize the stiffness of the point of application of forces, thereby maintaining structural integrity [146]. However, plastic strains can cause

non-uniformity in the expanded portion of the stent. Hence, James *et al.* [18] used topology optimization to design a bi-stable stent that snaps-through to a stable expanded configuration, relying on the geometric non-linearity of the structure.

Origami-based designs have emerged as a powerful tool in developing deployable devices for MIS [19]. According to Edmondson *et al.* [147], “Origami can be viewed as a compliant mechanism when folds are treated as joints and panels as links.” A pair of origami-inspired surgical forceps was developed to ease the fabrication and sterilization process of robotic forceps. Increase in flexibility while maintaining rigidity was achieved by utilizing multi-layer lamina emergent mechanisms (MLEMs) in the design process. (MLEMs are a type of CM made from multiple sheets (lamina) of material with motion out of plane of fabrication, to achieve specific design objectives [148]). Subsequently, small grippers (3 mm in diameter) were developed for the Intuitive Surgical’s da Vinci robotic surgical systems which can be deployed inside the body during surgery [149]. Salerno *et al.* [150] integrated an origami parallel module to generate rotations and translation of a compliant gripper. Recently, Kuribayashi *et al.* [151] designed a self-deployable origami stent graft using hill and valley folds. Bobbert *et al.* [152] fused the origami, kirigami, and multi-stability principles to fabricate deployable meta-implants. It was also shown that the mechanical properties of the implant can gradually increase, depending on the design of kirigami cut patterns that determine the porous structures allowing bone regeneration. Halverson *et al.* [17] developed a disc implant based on CORE to mimic the biomechanics of human spine. Later, Nelson *et al.* [153] demonstrated a deployable CORE joint (D-CORE) using curved-folding origami techniques to enable transition from a flat state to a deployed functioning state. Origami works well with flexible non-metallic materials, thus making them ideal for MRI-guided procedures which is hazardous in the presence of magnetic materials. Recently, an MR-conditional SMA-based origami joint using CORE for potential applications in endoscopy was demonstrated [154].

4 Discussion

This study set out with the aim of assessing the utility of CMs in designing surgical devices. There are some challenges that hinder the further development and implementation of these devices in clinical practice. A drawback concerning CMs is the adverse effect of stress concentrations and fatigue, especially in flexure-based designs under cyclic loading. This is a major challenge in the medical field where device failure is not acceptable. To tackle this issue, there is a growing interest towards developing multi-material CMs [155–158] and functional grading of CMs [36, 159], to enhance structural integrity. The emerging concept of the so-called 4D printing ushers in many more possibilities for using CMs in surgical applications [160]. This technology can strengthen mechanical properties and create multi-material programmable structures made of elastomers and soft ac-

tive materials like shape memory polymers which react to environment stimuli such as temperature, moisture and magnetic field. Soft robotics is another emerging field of interest which utilizes flexibility to function but is not classified under CMs. Inspired by the softness and body compliance of biological systems, continuum devices based on soft robotics systems are designed using compliant materials [161].

The behaviour of CMs with geometric nonlinearity caused by large deflections is disregarded in many studies described in Section 3. Researchers have investigated this behaviour of CMs using topology synthesis and other non-linear modelling methods. It is beyond the scope of this paper to discuss these approaches, and readers are advised to refer to the following works: [162–166]. An interesting finding of this study is the pivotal role of CMs in developing a new class of force sensors for surgical procedures. However, much uncertainty still exists on the underlying convoluted issues of hysteresis, plastic deformation, among others as discussed in Section 3.4. There is scope for improvement by analysing and understanding the deformation of flexible members of CMs under these complex conditions.

This review highlights the merits of CMs over conventional rigid body mechanisms due to elimination of joint friction, backlash, wear, and need for lubrication. This aspect is leveraged by integration of CMs with modern actuators such as magnets, SMAs, and piezoelectric materials [167]. However, a major challenge lies in analysing an overall system of CM consisting of multiple flexible members. While the monolithic nature of most of the CMs simplifies the fabrication and assembly processes, the flip side is that the whole design may fail if even one part of the mechanism breaks. It is infeasible to restore and modify CM-based designs for quick testing and improvement. Since the key functioning of CMs depends on the stiffness and the resulting deformation, accurate fabrication is critical, which can lead to higher production costs and lead time.

From a clinical standpoint, the protection of instruments from contamination due to contact with fluids is important. As a potential solution, some researchers have suggested soft elastic coating of the instrument [61, 130, 168]. However, further analysis of the implications of *in-vivo* operating conditions on the instrument’s performance, while maintaining sterilization, is necessary.

5 Conclusions

An overview of the design aspects of CMs in surgical interventions is presented in this paper, discussing design methodology, material selection and failure prevention, fabrication, and actuation methods. CMs provide many advantages such as reduction of assembly steps, high precision, accuracy and repeatability with the elimination of backlash, friction and wear. This study has identified the virtues of elastic deformation of compliant members in achieving desired functions tailored for diverse surgical applications including but not limited to laparoscopy, endoscopy, ablation, ENT surgery, vitreoretinal surgery, to robot-assisted surgical interventions. The challenges associated with these applica-

tions related to biocompatibility of surgical instrument, fatigue, stress concentration, energy efficiency, fabrication and complex modelling methods of CMs are discussed. The domain of CMs is a niche area of research that has seen tremendous growth in the last few decades and has raised many questions in need of further investigation. The analysis undertaken here extends our existing knowledge of CMs and offers valuable insights for future research. This would help in paving the way towards seamless integration of CMs in designing safe, dexterous, efficient and cutting-edge surgical devices.

Acknowledgements

The authors would like to thank Jyoti Sonawane for her assistance in the literature review. This research has received funding from the European Research Council (ERC) under the European Union's Horizon 2020 Research and Innovation programme (ERC Proof-of-Concept Grant Agreement #790088—project INSPIRE) and the Netherlands Organization for Scientific Research (Innovational Research Incentives Scheme Vidi: SAMURAI project #14855).

References

- [1] Howell, L. L., 2001. *Compliant mechanisms*. John Wiley & Sons.
- [2] Bloom, D. A., McGuire, E. J., and Lapides, J., 1994. "A brief history of urethral catheterization". *The Journal of Urology*, **151**(2), pp. 317–325.
- [3] Fowler, R., Howell, L., and Magleby, S., 2011. "Compliant space mechanisms: a new frontier for compliant mechanisms". *Mech. Sci.*, **2**(2), pp. 205–215.
- [4] Howell, L. L., Magleby, S. P., Olsen, B. M., and Wiley, J., 2013. *Handbook of Compliant Mechanisms*. Wiley Online Library.
- [5] Howell, L. L., 2013. "Compliant mechanisms". In *21st Century Kinematics*. Springer, pp. 189–216.
- [6] Gallego, J. A., and Herder, J., 2009. "Synthesis methods in compliant mechanisms: An overview". In ASME 2009 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference, pp. 193–214.
- [7] Kota, S., Lu, K.-J., Kreiner, Z., Trease, B., Arenas, J., and Geiger, J., 2005. "Design and application of compliant mechanisms for surgical tools". *Journal of biomechanical engineering*, **127**(6), pp. 981–989.
- [8] Burgner-Kahrs, J., Rucker, D. C., and Choset, H., 2015. "Continuum robots for medical applications: A survey". *IEEE Transactions on Robotics*, **31**(6), pp. 1261–1280.
- [9] Entsfellner, K., Kuru, I., Maier, T., Gumprecht, J. D., and Lüth, T., 2014. "First 3D printed medical robot for ENT surgery - Application specific manufacturing of laser sintered disposable manipulators". In 2014 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS), pp. 4278–4283.
- [10] Lan, C.-C., and Wang, J.-Y., 2011. "Design of adjustable constant-force forceps for robot-assisted surgical manipulation". In 2011 IEEE International Conference on Robotics and Automation (ICRA), pp. 386–391.
- [11] Li, L., and Chew, Z. J., 2018. "Microactuators: Design and technology". In *Smart Sensors and MEMS*. Elsevier, pp. 313–354.
- [12] Lobontiu, N., 2017. *System dynamics for engineering students: Concepts and applications*. Academic Press.
- [13] Gafford, J. B., Kesner, S. B., Wood, R. J., and Walsh, C. J., 2013. "Microsurgical Devices by Pop-Up Book MEMS". In ASME 2013 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference.
- [14] Russo, S., Ranzani, T., Gafford, J., Walsh, C. J., and Wood, R. J., 2016. "Soft pop-up mechanisms for micro surgical tools: Design and characterization of compliant millimeter-scale articulated structures". In 2016 IEEE International Conference on Robotics and Automation (ICRA), pp. 750–757.
- [15] Gassert, R., Moser, R., Burdet, E., and Bleuler, H., 2006. "MRI/fMRI-compatible robotic system with force feedback for interaction with human motion". *IEEE/ASME Transactions on Mechatronics*, **11**(2), pp. 216–224.
- [16] Hong, M. B., and Jo, Y.-H., 2012. "Design and evaluation of 2-DOF compliant forceps with force-sensing capability for minimally invasive robot surgery". *IEEE Transactions on Robotics*, **28**(4), pp. 932–941.
- [17] Halverson, P. A., 2010. "Modeling, design, and testing of contact-aided compliant mechanisms in spinal arthroplasty". *PhD Thesis. Brigham Young University-Provo*.
- [18] James, K. A., and Waisman, H., 2016. "Layout design of a bi-stable cardiovascular stent using topology optimization". *Computer Methods in Applied Mechanics and Engineering*, **305**, pp. 869–890.
- [19] Johnson, M., Chen, Y., Hovet, S., Xu, S., Wood, B., Ren, H., Tokuda, J., and Tse, Z. T. H., 2017. "Fabricating biomedical origami: A state-of-the-art review". *International journal of computer assisted radiology and surgery*, **12**(11), pp. 2023–2032.
- [20] Gallego, J. A., and Herder, J., 2009. "Classification for literature on compliant mechanisms: a design methodology based approach". In ASME 2009 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference, pp. 289–297.
- [21] Hegde, S., and Ananthasuresh, G., 2010. "Design of Single-Input-Single-Output Compliant Mechanisms for Practical Applications Using Selection Maps". *Journal of Mechanical Design*, **132**(8).
- [22] Hegde, S., and Ananthasuresh, G., 2012. "A Spring-mass-lever Model, Stiffness and Inertia Maps for Single-input, Single-output Compliant Mechanisms". *Mechanism and machine theory*, **58**, pp. 101–119.

- [23] Amellal, K., Tzoganakis, C., Penlidis, A., and Rempel, G. L., 1994. "Injection molding of medical plastics: a review". *Advances in Polymer Technology: Journal of the Polymer Processing Institute*, **13**(4), pp. 315–322.
- [24] Frecker, M. I., Dziedzic, R., and Haluck, R., 2002. "Design of multifunctional compliant mechanisms for minimally invasive surgery". *Minimally Invasive Therapy & Allied Technologies*, **11**(5-6), pp. 311–319.
- [25] Frecker, M. I., Powell, K. M., and Haluck, R., 2005. "Design of a multifunctional compliant instrument for minimally invasive surgery". *Journal of Biomechanical Engineering*, **127**(6), pp. 990–993.
- [26] Aguirre, M. E., and Frecker, M., 2006. "Design of a 1.0 mm multifunctional forceps-scissors instrument for minimally invasive surgery". In ASME 2006 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference, pp. 557–563.
- [27] Aguirre, M. E., and Frecker, M., 2008. "Design innovation size and shape optimization of a 1.0 mm multifunctional forceps-scissors surgical instrument". *Journal of Medical Devices*, **2**(1), p. 015001.
- [28] George B, L., and Bharanidaran, R., 2020. "A novel design of compliant forceps with serpentine flexures". *Australian Journal of Mechanical Engineering*, pp. 1–8.
- [29] Cronin, J. A., Frecker, M. I., and Mathew, A., 2008. "Design of a compliant endoscopic suturing instrument". *Journal of Medical Devices*, **2**(2), p. 025002.
- [30] Doria, M., and Birglen, L., 2009. "Design of an under-actuated compliant gripper for surgery using nitinol". *Journal of Medical Devices*, **3**(1).
- [31] Disch, A., Lutze, T., Schauer, D., Mueller, C., and Reinecke, H., 2008. "Innovative polymer-based shaft instruments for minimally invasive surgery". *Minimally Invasive Therapy & Allied Technologies*, **17**(5), pp. 275–284.
- [32] O'Hanley, H., Rosario, M., Chen, Y., Maertens, A., Walton, J., and Rosen, J., 2011. "Design and Testing of a Three Fingered Flexural Laparoscopic Grasper". *Journal of Medical Devices*, **5**(2), p. 027508.
- [33] Aguirre, M. E., and Frecker, M., 2010. "Design and optimization of hybrid compliant narrow-gauge surgical forceps". In ASME 2010 Conference on smart materials, adaptive structures and intelligent systems, pp. 779–788.
- [34] Aguirre, M., Hayes, G., Meirom, R., Frecker, M. I., Muhlstein, C., and Adair, J. H., 2011. "Optimal design and fabrication of narrow-gauge compliant forceps". *Journal of Mechanical Design*, **133**(8).
- [35] Aguirre, M. E., and Frecker, M., 2011. "Design of a multi-contact-aided compliant mechanism". In ASME 2011 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference, pp. 255–259.
- [36] Jovanova, J., Nastevska, A., Frecker, M., and Aguirre, M. E., 2018. "Analysis of a functionally graded compliant mechanism surgical grasper". In 2018 International Conference on Reconfigurable Mechanisms and Robots (ReMAR), pp. 1–8.
- [37] Piccin, O., Kumar, N., Meylheuc, L., Barbé, L., and Bayle, B., 2012. "Design, development and preliminary assessment of grasping devices for robotized medical applications". In ASME 2012 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference, pp. 65–73.
- [38] Forbrigger, C., Lim, A., Onaizah, O., Salmanipour, S., Looi, T., Drake, J., and Diller, E. D., 2019. "Cableless, magnetically driven forceps for minimally invasive surgery". *IEEE Robotics and Automation Letters*, **4**(2), pp. 1202–1207.
- [39] Canfield, S., Edinger, B., Frecker, M. I., and Koopmann, G. H., 1999. "Design of a piezoelectric inchworm actuator and compliant end effector for minimally invasive surgery". In Smart Structures and Materials 1999: Smart Structures and Integrated Systems, Vol. 3668, International Society for Optics and Photonics, pp. 835–843.
- [40] Shuib, S., Yusoff, R., Hassan, A., Ridzwan, M., and Ibrahim, M., 2007. "A disposable compliant-forceps for hiv patients". *Journal of Medical Science*, **7**(4), pp. 591–596.
- [41] Sun, Y., Liu, Y., Xu, L., and Lueth, T. C., 2019. "Design of a disposable compliant medical forceps using topology optimization techniques". In 2019 IEEE International Conference on Robotics and Biomimetics (ROBIO), pp. 924–929.
- [42] Sun, Y., Liu, Y., Xu, L., Zou, Y., Faragasso, A., and Lueth, T. C., 2020. "Automatic design of compliant surgical forceps with adaptive grasping functions". *IEEE Robotics and Automation Letters*, **5**(2), pp. 1095–1102.
- [43] Haddab, Y., Chaillet, N., and Bourjault, A., 2000. "A micrograsper using smart piezoelectric actuators". In Proceedings of 2000 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS), Vol. 1, pp. 659–664.
- [44] Zubir, M. N. M., and Shirinzadeh, B., 2009. "Development of a high precision flexure-based micrograsper". *Precision Engineering*, **33**(4), pp. 362–370.
- [45] Gonenc, B., Handa, J., Gehlbach, P., Taylor, R. H., and Iordachita, I., 2013. "Design of 3-DOF force sensing micro-forceps for robot assisted vitreoretinal surgery". In 2013 35th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC), pp. 5686–5689.
- [46] Yang, M., Culkar, K. M., Powell, K., Frecker, M. I., and Zahn, J. D., 2004. "Design and fabrication of a UV-LIGA compliant micrograsper for ophthalmic surgery". In ASME 2004 International Mechanical Engineering Congress and Exposition, pp. 159–164.
- [47] Tan, U.-X., Latt, W. T., Shee, C. Y., and Ang, W. T., 2010. "A low-cost flexure-based handheld mechanism for micromanipulation". *IEEE/ASME Transactions on*

- Mechatronics*, **16**(4), pp. 773–778.
- [48] Goldfarb, M., and Celanovic, N., 1999. “A flexure-based gripper for small-scale manipulation”. *Robotica*, **17**(2), pp. 181–187.
 - [49] Dewaele, F., Kalmar, A. F., De Ryck, F., Lumen, N., Williams, L., Baert, E., Vereecke, H., Kalala Okito, J. P., Mabilde, C., Blanckaert, B., et al., 2014. “A novel design for steerable instruments based on laser-cut nitinol”. *Surgical Innovation*, **21**(3), pp. 303–311.
 - [50] Jelínek, F., Arkenbout, E. A., Henselmans, P. W., Pessers, R., and Breedveld, P., 2015. “Classification of joints used in steerable instruments for minimally invasive surgery—a review of the state of the art”. *Journal of Medical Devices*, **9**(1).
 - [51] Simi, M., Tolou, N., Valdastrì, P., Herder, J., Mencassi, A., and Dario, P., 2012. “Modeling of a compliant joint in a magnetic levitation system for an endoscopic camera”. *Mechanical Sciences*, **3**(1).
 - [52] Ramu, G., and Ananthasuresh, G., 2009. “A flexure-based deployable stereo vision mechanism and temperature and force sensors for laparoscopic tools”. In Proceedings of the 14th National Conference on Machines and Mechanisms (NaCoMM).
 - [53] Peirs, J., Van Brussel, H., Reynaerts, D., and De Gersem, G., 2002. “A flexible distal tip with two degrees of freedom for enhanced dexterity in endoscopic robot surgery”. In Proceedings of the 13th Micromechanics Europe Workshop, pp. 271–274.
 - [54] Peirs, J., Reynaerts, D., Van Brussel, H., De Gersem, G., and Tang, H.-W., 2003. “Design of an advanced tool guiding system for robotic surgery”. In 2003 IEEE International Conference on Robotics and Automation (ICRA), Vol. 2, pp. 2651–2656.
 - [55] Yoneyama, T., Watanabe, T., Kagawa, H., Hamada, J., Hayashi, Y., and Nakada, M., 2011. “Force detecting gripper and flexible micro manipulator for neurosurgery”. In 2011 Annual International Conference of the IEEE Engineering in Medicine and Biology Society, pp. 6695–6699.
 - [56] Kanada, Y., Yoneyama, T., Watanabe, T., Kagawa, H., Sugiyama, N., Tanaka, K., and Hanyu, T., 2013. “Force feedback manipulating system for neurosurgery”. *Procedia CIRP*, **5**, pp. 133–136.
 - [57] Kato, T., Okumura, I., Song, S.-E., Golby, A. J., and Hata, N., 2014. “Tendon-driven continuum robot for endoscopic surgery: Preclinical development and validation of a tension propagation model”. *IEEE/ASME Transactions on Mechatronics*, **20**(5), pp. 2252–2263.
 - [58] Kato, T., Okumura, I., Kose, H., Takagi, K., and Hata, N., 2014. “Extended kinematic mapping of tendon-driven continuum robot for neuroendoscopy”. In 2014 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS), pp. 1997–2002.
 - [59] Krieger, Y. S., Roppenecker, D. B., Kuru, I., and Lueth, T. C., 2017. “Multi-arm snake-like robot”. In 2017 IEEE International Conference on Robotics and Automation (ICRA), pp. 2490–2495.
 - [60] Hu, Y., Zhang, L., Li, W., and Yang, G.-Z., 2019. “Design and Fabrication of a 3-D Printed Metallic Flexible Joint for Snake-Like Surgical Robot”. *IEEE Robotics and Automation Letters*, **4**(2), pp. 1557–1563.
 - [61] Thomas, T. L., Venkiteswaran, V. K., Ananthasuresh, G. K., and Misra, S., 2020. “A Monolithic Compliant Continuum Manipulator: A Proof-of-Concept Study”. *Journal of Mechanisms and Robotics*, **12**(6), 05. 061006.
 - [62] Eastwood, K. W., Francis, P., Azimian, H., Swarup, A., Looi, T., Drake, J. M., and Naguib, H. E., 2018. “Design of a contact-aided compliant notched-tube joint for surgical manipulation in confined workspaces”. *Journal of Mechanisms and Robotics*, **10**(1).
 - [63] Wei, D., Wenlong, Y., Dawei, H., and Zhijiang, D., 2012. “Modeling of flexible arm with triangular notches for applications in single port access abdominal surgery”. In 2012 IEEE International Conference on Robotics and Biomimetics (ROBIO), pp. 588–593.
 - [64] Wenlong, Y., Wei, D., and Zhijiang, D., 2013. “Mechanics-based kinematic modeling of a continuum manipulator”. In 2013 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS), pp. 5052–5058.
 - [65] Du, Z., Yang, W., and Dong, W., 2015. “Kinematics modeling of a notched continuum manipulator”. *Journal of Mechanisms and Robotics*, **7**(4).
 - [66] Du, Z., Yang, W., and Dong, W., 2015. “Kinematics modeling and performance optimization of a kinematic-mechanics coupled continuum manipulator”. *Mechatronics*, **31**, pp. 196–204.
 - [67] York, P. A., Swaney, P. J., Gilbert, H. B., and Webster, R. J., 2015. “A wrist for needle-sized surgical robots”. In 2015 IEEE International Conference on Robotics and Automation (ICRA), pp. 1776–1781.
 - [68] Swaney, P. J., York, P. A., Gilbert, H. B., Burgner-Kahrs, J., and Webster, R. J., 2017. “Design, fabrication, and testing of a needle-sized wrist for surgical instruments”. *Journal of Medical Devices*, **11**(1).
 - [69] Eastwood, K. W., Azimian, H., Carrillo, B., Looi, T., Naguib, H. E., and Drake, J. M., 2016. “Kinetostatic design of asymmetric notch joints for surgical robots”. In 2016 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS), pp. 2381–2387.
 - [70] Kutzer, M. D., Segreti, S. M., Brown, C. Y., Armand, M., Taylor, R. H., and Mears, S. C., 2011. “Design of a new cable-driven manipulator with a large open lumen: Preliminary applications in the minimally-invasive removal of osteolysis”. In 2011 IEEE International Conference on Robotics and Automation (ICRA), pp. 2913–2920.
 - [71] Liu, W. P., Lucas, B. C., Guerin, K., and Plaku, E., 2011. “Sensor and sampling-based motion planning for minimally invasive robotic exploration of osteolytic lesions”. In 2011 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS), pp. 1346–1352.

- [72] Murphy, R. J., Moses, M. S., Kutzer, M. D., Chirikjian, G. S., and Armand, M., 2013. "Constrained workspace generation for snake-like manipulators with applications to minimally invasive surgery". In 2013 IEEE International Conference on Robotics and Automation (ICRA), pp. 5341–5347.
- [73] Murphy, R. J., Kutzer, M. D., Segreti, S. M., Lucas, B. C., and Armand, M., 2014. "Design and kinematic characterization of a surgical manipulator with a focus on treating osteolysis". *Robotica*, **32**(6), pp. 835–850.
- [74] Moses, M. S., Murphy, R. J., Kutzer, M. D., and Armand, M., 2015. "Modeling cable and guide channel interaction in a high-strength cable-driven continuum manipulator". *IEEE/ASME Transactions on Mechatronics*, **20**(6), pp. 2876–2889.
- [75] Alambeigi, F., Murphy, R. J., Basafa, E., Taylor, R. H., and Armand, M., 2014. "Control of the coupled motion of a 6 DoF robotic arm and a continuum manipulator for the treatment of pelvis osteolysis". In 2014 36th Annual International Conference of the IEEE Engineering in Medicine and Biology Society, pp. 6521–6525.
- [76] Wilkening, P., Alambeigi, F., Murphy, R. J., Taylor, R. H., and Armand, M., 2017. "Development and experimental evaluation of concurrent control of a robotic arm and continuum manipulator for osteolytic lesion treatment". *IEEE Robotics and Automation Letters*, **2**(3), pp. 1625–1631.
- [77] Gao, A., Murphy, R. J., Liu, H., Iordachita, I. I., and Armand, M., 2016. "Mechanical model of dexterous continuum manipulators with compliant joints and tendon/external force interactions". *IEEE/ASME Transactions on Mechatronics*, **22**(1), pp. 465–475.
- [78] Liu, H., Farvardin, A., Grupp, R., Murphy, R. J., Taylor, R. H., Iordachita, I., and Armand, M., 2015. "Shape tracking of a dexterous continuum manipulator utilizing two large deflection shape sensors". *IEEE Sensors Journal*, **15**(10), pp. 5494–5503.
- [79] Gao, A., Carey, J. P., Murphy, R. J., Iordachita, I., Taylor, R. H., and Armand, M., 2016. "Progress toward robotic surgery of the lateral skull base: Integration of a dexterous continuum manipulator and flexible ring curette". In 2016 IEEE International Conference on Robotics and Automation (ICRA), pp. 4429–4435.
- [80] Coemert, S., Gao, A., Carey, J. P., Traeger, M. F., Taylor, R. H., Lueth, T. C., and Armand, M., 2016. "Development of a snake-like dexterous manipulator for skull base surgery". In 2016 38th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC), pp. 5087–5090.
- [81] Coemert, S., Alambeigi, F., Deguet, A., Carey, J., Armand, M., Lueth, T., and Taylor, R., 2016. "Integration of a snake-like dexterous manipulator for head and neck surgery with the da vinci research kit". In Proc. Hamlyn Symp. Med. Robot., pp. 58–59.
- [82] Alambeigi, F., Wang, Y., Murphy, R. J., Iordachita, I., and Armand, M., 2016. "Toward robot-assisted hard osteolytic lesion treatment using a continuum manipulator". In 2016 38th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC), pp. 5103–5106.
- [83] Alambeigi, F., Sefati, S., Murphy, R. J., Iordachita, I., and Armand, M., 2016. "Design and characterization of a debriding tool in robot-assisted treatment of osteolysis". In 2016 IEEE International Conference on Robotics and Automation (ICRA), pp. 5664–5669.
- [84] Alambeigi, F., Wang, Y., Sefati, S., Gao, C., Murphy, R. J., Iordachita, I., Taylor, R. H., Khanuja, H., and Armand, M., 2017. "A curved-drilling approach in core decompression of the femoral head osteonecrosis using a continuum manipulator". *IEEE Robotics and Automation Letters*, **2**(3), pp. 1480–1487.
- [85] Alambeigi, F., Bakhtiarinejad, M., Azizi, A., Hegeman, R., Iordachita, I., Khanuja, H., and Armand, M., 2018. "Inroads toward robot-assisted internal fixation of bone fractures using a bendable medical screw and the curved drilling technique". In 2018 7th IEEE International Conference on Biomedical Robotics and Biomechatronics (BioRob), pp. 595–600.
- [86] Alambeigi, F., Bakhtiarinejad, M., Sefati, S., Hegeman, R., Iordachita, I., Khanuja, H., and Armand, M., 2019. "On the use of a continuum manipulator and a bendable medical screw for minimally invasive interventions in orthopedic surgery". *IEEE Transactions on Medical Robotics and Bionics*, **1**(1), pp. 14–21.
- [87] Dupont, P. E., Lock, J., Itkowitz, B., and Butler, E., 2009. "Design and control of concentric-tube robots". *IEEE Transactions on Robotics*, **26**(2), pp. 209–225.
- [88] Gilbert, H. B., Rucker, D. C., and Webster III, R. J., 2016. "Concentric tube robots: The state of the art and future directions". In *Robotics Research*. Springer, pp. 253–269.
- [89] Alfallahi, H., Renda, F., and Stefanini, C., 2020. "Concentric tube robots for minimally invasive surgery: Current applications and future opportunities". *IEEE Transactions on Medical Robotics and Bionics*, **2**(3), pp. 410–424.
- [90] Liu, J., Hall, B., Frecker, M., and Reutzel, E. W., 2013. "Compliant articulation structure using superelastic nitinol". *Smart Materials and Structures*, **22**(9), p. 094018.
- [91] Chandrasekaran, K., and Thondiyath, A., 2017. "Design of a two degree-of-freedom compliant tool tip for a handheld powered surgical tool". *Journal of Medical Devices*, **11**(1), p. 014502.
- [92] Chandrasekaran, K., Sathuluri, A., and Thondiyath, A., 2017. "MagNex—Expendable robotic surgical tooltip". In 2017 IEEE International Conference on Robotics and Automation (ICRA), pp. 4221–4226.
- [93] Arata, J., Kogiso, S., Sakaguchi, M., Nakadate, R., Oguri, S., Uemura, M., Byunghyun, C., Akahoshi, T., Ikeda, T., and Hashizume, M., 2015. "Articulated minimally invasive surgical instrument based on compliant mechanism". *International Journal of Computer Assisted Radiology and Surgery*, **10**(11), pp. 1837–1843.

- [94] Arata, J., Fujisawa, Y., Nakadate, R., Kiguchi, K., Harada, K., Mitsuishi, M., and Hashizume, M., 2019. "Compliant four degree-of-freedom manipulator with locally deformable elastic elements for minimally invasive surgery". In 2019 International Conference on Robotics and Automation (ICRA), pp. 2663–2669.
- [95] Wu, Z., Bandara, D., Kiguchi, K., and Arata, J., 2019. "Design Strategy for a Surgical Manipulator based on a Compliant Mechanism: Rigidity and Range of Motion: Finding the Optimized Balance". In 2019 IEEE International Conference on Robotics and Biomimetics (ROBIO), pp. 2220–2224.
- [96] Hanks, B. W., Frecker, M., and Moyer, M., 2016. "Design of a compliant endoscopic ultrasound-guided radiofrequency ablation probe". In ASME 2016 international design engineering technical conferences and computers and information in engineering conference.
- [97] Hanks, B., Frecker, M., and Moyer, M., 2017. "Optimization of a compliant endoscopic radiofrequency ablation electrode". In ASME 2017 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference.
- [98] Zoppi, M., Sieklicki, W., and Molino, R., 2008. "Design of a microrobotic wrist for needle laparoscopic surgery". *Journal of Mechanical Design*, **130**(10).
- [99] Sieklicki, W., Zoppi, M., and Molino, R., 2009. "Superelastic compliant mechanisms for needlescopic surgical wrists". In 2009 ASME/IFTOMM International Conference on Reconfigurable Mechanisms and Robots, pp. 392–399.
- [100] Tuijthof, G. J., Herder, J. L., van Dijk, C. N., and Pistecky, P. V., 2004. "A compliant instrument for arthroscopic joint fusion". In ASME 2004 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference, pp. 397–405.
- [101] Nai, T. Y., Tuijthof, G. J. M., and Herder, J. L., 2010. "Design of a compliant steerable arthroscopic punch". *Journal of Medical Devices*, **4**, p. 027525.
- [102] Nai, T. Y., Herder, J. L., and Tuijthof, G. J., 2011. "Steerable mechanical joint for high load transmission in minimally invasive instruments". *Journal of Medical Devices*, **5**(3), p. 034503.
- [103] Swaney, P. J., Burgner, J., Gilbert, H. B., and Webster, R. J., 2012. "A flexure-based steerable needle: high curvature with reduced tissue damage". *IEEE Transactions on Biomedical Engineering*, **60**(4), pp. 906–909.
- [104] Yim, S., and Sitti, M., 2011. "Design and analysis of a magnetically actuated and compliant capsule endoscopic robot". In 2011 IEEE International Conference on Robotics and Automation (ICRA), pp. 4810–4815.
- [105] Grames, C. L., Tanner, J. D., Jensen, B. D., Magleby, S. P., Steger, J. R., and Howell, L. L., 2015. "A meso-scale rolling-contact gripping mechanism for robotic surgery". In ASME 2015 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference.
- [106] Choi, D. Y., and Riviere, C. N., 2006. "Flexure-based manipulator for active handheld microsurgical instrument". In 2005 IEEE Engineering in Medicine and Biology 27th Annual Conference, pp. 2325–2328.
- [107] Awtar, S., Trutna, T. T., Nielsen, J. M., Abani, R., and Geiger, J., 2010. "Flexdex™: a minimally invasive surgical tool with enhanced dexterity and intuitive control". *Journal of Medical Devices*, **4**(3).
- [108] Salamon, B. A., and Midha, A., 1998. "An Introduction to Mechanical Advantage in Compliant Mechanisms". *Journal of Mechanical Design*, **120**(2), 06, pp. 311–315.
- [109] Herder, J., and Van Den Berg, F., 2000. "Statically balanced compliant mechanisms (SBCM's), an example and prospects". In Proceedings of the 26th ASME DETC Biennial, Mechanisms and Robotics Conference.
- [110] Drenth, J., and Herder, J. L., 2004. "Numerical optimization of the design of a laparoscopic grasper, statically balanced with normal springs". In ASME 2004 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference, pp. 923–933.
- [111] Powell, K. M., and Frecker, M. I., 2005. "Method for optimization of a nonlinear static balance mechanism with application to ophthalmic surgical forceps". In ASME 2005 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference, pp. 441–447.
- [112] de Lange, D. J., Langelaar, M., and Herder, J. L., 2008. "Towards the design of a statically balanced compliant laparoscopic grasper using topology optimization". In ASME 2008 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference, pp. 293–305.
- [113] Tolou, N., and Herder, J. L., 2009. "Concept and modeling of a statically balanced compliant laparoscopic grasper". In ASME 2009 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference, pp. 163–170.
- [114] Hoetmer, K., Herder, J. L., and Kim, C. J., 2009. "A building block approach for the design of statically balanced compliant mechanisms". In ASME 2009 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference, pp. 313–323.
- [115] Hoetmer, K., Woo, G., Kim, C., and Herder, J., 2010. "Negative stiffness building blocks for statically balanced compliant mechanisms: design and testing". *Journal of Mechanisms and Robotics*, **2**(4).
- [116] Lassooij, J., Tolou, N., Tortora, G., Caccavaro, S., Menciassi, A., and Herder, J., 2012. "A statically balanced and bi-stable compliant end effector combined with a laparoscopic 2DoF robotic arm". *Mechanical Sciences*, **3**(2), pp. 85–93.
- [117] Stapel, A., and Herder, J. L., 2004. "Feasibility study of a fully compliant statically balanced laparoscopic grasper". In ASME 2004 International Design En-

- Engineering Technical Conferences and Computers and Information in Engineering Conference, pp. 635–643.
- [118] Lamers, A., Sánchez, J. A. G., and Herder, J. L., 2015. “Design of a statically balanced fully compliant grasper”. *Mechanism and machine theory*, **92**, pp. 230–239.
- [119] Reddy, A. N., Maheshwari, N., Sahu, D. K., and Ananthasuresh, G., 2010. “Miniature compliant grippers with vision-based force sensing”. *IEEE Transactions on Robotics*, **26**(5), pp. 867–877.
- [120] Bhargav, S. D., Chakravarthy, S., and Ananthasuresh, G., 2012. “A compliant end-effector to passively limit the force in tele-operated tissue-cutting”. *Journal of Medical Devices*, **6**(4).
- [121] Berkelman, P. J., Whitcomb, L. L., Taylor, R. H., and Jensen, P., 2003. “A miniature microsurgical instrument tip force sensor for enhanced force feedback during robot-assisted manipulation”. *IEEE Transactions on Robotics and Automation*, **19**(5), pp. 917–921.
- [122] Seibold, U., Kubler, B., and Hirzinger, G., 2005. “Prototype of instrument for minimally invasive surgery with 6-axis force sensing capability”. In *Proceedings of the 2005 IEEE International Conference on Robotics and Automation (ICRA)*, pp. 496–501.
- [123] Tada, M., and Kanade, T., 2004. “An MR-compatible optical force sensor for human function modeling”. In *International Conference on Medical Image Computing and Computer-Assisted Intervention*, Springer, pp. 129–136.
- [124] Tokuno, T., Tada, M., and Umeda, K., 2008. “High-precision MRI-compatible force sensor with parallel plate structure”. In *2008 2nd IEEE RAS & EMBS International Conference on Biomedical Robotics and Biomechanics*, pp. 33–38.
- [125] Chapuis, D., Gassert, R., Sache, L., Burdet, E., and Bleuler, H., 2004. “Design of a simple MRI/fMRI compatible force/torque sensor”. In *2004 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS)*, Vol. 3, pp. 2593–2599.
- [126] Gassert, R., Dovat, L., Lamercy, O., Ruffieux, Y., Chapuis, D., Ganesh, G., Burdet, E., and Bleuler, H., 2006. “A 2-DOF fMRI compatible haptic interface to investigate the neural control of arm movements”. In *Proceedings 2006 IEEE International Conference on Robotics and Automation (ICRA)*, pp. 3825–3831.
- [127] Gassert, R., Chapuis, D., Bleuler, H., and Burdet, E., 2008. “Sensors for applications in magnetic resonance environments”. *IEEE/ASME Transactions On Mechatronics*, **13**(3), pp. 335–344.
- [128] Puangmali, P., Althoefer, K., and Seneviratne, L. D., 2009. “Novel design of a 3-axis optical fiber force sensor for applications in magnetic resonance environments”. In *2009 IEEE International Conference on Robotics and Automation (ICRA)*, pp. 3682–3687.
- [129] Su, H., and Fischer, G. S., 2009. “A 3-axis optical force/torque sensor for prostate needle placement in magnetic resonance imaging environments”. In *2009 IEEE International Conference on Technologies for Practical Robot Applications*, pp. 5–9.
- [130] Polygerinos, P., Seneviratne, L. D., Razavi, R., Schaeffer, T., and Althoefer, K., 2012. “Triaxial catheter-tip force sensor for mri-guided cardiac procedures”. *IEEE/ASME Transactions on mechatronics*, **18**(1), pp. 386–396.
- [131] Gao, A., Zhou, Y., Cao, L., Wang, Z., and Liu, H., 2018. “Fiber bragg grating-based triaxial force sensor with parallel flexure hinges”. *IEEE Transactions on Industrial Electronics*, **65**(10), pp. 8215–8223.
- [132] Peirs, J., Clijnen, J., Reynaerts, D., Van Brussel, H., Herijgers, P., Corteville, B., and Boone, S., 2004. “A micro optical force sensor for force feedback during minimally invasive robotic surgery”. *Sensors and Actuators A: Physical*, **115**(2-3), pp. 447–455.
- [133] Fifanski, S. K., Rivera Gutiérrez, J. L., Clogenson, M., Baur, C., Bertholds, A., Llosas, P., and Henein, S., 2016. “Flexure-based multi-degrees-of-freedom in-vivo force sensors for medical instruments”. *Proceedings of Euspen 2016*, **1**(CONF), pp. 333–334.
- [134] Tan, U.-X., Yang, B., Gullapalli, R., and Desai, J. P., 2010. “Triaxial MRI-compatible fiber-optic force sensor”. *IEEE Transactions on Robotics*, **27**(1), pp. 65–74.
- [135] Kesner, S. B., and Howe, R. D., 2011. “Design principles for rapid prototyping forces sensors using 3-D printing”. *IEEE/ASME Transactions on mechatronics*, **16**(5), pp. 866–870.
- [136] Kumar, N., Piccin, O., Meylheuc, L., Barbé, L., and Bayle, B., 2014. “Design, development and preliminary assessment of a force sensor for robotized medical applications”. In *2014 IEEE/ASME International Conference on Advanced Intelligent Mechatronics*, pp. 1368–1374.
- [137] Krishnan, G., and Ananthasuresh, G., 2008. “Evaluation and design of displacement-amplifying compliant mechanisms for sensor applications”. *Journal of Mechanical Design*, **130**(10).
- [138] Turkseven, M., and Ueda, J., 2011. “Design of an MRI compatible haptic interface”. In *2011 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS)*, pp. 2139–2144.
- [139] Turkseven, M., and Ueda, J., 2012. “Analysis of an MRI compatible force sensor for sensitivity and precision”. *IEEE Sensors Journal*, **13**(2), pp. 476–486.
- [140] Chakravarthy, S., Balakuntala, M. V., Rao, A. M., Thakur, R. K., and Ananthasuresh, G., 2018. “Development of an integrated haptic system for simulating upper gastrointestinal endoscopy”. *Mechatronics*, **56**, pp. 115–131.
- [141] Stratton, E., Howell, L., and Bowden, A., 2010. “Force-displacement model of the FlexSuRe™ spinal implant”. In *ASME 2010 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference*, pp. 37–46.
- [142] Martin, T., Gengenbach, U., Guth, H., Ruther, P., Paul, O., and Bretthauer, G., 2012. “Silicon linkage with novel compliant mechanism for piezoelectric actua-

- tion of an intraocular implant". *Sensors and Actuators A: Physical*, **188**, pp. 335–341.
- [143] Krucinski, S., Vesely, I., Dokainish, M., and Campbell, G., 1993. "Numerical simulation of leaflet flexure in bioprosthetic valves mounted on rigid and expansile stents". *Journal of Biomechanics*, **26**(8), pp. 929–943.
- [144] Chen, Y., Zion, T. T., Wang, W., Kwong, R. Y., Stevenson, W. G., and Schmidt, E. J., 2015. "Intracardiac MR imaging & MR-tracking catheter for improved MR-guided EP". *Journal of Cardiovascular Magnetic Resonance*, **17**(1), pp. 1–2.
- [145] Herrmann, H. C., Mankame, N., and Ananthasuresh, S. G., 2009. Percutaneous heart valve, Nov. 24. US Patent 7,621,948.
- [146] Guimarães, T., Oliveira, S., and Duarte, M., 2008. "Application of the topological optimization technique to the stents cells design for angioplasty". *Journal of the Brazilian Society of Mechanical Sciences and Engineering*, **30**(3), pp. 261–268.
- [147] Edmondson, B. J., Bowen, L. A., Grames, C. L., Magleby, S. P., Howell, L. L., and Bateman, T. C., 2013. "Oriceps: Origami-inspired forceps". In ASME 2013 conference on Smart Materials, Adaptive Structures and Intelligent Systems.
- [148] Gollnick, P. S., Magleby, S. P., and Howell, L. L., 2011. "An introduction to multilayer lamina emergent mechanisms". *Journal of Mechanical Design*, **133**(8).
- [149] Medgadget Inc., 10 March 2016, accessed 7 August 2020. "Origami used to miniaturize, improve surgical tools, medical implants". <https://www.medgadget.com/2016/03/origami-used-to-miniaturize-improve-surgical-tools-medical-implants.html>.
- [150] Salerno, M., Zhang, K., Menciassi, A., and Dai, J. S., 2016. "A novel 4-DOF origami grasper with an SMA-actuation system for minimally invasive surgery". *IEEE Transactions on Robotics*, **32**(3), pp. 484–498.
- [151] Kuribayashi, K., Tsuchiya, K., You, Z., Tomus, D., Umamoto, M., Ito, T., and Sasaki, M., 2006. "Self-deployable origami stent grafts as a biomedical application of Ni-rich TiNi shape memory alloy foil". *Materials Science and Engineering: A*, **419**(1-2), pp. 131–137.
- [152] Bobbert, F., Janbaz, S., van Manen, T., Li, Y., and Zadpoor, A., 2020. "Russian doll deployable meta-implants: Fusion of kirigami, origami, and multistability". *Materials & Design*, p. 108624.
- [153] Nelson, T. G., Lang, R. J., Magleby, S. P., and Howell, L. L., 2016. "Curved-folding-inspired deployable compliant rolling-contact element (D-CORE)". *Mechanism and Machine Theory*, **96**, pp. 225–238.
- [154] Taylor, A. J., Slutzky, T., Feuerman, L., Ren, H., Tokuda, J., Nilsson, K., and Tse, Z. T. H., 2019. "MR-Conditional SMA-Based Origami Joint". *IEEE/ASME Transactions on Mechatronics*, **24**(2), pp. 883–888.
- [155] Bejgerowski, W., Gerdes, J. W., Gupta, S. K., and Bruck, H. A., 2011. "Design and fabrication of miniature compliant hinges for multi-material compliant mechanisms". *The International Journal of Advanced Manufacturing Technology*, **57**(5-8), p. 437.
- [156] Tang, L., Chen, Y., and He, X., 2007. "Multi-material compliant mechanism design and haptic evaluation". *Virtual and Physical Prototyping*, **2**(3), pp. 155–160.
- [157] Gouker, R. M., Gupta, S. K., Bruck, H. A., and Holzschuh, T., 2006. "Manufacturing of multi-material compliant mechanisms using multi-material molding". *The International Journal of Advanced Manufacturing Technology*, **30**(11-12), pp. 1049–1075.
- [158] Vogtmann, D. E., Gupta, S. K., and Bergbreiter, S., 2011. "Multi-material compliant mechanisms for mobile millirobots". In 2011 IEEE International Conference on Robotics and Automation (ICRA), pp. 3169–3174.
- [159] Conlan-Smith, C., Bhattacharyya, A., and James, K. A., 2018. "Optimal design of compliant mechanisms using functionally graded materials". *Structural and Multidisciplinary Optimization*, **57**(1), pp. 197–212.
- [160] Akbari, S., Sakhaei, A. H., Kowsari, K., Yang, B., Serjouei, A., Yuanfang, Z., and Ge, Q., 2018. "Enhanced multimaterial 4d printing with active hinges". *Smart Materials and Structures*, **27**(6), p. 065027.
- [161] Rus, D., and Tolley, M. T., 2015. "Design, fabrication and control of soft robots". *Nature*, **521**(7553), pp. 467–475.
- [162] Joo, J., Kota, S., and Kikuchi, N., 2001. "Large deformation behavior of compliant mechanisms". In ASME 2001 Design Engineering Technical Conference and Computers and Information in Engineering Conference, Vol. 1, pp. 9–12.
- [163] Saxena, A., and Ananthasuresh, G., 2001. "Topology synthesis of compliant mechanisms for nonlinear force-deflection and curved path specifications". *Journal of Mechanical Design*, **123**(1), pp. 33–42.
- [164] Joo, J., and Kota, S., 2004. "Topological synthesis of compliant mechanisms using nonlinear beam elements". *Mechanics Based Design of Structures and Machines*, **32**(1), pp. 17–38.
- [165] Jung, D., and Gea, H. C., 2004. "Compliant mechanism design with non-linear materials using topology optimization". *International Journal of Mechanics and materials in Design*, **1**(2), pp. 157–171.
- [166] Hao, G., Yu, J., and Li, H., 2016. "A brief review on nonlinear modeling methods and applications of compliant mechanisms". *Frontiers of Mechanical Engineering*, **11**(2), pp. 119–128.
- [167] Kota, S., 2001. "Compliant systems using monolithic mechanisms". *Smart Materials Bulletin*, **2001**(3), pp. 7–10.
- [168] Yip, M. C., Yuen, S. G., and Howe, R. D., 2010. "A robust uniaxial force sensor for minimally invasive surgery". *IEEE Transactions on Biomedical Engineering*, **57**(5), pp. 1008–1011.
- [169] Hopkins, J. B., and Culpepper, M. L., 2010. "Synthe-

- sis of multi-degree of freedom, parallel flexure system concepts via Freedom and Constraint Topology (FACT)–Part I: Principles”. *Precision Engineering*, **34**(2), pp. 259–270.
- [170] Hopkins, J. B., and Culpepper, M. L., 2010. “Synthesis of multi-degree of freedom, parallel flexure system concepts via freedom and constraint topology (fact). part ii: Practice”. *Precision Engineering*, **34**(2), pp. 271–278.
- [171] Su, H.-J., Dorozhkin, D. V., and Vance, J. M., 2009. “A screw theory approach for the conceptual design of flexible joints for compliant mechanisms”. *Journal of Mechanisms and Robotics*, **1**(4).
- [172] Kim, C. J., Moon, Y.-M., and Kota, S., 2008. “A building block approach to the conceptual synthesis of compliant mechanisms utilizing compliance and stiffness ellipsoids”. *Journal of Mechanical Design*, **130**(2).
- [173] Krishnan, G., Kim, C., and Kota, S., 2011. “An intrinsic geometric framework for the building block synthesis of single point compliant mechanisms”. *Journal of Mechanisms and Robotics*, **3**(1).
- [174] Frecker, M. I., Ananthasuresh, G. K., Nishiwaki, S., Kikuchi, N., and Kota, S., 1997. “Topological Synthesis of Compliant Mechanisms Using Multi-Criteria Optimization”. *Journal of Mechanical Design*, **119**(2), 06, pp. 238–245.
- [175] Sigmund, O., 1997. “On the design of compliant mechanisms using topology optimization”. *Journal of Structural Mechanics*, **25**(4), pp. 493–524.
- [176] Saxena, A., and Ananthasuresh, G., 2000. “On an optimal property of compliant topologies”. *Structural and multidisciplinary optimization*, **19**(1), pp. 36–49.
- [177] Howell, L. L., and Midha, A., 1995. “Parametric Deflection Approximations for End-Loaded, Large-Deflection Beams in Compliant Mechanisms”. *Journal of Mechanical Design*, **117**(1), 03, pp. 156–165.
- [178] Su, H.-J., 2009. “A pseudorigid-body 3r model for determining large deflection of cantilever beams subject to tip loads”. *Journal of Mechanisms and Robotics*, **1**(2).
- [179] Venkiteswaran, V. K., and Su, H.-J., 2016. “Pseudo-rigid-body models for circular beams under combined tip loads”. *Mechanism and Machine Theory*, **106**, pp. 80–93.
- [180] Baichapur, G. S., Gugale, H., Maheshwari, A., Bhargav, S. D., and Ananthasuresh, G., 2014. “A Vision-based Micro-Newton Static Force Sensor using a Displacement-amplifying Compliant Mechanism (DaCM)”. *Mechanics Based Design of Structures and Machines*, **42**(2), pp. 193–210.
- [181] Khan, S., and Ananthasuresh, G., 2014. “Improving the sensitivity and bandwidth of in-plane capacitive micro-accelerometers using compliant mechanical amplifiers”. *Journal of Microelectromechanical Systems*, **23**(4), pp. 871–887.
- [182] Khan, S., and Ananthasuresh, G., 2014. “A Micromachined Wideband In-plane Single-axis Capacitive Accelerometer with a Displacement-amplifying Compliant Mechanism”. *Mechanics Based Design of Structures and Machines*, **42**(3), pp. 355–370.
- [183] Stratasys Ltd., accessed 7 August 2020. “Biocompatible”. <https://www.stratasys.com/materials/search/biocompatible>.
- [184] Liu, N., Bergeles, C., and Yang, G.-Z., 2016. “Design and analysis of a wire-driven flexible manipulator for bronchoscopic interventions”. In 2016 IEEE International Conference on Robotics and Automation (ICRA), pp. 4058–4063.
- [185] Ebert-Uphoff, I., Gosselin, C. M., Rosen, D. W., and Laliberte, T., 2005. “Rapid prototyping for robotics”. *Cutting Edge Robotics*, pp. 17–46.
- [186] Tan, K., Chua, C., Leong, K., Cheah, C., Gui, W., Tan, W., and Wiria, F., 2005. “Selective laser sintering of biocompatible polymers for applications in tissue engineering”. *Bio-medical Materials and Engineering*, **15**(1, 2), pp. 113–124.
- [187] Sun, Y., Liu, Y., and Lueth, T. C., 2019. “Fe-analysis of bio-inspired compliant mechanisms in matlab for medical applications”. In 2019 IEEE International Conference on Cyborg and Bionic Systems (CBS), pp. 54–59.
- [188] Yap, C. Y., Chua, C. K., Dong, Z. L., Liu, Z. H., Zhang, D. Q., Loh, L. E., and Sing, S. L., 2015. “Review of selective laser melting: Materials and applications”. *Applied physics reviews*, **2**(4), p. 041101.
- [189] Coemert, S., Traeger, M. F., Graf, E. C., and Lueth, T. C., 2017. “Suitability evaluation of various manufacturing technologies for the development of surgical snake-like manipulators from metals based on flexure hinges”. *Procedia CIRP*, **65**, pp. 1–6.
- [190] Suh, J.-w., Kim, K.-y., Jeong, J.-w., and Lee, J.-j., 2015. “Design considerations for a hyper-redundant pulleyless rolling joint with elastic fixtures”. *IEEE/ASME Transactions on Mechatronics*, **20**(6), pp. 2841–2852.
- [191] Haga, Y., Muyari, Y., Goto, S., Matsunaga, T., and Esashi, M., 2011. “Development of minimally invasive medical tools using laser processing on cylindrical substrates”. *Electrical Engineering in Japan*, **176**(1), pp. 65–74.
- [192] Fischer, H., Vogel, B., Pfleging, W., and Besser, H., 1999. “Flexible distal tip made of nitinol (niti) for a steerable endoscopic camera system”. *Materials Science and Engineering: A*, **273**, pp. 780–783.
- [193] Theisen, W., and Schuermann, A., 2004. “Electro discharge machining of nickel–titanium shape memory alloys”. *Materials Science and Engineering: A*, **378**(1-2), pp. 200–204.
- [194] Prakash, C., Kansal, H. K., Pabla, B., Puri, S., and Aggarwal, A., 2016. “Electric discharge machining—a potential choice for surface modification of metallic implants for orthopedic applications: A review”. *Proceedings of the Institution of Mechanical Engineers, Part B: Journal of Engineering Manufacture*, **230**(2), pp. 331–353.

- [195] Garford, J. B., Kesner, S. B., Wood, R. J., and Walsh, C. J., 2013. "Force-sensing surgical grasper enabled by pop-up book MEMS". In 2013 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS), pp. 2552–2558.
- [196] Pathak, R. K., Kumar, A. R., and Ananthasuresh, G., 2013. "Simulations and experiments in punching spring-steel devices with sub-millimeter features". *Journal of Manufacturing Processes*, **15**(1), pp. 108–114.
- [197] Scharvogel, M., and Winkelmüller, W., 2011. "Metal injection molding of titanium for medical and aerospace applications". *JOM*, **63**(2), pp. 94–96.
- [198] Goodship, V., 2004. *Practical Guide to Injection Moulding*. Smithers Rapra.
- [199] Abbott, D. J., Becke, C., Rothstein, R. I., and Peine, W. J., 2007. "Design of an endoluminal NOTES robotic system". In 2007 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS), pp. 410–416.
- [200] Breedveld, P., Sheltes, J., Blom, E. M., and Verheij, J. E., 2005. "A new, easily miniaturized steerable endoscope". *IEEE engineering in medicine and biology magazine*, **24**(6), pp. 40–47.
- [201] Johnson, P. J., Serrano, C. M. R., Castro, M., Kuenzler, R., Choset, H., Tully, S., and Duvvuri, U., 2013. "Demonstration of transoral surgery in cadaveric specimens with the medrobotics flex system". *The Laryngoscope*, **123**(5), pp. 1168–1172.
- [202] Le, H. M., Do, T. N., and Phee, S. J., 2016. "A survey on actuators-driven surgical robots". *Sensors and Actuators A: Physical*, **247**, pp. 323–354.
- [203] Morgan, N., 2004. "Medical shape memory alloy applications—the market and its products". *Materials Science and Engineering: A*, **378**(1-2), pp. 16–23.
- [204] Tarniță, D., Tarniță, D., Bîzdoacă, N., Mîndrilă, I., and Vasilescu, M., 2009. "Properties and medical applications of shape memory alloys". *Rom J Morphol Embryol*, **50**(1), pp. 15–21.
- [205] Nakamura, Y., Matsui, A., Saito, T., and Yoshimoto, K., 1995. "Shape-memory-alloy active forceps for laparoscopic surgery". In Proceedings of 1995 IEEE International Conference on Robotics and Automation (ICRA), Vol. 3, pp. 2320–2327.
- [206] Liaw, H. C., Shirinzadeh, B., and Smith, J., 2007. "Robust adaptive motion tracking control of piezoelectric actuation systems for micro/nano manipulation". In Proceedings 2007 IEEE International Conference on Robotics and Automation (ICRA), pp. 1110–1115.
- [207] Xu, Q., and Tan, K. K., 2016. *Advanced control of piezoelectric micro-/nano-positioning systems*. Springer.
- [208] Heunis, C. M., Sikorski, J., and Misra, S., 2018. "Flexible instruments for endovascular interventions: Improved magnetic steering, actuation, and image-guided surgical instruments". *IEEE Robotics Automation Magazine*, **25**, pp. 71–82.
- [209] De Greef, A., Lambert, P., and Delchambre, A., 2009. "Towards flexible medical instruments: Review of flexible fluidic actuators". *Precision Engineering*, **33**(4), pp. 311–321.
- [210] Aguirre, M., Steinórsson, Á. T., Horeman, T., and Herder, J., 2015. "Technology demonstrator for compliant statically balanced surgical graspers". *Journal of Medical Devices*, **9**(2), p. 020926.