

A working lexicon for the tele-intensive care unit: we need to define tele-intensive care unit to grow and understand it.

H Neal Reynolds

Herb Rogove

Joseph Bander

Matthew McCambridge MD

Elizabeth Cowboy

See next page for additional authors

Follow this and additional works at: <https://scholarlyworks.lvhn.org/administration-leadership-health-care-informatics>



Part of the [Medicine and Health Sciences Commons](#)

This Article is brought to you for free and open access by LVHN Scholarly Works. It has been accepted for inclusion in LVHN Scholarly Works by an authorized administrator. For more information, please contact LibraryServices@lvhn.org.

Authors

H Neal Reynolds, Herb Rogove, Joseph Bander, Matthew McCambridge MD, Elizabeth Cowboy, and Michael Niemeier

A Working Lexicon for the Tele-Intensive Care Unit: We Need to Define Tele-Intensive Care Unit to Grow and Understand It

H. Neal Reynolds, M.D.,¹ Herb Rogove, D.O.,² Joseph Bander, M.D.,³
Matt McCambridge, M.D.,⁴ Elizabeth Cowboy, M.D.,⁵
and Michael Niemeier, M.D.⁶

¹R Adams Cowley Shock Trauma Center, University of Maryland
School of Medicine, Baltimore, Maryland.

²C30 Medical Group, Ojai, California.

³St. Joseph Mercy Health System, Ann Arbor, Michigan.

⁴Lehigh Valley Health Network, Allentown, Pennsylvania.

⁵Advanced ICU Care, St. Louis, Missouri.

⁶Methodist Hospital, Clarian Health, Indianapolis, Indiana.

*“Mr. Edison, I was informed, had been up the two previous
nights discovering “a bug” in his phonograph—an expression
for solving a difficulty, and implying that some imaginary in-
sect has secreted itself inside and is causing all the trouble.”*

—*Pall Mall Gazette*, March 11, 1889

Abstract

Telemedicine in the intensive care unit (Tele-ICU) has grown exponentially since the first formalized program in 2000. Initially, there was limited product choice, and certain capabilities have been engineered into the process with the implication of necessity. New technology is evolving, and new vendors are entering the market place, which should yield a multitude of technologies from which to select. To date, there has been no organized lexicon designed to facilitate communication, comparison, or evaluation. This article is designed as a starting point to develop a lexicon applicable to all technologies for the Tele-ICU with the goal of facilitating clinical comparisons and administrative choices.

Key words: telemedicine, technology, telecommunications

Introduction

Technical language evolves sometimes by accident, sometimes in an orderly fashion. Edison’s computer “bug” was not an insect but understood to be some vague process-causing malfunction deep within the electronic device, and so “bug” became the accepted lexicon for computer science. From time to time, new definitions should be established. Time is past due to develop a lexicon for the utilization of telemedicine in the intensive care unit (ICU). There exists significant technical lexicon for telemedicine¹ and a clear understanding that technical “informational literacy” will help clinicians understand conceptual differ-

ences in the telemedicine technology.² In general, lexicon has evolved in the larger field of telemedicine to include “tele-pathology,” “tele-dermatology,” tele-radiology,” and so on.³ However, to date, we could not discover any published lexicon describing the structure and function of the telemedicine in the ICU (Tele-ICU). Therefore, in creating this lexicon document, new meanings and concepts will be proposed, in some cases, without previous references. The intent of this article is to grow a technical language that will facilitate evolution of discussion. A high-level, technical language, when uniformly used across all vendors and devices, should organize the thought process for comparison yet should be comprehensible by the relative novice. Additionally, the goal would be to develop definitions that are intuitively meaningful. As a consequence, visual image and concrete examples will be an important aspect of these definitions.

As defined by the American Telemedicine Association (ATA), “Telemedicine is the use of medical information exchanged from one site to another via electronic communications to improve patients’ health status.” The ATA suggests that telemedicine “*may assist in addressing the shortage of healthcare providers by promoting new models of practice that improve the effectiveness and efficiency of the care process,*” improve distribution of services, and partially mitigate the physician manpower shortage.⁴ Others describe telemedicine as a “branch of e-health that uses communications networks for the delivery of healthcare services and medical education from one geographic location to another.”⁵

The introduction of telemedicine into the ICU dates back 20–30 years with the early efforts of Grundy et al.^{6,7} In the United States, telemedicine in the ICU has gone through an explosive growth phase since the first, commercial system installation in Norfolk, Virginia, in 2000.^{8,9} During the decade 2000–2010, a single design has largely driven growth in the Tele-ICU arena. More recently, there has been significant growth with other vendors and technologies.

The impetus driving telemedicine in the ICU evolved from a recognized manpower maldistribution^{10,11} critical care manpower shortage,^{12–16} growing clinical support for intensivists,^{17–20} influence from the Leapfrog group,²¹ successes with a systematic approach to telemedicine in the ICU,^{8,9,22} and support for increasing telemedicine as part of healthcare reform.^{23,24} The intensivist, when employing telemedicine technologies, can have a significant and positive impact on outcomes of the critically ill patient.^{8,9,25,26} Elements of the visual examination and visual review of graphical waveforms improves accuracy of decision making.^{27,28} There are multiple reports suggesting better compliance with evidence-based medical protocols when a centralized telemedicine process is in

place.²⁹⁻³⁵ Telemedicine may have even greater impact in the rural environment.³⁶⁻³⁹ Therefore, we believe that Tele-ICU is here to stay, although evolution is inevitable.

The Tele-ICU is defined as “Networks of audio-visual communications and data systems to link hospital ICUs to intensivists and other critical care professionals at remote locations.”^{40,41} As with other focused fields of endeavor, understanding the Tele-ICU literature requires knowledge of the process. Further, understanding the comparative basis of Tele-ICU models will be vital. To date, there have been no head-to-head evaluations of the various alternative technologies, networking models, or staffing structures for the Tele-ICU. It remains unclear what informational elements are absolutely mandatory to facilitate best diagnostic and therapeutic decisions via telemedicine or tele-presence or what elements are unnecessary, superfluous, or over engineered. Once the lexicon is developed, the next step may be to perform a structured comparative evaluation of systems such as suggested by the Office for the Advancement of TeleHealth, Health and Human Resources Services Administration, and the National Center for Research Resources.⁴²

Background and Process

To achieve the goal of lexicon development, national experts in the field ICU telemedicine have been recruited to develop and review the current article. For the purposes of this article, “expert” candidates were selected as individuals who are physicians working clinically, either trained in critical care medicine or actively working in critical care medicine, with at least 5 years of experience in telemedicine, and have a national/international presence by either committee work, editorial positions, or original contributions to the related literature. As is the nature when employing clinicians, each expert may have clinical experience or bias with specific technologies and vendors. Therefore, experts with multiple perspectives were recruited to develop a balance of opinions. All potential conflict of interest is described in each author’s financial disclosure in the appendix. Specific acronyms were avoided in the development of the lexicon.

The specific lexicon developed was by group consensus so as to

- Generically define major distinguishing characteristics between existing systems.
- Create verbage describing current variable usage patterns.
- Avoid fine technical details likely to be outside the purview of administrators and clinicians.
- Create a language that is flexible enough for future evolutions of the Tele-ICU.

Due to manuscript length limitations, some definitions were not included with the hope that the current foundation will serve as a living document for future growth of the language.

In developing descriptors, it was necessary to be relatively rigid while recognizing that there may be systems with multiple or overlapping characteristics. To give life to this effort, once the descriptors have been developed, then existing vendors are described by using the new definitions.

The following descriptors have been created:

General Descriptors:

1. Intensivist or Tele-ICU Intensivist
2. Virtual Presence or Bedside presence
3. Surrogate examiner and Surrogate “Hands”
4. Tele-ICU or Tele-Critical Care
5. Tele-Expertise or Tele-Surveillance

Structural Descriptors:

1. Centralized or De-Centralized Tele-ICU
2. Open, Closed, and Hybrid Communications Architecture
3. Fixed versus Portable technology at the ICU end
 - a. Dependent portability
 - b. Semi-autonomous, independent portability
 - c. Autonomous, independent portability
4. Fixed, Portable, Mobile technology at the Tele-ICU (user) end
5. Tele ICU Clinical Information Systems (CIS)
 - a. Independent Tele-ICU only CIS or
 - b. Hospital Wide Integrated CIS.

Care Models:

1. Continuous
2. Pre-Emptive/Scheduled
3. Reactive

General Descriptors

INTENSIVIST OR TELE-ICU INTENSIVIST

Intensivist. A physician with additional, sub-specialty fellowship training in the defined field of Critical Care Medicine who can provide both cognitive skills and psychomotor skills in the care of the critically ill patient.⁴³ Intensivists may have background training in either internal medicine, anesthesia, surgery, or emergency medicine.

Tele-ICU intensivists. An Intensivist, as just described, who provides cognitive skills via virtual presence⁴⁴ and directs, requests, or delegates the performance of psychomotor skills at the bedside through a “surrogate examiner” or via “surrogate hands.”

REMOTE PRESENCE OR ONSITE PRESENCE

Onsite presence. Traditional visitation at the bedside by a care provider that includes elements described in Remote Presence plus “hands-on” contact such as palpation of the abdomen, assessment of skin temperature, auscultation of heart and breath sounds, and so on.

Remote presence. The use of audio-video technology to place the remote care provider (physician, nurse, or intermediate level care provider) virtually at the bedside for real-time, two way verbal and visual communication, review of bedside graphics, and visual patient examination. There is no direct “hands-on” contact with the patient, but there may be surrogate “hands-on” examination. Remote presence examination can include auscultation of the heart, lungs,

abdomen, vascular bruits, and so on with the use of electronic stethoscope as well as ophthalmologic funduscopic exam, otologic exam, and/or real-time ultrasound examination.

SURROGATE EXAMINER/SURROGATE HANDS

Generally, the Tele-ICU Intensivist is remote from the hospital ICU or if in proximity to the ICU, the Intensivist would generally be restricted from leaving the Tele-ICU. In the normal process of care, it may be necessary to perform certain elements of examination such as palpation for crepitus, for warmth or coolness, or fluctulance, all of which can be performed by a bedside nurse. Higher skill level may be required to perform airway intubation, central line insertion, closed thoracostomy tubes, arterial line insertion, lumbar puncture, and so on. The remote Tele-ICU intensivist will delegate those psychomotor skills to a trained, on-site individual (see “Surrogate Hands”).

Surrogate examiner. This could be the same person as the individual performing bedside procedures but more likely is the bedside nurse, resident, or fellow. Necessary points of examination are performed at the direction of and/or during observation by the Tele-ICU intensivist who can observe the examination, provide immediate feedback, or coach elements of the examination.

Surrogate “Hands”. Can be any of several individuals who will be available to perform procedures. Options for “surrogate hands” include the following:

1. Medical/Surgical Residents in a training program
2. Critical care Fellows in a fellowship program
3. In-house Hospitalists
4. In house Anesthesiologists/Anesthetists
5. Cross coverage by Emergency Department physicians
6. Respiratory Therapists with airway management skills
7. Dedicated “Line Teams” for insertion of central lines
8. Acute care nurse practitioner or physician assistant

TELE-ICU OR TELE-CRITICAL CARE

Tele-ICU. Refers to “where” the remote critical care services originate (the intensivist end). The Society of Critical Care Medicine (SCCM) and the internal medicine committee of the SCCM have adopted the phrase “Tele-ICU.” Adoption of the term, although somewhat imperfect, was partially driven by a growing acceptance and usage by The Center for Medicare and Medicaid Services. The same terminology has been adopted by the ATA Tele-ICU “Discussion Group.” The four letters “eICU” are trademarked and will not be used in this article. “Tele-ICU” implies a specific physical existence and location. Though generally accurate, there are multiple applications that do not operate from a single, discrete site. Despite some ambiguity, the term “Tele-ICU” will be used to define the discrete electronic ICU or distributed, open architecture electronic ICU programs.

Tele-critical care. Refers to where the critical care services are provided (the patient end). A more general phrase suggesting that the




Fig. 1. Centralized Tele-ICU with an established central monitoring facility. The central monitoring facility is a real and defined site, staffed variable periods per day ranging from 12 to 24 h.

delivery of telemedicine critical care services is not site specific. Critical care could be provided in the ICU, the general medical/surgical floors during a rapid response event, the emergency department, or in the Post Anesthesia Care Unit (PACU) after surgery.

Structural Descriptors

CENTRALIZED OR DE-CENTRALIZED TELE-ICU

The centralized tele-ICU. A hub-and-spoke model from which critical care services originate.³² The hub (or center) is an established, remote site with staffing to include intensivists, nurses, and clerical and technical staff. The established hub is connected to multiple medical facilities and/or multiple ICUs (Fig. 1).

The de-centralized tele-ICU. A reverse, hub-and-spoke model.⁴⁵ In this model, there is no defined established central monitoring facility. The de-centralized model typically involves computers equipped with camera, speakers, microphones located at sites of convenience such as physician offices, homes, or mobile sites utilizing laptop computers or smart phones. Therefore, the de-centralized Tele-ICU is not a specific or single site but rather a process (Fig. 2).

COMMUNICATIONS ARCHITECTURE: OPEN, CLOSED, OR HYBRID

Open architecture. A flexible, undedicated communications system⁴⁶ that supports connectivity by one or multiple care providers from one or multiple sites to one or multiple sites. Open architecture generally implies connectivity via the Internet. Open architecture networks may take any of the following forms:

- a. Single Physician, To a Single Site with Multiple Patients. One physician (from home or office) providing critical care services to one ICU and providing care for any or all the patients within that ICU.




Fig. 2. Decentralized Tele-ICU has no established central monitoring facility. This diagram indicates a single medical facility at the Hub. The open architecture De-Centralized Tele-ICU could connect with multiple facilities.

- b. Single Physician, To Multiple Sites, Multiple Patients. An individual physician could be virtually present at several different medical sites from a home or office and provide services for some or all patients at each site.
- c. Multiple Physicians, From Multiple Sites, to a Single Patient. Multiple physicians to simultaneously evaluate a single patient from different sites (Fig. 3).
- d. Multiple Physicians, From Multiple Sites, To Multiple Sites, Multiple Patients. Open architecture can support capability to access patients located at multiple sites from one or more different remote care provider sites.

Each care provider has full access to all information about the patient to include audio-video connection, access to the CIS, and




Fig. 3. Open architecture system in which multiple physicians or consultants can access the patient (via audio-video) and patient clinical information simultaneously. The “cloud” is meant to indicate the Internet.

digitized radiographic information. Functionally, each physician is within the “system.”

Closed architecture. A system that supports point-to-point communications.^{47,48} Specifically, a closed architecture Tele-ICU would be point-to-point communication to a patient from a central monitoring facility. For purity of definition, there would be no option for physicians outside the closed system to have full access to patient audio-video, clinical or radiographic information. Medical consultants would be functionally external to the Tele-ICU system and could only evaluate a patient via traditional communications such as the telephone system (Fig. 4).

The closed architecture model generally employs dedicated high-speed lines. Functionally, the closed architecture exists outside the Internet, although internal data may be transmitted over high-speed cables or fibers by using Internet Protocol (IP).

Combined hybrid open and closed architecture. By combining the capabilities of the open and closed architecture,⁴⁹ it may be possible to achieve greater connectivity, maximize inclusivity of consulting or community physicians, keeping costs within reach, and getting high-level reliability of dedicated lines. Conceptually, the hybrid architecture could look as shown next in Figure 5.

Utilizing Internet connectivity from consultant office sites, including wireless mobile technology, would enable a larger spectrum of clinicians to participate in patient care and render patient care with the benefit of ALL clinical information. Such clinicians are functionally within the Tele-ICU system. Meanwhile, the Tele-ICU can maintain the security and reliability of dedicated lines from the Tele-ICU to the patient bedside.

TECHNOLOGY AT THE ICU END: FIXED OR PORTABLE

Fixed at the ICU end. Audio-video devices are permanently attached to some site in proximity of the patient,^{8,9,44} facing the patient




Fig. 4. Closed architecture with direct physician to patient communications limited to care providers within the central monitoring facility. Generally, subspecialties are excluded from direct patient visualization and communication limited to traditional telephonic modalities and are functionally outside the Tele-ICU system.




Fig. 5. Combined open and closed architecture. A network of physicians would be able to access the patient audio-video information and other technical data from the Internet. However, the Tele-ICU, as such could be external to the Internet. The “cloud” is meant to indicate the Internet.

generally from the foot of the bed (Fig. 6). The camera may be connected to the hospital network, have an IP address, and be Web accessible. To provide the audio, cameras can be equipped with Voice over Internet Protocol technology to support the two-way audio. Commercially available packages usually provide software for the remote control of Web-enabled devices.

Portable technology at the ICU end. Systems are brought to or arrive at the bedside. Portable units, generically referred to as “Carts,” provide two way, real-time, audio and video communications (Fig. 7). Portable technology may be wireless and perform within a wireless environment or require access to a Local Area Network drop access point via a cable and jack-type connection to the network. Portable systems can move or be moved outside the ICU room to participate in family conferences, collaborative rounds, physician-to-physician conversations, or “Rapid Response” activities.



Fig. 6. Simple, fixed cameras located on the ceiling over an ICU bed (left). Sophisticated camera, speaker, and microphone with a screen to show remote Tele-ICU physician (middle). Commercially available software for the remote control of cameras via the Internet (far right).

Portability at the ICU end may be further subcategorized as follows:

1. **Dependent Portability:** A cart system that is dependent on personnel who should push the device to the bedside, activate the device, sometimes re-position to achieve optimal view, and then return the device to a storage site. In addition, the device will need to be disconnected and re-connected to power for battery charge.
2. **Semi-Autonomous Independent Portability:** Movement of the device is under the direct, real-time control of the remote clinician. Semi-autonomous devices, generally called “Robots,” operate within a wireless hospital environment and will not require physical connection to the hospital network, can be driven to any site within the ICU, and can be re-positioned by the remote operator completely independent of local personnel.
3. **Autonomous Independent Portability:** At this point, fully autonomous devices remain in development. Movement to the bedside would be based on pre-programming within the ICU and would not require ICU personnel for assistance. In addition, the remote operator would not be required to have geographic knowledge of the ICU environment. When the remote care provider is summoned to a particular bed, then the remote operator would send a command for the device to go to that particular bed.

**TECHNOLOGY AT THE TELE-ICU END:
FIXED, PORTABLE, OR MOBILE (MHEALTH)**

Fixed technology at the tele-ICU end. The “Tele-ICU end” refers to the site from which the remote care provider originates the tele-medicine service (Fig. 8). “Fixed Technology” will include equipment so bulky, complex, or heavy that would not be easily moved.^{8,9,22,45}

Portable technology at the tele-ICU end. Technology or equipment that allows the remote care provider freedom of movement while supporting two way audio-video communications and access to clinical data, real-time vital signs, or radiographic information, and so on (Fig. 9). For the purposes of this lexicon, “portable,” requiring

laptop devices has been separately defined from “mobile” as described next.^{50,51} Typical requirements include a portable “laptop” computer, necessary communications software, built-in or attached camera, speakers, microphone, with hard-wired, Wi-Fi, or Broadband Internet access. The portable care provider may access



Fig. 7. Portable devices to include (from left to right) Global Media cart, Polycom cart, Tandberg cart, Intouch Health cart, and Intouch Health “Robot.” All the devices just described are dependent on local personnel for movement except the “Semi-Autonomous” Intouch Health robot on the far right.

patients remotely from home, office, hotel lobby, roadside, and so on.

Mobile or “mHealth” technology at the “Tele-ICU” end. mHealth is the practice of healthcare delivery by ultra-mobile devices.^{52,53} mHealth includes hand-held smart-phones that can provide the mobile remote clinician with audio-visual patient access and/or graphical data access (Fig. 10).

TELE-ICU CIS: TELE-ICU ONLY CIS OR HOSPITAL WIDE INTEGRATED CIS

For purposes of description, the following will be defined as if pure. However, systems may evolve from independent to integrated with the gradual development of various interfaces.

Tele-ICU only CIS. A Tele-ICU system that uses a separate CIS in the ICU only without integration to the hospital wide



Fig. 8. Two types of “Fixed Technology at the Tele-ICU End.” On the left is the Centralized Monitoring Facility, iMDSoft Installation at Lehigh Valley Hospital Center, Lehigh Valley Pennsylvania. On the right is a de-centralized physician work station at a home Tele-ICU site.

CIS.^{8,9,55} The electronic database is created independently once the patient is in the ICU and does not move with the patient as the patient transitions back out to the general medical/Surgical floors.

Hospital wide integrated CIS. A Tele-ICU system that uses the facility’s existing CIS.²² The electronic database is created at any entry point to the hospital and moves with the patient as the patient transitions from the Floor/ED/PACU to the ICU or when moving back out to the general medical/surgical floors.

CARE MODELS: CONTINUOUS, PRE-EMPTIVE/ SCHEDULED, REACTIVE

Care models may vary widely. The following definitions are presented assuming care models are distinct. However, some programs may offer all or several different forms of care models. The intent is to develop an understanding of how the care is being delivered and to be able to describe the care model succinctly.

Continuous care model. The word “continuous,” implying absolute 24×7 monitoring, requires some consideration. Continuous may be 24×7 or limited to 8, 12, or 18 h per day depending on institutional needs and contractual agreement between medical facilities.^{8,9,22} The concept of continuous care may be further nuanced based on different staffing models. Certain arrangements include 24×7 critical care nursing monitoring with the intensivists providing support during the 8, 12, or 18 h per day. See concept diagram in Figure 11.

Within the Continuous Model, there are other nuances. The level of coverage provided may vary from patient to patient. Specifically, by contractual agreement, private attending physician staff may be given the right to completely opt out of the remote Tele-ICU services or transfer a very high level of responsibility to the Tele-ICU. By opting out, the private attending patient may only receive emergency care, as the attending is summoned to the bedside. Legal counsel may mandate that programs offer the patient the option to “opt out” of Tele-ICU services.

The pre-emptive/scheduled care model. A Tele-ICU program in which the virtual visits occur at a defined time and are not in response to a physician




Fig. 9. Portable laptop devices for two-way audio-video communications and virtual presence.

page or solicitation.^{56,57} The pre-emptive virtual visit may be routinely timed to occur several hours after the intensivist has left the ICU premises, for example, routinely at 10:00 pm. The scheduled visitation model, although only minimally different than the Pre-emptive model, includes the scheduled performance of routine daily ICU rounds more typically during the morning hours, even though it may include both morning and afternoon rounds and team/collaborative rounds.

The reactive tele-ICU care model. Includes a virtual visit for the purpose of responding to an acute physiologic issue or for the performance of a critical care consultation.⁵⁸ Reactive virtual care would be unscheduled and unpredictable.

APPLICATION OF THE LEXICON TO EXISTING TECHNOLOGY

To give life to the lexicon just described, known current and future programs and products were described with the developed terminology. Other than Skype, only U.S. products were reviewed. There was no attempt to examine international products. There was no malice or fore-thought if a U.S. product was overlooked. The lexicon was applied either from direct communication with vendors, personal experience of one or more authors with the vendors, and/or review of any written and published literature. Estimates describing market penetration were not

independently verified. Although the descriptors and lexicon may not fit with absolute precision, the effort was to give a generally fair description and, therefore, a fair, real-world demonstration of the lexicon in action.

All known U.S. vendors, providing Tele-ICU programs or CIS that have or could evolve to a Tele-ICU program, were queried. In some cases, there is no program and no stated intention to enter the market. In some cases, there is a clear intention to enter the market pending development of product and complete resolution of patent-related litigation. Generally, the products have been listed in order of decreasing market penetration and/or closeness to providing an actual “Tele-ICU” (Table 1).

Discussion

The intent of this lexicon is twofold. The first intention is to stimulate and broaden the range of technological discussions. With the recent resolution of a patent infringement suit, multiple vendors are likely to enter the Tele-ICU market, particularly considering the growing physician shortage, size of the Tele-ICU market, maldistribution of physicians, and recent emphasis on telemedicine as part of healthcare reform. As new vendors move into the market, there will be greater variability in products with more available choices to make when establishing a Tele-ICU program. Therefore, it is incumbent on the purchaser to have a higher level of sophistication in making choices rather than relying on the word of vendors. To make such choices, it is not necessary for the administrator or clinician to be knowledgeable at the micro level such as switching theory, bytes versus bits, Cable Cat, or specifics of data transmission speed. Rather, it is important to have the high-level view of systems with the advantages and limitations. In this article, we have begun to develop some of the high-level definitions and lexicon for the Tele-ICU.

Second, this lexicon aims at setting the stage for comparative analysis and to ask the question, “What system provides the best patient outcomes at an affordable cost.” Some of the current Tele-ICU systems have evolved substantial capability with the implication that these capabilities are necessary. Now, it is time to begin the high-level discussions regarding where the technology should go next.

Almost all the technologies discussed in this article are currently available and deployable. However, certain applications, such as the “autonomous independent portability,” have not yet been deployed as described. Further, higher-level hybrid models are just evolving, and final forms are yet to be defined.

Conclusion

Telemedicine has become an ever-increasing part of the medical delivery system driven by the evolving and aging demographics, increased demand for ICU care, cost containment initiatives, limited manpower supply, and mal distribution of care providers. The Tele-ICU technology



Fig. 10. Mobile devices with access to graphical information to include Airstrip Technologies (San Antonio, TX) with laboratory, radiographic, and vital signs data (three left-handed images) or Global Care Quest (right image).




Fig. 11. Various paradigms for “Continuous Care.” Top panel describes a true 24×7 continuous care and monitoring program with both monitoring nursing staff and intensivists continuously active. Least involvement shown in lowest panel with remote care and monitoring enabled when local hospital staff is least available.

should continue as an integral part of the solution for the Critical Care Manpower shortage. However, the ultimate technology and supporting scientific studies will need to continue to evolve.

We believe that the future of the Tele-ICU will include a more open architecture that permits the mobile consulting physician full access to patient information from any site. Additionally, the open architecture will provide cost savings for physician coverage, as professional fees could be reduced to an “on-call” standing charge and as needed consultation charge rather than the 24×7 hourly physician costs. Similarly, the “on-call” structure improves the feasibility to have multiple or essentially all subspecialists available. Open architecture should integrate with the hospital system clinical information to enhance access for the mobile physician. Hybrids are expected to evolve that may include combined de-centralized sites with centralized sites and systems offering different technologies at different facilities. There will likely be less dependence on costly fixed and dedicated “high speed” lines and, as wireless broad band increases in availability and speed, greater utilization of the open network provided by the Internet or extremely high-speed wireless networks.

Cost of the technology is expected to come down as more vendors enter the market and provide a wider spectrum of solutions. Cost of personnel and manpower supply will continue to be problematic unless models evolve that permit satisfactory and acceptable physician “personal life choices” while still providing quality patient care. It is conceivable that the central monitoring site could be populated only with nurses and mid-level practitioners with all the physician back up located at de-centralized sites such as home, office, or mobile devices.

Finally, introduction of the Tele-ICU technology has often been limited or delayed due to resistance from local or community

Table 1. Summary Description of Current Tele-ICU Technology

Philips VISICU: (Baltimore, MD, 866-484-7428). Covering about 7,000 ICU beds with:
Continuous monitoring care model
Centralized Tele-ICU
Closed architecture
Fixed technology at Tele-ICU end
Generally fixed technology at the ICU end
Proprietary Tele-ICU only CIS
Intouch Health: (Goleta, CA, 805-562-8686). Installed at about 400 hospitals with:
Reactive, pre-emptive, or scheduled monitoring care model
De-centralized or centralized Tele-ICU
Open architecture
Portable or fixed technology capability at the Tele-ICU end
Semi-autonomous, independent, portable technology at the ICU end
Currently, no CIS
Cerner: (Kansas City, MO, 816-221-1024). Covering about 300 beds with:
Continuous monitoring care model
Centralized Tele-ICU
Closed architecture
Fixed technology at the Tele-ICU end
Fixed and portable technology at the ICU end
Hospital-wide integrated CIS
iMDsoft: (Needham, MA, 781-449-5567). One U.S. installation covering 36 beds with:
Continuous monitoring care model
Centralized Tele-ICU
Closed architecture
Fixed technology at the Tele-ICU end
Fixed technology at the ICU end
Tele ICU only CIS at current installation but may be integrated hospital-wide depending on deployment
Vidyo: (Hackensack, NJ, 866-998-4396). ICU deployment in one 20-bed unit at this time. Product sold through reseller.
Reactive, pre-emptive, or scheduled monitoring care model
De-centralized Tele-ICU
Open architecture
Fixed technology at the Tele-ICU end
Portable technology at the ICU end
No CIS
Global Media: (Scottsdale, AZ, 800-886-3692). Covering ±500 beds with 450 devices installed. Best fit description:
Reactive, pre-emptive, or scheduled monitoring care model

continued →

Table 1. Summary Description of Current Tele-ICU Technology *continued*

De-centralized Tele-ICU
Open architecture
Fixed technology at the Tele-ICU end
Portable technology at the ICU end
No CIS
Global Care Quest: (El Segundo, CA, 888-767-8352). Covering < 100 beds with:
Reactive, pre-emptive, or scheduled monitoring care model
De-centralized Tele-ICU
Open architecture
Mobile and fixed technology at the Tele-ICU end
Portable and fixed technology at the ICU end
Tele-ICU only CIS or hospital-wide integrated CIS optional
Tandberg: (Oslo, Norway and New York). Product sold through distributor and to secondary retailer, data on deployment and utilization limited. Best fit description:
Reactive, pre-emptive, or scheduled monitoring care model
Centralized Tele-ICU
Open architecture
Fixed technology at the user end
Dependent portable technology at the ICU end
No CIS
Polycom: (Pleasanton, CA, 800-polycom, 925-924-6000). ICU deployment is unknown, similar to Tandberg. Best-fit description:
Reactive, pre-emptive, or scheduled monitoring care model
Centralized Tele-ICU
Open architecture
Fixed technology at the user end
Dependent portable technology at the ICU end
No CIS
Clinicom: (San Diego, CA, 800-350-8202). No current Tele-ICU installations. Stated intentions to enter the Tele-ICU market, primarily with the Department of Defense.
EPIC: (Verona, WI, 608-271-9000). No current Tele-ICU installations. No stated current intentions to enter the Tele-ICU market.
McKesson: (San Francisco, CA, 415-983-8300). No current Tele-ICU technology. No stated current intentions to enter the Tele-ICU market.
Eclipsys: (Atlanta, GA, 800-869-8300). No current Tele-ICU technology. No stated current intentions to enter the Tele-ICU market.
Skype: (Luxembourg). Unknown penetration into medical market, 521 million accounts and 40,000 +/- daily users, not FDA approved, not HIPAA compliant. ⁵⁹⁻⁶¹
Google Video Chat: Unknown penetration into medical market. ⁶²
Others: Extreme Mobile Technologies (West Chester, OH), LifeSize (Austin, TX) function in a similar manner as Polycom, Tandberg.

physicians. With open architecture networking, the local community of physicians can have access to the system and collaborate/participate in the remote care of their own patients. Although costs of technology, manpower shortage, and cultural adaption are currently major issues, continued experience and published peer-reviewed literature will position the Tele-ICU as an essential and integrative process.

Disclosure Statement

H. Neal Reynolds, M.D.—Unfunded Board of Advisors, In Touch Health.

Herb Rogove, D.O.—President and CEO of the private company, C₃O providing remote clinical support.

Joseph Bander, M.D.—None.

Matthew McCambridge, M.D.—None.

Elizabeth Cowboy, M.D.—Employee of private corporation, Advanced ICU.

Michael Niemeier, M.D.—None.

REFERENCES

1. Ferrer-Roca O. *Standards in telemedicine in e-health systems. Quality and reliability. Models and standards*, Chapter 17. Medical Information Science Reference. Hershey, PA: IGI Global, 2011:220-240.
2. Kisilauka M. Knowledge management prerequisites for building an information society in healthcare. *Intern J Med Inform* 2006;75:322-329
3. Bansal M. *Telemedicine in medical informatics, a primer*, Chapter 13, 239-273. Tata McGraw Hill, 20034.
4. Speedie SM, Johnston B, Cox R, Antoniotti N, Lindsey J. Telehealth and Healthcare Provider Shortages. ATA Whitepaper #1. Position Statement and Recommendations from the American Telemedicine Association. Available at www.americantelemed.org/files/public/policy/Telehealth_and_the_Provider_Shortage.pdf (last accessed September 18, 2011).
5. Sood S, Mbarika V, Jugoo S, Dookhy R, Doarn C, Prakash N, Merrell R. What is Telemedicine? A Collection of 104 Peer-Reviewed Perspectives and Theoretical Underpinnings. *Telemed E Health* 2007;13:573. Available at www.libertonline.com/doi/abs/10.1089/tmj/2006.0073 (last accessed April 25, 2011).
6. Grundy BL, Jones PK, Lovitt A. Telemedicine in critical care: Problems in design, implementation, and assessment. *Crit Care Med* 1982;10:471-475.
7. Grundy BL, Crawford P, Jones PK, et al. Telemedicine in critical care: An experiment in healthcare delivery. *JACEP* 1977;6:439-444.
8. Rosenfeld BA, Dorman T, Breslow MJ, et al. Intensive care unit telemedicine: Alternate paradigm for providing continuous intensivist care. *Crit Care Med* 2000;28:3925-3931.
9. Breslow MJ, Rosenfeld BA, Doerfler M, Burke G, Yates G, et al. Effect of a multiple-site intensive care unit telemedicine program on clinical and economic outcomes: An alternative paradigm for intensivist staffing. *Crit Care Med* 2004;32:31-38.
10. Kirch DG. *How to fix the doctor shortage*. Wall Street Journal, Opinion Section, January 4, 2010.
11. Ricketts TC. Workforce issues in rural areas: A focus on policy equity. *Am J Public Health* 2005;95:42-48.
12. Angus DC, Kelley Ma, Schmitz RJ, Whiet A, Popovich J. Current and projected workforce requirements for care of the critical ill and patients with pulmonary disease: Can we meet the requirements of an aging population? *JAMA* 2000;284:2762-2770.
13. Health Resources and Services Administration Report to Congress. The Critical Care Workforce: A Study of the Supply and Demand for Critical Care Physicians.

- Requested by: Senate Report 108-81. Available at <http://bhpr.hrsa.gov/health/workforce/reports/criticalcare/default.htm> (last accessed September 2006).
14. Kelley MA, Angus DC, Chalfin DB, et al. The critical care crisis in the United States: A report from the profession. *Crit Care Med* **2004**;32:1219–1222.
 15. Grover A. Critical care workforce: A policy perspective. *Crit Care Med* **2006**;34[Suppl.]:S7–S11.
 16. Krell K. Critical care workforce. *Crit Care Med* **2008**;36:1350–1353.
 17. Reynolds HN, Haupt MT, Carlson R. Impact of critical care physician staffing on patients with septic shock in a University Medical Care Unit. *JAMA* **1988**;260:3446–3450.
 18. Li TC, Phillips MC, Shaw L, et al. On site physician staffing in a community hospital intensive care unit: Impact on test and procedure use and on patient outcome. *JAMA* **1984**;252:2023–2027.
 19. Young M, Birkmeyer J. Potential reduction in mortality rates using an intensivist model to manage intensive care units. *Eff Clin Pract* **2000**;3:284–289.
 20. Engoren M. The effect of prompt physician visits on intensive care unit mortality and cost. *Crit Care Med* **2005**;33:727–732.
 21. Leapfrog. Four Leaps in Hospital Quality, Safety and Affordability. Available at www.leapfroggroup.org (last accessed September 18, 2011).
 22. McCambridge M, Jones K, Paxton H, et al. Association of health information technology and teleintensivist coverage with decreased mortality and ventilator use in critically ill patients. *Arch Int Med* **2010**;170:648–653.
 23. Bashshur RL, Shannon GW, et al. National telemedicine initiatives: Essential to healthcare reform (Policy White Paper for the American Telemedicine Association). *Telemed E Health* **2009**;15:1–11.
 24. New England Healthcare Institute. *Tele-ICUs: Remote management in intensive care units*. Cambridge: Massachusetts Technology Collaborative and Health Technology Center, **2007**.
 25. Kohl BA, Gutsche JT, Kim P, et al. Effect of telemedicine on mortality and length of stay in a university ICU. *Crit Care Med* **2007**;35:A22.
 26. Breslow MJ. Remote ICU Care Programs: Current status. *J Crit Care* **2007**;22:66–76.
 27. Vespa PM, et al. Intensive care unit robotic telepresence facilitates rapid physician response to unstable patients and decreased costs in a neurointensive care. *Surg Neurol* **2007**;67:331–337.
 28. Meyer BC, Raman R, Hemmen T, et al. Efficacy of site-independent telemedicine in the STROkE Doc, trial: A randomized, blinded, prospective study. *Lancet Neurol* **2008**;7:787–795.
 29. Youn BA. Utilizing robots and an ICU telemedicine program to provide intensivist support for rapid response teams. *Chest* **2006**;130:1025.
 30. Patel B, Kao L, Thomas E, et al. Improving compliance with surviving sepsis campaign guidelines via remote electronic ICU monitoring. *Crit Care Med* **2007**;35:A275.
 31. Giessel GM, Leedom B. Centralized, remote ICU intervention improves best practices compliance. *Chest* **2007**;132:444a.
 32. Youn BA. ICU process improvement: Using telemedicine to enhance compliance and documentation for the ventilator bundle. *Chest* **2006**;130:226S.
 33. Badawi O, Shemmeri E. Greater collaboration between remote intensivists and on-site clinicians improves best practice compliance. *Crit Care Med* **2006**;34:A20.
 34. Rincon T, Bourke G, Ikeda D. Centralized, remote care improves sepsis identification, bundle compliance and outcomes. *Chest* **2007**;132:557S–558S.
 35. Aaronson M, Zawada ET, Herr P. Role of a telemedicine intensive care unit program (TISP) on glycemic control (GC) in seriously ill patients in a rural health system. *Chest* **2006**;130(suppl):226S.
 36. Zawada ET, Kapsala D, Herr P, et al. Prognostic outcomes after the initiation of an electronic telemedicine intensive care unit in a rural health system. *South Dakota Med* **2006**;59:391–393.
 37. Zawada E, Aaronson ML, Herr P, Erickson EK. Relationship between levels of consultative management and outcomes in a telemedicine intensivist staffing program (TISP) in a rural health system. *Chest* **2006**;130:226S.
 38. Zawada E, Herr P, Erickson D, Hitt J. Financial benefit of a teleintensivist program to a rural health system. *Chest* **2007**;132:444.
 39. Burgiss SG. *Telehealth Technical Assistance Manual*. National Rural Health Association, **2006**. Available at <http://199.237.254.34/pubs/pdf/Telehealth.pdf> (last accessed September 18, 2011).
 40. Tele-ICU White paper from SCCM. Available at www.sccm.org/publications/Critical_Connections/Archives/February_2010/Pages/Telemedicine.aspx (last accessed April 25, 2011).
 41. NEHI Tele-ICU white paper. Available at www.masstech.org/ehealth/cmyk_tele_ICU.pdf (last accessed April 25, 2011).
 42. Puskin DS, Cohen Z, Ferguson AS, Krupinski E, Spaulding R. National Center for Research Resources Conference—future of telehealth: Essential tools and technologies for clinical research and care, implementation and evaluation of telehealth tools and technologies. *Telemed E Health* **2010**;16:96–102.
 43. Brill R, Spevitz A, Branson R, et al. Critical care delivery in the intensive care unit: Defining clinical roles and best practice models. *Crit Care Med* **2001**;29:2007–2019.
 44. Celi LA, Hassan E, Marquardt C, et al. The e-ICU: Its not just telemedicine. *Crit Care Med* **2001**;29(suppl) N:183–189.
 45. Reynolds HN. The Tele-ICU: Why, How and What. What the future may look like. In: Latifi R, ed. *Telemedicine for trauma, emergencies, and disaster management*, Chapter 21. Norwood, MA: Artech House Publishers, **2011**.
 46. Wang HJ, Raman S, ChEn-Nee C, et al. ICEBERG: An Internet core network architecture for integrated communications. *J IEEE* **2000**;7:10–19.
 47. Lambrecht C. Emergency physicians role in a clinical telemedicine network. *Ann Emerg Med* **1997**;30:670–674.
 48. Lowitt MH, Kessler II, Kauffman L, et al. Teledermatology and in-person examinations. A comparison of patient and physician perspectives and diagnostic agreement. *Arch Derm* **1998**;134:470–476.
 49. Gbaguidi C, Hubaux J, Hamdi M, et al. A Programmable architecture for provision of hybrid services. *Commun Mag IEEE* **1999**;37:110–116.
 50. Kontaxakis G, Walter S, Sakas G. EU-TeleInVivo: an integrated portable telemedicine workstation featuring acquisition, processing and transmission over low-bandwidth lines of 3D ultrasound volume images. *Information Technology Applications in Biomedicine, 2000 proceedings 2000 IEEE EMBS International Conference*. **2000**;158–163. Available at http://ieeexplore.ieee.org/xpl/freeabs_all.jsp?arnumber=892378. (last accessed April 25, 2011).
 51. Rosser JC, Probst RL, Rodas EB, et al. Evaluation of the effectiveness of portable low-bandwidth telemedical applications for postoperative followup: Initial results. *J Am Coll Surg* **2000**;191:196–203.
 52. Ackerman M, Craft R, Ferrante F, et al. Telemedicine technology. *Telemed J E Health* **2002**;8:71–78.
 53. Figueredo MVM, Dias JS. Mobile telemedicine system for home care and patient monitoring. *Engineering in Medicine and Biology Society, IEMBS 26th Annual International Conference of the IEEE, San Francisco, CA*. **2004**;3387–3390.
 54. Istepanian RH, Woodward B, Gorilas E, Balos PA. Design of mobile telemedicine systems using GSM and IS-54 cellular telephone standards. *J Telemed Telecare* **1998**;4:80–82.
 55. Ries M. Tele-ICU: A new paradigm in critical care. *Inter Anesthesiol Clin* **2009**;47:153–170.
 56. Reynolds HN. Underserved does not necessarily mean rural: Remote telepresence in downtown Baltimore. *Telemed E Health* **2008**;14(Suppl 1):103.
 57. Reynolds EM, Reynolds HN. The non-adopter phenomenon. A case where being "Senior" may enhance technological adoption. *Telemed E Health* **2010**;16(Suppl. 1):S–81.

58. Petlin J, Nelson ME, Goodman J. Deployment and early experience with remote-presence patient care in a community hospital. *Surg Endosc* 2007;21:53-56.
59. Skype Corporate description. Available at <http://about.skype.com>. (last accessed April 27, 2011).
60. Skype Medical Consultation. Available at <http://directory.skype.com/en/skypeprime/listing/baladina.kavoa/Medical-Consultation> (last accessed April 27, 2011).
61. Skype definition. Available at http://wikipedia.org/wiki/skype_limited (last accessed April 27, 2011).
62. Google video Chat. Available at www.google.com/chat/video (last accessed April 27, 2011).

Address correspondence to:

H. Neal Reynolds, M.D.

*R Adams Cowley Shock Trauma Center
University of Maryland School of Medicine*

22 S. Greene St.

Baltimore, MD 21201

E-mail: hneal.reynolds@gmail.com; lflowers@umm.edu

Received: March 2, 2011

Revised: May 5, 2011

Accepted: May 14, 2011