

On the Relationship between Dynamic Contrast-Enhanced Ultrasound Parameters and the Underlying Vascular Architecture Extracted from Acoustic Angiography

*Original*

On the Relationship between Dynamic Contrast-Enhanced Ultrasound Parameters and the Underlying Vascular Architecture Extracted from Acoustic Angiography / Panfilova, A.; Shelton, S. E.; Caresio, C.; van Sloun, R. J. G.; Molinari, F.; Wijkstra, H.; Dayton, P. A.; Mischi, M.. - In: ULTRASOUND IN MEDICINE AND BIOLOGY. - ISSN 0301-5629. - ELETTRONICO. - 45:2(2019), pp. 539-548. [10.1016/j.ultrasmedbio.2018.08.018]

*Availability:*

This version is available at: 11583/2954677 since: 2022-02-04T11:36:57Z

*Publisher:*

Elsevier USA

*Published*

DOI:10.1016/j.ultrasmedbio.2018.08.018

*Terms of use:*

openAccess

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

*Publisher copyright*

Elsevier postprint/Author's Accepted Manuscript

© 2019. This manuscript version is made available under the CC-BY-NC-ND 4.0 license  
<http://creativecommons.org/licenses/by-nc-nd/4.0/>. The final authenticated version is available online at:  
<http://dx.doi.org/10.1016/j.ultrasmedbio.2018.08.018>

(Article begins on next page)

# On the relationship between dynamic contrast - enhanced ultrasound parameters and the underlying vascular architecture extracted from acoustic angiography

Anastasiia Panfilova<sup>a,\*</sup>, Sarah E. Shelton<sup>b</sup>, Cristina Caresio<sup>c</sup>, Ruud JG van Sloun<sup>a</sup>, Filippo Molinari<sup>c</sup>, Hessel Wijkstra<sup>a,d</sup>, Paul A. Dayton<sup>b</sup>, Massimo Mischi<sup>a</sup>

<sup>a</sup>*Department of Electrical Engineering, Technical University of Eindhoven, Eindhoven, P.O.Box 513, 5600 MB Eindhoven, SPS Group, The Netherlands*

<sup>b</sup>*Joint Department of Biomedical Engineering, University of North Carolina at Chapel Hill and North Carolina State University, 152 MacNider Hall, Campus Box 7575, Chapel Hill, NC 27599, United States of America*

<sup>c</sup>*Politecnico di Torino, Torino Italy, Corso Duca degli Abruzzi 24, 10129 Torino, Italy*

<sup>d</sup>*AMC University Hospital, Urology Dept, Amsterdam, Postbus 22660, 1100 DD Amsterdam, the Netherlands*

---

## Abstract

Dynamic contrast-enhanced ultrasound (DCE-US) has been proposed as a powerful tool for cancer diagnosis by estimation of perfusion and dispersion parameters reflecting angiogenic vascular changes. This work aims at identifying which vascular features are mainly reflected by the estimated perfusion and dispersion parameters through comparison with Acoustic Angiography (AA). AA is a high resolution technique that allows quantification of vascular morphology. 3D AA and 2D DCE-US bolus acquisitions monitored growth of fibrosarcoma tumors in 9 rats. AA-derived vascular properties were analyzed

---

\*Corresponding Author: Anastasiia Panfilova, Department of Electrical Engineering, Technical University of Eindhoven, De Groene Loper 19, Eindhoven, The Netherlands; Email, A.P.Panfilova@tue.nl; Phone, +31616070943

along with DCE-US perfusion and dispersion in order to investigate the differences between tumor and control, and their evolution in time. AA-derived microvascular density and DCE-US perfusion showed good agreement, confirmed by their spatial distributions. No vascular feature was correlated with dispersion. Yet, dispersion provided better cancer classification than perfusion. We therefore hypothesize that dispersion characterizes vessels that are smaller than those visible with AA.

*Keywords:* Acoustic angiography, Dynamic contrast-enhanced ultrasound, Cancer, Dispersion, Perfusion, Ultrasound contrast agents

---

## 1 **Introduction**

2 Malignant tissue shows a set of alterations from benign tissue that can be  
3 used as markers to detect it (Koumoutsakos et al., 2013). Of particular in-  
4 terest for cancer imaging are the altered vascular architecture and the conse-  
5 quent changes in blood supply. Angiogenic vessels grow to nourish the tumor  
6 and support its proliferation. These vessels have been found to be tortuous,  
7 to grow chaotically, without the typical vessel hierarchy, and with a high oc-  
8 currence of arteriovenous shunts. Many of these properties can be recognized  
9 with contrast-enhanced ultrasound techniques, which have shown promising  
10 results for distinguishing malignant tissue from benign (Brock et al., 2013;  
11 Gessner et al., 2013; Kuenen et al., 2013b, 2011; Mischi et al., 2012; Quaia,  
12 2011; Shelton et al., 2015).

13 Dynamic contrast-enhanced ultrasound (DCE-US) captures the contrast-  
14 agent passage through the vascular bed after its injection in the patient’s  
15 bloodstream. Specifically, it registers the local evolution of gray-level inten-  
16 sity at each pixel, referred to as the time intensity curve (TIC), which reflects  
17 the varying ultrasound contrast agent (UCA) concentration. The recorded  
18 intensities are then converted into UCA concentration with a linearization  
19 function specific to the employed ultrasound scanner (Rognin et al., 2008),  
20 yielding an indicator dilution curve (IDC) for every pixel in the video. Vari-  
21 ous characteristics of IDCs have been proven to be useful for distinguishing  
22 malignant from benign tissue (Mischi et al., 2012).

23 Several approaches have been adopted to extract information from IDCs  
24 derived from DCE-US bolus acquisitions. Some heuristic features of the  
25 IDCs, such as the wash-in time and the peak intensity, are related to cancer

26 (Mischi et al., 2012; Zhao et al., 2010). Multiple other techniques employ  
27 IDC fitting by analytical models, such as the lognormal, gamma, and local  
28 density random walk (LDRW) model (Strouthos et al., 2010). Functional  
29 parameters of the curves (e.g. area-under-the-curve) are extracted and dis-  
30 played in colormaps, aiming to obtain a clearly distinguishable malignant  
31 region. All these approaches mainly attempt to quantify perfusion, which  
32 is motivated by the presence of ample arteries feeding the tumor, increased  
33 microvascular density (MVD), and presence of arteriovenous shunts. Despite  
34 this, clinical evidence has shown that cancerous lesions in the prostate can  
35 also be iso- or hypo-perfused (Brock et al., 2013). Indeed, it is known that  
36 tumor tissue has higher resistance to blood flow (Narang and Varia, 2011).  
37 This induces a counterbalancing factor that complicates predictions about  
38 the level of blood supply within the tumor, as compared to surrounding tis-  
39 sue (Cosgrove and Lassau, 2010). Furthermore, the MVD inside the tumor  
40 can be strongly heterogeneous, creating highly perfused regions as well as  
41 hypoxic, avascular regions. Therefore, assessment of perfusion alone is in-  
42 sufficient for reliable cancer diagnostics. These findings have motivated the  
43 development of contrast ultrasound dispersion imaging (CUDI), a method  
44 which enables assessment of UCA dispersion, in addition to quantification of  
45 perfusion (Kuenen et al., 2011; Mischi et al., 2012).

46 CUDI aims at quantifying the UCA dispersion due to the architecture  
47 of the vascular tree and complex multipath trajectories available for UCA  
48 transport. The main hypothesis that lies in the foundation of the method  
49 states that dispersion reflects structural vascular changes induced by angio-  
50 genesis. The first CUDI approach involved modelling of the IDCs in time

51 domain with a LDRW model and extraction of a dispersion-related parame-  
52 ter from the fitted model (Kuenen et al., 2011). An important complication  
53 associated with this approach was poor signal to noise ratio, hindering the  
54 fitting procedure and decreasing its reliability. This problem has been mit-  
55 igated by spatiotemporal similarity analysis (Kuenen et al., 2013a; Mischi  
56 et al., 2012). In a promising implementation, this approach involves calcu-  
57 lation of an average correlation coefficient measuring the similarity of a TIC  
58 at a pixel and its surrounding pixels (Kuenen et al., 2013b). A theoretical  
59 description of the problem within the framework of the LDRW model has  
60 shown that the correlation coefficient between IDCs is monotonically related  
61 to the dispersion coefficient (Kuenen et al., 2013a). Moreover, this approach  
62 has demonstrated its superior performance compared to perfusion-related  
63 parameters at localizing prostate cancer in a clinical setting (Kuenen et al.,  
64 2013b). This method has been validated against cell differentiation reflected  
65 with the Gleason score for prostate cancer (Schalk, 2017). Another study  
66 identified that regions of low dispersion correlated with those of high MVD,  
67 quantified by immunohistology (Saidov et al., 2016). However, in this study  
68 detailed characterization of the vascular architecture (e.g. tortuosity and  
69 vessel size) was not available.

70 Acoustic angiography (AA) can provide accurate characterization of the  
71 vascular architecture: it is a high-resolution technique, capable of imaging  
72 individual microvessels (Gessner et al., 2013; Shelton et al., 2015). AA per-  
73 mits imaging vessels at a high resolution of 100-200  $\mu\text{m}$  at 2 cm depth with  
74 minimal signal from tissue. While transmitting ultrasonic waves at frequen-  
75 cies in the order of a few MHz, close to the UCA bubble resonance frequency,

76 it records the nonlinear response of the contrast agents in a high frequency  
77 range centered at 30 MHz. This technique grants the possibility to quantify  
78 vessel density and morphology measures such as the sum of angles metric  
79 (SOAM) and distance metric (DM) (Rao et al., 2016; Shelton et al., 2015).  
80 These parameters have been reported to be significantly different for malig-  
81 nant and benign tissue (Gessner et al., 2013; Shelton et al., 2015). Thereby,  
82 AA gives the opportunity to validate whether these features are reflected in  
83 DCE-US due to the different character of UCA perfusion and dispersion in  
84 these vessels.

85 The aim of this work is to determine whether DCE-US is able to char-  
86 acterize the underlying vascular architecture. It involves DCE-US and AA  
87 imaging of fibrosarcoma tumors and control regions in a longitudinal study of  
88 9 rats. AA and DCE-US acquisitions were performed every 3 days, at 4 time  
89 points, starting with the day when the tumors could be palpated. An overall  
90 comparison of the tumor’s and control’s vascular properties was performed.  
91 Additionally, a longitudinal study of these properties was conducted, aiming  
92 to find similar trends in features extracted from the two different techniques  
93 of DCE-US and AA.

## 94 **Materials and Methods**

### 95 *Rat Models*

96 Fibrosarcoma tumor implantation was performed in rats according to a  
97 previously applied protocol (Streeter et al., 2011). The tumor models were  
98 established from propagated tumor tissue provided by the Dewhirst Lab at  
99 Duke University. Before surgery the (Fischer 344) rats were anesthetized

100 with isoflurane; their left flank was then shaved and disinfected. An incision  
101 ( $\sim 2$  mm) was made above the quadriceps muscle, and a sample of tumor  
102 tissue ( $\sim 1$   $mm^3$ ) was positioned under the skin. The incision was closed  
103 with 1-2 staples. This procedure was performed at 3 different time points  
104 with 9 rats in total. Rats belonging to the same series were operated on the  
105 same day.

106 On day 8 after implantation, the first ultrasound acquisition was per-  
107 formed if the tumors were palpable. Otherwise, we waited for 2-3 days for  
108 subsequent assessment. When the tumors were palpable, UCA was injected  
109 in the rats' tail vein through a 24 gauge catheter while the animals were anes-  
110 thetized with vaporized isoflurane in oxygen. DCE-US was performed on the  
111 tumor-bearing flank for assessment of perfusion and dispersion. The AA ac-  
112 quisition protocol immediately followed the DCE-US acquisition to minimize  
113 the amount of time each animal spent under anesthesia. The beginning of  
114 the DCE-US and AA acquisitions were different between the series, start-  
115 ing with day 8, day 11, and day 13, respectively. For all but one animal,  
116 subsequent imaging acquisitions were performed with an interval of 3 days,  
117 amounting to 4 time points in total. One rat was an exception since we  
118 were not able to inject the contrast (for both modalities) in its tail vein, and  
119 managed to image only at the first and third time points. All experiments  
120 were performed at the University of North Carolina at Chapel Hill, approved  
121 by the Institutional Animal Care and Use Committee at the University of  
122 North Carolina at Chapel Hill.



123 *Image acquisition*

124 *DCE-US bolus injection protocol*

125 A UCA bolus of  $2 \times 10^8$  microbubbles was injected in the rats' tail vein.  
126 The contrast agent used in this study was made in-house; it has a lipid  
127 shell and perfluorocarbon core, similar to Definity<sup>®</sup> (Latheus Medical Imag-  
128 ing/U.S.A, N. Billerica). A 15L8-S probe was utilized with a Siemens Se-  
129 quoia scanner in Cadence Pulse Sequencing mode at an insonifying central  
130 frequency of 7 MHz. The acquired DCE-US recordings were stored in DI-  
131 COM format.

132 *AA continuous infusion protocol*

133 A continuous infusion of microbubbles was administered using a syringe  
134 pump (PHD 2000, Harvard Apparatus) at a rate of  $1.5 \times 10^8$  microbubbles  
135 per minute. AA imaging was performed with a dual-frequency single-element  
136 transducer transmitting at 4 MHz, and receiving around 30 MHz. The 3D  
137 AA images were acquired plane by plane, with a step size of 100  $\mu\text{m}$ .

138 *DCE-US bolus data processing*

139 *Preprocessing*

140 All the bolus recordings were filtered with a Gaussian filter, as previously  
141 performed in (Mischi et al., 2012), using a kernel of 0.13 mm equal to 1.6  
142 pixels. This value improved the signal-to-noise-ratio at the cost of additional  
143 spatial correlation between TICs at neighbouring pixels. The TIC power  
144 of every pixel was evaluated as the root mean square of the TIC after the  
145 baseline was removed. Regions with a level of TIC power below -22 dBs of  
146 the maximum TIC power over all images were excluded from further analysis

147 (shown in black in Fig. 1 a.). This limited the effect of random noise on the  
148 parameters of interest (Kuenen et al., 2014). Characteristic of DCE-US is  
149 multiplicative noise: noise proportional in its power to the signal amplitude.  
150 By eliminating regions with low TIC power, we avoided erroneous parameter  
151 estimation from regions with low signal power where random noise dominates.  
152 After this, the intensity values of the remaining regions were linearized by  
153 inverting the logarithmic compression function implemented in the adopted  
154 scanner, yielding the IDCs.

#### 155 *Assessment of dispersion*

156 An average correlation coefficient was calculated for every pixel between  
157 its own IDC and those at its surrounding pixels within a ring-shaped kernel  
158 (Mischi et al., 2012) with an inner radius of 0.6 mm and an outer radius  
159 of 2 mm. The inner radius was chosen equal to the lateral resolution of  
160 the preprocessed bolus recordings at  $\sim 2$  cm depth as identified with local  
161 autocorrelation analysis. Details about the latter procedure can be found in  
162 (Mischi et al., 2012). The lateral resolution was taken as a reference since  
163 it was worse than the axial resolution. The outer radius of the kernel was  
164 set equal to the size of 2 mm, which a tumor can usually reach without  
165 neovascularization (Folkman, 1971). The time window over which the IDCs  
166 were correlated to each other was selected to maximize the area under the  
167 receiver operating characteristic curve for tumor classification, resulting in a  
168 value of 17 seconds as proposed in previous work (Panfilova et al., 2016). This  
169 is the only informative segment of the IDC (Fig. 2) due to early recirculation,  
170 as often observed in small animals (Stapleton et al., 2009). In this work, the  
171 beginning of the analyzed time window was set with 3 seconds before the

172 appearance time, ensuring the wash-in phase to be entirely captured.

### 173 *Assessment of perfusion*

174 Wash-in-rate was adopted to assess perfusion and computed as the slope  
175 of a line fitted to the IDC in the 2-second interval after appearance time, as  
176 illustrated in Fig. 2. The value of 2 seconds was chosen to reflect the rise of  
177 UCA concentration in the initial part of the IDCs in all acquired clips.

### 178 *AA data processing*

179 The AA volumes were interpolated to reduce the inter-plane distance to  
180 50 microns and make the pixels isotropic. Visible vessels were manually seg-  
181 mented and characterized in terms of vessel dimensions: vessel length (VL)  
182 and mean radius (MR). VL was computed as the length of the vessel segment  
183 identified between successive branching points, and MR was computed as the  
184 mean radius of this vessel segment along its length. Vessel tortuosity was as-  
185 sessed with the distance metric and the sum of angles metric (Bullitt et al.,  
186 2003). The DM was computed as the ratio of vessel length to the Euclidian  
187 distance between its beginning and end. The SOAM was calculated as the  
188 sum of angles between successive points on the vessel centerline divided by  
189 VL, using the same formula as in (Bullitt et al., 2003), but excluding the  
190 torsional angle. Besides these individual vessel properties, MVD was calcu-  
191 lated as a global characteristic of the tumor at a given timepoint, defined as  
192 the number of visible vessel segments divided by tumor volume. The volume  
193 vascular density (VVD) was computed with a moving 3D isotropic kernel in  
194 the central slice of the tumor ( $\sim 1$  mm in thickness). Otsu’s method (Vala  
195 and Baxi, 2013) was used to select a threshold to separate noise from vessel

196 signal within the central slices; the percentage of pixels with vessel signal  
197 from the overall number of pixels in the 3D kernel was calculated.

### 198 *Statistical analysis*

199 The DCE-US parameters were spatially downsampled by a factor 7 in  
200 both directions, equal to the resolution of the preprocessed images. This  
201 was performed to exclude spatial correlation and prepare the data for the  
202 statistical tests that require sample independence.

### 203 *Comparison between tumor and control*

204 Dispersion and perfusion values were divided into two groups. The tumor  
205 group was composed of the manually selected tumor regions (inside the red  
206 contour, Fig. 1 a.) from all rats at all time points binned together. The  
207 control group was taken from pixels outside the tumor contour, dilated by  
208  $\sim 1$  mm (in blue, Fig. 1 a.). The region between the red and blue contour  
209 was excluded from analysis to avoid erroneous pixel assignment to tumor  
210 or control, since DCE-US information was not considered sufficiently com-  
211 prehensive for such accurate tumor delineation, as required by e.g. ablation  
212 therapy and surgery. The AA parameters were extracted in a similar fashion:  
213 vessels were taken from within the tumor region and outside it in the same  
214 flank (Fig. 3). Vessel segments on the border of the selected contour, whose  
215 belonging to a tumor or control group was debatable were disregarded from  
216 analysis.

217 An Anderson-Darling goodness of fit hypothesis test was performed on  
218 all the parameter distributions to check for data normality. Since all the dis-  
219 tributions were identified as non-Gaussian, a Mann Whitney non-parametric

220 test was performed to establish the significance (p-value) of the difference  
221 between tumor and control. No additional subsampling or upsampling was  
222 performed to make the control and tumor data sets balanced, since the Mann  
223 Whitney test can be applied to data sets with distributions of different size  
224 (Mann and Whitney, 1947).

225 The Cohen’s d was used as a measure of the ‘effect size’ (Sullivan and  
226 Feinn, 2012) that the tumor has on the underlying vasculature, calculated as  
227 the difference between the means of two distributions divided by the standard  
228 deviation of the control. The values of the Cohen’s d term allow to classify  
229 the difference between two distributions according to 4 categories: small,  
230 medium, large, and very large for values of 0.2, 0.5, 0.8, 1.3, respectively.

### 231 *Longitudinal study of tumor and control*

232 A longitudinal study of the tumor evolution was performed with the  
233 Kruskal-Wallis post hoc test, evaluating the differences among the distribu-  
234 tions of dispersion and perfusion, and vascular features of tumor and control  
235 at 4 time points. The Kruskal-Wallis test (Kruskall and Wallis, 1952) does  
236 not require equal sample sizes, which is an advantage considering that our  
237 data set is unbalanced and incomplete: data is missing for one tumor at two  
238 time points as well as control at several time points for the large tumors.  
239 Moreover, the number of visible vessels is different for every image acqui-  
240 sition. For all rats, all parameter values were binned together according to the  
241 time point of the acquisition.

242 The statistic test calculation is influenced by the number of observations  
243 and can result in different outcomes for different sample sizes (Kruskall and  
244 Wallis, 1952). Since the number of pixels provided more samples for disper-

245 sion and perfusion compared to the number of vessels extracted with AA,  
246 these pixels were randomly subsampled to yield the same number of samples  
247 as vessels per each representative dataset of tumor and control at each time  
248 point. The only parameter that remained different in terms of group size is  
249 the MVD; being a global parameter that characterizes the entire tumor and  
250 control at a specific time point.

251 After the post-hoc tests were performed, the Pearson correlation coeffi-  
252 cient was computed between the medians of the parameters showing similar  
253 longitudinal trends.

#### 254 *Mapping of vascular properties on the bolus acquisition plane*

255 During the DCE-US bolus acquisitions the operator always tried to im-  
256 age the largest cross-section of the tumor, and keep the same orientation of  
257 the probe as used for AA. However, it was noticed that these precautions  
258 were not sufficient to reliably identify the DCE-US plane within AA: even a  
259 movement of the order of  $\sim 1$  mm alters the imaged vascular pattern of a tu-  
260 mor. It was noticed that the perfusion maps highlight larger vessels, clearly  
261 visible in the AA (Fig. 2 b. and d.). These vessels were used as markers  
262 to locate the bolus recording plane in the AA volume. For this, a dedicated  
263 tool was developed, allowing to freely scroll through the AA volume planes  
264 and change their orientation.

265 The selection of the plane was performed by visual inspection, choosing an  
266 image containing as many as possible vessel markers present in the perfusion  
267 maps. A slice in the AA volume of  $\sim 1$  mm thickness was selected and  
268 an extension of the skeletonization algorithm described in (Meiburger et al.,  
269 2016) was applied to extract MVD (Fig. 2 e.), MR (Fig. 2 f.), VL, and

270 SOAM. This slice thickness was chosen to be of the order of the elevational  
271 resolution in the bolus recordings and sufficiently large to register vessel  
272 segments. This allowed a qualitative comparison of the spatial distribution  
273 of the vascular features with those of dispersion and perfusion in the same  
274 plane.

275 All the image processing and statistical analysis was performed with Mat-  
276 lab software (the MathWorks, Natick, MA).

## 277 **Results**

### 278 *Statistical analysis*

#### 279 *Comparison between tumor and control*

280 For all the extracted parameters, tumor and control have significantly dif-  
281 ferent distributions ( $p < 0.001$ ). However, the magnitude of the differences,  
282 expressed in Cohen's  $d$ , spans a wide range (Fig. 4), showing a marginal  
283 effect size for the DM (Fig. 4 c.), and small to very large differences for the  
284 rest of the parameters.

#### 285 *Longitudinal study of tumor and control*

286 Since the DM showed almost no difference between tumor and control,  
287 it was excluded from the longitudinal analysis. Boxplots with all parame-  
288 ter values binned according to the time points are shown in Fig. 5, while  
289 Fig. 6 illustrates the results of the post hoc Kruskal-Wallis test, color-coded  
290 according to the significance level of the intra-distribution differences.

291 The dispersion median is relatively constant in time for both tumor and  
292 control, showing a significant difference for control and tumor distributions

293 (Fig. 5 a., Fig. 6 a.). Tumor perfusion is significantly different from the  
294 control at all time points (Fig. 5 b., Fig. 6 b.), peaking for the tumors at  
295 the second time point. Interestingly, the longitudinal trend of the control's  
296 perfusion seems to mimic the tumor's trend in time, however, at a smaller  
297 magnitude, not identified as significant with the post hoc test.

298 The VVD is stably higher for tumor, while the MVD seems to follow a  
299 similar trend to that of perfusion, peaking for tumors at the second time  
300 point. However, the result of the MVD post hoc test is difficult to compare  
301 to others since the number of samples is different: only one value of MVD  
302 per time point is available, while the other parameters were subsampled  
303 according to the number of segmented vessels in the AA volume at a given  
304 time point.

305 The post hoc results, illustrated by colormaps in Fig. 6, are comparable  
306 for dispersion, the VVD, the VL, and the SOAM. However, no significant  
307 correlation between the medians of the dispersion levels and the mentioned  
308 AA parameters has been identified. As for perfusion, the mean perfusion in  
309 tumors and their MVD showed a significant correlation coefficient of 0.572  
310 ( $p < 0.001$ ) and inclusion of both control and tumor values resulted in a  
311 Pearson correlation coefficient of 0.67 ( $p < 0.001$ ).

### 312 *Mapping of vascular properties on the bolus acquisition plane*

313 The spatial parametric maps of the AA skeleton confirmed our observa-  
314 tion that there is a correlation between regions of high perfusion and elevated  
315 MVD (Fig. 1 b. and f.). No spatial correspondance was found between dis-  
316 persion and the other AA - derived parameters.



317 **Discussion**

318 Dispersion shows a large difference (Cohen’s  $d = 1.68$ ) between tumor  
319 and control, exhibiting stable performance at tumor detection as it develops.  
320 Perfusion shows a lower discrimination power than dispersion, that is high for  
321 younger tumors, peaking at time point 2, and decreases with tumor growth.  
322 Interestingly, the perfusion level in the control around the tumor is also  
323 elevated (Fig. 5 b.), showing a similar trend as in the tumor itself. This  
324 may reflect that the overall perfusion of tissue around the tumor is increased  
325 and influenced by the tumor. This effect has been shown for the SOAM,  
326 which exhibits intermediate values between that of tumor and control in  
327 tissue adjacent to the tumor (Rao et al., 2016). Moreover, it has been shown  
328 for the fibrosarcoma model that the vascular source is often located in the  
329 periphery of the tumor (Ponce et al., 2007; Tozer et al., 1990; Viglianti et al.,  
330 2004).

331 Dispersion of the control stays stable over time, indicating that dispersion-  
332 related changes mainly occur within the tumor itself, and not in the sur-  
333 roundings. The spatial perfusion and dispersion maps are complementary,  
334 showing different patterns of highlighted regions (Fig. 1 b. and c.). Perfusion  
335 highlights large vessels, as well as regions with high MVD.

336 The SOAM indicates that the tumor has more tortuous vessels, exhibiting  
337 a similar trend to that of dispersion (Fig. 5 a., g.) and comparable results  
338 for the post-hoc test (Fig. 6 a., g.). Nevertheless, the effect size difference, as  
339 indicated by Cohen’s  $d$ , is much lower for the SOAM than for dispersion. In  
340 general, the control regions in this experiment show a higher tortuosity than  
341 we previously observed for these rats, expressed by the DM in (Shelton et al.,

2015). Direct comparison of the SOAM in this work and in (Shelton et al., 2015) is not available since the calculation of the SOAM has been adjusted since then. The unusually high tortuosity for control may be caused by the presence of the bowel region in some of the AA images, which was excluded from analysis in earlier studies, and may have elevated tortuosity. Previous data also shows that the SOAM exhibited an intermediate level of tortuosity in tissue up to 1 cm away from a tumor, with a mean tortuosity between that of tumors and non tumor-bearing animals (Rao et al., 2016). The discrimination power of the SOAM in our data set increases for smaller vessels (Cohen’s  $d= 0.14$  for vessels with a radius  $> 0.11$  mm,  $0.28$  with an intermediate radius, and  $0.43$  with a radius  $< 0.09$  mm). Therefore, its relation with the extracted DCE-US features can not be fully appreciated due to the finite resolution of AA. Similarly, a previous study has shown that the difference in MVD between tumor and control increases for smaller vessels (Sedelaar et al., 2001). Therefore, it may be that the SOAM, MVD, and other metrics extracted in this study are related to dispersion; however, mainly smaller vessels’ properties have a significant influence on it. Supporting this hypothesis is the former observation that regions with increased MVD correspond to those with low dispersion (Saidov et al., 2016), as derived from immunohistology. The immunohistology derived MVD was based on evaluation of tomato lectin binding to the endothelial cells and therefore characterized the presence of vessels of all sizes.

Spots of increased vascular density or large vessels were detected with perfusion colormaps. The correlation between median perfusion level and MVD is the only significant inter-parameter agreement found in this work.

367 The Kruskal-Wallis test is ideally constructed for a study design when  
368 subjects are randomly assigned to different groups, so that each subject ap-  
369 pears in one group only (Kruskall and Wallis, 1952). Moreover, the subjects  
370 within the group must be independent. We realize that these assumptions  
371 are not strictly valid in this study, since we observe the tumor evolution in  
372 the same rats over time and since the vessels selected from the same rat are,  
373 strictly speaking, not independent. However, we do not expect these limita-  
374 tions to be crucial for deriving a meaningful conclusion about the significant  
375 trends in time.

376 Imaging initialization was different among 3 series of experiments, start-  
377 ing with day 8, 11, and 13 after tumor implantation, as explained before. We  
378 consider that combining all the rats together according to the number of the  
379 acquisition is justified as the imaging was initialized according to the same  
380 strategy: when the tumors became palpable. However, since we waited for  
381 2-3 days for subsequent assessment if tumors were not palpable on day 8, in  
382 future work it may be beneficial to assess the tumors every day or evaluate  
383 all tumors in a single cohort. This would ensure that the development of the  
384 imaged tumors is more consistent.

385 It is often observed that the wash-out phase is masked by recirculation in  
386 small animals. (Stapleton et al., 2009) shows that for a range of administered  
387 UCA doses the wash-out phase is more prominent in mice. Different UCA  
388 doses should therefore be investigated in our future work, since a prominent  
389 wash-out phase, in our experience, enhances the performance of CUDI (Kue-  
390 nen et al., 2013b). A clear wash-out would also allow evaluating the wash-out  
391 as a complementary perfusion parameter.

392 An important limitation of this study is the 2D character of the extracted  
393 parameters of dispersion and perfusion. The results of the post hoc tests,  
394 therefore, must still be taken with caution since it was performed for 3D  
395 vascular features evaluated in the whole tumor volume and 2D dispersion  
396 and perfusion that leave us blind to out of plane information and restrict us  
397 to the central tumor slice, which is not always representative of the whole  
398 tumor (Streeter et al., 2011). We mitigated this limitation by performing  
399 an additional spatial comparison of the parameter maps in the same plane,  
400 matched with the help of large vessels identified in the perfusion maps. The  
401 agreement between perfusion and MVD, noticed in the longitudinal trends,  
402 was also identified in the spatial distribution of these parameters in the same  
403 plane, raising more confidence to the finding that perfusion and MVD are  
404 correlated.

405 An improved study design should either include 3D DCE-US (Schalk  
406 et al., 2015), giving more accurate overall tumor characteristics, or a regis-  
407 tration procedure, allowing to fix the orientation of the probes and identify  
408 the location of the DCE-US plane within the AA volume. The finding that  
409 perfusion highlights large vessels can be used to further improve registration.

410 The absence of any parameters correlated with dispersion may pinpoint  
411 to the limitation of AA as a validation method for CUDI: while enabling very  
412 high resolution ultrasound imaging, it may not be sufficient to find out which  
413 vascular properties substantially influence dispersion, since dispersion may  
414 be mainly defined by properties of subresolution vessels. In this respect, it is  
415 possible to direct our attention to superlocalization methods that overcome  
416 the limit of diffraction: they are able to track single bubbles and determine

417 their exact positions by finding the centers of their point spread functions  
418 (Cox and Beard, 2015; Errico et al., 2015). Another possible reason for the  
419 absence of vascular parameters that correlate with dispersion is that the  
420 adopted dispersion parameter, is in fact related to both dispersion and flow  
421 velocity (Kuenen et al., 2013a). Different vascular parameters may contribute  
422 to the separate terms of dispersion and flow velocity, while we assessed their  
423 combination. In this regard, it would also be of interest to apply another  
424 analysis to the DCE-US bolus recordings that allows to separate dispersion  
425 and velocity contributions (van Sloun et al., 2017).

## 426 **Conclusions**

427 In this work, dispersion demonstrated its superior performance at tumor  
428 classification compared to perfusion, as previously found for prostate cancer  
429 (Kuenen et al., 2013a,b; Mischi et al., 2012). Perfusion colormaps highlight  
430 large vessels and regions of elevated MVD. The vascular factors that deter-  
431 mine the dispersion level remain yet to be found, as well as the role of vessels  
432 with a diameter below 100-200  $\mu$  in defining perfusion levels.

## 433 **Acknowledgements**

434 This work was supported by the European Research Council Starting  
435 Grant (#280209) and the Impulse2 programme within TU/e and Philips.  
436 This work was also supported by R01CA170665, R43CA165621, and R01CA189479  
437 from the National Institutes of Health. Disclosure: P.A. Dayton declares that  
438 he is an inventor on a patent enabling the acoustic angiography technology

439 and a co-founder of SonoVol, Inc, a company which has licensed this tech-  
440 nology.

#### 441 **Abbreviations**

442 **AA** Acoustic angiography

443 **CUDI** Contrast ultrasound dispersion imaging

444 **DCE-US** Dynamic contrast-enhanced ultrasound

445 **DM** Distance metric

446 **IDC** Indicator dilution curve

447 **LDRW** Local density random walk

448 **MR** Mean radius

449 **MVD** Microvascular density

450 **SOAM** Sum of angles metric

451 **TIC** Time intensity curve

452 **UCA** Ultrasound contrast agent

453 **VL** Vessel length

454 **VVD** Volume vascular density

- 455 Brock M, Eggert T, Rein JP, Roghmann F, Braun K, Bjrn L, Sommerer F,  
456 J. N, Bodman C. Multiparametric ultrasound of the prostate: Adding con-  
457 trast enhanced ultrasound to real-time elastography to detect histopatho-  
458 logically confirmed cancer. *J Urol*, 2013;189:93–98.
- 459 Bullitt E, Gerig G, Pizer SM, Lin W, Aylward SR. Measuring tortuosity of  
460 the intracerebral vasculature from mra images. *IEEE Trans Med Imaging*,  
461 2003;22:1163–1171.
- 462 Cosgrove D, Lassau N. Imaging of perfusion using ultrasound. *Eur J Nucl*  
463 *Med Mol Imaging*, 2010;37:65–85.
- 464 Cox B, Beard P. Super-resolution ultrasound. *Nature*, 2015;527:451–452.
- 465 Errico C, Pierre J, Pezet S, Desailly Y, Lenkei Z, Couture O, M. T. Ultra-  
466 fast ultrasound localization microscopy for deep super-resolution vascular  
467 imaging. *Nature*, 2015;527:499–502.
- 468 Folkman J. Tumor angiogenesis: theraputic implications. *N Engl J Med*,  
469 1971;285:1182–1186.
- 470 Gessner R, Frederick C, Foster F, Dayton P. Acoustic angiography: A new  
471 imaging modality for asesing microvascular architecture. *Int J Biomed*  
472 *Imaging*, 2013;2013.
- 473 Koumoutsakos P, Pivkin I, Milde F. The fluid mechanic of cancer and its  
474 therapy. *Annu Rev Fluid Mech*, 2013;45:325–355.
- 475 Kruskall WH, Wallis WA. Use of ranks in one-criterion variance analysis. *J*  
476 *Am Stat Assoc*, 1952;47:583–621.

- 477 Kuenen M, Saidov T, Wijkstra H, Mischi M. Contrast-ultrasound disper-  
478 sion imaging for prostate cancer localization by improved spatiotemporal  
479 similarity analysis. *Ultrasound Med Biol*, 2013a;39:1631–1641.
- 480 Kuenen M, Saidov T, Wijkstra H, Mischi M. Spatiotemporal correlation  
481 of ultrasound-contrast-agent dilution curves for angiogenesis localization  
482 by dispersion imaging. *IEEE Trans Ultrason Ferroelectr Freq Control*,  
483 2013b;60:2665–2669.
- 484 Kuenen MPJ, Herold IHF, Korsten HHM, de la Rosette JJMCH, Wijkstra  
485 H. Maximum-likelihood estimation for indicator dilution analysis. *IEEE*  
486 *Trans Biomed Eng*, 2014;61:821–831.
- 487 Kuenen MPJ, Mischi M, Wijkstra H. Contrast-ultrasound diffusion imag-  
488 ing for localization of prostate cancer. *IEEE Trans Med Imaging*,  
489 2011;30:1493–1502.
- 490 Mann HB, Whitney D. On a test of whether one of two random variables is  
491 stochastically larger than the other. *Ann Math Stat*, 1947;18:50–60.
- 492 Meiburger KM, Nam S, Chung E, Suggs LJ, Emelianov SY, Molinari F.  
493 Skeletonization algorithm-based blood vessel quantification using in vivo  
494 3d photoacoustic imaging. *Phys Med Biol*, 2016;61:7994–8009.
- 495 Mischi M, Kuenen MPJ, Wijkstra H. Angiogenesis imaging by spatiotempo-  
496 ral analysis of ultrasound-contrast-agent dispersion kinetics. *IEEE Trans*  
497 *Ultrason Ferroelectr Freq Control*, 2012;59:621–629.
- 498 Narang AS, Varia S. Role of tumor vascular architecture in drug delivery.  
499 Vol. 63, 2011.



500 Panfilova A, Shelton S, Sloun R, Demi L, Wijkstra H, Dayton P, Mischi M.  
501 Does contrast ultrasound dispersion imaging reveal changes in tortuosity?  
502 a comparison with acoustic angiography, 2016.

503 Ponce A, Viglianti B, Yu D, Yarmolenko P, Michelich C, Woo J, Bally M, De-  
504 whirst M. Magnetic resonance imaging of temperature-sensitive liposome  
505 release: Drug dose painting and antitumor effects. *J. Natl Cancer Inst*,  
506 2007;99:53–63.

507 Quaia E. Assessment of tissue perfusion by contrast-enhanced ultrasound.  
508 *Eur Radiol*, 2011;21:604–615.

509 Rao S, Shelton S, Dayton P. The 'fingerprint' of cancer extends beyond solid  
510 tumor boundaries: assessment with a novel ultrasound imaging approach.  
511 *IEEE Trans Biomed Eng*, 2016;63.

512 Rognin NG, Frinking P, Costa M, Arditi M. In-vivo perfusion quantifica-  
513 tion by contrast ultrasound: Validation of the use of linearized video data  
514 vs. raw rf data. In: 2008 IEEE Int Ultrason Symp Proceedings. Piscat-  
515 away:IEEE, 2008. pp. 1690–1693.

516 Saidov T, Heneweer C, Keunen M, Broich-Oppert J, Wijkstra H, Rosette J,  
517 Mischi M. Fractal dimension of tumor microvasculature by dce-us: prelim-  
518 inary study in mice. *Ultrasound Med Biol*, 2016;42:2852–2863.

519 Schalk S. Towards 3d prostate cancer localization by contrast ultrasound  
520 dispersion imaging. Ph.D. thesis, 2017.

521 Schalk SG, Demi L, Smeenge M, Millis DM, Wallace K, de la Rosette J,  
522 Wijkstra H, Mischi M. 4d spatiotemporal analysis of ultrasound contrast

523 agent dispersion for prostate cancer localization: a feasibility study. *IEEE*  
524 *Trans. Ultrason. Ferroelectr. Freq. Control*, 2015;62:839–851.

525 Sedelaar JM, Leenders G, Hulsbergen-van de Kaa C, Poel H, Laak J, De-  
526 bruyne F, Wijkstra H, Rosette J. Microvessel density: Correlation be-  
527 tween contrast ultrasonography and histology of prostate cancer. *Eur Urol*,  
528 2001;40:285–293.

529 Shelton S, Lee Y, Lee M, Cherin E, Foster F, Aylward S, Dayton P. Quantifi-  
530 cation of microvascular tortuosity during tumor evolution utilizing acoustic  
531 angiography. *Ultrasound Med Biol*, 2015;41:1869–1904.

532 Stapleton S, Goodman H, Yu-Qing Z, Cherin E, Henkelman R, Burns P,  
533 Foster F. Acoustic and kinetic behavior of definity in mice exposed to high  
534 frequency ultrasound. *Ultrasound Med Biol*, 2009;35:296–307.

535 Streeter J, R.C. G, Tsuruta J, Feingold S, Dayton P. Assessment of molec-  
536 ular imaging of angiogenesis with three-dimensional ultrasonography. *Mol*  
537 *Imaging*, 2011;10:460–468.

538 Strouthos C, Lampaskis M, Sboros V, McNeilly A, Averkiou M. Indicator  
539 dilution models for the quantification of microvascular blood flow with  
540 bolus administration of ultrasound contrast agents. *IEEE Trans Ultrason*  
541 *Ferroelectr Freq Control*, 2010;57:1296–1310.

542 Sullivan MG, Feinn R. Using effect size - or why the p-value is not enough.  
543 *J Grad Med Educ*, 2012;4:279–282.

- 544 Tozer GM, Lewis S, Michalowski A, Aber V. The relationship between re-  
545 gional variations in blood flow and histology in a transplanted rat fibrosar-  
546 coma. *Br. J. Cancer*, 1990;61:250–257.
- 547 Vala HJ, Baxi A. A review on otsu image segmentation algorithm. *Int J Adv*  
548 *Res Comput Sci Softw Eng*, 2013;2:387–389.
- 549 van Sloun RJ, Demi L, Postema AW, de la Rosette JJ, Wijkstra H, Mischi  
550 M. Ultrasound-contrast-agent dispersion and velocity imaging for prostate  
551 cancer localization. *Med Image Anal*, 2017;35:610–619.
- 552 Viglianti B, Abraham S, Michelich C, Yarmolenko P, MacFall J, Bally  
553 M, Dewhirst M. In vivo monitoring of tissue pharmacokinetics of lipo-  
554 some/drug using mri: Illustration of targeted delivery. *Magnet Reson Med*,  
555 2004;51:1153–1162.
- 556 Zhao H, Xu R, Ouyang Q, Chen L, Dong B, Huihua Y. Contrast-enhanced  
557 ultrasound is helpful in the differentiation of malignant and benign breast  
558 lesions. *Eur J Radiol*, 2010;73:288–293.

559 **Figure Captions**

560 **Figure 1:** DCE-US and AA images of the same plane, and maps of the  
561 extracted features. a: maximum intensity projection of the DCE-US  
562 video. The tumor is encircled by a red contour, while the region out-  
563 side the blue contour belongs to the control, separated by a margin  
564 which was not included in the analysis. Regions with power below the  
565 threshold of -22 dBs of the maximum intensity are displayed in black.  
566 b-c: perfusion and dispersion colormaps, respectively. Regions with  
567 power below -22 dBs of the maximum intensity are displayed in white.  
568 d: Selected AA slice. e-f: vascular skeleton, colorcoded according to  
569 the values of the microvascular desity, and mean radius, respectively  
570 (yellow indicates low values, while red inicates high values). The num-  
571 bers in b and d illustate the vessels identified in the perfusion maps,  
572 used as markers to locate the right plane in AA volumes.

573 **Figure 2:** A typical preprocessed indicator dilution curve. T1 shows the  
574 appearance time, T0 is taken 3 seconds before appearance time. The  
575 interval from T0 to T2 shows the interval of the IDC used for disper-  
576 sion analysis. The tangens of the angle alpha of the line fitted to the  
577 indicator dilution curve in the 2 seconds after appearance time is the  
578 wash-in-rate.

579 **Figure 3:** AA maximum intensity projection. The tumor region is indicated  
580 by the red contour, surrounded by the control region.

581 **Figure 4:** Boxplots of tumor and control parameters, binned together from  
582 all time points. a: dispersion, b: perfusion, c: distance metric, d: sum

583 of angles metric, e: vessel length, f: vessel radius, g: microvascular  
584 density, h: volume vascular density. Cohen's d measure is indicated  
585 above the plots.

586 **Figure 5:** Boxplots of tumor (T1, T2, T3, T4) and control (C1, C2, C3, C4)  
587 parameters, binned together at different time points. a: dispersion,  
588 b: perfusion, c: volume vascular density, d: microvascular density, e:  
589 vessel radius, f: vessel length, g: sum of angles metric.

590 **Figure 6:** Results of the post hoc Kruskal-Wallis test performed on tumor  
591 and control parameters at four time points (indicated by T1, T2, T3,  
592 T4 and by C1, C2, C3, C4, respectively). The colors of the rows indi-  
593 cate whether the distribution is significantly different from the others,  
594 green and yellow representing different significance levels. a: disper-  
595 sion, b: perfusion, c: volume vascular density, d: microvascular density,  
596 e: vessel radius, f: vessel length, g: sum of angles metric.