



## Complex target SELEX-based identification of DNA aptamers against Bungarus caeruleus venom for the detection of envenomation using a paper-based device

Anand, Anjali; Chatterjee, Bandhan ; Dhiman, Abhijeet; Goel, Renu ; Khan, Eshan; Malhotra, Anita; Santra, Vishal; Salvi, Nitin ; Bhatnagar, Ira ; Kumar, Amit; Asthana, Amit ; Sharma, Tarun Kumar

### Biosensors and Bioelectronics

DOI:

<https://doi.org/10.1016/j.bios.2021.113523>

Published: 01/12/2021

Peer reviewed version

[Cyswllt i'r cyhoeddiad / Link to publication](#)

*Dyfyniad o'r fersiwn a gyhoeddwyd / Citation for published version (APA):*

Anand, A., Chatterjee, B., Dhiman, A., Goel, R., Khan, E., Malhotra, A., Santra, V., Salvi, N., Bhatnagar, I., Kumar, A., Asthana, A., & Sharma, T. K. (2021). Complex target SELEX-based identification of DNA aptamers against Bungarus caeruleus venom for the detection of envenomation using a paper-based device. *Biosensors and Bioelectronics*, 193, [113523]. <https://doi.org/10.1016/j.bios.2021.113523>

#### Hawliau Cyffredinol / General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal ?

#### Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

# Journal Pre-proof

Complex target SELEX-based identification of DNA aptamers against *Bungarus caeruleus* venom for the detection of envenomation using a paper-based device

Anjali Anand, Bandhan Chatterjee, Abhijeet Dhiman, Renu Goel, Eshan Khan, Anita Malhotra, Vishal Santra, Nitin Salvi, M.V. Khadilkar, Ira Bhatnagar, Amit Kumar, Amit Asthana, Tarun Kumar Sharma

PII: S0956-5663(21)00560-1

DOI: <https://doi.org/10.1016/j.bios.2021.113523>

Reference: BIOS 113523

To appear in: *Biosensors and Bioelectronics*

Received Date: 7 April 2021

Revised Date: 30 June 2021

Accepted Date: 18 July 2021

Please cite this article as: Anand, A., Chatterjee, B., Dhiman, A., Goel, R., Khan, E., Malhotra, A., Santra, V., Salvi, N., Khadilkar, M.V., Bhatnagar, I., Kumar, A., Asthana, A., Sharma, T.K., Complex target SELEX-based identification of DNA aptamers against *Bungarus caeruleus* venom for the detection of envenomation using a paper-based device, *Biosensors and Bioelectronics*, <https://doi.org/10.1016/j.bios.2021.113523>.

This is a PDF file of an article that has undergone enhancements after acceptance, such as the addition of a cover page and metadata, and formatting for readability, but it is not yet the definitive version of record. This version will undergo additional copyediting, typesetting and review before it is published in its final form, but we are providing this version to give early visibility of the article. Please note that, during the production process, errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

© 2021 Elsevier B.V. All rights reserved.



**CRedit authorship contribution statement**

Anjali Anand: Methodology, data curation, analysis and writing-original draft. Bandhan Chatterjee: Data curation and writing-original draft, Abhijeet Dhiman: Data curation. Renu Goel: Data Curation and MS data analysis. Eshan Khan: Data Curation. Anita Malhotra: Resources, venom collection Writing-review and editing. Vishal Santra: Resources, venom collection writing-review and editing. Nitin Salvi: venom collection Writing-review and editing. M.V. Khadilkar: venom collection writing-review and editing. Ira Bhatnagar: Data curation. Amit Kumar: CD data analysis writing-review and editing. Amit Asthana: review & editing, Supervision. Tarun K. Sharma: Conceptualization, Writing – review & editing, Supervision.

1 **Complex target SELEX-based identification of DNA aptamers against *Bungarus caeruleus***  
2 **venom for the detection of envenomation using a paper-based device**

3 *Anjali Anand<sup>1†</sup>, Bandhan Chatterjee<sup>1†</sup>, Abhijeet Dhiman<sup>1</sup>, Renu Goel<sup>2</sup>, Eshan Khan<sup>3</sup>, Anita*  
4 *Malhotra<sup>4</sup>, Vishal Santra<sup>5,6</sup>, Nitin Salvi<sup>7</sup>, M.V. Khadilkar<sup>7</sup>, Ira Bhatnagar<sup>8</sup>, Amit Kumar<sup>3</sup>, Amit*  
5 *Asthana<sup>8,9\*</sup> and Tarun Kumar Sharma<sup>1\*</sup>*

6 <sup>1</sup>Aptamer Technology and Diagnostics Laboratory, Multidisciplinary Clinical and Translational  
7 Research Group, Translational Health Science and Technology Institute (THSTI), Faridabad,  
8 Haryana, 121001, India.

9 <sup>2</sup>Translational Health Science and Technology Institute (THSTI), Faridabad, Haryana, 121001,  
10 India

11 <sup>3</sup>Discipline of Biosciences and Biomedical Engineering, Indian Institute of Technology Indore,  
12 Simrol, Indore, 453552, India.

13 <sup>4</sup>School of Natural Sciences, College of Environment Sciences and Engineering, Bangor  
14 University, Bangor, LL57 2UW, UK.

15 <sup>5</sup>Simultala Conservationists (Foundation for Wildlife), Nalikul, Hooghly, 712407, West Bengal,  
16 India.

17 <sup>6</sup>Captive and Field Herpetology, 13 Hirfron, Anglesey, LL65 1YU, Wales, UK.

18 <sup>7</sup>Premium Serums and Vaccines Pvt. Ltd, Maharashtra, 410504, India.

19 <sup>8</sup>CSIR- Centre for Cellular & Molecular Biology, CCMB-Annexe-II, Medical Biotechnology  
20 Complex, Uppal Road, Uppal, Hyderabad - 500 039, Telangana, India.

21 <sup>9</sup> Department of Pharmaceutical Analysis, National Institute of Pharmaceutical Education And  
22 Research (NIPER), Balanagar, Hyderabad, Telangana, 500037, India

23 \*Email: Tarun K. Sharma, tarun@thsti.res.in; Telephone: +91-129-2876370.

24 Amit Asthana, amit.asthana@niperhyd.ac.in; Telephone: +91 9666244410

25 † Contributed equally

## 26 **Abstract**

27 Complex target SELEX always have been an intriguing approach to the scientific community, as  
28 it offers the potential discovery of novel biomarkers. We herein have successfully performed  
29 SELEX on *Bungarus caeruleus* venom to develop a panel of highly affine aptamers that  
30 specifically recognizes the *B. caeruleus* (common krait) venom and was able to discriminate the  
31 *B. caeruleus* venom from Cobra, Russell's, and Saw-scaled viper's venom. The aptamers  
32 generated against the crude venom also lead to the identification of the specific component of the  
33 venom, which is  $\beta$ -Bungarotoxin, a toxin uniquely present in the *B. caeruleus* venom. The best  
34 performing aptamer candidates were used as a molecular recognition element in a paper-based  
35 device and were able to detect as low as 2 ng krait venom in human serum background. The  
36 developed aptamer-based paper device can be used for potential point-of-care venom detection  
37 applications due to its simplicity and affordability.

38 **Keywords:** Aptamers; SELEX; Krait; Snake venom; Bungarotoxin; Mass-Spectrometry; Paper  
39 device.

40

41

## 42 **1. Introduction**

43 The development of aptamers against complex targets or disease conditions where specific  
44 biomarkers are not known has always been an intriguing challenge for the scientific community  
45 (Shamah et al., 2008). Complex targets, like cells, with a myriad of possible targets expressed on  
46 their membranes (Takahashi, 2018), or heterogeneous mixture of proteins like secretome of  
47 cancer cells, infectious pathogens (Bonin-Debs et al., 2004; Ranganathan and Garg, 2009) and/or  
48 snake venoms (Tan et al., 2018; Warrell, 2012) provide an exciting possibility of developing  
49 aptamers against an unknown component of the heterogeneous mixture. This can also result in  
50 the aptamer-based discovery of novel biomarkers, which can then be used for diagnostic  
51 purposes (Berezovski et al., 2008).

52 In recent years, complex target SELEX has proven its utility to identify disease-specific  
53 biomarkers from the secretome of cancer cells. Using this approach, an aptamer that could  
54 discriminate between the secretome of pancreatic cancer and non-cancerous cells was identified.  
55 By utilizing biochemical purification methods and mass-spectrometric analysis aptamer target  
56 was identified as cyclophilin B, a biomarker uniquely present in the secretome of pancreatic  
57 cancer cells (Ray et al., 2012). Later, the same biomarker was identified in the serum of  
58 pancreatic cancer patients. Besides, using a complex target SELEX strategy, aptamers against  
59 protozoan and bacterial pathogens have also been designed leading to the discovery of unique  
60 pathogen-specific biomarkers (Shamah et al., 2008; Ulrich and Wrenger, 2009).

61 Snake venom is a rather more complex mixture as it consists of several different proteins,  
62 peptides, amines, carbohydrates, and various other substances (Warrell, 2010). The development  
63 of binders (eg. antibodies) against a specific venom component requires the same in its purified  
64 form. However, the purification of the aforementioned individual components from crude venom

65 is very challenging. Further, cloning, expression and purification of various proteins present in  
66 venom also pose challenges as gene cloning procedure require mRNA from venom glands for  
67 that one need to sacrifice the snake which is not permissible as per wildlife act in several  
68 countries.

69 A large body of literature suggests that antibodies developed in response to inactivated crude  
70 venom of particular species are highly cross-reactive due to the high molecular weight of  
71 proteins and common epitopes present in the venom of other species (Fernandes et al., 2000;  
72 Ledsgaard et al., 2018; Stábeli et al., 2005). On the other hand, lethal toxins may be species-  
73 specific but they are poorly immunogenic owing to their low molecular weight (Knudsen and  
74 Laustsen, 2018; Laustsen et al., 2017). Considering the complexity of snake venom, complex  
75 target SELEX offers an exciting opportunity to screen species-specific aptamers from a large  
76 aptamer library that can target species-specific uniquely present small molecular weight toxins.  
77 Such aptamers can be a valuable tool for assessing the true burden of snake envenomation and  
78 conducting forensic investigations (Brunda et al, 2006; Theakston and Laing, 2014). Snake  
79 envenomation is a major public health concern around the world, particularly in tropical  
80 countries where they are a major contributor to mortality and morbidity (Murray et al., 2015;  
81 Suraweera et al., 2020). In large parts of India, the majority of mortality is attributable to the  
82 “Big Four” venomous species, which comprise the Indian or spectacled cobra (*Naja naja*),  
83 common krait (*Bungarus caeruleus*), saw-scaled viper (*Echis carinatus*) and Russell's viper  
84 (*Daboia russelii*) (Choudhury et al., 2017; David A Warrell, 1999; Puzari and Mukherjee, 2020)  
85 with an estimated annual average death count of 58,000 (Suraweera et al., 2020). Thus, for  
86 effective snakebite management, rapid and accurate identification of the envenoming species is  
87 the key to reduce the severity and fatality associated with the envenoming (Mohapatra et al.,

2011). The current diagnosis regimen primarily involves clinical examination that mainly relies upon symptoms (Ariaratnam et al., 2009; Sano-Martins et al., 1994; Warrell, 2012) and further confirmation involves antibody-based detection in some cases (Warrell, 2010). However, using an antibody for molecular recognition has many limitations associated with it; for instance, the development of antibodies depends on biological systems (mainly animals), provides batch-to-batch variation, high cost of development, and requirement of refrigeration for storage to name a few (Bandhan Chatterjee et al., 2020; Kaur et al., 2018). Many of these limitations posed by the antibodies can be readily ameliorated by aptamers, the chemical surrogates of antibodies.

Aptamers are single-stranded nucleic acid molecules that form a variety of secondary structures functionally mimicking the epitope binding sites of antibodies. These polynucleotides are chemically synthesized so are relatively very cheap and lacking any batch-to-batch variation. Once developed against a target, they can be rapidly mass-produced in a short time and do not require any stringent storage conditions (Dhiman et al., 2017; Kaur et al., 2019; Taneja et al., 2020).

In the recent past, several assays and diagnostic tests based on different techniques and principles were evolved which includes, bioassays, immunodiffusion, immunoelectrophoresis, immunofluorescence, haemagglutination, radioimmunoassay (RIA), enzyme-linked immunosorbent assay (ELISA) and optical immunoassay (Gopalakrishnakone et al., 2015). However, there exists a major gap in the non-availability of a low-cost point-of-care (POC) device that can diagnose snake venom using body fluid is the major hurdle in the effective treatment of snakebite through the administration of specific antivenom (Michael, 2013). Microfluidic, paper-based and lateral flow devices are the possible solution for POC application for the rapid and affordable detection of snake-venom. There are few reports on microfluidic



111 way of profiling snake-venom. However, none for the rapid detection of snake bite (Slagboom et  
112 al., 2018; Zancolli et al., 2017). There are few reports on the development of lateral flow-based  
113 immunochromatographic assay and dot ELISA-based (Shaikh et al., 2017) specific snake venom  
114 detection. Both the methods employed antibodies of rat and rabbit origin and are time-  
115 consuming. Although, a preliminary data on the paper-based sensor was reported by our group  
116 but this sensor had utilized antibodies as a molecular recognition element that evinced batch-to-  
117 batch variation thus limits its scale-up and POC application (Michael, 2013). This limitation can  
118 easily be overcome by aptamers.

119 Using a complex target SELEX strategy, we herein, report the development of a panel of highly  
120 affine aptamers against the crude venom of *B. caeruleus* (Common Indian Krait). To the best of  
121 our knowledge, this is the first report of successful SELEX on crude snake venom. The SELEX  
122 was performed against the crude venom of *B. caeruleus*, which is a heterogeneous mixture of  
123 proteins, peptides, enzymes, amines, carbohydrates, and various other substances. The developed  
124 aptamers have shown highly selective binding for the venom of *B. caeruleus* and were able to  
125 discriminate it from the venom of other snakes as well as red scorpion venom. The best  
126 performing aptamer candidate was able to detect *B. caeruleus* venom in both buffers as well as  
127 human serum background with a limit of detection of as low as 2 ng. The molecular target was  
128 also identified using biochemical and mass-spectrometric analysis. Further, to demonstrate its  
129 possible POC application we have also adapted aptamers onto a paper-based device.

## 130 **2. Materials and methods**

### 131 ***2.1 Reagents and Chemicals***

132 All routine reagents were procured from Sigma Aldrich, USA. Oligonucleotides used in the  
133 study were procured from Integrated DNA Technologies (IDT, USA). Ninety-six well plates  
134 (MaxiSorp™) were procured from Thermo Fischer Scientific, U.S.A. 3,3',5,5'-  
135 tetramethylbenzidine (TMB) (BD OptEIA™) was procured from BD Biosciences, USA.

## 136 ***2.2 Snake venom collection and procurement***

137 Snake venom was obtained from wild specimens collected under permit numbers 5141/WL/4R-  
138 6/2017, A.33011/5/2011-CWLW/305, and W.L./Research Study/WLM/2341 issued by West  
139 Bengal, Mizoram and Himachal Pradesh Forest Departments, India and their specific identity  
140 were confirmed by professional herpetologists. The specimens were handled according to  
141 relevant guidelines or regulations and were released after milking. For the current work, ethical  
142 permission was obtained from Institutional Ethics Committee, Bangor University, U.K. and  
143 Translational Health Science and Technology Institute, India. In addition to this, venom from the  
144 'Big Four' species was also procured from KV Institute, Uttar Pradesh, India and the Irula Snake  
145 Catchers Industrial Co-operative Society (ISCICS), Vadanemmel village, Kancheepuram  
146 District, Tamil Nadu, India. Biosafety permissions to handle snake venom at the institute were  
147 also obtained from the Institutional Biosafety Committee, THSTI, Faridabad-121001, Haryana,  
148 India.

## 149 ***2.3 Aptamer development through SELEX***

150 The G-quadruplex biased and completely random libraries as described in recent work from our  
151 group (Kalra et al., 2018) were used to screen aptamers against venom of *B. caeruleus*. These  
152 libraries have identical 18 nucleotide long primer binding sites for DNA amplification. For PCR  
153 amplification DRF (Forward- 5' GTC TTG ACT AGT TAC GCC 3') and DRR (Reverse - 5'  
154 GAG GCG CCA ACT GAA TGA-3') primers were used to prime the template and nascent

155 strands, respectively. For single-stranded (ss)DNA generation, PCR was performed using 5'  
156 FAM-labeled DRF (Fluorescent in nature) and 3' rA-modified DRR (5' GAG GCG CCA ACT  
157 GAA TGrA-3') primers. The process of ssDNA generation was same as described recently (B.  
158 Chatterjee et al., 2020; Dhiman et al., 2018; Kalra et al., 2018). To develop aptamers against the  
159 crude venom of *B. caeruleus*, a nitrocellulose membrane (NCM)-based subtractive Systematic  
160 Evolution of Ligands through EXponential enrichment (SELEX) approach was adopted.

161 Experimentally, a mixture of DNA libraries containing 1500 pmol of each library in selection  
162 buffer (SB, 10 mM Tris pH 7.5 supplemented with 10 mM MgCl<sub>2</sub>, 50 mM KCl and 25 mM  
163 NaCl) was heated at 92 °C, followed by snap chilling on ice, and then brought to room  
164 temperature. Such prepared libraries were then incubated for 1 hour at room temperature (RT)  
165 with a cocktail of venom from *N. naja*, *D. russelii* and *E. carinatus* dissolved in Nuclease Free  
166 Water (NFW) immobilized on NCM, as a counter-selection step. The unbound sequences were  
167 then incubated with NCM-immobilized *B. caeruleus* venom for 1 hour at RT . After incubation,  
168 the membrane was washed with washing buffer (SB supplemented with 0.5% Tween-20) to  
169 remove the unbound sequences. Venom-bound sequences were then eluted by heating the  
170 membrane in NFW at 92 °C for 10 minutes, and the resultant solution was used for PCR  
171 amplification of the eluted sequences. Amplification of aptamers followed by generation of  
172 ssDNA population was performed as described recently by our group (Bandhan Chatterjee et al.,  
173 2020; Dhiman et al., 2018; Taneja et al., 2020).

174 The selection pressure was gradually increased with every successive round by increasing the  
175 amount of the counter-selection venoms and simultaneously reducing the amount of target  
176 venom. The number of washes and the strength of T-20 in the washing buffer were also  
177 increased up to 1.5% T-20 as selection progresses. Additionally, salmon sperm DNA and dextran

178 sulfate were added to reduce the non-specific interactions during SELEX. In addition, counter  
179 selection was also given with human serum, and the volume of human serum was gradually  
180 increased from round 1-8. After eight rounds of SELEX, the aptamer pools from rounds 2, 4, 6  
181 and 8 were evaluated using an Aptamer Linked Immobilized Sorbent Assay (ALISA) to monitor  
182 the progress of aptamer selection. The aptamer pool of the round displaying the highest binding  
183 towards the target venom was sent for Next Generation Sequencing (NGS) at Eurofins Genomics  
184 service, India.

#### 185 ***2.4 Aptamer Linked Immobilized Sorbent Assay (ALISA)***

186 ALISA was used to evaluate the binding of the developed aptamer candidates towards *B.*  
187 *caeruleus* venom. Experimentally, 500 ng snake venom (krait or other species) was coated onto a  
188 96-well plate using standard 100 mM carbonate-bicarbonate buffer, pH 9.6 at 37 °C for 1.5  
189 hours. After this, the coating solution was discarded and marginal sites were blocked with 5%  
190 skimmed milk supplemented with 0.25% T-20 in SB.

191 After blocking, wells were washed with SB once and then 100 picomoles of 5' biotin-labeled  
192 aptamer were added to each well for 1 hr at RT. Following this, the plates were washed with SB  
193 and SB supplemented with 1% T-20 (v/v). Next, biotinylated aptamer bound to venom was  
194 probed with 1:3000 (v/v) streptavidin-horseradish peroxidase (Sigma Aldrich U.S.A.) and was  
195 further incubated for 1 hr at RT. Finally, 100  $\mu$ L of TMB substrate was added to each well and  
196 incubated for 3-5 min at RT. The reaction was quenched using 5% H<sub>2</sub>SO<sub>4</sub> and optical density  
197 (OD) was measured at 450 nm.  $\Delta$ OD<sub>450</sub> obtained, by subtracting OD<sub>450</sub> of appropriate negative  
198 controls (aptamer control and antigen control) was plotted.

199

#### 200 ***2.5 Assessment of cross-reactivity of aptamers***

201 To assess the cross-reactivity of SELEX-derived aptamers, 10 aptamer candidates were screened  
202 by ALISA for their ability to bind with the venom of the ‘Big Four’ species. Based on the  
203 outcome of this experiment, two aptamer candidates were selected and their binding was  
204 assessed for eight different snake venoms including ‘Big Four’, namely venom obtained from *N.*  
205 *naja*, *N. kaouthia*, *N. oxiana*, *D. russelii*, *B. caeruleus*, *B. fasciatus*, *B. niger*, and *E. carinatus*. In  
206 addition to snake venoms, we have also assessed cross-reactivity of aptamers for red scorpion  
207 (*Hottentotta tamulus*, *H. tamulus*) venom as well. These species were chosen based on their  
208 distribution and importance in terms of snake and scorpion bite in the Indian subcontinent. The  
209 ALISA was performed as described in the previous section.

### 210 **2.6 Circular Dichroism (CD)**

211 The CD experiments were performed on J-815 Spectropolarimeter (JASCO, Tokyo Japan) to  
212 determine the secondary structure of two selected aptamer candidates. A quartz cuvette with  
213 0.2 cm path length was used to record the spectra of samples containing 20  $\mu\text{M}$  of each aptamer  
214 in binding buffer, using an average of three scans.

### 215 **2.7 Truncation of aptamers**

216 Based on the NUPACK (<http://www.nupack.org>) predicted secondary structure truncated  
217 variants of two selected aptamers were designed and their binding was assessed and compared  
218 with their respective parents’ counterparts in an ALISA.

### 219 **2.8 Determination of apparent dissociation constant ( $K_d$ )**

220 The dissociation constant ( $K_d$ ) of the selected aptamers (B6, B8 and combination of both) was  
221 gauged with ALISA for *B. caeruleus* venom. The ALISA was performed as described in the  
222 earlier section. Briefly, a fixed venom amount was subjected to various aptamer concentrations

223 ranging from 2 to 500 nM. The absorbance at 450 nm was plotted as a function of aptamer  
224 concentration and  $K_d$  was measured using the following equation in Graph-pad Prism version 7:

$$225 \quad Y = Bmax \frac{X}{Kd + X}$$

226 Here, Y represents the aptamer binding; X is aptamer concentration and Bmax is maximum  
227 binding.

### 228 ***2.9 Limit of detection of selected aptamers***

229 The limit of detection (LOD) of venom of *B. caeruleus* for selected aptamers (B6, B8 and  
230 combination of both) was determined by ALISA. For this ALISA was performed with different  
231 amounts of *B. caeruleus* venom ranging from 2-1000 ng/well. Rest protocol was followed as  
232 described under section 2.4.

### 233 ***2.10 Evaluation of aptamer candidates to detect geographically distinct Krait venom***

234 Venom variability among geographically distinct populations of same species might pose a  
235 challenge in the species-specific diagnosis of venom (Casewell et al., 2020; Chippaux et al.,  
236 1991a). To address this challenge, we have also assessed the aptamer binding against krait  
237 venom obtained from three geographically distinct populations (Central, East and Southern  
238 India) using ALISA as described in the aforementioned section. To determine the effect of  
239 venom collection time along with geography we have performed ALISA with 5 samples of *B.*  
240 *caeruleus* venom collected at three different sites between years 2017-2021 (Detailed  
241 information given in result section and **Figure S3** in supplementary material).

242

243

### 244 **2.11 Fabrication of paper-based devices.**

245 All paper microfluidic devices used in the current work were fabricated on Whatman®  
246 qualitative filter paper (Grade 4, thickness 0.205 mm, Cytiva, India). All the paper-based devices  
247 used in this work were designed using the CorelDraw X8 software (Corel Corporation, Ottawa,  
248 Canada). The design was printed using a wax printer (ColorQube 8570, Xerox India Ltd., India).  
249 The printed devices were kept on the hot plate at 120°C for 2 minutes to melt the wax so that it  
250 can percolate down to the other side of the paper and forms the hydrophobic barrier. The printed  
251 wax devices were then cut using a computer-controlled laser cutting and engraving machine  
252 (Model CMA 6040, GD Han's Yueming Laser Group Co., Ltd, China) or with simple scissors for  
253 further use. All the paper-based devices used in this work are used without any further  
254 modifications.

### 255 **2.12 Development of aptamer and paper-based devices**

256 To detect the *B. caeruleus* venom on a paper-based device, initially, a known amount of snake  
257 venom was coated on the surface of the hydrophilic reaction zone, the white circled area (CA) of  
258 a paper-strip using the pipette. Uncoated CA was served as a venom control (VC). After coating,  
259 the strip was dried at RT for 5 minutes. Thereafter, the CA was blocked with 5% skimmed milk  
260 supplemented with 0.25% T-20 in SB for 30 min. CA was then washed one time with SB  
261 followed by the addition of 100 picomoles of biotinylated B6 and B8 aptamers in reaction  
262 volume of 10 µl followed by an incubation of 30 min at RT. Thereafter, the CA was washed  
263 twice with SB supplemented with 1% T-20 followed by the addition of streptavidin-HRP  
264 (1:2000; 10 µl/CA) and incubated for 30 min at RT. Next to this, paper-based device was washed  
265 as described above. Finally, the TMB substrate (10 µl/CA) was used as the substrate for HRP,  
266 which gives blue color. The blue CA indicating the presence of *B. caeruleus* venom while no

267 such color was observed in absence of snake venom. Finally, the image of the paper strip was  
268 captured by using an android smart phone camera (Samsung A series).

269

270

### 271 ***2.13 Limit of detection***

272 To determine the low-end detection limit of *B. caeruleus* venom on paper-based device, a range  
273 (1000-2 ng) of venom was spiked in the coating solution. The rest of the steps such as blocking,  
274 washing, incubation time, amount of biotinylated B6/B8 aptamers, and streptavidin-HRP were  
275 followed as mentioned previously. Further, to study the effect of sample matrix on krait-specific-  
276 aptamer-venom interaction, human serum sample obtained from a healthy individual was diluted  
277 (1:100 in coating solution), and then it was spiked with a range of (1000-2 ng) of *B.*  
278 *caeruleus* venom. Following this, the appearance of blue color was used to determine the lowest  
279 possible amount of venom that can be detected in the serum.

### 280 ***2.14 Sample preparation for LC-MS/MS***

281 Ten micrograms of crude venom were resolved by SDS-PAGE. The gel was stained with  
282 Coomassie brilliant blue and destained with water. The specific band of ~55kDa was cut from  
283 the gel. Forty mM ammonium bicarbonate (ABC) in 40% acetonitrile (ACN) was used to destain  
284 the excised bands. Reduction and alkylation were done by 5 mM dithiothreitol (DTT) at 60 °C  
285 for 45 min and 10 mM iodoacetamide (IAA) on the gel bands respectively. The gel pieces were  
286 dehydrated using 100% ACN and dried further for 10 mins at room temperature (RT). The in-  
287 gel digestion was carried out as explained previously (Goel et al., 2013). Trypsin (Gold mass-  
288 spectrometry trypsin; Promega, Madison, WI) was added in ice-cold tubes and kept at 37 °C for  
289 10-12 h. Peptides were removed from the gel pieces by adding 50% ACN with 0.1% formic acid



290 (FA) in the tubes. Finally, the same step was carried with 100% ACN with 0.1% FA. The  
291 peptides were lyophilized and kept at  $-80\text{ }^{\circ}\text{C}$  until LC-MS/MS analysis.

### 292 ***2.15 LC-MS/MS analysis***

293 Digested samples were reconstituted in 0.1% FA and analyzed by reverse-phase high-pressure  
294 liquid chromatography-electrospray ionization tandem mass spectrometry using an Eksper-  
295 nanoLC 415 system (Eksigent; Dublin, CA) which is directly connected to a Sciex 5600 Triple-  
296 TOF (SCIEX; Concord, Canada) mass spectrometer.

297 Mobile phase A consisted of 2% acetonitrile/98% of 0.1% formic acid (v/v) in water, and 98%  
298 acetonitrile/2% of 0.1% formic acid (v/v) in water was used as mobile phase B during reverse-  
299 phase high-pressure liquid chromatography. The analytical column (75 $\mu\text{m}$  x 15 cm) from  
300 Eksigent used for the peptides separation and retention time drift was maintained by keeping the  
301 temperature constant at  $35^{\circ}\text{C}$ . The acquired raw files were saved in .wiff format. Autocalibration  
302 of MS and MS-MS were done with 25 fmol b-gal, and 20  $\mu\text{m}$  SilicaTip electrospray PicoTip  
303 emitter (New Objective Cat. No. FS360-20-10-N-5-C7-CT) was used to inject the peptides into  
304 the mass spectrometer.

305 A high-resolution TOF-MS scan over a mass range 350–1250 m/z was used for Data-Dependent  
306 Acquisition (DDA) and intensity greater than 150 cps and charge state between +2 to +5, was  
307 used to select the parent ion. Once the parent ion and isotopes were fragmented by MS/MS, they  
308 were excluded for 12 s from further MS/MS fragmentation. Rolling collision energy was used to  
309 trigger collision-induced dissociation. Accumulation time of 250 ms and 70 ms ions was used for  
310 MS and MS/MS respectively.

### 311 ***2.16 Database search***

312 MS/MS spectra were searched in Protein Pilot software v. 5.0.1 (SCIEX). The following settings  
313 were used for Paragon search: Sample type: Identification; Dithiothreitol, Iodoacetamide were  
314 used for reduction and alkylation respectively, Digestion: Trypsin; TripleTOF 5600 as  
315 instrument type: Species: *B. caeruleus*; Thorough ID for Search effort. Carbamidomethylation  
316 was used as a fixed modification. A confidence score of  $> 0.05$  was used for peptide  
317 identification for consideration for further analysis. *B. caeruleus* sequences were fetched from the  
318 UniProt website ([www.uniprot.org](http://www.uniprot.org)) and contamination list was added to this. False discovery  
319 rate analysis was also performed. A peptide and product ion tolerance of 0.05 Da was used in  
320 searches. The output file from this search contains protein identification, UniProt accession  
321 number, cleaved and modified peptide sequences, relative intensity, precursor and fragment ion  
322 charge and unused Protscore. Wiff files were also processed with MaxQuant using the protein  
323 identification parameter (Cox and Mann, 2008). All other parameters were same as used in the  
324 Protein Pilot software.

### 325 ***2.17 Competitive Aptamer Linked Immobilized Sorbent Assay (ALISA)***

326 To substantiate the mass spectrometry data and to map the possible binding target, a competitive  
327 ALISA was performed. In this assay, 5' biotin- labeled aptamers (B6 and B8) were challenged  
328 with a range (50-6400 pmol) of previously reported  $\beta$ -bungarotoxin specific unlabelled aptamer  
329 ( $\beta$ B-1; 5'-GTTTTCCCCTTGTCGCTTTTGGTTCGTTCTGCCTCTATCT-3') in an ALISA (Ye  
330 et al., 2014). The reverse competitive ALISA was also performed with a range (50-6400 pmol)  
331 of unlabelled B6 and B8 aptamers challenged with 5' biotin-labeled  $\beta$ B-1 aptamer. The binding  
332 of aptamers was quantified in terms of % binding considering unchallenged aptamer OD<sub>450</sub> value  
333 as 100% binding.

## 334 **3. Results**

### 335 3.1. Generation of Aptamers

336 As described in **Figure 1a**, a NCM based subtractive SELEX approach was used to develop  
337 aptamers against the *B. caeruleus* venom. A total of 8 rounds of SELEX were performed. After  
338 eight iterative rounds of selection, the binding of the SELEX representative population from  
339 various rounds (2, 4, 6 and 8) was assessed for their binding with *B. caeruleus* venom. It is  
340 evident from **Figure 1b** that the aptamer population from round 6 (R6) displayed the highest  
341 binding, having approximately five times higher binding than the initial library mix. Thus, the  
342 pool of R6 was further amplified and was subjected to NGS. Based on the NGS data, ten unique  
343 aptamer sequences (**Table S1**) with the highest copy numbers were selected for further study.


344 The top ten aptamers were evaluated using ALISA for their ability to bind with the venom of *B.*  
345 *caeruleus*. The binding aptitude was experimentally recorded in terms of  $\Delta OD_{450}$  ( $OD_{450}$  of the  
346 test set- $OD_{450}$  of antigen control). Higher  $\Delta OD_{450}$  reflects the higher binding of DNA aptamers  
347 for the venom of *B. caeruleus*. All the selected aptamers displayed impressive binding towards  
348 the *B. caeruleus* venom (**Figure 1c**).

349

350

351

352




353 **Figure 1: (a) A schematic representation of the SELEX strategy.** (1) Counter-selection: DNA  
 354 libraries (DLs; G-quadruplex biased and unbiased) were incubated with the nitrocellulose  
 355 membrane preimmobilized with *Naja naja* (N), *Daboia russelii* (D), and *Echis carinatus* (E)  
 356 venoms (NDE) to remove any sequences with the binding tendency to these, (2) Positive  
 357 selection: Unbound sequences are then incubated with *Bungarus caeruleus* venom (B) to select  
 358 sequences with an affinity towards it; (3) Elution of binders; (4) Amplification: PCR products  
 359 were made single-stranded and used for the next round of SELEX. The whole process of counter  
 360 selection and the selection is repeated eight times. (5) Sequencing: Finally, the aptamer pool  
 361 from each round with the highest affinity was subjected to NGS to identify aptamers sequences.  
 362 **(b)** The binding propensity of the mixture of DNA libraries in comparison to various round  
 363 sequence pools. **(c)** The relative binding aptitude of the developed aptamers with the *B.*  
 364 *caeruleus* venom. All the developed aptamers depicted good binding ability with the *B.*  
 365 *caeruleus* venom. Here higher  $\Delta OD_{450}$  reflects the higher binding propensity of the aptamers.

366 Bars represent mean  $\pm$  SD. To compare the binding one-way ANOVA with multiple comparison  
 367 was applied. \*\*\*\* represent statistical significance at p value (\*p < 0.01, \*\*p < 0.001, \*\*\*p <  
 368 0.0001, \*\*\*\*p < 0.0001).

### 369 3.2. Assessment of selectivity of aptamers


370 The selectivity of the developed aptamers was evaluated with ALISA. For this, the aptamers  
 371 were checked with the venom of Common krait (*Bungarus caeruleus*), Indian cobra (*Naja naja*),  
 372 Russell's viper (*Daboia russelii*), and Saw-scaled viper (*Echis carinatus*). Except for B1, B2, B3  
 373 and B10, most of the developed aptamers displayed excellent selectivity for *B. caeruleus* venom  
 374 (Figure 2a-j). The B2 and B3 aptamers displayed the highest level of cross-reactivity against *N.*  
 375 *naja* and *D. russelii* venom while B1 and B10 evinced high cross-reactivity with *D. russelii*  
 376 venom but marginal cross-reactivity was observed for *N. naja* venom as well. Owing to the high  
 377 cross-reactivity these four aptamers (B1, B2, B3, and B10) were not considered for further  
 378 studies.



386 **Figure 2(a-j):** Relative binding of 10 SELEX derived aptamers against the venom of 'Big Four'.  
 387 All the developed aptamers largely displayed selectivity against *B. caeruleus* venom. To

388 compare the binding one-way ANOVA with multiple comparison was applied. \*\*\*\* represent  
 389 statistical significance at p value (\*p < 0.01, \*\*p < 0.001, \*\*\*p < 0.0001, \*\*\*\*p < 0.0001) Bars  
 390 represent mean  $\pm$  SD

391 Based on this data, two aptamer candidates (B6 and B8) were then assessed for their ability to  
 392 bind a range of snake venom along with red scorpion venom. For this study, the aptamers were  
 393 subjected to the venom of *B. caeruleus*, *N. naja*, *N. kaouthia*, *N. oxiana*, *D. russelii*, *B. fasciatus*,  
 394 *B. niger*, *E. carinatus* and *Hottentotta tamulus* (*H. tamulus*, Red scorpion). Both the tested  
 395 aptamers (B6 and B8) displayed the highest binding and selectivity towards the *B. caeruleus*  
 396 venom (**Figure 3a-3b**). However, marginal cross-reactivity was observed with the venom of *B.*  
 397 *fasciatus* which is another krait species



406 **Figure 3(a-b):** The relative binding ( $\Delta OD_{450}$ ) of 2 best-performing aptamers with the venom of  
 407 *N. naja*, *N. kaouthia*, *N. oxiana*, *D. russelii*, *B. caeruleus*, *B. fasciatus*, *B. niger*, *E. carinatus* and  
 408 Indian Red Scorpion (*H. tamulus*). Both the aptamers, B6 and B8 displayed the highest level of

409 binding and selectivity. To compare the binding one-way ANOVA with multiple comparison  
410 was applied. \*\*\*\* represent statistical significance at p value (\*p < 0.01, \*\*p < 0.001, \*\*\*p <  
411 0.0001, \*\*\*\*p < 0.0001) Bars represent mean  $\pm$  SD.

### 412 **3.3. Secondary structure of the aptamers**

413 The secondary structures of the B6 and B8 aptamers were first predicted with a NUPACK web  
414 server (<http://www.nupack.org/>). The NUPACK predicted structure of B6 and B8 displayed  
415 stem-loop-like structures. B6 aptamer evinced two stem-loop-like motifs while a single stem-  
416 loop was observed in the case of B8 aptamer (**Figure 4a and 4b**) with minimum free energy  
417 (MFE) of -3.89 and -0.76 kcal/mol for B6 and B8 aptamers respectively. These *in-silico*  
418 predictions were further validated with CD studies. The CD studies revealed the presence of  
419 negative peaks at ~257 nm and ~256 nm, and positive peaks at ~302 nm and ~292 nm for B6 and  
420 B8 respectively (**Figure 4c and 4d**). These peaks confirm the presence of stem-loop structure  
421 (Kypr et al., 2009; Sharma et al., 2017).

### 422 **3.4. Truncation of aptamers**

423 It is evident from the literature that all nucleotides present in a particular aptamer do not take part  
424 in its interaction with the cognate target and only a specific portion do. However, the flanking  
425 region may provide the overall stability to the structured portion of the aptamer thus truncation  
426 of such sequences may diminish the aptamer binding (Dhiman et al., 2018; Hasegawa et al.,  
427 2016). To elucidate the structure-activity relationship, two NUPACK web derived secondary  
428 structure-guided truncated variants of B6 (13-mer B6-T1 and 10-mer B6-T2) and B8 aptamer  
429 (10-mer, B8-T1) were designed and synthesized with 5' biotin label (**Table S2**). When  
430 comparing the binding of these truncated variants with their respective parent aptamer,

431 considering parent aptamer binding as 100%, it is evident (**Figure 4e and 4f**) that truncated  
432 variants of B6 (13-mer B6-T1 and 10-mer B6-T2) evinced ~90% reduction in binding while 10-  
433 mer B8-T1, a truncated variant of B8, lost around 75% binding. This data suggests the  
434 contribution of the truncated portion of aptamers in their binding to the venom of *B. caeruleus*.  
435 Therefore, for subsequent experiments, the parent aptamers were used.

436

437

438

439

440

441

442

443

444


445

446

447

448

449





450 **Figure 4: The secondary structure of different aptamers.** The secondary structure is predicted  
451 by NUPACK for aptamer **a)** B6 and **b)** B8. Both the aptamers have a stem and loop structure.  
452 The CD spectrum of the **c)** B6 and **d)** B8 aptamer, **(e-f)** Comparison of binding of truncated  
453 aptamers with their respective parent aptamers. To compare the binding one-way ANOVA with  
454 multiple comparison was applied. \*\*\*\* represent statistical significance at p value (\*p < 0.01,  
455 \*\*p < 0.001, \*\*\*p < 0.0001, \*\*\*\*p < 0.0001) Bars represent mean  $\pm$  SD.

456


### 457 **3.5. Determination of apparent dissociation constant ( $K_d$ ) of aptamer candidates**

458 The affinity of the developed B6 and B8 aptamers were measured in terms of the dissociation  
459 constant ( $K_d$ ) of the aptamers.  $K_d$  was determined by exposing B6 and B8 aptamers at  
460 concentrations ranging from 2-500 nM to a constant amount of *B. caeruleus* venom (**Figure 5a**).  
461 The study reveals the respective  $K_d$  value of B6 and B8 as 148.6 nM and 19.87 nM. The  $K_d$   
462 values suggest that the aptamer B8 is ~7.4-fold more affine than B6 aptamer. By combining B6  
463 and B8 aptamers in equimolar concentrations ranging from 2-500 nM, the affinity of the two  
464 aptamers was assessed in combination. It is evident from **Figure S1a** that the affinity of B6 and  
465 B8 combination is ~2 fold better than the B6 aptamer alone. This data also suggests that the B8  
466 aptamer, which has a higher affinity than the B6 aptamer alone and the B6 and B8 combination,  
467 is a substantial contributor to the improved affinity.

### 468 **3.6. Limit of detection (LOD) of B6 and B8 aptamers**

469 The selected B6 and B8 aptamers were evaluated for their ability to detect the lowest possible  
470 amount of *B. caeruleus* venom in an ALISA. Aptamer response was plotted as a function of  
471 venom amount. It is evident from **Figure 5b**, that the aptamer response increased from 2 to 64 ng

472 venom. However, this response achieves the saturation plateau at 125 ng, and beyond that, not  
 473 much increase in aptamer response was observed even after increasing the amount of venom.  
 474 This results show that these aptamers can detect as low as 2 ng venom (equivalent to 20 pg/ $\mu$ L)  
 475 of *B. caeruleus*. However, B8 evinced a better signal intensity in comparison to B6 aptamer. In  
 476 addition to this, we have also determined the LOD of B6 and B8 in combination. **Figure S1b**  
 477 indicates that the OD value of B6 and B8 combination was marginally increased only at higher  
 478 venom amount (125-1000 ng) in comparison to B6 alone. Overall, no significant improvement  
 479 was observed in terms of LOD at lower venom amount (2-64 ng) when combination of B6 and  
 480 B8 aptamers were used in comparison to B6 and B8 alone.



486 **Figure 5: a)** The Apparent dissociation constant curve derived through non-linear regression  
 487 representing binding affinity ( $K_d$ ) of B6 and B8 aptamers for *B. caeruleus* venom **b)** The  
 488 response of the aptamers when subjected to the various venom amount. Both the aptamers were  
 489 able to detect up to 20 pg/ $\mu$ L of *B. caeruleus* venom. Limit of detection of **(c)** B6 and B8  
 490 aptamers in human serum spiked with *B. caeruleus* venom. Both the aptamers were able to detect  
 491 up to 20 pg/ $\mu$ L of *B. caeruleus* venom.

### 492 3.7. Evaluation of aptamer performance in serum background

493 As the current study also aims to develop aptamers for diagnostics and epidemiological  
494 applications for krait envenomation, the ability of aptamers to detect *B. caeruleus* venom in  
495 human serum background was also assessed. Both B6 and B8 displayed impressive binding with  
496 the *B. caeruleus* venom in human serum. The aptamers showed minimal interference by the  
497 serum components, marking their potential to be used in clinical settings. The aptamers charted a  
498 LOD of ~2 ng ( equivalent to ~20 pg/ $\mu$ L) for both B6 and B8 (**Figure 5c**).

### 499 **3.8. Evaluation of the effect of venom variability in terms of geographical variation on the** 500 **aptamers binding**

501 The aptamers were also assessed for their ability to detect the *B. caeruleus* venom from different  
502 geographical locations to evaluate the effect of geographical variations (Casewell et al., 2020;  
503 Kalita and Mukherjee, 2019; Senji Laxme et al., 2021). The aptamers displayed differential  
504 binding with the venom from different geographical locations (three different regions of India;  
505 Central, Eastern and Southern). The binding pattern of both the aptamers was also varied for  
506 each geographical variation of venom. The aptamer B8 displayed a better binding propensity  
507 with each geographic variety of the venom than B6, which is in agreement with the earlier results  
508 in this study. Interestingly when a mixture of both the aptamers was gauged for their binding  
509 aptitude towards the venom obtained from various geographical regions, its binding response  
510 was marginally improved from the B6 alone suggesting the contribution of B8 aptamer in  
511 improvement of binding to venom (**Figure S2**).


512 To evaluate the effect of venom collection time and geography we have assessed the aptamer  
513 binding against 5 different venom samples of *B. caeruleus*. These samples were collected  
514 between years 2017-2021 at Hooghly, West Bengal, India, K.V. Institute Uttar Pradesh, India  
515 and Irula Snake Catchers Industrial Co-operative Society (ISCICS), Vadanemmeli village,

516 Kancheepuram District, Tamil Nadu, India. ALISA data clearly demonstrate that developed  
517 aptamers were able to bind to all 5 samples and as expected, B8 was found to be superior to B6  
518 (**Figure S3**). The binding of B6 and B8 combination was superior to B6 and comparable with B8  
519 aptamer. This data indicates that higher binding of B6 and B8 combination than B6 alone is due  
520 to the contribution of B8 aptamer in improving the affinity. This data also indicates that aptamers  
521 binding, particularly B8 was minimally affected by the time of venom collection and  
522 geographical variation.

### 523 **3.9. ALISA-based detection of snake venom on a paper-based device.**

524 Next, we assessed the selectivity and sensitivity of B6 and B8 aptamers in an indirect format  
525 assay comparable to ALISA but in a simple (instrument less), fast (> 2hr), cost-effective, and  
526 direct visual paper-based device with high suitability for point-of-care testing (POCT). **Figure**  
527 **6a** shows two sets of experiments with B6 and B8 aptamers. The first two sub-set in both the  
528 images with blue color indicates the B6 and B8 aptamer interact specifically with venom. The  
529 other two sets represent venom control (VC) with no-colour i.e. negative control, indicating that  
530 both the aptamer do not show any non-specific binding with paper surface or with blocking and  
531 washing buffers. Next, we assessed the selectivity of both the aptamers in a similar fashion by  
532 immobilizing equal quantity of venoms from different snakes such as *N. naja*, *N. kaouthia*, *N.*  
533 *oxiana*, *D. russelii*, *B. fasciatus*, *B. niger*, *E. carinatus*, and the scorpion venom namely *H.*  
534 *tamulus* (the Indian Red scorpion). No blue color was observed in any of the snake and scorpion  
535 venom except *B. caeruleus* indicating high selectivity of B6 and B8 aptamers  
536 respectively (**Figure 6b**). Further to establish the sensitivity of the present assay different  
537 quantity of *B. caeruleus* venom was immobilized ranging from 1000 ng to 2 ng ("0" served as  
538 VC). It was observed that the intensity of blue color was highest at 1000 ng and visible up to 2

539 ng in both the cases B6 and B8 aptamers respectively (**Figure 6c**). These results demonstrate the  
 540 high capability of B6 and B8 aptamers for *B. caeruleus* venom detection *via* assay performed  
 541 with paper-based device. The present aptasensor offer an instrument less assay for venom  
 542 detection and one can easily see the output just by observing the change in color of the test zone.  
 543 Further, being a simple and handy device it has a high potential to be used as a POC assay for  
 544 snake venom detection.



545  
 546  
 547  
 548 **Figure 6:** Paper-based device for detection of snake venom (a) The reaction zones with B6 and  
 549 B8 showing TMB colour in the images, while no colour appears in venom control (VC) with no  
 550 venom samples, (b) the selectivity test on paper-based devices indicates that both B6 and B8 do  
 551 not have any cross-reactivity with other venom samples. Analytical sensitivity of paper-based  
 552 device for venom detection (c) in crude venom and (d) in Human serum background.

553

### 554 **3.10. Determining limit of Venom detection in Human Serum on paper-based device**

555 The sensitivity of venom detection in human serum background was established using a paper-  
556 based device. A range of *B. caeruleus* venom (1000-2 ng) spiked into human serum background  
557 (obtained from a healthy individual) was evaluated to assess the effect of the clinical sample  
558 matrix on sensor sensitivity. The result indicated that aptamers B6 and B8 were able to detect as  
559 low as 2 ng of venom in serum (**Figure 6d**) and also show a significant difference when  
560 compared to the serum sample without antigen designated as “0” in **figure 6d**. Notably, the  
561 aptasensor exhibited a similar low-end detection limit of 2 ng venom in both the cases *i.e.*, in  
562 coating solution as well as in human serum background. Overall, these results exhibited the high  
563 selectivity and sensitivity of B6 and B8 aptamers against *B. caeruleus* venom with paper-based  
564 devices. However, in relative terms B8 evinced better color intensity on paper in comparison to  
565 its B6 counterpart.

566

### 567 **3.11. LC-MS/MS analysis**

568 The component of the venom binding with the B6 and B8 aptamers was first identified with  
569 western blot, developed by interacting the best performing aptamer candidates (B6 and B8) with  
570 the resolved protein components on the NCM (**Figure S4**). Results suggest that the developed  
571 aptamers bind to a venom component with a molecular mass of around ~ 55 kd (**Figure S4**).  
572 Since, B8 was emerged as a best performing aptamer candidate having higher binding and  
573 affinity in comparison to B6 thus it was subjected to detailed MS analysis. To investigate the  
574 venom component binding with the B8 aptamer, the aforementioned protein was excised from

575 the gel and were subjected to LC-MS/MS analysis. LC-MS/MS analysis identified Basic  
576 phospholipase A2  $\beta$ -bungarotoxin A2 chain by using ProteinPilot and MaxQuant search engines.  
577 Both software identified the peptides TAALCFGDSEYIGAHK, TIICYGAAGTCGR &  
578 TWGHYADYGCYCGAGGSGTPVDALDR, which correspond to  $\beta$ -bungarotoxin A2 (**Figure**  
579 **S5**). This data suggests the target of the aptamer is  $\beta$ -bungarotoxin, however, B8 also binds with  
580 a different unique epitope as reported above.

### 581 **3.12. Competitive Aptamer Linked Immobilized Sorbent Assay.**

582 To substantiate the MS findings we have performed competitive ALISA where previously  
583 reported aptamer,  $\beta$ B-1 was challenged with B6 and B8 aptamers (aptamers developed in the  
584 current study) and vice versa. It is important to note that  $\beta$ -bungarotoxin of *B. multicinctus* and  
585 *B. caeruleus* is highly similar (~99% similarity, **Figure S6**). As expected, binding of  $\beta$ B-1  
586 aptamer was reduced when it was challenged with B6 and B8 aptamers in a competitive ALISA  
587 (**Figure S7**). Interestingly when order of the aptamers was reversed i.e. when B6 and B8  
588 aptamers were challenged with  $\beta$ B-1 aptamer no significant reduction in B8 aptamer binding was  
589 observed. However, on the other hand, binding of B6 aptamer was reduced. This data suggests  
590 that B8 aptamer has better affinity than the the previously reported  $\beta$ B-1 aptamer (Ye et al.,  
591 2014).

592 Further, these observations are strongly supported by the  $K_d$  data of B6 and B8 aptamers (**Figure**  
593 **S7**). As  $K_d$  value of B8 aptamer ( $K_d \sim 19.87$  nM) was superior than the previously reported  $\beta$ B-1  
594 aptamer ( $K_d \sim 65.9$  nM) of *B. multicinctus* thus, it was able to displace the  $\beta$ B-1 aptamer in  
595 competitive ALISA format while  $\beta$ B-1 was unable to do so due to its comparatively lower  
596 affinity than B8 aptamer. On the other hand, as affinity of B6 is slightly lower (~ 2 fold lower  
597 than  $\beta$ B-1 aptamer) thus it was able to displace  $\beta$ B-1 only at the higher concentrations while B6

598 was displaced by  $\beta$ B-1 even at the lower concentrations owing to the comparatively higher  
599 affinity of  $\beta$ B-1. This data clearly indicates that cognate target of B6 and B8 aptamers is  $\beta$ -  
600 bungarotoxin.

#### 601 **4. Discussion**

602 We herein report the success of the employed NCM-assisted SELEX for a complex target i.e.,  
603 crude venom of *B. caeruleus*. The feat of the SELEX can be simply gauged by the observed high  
604 binding propensity of the developed aptamers with *B. caeruleus* venom. The pool of various  
605 SELEX rounds showed an incremental binding until R6. It then showed a loss of binding  
606 propensity during the 8<sup>th</sup> round of SELEX, which can be attributed to the loss of binders,  
607 reduction in the aptamer pool complexity, or a combination of both, during the SELEX (Wang et  
608 al., 2019). Based on NGS data, the ten aptamer candidates showing the highest multiplicity (thus  
609 enrichment during SELEX) were assessed for their binding against *B. caeruleus* venom. These  
610 ten SELEX derived aptamer candidates were having varied length. We have used PCR master  
611 mix containing Taq DNA polymerase, having a high error rate ( $2.2 \times 10^{-5}$  per nt per cycle) to  
612 promote diversity of the aptamer library during amplification cycle. Some SELEX derived  
613 aptamers are slightly larger than the original length of the library (central random region of  
614 library). This is possibly because of the extendase activity of DNA polymerase (hu, 1993).  
615 Slightly smaller length aptamer possibly generated because of the mutation introducing nature of  
616 polymerase used or mispriming event or polymerase pausing that could happen during PCR  
617 (Kalra et al., 2018; Li et al., 1990; Ranu, 1994; Westberg et al., 1999).

618 All the developed aptamers displayed a high binding aptitude towards the *B. caeruleus* venom. In  
619 terms of selectivity, though, there was variation in terms of binding tendencies towards the  
620 venom of other tested species. The observed cross-recognition by certain aptamers can be



621 attributed to similarity in the structure of epitopes recognized by them (CHADWICK, 2008;  
622 Dasgupta, 2019). Conversely, the selectivity can be explained by the complete absence or sparse  
623 presence of the same or similar epitope in the venom of other species (CHADWICK, 2008;  
624 Dasgupta, 2019). As the composition of the venom varies with snake species, the complete  
625 absence or sparsity of such epitopes seems normal (Choudhury et al., 2017; Gutiérrez et al.,  
626 2017; Warrell, 2012, 2010). Nevertheless, aptamer B6 and B8 displayed the highest performance  
627 among the initially selected ten aptamers when considered both binding proclivity and selectivity  
628 towards the *B. caeruleus* venom. These aptamers were also able to recognize *B. caeruleus* venom  
629 in a human serum background. This underlines the potential of the developed aptamer to be used  
630 as a tool to detect the presence of snake venom. Further development of aptamer-based  
631 diagnostics can obviate many limitations with the current antibody-based assays, notably, the  
632 difficulty in generating antibodies against low immunogenic but highly toxic components of the  
633 venom (Berm et al., 2018; Warrell, 2010). The current practice of developing antibodies in  
634 mice, chicken-egg yolk, or equines largely produces antibodies against large molecular weight  
635 antigens but fails to produce quality antibodies against small molecular weight neurotoxins  
636 (Pereira et al., 2019). However, aptamer development does not depend on the immunogenicity of  
637 the target, and as shown in the present study they can easily be generated against any target  
638 regardless of its molecular weight and structural complexity and may offer a reliable diagnostic  
639 tool (Chopra et al., 2014; Dhiman et al., 2017; Liu et al., 2018; Parashar, 2016; Sharma et al.,  
640 2017; Toh et al., 2014) The aptamers developed here are also important from the point of view  
641 that this is the first time a SELEX has been performed on crude snake venom, rather than  
642 purified components of the venom (Ye et al., 2014).

643 Discovery and screening of biomarkers with aptamers have been gaining attention in recent  
644 times (Berezovski et al., 2008; Gold et al., 2010; Jin et al., 2016; Kim et al., 2009; Ma et al.,  
645 2019; Ostroff et al., 2010; Ulrich and Wrenger, 2009), owing to their obvious advantages over  
646 antibodies (Gold et al., 2010; Kaur et al., 2018). One specific advantage is that once developed  
647 against a specific protein (potential biomarker), aptamers can be used in high throughput protein  
648 profiling, which then can be used to profile biomarkers or disease identification (Gold et al.,  
649 2010). We, herein also identified the target of the best performing aptamer. The mass study  
650 revealed the target as basic phospholipase A2  $\beta$ -bungarotoxin A2 chain, a biomarker common in  
651 krait venoms (Oh et al., 2017). To ascertain the finding of mass study, a competitive ALISA was  
652 performed using previously reported  $\beta$ -bungarotoxin binding aptamer ( $\beta$ B-1). Aptamer affinity  
653 data in combination with competitive ALISA further attest the  $\beta$ -bungarotoxin as a cognate  
654 target of best performing aptamer candidates generated in the current study.

655 Another interesting finding of this study is that the developed aptamers were specific to the  
656 venom of *B. caeruleus* and did not evince any significant cross-reactivity with the venom of the  
657 congeneric *B. niger* and *B. fasciatus*, despite  $\beta$ -bungarotoxin being an integral component of  
658 their venom (Silva et al., 2016). However, the composition of the  $\beta$ -bungarotoxin itself varies  
659 with the species (Khow et al., 2002; Setiyawan, 2013; Yanoshita et al., 2006) and with the  
660 geographical variation (Hia et al., 2020; Rusdi et al., 2019) within the same species. As revealed  
661 with mass study, the aptamer binds with the basic phospholipase A2 beta-bungarotoxin A2  
662 chain, which might be missing from the venom of *B. niger* and *B. fasciatus*, or simply they have  
663 subtle changes in the aminoacid sequences giving rise to different epitopes (Oh et al., 2017;  
664 Rusmili et al., 2014). Similarly, the differential binding of the B6 and B8 aptamer with the  
665 venom from varied geographical sources can also be explained, as geographical variations affect

666 the venom composition via a variety of mechanisms including but not limited to seasonal  
667 variation, diet, habitat, age-dependent change, and sexual dimorphism (Chippaux et al., 1991a;  
668 Kalita et al., 2018). This might lead to the reduced amount of  $\beta$ -bungarotoxin present in the  
669 crude venom or minor variations in amino acid sequences leading to the generation of different  
670 epitopes or altered structure of the existing epitopes (Oh et al., 2017). Taken together, this study  
671 has generated a panel of aptamers that can provide a constant source of uniform-quality reagent  
672 for the detection of snake envenomation for clinical diagnostics and epidemiological purposes.  
673 One potential application where these aptamers can be utilized is the forensic investigation to  
674 ascertain the cause of envenomation or death. For example, the Government of Andhra Pradesh,  
675 India recently launched a scheme called “Apathbandhu” to provide monetary benefits to  
676 dependants of snakebite victims who die as a result of envenomation. The aptamers developed in  
677 this study can potentially be used as a tool to in autopsy specimens to confirm the real cause of  
678 death and to prevent false claims (Brunda et al, 2006). In addition, compared to their  
679 conventional counterpart, ALISA experiments performed using the aptamers with paper-based  
680 devices shows promising results. As they can be accomplished with minimal use of sample and  
681 reagents, are less time consuming and have shown equal, if not better, sensitivity and selectivity,  
682 they will be of particular benefit in resource-constrained settings.

683 One limitation of this study is that due to the lack of availability of the venom of other krait  
684 species (*B. sindanus* and *B. walli* which are likely to be co-distributed with *B. caeruleus* in large  
685 parts of western, central and eastern regions of India) performance of B6 and B8 aptamers were  
686 not evaluated against these venoms. However, we are planning to evaluate the performance of  
687 our aptamers against the aforementioned venoms in near future as soon as we have access to  
688 these venoms.

## 689 **5. Conclusions**

690 In conclusion, we have successfully performed SELEX on a complex target, i.e., crude venom,  
691 and developed highly affine aptamers, stringently selective against its target, *B. caeruleus*  
692 venom. The developed aptamers were specific against *B. caeruleus* venom, and do not evince  
693 any cross-reactivity with the venom of other snake and scorpion species tested. The best  
694 performing aptamer candidate, B8, shows high binding in a serum background as well,  
695 demonstrating its potential to be used as a diagnostic tool in clinical, epidemiological studies,  
696 and forensic investigation. The success of this study and the demonstration of successful transfer  
697 of conventional ALISA method to a paper-based ALISA paves the way for similar studies to  
698 identify highly selective aptamers using complex targets, like crude venom. Further, an improve  
699 paper-based point-of-care (POC) device for the quantitative and qualitative detection of venom  
700 would allow for more affordable, rapid venom identification and better treatment modalities.

### 701 **Supplementary Material**

702 Supplementary material is available online.

### 703 **LC-MS/MS data availability**

704 The raw data derived from this study is available from two public data repositories. The MS/MS  
705 data described in this study is freely available from Massive  
706 (<ftp://massive.ucsd.edu/MSV000085338/>) and Proteome Xchange database by using the  
707 following ID PXD019262.

### 708 **Acknowledgments**

709 A.A. is thankful to the Department of Biotechnology (DBT), Govt. of India and Biotechnology  
710 Industrial Research Assistance Council (BIRAC) for Junior Research Fellowship and Technical  
711 Assistant respectively. TKS acknowledge the funding support through DBT-Innovative Young

712 Biotechnologist Award (IYBA) grant (BT/010/IYBA/2016/10) and BIRAC-PACE grant  
713 (BT/AIR0340/PACE-13/17). B.C. is grateful to BIRAC, Govt. of India for the post-doctoral  
714 position through BIRAC-PACE grant to T.K.S.; The aptamer sequences mentioned in the  
715 manuscript are deposited to Indian Patent Office (Patent No.201911036350).

#### 716 **Funding**

717 This work was supported by the THSTI Core grant and Department of Biotechnology, Govt. of  
718 India for Innovative Young Biotechnologist Award (IYBA), grant number  
719 (BT/010/IYBA/2016/10) and BIRAC-PACE grant (BT/AIR0340/PACE-13/17). Venoms used in  
720 this study were collected with funds provided by European Union Seventh Framework  
721 Programme (grant agreement number PIRSES-GA-2013-612131) to BITES consortium led by  
722 AM, and a small grant from The Rufford Foundation to VS.

#### 723 **Conflict of Interest**

724 The authors declare no conflict of interest.

#### 725 **References**

- 726 Ariaratnam, C.A., Rezvi Sheriff, M.H., Arambepola, C., Theakston, R.D.G., Warrell, D.A.,  
727 2009. *Am. J. Trop. Med. Hyg.* 81, 725–731.
- 728 Berezovski, M. V, Lechmann, M., Musheev, M.U., Mak, T.W., Krylov, S.N., 2008. *J. Am.*  
729 *Chem. Soc.* 130, 9137–9143.
- 730 Berm, E., Fuglsang-madsen, A., Føns, S., Lomonte, B., Laustsen, A.H., 2018 1–37.
- 731 Bonin-Debs, A.L., Boche, I., Gille, H., Brinkmann, U., 2004. *Expert Opin. Biol. Ther.* 4, 551–  
732 558.
- 733 Brunda et al, 2006. *JOURNAL OF AOAC Int.* 89, 1360–1366.
- 734 Casewell, N.R., Jackson, T.N.W., Laustsen, A.H., Sunagar, K., 2020. *Trends Pharmacol. Sci.* 41,

- 735 570–581.
- 736 CHADWICK, S.J., 2008, in: *Managing the Allergic Patient*. Elsevier, pp. 19–72.
- 737 Chatterjee, Bandhan, Jyoti Das, S., Anand, A., Kumar Sharma, T., 2020a. *Mater. Sci. Energy*  
738 *Technol.* 3, 127–35.
- 739 Chatterjee, B., Kalyani, N., Anand, A., Khan, E., Das, S., Bansal, V., Kumar, A., Sharma, T.K.,  
740 2020. *Microchim. Acta* 187.
- 741 Chippaux, J.P., Williams, V., White, J., 1991a. *Toxicon* 29, 1279–1303.
- 742 Chopra, A., Shukla, R., Sharma, T.K., 2014. *Aptamers Synth. Antibodies* 1, 1–11.
- 743 Choudhury, M., McCleary, R.J.R., Keshewani, M., Kini, R.M., Velmurugan, D., 2017. *Toxicon*  
744 135, 33–42.
- 745 Dasgupta, A., 2019. *Biot. Other Interf. Immunoassays* 1–15.
- 746 David A Warrell, 1999WHO/SEARO guidelines for the clinical management of snake bites in  
747 the Southeast Asian region. *Southeast Asian J. Trop. Med. Public Health*.
- 748 Dhiman, A., Haldar, S., Kumar, S., Sharma, N., Bansal, A., Ahmad, Y., Kumar, A., Kumar, T.,  
749 Sivaswami, J., 2018. *Tuberculosis* 112, 27–36.
- 750 Dhiman, A., Kalra, P., Bansal, V., Bruno, J.G., Sharma, T.K., Kumar, T., 2017. *Sensors*  
751 *Actuators B. Chem.* 246, 535–553.
- 752 Fernandes, I., Lima, E.X., Takehara, H.A., Moura-Da-Silva, A.M., Tanjoni, I., Gutiérrez, J.M.,  
753 2000. *Toxicon* 38, 633–644.
- 754 Goel, R., Murthy, K.R., Srikanth, S.M., Pinto, S.M., Bhattacharjee, M., Kelkar, D.S.,  
755 Madugundu, A.K., Dey, G., Mohan, S.S., Krishna, V., Prasad, T.S.K., Chakravarti, S.,  
756 Harsha, H.C., Pandey, A., 2013. *Clin. Proteomics* 10, 1.
- 757 Gold, L., Ayers, D., Bertino, J., Bock, C., Bock, A., Brody, E.N., Carter, J., Dalby, A.B., Eaton,

758 B.E., Fitzwater, T., Flather, D., Forbes, A., Foreman, T., Fowler, C., Gawande, B., Goss,  
759 M., Gunn, M., Gupta, S., Halladay, D., Heil, J., Heilig, J., Hicke, B., Husar, G., Janjic, N.,  
760 Jarvis, T., Jennings, S., Katilius, E., Keeney, T.R., Kim, N., Koch, T.H., Kraemer, S.,  
761 Kroiss, L., Le, N., Levine, D., Lindsey, W., Lollo, B., Mayfield, W., Mehan, M., Mehler,  
762 R., Nelson, S.K., Nelson, M., Nieuwlandt, D., Nikrad, M., Ochsner, U., Ostroff, R.M., Otis,  
763 M., Parker, T., Pietrasiewicz, S., Resnicow, D.I., Rohloff, J., Sanders, G., Sattin, S.,  
764 Schneider, D., Singer, B., Stanton, M., Sterkel, A., Stewart, A., Stratford, S., Vaught, J.D.,  
765 Vrkljan, M., Walker, J.J., Watrobka, M., Waugh, S., Weiss, A., Wilcox, S.K., Wolfson, A.,  
766 Wolk, S.K., Zhang, C., Zichi, D., 2010. PLoS One 5.

767 Gopalakrishnakone, P., Faiz, A., Fernando, R., Gnanathasan, C.A., Habib, A.G., Yang, C.C.,  
768 2015. Toxinology Clin. Toxinology Asia Pacific Africa 1–636.

769 Gutiérrez, J.M., Calvete, J.J., Habib, A.G., Harrison, R.A., Williams, D.J., Warrell, D.A., 2017.  
770 Nat. Rev. Dis. Prim. 3, 17063.

771 Hasegawa, H., Savory, N., Abe, K., Ikebukuro, K., 2016. Molecules 21, 421.

772 Hia, Y.L., Tan, K.Y., Tan, C.H., 2020. Acta Trop. 207, 105460.

773 hu, G., 1993. DNA Cell Biol. 12, 763–770.

774 Jin, C., Qiu, L., Li, J., Fu, T., Zhang, X., Tan, W., 2016. Analyst 141, 461–466.

775 Kalita, B., Mackessy, S.P., Mukherjee, A.K., 2018. Expert Rev. Proteomics 15, 837–849.

776 Kalita, B., Mukherjee, A.K., 2019. J. Proteins Proteomics 10, 149–164.

777 Kalra, P., Mishra, S.K., Kaur, S., Kumar, A., Prasad, H.K., Sharma, T.K., Tyagi, J.S., 2018. Mol.  
778 Ther. Nucleic Acid 13, 99–109.

779 Kaur, H., Bruno, J.G., Kumar, A., Sharma, T.K., 2018. Theranostics 8, 4016–4032.

780 Kaur, H., Chatterjee, B., Bruno, J.G., Sharma, T.K., 2019, in: Advances in Biochemical

- 781 Engineering/Biotechnology. Springer, Berlin, Heidelberg, pp. 1–12.
- 782 Khow, O., Chanhom, L., Omori-Satoh, T., Sitprija, V., 2002. *Toxicon* 40, 463–469.
- 783 Kim, Y., Liu, C., Tan, W., 2009. *Biomark. Med.* 3, 193–202.
- 784 Knudsen, C., Laustsen, A.H., 2018. *Trop. Med. Infect. Dis.* 3, 1–11.
- 785 Kypr, J., Kejnovská, I., Renčiuk, D., Vorlíčková, M., 2009. *Nucleic Acids Res.* 37, 1713–1725.
- 786 Laustsen, A.H., Engmark, M., Clouser, C., Timberlake, S., Vigneault, F., Gutiérrez, J.M.,  
787 Lomonte, B., 2017. *PeerJ* 2017.
- 788 Ledsgaard, L., Jenkins, T.P., Davidsen, K., Krause, K.E., Martos-Esteban, A., Engmark, M.,  
789 Andersen, M.R., Lund, O., Laustsen, A.H., 2018. *Toxins (Basel)*. 10, 1–19.
- 790 Li, H., Cui, X., Arnheim, N., 1990. *Proc. Natl. Acad. Sci. U. S. A.* 87, 4580–4584.
- 791 Liu, C.C., Yu, J.S., Wang, P.J., Hsiao, Y.C., Liu, C.H., Chen, Y.C., Lai, P.F., Hsu, C.P., Fann,  
792 W.C., Lin, C.C., 2018. *PLoS Negl. Trop. Dis.* 12, 1–23.
- 793 Ma, X., Lakshmi Priya, T., Gopinath, S.C.B., 2019. *J. Anal. Methods Chem.* 2019.
- 794 Michael, I., 2013.
- 795 Mohapatra, B., Warrell, D.A., Suraweera, W., Bhatia, P., Dhingra, N., Jotkar, R.M., Rodriguez,  
796 P.S., Mishra, K., Whitaker, R., Jha, P., 2011. *PLoS Negl. Trop. Dis.* 5, 1–8.
- 797 Murray, C.J.L., Barber, R.M., Foreman, K.J., Ozgoren, A.A., Abd-Allah, F., Abera, S.F.,  
798 Aboyans, V., Abraham, J.P., Abubakar, I., Abu-Raddad, L.J., Abu-Rmeileh, N.M., Achoki,  
799 T., Ackerman, I.N., Ademi, Z., Adou, A.K., Adsuar, J.C., Afshin, A., Agardh, E.E., Alam,  
800 S.S., Alasfoor, D., Albittar, M.I., Alegretti, M.A., Alemu, Z.A., Alfonso-Cristancho, R.,  
801 Alhabib, S., Ali, R., Alla, F., Allebeck, P., Almazroa, M.A., Alsharif, U., Alvarez, E., Alvis-  
802 Guzman, N., Amare, A.T., Ameh, E.A., Amini, H., Ammar, W., Anderson, H.R., Anderson,  
803 B.O., Antonio, C.A.T., Anwari, P., Arnlöv, J., Arsenijevic, V.S.A., Artaman, A., Asghar,



804 R.J., Assadi, R., Atkins, L.S., Avila, M.A., Awuah, B., Bachman, V.F., Badawi, A., Bahit,  
805 M.C., Balakrishnan, K., Banerjee, A., Barker-Collo, S.L., Barquera, S., Barregard, L.,  
806 Barrero, L.H., Basu, A., Basu, S., Basulaiman, M.O., Beardsley, J., Bedi, N., Beghi, E.,  
807 Bekele, T., Bell, M.L., Benjet, C., Bennett, D.A., Bensenor, I.M., Benzian, H., Bernabé, E.,  
808 Bertozzi-Villa, A., Beyene, T.J., Bhala, N., Bhalla, A., Bhutta, Z.A., Bienhoff, K., Bikbov,  
809 B., Biryukov, S., Blore, J.D., Blosser, C.D., Blyth, F.M., Bohensky, M.A., Bolliger, I.W.,  
810 Başara, B.B., Bornstein, N.M., Bose, D., Boufous, S., Bourne, R.R.A., Boyers, L.N.,  
811 Brainin, M., Brayne, C.E., Brazinova, A., Breitborde, N.J.K., Brenner, H., Briggs, A.D.,  
812 Brooks, P.M., Brown, J.C., Brugha, T.S., Buchbinder, R., Buckle, G.C., Budke, C.M.,  
813 Bulchis, A., Bulloch, A.G., Campos-Nonato, I.R., Carabin, H., Carapetis, J.R., Cárdenas,  
814 R., Carpenter, D.O., Caso, V., Castañeda-Orjuela, C.A., Castro, R.E., Catalá-López, F.,  
815 Cavalleri, F., Çavlin, A., Chadha, V.K., Chang, J.C., Charlson, F.J., Chen, H., Chen, W.,  
816 Chiang, P.P., Chimed-Ochir, O., Chowdhury, R., Christensen, H., Christophi, C.A., Cirillo,  
817 M., Coates, M.M., Coffeng, L.E., Coggeshall, M.S., Colistro, V., Colquhoun, S.M., Cooke,  
818 G.S., Cooper, C., Cooper, L.T., Coppola, L.M., Cortinovis, M., Criqui, M.H., Crump, J.A.,  
819 Cuevas-Nasu, L., Danawi, H., Dandona, L., Dandona, R., Dansereau, E., Dargan, P.I.,  
820 Davey, G., Davis, A., Davitoiu, D. V., Dayama, A., De Leo, D., Degenhardt, L., Del Pozo-  
821 Cruz, B., Dellavalle, R.P., Deribe, K., Derrett, S., Des Jarlais, D.C., Dessalegn, M.,  
822 Dharmaratne, S.D., Dherani, M.K., Diaz-Torné, C., Dicker, D., Ding, E.L., Dokova, K.,  
823 Dorsey, E.R., Driscoll, T.R., Duan, L., Duber, H.C., Ebel, B.E., Edmond, K.M., Elshrek,  
824 Y.M., Endres, M., Ermakov, S.P., Erskine, H.E., Eshrati, B., Esteghamati, A., Estep, K.,  
825 Faraon, E.J.A., Farzadfar, F., Fay, D.F., Feigin, V.L., Felson, D.T., Fereshtehnejad, S.M.,  
826 Fernandes, J.G., Ferrari, A.J., Fitzmaurice, C., Flaxman, A.D., Fleming, T.D., Foigt, N.,

827 Forouzanfar, M.H., Fowkes, F.G.R., Paleo, U.F., Franklin, R.C., Fürst, T., Gabbe, B.,  
828 Gaffikin, L., Gankpé, F.G., Geleijnse, J.M., Gessner, B.D., Gething, P., Gibney, K.B.,  
829 Giroud, M., Giussani, G., Dantes, H.G., Gona, P., González-Medina, D., Gosselin, R.A.,  
830 Gotay, C.C., Goto, A., Gouda, H.N., Graetz, N., Gugnani, H.C., Gupta, Rahul, Gupta,  
831 Rajeev, Gutiérrez, R.A., Haagsma, J., Hafezi-Nejad, N., Hagan, H., Halasa, Y.A., Hamadeh,  
832 R.R., Hamavid, H., Hammami, M., Hancock, J., Hankey, G.J., Hansen, G.M., Hao, Y.,  
833 Harb, H.L., Haro, J.M., Havmoeller, R., Hay, S.I., Hay, R.J., Heredia-Pi, I.B., Heuton, K.R.,  
834 Heydarpour, P., Higashi, H., Hajar, M., Hoek, H.W., Hoffman, H.J., Hosgood, H.D.,  
835 Hossain, M., Hotez, P.J., Hoy, D.G., Hsairi, M., Hu, G., Huang, C., Huang, J.J., Husseini,  
836 A., Huynh, C., Iannarone, M.L., Iburg, K.M., Innos, K., Inoue, M., Islami, F., Jacobsen,  
837 K.H., Jarvis, D.L., Jassal, S.K., Jee, S.H., Jeemon, P., Jensen, P.N., Jha, V., Jiang, G., Jiang,  
838 Y., Jonas, J.B., Juel, K., Kan, H., Karch, A., Karema, C.K., Karimkhani, C., Karthikeyan,  
839 G., Kassebaum, N.J., Kaul, A., Kawakami, N., Kazanjan, K., Kemp, A.H., Kengne, A.P.,  
840 Keren, A., Khader, Y.S., Khalifa, S.E.A., Khan, E.A., Khan, G., Khang, Y.H., Kieling, C.,  
841 Kim, D., Kim, S., Kim, Y., Kinfu, Y., Kinge, J.M., Kivipelto, M., Knibbs, L.D., Knudsen,  
842 A.K., Kokubo, Y., Kosen, S., Krishnaswami, S., Defo, B.K., Bicer, B.K., Kuipers, E.J.,  
843 Kulkarni, C., Kulkarni, V.S., Kumar, G.A., Kyu, H.H., Lai, T., Lalloo, R., Lallukka, T.,  
844 Lam, H., Lan, Q., Lansingh, V.C., Larsson, A., Lawrynowicz, A.E.B., Leasher, J.L., Leigh,  
845 J., Leung, R., Levitz, C.E., Li, B., Li, Yichong, Li, Yongmei, Lim, S.S., Lind, M.,  
846 Lipshultz, S.E., Liu, S., Liu, Y., Lloyd, B.K., Lofgren, K.T., Logroscino, G., Looker, K.J.,  
847 Lortet-Tieulent, J., Lotufo, P.A., Lozano, R., Lucas, R.M., Lunevicius, R., Lyons, R.A., Ma,  
848 S., Macintyre, M.F., Mackay, M.T., Majdan, M., Malekzadeh, R., Marcenes, W., Margolis,  
849 D.J., Margono, C., Marzan, M.B., Masci, J.R., Mashal, M.T., Matzopoulos, R., Mayosi,

850 B.M., Mazorodze, T.T., McGill, N.W., McGrath, J.J., McKee, M., McLain, A., Meaney,  
851 P.A., Medina, C., Mehndiratta, M.M., Mekonnen, W., Melaku, Y.A., Meltzer, M., Memish,  
852 Z.A., Mensah, G.A., Meretoja, A., Mhimbira, F.A., Micha, R., Miller, T.R., Mills, E.J.,  
853 Mitchell, P.B., Mock, C.N., Ibrahim, N.M., Mohammad, K.A., Mokdad, A.H., Mola,  
854 G.L.D., Monasta, L., Hernandez, J.C.M., Montico, M., Montine, T.J., Mooney, M.D.,  
855 Moore, A.R., Moradi-Lakeh, M., Moran, A.E., Mori, R., Moschandreas, J., Moturi, W.N.,  
856 Moyer, M.L., Mozaffarian, D., Msemburi, W.T., Mueller, U.O., Mukaigawara, M.,  
857 Mullany, E.C., Murdoch, M.E., Murray, J., Murthy, K.S., Naghavi, M., Naheed, A., Naidoo,  
858 K.S., Naldi, L., Nand, D., Nangia, V., Narayan, K.M.V., Nejjari, C., Neupane, S.P.,  
859 Newton, C.R., Ng, M., Ngalesoni, F.N., Nguyen, G., Nisar, M.I., Nolte, S., Norheim, O.F.,  
860 Norman, R.E., Norrving, B., Nyakarahuka, L., Oh, I.H., Ohkubo, T., Ohno, S.L., Olusanya,  
861 B.O., Opio, J.N., Ortblad, K., Ortiz, A., Pain, A.W., Pandian, J.D., Panelo, C.I.A.,  
862 Papachristou, C., Park, E.K., Park, J.H., Patten, S.B., Patton, G.C., Paul, V.K., Pavlin, B.I.,  
863 Pearce, N., Pereira, D.M., Perez-Padilla, R., Perez-Ruiz, F., Perico, N., Pervaiz, A.,  
864 Pesudovs, K., Peterson, C.B., Petzold, M., Phillips, M.R., Phillips, B.K., Phillips, D.E., Piel,  
865 F.B., Plass, D., Poenaru, D., Polinder, S., Pope, D., Popova, S., Poulton, R.G., Pourmalek,  
866 F., Prabhakaran, D., Prasad, N.M., Pullan, R.L., Qato, D.M., Quistberg, D.A., Rafay, A.,  
867 Rahimi, K., Rahman, S.U., Raju, M., Rana, S.M., Razavi, H., Reddy, K.S., Refaat, A.,  
868 Remuzzi, G., Resnikoff, S., Ribeiro, A.L., Richardson, L., Richardus, J.H., Roberts, D.A.,  
869 Rojas-Rueda, D., Ronfani, L., Roth, G.A., Rothenbacher, D., Rothstein, D.H., Rowley, J.T.,  
870 Roy, N., Ruhago, G.M., Saeedi, M.Y., Saha, S., Sahraian, M.A., Sampson, U.K.A.,  
871 Sanabria, J.R., Sandar, L., Santos, I.S., Satpathy, M., Sawhney, M., Scarborough, P.,  
872 Schneider, I.J., Schöttker, B., Schumacher, A.E., Schwebel, D.C., Scott, J.G., Seedat, S.,

873 Sepanlou, S.G., Serina, P.T., Servan-Mori, E.E., Shackelford, K.A., Shaheen, A., Shahraz,  
874 S., Levy, T.S., Shangguan, S., She, J., Sheikhabahaei, S., Shi, P., Shibuya, K., Shinohara, Y.,  
875 Shiri, R., Shishani, K., Shiue, I., Shrime, M.G., Sigfusdottir, I.D., Silberberg, D.H., Simard,  
876 E.P., Sindi, S., Singh, A., Singh, J.A., Singh, L., Skirbekk, V., Slepak, E.L., Sliwa, K.,  
877 Soneji, S., Søreide, K., Soshnikov, S., Sposato, L.A., Sreeramareddy, C.T., Stanaway, J.D.,  
878 Stathopoulou, V., Stein, D.J., Stein, M.B., Steiner, C., Steiner, T.J., Stevens, A., Stewart,  
879 A., Stovner, L.J., Stroumpoulis, K., Sunguya, B.F., Swaminathan, S., Swaroop, M., Sykes,  
880 B.L., Tabb, K.M., Takahashi, K., Tandon, N., Tanne, D., Tanner, M., Tavakkoli, M.,  
881 Taylor, H.R., Te Ao, B.J., Tediosi, F., Temesgen, A.M., Templin, T., Ten Have, M.,  
882 Tenkorang, E.Y., Terkawi, A.S., Thomson, B., Thorne-Lyman, A.L., Thrift, A.G., Thurston,  
883 G.D., Tillmann, T., Tonelli, M., Topouzis, F., Toyoshima, H., Traebert, J., Tran, B.X.,  
884 Trillini, M., Truelsen, T., Tsilimbaris, M., Tuzcu, E.M., Uchendu, U.S., Ukwaja, K.N.,  
885 Undurraga, E.A., Uzun, S.B., Van Brakel, W.H., Van De Vijver, S., Van Gool, C.H., Van  
886 Os, J., Vasankari, T.J., Venketasubramanian, N., Violante, F.S., Vlassov, V. V., Vollset,  
887 S.E., Wagner, G.R., Wagner, J., Waller, S.G., Wan, X., Wang, H., Wang, J., Wang, L.,  
888 Warouw, T.S., Weichenthal, S., Weiderpass, E., Weintraub, R.G., Wenzhi, W., Werdecker,  
889 A., Westerman, R., Whiteford, H.A., Wilkinson, J.D., Williams, T.N., Wolfe, C.D.,  
890 Wolock, T.M., Woolf, A.D., Wulf, S., Wurtz, B., Xu, G., Yan, L.L., Yano, Y., Ye, P.,  
891 Yentür, G.K., Yip, P., Yonemoto, N., Yoon, S.J., Younis, M.Z., Yu, C., Zaki, M.E., Zhao,  
892 Y., Zheng, Y., Zonies, D., Zou, X., Salomon, J.A., Lopez, A.D., Vos, T., 2015. *Lancet* 386,  
893 2145–2191.

894 Oh, A.M.F., Tan, C.H., Ariarane, G.C., Quraishi, N., Tan, N.H., 2017. *J. Proteomics* 164, 1–18.

895 Ostroff, R.M., Bigbee, W.L., Franklin, W., Gold, L., Mehan, M., Miller, Y.E., Pass, H.I., Rom,

- 896 W.N., Siegfried, J.M., Stewart, A., Walker, J.J., Weissfeld, J.L., Williams, S., Zichi, D.,  
897 Brody, E.N., 2010. PLoS One 5.
- 898 Parashar, A., 2016. Int. J. Bioassays 5, 4799.
- 899 Pereira, E.P.V., van Tilburg, M.F., Florean, E.O.P.T., Guedes, M.I.F., 2019. Int.  
900 Immunopharmacol.
- 901 Puzari, U., Mukherjee, A.K., 2020. Anal. Chim. Acta.
- 902 Ranganathan, S., Garg, G., 2009. Genome Med. 1, 1–7.
- 903 Ranu, R.S., 1994. Anal. Biochem.
- 904 Ray et al, 2012. J. Clin. Invest. 122, 1734–1741.
- 905 Rusdi, M., Rusmili, A., Othman, I., Asnawi, S., Abidin, Z., Yusof, A., Ratanabanangkoon, K.,  
906 Chanhome, L., Hodgson, W.C., Id, J.C., 2019. PLoS One.
- 907 Rusmili, M.R.A., Yee, T.T., Mustafa, M.R., Hodgson, W.C., Othman, I., 2014. Biochem.  
908 Pharmacol. 91, 409–416.
- 909 Sano-Martins, I.S., Fan, H.W., Castro, S.C.B., Tomy, S.C., Franca, F.O.S., Jorge, M.T.,  
910 Kamiguti, A.S., Warrell, D.A., Theakston, R.D.G., Biasg, 1994. Toxicon 32, 1045–1050.
- 911 Senji Laxme, R.R., Attarde, S., Khochare, S., Suranse, V., Martin, G., Casewell, N.R., Whitaker,  
912 R., Sunagar, K., 2021. PLoS Negl. Trop. Dis. 15, e0009150.
- 913 Setiyawan, 2013. J. Chem. Inf. Model. 53, 1689–1699.
- 914 Shaikh, I.K., Dixit, P.P., Pawade, B.S., Waykar, I.G., 2017. Toxicon 139, 66–73.
- 915 Shamah, S.M., Healy, J.M., Cload, S.T., 2008. Acc. Chem. Res. 41, 130–138.
- 916 Sharma, M., Gogoi, N., Dhananjaya, B.L., Menon, J.C., Doley, R., 2014. Toxin Rev. 33, 7–15.
- 917 Sharma, T.K., Bruno, J.G., Dhiman, A., 2017. Biotechnol. Adv. 35, 275–301.
- 918 Silva, A., Maduwage, K., Sedgwick, M., Pilapitiya, S., Weerawansa, P., Dahanayaka, N.J.,

- 919 Buckley, N.A., Johnston, C., Siribaddana, S., Isbister, G.K., 2016. *PLoS Negl. Trop. Dis.*  
920 10, 1–18.
- 921 Slagboom, J., Otvos, R.A., Cardoso, F.C., Iyer, J., Visser, J.C., van Doodewaerd, B.R.,  
922 McCleary, R.J.R., Niessen, W.M.A., Somsen, G.W., Lewis, R.J., Kini, R.M., Smit, A.B.,  
923 Casewell, N.R., Kool, J., 2018. *Toxicon*.
- 924 Stábeli, R.G., Pimenta Magalhães, L.M., Selistre-De-Araujo, H.S., Oliveira, E.B., 2005. *Toxicon*  
925 46, 308–317.
- 926 Suraweera, W., Warrell, D., Whitaker, R., Menon, G., Rodrigues, R., Fu, S.H., Begum, R., Sati,  
927 P., Piyasena, K., Bhatia, M., Brown, P., Jha, P., 2020. *Elife* 9, 1–37.
- 928 Takahashi, M., 2018. *Biochimie* 145, 63–72.
- 929 Tan, K.Y., Tan, N.H., Tan, C.H., 2018. *Sci. Rep.* 8, 1–14.
- 930 Taneja, V., Goel, M., Shankar, U., Kumar, A., Khilnani, G.C., Prasad, H.K., Prasad, G.B.K.S.,  
931 Gupta, U.D., Sharma, T.K., 2020. *ACS Comb. Sci.* 22, 656–666.
- 932 Theakston, R.D.G., Laing, G.D., 2014. *Toxins (Basel)*. 6, 1667–1695.
- 933 Toh, S.Y., Citartan, M., Gopinath, S.C.B., Tang, T.H., 2014. *Biosens. Bioelectron.* 64, 392–403.
- 934 Ulrich, H., Wrenger, C., 2009. *Cytom. Part A* 75, 727–733.
- 935 Wang, T., Chen, C., Larcher, L.M., Barrero, R.A., Veedu, R.N., 2019. *Biotechnol. Adv.* 37, 28–  
936 50.
- 937 Warrell, D.A., 2012. *Infect. Dis. Clin. North Am.* 26, 207–223.
- 938 Warrell, D.A., 2010. *Lancet* 375, 77–88.
- 939 Westberg, J., Holmberg, A., Uhlén, M., Pettersson, B., 1999. *Electrophoresis* 20, 502–510.
- 940 Yanoshita, R., Ogawa, Y., Murayama, N., Omori-Satoh, T., Saguchi, K.I., Higuchi, S., Khow,  
941 O., Chanhom, L., Samejima, Y., Sitprija, V., 2006. *Toxicon* 47, 416–424.

942 Ye, F., Zheng, Y., Wang, X., Tan, X., Zhang, T., Xin, W., Wang, Jie, Huang, Y., Fan, Q., Wang,  
943 Jinglin, 2014. PLoS One 9, 2–9.

944 Zancolli, G., Sanz, L., Calvete, J.J., Wüster, W., 2017. Toxins (Basel).

945

946

947

Journal Pre-proof

### Highlights

- A novel DNA aptamer against venom *Bungarus caeruleus* (Krait) has been identified for the first time.
- Phospholipase A2-Beta-bungarotoxin is identified as a target of aptamer using Mass spectrometry.
- Aptamer was adapted on to a paper-based device.
- Best performing aptamer can detect as low as 2ng krait venom in serum background.



**Declaration of interests**

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

The authors declare the following financial interests/personal relationships which may be considered as potential competing interests:

Journal Pre-proof