

OPEN

AIDS Journal of Acquired Immune Deficiency Syndromes Publish Ahead of Print
DOI: 10.1097/QAI.0000000000003092

Laboratory Evaluation of the VISITECT[®] Advanced Disease Semi-quantitative Point-of-care CD4 Test

Kwana Lechiile, BSc^{1,§,*}, Tshepo B Leeme, MBBS, MSc^{1,§}, Mark W Tenforde, MD, PhD², Mbabi Bapabi, BSc¹, Julita Magwenzi, BSc¹, Oitshepile Maithamako, BSc¹, Fredah Mulenga, BSc¹, Terence Mohammed, BSc¹, Julia Ngidi, BSc³, Margaret Mokomane, PhD⁴, David S Lawrence, MBChB, MSc, MRCP^{1,5}, Madisa Mine, PhD³, Joseph N Jarvis, MBBS, MSc, MRCP, PhD^{1,5}

¹Botswana Harvard AIDS Institute Partnership, Gaborone, Botswana

²Botswana-UPenn Partnership, Gaborone, Botswana

³Botswana National Health Laboratory, Gaborone, Botswana

⁴University of Botswana, Gaborone, Botswana

⁵Department of Clinical Research, Faculty of Infectious and Tropical Diseases, London School of Hygiene and Tropical Medicine, London, UK

^sThese authors contributed equally

*Corresponding author: Ms Kwana Lechiile, Botswana Harvard AIDS Institute Partnership,
Private Bag BO320, Gaborone, Botswana. Telephone Number: +2673902671 Email:
klechiile@bhp.org.bw

Conflict of Interest and Source of Funding: JNJ is funded by the National Institute for Health Research (NIHR) through a Global Health Research Professorship (RP-2017-08-ST2-012) using UK aid from the UK Government to support global health research. For the remaining authors none were declared.

This is an open-access article distributed under the terms of the Creative Commons Attribution-Non Commercial-No Derivatives License 4.0 (CCBY-NC-ND), where it is permissible to download and share the work provided it is properly cited. The work cannot be changed in any way or used commercially without permission from the journal.

Abstract

Background: Advanced HIV disease (AHD; CD4 counts <200 cells/ μ L) remains common in many low- and middle-income settings. An instrument-free point-of-care test to rapidly identify patients with AHD would facilitate implementation of the World Health Organization (WHO)

recommended package of care. We performed a laboratory-based validation study to evaluate the performance of the VISITECT[®] CD4 Advanced Disease assay in Botswana.

Setting: A laboratory validation study.

Methods: Venous blood samples from people living with HIV having baseline CD4 testing in Gaborone, Botswana, underwent routine testing using flow cytometry, followed by testing with the VISITECT[®] CD4 Advanced Disease assay by a laboratory scientist blinded to the flow cytometry result with a visual read to determine if the CD4 count was below 200 cells/ μ L. A second independent investigator conducted a visual read blinded to the results of both flow cytometry and the initial visual read. The sensitivity and specificity of the VISITECT[®] for detection of AHD were determined using flow cytometry as a reference standard, and inter-rater agreement in VISITECT[®] visual reads assessed.

Results: 1053 samples were included in the analysis. The VISITECT test correctly identified 112/119 samples as having a CD4 count <200 cells/ μ L, giving a sensitivity of 94.1% (95% confidence interval [CI] 88.3-97.6%) and specificity of 85.9% (95% CI 83.5-88.0%) compared to flow cytometry. Inter-rater agreement between the two independent readers was 97.5%, *Kappa* 0.92 ($p < 0.001$).

Conclusions: The VISITECT[®] CD4 Advanced Disease reliably identified individuals with low CD4 counts and could facilitate implementation of the WHO recommended package of interventions for AHD.

Short title: VISITECT CD4 Evaluation

Key words: HIV; AIDS; CD4 count; Point-of-care; Advanced HIV Disease

Word count: 2222 words

Introduction

CD4 cell count testing continues to have an important role in the care of people living with human immunodeficiency virus (HIV)^{1,2}. The removal of the CD4-based antiretroviral therapy (ART) initiation criteria³ and the scaling up of viral load testing in line with World Health Organization (WHO) recommendations⁴ has led to a scaling back of CD4 testing at treatment initiation and for monitoring⁵. However, CD4 count measurement remains essential for the identification of patients living with advanced HIV disease (AHD)⁶, defined as a CD4 cell count <200 cells/ μ L, and is required to determine whether individuals need screening and prophylaxis for opportunistic infections (OIs) and enhanced monitoring⁷. CD4 testing is the gateway to the WHO package of evidence-based interventions for AHD⁸, which includes cryptococcal antigen (CrAg) screening, tuberculosis (TB) screening and cotrimoxazole prophylaxis.

Recent data show that baseline CD4 count testing is declining in many high HIV-prevalence settings in Low- and Middle-Income Countries (LMICs)⁹⁻¹³ as funding and support for CD4 testing infrastructure is withdrawn¹². This reduction in baseline CD4 testing makes effective implementation of the WHO-recommended package of AHD interventions challenging; clinical staging has extremely low sensitivity for identification of AHD¹⁴, and failure to identify vulnerable individuals with low CD4 counts is likely to have negative implications for overall

HIV care in many LMICs, particularly in sub-Saharan Africa (SSA), where the burden of AHD and incidence of opportunistic infections remains high¹⁵⁻¹⁷, and are associated with ongoing HIV-related mortality¹⁸. The proportions of individuals presenting with advanced HIV disease is still reported to be in the range of 20-35% in recent studies conducted in the southern and east African region^{12,15-18}, and the overall number of individuals with very low CD4 counts has remained stable for the past decade^{17,19}.

A further challenge to effective implementation of the WHO AHD package has been the widespread uptake of rapid or same-day ART initiation²⁰. With rapid ART initiation, even if baseline CD4 tests are sent, the current turn-around times with flow-cytometry²¹ testing in centralized laboratories, which remains the mainstay of CD4 testing in most LMICs, means that ART initiation occurs in most cases prior to the availability of CD4 results. As there is a relatively short window of opportunity in which to implement key life-saving interventions such as CrAg screening²², TB screening²³, and co-trimoxazole prophylaxis²⁴, many opportunities to reduce morbidity and mortality are missed⁹.

Low-cost rapid CD4 tests that can be performed at or near the point-of-care (POC) by any healthcare worker, and accurately identify individuals with AHD, would help overcome the barriers currently posed to implementation of WHO-recommended package of AHD care by the scaling back of centralized CD4 testing and adoption of rapid ART initiation guidelines. The VISITECT[®] CD4 Advanced Disease test is an instrument-free semi-quantitative diagnostic lateral flow assay (LFA) detecting the CD4 protein on the surface of CD4+ T lymphocytes designed to be used at point of care or primary healthcare level using capillary blood or venous blood, providing a dichotomous “high” or “low” result if CD4 count is higher or lower than 200 cells/ μ L within 40 minutes (Figure 1a)⁶. We performed a laboratory-based validation study to

evaluate the performance of the VISITECT[®] CD4 Advanced Disease against the FACSCalibur and AQUIOS flow cytometry platforms in Botswana using venous blood collected from HIV-positive individuals, assessing diagnostic accuracy and inter-rater agreement in interpretation of the LFA.

Methods

Participants

Between November 2020 and April 2021, we screened sequential venous EDTA blood samples sent for CD4 testing at the Botswana-Harvard HIV Reference Laboratory (BHHRL) for study inclusion. The BHHRL performs nearly all CD4 measurements for patients attending clinics offering HIV care in greater Gaborone, with a total catchment population of approximately 300,000. We restricted our study to samples from individuals who were having “baseline” CD4 tests and had no record of previous CD4 count measurement in the electronic national HIV database dating back to 2004 using a unique laboratory identification number in order to avoid testing large numbers of CD4 samples with high CD4 counts from individuals established on ART undergoing annual CD4 monitoring. Samples from individuals under 5 years of age were excluded, as were samples that had been collected more than 24 hours prior to screening by the study team. Anonymized age and sex data were recorded for each sample.

Testing

Following routine CD4 testing at BHHRL using the FACSCalibur and AQUIOS flow cytometry instruments, patient samples were tested with VISITECT[®] CD4 Advanced Disease kit by a laboratory scientist (KL) blinded to the flow cytometry result. The EDTA blood samples were

inverted until homogenization took place prior to testing, and 30µl of sample was used for each test. Testing was batched, with a maximum of 6 tests performed at a time according to Method 1 of the manufacturer's instructions (Figure 1b). A visual read was conducted by the testing laboratory scientist (KL) to determine if the sample had a CD4 count above or below 200 cells/ul and results were recorded. A test line that appeared darker than the reference line indicated a CD4 count above range (AR), while test line that appeared the same intensity or lighter than the reference line or no test line visible indicated a CD4 count below range (BR) (Figure 1c). The absence of a control line and/or reference line indicated an invalid run and the test was repeated. For the last 4 months of the study a second independent investigator (TBL) conducted an additional visual read blinded to the results of both flow cytometry results and the initial visual read.

Data management and statistical analysis

Data from the VISITECT® CD4 Advanced Disease test were recorded by investigators in a laboratory testing log. Routine CD4 count results from the flow cytometry platforms, along with age and sex data, were subsequently extracted from the electronic medical records system and transcribed into the laboratory log. Data were then double entered into a CVS file and imported into STATA 14 (StataCorp, College Station, TX) for cleaning and analysis. In all analyses the VISITECT® visual reads of the first reader (KL) were used as the definitive VISITECT® CD4 result for consistency across all samples. The baseline characteristics of the cohort were summarised using descriptive statistics. Routine CD4 results from the flow cytometry testing were summarised according to VISITECT® CD4 result (AR or BR) and visualised using a scatter plot. We calculated sensitivity, specificity, negative and positive predictive values of the VISITECT® CD4 Advanced Disease assay against a reference standard of flow cytometry for a

CD4 count at or below or above 200 cells/ μ L. To determine interrater reliability, we calculated percent agreement between the two readers and Cohen's kappa statistic.

Ethics statement

This research project was approved by Institutional Review Boards at the Botswana Ministry of Health and Wellness (the Health Research and Development Committee) and University of Botswana. A waiver of individual consent was granted as the study used only anonymized residual laboratory samples and anonymized patient data.

Results

We screened 12,096 sequential patient samples during the study period to determine whether they were received baseline CD4 testing at the BHHRL. Of these, 1072 (8.9%) were baseline CD4 tests with no record of prior CD4 testing post-2004 in the national database. We excluded 19 samples from the analysis, nine as VISITECT[®] testing could not be conducted within 24 hours from the time of sample collection, five as no age was recorded, three as no CD4 result was available due to technical errors or the tests not being performed, and two as they were from individuals under 5 years old (Figure 2). Of the 1053 samples included in the final analysis, 654 (62.2%) were from females, and the median age of the cohort was 38 years (interquartile range [IQR] 31-46 years). One hundred and nineteen (11.3%) of the patients receiving baseline CD4 testing had advanced HIV disease (CD4 <200 cells/ μ L) based on the reference flow cytometry testing. Individuals presenting with AHD were more likely to be male than those without AHD (57% compared to 35%, $p < 0.001$), and were older with a median age of 40 years (IQR 34-46) compared to 38 years (IQR 31-46) in those without AHD ($p = 0.04$).

Diagnostic performance of the VISITECT[®] CD4 Advanced Disease assay

The VISITECT[®] CD4 Advanced Disease assay was above range in 809/1053 (76.8%) and below range in 244/1053 (23.2%) samples tested. The VISITECT test correctly identified 112/119 samples as having a CD4 count below 200 cells/ μ L, giving a sensitivity of 94.1% (95% confidence interval [CI] 88.3-97.6%) compared to flow cytometry, with an 85.9% (95% CI 83.5-88.0%) specificity (Table 1). The median CD4 count was 584 (IQR 442-765) in the above range category and 223 (IQR 147-319) in the below range category (Figure 3). In our cohort with a relatively low overall prevalence of AHD (11.3%), the positive predictive value of the VISITECT[®] test for a CD4 cell count <200 cells/ μ L was 45.9% (95% CI 39.5 -52.4%) with 112 of 244 samples classified as below range having a CD4 <200 cells/ μ L. The negative predictive value was 99.1% (95% CI 98.2-99.7%) with 802 of 809 samples classified as above range having a CD4 >200 cells/ μ L. Of the 132 samples with CD4 counts >200 cells/ μ L misclassified by the VISITECT[®] test as being below range, the majority (89/132, 67%) had CD4 counts below 350 cells/ μ L. Inter-rater agreement was assessed through paired readings of 472 samples (Table 2). Overall agreement was excellent, with 97.5% agreement between the two independent readers giving a *Kappa* statistic of 0.92 ($p < 0.001$) (Table 3).

Discussion

The VISITECT[®] CD4 Advanced Disease had excellent sensitivity and good specificity for detection of advanced HIV disease when compared to the gold standard of flow cytometry CD4 count testing on venous blood. Inter-rater agreement in the reading of results between the two independent investigators was excellent.

Comparable performance characteristics were observed in the only other reported validation study of the VISITECT[®] CD4 Advanced Disease assay, using a smaller sample of 708 venous blood samples in Malawi, Zimbabwe and Democratic Republic of Congo, which found a sensitivity of 95% and a specificity of 81.9%²⁵. This prior study also examined the performance of the VISITECT[®] assay on 433 finger-prick blood, reporting similar sensitivity and specificity results to those obtained using venous blood²⁵.

In many regions of sub-Saharan Africa, where AHD is still common despite wide roll out of ART^{7,15,20,26,27} and data show that the lack of support for CD4 infrastructure is resulting in the scaling back of centralized flow-cytometry CD4 testing⁵, the VISITECT[®] Advanced Disease test could provide an affordable and rapidly implementable means of effectively identifying individuals with AHD, enabling stratification of high risk individuals into appropriate differentiated service delivery streams. Our study implemented the VISITECT[®] Advanced Disease test in a central laboratory. Although instrument-free CD4 tests that do not require any specialized external reagents or electricity potentially have an important role in very low-resource and/or rural laboratories where flow cytometry is not easily accessible, a key benefit of the low-cost, instrument-free, VISITECT[®] POC assay is that can potentially be performed by any healthcare worker in the primary health care setting. Limited data suggest that VISITECT[®] semi-quantitative CD4 testing can be reliably performed by clinicians or lay staff at the point of care²⁸, with users generally rating the testing process easy to perform and results easy to read¹. However, the testing process requires a series of timed steps and accurate application of sample and buffers, requiring user focus and precision¹, and in the prior validation study some non-laboratory-based health-care workers reported concerns about their ability to conduct VISITECT[®] CD4 Advanced Disease testing whilst multi-tasking in their other work roles.

Further research is needed to determine how the VISITECT[®] CD4 assay can be best incorporated into CD4 testing algorithms; it is likely that differing approaches will be required depending on testing context, with POC testing by clinical or lay staff in low through put rural settings, and batched testing by dedicated lay or laboratory staff in higher volume clinics.

A further important consideration in implementation of the VISITECT[®] CD4 assay is the prevalence of AHD in the targeted population. A lower than anticipated proportion of our study sample had AHD, with 11.3% of samples tested having a CD4 cell count <200 cells/ μ L. Although the VISITECT[®] test had an excellent negative predictive value for AHD, the positive predictive value (PPV) in our study population was below 50%. The PPV would be higher in populations with higher AHD prevalence (for example, applying our observed test specificity to an equivalent population with a 25% AHD prevalence would yield a PPV of 69%), but even in this context a substantial proportion of identified as having AHD on the VISITECT[®] assay would have CD4 counts above 200 cells/ μ L. Our data demonstrate that the majority of these misclassified individuals have CD4 counts between 200 and 350 cells/ μ L; screening such individuals for opportunistic infections or stratifying them into more intensive clinical follow-up would be very unlikely to cause adverse patient outcomes but would have implications for resource use that need to be considered by program managers.

In conclusion, our study demonstrates that the VISITECT[®] CD4 Advanced Disease assay can rapidly and effectively identify individuals with AHD. This low-cost, instrument-free, POC test provides a valuable tool to facilitate the effective identification and management of individuals presenting with AHD in resource-limited settings.

Acknowledgments

Author contributions: *Conceptualization*: M.W.T., J.N.J., M.M. *Data curation*: K.L., T.B.L., M.B. *Data analysis*: T.B.L., J.N.J. *Funding*: M.W.T., J.N.J. *Project administration*: K.L., T.B.L., M.B., D.L., J.N.J. *Study conduct*: K.L., T.B.L., M.B., J.N.J. *Laboratory work and interpretation*: K.L., T.B.L., M.B. *manuscript draft*: K.L., T.B.L., *final manuscript preparation*: K.L., T.B.L., J.N.J.

We thank the Botswana Harvard Referral Reference Laboratory for allowing us to use their facility and patient samples to conduct this project, and the Ministry of Health for providing us with access to data through the Integrated Patient Management Systems. We also thank Omega Diagnostics for supplying the VISITECT[®] CD4 Advanced Disease assay for validation and supporting the costs of a laboratory technician. Omega Diagnostics had no role in the interpretation of results or in the decision to publish.

References

1. Scorgie F, Mohamed Y, Anderson D, Crowe SM, Luchters S, Chersich MF. Qualitative assessment of South African healthcare worker perspectives on an instrument-free rapid CD4 test. *BMC Health Serv Res.* 2019;19(1). doi:10.1186/s12913-019-3948-x
2. Ford N, Meintjes G, Vitoria M, Greene G, Chiller T. The evolving role of CD4 cell counts in HIV care. *Curr Opin HIV AIDS.* 2017;12(2):123-128.

doi:10.1097/COH.0000000000000348

3. Song A, Liu X, Huang X, et al. From CD4-based initiation to treating all HIV-infected adults immediately: An evidence-based meta-analysis. *Front Immunol.* 2018;9(FEB):212. doi:10.3389/fimmu.2018.00212
4. Roberts T, Cohn J, Bonner K, Hargreaves S. Scale-up of Routine Viral Load Testing in Resource-Poor Settings: Current and Future Implementation Challenges. Published online 2016. doi:10.1093/cid/ciw001
5. Pham MD, Romero L, Parnell B, Anderson DA, Crowe SM, Luchters S. Feasibility of antiretroviral treatment monitoring in the era of decentralized HIV care: a systematic review. *AIDS Res Ther.* 2017;14:3. doi:10.1186/s12981-017-0131-5
6. Omega Diagnostics Group PLC. VISITECT CD4 Advanced Disease.
7. Leeme TB, Mine M, Lechiile K, et al. Utility of CD4 count measurement in the era of universal antiretroviral therapy: an analysis of routine laboratory data in Botswana. *HIV Med.* 2021;22(1):1-10. doi:10.1111/hiv.12951
8. World Health Organisation. *HIV Treatment and Care: What's New in Treatment Monitoring: Viral Load and CD4 Testing: Information Note.*; 2017.
9. Nasuuna E, Tenforde MW, Muganzi A, Jarvis JN, Manabe YC, Kigozi J. Reduction in Baseline CD4 Count Testing following Human Immunodeficiency Virus “treat All” Adoption in Uganda. *Clin Infect Dis.* 2020;71(9):2497-2499. doi:10.1093/cid/ciaa261
10. Ngongo NM, Nani-Tuma HS, Mambimbi MM, et al. Progressive phasing out of baseline CD4+ cell count testing for people living with HIV in Kinshasa, Democratic Republic of the Congo. *AIDS.* 2021;35(5):841-843. doi:10.1097/QAD.0000000000002802
11. Renju J, Rice B, Songo J, et al. Influence of evolving HIV treatment guidance on CD4

- counts and viral load monitoring: A mixed-methods study in three African countries. *Glob Public Health*. 2021;16(2):288-304. doi:10.1080/17441692.2020.1805785
12. Zaniewski E, Dao Ostinelli CH, Chammartin F, et al. Trends in CD4 and viral load testing 2005 to 2018: multi-cohort study of people living with HIV in Southern Africa. *J Int AIDS Soc*. 2020;23(7). doi:10.1002/jia2.25546
 13. Anderegg N, Panayidou K, Abo Y, et al. Global Trends in CD4 Cell Count at the Start of Antiretroviral Therapy: Collaborative Study of Treatment Programs. *Clin Infect Dis*. 2018;66(6):893-903. doi:10.1093/cid/cix915
 14. Lebelonyane R, Mills LA, Mogorosi C, et al. Advanced HIV disease in the Botswana combination prevention project: prevalence, risk factors, and outcomes. *AIDS*. 2020;34(15):2223-2230. doi:10.1097/QAD.0000000000002627
 15. Chihana ML, Huerga H, Van Cutsem G, et al. Distribution of advanced HIV disease from three high HIV prevalence settings in Sub-Saharan Africa: a secondary analysis data from three population-based cross-sectional surveys in Eshowe (South Africa), Ndhiwa (Kenya) and Chiradzulu (Malawi). *Glob Health Action*. 2019;12(1). doi:10.1080/16549716.2019.1679472
 16. Mupfumi L, Moyo S, Shin SS, et al. High incidence of tuberculosis in the first year of antiretroviral therapy in the Botswana National antiretroviral therapy programme between 2011 and 2015. *AIDS*. 2019;33(15):2415-2422. doi:10.1097/QAD.0000000000002363
 17. Tenforde MW, Mokomane M, Leeme T, et al. Epidemiology of adult meningitis during antiretroviral therapy scale-up in southern Africa: Results from the Botswana national meningitis survey. *J Infect*. 2019;79(3):212-219. doi:10.1016/j.jinf.2019.06.013
 18. Tenforde MW, Gertz AM, Lawrence DS, et al. Mortality from HIV-associated meningitis

- in sub-Saharan Africa: a systematic review and meta-analysis. *J Int AIDS Soc.* 2020;23(1). doi:10.1002/jia2.25416
19. Lilian RR, Rees K, McIntyre JA, Struthers HE, Peters RPH. Same-day antiretroviral therapy initiation for HIV-infected adults in South Africa: Analysis of routine data. *PLoS One.* 2020;15(1). doi:10.1371/journal.pone.0227572
 20. Lebelonyane R, Bachanas P, Block L, et al. Rapid antiretroviral therapy initiation in the Botswana Combination Prevention Project: a quasi-experimental before and after study. *lancet HIV.* 2020;7(8):e545-e553. doi:10.1016/S2352-3018(20)30187-9
 21. Alvarez-Uria G, Reddy R, Reddy S, Naik PK, Midde M. Evaluation of a Low-Cost Strategy for Enumerating CD4 Lymphocyte Absolute Count and Percentage Using the FACSCalibur Flow Cytometer in HIV-Infected Patients from a Resource-Limited Setting. *ISRN AIDS.* 2012;2012:1-4. doi:10.5402/2012/494698
 22. Jarvis JN, Lawn SD, Vogt M, Bangani N, Wood R, Harrison TS. Screening for Cryptococcal Antigenemia in Patients Accessing an Antiretroviral Treatment Program in South Africa. Published online 2009:48. doi:10.1086/597262
 23. Gupta-Wright A, Corbett EL, van Oosterhout JJ, et al. Rapid urine-based screening for tuberculosis in HIV-positive patients admitted to hospital in Africa (STAMP): a pragmatic, multicentre, parallel-group, double-blind, randomised controlled trial. *Lancet (London, England).* 2018;392(10144):292-301. doi:10.1016/S0140-6736(18)31267-4
 24. Zachariah R, Harries AD, Luo C, Bachman G, Graham SM. Scaling-up co-trimoxazole prophylaxis in HIV-exposed and HIV-infected children in high HIV-prevalence countries. *Lancet Infect Dis.* 2007;7(10):686-693. doi:10.1016/S1473-3099(07)70239-6
 25. Ndlovu Id Z, Massaquoi Id L, Bangwen NE, et al. Diagnostic performance and usability

of the VISITECT CD4 semi-quantitative test for advanced HIV disease screening.

Published online 2020. doi:10.1371/journal.pone.0230453

26. Carmona S, Bor J, Nattey C, et al. Persistent High Burden of Advanced HIV Disease among Patients Seeking Care in South Africa's National HIV Program: Data from a Nationwide Laboratory Cohort. *Clin Infect Dis*. 2018;66(suppl_2):S111-S117. doi:10.1093/cid/ciy045
27. Osler M, Hilderbrand K, Goemaere E, et al. The Continuing Burden of Advanced HIV Disease Over 10 Years of Increasing Antiretroviral Therapy Coverage in South Africa. *Clin Infect Dis S118 • CID*. 2018;2018(2):66. doi:10.1093/cid/cix1140
28. Luchters S, Technau K, Mohamed Y, et al. Field Performance and Diagnostic Accuracy of a Low-Cost Instrument-Free Point-of-Care CD4 Test (Visitect CD4) Performed by Different Health Worker Cadres among Pregnant Women. Published online 2019. doi:10.1128/JCM

Table 1. Diagnostic performance of the VISITECT® versus Flow Cytometry Results for sensitivity and specificity.

Flow Cytometry Result	VISITECT®			Value [% (95% CI)]	
	<200 cells/ μ L	\geq 200 cells/ μ L	Total	Sensitivity	Specificity
<200	112	7	119	94.1 (88.3-97.6)	85.9 (83.5-88.0)
\geq 200	132	802	934		
Total	244	809	1053		

CI (Confidence Interval)

Table 2. Diagnostic performance of VISITECT® for interrater agreement

Scorer B	Scorer A		
	Below Average	Above Average	Total
Below Average	83	3	86
Above Average	9	377	386
Total	92	380	472

Table 3. Interrater agreement

Interrater reliability	% Agreement		Cohen's kappa	SE	P value
	Expected	Observed			
	69.4%	97.5%	0.9169	0.0460	<0.0001

SE: Standard Error

Figure 1. (a) Omega Diagnostics VISITECT® CD4 Advanced Disease test kit– supplied with foil pouches containing cassettes and desiccants, buffer solution, sterile retractable lancets, sampling devices, and alcohol swabs. (b) VISITECT® CD4 Technical Guidance 16 Batch Testing guide- Batch Testing Method 1 (maximum 6 samples). (c) Results interpretation of VISITECT®CD4 test




Figure 2. Participants schema




Figure 3. Distribution of CD4 cell counts obtained by flow-cytometry classified by the VISITECT® CD4 Advanced Disease test. The grey bars indicate the median CD4 count in each group.

