



ELSEVIER

Contents lists available at ScienceDirect

## Data in Brief

journal homepage: [www.elsevier.com/locate/dib](http://www.elsevier.com/locate/dib)

CrossMark

## Data Article

A data comparison between a traditional and the single-step  $\beta$ -galactosidase assay

Jorrit Schaefer, Goran Jovanovic, Ioly Kotta-Loizou, Martin Buck\*

Department of Life Sciences, Faculty of Natural Sciences, Imperial College London, London SW7 2AZ, UK

## ARTICLE INFO

## Article history:

Received 29 March 2016

Received in revised form

11 May 2016

Accepted 25 May 2016

Available online 1 June 2016

## Keywords:

LacZ

 $\beta$ -galactosidase $\beta$ -Galactosidase

B-Gal

 $\beta$ -Gal

Microplate reader

## ABSTRACT

This article describes reproducibility of a single-step automated  $\beta$ -galactosidase, and the equivalence of its data to the traditional assay ("Experiments in Molecular Genetics" [1]). This was done via a pairwise comparison of both methods using strains with Miller Unit [MU] values ranging from 0 to over 2000. The data presented in this article is associated with the research article entitled "A single-step method for mid to high throughput  $\beta$ -galactosidase assays in *Escherichia coli* using a microplate reader" [2].

© 2016 The Authors. Published by Elsevier Inc. This is an open access article under the CC BY license (<http://creativecommons.org/licenses/by/4.0/>).

## Specifications Table

Subject area	Biology
More specific subject area	Enzymatic assays
Type of data	Table
How data was acquired	FLUOstar Omega Microplate Reader (BMG LABTECH)
Data format	Raw/Analysed
Experimental factors	None applied
Experimental features	$\beta$ -galactosidase assay
Data source location	Imperial College Road, Imperial College, London SW7 2AZ, UK
Data accessibility	Data is with this article

DOI of original article: <http://dx.doi.org/10.1016/j.ab.2016.03.017>

\* Corresponding author.

E-mail address: [m.buck@imperial.ac.uk](mailto:m.buck@imperial.ac.uk) (M. Buck).

<http://dx.doi.org/10.1016/j.dib.2016.05.063>

2352-3409/© 2016 The Authors. Published by Elsevier Inc. This is an open access article under the CC BY license (<http://creativecommons.org/licenses/by/4.0/>).

## Value of the Data

- Validates a single-step automated assay to measure  $\beta$ -galactosidase activity in *E. coli* with a microplate reader.
- Establishes the method provides consistent data, with a coefficient of variation similar to that of the traditional assay.
- Student's *t*-test demonstrates equivalence of the single-step and the traditional assay data.

## 1. Data

The data presented in this article is a pairwise comparison of the MU activities of the traditional  $\beta$ -galactosidase assay and a single-step automated method [2]. 4 different strains with MU activities (0–10, 40–70, 500–900, and 1800–2400 respectively) were assayed using both methods.

## 2. Experimental design, materials and methods

### 2.1. Cell culture and sample preparation

Samples were grown overnight at 37 °C in LB media. For the traditional assay, samples were treated as described by Miller [1]. For the single-step automated microplate assay, 80  $\mu$ L of culture was transferred into 96-well Costar<sup>®</sup> flat bottom microplates (Sigma-Aldrich). 120  $\mu$ L of a custom mix was added to each well used, including the blank (containing 80  $\mu$ L LB). The custom mix consists of 60 mM Na<sub>2</sub>HPO<sub>4</sub>, 40 mM NaH<sub>2</sub>PO<sub>4</sub>, 10 mM KCl, 1 mM MgSO<sub>4</sub>, 36 mM  $\beta$ -mercaptoethanol, 166  $\mu$ L/ml T7 lysozyme, 1.1 mg/ml ONPG, 6.7% PopCulture<sup>®</sup> Reagent. The Costar microtitre plates were then placed in a FLUOstar Omega Microplate Reader (BMG LABTECH).

### 2.2. Kinetic OD<sub>600</sub> and OD<sub>420</sub> measurements

The FLUOstar Omega Microplate Reader was set to 30 °C, with OD<sub>600</sub> and OD<sub>420</sub> readings taken every 60 s for 1 h (20 flashes per well each cycle). Wells were set to shake at 500 rpm (double orbital shaking) between readings to ensure homogenisation of the sample and more efficient lysis. A script to run these parameters on a FLUOstar Omega Microplate Reader can be found in [Supplementary Data A](#).

### 2.3. MARS Data Analysis

The MARS Data Analysis software package was used to process the OD<sub>600</sub> and OD<sub>420</sub> data generated by the by FLUOstar Omega Microplate Reader. Miller Units values are expressed as follows:

$$\text{Miller Units} = \frac{1000 * (\text{OD}_{420}/\text{min})}{\text{OD}_{600} * \text{volume used (ml)}} = \frac{1000 * (\text{OD}_{420}/\text{min})}{(\text{OD}_{600} * 2.5) * 0.080} = \frac{5000 * (\text{OD}_{420}/\text{min})}{\text{OD}_{600}}$$

The kinetic OD<sub>420</sub> readings were converted into the slope of OD<sub>420</sub> over time (OD<sub>420</sub>/min), multiplied by 5000, and adjusted for the initial OD<sub>600</sub> reading at the first time point. The obtained MU values were then compared to those of the traditional assay (Table 1). More detailed lab notes on running this assay and subsequent processing in MARS Data Analysis are available in [Supplementary Data B](#).

### 2.4. Data comparison between the single-step $\beta$ -galactosidase assay and the traditional assay

See [Table 1](#).

**Table 1**

The single-step automated assay was consistent over multiple orders of magnitude and generated data equivalent to that of the traditional assay. The coefficients of variation of the Miller Unit [MU] measurements using the one-step and the traditional assay were similar for all  $\beta$ -galactosidase expressing strains suggesting that the two methods have comparable precision and reproducibility. Additionally, no statistically significant difference was observed between the single-step and traditional  $\beta$ -galactosidase assay (two-tailed Student's *t*-test assuming unequal variance;  $P > 0.05$ ).

	One-step assay			Traditional assay			T-test
	[MU]	Average	Coefficient of variation	[MU]	Average	Coefficient of variation	<i>P</i> value
MG1655 $\Delta lacZ$	–13 12 –5	<b>–2</b>	–	1.1 1.3 1	<b>1.1</b>	–	<b>0.71</b>
MG1655 $\Delta lacZ \Phi P_{pspA-lacZ}$	87 48 72	<b>69</b>	0.23	52 35 37	<b>42</b>	0.18	<b>0.12</b>
MG1655 $\Delta lacZ \Phi P_{pspA-lacZ} \Delta pspA$	691 642 613	<b>649</b>	0.05	665 677 686	<b>676</b>	0.01	<b>0.35</b>
MC1061 $\Delta lacZ \Phi P_{pspA-lacZ} \Delta pspA$	2985 2464 2351	<b>2600</b>	0.11	2070 2473 2062	<b>2202</b>	0.09	<b>0.17</b>

## Acknowledgements

This work was supported by BBSRC (BB/J00717X/1) and MRC (MR/M017672/1) project grants.

## Transparency document. Supplementary material

Supplementary data associated with this article can be found in the online version at <http://dx.doi.org/10.1016/j.dib.2016.05.063>.

## Appendix A. Supplementary material

Supplementary data associated with this article can be found in the online version at <http://dx.doi.org/10.1016/j.dib.2016.05.063>.

## References

- [1] J.H. Miller, *Experiments in Molecular Genetics*, CSH Laboratory Press, Cold Spring Harbor, NY, 1972.
- [2] J. Schaefer, G. Jovanovic, I. Kotta-Loizou, M. Buck, Single-step method for  $\beta$ -galactosidase assays in *Escherichia coli* using a 96-well microplate reader, *Anal. Biochem.* (2016), <http://dx.doi.org/10.1016/j.ab.2016.03.017>.