

MEETING ABSTRACT

Open Access

Analysis and reduction of eddy current effects induced by tesseral end zonal gradient coils in different collimator geometries for SPECT/MRI integration

Amine M Samudi^{1*}, Karen Van Audenhaege², Gunter Vermeeren¹, Emmeric Tanghe¹, Luc Martens¹, Roel Van Holen², Wout Joseph¹

From PSMR 2015: 4th Conference on PET/MR and SPECT/MR
La Biodola, Isola d'Elba, Italy. 17-21 May 2015

¹INTEC, Ghent University/iMinds, Ghent, Belgium

SPECT and MRI each have their respective advantages and limitations. Combining these two technologies in a synergistic manner allows researchers to exploit the strengths of both techniques but also result in disturbing eddy currents. In this paper, we studied the temporal variation of the induced magnetic field due to the transverse and longitudinal gradient coils, in a full-ring multi-pinhole collimator. We also investigated the effect of the ring geometry (hexagonal or pentagonal) on the resulting eddy currents and reduced the eddy currents by adding gaps between the collimators. We modeled x, y, and z-gradient coils and different arrangements of the SPECT collimators using FEKO. We arranged the collimators in pentagonal and hexagonal arrangements and we added gaps between the collimators in the pentagonal arrangement. The setup was simulated with a broadband simulation from 0 to 10 kHz with a step of 400 Hz to cover the frequency range of the gradient on-off switching (a sinusoidal ramp from 500 mT/m to 0 mT/m within 0.25 ms). The collimator design contains 20 loft-holes with 500- μ m-diameter pinhole openings. The density of the collimator is equal to 17.31 ± 0.10 g/cm³, and the conductivity equal to 108 n Ω .m. Simulations showed that the hexagonal geometry induces larger eddy currents. By adding relatively small gaps between the collimators (1.7 mm), the maximum value of the induced magnetic field is reduced by 50.6 % and 75.8 % for transverse and longitudinal gradient coils, respectively. As a result, the maximum value of the induced field is now less than 2 % of the applied gradient field.

Authors' details

¹INTEC, Ghent University/iMinds, Ghent, Belgium . ²ELIS, Ghent University/iMinds, Ghent, Belgium .

Published: 18 May 2015

doi:10.1186/2197-7364-2-S1-A51

Cite this article as: Samudi et al.: Analysis and reduction of eddy current effects induced by tesseral end zonal gradient coils in different collimator geometries for SPECT/MRI integration. *EJNMMI Physics* 2015 **2**(Suppl 1):A51.

ERRATUM

Open Access



Erratum to: Analysis and reduction of eddy current effects induced by tesseral end zonal gradient coils in different collimator geometries for SPECT/MRI integration

Amine M. Samoudi^{1*}, Karen Van Audenhaege², Gunter Vermeeren¹, Emmeric Tanghe¹, Luc Martens¹, Roel Van Holen² and Wout Joseph¹

Any correspondence should be addressed to George Loudos, Supplement 'Proceedings of the 4th PSMR Conference on PET/MR and SPECT/MR' organizer: gloudos@teiath.gr

Erratum

Unfortunately, the original version of this article [1] contained an error. The author's name, 'Amine M Samoudi', was spelt incorrectly. This can be found corrected in the author list above.

Author details

¹INTEC, Ghent University/iMinds, Ghent, Belgium. ²ELIS, Ghent University/iMinds, Ghent, Belgium.

Published online: 09 November 2015

Reference

1. Samoudi AM, Karen Van A, Gunter V, Emmeric T, Luc M, Roel Van H, et al. Analysis and reduction of eddy current effects induced by tesseral end zonal gradient coils in different collimator geometries for SPECT/MRI integration. *EJNMMI Physics*. 2015;2(1):A51.

Submit your manuscript to a SpringerOpen[®] journal and benefit from:

- ▶ Convenient online submission
- ▶ Rigorous peer review
- ▶ Immediate publication on acceptance
- ▶ Open access: articles freely available online
- ▶ High visibility within the field
- ▶ Retaining the copyright to your article

Submit your next manuscript at ▶ springeropen.com