MEETING ABSTRACT **Open Access**

Compatibility of metal additive manufactured tungsten collimator for SPECT/MRI integration

Amine M Samudi[1*], Karen Van Audenhaege[2], Gunter Vermeeren[1], Luc Martens[1], Roel Van Holen[2], Wout Joseph[1]

[1]INTEC, Ghent University/iMinds, Ghent, Belgium

We optimized the MR-compatibility of a novel tungsten collimator, produced with metal additive manufacturing that is part of a microSPECT insert for a preclinical SPECT/MRI scanner. We characterized the current density due to the gradient field and adapted the collimators by smart design to reduce the induced eddy currents. The z-gradient coil and the collimator were modeled with SEMCAD. The gradient strength was 510 mT/m, the gradient efficiency was about 3.4 mT/m/A. The setup was simulated with a working frequency of 10 kHz. The system consists of 7 identical collimators and digital silicon photomultipliers assembled in a ring. We evaluated the global reduction in current density J (reduction) based on the sum of all current densities in the collimator. We applied the following optimizations on the collimator: 1. We reduced the excessive material in the flanges. 2. We applied horizontal slits of 2 mm in the collimator surface. 3. We reduced material in the core; the photons are attenuated before they reach the core. The collimator will need a supporting structure. 4. The supporting structure can be avoided by using two vertical slits in the middle of the collimator. 5. We used a Z-shaped slit instead of the vertical slit. Results of simulations show that smaller flanges reduce the current density with 23%. The horizontal slits reduce the eddy currents with 6%. Using less material in the core or applying vertical slits results in the same reduction of current density. However, the vertical slits are cheaper because a hollow collimator requires supporting structures during production. Both can be combined if z-shaped slits are used to prevent attenuation problems. The reduction is then 27%. Finally, when all previous adaptations are combined, the reduction in eddy currents is about 56.3%.

Authors' details
[1]INTEC, Ghent University/iMinds, Ghent, Belgium . [2]ELIS, Ghent University/iMinds, Gent, Belgium .

EJNMMI Physics
a SpringerOpen Journal

CrossMark

Erratum to: Compatibility of metal additive manufactured tungsten collimator for SPECT/MRI integration

Amine M. Samoudi[1*], Karen Van Audenhaege[2], Gunter Vermeeren[1], Luc Martens[1], Roel Van Holen[2] and Wout Joseph[1]

Any correspondence should be addressed to George Loudos, Supplement 'Proceedings of the 4th PSMR Conference on PET/MR and SPECT/MR' organizer: gloudos@teiath.gr

Erratum

Unfortunately, the original version of this article [1] contained an error. The author's name, 'Amine M Samoudi', was spelt incorrectly. This can be found corrected in the author list above.

Author details
[1]INTEC, Ghent University/iMinds, Ghent, Belgium. [2]ELIS, Ghent University/iMinds, Ghent, Belgium.

Published online: 09 November 2015

Reference
1. Samudi AM, Van Audenhaege K, Vermeeren G, Luc M, Van Holen R, Joseph W. Compatibility of metal additive manufactured tungsten collimator for SPECT/MRI integration. EJNMMI Physics. 2015;2 Suppl 1:A52.