

Using Artificial Intelligence for Optimization of the Processes and Resource Utilization in Radiotherapy

Revathy Krishnamurthy, MD¹; Naveen Mummudi, DMRT MD¹; Jayant Sastri Goda, MD, DNB¹; Supriya Chopra, MD, DNB¹; Ben Heijmen, PhD²; and Jamema Swamidas, PhD¹

The radiotherapy (RT) process from planning to treatment delivery is a multistep, complex operation involving numerous levels of human-machine interaction and requiring high precision. These steps are labor-intensive and time-consuming and require meticulous coordination between professionals with diverse expertise. We reviewed and summarized the current status and prospects of artificial intelligence and machine learning relevant to the various steps in RT treatment planning and delivery workflow specifically in low- and middle-income countries (LMICs). We also searched the PubMed database using the search terms (Artificial Intelligence OR Machine Learning OR Deep Learning OR Automation OR knowledge-based planning AND Radiotherapy) AND (list of Low- and Middle-Income Countries as defined by the World Bank at the time of writing this review). The search yielded a total of 90 results, of which results with first authors from the LMICs were chosen. The reference lists of retrieved articles were also reviewed to search for more studies. No language restrictions were imposed. A total of 20 research items with unique study objectives conducted with the aim of enhancing RT processes were examined in detail. Artificial intelligence and machine learning can improve the overall efficiency of RT processes by reducing human intervention, aiding decision making, and efficiently executing lengthy, repetitive tasks. This improvement could permit the radiation oncologist to redistribute resources and focus on responsibilities such as patient counseling, education, and research, especially in resource-constrained LMICs.

JCO Global Oncol 8:e2100393. © 2022 by American Society of Clinical Oncology

Creative Commons Attribution Non-Commercial No Derivatives 4.0 License 

INTRODUCTION

The anticipated increase in cancer burden over the next few years could potentially overwhelm the oncology care system, especially in resource-constrained low- and middle-income countries (LMICs).^{1,2} Although radiotherapy (RT) is an indispensable component of cancer care, access to it worldwide is very inequitable, with the current density of RT machines per million population ranging from 0 to 11.6, depending on the economic situation of a country.³ The RT process starts with a series of visits to radiation oncology (RO) clinic, culminating in the final diagnosis, staging, and prognostication after which a radiation treatment protocol is assigned. Once a protocol is assigned, the subsequent RT treatment process can be categorized into imaging, target and organs-at-risk (OARs) segmentation, treatment plan generation, onboard imaging, treatment delivery, and quality assurance (QA) checks. These steps are labor-intensive and time-consuming, requiring multiple levels of human-machine interaction and a high degree of precision.⁴ The patient continues to visit the clinic on conclusion of therapy for toxicity management and follow-up. This workflow is summarized in [Figure 1](#).

The RT workflow requires meticulous coordination between trained medical professionals with diverse expertise, i.e., radiation oncologists, medical physicists, dosimetrists, and radiation therapists.⁵ Understaffing and workforce burnout is, unfortunately, a common problem plaguing RO in LMICs heightened by the ongoing COVID-19 pandemic.⁶ Training the highly specialized RO workforce requires a high cost and time commitment. The role of artificial intelligence (AI) and machine learning (ML) in optimizing RT processes to achieve the best human, technological, and financial resource utilization is worth exploring.

ARTIFICIAL INTELLIGENCE AND MACHINE LEARNING

Russell and Norvig⁷ have defined AI as “the designing and building of intelligent agents that receive precepts from the environment and take actions that affect that environment.” A more perceptive definition given by Goel is “the science of building artificial minds by understanding how natural minds work and understanding how natural minds work by building artificial minds.” ML is a branch of AI that allows computer systems to progressively learn, train, and improve on the knowledge

Author affiliations and support information (if applicable) appear at the end of this article.

Accepted on September 23, 2022 and published at ascopubs.org/journal-go on November 17, 2022; DOI <https://doi.org/10.1200/GO.21.00393>

CONTEXT

Key Objective

We review the role of artificial intelligence (AI) and machine learning (ML) in improving the efficiency of various radiotherapy processes and the challenges in their clinical integration.

Knowledge Generated

AI and ML can improve the accuracy, robustness, and speed of radiotherapy processes by reducing or eliminating human interference, aiding in decision making, and efficiently executing time-consuming, repetitive tasks. As its clinical utility remains yet to be proven, multi-institutional collaborative effort between various stakeholders is urgently needed, before the revolutionary impact of AI and ML bears fruition.

Relevance

The anticipated increase in cancer burden over the next few years coupled with cancer care becoming more personalized and tailored could potentially pressurize the oncology care system in the years to come, especially in low- and middle-income countries.

gained from a variety of input data without being overtly programmed.⁸ AI and ML can improve the accuracy, robustness, and speed of RT processes by reducing or eliminating human interference, aiding in decision making, and efficiently executing lengthy, repetitive tasks.⁹ Using ML can free up time for more rewarding tasks such as education, research, patient counseling, and quality checks. The following review focuses on the potential use of ML and AI to transform the existing RT workflow and create a sustainable model that

can be adopted in LMICs to supplement human efforts in labor-intensive tasks: segmentation, planning, and QA.

ROLE OF ML IN IMAGE SEGMENTATION

Manual segmentation (or contouring) of the target and OARs is a time-consuming and highly subjective task that lies at the core of RT planning. Historical solutions offered by technology to this conundrum include edge- and region-based methods and atlas-based methods of autosegmentation.¹⁰




FIG 1. Utility of AI/ML in RT processes. AI, artificial intelligence; CNN, convolutional neural networks; ML, machine learning; QA, quality assurance; RT, radiotherapy.

Deep learning (DL), a subset of ML, is essentially a neural network with three or more layers. These can comprise simple feed-forward models such as artificial neural network or complex models such as convolutional neural networks (CNN) and recurrent neural networks. CNN has been increasingly used to learn complex nonlinear relationships within the imaging data to speed up and improve OARs delineation in mediastinum, pelvis, thorax, brain, and head/neck.¹¹ Image segmentation on the basis of DL uses either patches or regions of an image or the entire image as input to estimate the likelihood that a given image sample belongs to the object being segmented. The likelihood map can be further enhanced by methods combining DL and deformable models.¹² Multiple papers published using ML for auto-segmentation of OARs have demonstrated no clinically meaningful difference between segmentation by model and clinicians or radiographers with very high values of dice similarity coefficient, while recording a significant reduction in time needed for the segmentation.¹³⁻¹⁷ For example, the average segmentation time for abdominal OARs liver, stomach, duodenum, and kidneys was 7.1 minutes with automation versus 22.6 minutes when done manually.¹³ Lesion segmentation is more complex than OARs segmentation because of the heterogeneity in shape, size, and location of the target and, therefore, is still in nascent stages.¹² Computer-aided diagnosis methods, including conventional radiomics and CNN-based algorithms that enhance lesion detection in diagnostic radiology, can potentially be used in lesion segmentation during RT planning.¹⁸

ROLE OF ML IN RADIOTHERAPY TREATMENT PLAN GENERATION AND ADAPTIVE PLANNING

The advent of intensity-modulated RT and volumetric modulated arc therapy that offer exceptionally conformal RT delivery has increased manifold the intricacy and complexity of RT planning. High-precision treatment procedures such as stereotactic body ablative RT often consume hours or even days of human effort for planning.¹⁹ Knowledge-based planning (KBP), which uses data from previous good cases to inform current patient planning parameters, has emerged as a powerful tool to accelerate the process of RT planning.¹⁹ Efforts are ongoing to establish indigenous KBP models for cancers common in LMICs, such as cervical cancers, and validate them in various geoethnic populations to test efficacy in patients with different anatomies on the basis of geographical locations.^{20,21} Supervised DL algorithms have been used for beam direction optimization, where the possible subsequent beam distribution is predicted on the basis of patient anatomy.²² Use of DL in the prediction of spatial dose distribution has been extensively explored, with different architectures of CNN being used to predict the geometric and planning parameters of historical patients.²² A significant gain in time has been reported with ML over non-ML methods such as column generation to select beam orientations, calculate the dose influence matrices, and finally solve the fluence map optimization with comparable

dosimetry.²³ ML algorithms have also been used to enhance KBP further to generate treatment plans.²⁴ Recent studies have even attempted to emulate the decision-making strategy of human planners when solving a specific dosimetric trade-off problem, thereby potentially reducing the element of subjectivity.²⁵

Another unique approach is the use of Pareto surface-based techniques for multicriteria optimization, where a database of plans is created for a single patient and the plan that achieves the best balance between different treatment planning goals is chosen by the planner and the physician.^{26,27} The Erasmus i-cycle (created in an academic university) is a vendor-neutral algorithm using multicriteria optimization that is in clinical use for external beam therapy and is being developed for CyberKnife, proton therapy, and brachytherapy (BiCyle).²⁸⁻³²

ROLE OF ML IN RADIOTHERAPY ONBOARD IMAGING AND TREATMENT DELIVERY

ML techniques, including DL approaches, have dealt with intra- and interfraction patient and organ motion during RT treatment delivery to aid tumor gating and motion tracking.³³ Frameworks have been built using neural networks trained on collected patient breathing data to predict the breathing pattern while delivering RT.³⁴ ML has been used to aid motion tracking by assisting in the detection of the tumor (marker-less tracking) or surrogate markers.³⁵ ML has been used to help avert setup errors and patient safety hazards by tracking the treatment room components and the patient's body in real time using 3D cameras to fine tune a CNN for object recognition.³⁶ A group of scientists have developed a computer vision-based pneumatic soft robot actuator to better estimate a patient's head pitch motion and to manipulate the patient head position on the basis of sensed head pitch motion, thereby potentially eliminating the need for immobilization with a thermoplastic mask.³⁷

The role of ML in online adaptive RT planning has been extensively explored, mainly in deformable registration and dose warping, facilitating high registration accuracy and efficient execution even if graphical processing units are unavailable.³⁸ A proof-of-concept study investigates online multileaf collimator tracking to generate appropriate safety margins for online adaptation of the treatment plan on the basis of the patient's motion and the ability of the machine to follow these excursions.³⁹ Algorithms can assist physicians in supervising variations during treatment course by evaluating daily setup variations and anatomic changes, for early identification of adaptive replanning requirement.⁴⁰

ROLE OF ML IN RADIOTHERAPY QUALITY ASSURANCE

Implementing regular and meticulous QA in RT is expected to lead to more accurate treatment delivery and better clinical outcomes. ML has excellent potential to enhance the efficacy and efficiency of RT QA processes as they are often repetitive and time-consuming. ML techniques have been used to predict gamma passing rates and the probability of the plan failing patient-specific intensity-modulated RT QA by

analyzing plan complexity; multiple components of the delivery system such as multileaf collimator, imaging system, and mechanical and dosimetric parameters, and plan delivery log files over time.⁴¹⁻⁴³ Another approach has identified RT treatment delivery errors using radiomics-based feature extraction from patient-specific gamma images.⁴⁴ A study has applied artificial neural network–based time series prediction modeling to predict the performance of beam symmetry of linear accelerators over time.⁴⁵ Although these *in silico* approaches of various ML tools have augmented RT QA procedures, it is pertinent to establish its real utility in the clinical context before implementation.

OTHER APPLICATIONS OF AI IN RADIATION ONCOLOGY

Data Annotation, Radiomics, and Response Prediction

Radiomics is a method that extracts a large number of features from medical images using data characterization algorithms.⁴⁶ Many institutions and health networks, including from India, are working to create repositories of annotated medical data and medical images including outcomes of treatment for furthering radiomic research in large image data sets.^{47,48} Distributed learning approaches with AI support have been used to conduct population-based studies on routine data and build decision support models.^{49,50} Image banking combined with predictive/prescriptive AI is a cost-effective and efficient alternative to identify signatures for response, toxicity, and outcome prediction after cancer treatment.⁵¹⁻⁵³

Natural Language Processing

Natural language processing (NLP) is a branch of AI that enables computers to interpret human language. NLP has already found application in the medical world to facilitate data extraction from free text in electronic medical records. The specific utility of NLP being explored in RO is standardization of contours and plans nomenclature to enable efficient data extraction.⁵⁴

ORIGINAL RESEARCH FROM LMICS USING AUTOMATION AND ML TO OPTIMIZE RT PLANNING, TREATMENT DELIVERY, AND QA PROCESSES

We searched the PubMed database using the search term (Artificial Intelligence[tw] OR Machine Learning[tw] OR Deep Learning[tw] OR Automation[tw] automated[tw] OR knowledge-based planning[tw] AND Radiotherapy[tw]) AND (list of Low- and Middle-Income Countries as defined by the World Bank at the time of writing this review). The search yielded a total of 90 results, of which results with first authors from the LMICs were chosen. The reference lists of retrieved articles were also reviewed to search for more studies. No language restrictions were imposed. A total of 20 research items with unique study objectives conducted with the aim of enhancing RT processes were studied in detail and are presented in [Table 1](#).

The majority of studies have focused on the utilization of CNN and other networks for autosegmentation.⁵⁵⁻⁶² The striking reduction in time burden seen with incorporating these

algorithms while maintaining the accuracy of contours can prove to be pivotal in resource allocation in LMICs. The studies on autoplanning have mainly used KBP.^{20,21,62,63} Developing indigenous models with local data rather than adopting western models to fit into the local workflow seems to be the standard approach, which is undoubtedly remarkable. Studies from LMICs using AI/ML to assist in online adaptive planning, treatment delivery, and QA are few and far between, and more work in this area must be encouraged.^{64,66-68}

CHALLENGES OF INTEGRATING AUTOMATION AND ML INTO CLINICAL PROCESSES

Although AI has already become pervasive in our day-to-day activities and has the potential to influence how medicine is practiced, many challenges remain before its complete integration into RT processes, as listed below.⁶⁹

1. Clinical utility yet to be determined: Most ML-based solutions are still in the stage of technological incubation, with the onus on the RO team to establish their clinical value.
2. Risk analysis of automation and AI: QA studies for automated treatment planning tools have been conducted, which stress the need for comprehensive manual review of the plans by physicians and physicists before implementation.⁷⁰
3. Black-box nature of AI algorithms: In the case of failure of AI-based solutions, there is no straightforward framework to fix the outcome or predict errors. This lack of transparency and difficulty in understanding the outputs and predicting failures may make physicians hesitant and distrustful to rely on AI in patient-related decisions, further delaying the adoption of AI into clinical practice. It is essential to train the RO staff to correctly use the ML model and accurately interpret the intended utility, scope, and limitations.
4. Interpretation: It is necessary to remember that even if some algorithms can perform at near-human ability, the way they perceive and interpret the inputs differs from the human mind.
5. Training data: A machine learning algorithm's accuracy and generalizability are influenced heavily by the quality and quantity of the training data more than the mathematical parameters. Since individual institutional data sets are bound to be minor, data sharing across multiple institutions can make the ML/DL algorithms more robust. Distributed learning is an emerging approach to securely transferring data sets between institutions.⁷¹
6. Patient privacy and anonymity: The potential of distributed learning to provide evidence-based personalized care in LMICs is immense. However, care should be taken to uphold the rules of ethics, standardization, and stringent privacy regulations.⁷²

In conclusion, integrating AI and ML in RT processes may allow radiation oncologists to spend more time on patient

TABLE 1. Original Research From LMICs Using Automation and ML to Optimize RT Planning, Delivery, and QA Processes

S. No.	Author(s), Country, and Year of Publication	RT Process Optimized	Purpose of the Study	Study Design	Automation Tool/AI Framework/ML Algorithm/Network Architecture Used	Significant Findings	Limitations of the Study
1	Liu et al, China 2021 ⁵⁵	Segmentation	Automatic segmentation of CTV and OARs after breast conservation surgery	Processing time measured for AI tool and pre- and post-AI assistance	Network architecture: CNN	Contouring time 20 and 30 minutes for CTV and OAR, respectively, reduced to 10 and 5 minutes with AI assistance	Single center, ground truth approved by only two oncologists, subjective scoring system, CT images with pacemaker and contrast not used
2	Diniz et al, Brazil 2020 ⁵⁶	Segmentation	Esophagus segmentation from planning CT images using an atlas-based DL approach	Training and testing of residual U-Net	Network: U-Net: a type of deep CNN	Efficiency and accuracy surpassing all earlier reports in the literature	Technical concerns as enumerated by authors, dissimilarity in training data
3	Men et al, China 2017 ⁵⁷	Segmentation	Automatic segmentation of target and OARs in rectal cancer	Comparison of performances of two networks	Network: Deep dilated CNN v U-Net	45 seconds per patient for segmentation of CTV and all OARs with good accuracy	Segmentation of intestine and colon not as good as other organs
4	Xue et al, China 2020 ⁵⁸	Segmentation	DL-based automatic detection and segmentation of brain metastases	Evaluation of a new DL model	Network: Cascaded 3D fully convolution network	Detection accuracy of 100% and strong correlation of segmentation with manual segmentations. Can simplify stereotactic RT planning and follow-up	Fixed size of the located nidus region used—more flexibility of method needed
5	Yang et al, China 2014 ⁵⁹	Segmentation	ML algorithm for autosegmentation of parotids on MRI	Validation of the automated parotid segmentation algorithm	ML algorithm: Support vector machine training	Potential to perform longitudinal or large-scale clinical studies to understand toxicity of parotid gland and treat radiation-induced xerostomia	CT-based method might have been more practical than MRI in LMICs
6	Liu et al, China 2020 ⁶⁰	Segmentation	Autosegmentation of CTV for post-MRM RT in breast cancer	Construction and validation model for autosegmentation of chest wall CTV	Network: CNN	It took 3.45 seconds to delineate chest wall CTV using the model	Single center, ground truth approved by only two oncologists, subjective scoring system
7	Pan et al, China 2021 ⁶¹	Segmentation	Autosegmentation of hippocampi on MRI	Validation of the new DL model	Network: CNN	Positive impact on improving delineation accuracy and reducing work load	Selection bias inherent to retrospective studies
8	Xia et al, China 2021 ⁶²	Segmentation + plan generation	AI-based full-process solution for rectal cancer RT	Combination of knowledge-based and script-based planning	Network: CNN	Total time for full-process planning without contour modification = 7 minutes, physician accepted 80% of autoplans without further operation	Solution may not be suitable for sites with complex target segmentation and beam angle selection protocols
9	Sarkar et al, India 2019 ⁶³	Plan generation	Non-DVH predictive method of KBP for VMAT SRS and SRT	Validation of the library plan-based model in new cases	ML algorithm: Not specified (commercial)	Model completes SRS/SRT plans in 1.5 runs	Plan quality dependent on quality of library plans

(Continued on following page)

TABLE 1. Original Research From LMICs Using Automation and ML to Optimize RT Planning, Delivery, and QA Processes (Continued)

S. No.	Author(s), Country, and Year of Publication	RT Process Optimized	Purpose of the Study	Study Design	Automation Tool/AI Framework/ML Algorithm/Network Architecture Used	Significant Findings	Limitations of the Study
10	Swamidas et al, India 2021 ²⁰	Plan generation	KBP for VMAT of cervical cancer	KBP model configured from institutional database compared with CP	ML algorithm: DL-based autosegmentation algorithm (commercial)	Comparable and for some organs, superior performance than CP	Modifications to the model not possible: Would need reconfiguration and revalidation for a set of dose constraints
11	Huang et al, China 2020 ⁷³	Plan generation	Develop an autoplanning platform to be interfaced with a commercial treatment planning system	Comparison of autoplans with manual VMAT plans for head and neck and prostate	ML algorithm: MCO	Autoplans clinically acceptable with better OAR sparing	Manual check needed to avoid system being trapped in a dead loop to achieve all dosimetric goals
12	Swamidas et al, India 2021 ²¹	Plan generation	Geoethnic validation of KBP models for image-guided VMAT in locally advanced cervical cancer	Compare and validate KBP models developed in two geoethnic patient populations	ML algorithm: DL-based autosegmentation algorithm (commercial)	Protocol compliance good in both clinical and KBP plans, across the institutions	Only abstract available
13	Li et al, China 2020 ⁷⁴	Plan generation	Automated estimation of BED	Comparison of dose parameters between pre- and post-programed models	NestNet, a variant of U-Net used to predict the BED distribution	Model predicts biologic dose distributions accurately—may pave the way for completely automated BED-based planning	Only anatomic information used for optimization; beam data, and calculation parameters not used
14	Wang et al, China 2013 ⁷⁵	Plan evaluation	Treatment plan quality evaluation	Semiautomated plan quality evaluation program compared with manual recording	Automation tool: JAVA and MATLAB-based	Evaluation time reduced from 10-20 to 2 minutes while maintaining accuracy	Difference of 0.2% in volume or 0.6 Gy in dose between the semiautomated program and manual recording
15	Jiang et al, China 2020 ⁷⁶	Plan evaluation	CNN-based dosimetry evaluation of esophageal RT planning	Model predicting DVH from DTH	DL: 1D-CN	Average prediction error on the planned target volume, left lung, right lung, heart, and spinal cord is 2.94%	Model takes multiple steps to predict DVH instead of end-to-end prediction, beam orientation not taken into account
16	Osman and Tamam, Sudan 2022 ⁶⁴	Dose distribution prediction	KBP for IMRT in head and neck cancer	KBP model used to generate model to predict dose distribution, which was compared with ground-truth plan	Attention-gated 3D U-Net architecture model	Attention-gated convolutions technique reduces network's redundancy and improves prediction performance compared with a baseline standard U-Net model	Model trained on a fixed beam configuration making it less generalizable to other sites and techniques
17	Nouri et al, Iran 2017 ⁶⁵	Treatment delivery	Evaluates accuracy of the neural network estimating tumor positions in real-time RT	Comparison of neural network training methods	AI algorithms: Neural network, GA, and PSO	Neural network algorithm can precisely trace the location of tumor. Accuracies of 0.8%, 12%, and 14% in the neural network, GA, and PSO, respectively	Clinical implication not enumerated

(Continued on following page)

TABLE 1. Original Research From LMICs Using Automation and ML to Optimize RT Planning, Delivery, and QA Processes (Continued)

S. No.	Author(s), Country, and Year of Publication	RT Process Optimized	Purpose of the Study	Study Design	Automation Tool/AI Framework/ML Algorithm/Network Architecture Used	Significant Findings	Limitations of the Study
18	Liang et al, China 2020 ⁶⁶	Onboard imaging	To evaluate and quantify intrafraction target motion in liver	A fully automated framework for analyzing intrafraction motion using orthogonal kV X-ray projections	U-Net to segment the fiducial markers	Precision and recall of the fiducial detection model were in excess of 95% demonstrating high model performance	Only select tumor locations were studied limiting generalizability
19	Li et al, China 2019 ⁶⁷	QA	Assess accuracy of ML to predict and classify QA results for VMAT plans	GPR prediction accuracy of PL and classification performance of PL and RF were evaluated	ML algorithms: PL regression model to predict GPR, RF model to classify QA as pass or fail	ML can assist VMAT QA and reduce QA workload	Exploratory study: Only GYN and head and neck VMAT plans used to train the ML model
20	Abubakar et al, Nigeria 2021 ⁶⁸	Offline image guidance protocols	Develop automated method for optimization and reduction of PTV margin using logged setup errors	Extract setup errors in three translational directions	Automated algorithm: MATLAB-based	Algorithm calculates weekly offline setup error correction values automatically: PTV margins can be accordingly reduced up to 48% for head and neck cancers	Does not consider rotational errors

Abbreviations: 1D-CN, one-dimensional convolutional network; AI, artificial intelligence; BED, biologically equivalent dose; CNN, convolutional neural networks; CP, clinical plans; CT, computed tomography; CTV, clinical target volume; DL, deep learning; DTH, distance to target histogram; DVH, dose volume histogram; GA, genetic algorithm; GPR, gamma passing rate; GYN, gynecological; IMRT, intensity-modulated RT; JAVA and MATLAB, computer programming languages; KBP, knowledge-based planning; LMICs, low- and middle-income countries; MCO, multicriteria optimization; ML, machine learning; MRI, magnetic resonance imaging; MRM, modified radical mastectomy; OARs, organs at risk; PL, poisson lasso; PSO, particle swarm optimization; PTV, planning target volume; QA, quality assurance; RF, random forest; RT, radiotherapy; S. No., serial number; SRS, stereotactic radiosurgery; SRT, stereotactic radiotherapy; VMAT, volumetric modulated arc therapy.

consultation, teaching, and research in resource-constrained setups with a heavy workload. Given the transformative impact that AI-based technology can bring to clinical processes and workflow, it is essential to integrate these concepts early in medical education and RO residency to facilitate a better understanding of the methods and encourage innovation. We

have in front of us a means to revolutionize the practice of RT as we know it. It is our responsibility toward the future generation to understand, plan, prioritize, conduct meaningful research, integrate and constantly improve AI and ML in RT without bias or prejudice, and deliver it to settings where the impact would be maximum.

AFFILIATIONS

¹Department of Radiation Oncology, Tata Memorial Centre, Homi Bhabha National Institute, Mumbai, India

²Division of Medical Physics, Department of Radiation Oncology, Erasmus MC Cancer Institute, Erasmus University Rotterdam, Rotterdam, the Netherlands

CORRESPONDING AUTHOR

Naveen Mummudi, Department of Radiation Oncology, Tata Memorial Centre, Homi Bhabha National Institute, Parel, Mumbai, 400016, India; e-mail: mbnaveenk@gmail.com.

AUTHOR CONTRIBUTIONS

Conception and design: All authors

Collection and assembly of data: All authors

Data analysis and interpretation: All authors

Manuscript writing: All authors

Final approval of manuscript: All authors

Accountable for all aspects of the work: All authors

AUTHORS' DISCLOSURES OF POTENTIAL CONFLICTS OF INTEREST

The following represents disclosure information provided by authors of this manuscript. All relationships are considered compensated unless otherwise noted. Relationships are self-held unless noted. I = Immediate Family Member, Inst = My Institution. Relationships may not relate to the subject matter of this manuscript. For more information about ASCO's conflict of interest policy, please refer to www.asco.org/rwc or ascopubs.org/go/authors/author-center.

Open Payments is a public database containing information reported by companies about payments made to US-licensed physicians (Open Payments).

Supriya Chopra

This author is an Associate Editor for *JCO Global Oncology*. Journal policy recused the author from having any role in the peer review of this manuscript.

Honoraria: Varian Medical Systems, KM Pharmaceuticals (KORTUC) (Inst)

Consulting or Advisory Role: KM Pharmaceuticals (KORTUC) (Inst)

Speakers' Bureau: Varian Medical Systems (Inst)

Research Funding: Varian Medical Systems, Elekta (Inst)

No other potential conflicts of interest were reported.

REFERENCES

- Sung H, Ferlay J, Siegel RL, et al: Global Cancer Statistics 2020: GLOBOCAN estimates of incidence and mortality worldwide for 36 cancers in 185 countries. *CA Cancer J Clin* 71:209-249, 2021
- Zubizarreta EH, Fidarova E, Healy B, et al: Need for radiotherapy in low- and middle-income countries—The silent crisis continues. *Clin Oncol* 27:107-114, 2015
- Abdel-Wahab M, Gondhwiardjo SS, Rosa AA, et al: Global radiotherapy: Current status and future directions-white paper. *JCO Glob Oncol* 7:827-842, 2021
- Bortfeld T, Jeraj R: The physical basis and future of radiation therapy. *Br J Radiol* 84:485-498, 2011
- Valentini V, Boldrini L, Mariani S, Massaccesi M: Role of radiation oncology in modern multidisciplinary cancer treatment. *Mol Oncol* 14:1431-1441, 2020
- Wadasadawala T, Kumar A, Laskar SG, et al: Multinational study to assess stress levels among the health care workers of radiation oncology community at the outset of the COVID-19 pandemic. *JCO Glob Oncol* 7:464-473, 2021
- Russell SJ, Norvig P: *Artificial Intelligence: A Modern Approach* (ed 4). Hoboken, Pearson, 2021
- Nichols JA, Herbert Chan HW, Baker MAB: Machine learning: Applications of artificial intelligence to imaging and diagnosis. *Biophys Rev* 11:111-118, 2019
- Davenport T, Kalakota R: The potential for artificial intelligence in healthcare. *Future Healthc J* 6:94-98, 2019
- Kalet IJ, Paluszynski W: Knowledge-based computer systems for radiotherapy planning. *Am J Clin Oncol* 13:344-351, 1990
- Boldrini L, Bibault JE, Masciocchi C, et al: Deep learning: A review for the radiation oncologist. *Front Oncol* 9:977, 2019
- Sahiner B, Pezeshk A, Hadjiiski LM, et al: Deep learning in medical imaging and radiation therapy. *Med Phys* 46:e1-e36, 2019
- Racine D, Becce F, Viry A, et al: Task-based characterization of a deep learning image reconstruction and comparison with filtered back-projection and a partial model-based iterative reconstruction in abdominal CT: A phantom study. *Phys Med* 76:28-37, 2020
- Nikolov S, Blackwell S, Zverovitch A, et al: Clinically applicable segmentation of head and neck anatomy for radiotherapy: Deep learning algorithm development and validation study. *J Med Internet Res* 23:e26151, 2021
- Porter E, Fuentes P, Siddiqui Z, et al: Hippocampus segmentation on noncontrast CT using deep learning. *Med Phys* 47:2950-2961, 2020
- van Velzen SGM, Bruns S, Wolterink JM, et al: AI-based quantification of planned radiotherapy dose to cardiac structures and coronary arteries in breast cancer patients. *Int J Radiat Oncol Biol Phys* 112:611-620, 2022
- Rigaud B, Anderson BM, Yu ZH, et al: Automatic segmentation using deep learning to enable online dose optimization during adaptive radiation therapy of cervical cancer. *Int J Radiat Oncol Biol Phys* 109:1096-1110, 2021
- Giger ML, Chan HP, Boone J: Anniversary paper: History and status of CAD and quantitative image analysis: The role of medical physics and AAPM. *Med Phys* 35:5799-5820, 2008
- Moore KL: Automated radiotherapy treatment planning. *Semin Radiat Oncol* 29:209-218, 2019
- Swamidass J, Pradhan S, Chopra S, et al: Development and clinical validation of knowledge-based planning for volumetric modulated arc therapy of cervical cancer including pelvic and para aortic fields. *Phys Imaging Radiat Oncol* 18:61-67, 2021

21. Swamidias J, Assenholt M, Serban M, et al: PO-1855 Protocol compliance of two knowledge based models in two geo-ethnic populations for cervical cancer. *Radiother Oncol* 161:S1581, 2021
22. Wang M, Zhang Q, Lam S, et al: A review on application of deep learning algorithms in external beam radiotherapy automated treatment planning. *Front Oncol* 10:580919, 2020
23. Sadeghnejad Barkousaraie A, Ogunmolu O, Jiang S, et al: A fast deep learning approach for beam orientation optimization for prostate cancer treated with intensity-modulated radiation therapy. *Med Phys* 47:880-897, 2020
24. Li X, Zhang J, Sheng Y, et al: Automatic IMRT planning via static field fluence prediction (AIP-SFFP): A deep learning algorithm for real-time prostate treatment planning. *Phys Med Biol* 65:175014, 2020
25. Shen C, Nguyen D, Chen L, et al: Operating a treatment planning system using a deep-reinforcement learning-based virtual treatment planner for prostate cancer intensity-modulated radiation therapy treatment planning. *Med Phys* 47:2329-2336, 2020
26. Craft DL, Hong TS, Shih HA, et al: Improved planning time and plan quality through multicriteria optimization for intensity-modulated radiotherapy. *Int J Radiat Oncol Biol Phys* 82:e83-e90, 2012
27. Wang M, Zhang Q, Lam S, et al: A review on application of deep learning algorithms in external beam radiotherapy automated treatment planning. *Front Oncol* 10:580919, 2020
28. Rossi L, Breedveld S, Heijmen BJM, et al: On the beam direction search space in computerized non-coplanar beam angle optimization for IMRT-prostate SBRT. *Phys Med Biol* 57:5441-5458, 2012
29. Rossi L, Sharfo AW, Aluwini S, et al: First fully automated planning solution for robotic radiosurgery—Comparison with automatically planned volumetric arc therapy for prostate cancer. *Acta Oncol* 57:1490-1498, 2018
30. van de Water S, Kooy HM, Heijmen BJM, et al: Shortening delivery times of intensity modulated proton therapy by reducing proton energy layers during treatment plan optimization. *Int J Radiat Oncol Biol Phys* 92:460-468, 2015
31. Breedveld S, Bennis ABA, Aluwini S, et al: Fast automated multi-criteria planning for HDR brachytherapy explored for prostate cancer. *Phys Med Biol* 64:205002, 2019
32. Oud M, Kolkman-Deurloo IK, Mens JW, et al: Fast and fully-automated multi-criterial treatment planning for adaptive HDR brachytherapy for locally advanced cervical cancer. *Radiother Oncol* 148:143-150, 2020
33. Lin T, Li R, Tang X, et al: Markerless gating for lung cancer radiotherapy based on machine learning techniques. *Phys Med Biol* 54:1555-1563, 2009
34. Terunuma T, Tokui A, Sakae T: Novel real-time tumor-contouring method using deep learning to prevent mistracking in X-ray fluoroscopy. *Radiol Phys Technol* 11:43-53, 2018
35. Mylonas A, Keall PJ, Booth JT, et al: A deep learning framework for automatic detection of arbitrarily shaped fiducial markers in intrafraction fluoroscopic images. *Med Phys* 46:2286-2297, 2019
36. Santhanam A, Min Y, Beron P, et al: Low SU-D-201-05: On the automatic recognition of patient safety hazards in a radiotherapy setup using a novel 3D camera system and a deep learning framework. *Med Phys* 43:3334-3335, 2016
37. Ogunmolu OP, Gu X, Jiang S, et al: Vision-based control of a soft robot for maskless head and neck cancer radiotherapy. August 21-24, 2016, Fort Worth, TX, 2016 IEEE Int Conf Autom Sci Eng:180-187, 2016
38. Xiao H, Ren G, Cai J: A review on 3D deformable image registration and its application in dose warping. *Radiat Med Prot* 1:171-178, 2020
39. Glitzner M, Fast MF, de Senneville BD, et al: Real-time auto-adaptive margin generation for MLC-tracked radiotherapy. *Phys Med Biol* 62:186-201, 2017
40. Guidi G, Maffei N, Meduri B, et al: A machine learning tool for re-planning and adaptive RT: A multicenter cohort investigation. *Phys Med* 32:1659-1666, 2016
41. Valdes G, Chan MF, Lim SB, et al: IMRT QA using machine learning: A multi-institutional validation. *J Appl Clin Med Phys* 18:279-284, 2017
42. Lam D, Zhang X, Li H, et al: Predicting gamma passing rates for portal dosimetry-based IMRT QA using machine learning. *Med Phys* 46:4666-4675, 2019
43. Osman AF, Maalef NJ, Jayesh K: Prediction of the individual multileaf collimator positional deviations during dynamic IMRT delivery priori with artificial neural network. *Med Phys* 47:1421-1430, 2020
44. Ma C, Wang R, Zhou S, et al: The structural similarity index for IMRT quality assurance: Radiomics-based error classification. *Med Phys* 48:80-93, 2021
45. Li Q, Chan MF: Predictive time series modeling using artificial neural networks for linac beam symmetry – An empirical study. *Ann N Y Acad Sci* 1387:84-94, 2017
46. Agarwal JP, Sinha S, Goda JS, et al: Tumor radiomic features complement clinico-radiological factors in predicting long-term local control and laryngectomy free survival in locally advanced laryngo-pharyngeal cancers. *Br J Radiol* 93:20190857, 2020
47. Kundu S, Chakraborty S, Mukhopadhyay J, et al: Research goal-driven data model and harmonization for de-identifying patient data in radiomics. *J Digit Imaging* 34:986-1004, 2021
48. Chopra S: Development of Clinically High Efficient Platforms for Individualised Treatment of Cervix Cancer. <https://clinicaltrials.gov/ct2/show/NCT05102240>
49. Field M, Vinod S, Aherne N, et al: Implementation of the Australian Computer-Assisted Theragnostics (AusCAT) network for radiation oncology data extraction, reporting and distributed learning. *J Med Imaging Radiat Oncol* 65:627-636, 2021
50. Choudhury A, Theophanous S, Lønne P-I, et al: Predicting outcomes in anal cancer patients using multi-centre data and distributed learning—A proof-of-concept study. *Radiother Oncol* 159:183-189, 2021
51. Yakar M, Etiz D, Metintas M, et al: Prediction of radiation pneumonitis with machine learning in stage III lung cancer: A pilot study. *Technol Cancer Res Treat* 20:15330338211016373, 2021
52. Jin C, Yu H, Ke J, et al: Predicting treatment response from longitudinal images using multi-task deep learning. *Nat Commun* 12:1851, 2021
53. Patel M, Zhan J, Natarajan K, et al: Machine learning-based radiomic evaluation of treatment response prediction in glioblastoma. *Clin Radiol* 76:628.e17-628.e27, 2021
54. Bitterman DS, Miller TA, Mak RH, et al: Clinical natural language processing for radiation oncology: A review and practical primer. *Int J Radiat Oncol Biol Phys* 110:641-655, 2021
55. Liu Z, Liu F, Chen W, et al: Automatic segmentation of clinical target volume and organs-at-risk for breast conservative radiotherapy using a convolutional neural network. *Cancer Manag Res* 13:8209-8217, 2021
56. Diniz JOB, Ferreira JL, Diniz PHB, et al: Esophagus segmentation from planning CT images using an atlas-based deep learning approach. *Comput Methods Programs Biomed* 197:105685, 2020
57. Men K, Dai J, Li Y: Automatic segmentation of the clinical target volume and organs at risk in the planning CT for rectal cancer using deep dilated convolutional neural networks. *Med Phys* 44:6377-6389, 2017
58. Xue J, Wang B, Ming Y, et al: Deep learning-based detection and segmentation-assisted management of brain metastases. *Neuro Oncol* 22:505-514, 2020

59. Yang X, Wu N, Cheng G, et al: Automated segmentation of the parotid gland based on atlas registration and machine learning: A longitudinal MRI study in head-and-neck radiation therapy. *Int J Radiat Oncol Biol Phys* 90:1225-1233, 2014
60. Liu Z, Liu F, Chen W, et al: Automatic segmentation of clinical target volumes for post-modified radical mastectomy radiotherapy using convolutional neural networks. *Front Oncol* 10:581347, 2020
61. Pan K, Zhao L, Gu S, et al: Deep learning-based automatic delineation of the hippocampus by MRI: Geometric and dosimetric evaluation. *Radiat Oncol* 16:12, 2021
62. Xia X, Wang J, Li Y, et al: An artificial intelligence-based full-process solution for radiotherapy: A proof of concept study on rectal cancer. *Front Oncol* 10:616721, 2021
63. Sarkar B, Munshi A, Ganesh T, et al: Standardization of volumetric modulated arc therapy-based frameless stereotactic technique using a multidimensional ensemble-aided knowledge-based planning. *Med Phys* 46:1953-1962, 2019
64. Osman AFI, Tamam NM: Attention-aware 3D U-Net convolutional neural network for knowledge-based planning 3D dose distribution prediction of head- and-neck cancer. *J Appl Clin Med Phys* 23:e13630, 2022
65. Nouri S, Hosseini Pooya SM, Soltani Nabipour J: Comparative analysis of neural network training methods in real-time radiotherapy. *J Biomed Phys Eng* 7:51-58, 2017
66. Liang Z, Zhou Q, Yang J, et al: Artificial intelligence-based framework in evaluating intrafraction motion for liver cancer robotic stereotactic body radiation therapy with fiducial tracking. *Med Phys* 47:5482-5489, 2020
67. Li J, Wang L, Zhang X, et al: Machine learning for patient-specific quality assurance of VMAT: Prediction and classification accuracy. *Int J Radiat Oncol Biol Phys* 105:893-902, 2019
68. Abubakar A, Zamri NAM, Shaukat SI, et al: Automated algorithm for calculation of setup corrections and planning target volume margins for offline image-guided radiotherapy protocols. *J Appl Clin Med Phys* 22:137-146, 2021
69. Netherton TJ, Cardenas CE, Rhee DJ, et al: The emergence of artificial intelligence within radiation oncology treatment planning. *Oncology* 99:124-134, 2021
70. Kisling K, Johnson JL, Simonds H, et al: A risk assessment of automated treatment planning and recommendations for clinical deployment. *Med Phys* 46:2567-2574, 2019
71. Kiser KJ, Fuller CD, Reed VK: Artificial intelligence in radiation oncology treatment planning: A brief overview. *J Med Artif Intell* 2:9, 2019
72. Yirmibesoglu Erkal E, Akpınar A, Erkal HŞ: Ethical evaluation of artificial intelligence applications in radiotherapy using the Four Topics Approach. *Artif Intell Med* 115:102055, 2021
73. Huang X, Quan H, Zhao B, et al: A plan template-based automation solution using a commercial treatment planning system. *J Appl Clin Med Phys* 21:13-25, 2020
74. Li Y, He K, Ma M, et al: Using deep learning to model the biological dose prediction on bulky lung cancer patients of partial stereotactic ablation radiotherapy. *Med Phys* 47:6540-6550, 2020
75. Wang J, Chen W, Studenski M, et al: A semi-automated tool for treatment plan-quality evaluation and clinical trial quality assurance. *Phys Med Biol* 58:N181-N187, 2013
76. Jiang D, Yan H, Chang N, et al: Convolutional neural network-based dosimetry evaluation of esophageal radiation treatment planning. *Med Phys* 47:4735-4742, 2020

