



## **Wearable cardioverter-defibrillator vest as a diagnostic and therapeutic tool after COVID-19**

**Authors:** Tomasz Chyży, Barbara Małecka, Jacek Bednarek, Małgorzata Mielnik, Maciej Dębski, Tomasz Miszański-Jamka, Krzysztof Boczar, Jacek Lelakowski, Andrzej Ząbek

**Article type:** Clinical vignette

**Received:** March 5, 2023

**Accepted:** April 26, 2023

**Early publication date:** May 30, 2023

This article is available in open access under Creative Common Attribution-Non-Commercial-No Derivatives 4.0 International (CC BY-NC-ND 4.0) license, allowing to download articles and share them with others as long as they credit the authors and the publisher, but without permission to change them in any way or use them commercially.

## **Wearable cardioverter-defibrillator vest as a diagnostic and therapeutic tool after COVID-19**

**Short title:** WCD vest after COVID-19

Tomasz Chyży<sup>1</sup>, Barbara Małecka<sup>1,2</sup>, Jacek Bednarek<sup>1</sup>, Małgorzata Mielnik<sup>3</sup>, Maciej Dębski<sup>4</sup>, Tomasz Miszalski-Jamka<sup>3</sup>, Krzysztof Boczar<sup>1</sup>, Jacek Lelakowski<sup>1,2</sup>, Andrzej Ząbek<sup>1,2</sup>

<sup>1</sup>Department of Electrophysiology, Institute of Cardiology, John Paul II Hospital, Kraków, Poland

<sup>2</sup>Institute of Cardiology, Jagiellonian University Medical College, Kraków, Poland

<sup>3</sup>Department of Radiology and Diagnostic Imaging, Institute of Cardiology, John Paul II Hospital, Kraków, Poland

<sup>4</sup>Department of Cardiology, Norfolk and Norwich University Hospital, University of East Anglia, Norwich, United Kingdom

### **Correspondence to:**

Andrzej Ząbek, MD,  
Department of Electrophysiology,  
Institute of Cardiology,  
John Paul II Hospital,  
Prądnicka 80, 31–202 Kraków, Poland,  
e-mail: andrzej\_j\_z@poczta.onet.pl

A 29-year-old female patient, previously in excellent health, presented with recurrent fainting caused by ventricular tachyarrhythmia three months after contracting COVID-19. An ambulatory Holter ECG showed a brief episode of non-sustained ventricular tachycardia (nsVT) at a rate of 270 bpm, which coincided with the presyncope (**Figure 1A**). The patient's resting ECG was normal, and an electrophysiology study did not reveal an inducible arrhythmia. Additionally, a cardiac magnetic resonance (CMR) scan did not demonstrate features of myocarditis (**Figure 1B**, Supplementary material, *Video S1*).

However, during hospitalization, the patient experienced recurrent nsVT, indicating an increased vulnerability to sustained ventricular arrhythmia (**Figure 1C**). It was assumed that this arrhythmia was related to the recent COVID-19 infection and would likely resolve

spontaneously. To monitor and treat the arrhythmia, the patient was offered a wearable cardioverter-defibrillator (WCD) instead of an implantable cardioverter-defibrillator (ICD). The patient was started on bisoprolol and electrolyte replacement, received education and training on using the WCD, and was subsequently discharged home (Figure 1D). The next day, the patient was readmitted with an electrical storm recorded by the WCD. The patient experienced 11 episodes of ventricular tachycardia with a cycle length ranging from 220 to 240 milliseconds and a duration varying from 40 to 220 seconds, including five episodes occurring at intervals of at least 5 minutes (Figure 1E–F). The patient did not lose consciousness during the electrical storm episodes and was able to abort the high-voltage therapy by simultaneously pressing two buttons on the WCD unit (Supplementary material, *Text S2*). The arrhythmia was successfully treated with amiodarone infusion.

Additionally, an ablation procedure was performed (Supplementary material, *Text S3*).

Six weeks later, the patient experienced another episode of an electrical storm which was recorded by the WCD. The patient was treated at the Emergency Department with oral propafenone and electrolyte replacement. The patient was admitted electively four months later to reassess the need for a WCD or ICD.

A Holter ECG showed 286 single ventricular ectopics and one pair. The arrhythmia frequency in the WCD memory was down-trending (Figure 1F). Given the reduction of the arrhythmia, the WCD prevention was discontinued. However, due to uncertainty surrounding the post-COVID-19 arrhythmia, an ICD implantation was offered but was refused by the patient. In the fourteen-month follow-up, the patient remained well with no reported syncope or palpitations. In young women, myocarditis has been reported as one of the cardiac complications of COVID-19, and it has been associated with an increased risk of sudden cardiac death (SCD) [1, 2] (Supplementary material, *Text S4*).

We observed a decrease in the arrhythmia burden during follow-up, indicating a self-limiting course of the disease. Using a WCD as a bridge therapy can be an effective temporary solution in cases where the patient is at risk of SCD but is not yet a candidate for a permanent ICD implantation [3]. In this case, the WCD was a valuable diagnostic and therapeutic tool to prevent SCD [4]. Eventually, the arrhythmia resolved, and the patient decided not to have an ICD implanted, which, in hindsight, turned out to be the right decision and the patient was not burdened with unnecessary therapy. This case highlights the importance of a shared decision-making process between the healthcare provider and the patient, where the patient's values, preferences, and concerns are taken into account when making treatment decisions. In

summary, we believe the WCD is a valuable option for patients with ventricular arrhythmia in the setting of a recent COVID-19 infection and myocarditis.

### **Supplementary material**

Supplementary material is available at [https://journals.viamedica.pl/kardiologia\\_polska](https://journals.viamedica.pl/kardiologia_polska).

### **Article information**








**Conflict of interest:** None declared.

**Funding:** None.

**Open access:** This article is available in open access under Creative Common Attribution-Non-Commercial-No Derivatives 4.0 International (CC BY-NC-ND 4.0) license, which allows downloading and sharing articles with others as long as they credit the authors and the publisher, but without permission to change them in any way or use them commercially. For commercial use, please contact the journal office at [kardiologiapolska@ptkardio.pl](mailto:kardiologiapolska@ptkardio.pl).

### **REFERENCES**

1. Annie FH, Alkhaimy H, Nanjundappa A, et al. Association between myocarditis and mortality in COVID-19 patients in a large registry. *Mayo Clin Proc Innov Qual Outcomes*. 2022; 6(2): 114–119, doi: [10.1016/j.mayocpiqo.2021.12.006](https://doi.org/10.1016/j.mayocpiqo.2021.12.006), indexed in Pubmed: [34977468](https://pubmed.ncbi.nlm.nih.gov/34977468/).
2. Lynge TH, Nielsen TS, Gregers Winkel Bo, et al. Sudden cardiac death caused by myocarditis in persons aged 1-49 years: a nationwide study of 14 294 deaths in Denmark. *Forensic Sci Res*. 2019; 4(3): 247–256, doi: [10.1080/20961790.2019.1595352](https://doi.org/10.1080/20961790.2019.1595352), indexed in Pubmed: [31489390](https://pubmed.ncbi.nlm.nih.gov/31489390/).
3. Sterliński M, Oręziak A, Przybylski A, et al. Experts of the Heart Rhythm Section of the Polish Cardiac Society: opinion on the use of wearable cardioverter-defibrillators in Poland. *Kardiol Pol*. 2019; 77(2): 238–243, doi: [10.5603/KP.2019.0031](https://doi.org/10.5603/KP.2019.0031), indexed in Pubmed: [30816992](https://pubmed.ncbi.nlm.nih.gov/30816992/).
4. Pilawa M, Kaźmierczak J, Kiedrowicz RM. An optimized long QT syndrome differentiating protocol: A new indication for wearable cardioverter-defibrillator vests. *Kardiol Pol*. 2022; 80(7-8): 857–858, doi: [10.33963/KP.a2022.0142](https://doi.org/10.33963/KP.a2022.0142), indexed in Pubmed: [35677993](https://pubmed.ncbi.nlm.nih.gov/35677993/).



**Figure 1.** **A.** Ambulatory Holter electrocardiography (ECG) monitoring: episode non-sustained ventricular tachycardia (nsVT) at a rate of 270 bpm. **B.** Cardiac magnetic resonance: Three chamber long-axis apical view of the left ventricle. Late gadolinium enhancement imaging indicates no post-inflammatory lesions with normal global left and right ventricular systolic function. **C.** ECG monitoring during hospitalization (record from the bedside monitor): an approximately 8-second episode of nsVT. **D.** Patient wearing a wearable cardioverter defibrillator. **E.** Trace of one of the ventricular tachycardia episodes (the arrows show the moments when two buttons on the device are pressed simultaneously to interrupt the high-voltage therapy). **F.** Trend of arrhythmia burden over the follow-up period (number of episodes and their duration)