

Original paper

## The role of ultra-low-dose computed tomography in the detection of pulmonary pathologies: a prospective observational study

Gobi K.<sup>A,B,C,D,E</sup>, Venkatesh Kasi Arunachalam<sup>A,B,D,E</sup>, Rajesh Kumar Varatharajaperumal<sup>C,F</sup>, Mathew Cherian<sup>A,E</sup>, Gopinath Periaswamy<sup>C,F</sup>, Rajesh S<sup>C,F</sup>

Kovai Medical Centre and Hospital, Coimbatore, India

### Abstract

**Purpose:** The aim of the study was to compare the image noise, radiation dose, and image quality of ultra-low-dose computed tomography (CT) and standard CT in the imaging of pulmonary pathologies.

**Material and methods:** This observational study was performed between July 2020 and August 2021. All enrolled patients underwent both ultra-low-dose and standard CTs. The image noise, image quality for normal pulmonary structures, presence or absence of various pulmonary lesions, and radiation dose were recorded for each of the scans. The findings of standard-dose CT were regarded as the gold standard and compared with that of ultra-low-dose CT.

**Results:** A total of 124 patients were included in the study. The image noise was higher in the ultra-low-dose CT compared to standard-dose CT. The overall image quality was determined to be diagnostic in 100% of standard CT images and in 96.77% of ultra-low-dose CT images with proportional worsening of the image quality as the body mass index (BMI) range was increased. Ultra-low-dose CT offered higher (> 90%) sensitivity for lesions like consolidation (97%), pleural effusion (95%), fibrosis (92%), and solid pulmonary nodules (91%). The effective radiation dose (mSv) was many times lower in ultra-low-dose CT when compared to standard-dose CT (mean ± SD: 0.50 ± 0.005 vs. 3.99 ± 1.57).

**Conclusions:** The radiation dose of ultra-low-dose chest CT was almost equal to that of a chest X-ray. It could be used for the screening and/or follow-up of patients with solid pulmonary nodules (> 3 mm) and consolidation.

**Key words:** ultra-low-dose CT, pulmonary nodules, radiation dose, third generation, dual source, dual energy.

### Introduction

High-resolution computed tomography (HRCT) of the thorax has become an indispensable diagnostic modality in the clinical context of lung pathologies. Consequently, the harmful effects related to radiation exposure have become an important factor to consider [1-3]. The high radiation dose of HRCT of the thorax and the location of most radiosensitive organs within the scan range make the situation much worse [3,4].

Some of the common indications for HRCT of the thorax include screening/diagnosis/follow-up of lung cancer, pulmonary infections including COVID-19 pneumonia, diffuse infiltrative lung disease, chronic obstructive lung disease, and acute respiratory distress syndrome [5-10].

It would be of great value to diagnose these conditions with the lowest possible radiation dose. To fulfil this need, a low-dose chest computed tomography (CT) protocol with a radiation dose of around 1.5 mSv had previously been implemented. Many previous studies have shown

### Correspondence address:

Venkatesh Kasi Arunachalam, Kovai Medical Centre and Hospital, Coimbatore, India, e-mail: [drkasivenkatesh@yahoo.co.in](mailto:drkasivenkatesh@yahoo.co.in)

### Authors' contribution:

A Study design · B Data collection · C Statistical analysis · D Data interpretation · E Manuscript preparation · F Literature search · G Funds collection

that this protocol offers good quality images for imaging some lung pathologies [3,6,7,11,12]. However, it is unable to offer complete protection from the risk of radiation [5].

Hence, an ultra-low-dose chest CT protocol offering a sub-millisievert radiation dose (~equal to 2 chest X-ray sessions) was considered [3,4,9]. This dramatic reduction in radiation dose was made possible after the introduction of several newer technologies [1,3,5,7,13-15].

This prospective study was done to test the ability of a non-contrast ultra-low-dose chest CT protocol to detect common pulmonary lesions using a third-generation Siemens dual-source CT scanner (Somatom Force).

## Material and methods

### Patient selection

The Local Ethics and Scientific Committee approved this study, and written informed consent was obtained from all participants. Patients of all age groups of both sexes, who underwent clinically indicated unenhanced HRCT thorax scans in the Radiology Department between July 2020 and August 2021 were included in this study. Images with extensive artifacts were excluded.

### Methodology

The study was done using a third-generation, dual-source, dual-energy CT scanner (Somatom force, Siemens Healthcare). All the images were obtained with patients in the supine position. No IV contrast medium was injected. All patients underwent both standard and ultra-low-dose CT scans without a change in position or time delay. But the order in which they were taken was interchanged randomly to avoid bias. Because a slice thickness of 0.6 mm in ultra-low-dose CT yielded much higher image noise, it was increased to 3 mm for comparison with standard-

dose CT images. All images were reconstructed in the transverse orientation, anonymized, and transferred for image analysis. The parameters of standard and ultra-low-dose CT are illustrated in Table 1.

### Image analysis

Image analysis was performed at lung window settings (window level, -500 HU; window width, 1500 HU), using a Syngo.via VB40 workstation. Image noise and image quality were assessed for all the patients.

### Objective image noise

Image noise was defined as the SD of attenuation measured in the air of the tracheal lumen above the aortic arch. CT attenuation was measured on 2 sets of lung window images (standard-dose CT, ultra-low-dose CT). Each region of interest (ROI) was delineated at the tracheal lumen above the aortic arch, and the size and location of the ROI were held constant across the 2-image series. The SD was measured 3 times, and the mean value was used in the analysis.

### Subjective visual assessment of image quality

All image sets (from which the patient information and image parameters had been deleted) were assigned random numbers, and image analysis was performed. Image quality was assessed by reference to both normal pulmonary structures and various pulmonary lesions.

### Normal pulmonary structures (airways, vessels, pleura, and sub-pleural lung)

On the 5-point scale: 1 – denotes excellent image quality without artifacts, 2 – denotes slight blurring that did

**Table 1.** Parameters of standard and ultra-low-dose computed tomography

Protocol	Standard HRCT thorax	Ultra-low-dose chest CT
Tube voltage	Automated tube voltage selection (reference kV: 120) CARE kV	100 kVp with 0.6-mm tin filter (100 Sn kVp)
Tube current time product	Automated tube current modulation (reference mAs: 122) CARE DOSE 4D	30 mAs
Collimation	192 × 0.6 mm	2 × 192 × 0.6
Gantry rotation time	0.5 seconds	0.25 seconds
Pitch	1.2	Ultra-long pitch (pitch = 3), TurboFlash mode
Data reconstruction	ADMIRE, strength level 3 with high spatial resolution kernel (Br59)	ADMIRE, strength level 3 with high spatial resolution kernel (BI57)
Slice thickness	0.6 mm	3 mm
Slice increment	0.6 mm	3 mm
Field of view	300 mm	300 mm
Pixel matrix	512 × 512	512 × 512

HRCT – high-resolution computed tomography, CT – computed tomography, ADMIRE – advance modelled iterative reconstruction

not compromise image assessment, 3 – denotes moderate blurring that slightly compromised assessment, 4 – denotes severe blurring rendering evaluation uncertain, 5 – denotes non-diagnostic image quality featuring strong artifacts.

Ratings of 1 to 5 were given to each of the above-mentioned normal pulmonary structures. The sum of these ratings was taken as the overall image quality score. The lower the score, the greater the diagnostic quality.

### Image quality assessment for various pulmonary lesions

Subjective visual assessment was performed on 12 categories of pulmonary lesions as follows: consolidation, ground-glass opacity, emphysema, bronchiectasis, bronchial wall thickening, pulmonary nodules, tree-in-bud appearance, interstitial septal thickening, pleural effusion, mosaic attenuation, fibrosis, and subsegmental atelectasis. The presence or absence of pulmonary lesions in each category was noted.

While assessing the solid pulmonary nodules, those that measured larger than 3 mm were selected and compared with standard and ultra-low-dose CT. The maximum number of nodules selected for any patient was 5, which were of the greatest dimensions. The image sets, from which patient information and image parameters had been deleted, were used to perform size measurements.

### Radiation dose estimations

The CT dose index volume ( $CTDI_{vol}$ ) and the dose length product (DLP) were recorded for each patient. The effective radiation dose (in millisieverts) was estimated by multiplying the DLP by a chest-specific conversion coefficient ( $0.014 \text{ mSv/mGy} \times \text{cm}$ ).

### Statistical analysis

The collected data were analysed using IBM-SPSS v27.0. Continuous variables were reported as mean  $\pm$  SD, and categorical variables were reported as frequencies or percentages. The measure of agreement in terms of nodule size among standard and ultra-low-dose CT was determined with the aid of Cohen kappa ( $\kappa$ ) statistics. A  $\kappa$  value of  $0 \leq \kappa \leq 0.20$  was regarded as poor,  $0.2 < \kappa \leq 0.4$  was fair,  $0.4 < \kappa \leq 0.6$  was moderate,  $0.6 < \kappa \leq 0.8$  was good, and  $\kappa > 0.8$  was excellent. Continuous variables, including radiation dose parameters, overall image quality scores, and image noise, were compared among the 2 groups (standard-dose CT and ultra-low-dose CT protocol) using a paired samples *t*-test. Bivariate analysis of discrete variables between both groups was done using the  $\chi^2$  test. The diagnostic accuracy of various pulmonary lesions in ultra-low-dose CT and standard-dose CT was measured and expressed using sensitivity, specificity, positive predictive value, and negative predictive value. A *p*-value of  $< 0.005$  was considered as statistically significant.

## Results

A total of 124 patients were included in the study (81 males and 43 females). The mean ( $\pm$  SD) age was  $55.78 \pm 15.63$  years (range, 12 to 85 years). The mean body mass index (BMI) was  $25.33 \pm 5.37$  (range, 16.29 to 39.00). Twenty-six patients had a BMI of 19.9 or less, 36 patients had a BMI of 20.0 to 24.9, 33 patients had a BMI of 25.1 to 29.9, and 29 patients had a BMI of  $\geq 30$ .

### Objective image noise

The mean  $\pm$  SD value of image noise (HU) in standard-dose CT was  $46.02 \pm 14.16$  and that of ultra-low-dose CT was  $81.73 \pm 17.71$ . Thus, the image noise was higher in ultra-low-dose CT, and there was a significant statistical difference between them.

### Overall image quality for normal pulmonary structures

The mean  $\pm$  SD value of the overall image quality score for standard-dose CT was  $5.0 \pm 0$  and that of ultra-low-dose CT was  $7.54 \pm 1.43$ . The mean score was higher for ultra-low-dose CT with a statistically significant difference as compared to the standard-dose CT. Therefore, it was clear that the standard-dose CT yielded higher image quality than the ultra-low-dose CT.

The overall image quality was determined to be diagnostic in 100% of standard CT images and 96.77% of ultra-low-dose CT images. Only 4 of the total 124 patients had non-diagnostic-quality images in ultra-low-dose CT scans; all these patients had a higher BMI ( $\geq 30$ ).

The mean  $\pm$  SD values of image quality in ultra-low-dose CT were  $6.57 \pm 0.57$ ,  $6.80 \pm 0.46$ ,  $7.81 \pm 1.01$ , and  $9.00 \pm 1.82$  in patients with BMI ranges of  $\leq 19.9$ , 20.0-24.9, 25.0-29.9, and  $\geq 30$ , respectively. On the other hand, the mean  $\pm$  SD values were the same ( $5.0 \pm 0$ ) in standard-dose CT among patients with all the above-mentioned BMI ranges.

It was observed that there was a proportional increase in the image quality score as the BMI range increased. Statistically significant differences were observed between both protocols among patients in all the BMI ranges.

### Subjective assessment of image quality for various pulmonary lesions

The findings of standard CT were used as a reference standard for comparison with the findings of ultra-low-dose CT. The imaging findings are summarized in Table 2.

- The order of sensitivity for various pulmonary lesions in ultra-low-dose CT was as follows: consolidation (97%) > pleural effusion (95%) > fibrosis (92%) > solid pulmonary nodules (91%) > subsegmental atelectasis (82%) > ground glass opacity (68%) > tree-in-bud appearance (50%) > pulmonary emphysema (41%) > bronchiecta-

**Table 2.** Analysis of various pulmonary lesions

S. No.	Lesion category	Presence of lesion in standard CT	Presence of lesion in ultra-low-dose CT	Sensitivity	Specificity	PPV	NPV
1	Solid pulmonary nodule	24	22	0.91	1	1	0.98
2	Tree-in-bud appearance	6	3	0.50	1	1	0.97
3	Ground glass opacity	44	30	0.68	1	1	0.85
4	Mosaic attenuation	11	1	0.09	1	1	0.92
5	Pulmonary emphysema	17	7	0.41	1	1	0.91
6	Bronchial wall thickening	25	9	0.36	1	1	0.86
7	Bronchiectasis	21	8	0.38	1	1	0.89
8	Consolidation	31	30	0.97	1	1	0.99
9	Fibrosis	13	12	0.92	1	1	0.99
10	Subsegmental atelectasis	39	32	0.82	1	1	0.92
11	Interstitial thickening	19	6	0.32	1	1	0.89
12	Pleural effusion	21	20	0.95	1	1	0.99

CT – computed tomography, PPV – positive predictive value, NPV – negative predictive value

sis (38%) > bronchial wall thickening (36%), interstitial thickening (32%), and mosaic attenuation (9%).


- The ultra-low-dose CT expressed higher sensitivity (> 90%) for the detection of solid pulmonary nodules of size > 3 mm and ≤ 20 mm (Figure 1), consolidation (Figure 2), pleural effusion (Figure 3), and fibrosis (Figure 4).
- No false-positive results were observed in ultra-low-dose CT.
- Specificity and positive predictive value (PPV) were 100% for all the 12 categories of pulmonary lesions in ultra-low-dose CT.
- In ultra-low-dose CT, most of the analysed pulmonary lesions yielded a higher negative predictive value (> 90%).

### Pulmonary nodule size assessment


A total of 65 solid nodules in 22 patients that were 3 mm to 20 mm in the longest dimension were measured in both standard CT and ultra-low-dose CT and compared. The measure of agreement was excellent ( $\kappa = 0.944$ ) and statistically significant ( $p < 0.001$ ) for standard CT and ultra-low-dose CT in size measurements of pulmonary nodules.

### Radiation dose estimation


The mean CT dose index volume (mGy), dose length product (mGy × cm), and effective radiation dose (mSv) showed statistically significant differences on both standard and




**Figure 1.** Axial sections of computed tomography (CT) showing detectability of solid pulmonary nodule (yellow arrow) by ultra-low dose CT. Solid pulmonary nodule was well seen on both standard and ultra-low dose CT scans



**Figure 2.** Axial sections of computed tomography (CT) showing detectability of consolidation (yellow arrow) by both standard ultra-low dose CT



**Figure 3.** Axial sections of computed tomography (CT) showing detectability of pleural effusion and non-detectability of emphysema by ultra-low dose CT. Pleural effusion was obvious on both standard and ultra-low dose CT scans. The mild paraseptal emphysematous changes (yellow arrow) as evident on standard CT were not appreciated on ultra-low dose CT



**Figure 4.** Axial sections of computed tomography (CT) showing detectability of fibrosis (yellow arrow) and non-detectability of bronchiectasis (yellow arrow) by ultra-low dose CT. Fibrosis is appreciated on both standard and ultra-low dose scans. But bronchiectatic changes were missed on ultra-low dose CT

**Table 3.** Comparison of radiation dose parameters between standard and ultra-low-dose computed tomography

	CT dose index volume (mGy)		Dose length product (mGy × cm)		Effective radiation dose (mSv)	
	Standard CT	Ultra-low-dose CT	Standard CT	Ultra-low-dose CT	Standard CT	Ultra-low-dose CT
Mean	7.87	0.10	285.44	3.63	3.99	0.050
Standard deviation	3.06	0	112.54	0.41	1.57	0.005
Mean difference	7.77		281.81		3.94	
t-value	28.25		27.88		27.88	
p-value	< 0.001		< 0.001		< 0.001	

ultra-low-dose CT protocols. The ultra-low-dose CT protocol was associated with a statistically significant reduction in radiation dose compared to standard CT (Table 3).

## Discussion

Owing to its greater diagnostic accuracy, there are several clinical indications for HRCT thorax, including screening for lung cancer and the diagnosis of benign lung parenchymal diseases such as pulmonary infection. The radiation risk incurred by patients is one of the major concerns that prevent its extensive use. Although low-dose CT was developed to overcome this, it also raises the concern of radiation risk in young patients and those undergoing repetitive CT examinations. It will be a boon if CT could be performed with the same diagnostic accuracy with a dose almost equivalent to that of a chest X-ray. This study was done to evaluate whether ultra-low-dose CT was able to serve this purpose.

In our study, the results showed good diagnostic accuracy of ultra-low-dose CT in the detection of > 3 mm sized solid pulmonary nodules with a sensitivity of 91%, a specificity of 100%, PPV of 100%, and negative predictive value (NPV) of 98%. Hence, it could identify fresh lesions in an otherwise normal case, as well as monitor changes in the number of already diagnosed cases (provided the nodule size was > 3 mm) (Figure 1).

As well as changes in the number, the size of the existing nodule is also an important parameter to be assessed in follow-up studies. Our observation in this study was that the concordance value was excellent ( $\kappa = 0.944$ ) and statistically significant ( $p < 0.001$ ) for both standard and ultra-low-dose CT in size measurements of a total of 65 pulmonary nodules. However, further studies with a higher sample size are needed to prove the efficacy of ultra-low-dose CT in size measurements.

Neroladaki *et al.* [1] in 2012, Huber *et al.* [16] in 2016, and Messerli *et al.* [6] in 2017 concluded that there was a good detection rate for solid pulmonary nodules in ultra-low-dose CT (mean effective radiation dose ~0.13 to 0.16 mSv). Gordic *et al.* [17] in 2013 reported similar results at a dose of 0.06 mSv. Kim *et al.* [18] in 2015 concluded that ultra-low-dose protocols with doses of 0.44 and 0.31 mSv also allowed size measurements.


Thus, ultra-low-dose CT can be utilized for screening and follow-up studies of primary or secondary lung cancers for the detection or exclusion of solid pulmonary nodules of size > 3 mm and monitoring of the number of > 3 mm sized solid pulmonary nodules. The average dose (0.05 mSv) offered by the ultra-low-dose protocol of our study deserves a particular mention because it was much lower than any of the previous protocols, with excellent diagnostic quality.

In addition to solid pulmonary nodules, diagnostic accuracy was also obtained for consolidation, pleural effusion, and fibrosis; all of them had sensitivity > 90%. Hence, their progression or resolution as a response to treatment could be identified using ultra-low-dose CT during follow-up. This is in concordance with the observation of Kim *et al.* [18] stating that ultra-low-dose CT could offer diagnostic-quality images for instances of increased attenuation.


Although lesions with good sensitivity like consolidation and pleural effusion are important findings of cardiogenic pulmonary oedema, ultra-low-dose CT has still not been recommended as an initial imaging modality. This is because other findings of this condition like smooth interstitial thickening and ground glassing (Figure 6) had a sensitivity of < 90%, and the diagnosis could be missed if the later ones were the sole CT manifestations, which may be the case in early stages. However, the ultra-low-dose CT could be used in follow-up imaging of this condition to look for resolution of consolidation.

In our study, there was a sensitivity of <90% for tree-in-bud nodules, ground glassing, interstitial thickening, bronchiectasis, bronchial wall thickening (Figure 5), subsegmental atelectasis, mosaic attenuation (Figure 5), and pulmonary emphysema. Hence, ultralow dose CT played no remarkable role in the evaluation of pulmonary infection with the above-mentioned presentations, diffuse infiltrative lung disease, and COPD.

Kim *et al.* [18] also quoted the limited role of ultra-low-dose CT in pulmonary emphysema, but for tree-in-bud nodules, ground glassing, interstitial thickening, bronchiectasis, and bronchial wall thickening the results were contradictory. This could be explained by the higher mean effective radiation dose (0.3 mSv) in their ultra-low-dose protocol in contrast to that used by us (0.05 mSv). Moreover, in their study the comparison was made between low-



**Figure 5.** Axial sections of computed tomography (CT) showing non-detectability of mosaic attenuation (yellow arrow) and bronchial wall thickening (blue arrow) by ultralow dose CT



**Figure 6.** Axial sections of computed tomography (CT) showing detectability of consolidation (yellow arrow) and non-detectability of ground-glassing (blue arrow) by ultralow dose CT

dose and ultra-low-dose CT, whereas in our study it was between standard-dose and ultra-low-dose CT.

Owing to the current COVID-19 pandemic, CT of the chest had been used as a screening modality in many institutions. Ground glassing was the most important CT manifestation, which unfortunately had a sensitivity of < 70% in ultra-low-dose CT, and hence this protocol was not recommended for imaging in this case. Shiri *et al.* [19] also suggested that ultra-low-dose imaging was not diagnostic of this condition.

In the context of image quality for normal pulmonary structures, it was found that ultra-low-dose CT offered diagnostic image quality in 96.77% of patients. Only 4 of the total 124 patients had non-diagnostic image quality, and all these patients had BMI > 30. This interpretation reinforced the conclusions of Lee *et al.* [20] and Ye *et al.* [21] that better results could be observed by using this protocol selectively for patients with BMI < 30.

In our study, the mean value of image noise in standard-dose CT was 46.02 HU whereas that of ultra-low-dose CT was 81.73 HU. Although the image noise offered by ultra-low-dose CT was statistically higher, the subjective image quality was diagnostic in most of the patients. Similar results were observed in a study by Kim *et al.* [18] in which statistically higher image noise was observed for ultra-low-dose CT when compared to low-dose CT (mean image noise: 76.8 HU and 102.3 HU vs. 47.7 HU). Lee *et al.* [20] also reported higher image noise for ultra-low-dose CT as compared to a CT protocol that was similar to our standard CT (mean image noise: 86.7 HU vs. 33.0 HU).

The CT dose index volume (mGy), dose length product (mGy × cm), and effective radiation dose (mSv) are the 3 parameters that were used for radiation dose estimation in our study. All 3 parameters were significantly lower for ultra-low-dose CT when compared to standard CT. This was evident from the much lower mean values

of CT dose index volume (0.1 mGy), dose length product (3.63 mGycm), and effective radiation dose (0.050 mSv) in ultra-low-dose CT when compared to 7.87 mGy, 285.44 mGycm, and 3.99 mSv, respectively, in standard CT. The mean effective radiation dose is around 30 times lower than a usual low-dose chest CT protocol. The mean radiation dose (0.05 mSv) in our ultra-low-dose CT protocol was inarguably low because it was very close to the dose of a chest X-ray (0.02 mSv) and also lower than previous studies of ultra-low-dose CT, e.g. Lee *et al.* [20] (0.29 mSv), Agostini *et al.* [22] (0.28 mSv), Macri *et al.* [3] (0.27 mSv), Yoon *et al.* [11] (0.17 mSv), Neroladiki *et al.* [1] (0.16 mSv), Messerli *et al.* [6] (0.13 mSv), and Burgard *et al.* [23] (0.1 mSv). This dramatic reduction in radiation dose was made possible by incorporating recent technologies like high-pitch dual-source scanning (turbo flash) and tin filtration in our ultra-low-dose protocol.

Despite the limited role of ultra-low-dose CT in diagnosing various clinical conditions, the drastically reduced radiation dose has made it a recommended tool wherever possible. In that context, ultra-low-dose CT could emerge as a revolutionary tool in thoracic imaging wherever indicated.

## Conclusions

The ultra-low-dose chest CT images acquired in a third-generation Siemens dual-source CT scanner employing 100 kVp and 30 mAs with turbo flash and tin filtration techniques offered a mean effective radiation dose of 0.05 mSv, which is almost equal to that of dose of a chest X-ray. Ultra-low-dose CT could be used for the detection and follow up of > 3 mm-sized solid pulmonary nodules. It is also useful for the follow-up imaging of consolidation and fibrosis to look for either progression or resolution. However, ultra-low-dose CT plays no remarkable role in evaluating pulmonary infection including COVID-19 infection, pulmonary oedema, diffuse infiltrative lung disease, and COPD.

## Conflict of interest

The authors report no conflict of interest.

## References

- Neroladiki A, Botsikas D, Boudabbous S, et al. Computed tomography of the chest with model-based iterative reconstruction using a radiation exposure similar to chest X-ray examination: preliminary observations. *Eur Radiol* 2013; 23: 360-366.
- Mayo-Smith WW, Hara AK, Mahesh M, et al. How I do it: managing radiation dose in CT. *Radiology* 2014; 273: 657-672.
- Macri F, Greffier J, Pereira FR, et al. Ultra-low-dose chest CT with iterative reconstruction does not alter anatomical image quality. *Diagn Interv Imaging* 2016; 97: 1131-1140.
- Greffier J, Pereira F, Hamard A, et al. Effect of tin filter-based spectral shaping CT on image quality and radiation dose for routine use on ultralow-dose CT protocols: a phantom study. *Diagn Interv Imaging* 2020; 101: 373-381.
- Ludwig M, Chipon E, Cohen J, et al. Detection of pulmonary nodules: a clinical study protocol to compare ultra-low dose chest CT and standard low-dose CT using ASIR-V. *BMJ Open* 2019; 9: e025661.
- Messerli M, Kluckert T, Knitel M, et al. Ultralow dose CT for pulmonary nodule detection with chest x-ray equivalent dose – a prospective intra-individual comparative study. *Eur Radiol* 2017; 27: 3290-3299.
- Newell JD, Fuld MK, Allmendinger T, et al. Very low-dose (0.15 mGy) chest CT protocols using the COPDGene 2 test object and a third-generation dual-source CT scanner with corresponding third-generation iterative reconstruction software. *Invest Radiol* 2015; 50: 40-45.
- The National Lung Screening Trial Research Team; Church TR, Black WC, Aberle DR. Results of initial low-dose computed tomographic screening for lung cancer. *N Engl J Med* 2013; 368: 1980-1991.
- Paks M, Leong P, Einsiedel P, et al. Ultralow dose CT for follow-up of solid pulmonary nodules: a pilot single-center study using Bland-Altman analysis. *Medicine (Baltimore)* 2018; 97: e12019.
- Ai T, Yang Z, Hou H, et al. Correlation of chest CT and RT-PCR testing in coronavirus disease 2019 (COVID-19) in China: a report of 1014 cases. *Radiology* 2020; 296: E32-E40.
- Yoon HJ, Chung MJ, Hwang HS, et al. Adaptive statistical iterative reconstruction-applied ultra-low-dose CT with radiography-comparable radiation dose: usefulness for lung nodule detection. *Korean J Radiol* 2015; 16: 1132.
- National Lung Screening Trial Research Team; Aberle DR, Berg CD, Black WC. The National Lung Screening Trial: overview and study design. *Radiology* 2011; 258: 243-253.
- Morsbach F, Gordic S, Desbiolles L, et al. Performance of turbo high-pitch dual-source CT for coronary CT angiography: first ex vivo and patient experience. *Eur Radiol* 2014; 24: 1889-1895.
- Solomon J, Mileto A, Ramirez-Giraldo JC, Samei E. Diagnostic performance of an advanced modeled iterative reconstruction algorithm for low-contrast detectability with a third-generation dual-source multidetector ct scanner: potential for radiation dose reduction in a multireader study. *Radiology* 2015; 275: 735-745.
- Gunn MLD, Kohr JR. State of the art: technologies for computed tomography dose reduction. *Emerg Radiol* 2010; 17: 209-218.
- Huber A, Landau J, Ebner L, et al. Erratum to: Performance of ultra-low-dose CT with iterative reconstruction in lung cancer screening: limiting radiation exposure to the equivalent of conventional chest X-ray imaging. *Eur Radiol* 2016; 26: 3653-3653.
- Gordic S, Morsbach F, Schmidt B, et al. Ultralow-dose chest computed tomography for pulmonary nodule detection: first perfor-

- mance evaluation of single energy scanning with spectral shaping. *Invest Radiol* 2014; 49: 465-473.
18. Kim Y, Kim YK, Lee BE, et al. Ultra-low-dose ct of the thorax using iterative reconstruction: evaluation of image quality and radiation dose reduction. *Am J Roentgenol* 2015; 204: 1197-1202.
  19. Shiri I, Akhavanallaf A, Sanaat A, et al. Ultra-low-dose chest CT imaging of COVID-19 patients using a deep residual neural network. *Eur Radiol* 2021; 31: 1420-1431.
  20. Lee SW, Kim Y, Shim SS, et al. Image quality assessment of ultra low-dose chest CT using sinogram-affirmed iterative reconstruction. *Eur Radiol* 2014; 24: 817-826.
  21. Ye K, Zhu Q, Li M, et al. A feasibility study of pulmonary nodule detection by ultralow-dose CT with adaptive statistical iterative reconstruction-V technique. *Eur J Radiol* 2019; 119: 108652.
  22. Agostini A, Floridi C, Borgheresi A, et al. Proposal of a low-dose, long-pitch, dual-source chest CT protocol on third-generation dual-source CT using a tin filter for spectral shaping at 100 kVp for CoronaVirus Disease 2019 (COVID-19) patients: a feasibility study. *Radiol Med (Torino)* 2020; 125: 365-373.
  23. Burgard CA, Gaass T, Bonert M, et al. Detection of artificial pulmonary lung nodules in ultralow-dose CT using an ex vivo lung phantom. *PLoS One* 2018; 13: e0190501.