



# Single-molecule bioelectronics

Jacob K. Rosenstein,<sup>1</sup> Serge G. Lemay<sup>2</sup> and Kenneth L. Shepard<sup>3\*</sup>

Experimental techniques that interface single biomolecules directly with microelectronic systems are increasingly being used in a wide range of powerful applications, from fundamental studies of biomolecules to ultra-sensitive assays. In this study, we review several technologies that can perform electronic measurements of single molecules in solution: ion channels, nanopore sensors, carbon nanotube field-effect transistors, electron tunneling gaps, and redox cycling. We discuss the shared features among these techniques that enable them to resolve individual molecules, and discuss their limitations. Recordings from each of these methods all rely on similar electronic instrumentation, and we discuss the relevant circuit implementations and potential for scaling these single-molecule bioelectronic interfaces to high-throughput arrayed sensing platforms. © 2014 Wiley Periodicals, Inc.

## How to cite this article:

*WIREs Nanomed Nanobiotechnol* 2015, 7:475–493. doi: 10.1002/wnan.1323

## INTRODUCTION

Living organisms are extraordinarily complex systems, possessing emergent features that arise from a deep hierarchy of chemical, mechanical, and electrical interactions among the tissues, cells, and molecules from which they are composed. Understanding each structural level poses unique practical challenges. At the smallest end of the scale, it can be challenging to acquire any useful information at all about an individual molecule, owing to its small physical size. Nanoscale dimensions make single molecules very susceptible to interactions with their environment and often lead to weak, stochastic signals that defy classical measurement techniques.

Despite these challenges, interfacing with single molecules offers unique opportunities to understand the basis of larger living systems, as well as to take advantage of the inherent spatial localization and heterogeneity that single-molecule data can offer. Equally valuable, this can lead to molecular assays using extremely small analyte volumes at low

concentration. Applications to DNA sequencing have tended to make headlines,<sup>1,2</sup> but at least as exciting are possibilities for single-molecule assays for nongenomic analytes for which no straightforward amplification protocol exists.

Single-molecule measurements reduce large continuous phenomena to discrete quantized systems. Molecules exist in integer numbers, and can be heterogeneous; molecules may exhibit discrete identifiable chemical or physical states, and they may shift between these states at discrete moments in time. Single molecules typically operate on faster timescales than ensemble systems, and the continuous output of ensemble measurements often represent the average of discrete molecular states, rather than a fundamentally continuous quantity.

Working with discrete variables can relax some of the experimental challenges brought on by weak signals; as long as discrete states can be distinguished from one another, there is complete information. However, the timing of signals remains a continuous variable, and in all recordings there will be a fundamental limit to the precision with which the timing of state transitions can be known.

Although many of the most popular single-molecule techniques are based on light microscopy, nonoptical pathways can offer considerable advantages. Electrochemical instrumentation can be cheap, fast, and small compared to most optical imaging systems; modern transistors have dimensions smaller than many biomacromolecules. Signal levels


\*Correspondence to: shepard@ee.columbia.edu

<sup>1</sup>School of Engineering, Brown University, Providence, RI, USA

<sup>2</sup>MESA+ Institute for Nanotechnology, University of Twente, Enschede, The Netherlands

<sup>3</sup>Departments of Electrical and Biomedical Engineering, Columbia University, New York, NY, USA

Conflict of interest: The authors have declared no conflicts of interest for this article.



**FIGURE 1** | An overview of available classes of single-molecule measurements. The signal path of all of these systems begins with conversion of a molecular state into a stream of ions, photons, electrons, or force, and the strength of this initial transduction often determines the overall signal quality. Eventually, the transduced signal is converted into an electronic current, which is then amplified and connected with other electronic systems. This review focuses on the highlighted platforms, which are physically small and entirely electrochemical.

are often larger for electrochemical measurements as compared to optical techniques, because chemical and electronic transport can be faster than photon generation in biochemical dyes. However, it can be more difficult to isolate single-molecule electrochemical signals from background activity.

Successful single-molecule transducers generally have the two complementary characteristics of *high gain* and *nanoscale spatial localization*. First, the transducers deliver some implicit amplification in the biophysical operation of a sensor, whereby single molecules influence nearby fluorophores, mobile ions, or electron channels. Second, this gain is spatially localized, reducing interference and background noise levels. Information regarding the state of a single molecule is converted into a stream of photons, ions, or electrons, and this transducer output is collected and amplified electronically, as illustrated in Figure 1. The weak signal levels mean that statistical errors and noise are primary concerns, and electronic measurement systems for single-molecule sensors tend to converge on a few circuit topologies that minimize measurement noise.

In this review, we discuss several varieties of single-molecule bioelectronic sensors, focusing on aqueous nonoptical platforms. We describe both the mechanisms of their interfaces with traditional microelectronics and the origin of the sensors' signal

amplitudes, temporal properties, noise performance, and potential for future large-scale integration. We exclude mass spectrometry<sup>3</sup> and atomic force microscopy,<sup>4</sup> reviews of which can be found elsewhere. In the *Ionic Channels* section, we discuss ion channels and nanopore sensors. In the *Nanowire and Nanotube Field-Effect Transistors* section, we review single-molecule field-effect transistors (FETs), primarily focusing on those based on single carbon nanotubes. The *Electron Tunneling Gaps* section presents tunneling gaps for electron transport through single molecules. In the *Redox Cycling and Electrocatalysis* section, single-molecule electrochemical methods based on oxidation-reduction reactions are reviewed. The *Low-Noise Current Amplifiers* section discusses the low-noise electronic current measurement systems traditionally used for single-molecule bioelectronics. The *Discussion* section concludes and comments on possibilities for future enhancements and applications.

## ION CHANNELS AND NANOPORES

### Single Ion Channel Recordings

Ion channel proteins are found in the cellular membranes of all organisms, and they are involved in a wide range of critical cellular processes.<sup>5</sup> In a historical

context, single-ion-channel recordings<sup>6</sup> may be considered the first successful single-molecule bioelectronic recordings of any type, and they are central to modern physiology.

As the field of electrophysiology progressed over the course of the 20th century, it was found that nerve cells can transmit signals via action potentials, electrical impulses which are propagated by a coordinated flux of dissolved ions in and out of a cell. Ions are not transported across cellular membranes uniformly; rather, the flux is concentrated through discrete protein ion channels which open and close ('gate') as a function of time. A wide range of ion channels have evolved which selectively transmit certain types of ions, and whose gating is sensitive to different environmental conditions such as electric fields,<sup>7</sup> ligand binding,<sup>8</sup> pH,<sup>9</sup> temperature,<sup>10</sup> or mechanical forces.<sup>11</sup>

The fact that the activity of single ion channels can be observed at all should not be taken for granted; even the existence of discrete ion channels was not definitively proven until the direct observation of single-channel gating currents<sup>12</sup> in the 1970s, and there are many proteins which still have not yet been observed as single channels. Even when a single channel can be isolated, recordings remain constrained by the noise floor of available experimental methods. In physiological conditions, single channels often have conductance on the order of femtosiemens (fS) or picosiemens (pS), and operate at membrane potentials of tens of millivolts (mV), producing single-channel currents on the order of picoamperes (pA) or lower. While single-channel recordings have been made of many higher-conductance ion channels, proteins that produce ion currents less than 0.1 pA are generally impractical to record at single-channel resolution.

For detailed treatments of ion channels, the reader is referred to numerous texts<sup>5,6</sup> and reviews<sup>7–11</sup> on these diverse molecules; here we simply wish to emphasize measurement considerations that make single channels electrically detectable. Several laboratory arrangements can facilitate recording ionic currents through single membrane proteins, including classical small-diameter glass pipettes,<sup>13</sup> microfabricated glass<sup>14</sup> and polymer<sup>15</sup> apertures, and scanning electrochemical microscopes (SECM).<sup>16</sup> From an electronic instrumentation perspective, even the largest single-channel currents are considered very small. Nonetheless, the picoampere currents through these channels represent a flux of millions ions per second, which is a huge amplification of the conformational change of a single transmembrane protein (this rate is thousands of times larger, for instance, than the number of photons emitted by the brightest organic fluorophores). The biophysical properties that make an

ion channel such an efficient modulator of localized ion flux are the same properties that allow its activity to be observed at single-molecule resolution.


## Nanopore Sensors

Nanopore sensors utilize similar physical principles as ion channel recordings, except that the ion channel itself is not under interrogation. Instead, ion flux is used as a proxy to detect other analyte molecules that may occupy the channel of the nanopore. Ideally, the shape of the ion channel (Figure 2(a) and (b)) does not fluctuate; rather, changes in its conductance can be attributed to the arrival of analyte molecules from the bulk solution (Figure 2(c)). This principle bears strong resemblance to Coulter counters<sup>21</sup> that are routinely used for automated cell counting and sorting. However, working with nanoscale molecules rather than microscale cells brings a new set of challenges.<sup>22</sup>

The basis of many of the limitations of nanopore sensors derives from the finite ionic currents that they produce; as with single-ion-channel recordings, nanoscale dimensions lead to weak signals which are challenging to measure. The simplest electrical model for a nanopore considers a cylindrical aperture through an insulating membrane, in which case the measured steady-state current ( $I_{\text{BIAS}}$ ) can be described by:

$$I_{\text{BIAS}} = V_{\text{BIAS}} \left( \frac{\pi d^2}{4b} + \frac{1}{d} \right) q \sum_i n_i \mu_i \quad (1)$$

where  $V_{\text{BIAS}}$  is the applied voltage,  $d$  is the diameter of the aperture,  $q$  is the elementary charge, and  $n_i$  and  $\mu_i$  are the number density and electrophoretic mobility, respectively, of dissolved cations and anions present in the electrolyte. Fits to measured data often find nanopores have an effective thickness  $b$  somewhat smaller than the actual membrane, as pores tend to have an 'hourglass' shape rather than a perfect cylinder. The  $1/d$  term arises from the hemispherical electrolyte 'access resistance' ( $R_A$ ) at the two sides of the pore<sup>23</sup> (Figure 3(d)). Although some implementations of nanopore sensors are actually repurposed ion channel proteins, nanopore sensors are not part of a living cell. As a consequence, they can be measured outside physiological conditions and, for practical reasons, are often tuned to achieve larger signal levels (high salt concentration and higher voltages). In addition, as nanopores are intended to be nongating and described primarily by their geometry, a range of pore constructions can be considered, including reconstituted ion channel proteins,<sup>18,28,29</sup> fabricated apertures in solid-state materials<sup>30–32</sup> (Figure 3), and




**FIGURE 2** | (a) Protein structure of alpha hemolysin ( $\alpha$ -HL), a bacterial toxin that self-assembles through cell membranes. (b) Top view of the  $\alpha$ -HL structure, showing its open center that allows ions to pass through cell membranes and can be used as a nanopore sensor. (c) Electric fields can induce single-stranded DNA molecules to pass through the  $\alpha$ -HL pore, transiently reducing the ionic conductance. (d) Enzymes such as DNA polymerase can be paired with nanopores to control the translocation velocity. (e) Polymerase ratchets a DNA molecule through a nanopore one base at a time, leaving behind a stepwise residual current. (f) The residual current is sensitive to the DNA base composition, such as the methylated-C substitution shown here. (g) Other biomolecular structures can also be used to construct nanopores, such as DNA origami [(a) and (b): Reprinted with permission from Ref 17]; (c): reprinted with permission from Ref 18; (d)–(f): Reprinted with permission from Ref 19; (g): Reprinted with permission from Ref 20].

other nanoscale constructs such as DNA origami<sup>33</sup> (Figure 2(g)). With these enhancements, the signals captured from nanopores can be considerably larger than for physiological ion channels. Bias currents greater than 100 pA are common in protein nanopores,<sup>34</sup> while solid-state nanopore currents typically exceed 1 nA.<sup>35</sup>

There are many potential applications of nanopore sensors, including single-molecule detection of toxins,<sup>36</sup> proteins,<sup>37</sup> and neurotransmitters.<sup>38</sup> Despite this richness of potential applications, in many discussions the word ‘nanopore’ has become synonymous with DNA sequencing. It has been the hope that nanopores can serve as the foundation of a future generation of single-molecule sequencing platforms, in which single intact nucleic acids

are threaded through nanopores while their base sequence is determined in real time.<sup>27</sup> There is real basis for this vision; the molecular-scale geometry of a nanopore can confine polymers to pass through the channel single-file, unfolded and in sequence. Recent commercial announcements suggest that the vision of nanopore sequencing has nearly been achieved.<sup>2</sup> However, while signal-to-noise ratios for counting single molecules can be quite robust, resolving features within individual molecules is often somewhat less reliable.<sup>39</sup>

Researchers continue to work on strategies to improve nanopores’ sensing performance, and these efforts can be categorized into several complementary goals: (1) improving chemical selectivity, (2) slowing transport or extending the time a molecule occupies



**FIGURE 3** | (a) An illustration of a double-stranded DNA molecule passing through a solid-state nanopore. Inset, a TEM image of a silicon nitride nanopore. (b) Current signals from translocations through SiNx nanopores can be as large as several nanoamperes. (c) Nanopore current blockades are sensitive to the orientation and position of a translocating molecule. (d) Simplified electrical equivalent circuit of a solid-state nanopore, highlighting the access resistance  $R_A$  and the membrane capacitance  $C_M$  [(a) and (b): Reprinted with permission from Ref 35; (c) and (d): Reprinted with permission from Ref 51].

a pore, (3) increasing signal levels, and (4) reducing noise levels.

Improving nanopores' chemical selectivity includes a wide scope of activities that emphasize chemical interactions between the pore and analyte molecules. Chemical probes can be attached to the openings<sup>40</sup> or inner walls of nanopores<sup>41</sup> that selectively bind target molecules, extending their translocation times relative to other molecules. There can also be significant chemical interactions between a molecule and the surfaces of the nanopore, which depend on the geometry and material of the pore.

The time that an analyte molecule spends in the channel determines the required temporal resolution of a nanopore recording, but improving time resolution involves a tradeoff with measurement noise. Electrophoretically driven DNA can travel through nanopores in less than 1  $\mu$ s per base, while noise-limited temporal resolution has often been on the order of tens of microseconds. All else being equal, slowing down translocation speeds would permit recordings to be made slower and with less error. Molecular motion can be slowed down by a number of mechanisms. One can decrease the temperature,<sup>42</sup> which both increases the viscosity of the electrolyte

and slows chemical kinetics. Electrolyte viscosity can also be increased with glycerol or other hygroscopic chemicals. Translocation times have been found to be extended for very small-diameter nanopores,<sup>43</sup> as polymers' conformations are limited by steric confinement and stronger van der Waals interactions with the surface of the pore. Molecular motion can also be modulated by improving the electrostatic conditions, through the modification of static charges on the walls of the nanopore, or introduction of tightly bound counterions<sup>44</sup> to modify the effective charge of the translocating molecule. Electronic control schemes involving biasing of integrated electrodes have also been proposed,<sup>45</sup> but fabricating these structures has proven difficult. Perhaps most promisingly, polymer translocation times have been extended dramatically through the use of enzymes such as polymerase, which can 'ratchet' a long DNA molecule through the pore one base at a time<sup>46</sup> (Figure 2(d)–(f)). This approach offers the intriguing combination of the sensitivity of nanopore sensors with the chemical specificity and slower kinetics of natural enzymes.

Increasing the signal levels from nanopores is largely a matter of modifying their geometries. The signal amplitude ( $\Delta I$ ) is distinct from the bias current

( $I_{\text{BIAS}}$ ), although often the signal amplitude is a fraction of the bias current, and the two cannot be adjusted independently. Increasing the bias voltage and electrolyte concentration are two obvious means of increasing the signal amplitude, but higher voltages can counterproductively increase the translocation velocities, and high salt conditions can be undesirable if they destabilize necessary polymer or enzyme structures. Increasing the diameter  $d$  of a nanopore can increase  $I_{\text{BIAS}}$ , but is less likely to improve  $\Delta I$ . The one geometrical change that tends to have unequivocally positive effects is the reduction of the effective thickness  $h$ ; this tends to improve both  $I_{\text{BIAS}}$  and  $\Delta I$ , while also improving capture efficiency and spatial resolution<sup>35</sup> (Figure 3(b)).

Efforts to decrease noise in nanopore recordings have several components, whose relative importance can change depending on the application. Interference from external energy sources is always a concern, but can generally be addressed with proper electronic isolation and electromagnetic shielding. Low-frequency  $1/f$  noise can be quite problematic in solid-state nanopores, although it can be mitigated to some degree with surface coatings, harsh acid exposure,<sup>41</sup> or high-voltage electrochemical cleaning.<sup>47</sup> However, fluctuating surface charges on the walls of the pore<sup>48,49</sup> may represent the limiting noise floor after other sources are reduced. Higher-frequency broadband noise is related to capacitive parasitics present in nanopore measurements, namely the capacitance of the thin dielectric membrane that supports the nanopore. Decreasing sensor parasitic capacitances can have a drastic impact on the noise amplitude for high-bandwidth nanopore recordings.<sup>50</sup> As the noise properties of nanopores themselves have improved, the noise contributions of the recording electronics have started to become a limiting factor, leading to increased interest in custom low-noise electronics for nanopore recordings.<sup>51–53</sup>

Both ion channels and nanopore sensors are able to produce impressively large signals from single-molecule activity thanks to their nanoscale dimensions and the fact that the positions of individual molecules affect the density of nearby mobile dissolved ions. Still, it is worth observing that the electrophoretic mobility of aqueous ions (e.g., at room temperature potassium ions have  $\mu_{\text{K}^+} \approx 7 \times 10^{-4} \text{ cm}^2 \text{ V}^{-1} \text{ s}^{-1}$ ) can be orders of magnitude slower than the mobility of electrons in solid-state materials ( $\mu_{\text{e}^-} \approx 1400 \text{ cm}^2 \text{ V}^{-1} \text{ s}^{-1}$  for silicon). If it could be arranged for single biomolecules to have electrostatic influence over the carrier density in an electronic conductor, the resulting electronic current signals could be much larger than for an ionic channel.

## NANOWIRE AND NANOTUBE FETS

There have been a number of demonstrations of electronic transistors whose gates are exposed to ionic solutions, allowing molecules from solution to have electrostatic influence over the transistor channel.<sup>54,55</sup> However, most available devices do not offer single-molecule sensitivity because their measured signals are the sum of influences from many uncorrelated molecules in solution. However, as nanoscale transistors approach single-molecule dimensions, opportunities emerge for direct field-effect detection of the electrical charge of individual analyte molecules.

Semiconducting nanowires have been used as field-effect biosensors,<sup>56</sup> where an electronic channel is sensitive to nearby charges in solution. Such nanowire FETs (NW-FETs) can be implemented using inorganic semiconductors either through bottom-up growth methods<sup>57</sup> or top-down lithography.<sup>58</sup> If these devices are appropriately functionalized they can be used for specific chemical detection.<sup>56</sup> In this arrangement, however, the entire exposed surface of the nanowire is equally sensitive, and background surface charge fluctuations can conceal signatures of analyte molecule binding. As a result, more elaborate frequency domain techniques have been used to try to discern single (or few) molecule behavior in a complex  $1/f$  noise background.<sup>59</sup> Alternately, nanowires can be constructed within or adjacent to a nanochannel to provide spatial localization, responding much like solid-state nanopores to local charges of single confined biomolecules.<sup>60</sup>

Another nanowire device that can reach single-molecule sensitivities is the carbon-nanotube FET (CNT-FET).<sup>61</sup> Carbon nanotubes can act as near-ideal one-dimensional electronic conductors, and can be either metallic or semiconducting depending on the particulars of the nanotube's carbon lattice.<sup>62</sup> Similarly to inorganic nanowire FETs, the geometry of a CNT-FET is particularly appealing for chemical sensors because the conducting channel can be directly exposed to aqueous electrolyte without any insulating layer. In addition, carbon nanotubes can be readily functionalized using organic chemistries.<sup>63</sup> Like a nanowire, a solution-gated CNT-FET can be sensitive to chemical conditions along its entire length, and detecting single molecules is aided by confining the transistor's sensitivity to a single point along the exposed channel. Fortunately, unlike the nanowire, this happens naturally when a nanotube is point-functionalized to tether a molecule under study because the tether directly affects the conducting channel of the device. In this configuration, with a

tethered probe molecule, we refer to the device as a single-molecule FET (smFET).

Fabrication of CNT-FETs begins with growth of carbon nanotubes using chemical vapor deposition. Long parallel nanotubes can be routinely grown, but it is difficult to predict their diameter and chirality. Atomic force microscopy can be used to identify nanotubes of a desired diameter (typically < 2 nm for this application), followed by photolithography to pattern metal contacts and etch away extraneous nanotubes. Several dozen transistors can be created along the length of a single carbon nanotube, increasing fabrication throughput<sup>64</sup>; however, the stochastic nature of the nanotubes' chirality and exact locations means that reliably fabricating these devices can be challenging.

The basic exposed gate device structure is shown in Figure 4(a). The electronic carrier density in the nanotube can be modulated by applying a voltage to the supporting silicon wafer, commonly referred to as the 'back gate'. The carbon nanotube can also be gated by controlling the electrochemical potential of the liquid electrolyte using an Ag/AgCl or platinum pseudo-reference electrode.<sup>68</sup> When a solution potential is applied, mobile ions in the electrolyte accumulate at the nanotube surface and induce carriers in the electronic channel. The distance over which the ions accumulate is described by the Debye length, which is < 1 nm for typical physiological buffer concentrations. This can be modeled as an electrochemical capacitance, and because the liquid gate capacitance is generally an order of magnitude larger than the back-gate capacitance, the nanotube is generally much more sensitive to the liquid gate.

Two techniques have been demonstrated that create a localized region of charge sensitivity along the nanotube. Each of these techniques locally disturb a nanotube's electronic structure through chemical modifications. The first approach electrochemically removes atoms from the carbon lattice,<sup>69</sup> disturbing its symmetry and creating a point defect with reduced conductance. In the second approach, a pyrene molecule tightly adsorbs to the surface of the nanotube through pi-pi stacking.<sup>70</sup> In each case, the chemical structure of the modification is then used to attach probe molecules that provide chemical specificity for analyte molecules of interest.

The point-defect approach is accomplished using conductance-controlled electrochemical etching of the tube in sulfuric acid. When the liquid gate voltage applied to a platinum solution electrode is reduced below a threshold voltage (approximately -1 V), the nanotube conductance initially decreases very slowly, followed by an abrupt jump. The applied voltage

is immediately removed when the abrupt jump is detected, stopping the oxidation. The oxidized device is exposed to potassium permanganate to create a carboxylic acid functional group on the defect site, which can be used to attach a known probe molecule.


The noncovalent functionalization is performed with exposure of the CNT-FETs to a solution of pyrene linker molecules, which stochastically adsorb to the nanotubes. Pyrene forms a tight noncovalent bond with the surface of the carbon nanotube, creating a localized electronic scattering site. The linker molecules also contain attachment chemistries for probe molecules or enzymes to be recorded.

Neither functionalization approach controls the probe location *a priori*, although it can be characterized afterward by scanning gate microscopy (SGM), which uses an electrically biased tip of an atomic force microscope as a liquid gate while recording the conductance of the nanotube. Figure 4(b) shows SGM images before and after functionalization; in a pristine carbon nanotube, transport is limited by Schottky barriers at the contacts, while after functionalization the gating sensitivity localized to a small region near the center of the device.

smFETs have been used to study single-molecule DNA hybridization<sup>71</sup> (Figure 4(c)) as well as lysozyme protein dynamics<sup>70</sup> (Figure 4(d)-(e)). For DNA studies, a single-stranded probe was attached to a point defect using a covalent amine to carboxylic acid coupling reaction, and the binding of complementary DNA target molecules resulted in two-level fluctuations in the tube conductance.<sup>71</sup> For enzyme recordings, lysozyme variants were genetically engineered with a cysteine residue which was then attached to a pyrene-maleimide linker molecule. Charge movements during the enzyme activity produced discrete conductance fluctuations in the smFET that were used to track the enzyme's activity.

As these signals are produced by the position or absence of individual analyte molecules, they tend to fluctuate between discrete states. The statistics of these transitions can resemble those of other systems, such as data from single-molecule fluorescence resonance energy transfer (FRET) studies<sup>72</sup> or random telegraph noise in nanoscale semiconductor devices.<sup>73</sup> These traces are naturally stochastic, but time traces can be analyzed using established techniques such as hidden Markov models.<sup>74</sup>

The signals produced by smFETs are remarkably large for single-molecule biosensors, thanks to the high mobility of the electrons travelling through the nanotube. The conductance of a CNT-FET can be hundreds of nS, producing signals as large as tens of nanoamperes (nA).



**FIGURE 4** | (a) An illustration of a liquid-gated field-effect transistor made from a single carbon nanotube. (b) Scanning gate microscopy images of a carbon-nanotube field-effect transistor (CNT-FET) before and after point functionalization. The CNT is chemically modified, so that its conductance is predominantly controlled by the electrostatics at a single defect site. (c) A time trace from a carbon nanotube modified with a DNA probe molecule. The CNT exhibits two discrete conductance levels, corresponding to the hybridization and melting of a complementary DNA molecule. (d) A single-molecule FET (smFET) trace with an attached enzyme, showing discrete conductance levels that correspond to the states within the activity of the enzyme. (e) A rendering of an enzyme tethered to an smFET [(a): Reprinted with permission from Ref 69; (b) and (c): Reprinted with permission from Ref 71; (d) and (e): Reprinted with permission from Ref 70].

Unfortunately, noise levels in smFETs tend also to be fairly high, limiting the usable dynamic range of the sensors. Most of the reported noise is flicker or  $1/f$  noise, which could either be a result of charges fluctuating in the electrolyte or in the insulator beneath the nanotube.<sup>75</sup> Studies have shown that  $1/f$  noise in carbon nanotube transistors is a strong function of device length,<sup>76</sup> suggesting simple scaling of the transistor geometry may reduce noise levels. The noise spectrum would be expected to reach a shot noise floor and a capacitive noise regime,<sup>77</sup> but this may be negligible until flicker noise levels are reduced.

## ELECTRON TUNNELING GAPS

In the single-molecule transducers discussed so far, the measured signal comes from a molecule's influence on adjacent ions (in the case of nanopores and ion channels) or an adjacent electron channel (in the case of NW-FETs). In the case of a CNT-FET, the electron channel is itself a single carbon nanotube molecule; one might intuit that another class of transducers might be possible in which electrons travel through the

analyte molecule itself, rather than through adjacent conducting channels.

Electron transport through single organic molecules has been investigated for several decades, often with ambition to use single molecules as building blocks in electronic circuits.<sup>78</sup> Although large-scale assembly of single-molecule circuits is not yet possible, scientific progress has been made toward understanding electron transport through molecular junctions.<sup>79</sup> The most common experiments utilize scanning probe microscopy (SPM), in which a sharp conductive tip is brought atomically close to a surface while the current through the tip is monitored. This technique can be used to measure electronic currents through single molecules, although the observed current is inevitably a function of the overall system, including the physics of the contacts between the metal tip and the molecule of interest.<sup>80,81</sup>

At the length scale of single molecules, electron transport is a quantum mechanical phenomenon, and electrons can interact with the energetic structure of the molecule. Metallic wires smaller than the Fermi wavelength of an electron have quantized conductance



$G_0 = 2e^2/h$  (approximately 77 microsiemens), and small organic molecules often exhibit conductances on the order of  $10^{-6} G_0$  to  $10^{-2} G_0$ . The upper ends of this range compare favorably to nanopores and NW-FETs, and could yield impressively strong signals if these structures can be reliably implemented.

Single-molecule electronic transport is often investigated using repeated formation and mechanical breaking of a nanoscale metallic wire.<sup>82</sup> Figure 5(a) shows this arrangement schematically; a nanoscale probe on a cantilever approaches a conducting surface, while electrically monitoring the current through the probe and optically measuring the deflection of the cantilever. When the tip is brought into contact with the surface and withdrawn (Figure 5(b)), it transiently forms a metal nanowire, with quantized conductance levels. Soon after the junction breaks, analyte molecules can span the gap, which is detected as nonzero conductances lower than  $G_0$  (Figure 5(c)). As this wire-pull can be repeated many times, histograms of these molecular currents can be made, and these histograms can correlate to the chemical structures and geometries of the molecules in solution.<sup>83</sup> In addition to using scanning probe platforms, tunneling gaps can be fabricated using mechanically controlled break junctions (MCBJ),<sup>84,85</sup> as well as with electromigration,<sup>86,87</sup> and these junctions can similarly measure single-molecule conductances.

Experimentally observed single-molecule conductances are highly variable, due to difficulties in reliable nanoelectrode formation, variable metal-to-molecule electrical contacts, and thermal fluctuations of the molecules themselves.<sup>85</sup> The mechanisms of the break junction procedure mean that measured currents can change by eight orders of magnitudes within seconds, making accurate instrumentation a challenge.<sup>88</sup> Molecular contacts vary with the atomic arrangement of the surface and the orientation of the molecule; although some molecules' conductance histograms can have well-defined peaks, they tend to have fairly wide variances. Covalent bonds to the metal surface can improve these contacts,<sup>89</sup> as can using surface-bound probe molecules which in turn bind the analyte.<sup>90</sup> In experiments which maintain tunneling probes at a constant distance for extended periods of time, stochastic current spikes are observed, which are attributable to molecules diffusing in and out of the sensing region, and reorienting within the gap (Figure 5(d)). Interestingly, statistical features of these stochastic spikes can be used to distinguish between molecules.<sup>91,92</sup>

Molecular conductance has been proposed as a mechanism for DNA sequencing,<sup>27,93,94</sup> using transverse tunneling currents through nucleotides in a

larger intact DNA molecule (Figure 5(e)). While this remains an open area of research, tunneling currents have been demonstrated to differentiate nucleotide monomers in solution,<sup>84,90</sup> and nucleotides within oligomers.<sup>95,96</sup> Electron transport can also detect amino acids<sup>91</sup> and proteins spanning nanogaps.<sup>87</sup>

## REDOX CYCLING AND ELECTROCATALYSIS


A final electronic single-molecule detection strategy becomes available when the analyte molecule itself is electrochemically active. Under the right conditions, electrochemical reduction or oxidation (*redox*) reactions of single analyte molecules can be converted into a measurable current at electrodes that connect to external electronic circuitry.

A typical reversible redox reaction involves the addition of one or a few electrons to a molecule in solution (reduction) or the reverse process of withdrawing electrons (oxidation), as summarized by the reaction



Here O and R represent different charge states of the analyte and are referred to as the oxidized and reduced forms of the molecule, respectively, while  $n$  is the number of electrons transferred (typically,  $n = 1$  or  $2$ ). Eq. (1) highlights the challenge inherent to single-molecule measurements based on electrochemistry: in order to electronically resolve individual electron-transfer processes, one would need to directly detect  $n$  electrons, an impossible feat in near-physiological conditions at room temperature. Thus, in order electrochemically detect a single redox molecule, one analyte molecule must participate in *many redox reactions*. To date, two main schemes have emerged to achieve this charge amplification, each suitable for a specific class of analytes.

The first approach, *redox cycling*, can be applied to small analyte molecules that undergo reversible electrochemical reactions, as typified by Eq. (1); this includes several important hormones and neurotransmitters. The core concept of redox cycling is to repeatedly reduce and oxidize each target molecule such that it oscillates between the O and R states. This repeated oxidation and reduction can be considered to be *amplification*, and the total amount of charge transferred is  $ne$  times the number of oxidation-reduction cycles, where  $-e$  is the electron charge. The number of oxidation and reduction reactions can in principle be arbitrarily large, but the usefulness of redox cycling will depend on achieving *rapid* cycling between the two states.



**FIGURE 5** | (a) A schematic of a typical scanning probe microscopy (SPM) experiment for single-molecule conductance measurements. A conductive cantilever is brought in contact with a conductive surface, under an electrical bias voltage ( $V_b$ ), while the current ( $I$ ) through the probe is monitored. The deflection of the cantilever can be monitored using a reflected laser to infer the applied mechanical force. (b) The probe is brought repeatedly into contact with the surface and withdrawn. As the probe withdraws, it forms a transient single-atom wire junction. (c) Monitoring the current through the probe during its withdrawal displays quantized conductance, and after the metal junction breaks, analyte molecules can enter the junction. Histograms from many wire-pull experiments can highlight the conductances of single analyte molecules. (d) If the SPM tip is held at a constant distance from the surface, analyte molecules can diffuse into the junction, creating transient spike trains. (e) A rendering of a proposed DNA sequencing platform utilizing nanoelectrodes functionalized with recognition molecules that record tunneling currents through a DNA molecule as it passes through a nanopore [(a) and (b): Reprinted with permission from Ref 24; (c): Reprinted with permission from Ref 25; (d): Reprinted with permission from Ref 26; (e): Reprinted with permission from Ref 27].

Efficient redox cycling can be achieved using two closely spaced electrodes biased at different potentials, such that the reduction reaction is favored at one electrode and the oxidation is favored at the second electrode. Each target molecule then acts as an electron shuttle, removing charge from one electrode and delivering it to the second electrode during each

oxidation–reduction cycle. The resulting current per molecule is

$$i_0 = \frac{ne}{\tau} \quad (3)$$

where  $\tau$  is the average time for performing one cycle.<sup>97</sup> Clearly, in order to achieve large signal levels the cycling time must be made as small as possible. As

under physiological conditions electric fields are predominantly screened by mobile salt ions, the only means of transport between electrodes for individual molecules is Brownian motion (diffusion). This dictates the geometry required for effective redox cycling: because the expected diffusion time between two (parallel, planar) electrodes is given by  $\tau/2 = z^2/2D$ , where  $D$  is the diffusion coefficient of the analyte and  $z$  is the distance between the electrodes, the spacing between the two electrodes should be made as small as possible. For a typical small molecule with  $n=1$  and  $D=0.5 \times 10^{-9} \text{ m}^2/\text{s}$ , coupled with an electrode spacing of 10 nm, Eq. (3) thus predicts an average cycle time of  $\sim 0.2 \mu\text{s}$  and a corresponding diffusion-limited current of  $\sim 1 \text{ pA}$ . This signal level is readily measurable with a bandwidth of order 1 kHz. For an electrode spacing of 1  $\mu\text{m}$ , however, the current drops to  $\sim 0.1 \text{ fA}$ , well below the noise floor of most instrumentation. This simple analysis thus highlights the key requirement for single-molecule detection by redox cycling: electrode spacing must be maintained on nanometer scales.

Early realizations of single-molecule detection by redox cycling focused on scanning probe electrochemical methods, in which a flat electrode tip was brought to within 10 nm of a metallic surface, with the tip and surface serving as the two electrodes.<sup>97–99</sup> Step-like features in the current were observed that were attributed to individual redox molecules entering the detection region and undergoing redox cycling before exiting again. These first experiments were more recently extended to recessed electrodes immersed in liquid mercury, which allows trapping a very small volume of solution between a solid electrode and the mercury bath.<sup>100</sup> In this case, current levels were observed that were consistent with single or small numbers of molecules being present in the trapped volume.


Over recent years, a parallel effort has also been underway to develop microfabricated solid-state devices that create the conditions required for efficient redox cycling.<sup>101</sup> These so-called nanogap transducers consist of a channel or chamber whose floor and roof host independently biased electrodes separated by  $<100 \text{ nm}$  (Figure 6(a)). Thanks to their reliance on modern lithographic techniques, these devices are suitable for parallelization and integration into more complex systems. Using nanogap devices, discrete signals whose dynamics were consistent with Brownian random walks were reported, providing quantitative evidence for single-molecule sensitivity<sup>101,65,66</sup> (Figure 6(b)). The current per molecule was however found to be typically a factor 4–10 lower than predicted from pure diffusion. This discrepancy was

attributed to transient adsorption of the analyte molecules to the electrodes, increasing the shuttling time beyond the time expected for free diffusive motion. While adsorption can be mitigated by adjusting the surface properties of the electrode, it remains unclear whether it can be fully eliminated; at the present time, currents on the order of  $\sim 10 \text{ fA}$  per molecule remain typical, a significant challenge for subsequent electronic circuitry.

Given these constraints, what sets the noise level and SNR for redox-cycling detection? An unavoidable limitation is of course the shot noise generated by the redox cycling process itself. It is important to note, however, that the sequential nature of redox cycling insures that it is not a Poisson process, and therefore that the normal prescription for shot noise,  $S = 2neI$ , where  $I$  is the total current, only represents an upper bound for the shot noise. Other important noise sources may involve the variable electrochemical potentials of single molecules. To our knowledge, and despite ongoing efforts, no detailed theoretical prediction yet exists for the noise spectral density resulting from redox cycling.

A second, potentially more damaging source of noise results from the reliance on Brownian motion for charge transport. Freely diffusing molecules can repeatedly enter and leave a nanogap detection volume, leading to corresponding telegraph-like noise in the measured redox cycling current.<sup>101,102</sup> Dominating this process are very short events that cannot be resolved within the bandwidth available for fA-level measurements, and which therefore manifest themselves as background noise. It has been argued<sup>103</sup> that this so-called *diffusion noise* can be suppressed using advective flow: molecules transiting through a nanogap detection volume of sufficient length necessarily undergo redox cycling for enough time to allow detection. Experimental verification of this analysis has not, however, yet been realized.

Another means of achieving electronic single-molecule detection may come through *electrocatalysis*. Rather than detecting redox molecules themselves, electrocatalysis offers an opportunity to detect molecules which catalyze redox reactions involving other nearby species. Many redox reactions are very slow at bare electrodes and require the presence of a suitable catalyst to achieve significant turnover. A conducting electrode immersed in a solution that contains an electrochemically active species may, therefore, not record a current until a catalyst is also present. This provides a route for detecting the presence of individual catalysts at the surface of the electrode via the catalytic electrochemical current, as illustrated in Figure 6(c). Although electrocatalysis




**FIGURE 6** | (a) An illustration of a redox cycling measurement, in which a single analyte molecule diffuses between two closely spaced electrodes and is alternatively oxidized and reduced. (b) Representative single-molecule data from a redox-cycling experiment. As molecules diffuse in and out of the region between the two electrodes, the measured current changes. (c) An illustration of electrocatalysis, in which surface-bound catalysts enable electrochemical reactions at rates far exceeding those with a bare electrode surface. (d) Current traces from the arrival of single nanoparticle catalysts at an electrode surface [(a): Reprinted with permission from Ref 65; (b): Reprinted with permission from Ref 66; (c) and (d): Reprinted with permission from Ref 67].

also quantifies redox currents at an electrode, it is different from redox cycling; rather than relying on the transport of redox molecules between two electrodes, the catalyst instead facilitates electrochemical reactions of other molecules in solution. One catalyst can interact with a great many redox molecules, without relying on transport between two electrodes.

This concept has been explored in recent years in the form of real-time study of the binding of (catalytic) nanoparticles to an electrode surface: each nanoparticle binding to the surface increases the catalytic ability of that surface, leading to a step in the electrochemical current detected at the electrode.<sup>67,104–106</sup> To date, the main focus has been on proof-of-concept demonstrations based on metal or metal-oxide nanoparticles, which can catalyze currents of order 10–100 pA upon attachment to a (noncatalytic) electrode (Figure 6(d)).

In theory, the same concept could be extended to observe individual catalytic *molecules*, such as redox enzymes, which are proteins that catalyze electrochemical reactions. The enzyme glucose oxidase is a particularly well-known example of this class of biocatalysts due to its broad deployment in glucose

sensing applications. Through appropriate chemical modification, redox enzymes can be ‘wired’ to an electrode surface, the electrode then providing (or withdrawing) electrons resulting from a catalytic reaction taking place in solution and providing a direct amperometric measurement of the enzymatic activity. This approach has been successfully applied in bulk toward fundamental understanding of the mechanistic properties of redox enzymes, a technique known as *protein film voltammetry*.<sup>107,108</sup> Attempts to scale this approach down to single molecules, however, have proven challenging. The difficulty arises from the fact that the catalytic rate of redox enzymes is rather slow, reaching only  $10^1$ – $10^4$  catalytic cycles per second even under optimal conditions, which offers little or no advantage over fluorescent methods.<sup>109</sup> This corresponds to maximum currents of only  $\sim 1$  fA per molecule for even the fastest redox enzymes, making electronic measurement impractical. So far, single-enzyme detection using this approach has not been demonstrated, and the lowest detection level has been in the range of 10–50 enzymes (hydrogenase molecules reducing protons to molecular hydrogen and collectively generating  $\sim 20$  fA of electrocatalytic current<sup>110</sup>).



**FIGURE 7** | (a) Electronic nanosensors are often arranged in a configuration that measures their current ( $I_{\text{SENSOR}}$ ) while applying a constant voltage ( $V_{\text{BIAS}}$ ) across the device. Variance in the measurement can be represented by an input-referred equivalent noise source, whose current ( $I_{\text{NOISE}}$ ) is added to the desired value. (b) A simplified schematic of a low-noise current preamplifier connected to a device modeled as a resistor,  $R_S$ , with parasitic capacitances ( $C_{SP}$  and  $C_{SS}$ ). The operational amplifier has its own input capacitance ( $C_I$ ) and feedback capacitance ( $C_F$ ). Noise sources are also noted in the figure including the input-referred voltage noise of the amplifier ( $v_{n-A}$ ), the thermal noise of the feedback resistor ( $i_{n-RF}$ ), and the noise of the sensor itself ( $i_{n-S}$ ). (c) Representative current noise power spectral density at the input of the current preamplifier. Below the frequency  $f_1$ , the noise is dominated by the flicker noise of the transducer. Between  $f_1$  and  $f_2$ , the thermal noise of the transducer and the feedback resistor of the preamplifier set the noise floor. Flicker noise in the amplifier or imperfect dielectrics in the transducer provides a  $\propto f$  slope between  $f_2$  and  $f_3$ . Above  $f_3$ , the input-referred noise acquires a  $\propto f^2$  slope due to the noise of the amplifier acting through  $\Sigma C_{IN}$ . This noise can be greatly reduced by scaling down the physical dimensions of the sensors, fluidics, interconnects, and electronic circuits. (d) A micrograph of a low-noise current preamplifier implemented in a modern CMOS process. (e) An illustration of a lipid bilayer reconstituted on the surface of an integrated circuit amplifier chip, reducing interconnects and relevant capacitances [(c) and (d): Reprinted with permission from Ref 51; (e): Reprinted with permission from Ref 117].

## LOW-NOISE CURRENT AMPLIFIERS

Almost without exception, the signals produced by single-molecule electronic sensors are amplified with voltage-clamp current preamplifiers, which apply a constant voltage across the sensor while measuring the resulting current (Figure 7(a)). This holds true for ion channels, nanopores, smFETs, electron tunneling junctions, and redox cycling, as discussed in this review, as well as many photodiode-based optical systems. All of these transducers share the properties of having very low electrical conductance, and producing weak current signals.<sup>111</sup>

These current preamplifiers are based on an operational amplifier with a high-impedance feedback network (Figure 7(b)). The amplifier maintains a constant voltage at the negative op amp terminal, while the output voltage is linearly proportional to the signal current  $I$ . This makes the system more tolerant of stray input capacitances ( $C_{IN} = C_I + C_{SS} + C_{SP}$ ); for a capacitor,  $I = C \frac{dV}{dt}$ , and thus the current through  $C_{IN}$  is nominally zero, although  $C_{IN}$  will interact with the amplifier to contribute higher-frequency noise density (Figure 7(c)).

The high-impedance feedback network may consist of a resistor, capacitor, or both in parallel. As all resistors contribute thermal current noise ( $S_I = 4kT/R \text{ A}^2/\text{Hz}$ ), the largest tolerable  $R_F$  is preferable. Feedback resistances ranging from several M $\Omega$  up to tens of G $\Omega$  are common, although these high-value resistors tend to be physically large and limit the largest measurable signal to  $I < V_{DD}/R_F$ , where  $V_{DD}$  is the supply voltage. Purely capacitive feedback yields the lowest noise, but the accumulated charge on  $C_F$  must be periodically discharged through a reset switch.

In a well-optimized system, at kHz bandwidths the input-referred noise spectral density tends to be on the order of 1–100 fA/ $\sqrt{\text{Hz}}$ , implying fA standard deviations for slow (<100 Hz) signals, pA standard deviations for moderately fast (<100 kHz) signals, and pA-to-nA deviations for MHz bandwidths.<sup>52,112</sup>

Any vision of a small parallelized single-molecule bioelectronic sensor array will need an appropriately sized multichannel electronic interface.<sup>113</sup> Discrete-component current amplifiers are generally on the scale of several cubic centimeters or more per channel, and systems with as many as 384 parallel low-noise current measurements are commercially available as

large rack-mount instruments.<sup>114</sup> Modern complementary metal oxide semiconductor (CMOS) technology offers a relatively straightforward path to shrinking the full electronic signal chain to submillimeter dimensions<sup>115</sup> (Figure 7(d)) This can lead to shorter interconnects, less capacitance (leading to lower noise), and large arrays of closely integrated sensors.<sup>116</sup> To take full advantage of these reduced interconnects can demand reconfiguring the fluidics and sensors around the integrated circuits themselves; e.g., introducing the amplifier into a nanopore fluid cell or assembling sensors directly onto the surface of the semiconductor chip. One such arrangement is shown in Figure 7(e), with lipid bilayers reconstituted directly onto microfabricated supports on the surface of a CMOS amplifier chip.<sup>117</sup>

There have been several demonstrated integrated CMOS current preamplifiers for single-molecule bioelectronic sensors,<sup>51,53,118–120</sup> in addition to the long and established history of integrated low-noise preamplifiers for photodiodes<sup>121</sup> and high-energy particle detectors<sup>122</sup> (additionally, there are ongoing commercial efforts which incorporate ASICs with nanopore measurements,<sup>2,39</sup> although at this time few specific technical details are publically available). Integrated amplifier designs face similar tradeoffs as discrete designs, with different constraints. Integrated semiconductor processes often do not offer junction field-effect transistors (JFETs), leaving MOS transistors as a next-best substitute. Integrated semiconductors also do not have high-value resistors available for the feedback network. There are several imperfect resistive-feedback substitutes, such as weakly inverted transistors or active low-transconductance stages,<sup>120</sup> but these devices introduce shot noise which is not present in discrete resistors. Capacitive feedback is still the lowest-noise topology, if an appropriate reset scheme can be implemented. The simplest arrangement is a single reset switch,<sup>112</sup> which is also the arrangement of discrete capacitive-feedback amplifiers such as the popular Axopatch 200B (Molecular Devices, Sunnyvale, CA, USA). Discharging the feedback network through a switched-capacitor network is also a possibility, as is arranging a more complicated system which alternates the polarity of the first stage in order to charge and discharge a feedback capacitor.<sup>123</sup>

Integrated electronics provide a path for higher throughput and smaller dimensions, but ultimately they can do no better than the signal-to-noise ratio of the single-molecule transducer. Reductions in parasitic capacitance and careful customization of preamplifier performance may see an additional order of magnitude reduction in noise floors from today's state-of-the-art; beyond this, we will need to look for

improvements in the specificity, sensitivity, and gain of the sensors themselves.

## DISCUSSION

The usefulness of single-molecule data and the low cost of nonoptical platforms will continue to drive investments and advances in single-molecule bioelectronic platforms. One path for improvement is simply scaling up single-molecule electronic platforms for higher throughput and higher levels of automation. Single-molecule data tend to be noisy and stochastic and often require more manual analysis than is tolerable outside a research environment. Parallelized systems, along with appropriate software and algorithms, will help to reduce that tradeoff, allowing more data to be collected and making lower-yield experiments more feasible.

Although in the abstract, single-molecule measurements represent the ultimate in sensitivity, in many applications they are impractical or unnecessary. Single-molecule sensing will see its growth as a key tool for applications that work with dilute analytes; applications with subtly heterogeneous samples; and studies of the biophysics of single-molecule interactions. One of the main limitations of single-biomolecule measurements is that they often support a relatively small range of analyte concentrations. At low concentrations, the analyte molecules are observed infrequently, and at high concentrations molecules appear too often to be differentiated from one another.

DNA sequencing remains the highest-profile commercial application in this space, and the scientific community has come to appreciate the value of long-read single-molecule sequence data.<sup>124</sup> Nanopores appear poised to be the first electronic single-molecule platforms to be widely used outside of a laboratory; there are a number of companies working to commercialize nanopore DNA sequencing, to varying degrees of success.<sup>2,125,126</sup> Ultimately, nanopore DNA sequencing may prove to be just the leading edge of a broader class of single-molecule electronic biosensors.

CMOS-integrated smFETs naturally lend themselves to large dense arrays, with close integration of the measurement electronics. In some sense, the smFET is a nanoscale equivalent to microscale ion-sensitive FETs (ISFETs), which have seen recent success in the DNA-sequencing market.<sup>127</sup> smFETs are smaller and faster than ISFETs, but their construction in large arrays remains unproven. The primary limitation to achieving arrays of smFETs comes from the initial growth and positioning of the carbon nanotubes,

and from high flicker noise. Fabricating these devices remains a challenging and manual process.

Redox-cycling has at least one practical advantage, in that it is a direct extension of established electrochemical methods. The lower gain of redox cycling and electrocatalysis may provide an advantage in dynamic range, allowing a single platform to scale from single-molecule to ensemble measurements. Unfortunately, achieving high cycling efficiency requires single-nanometer fluidic channels that are challenging to fabricate.

For tunneling gaps, at the moment their extreme sensitivity is as much of a detriment as an asset. Like many nanoscale devices, they suffer from fabrication and reliability challenges, but single-molecule tunneling measurements may yield insights into molecular electronics more broadly.<sup>81</sup> Tunneling sensors

may also be combined with nanopores<sup>86,128</sup> or nanochannels<sup>129</sup> that confine analyte molecules' locations and orientations. As tunneling currents depend strongly on the metal-to-molecule interface, future advances in tunneling gaps may incorporate as much chemistry as nanofabrication.<sup>130,26</sup>

In concept, electronic single-molecule platforms are not limited spatially by light diffraction, and are not limited temporally by photon emission. However, in practice, their implementation can be held back by optical lithography and the dimensions of metal interconnects and fluidic channels. Above all, yield and reliability remain the primary obstacles. Their future is bright, but nanoscale electronic single-molecule biosensors can be just as fickle as the single molecules that they measure.

## ACKNOWLEDGMENTS

This work was supported in part by the National Institutes of Health under grants U19-AI109761, R01-HG006882, R01-HG006879, and R01-GM107417.

## REFERENCES

1. Eid J, Fehr A, Gray J, Luong K, Lyle J, Otto G, Peluso P, Rank D, Baybayan P, Bettman B, et al. Real-time DNA sequencing from single polymerase molecules. *Science* 2009, 323:133.
2. Eisenstein M. Oxford nanopore announcement sets sequencing sector abuzz. *Nat Biotechnol* 2012, 30:295–296.
3. Fenn J, Mann M, Meng C, Wong SF, Whitehouse CM. Electrospray ionization for mass spectrometry of large biomolecules. *Science* 1989, 246:64.
4. Hansma H, Hoh J. Biomolecular imaging with the atomic force microscope. *Annu Rev Biophys Biomol Struct* 1994, 23:115.
5. Hille B. *Ion Channels of Excitable Membranes*. 3rd ed. 2001.
6. Sakmann B, Neher E. *Single-Channel Recording*. 2009.
7. Armstrong CM, Hille B. Voltage-gated ion channels and electrical excitability. *Neuron* 1998, 20:371–380.
8. Changeux JP, Devillers-Thiéry A, Chemouilli P. Acetylcholine receptor: an allosteric protein. *Science* 1984, 225:1335–1345.
9. Chen CC, England S, Akopian AN, Wood JN. A sensory neuron-specific, proton-gated ion channel. *Proc Natl Acad Sci U S A* 1998, 95:10240–10245.
10. Clapham DE. TRP channels as cellular sensors. *Nature* 2003, 426:517–524.
11. Martinac B. Mechanosensitive ion channels: molecules of mechanotransduction. *J Cell Sci* 2004, 117:2449–2460.
12. Neher E, Sakmann B. Single-channel currents recorded from membrane of denervated frog muscle fibres. *Nature* 1976, 260:799.
13. Hamill OP, Marty A, Neher E, Sakmann B, Sigworth FJ. Improved patch-clamp techniques for high-resolution current recording from cells and cell-free membrane patches. *Pflugers Arch* 1981, 391:85–100.
14. Fertig N, Meyer C, Blick R, Trautmann C, Behrends J. Microstructured glass chip for ion-channel electrophysiology. *Phys Rev E* 2001, 64:1–4.
15. Mayer M, Kriebel JK, Tosteson MT, Whitesides GM. Microfabricated teflon membranes for low-noise recordings of ion channels in planar lipid bilayers. *Biophys J* 2003, 85:2684–2695.
16. Novak P, Gorelik J, Vivekananda U, Shevchuk AI, Ermolyuk YS, Bailey RJ, Bushby AJ, Moss GWJ, Rusakov DA, Klenerman D, et al. Nanoscale-targeted patch-clamp recordings of functional presynaptic ion channels. *Neuron* 2013, 79:1067–1077.
17. Song L, Hobaugh MR, Shustak C, Cheley S, Bayley H, Gouaux JE. Structure of staphylococcal x-hemolysin, a heptameric transmembrane pore. *Science* 1996, 274:1859.
18. Kasianowicz JJ, Brandin E, Branton D, Deamer DW. Characterization of individual polynucleotide

- molecules using a membrane channel. *Proc Natl Acad Sci U S A* 1996, 93:13770–13773.
19. Laszlo AH, Derrington IM, Brinkerhoff H, Langford KW, Nova IC, Samson JM, Bartlett JJ, Pavlenok M, Gundlach JH. Detection and mapping of 5-methylcytosine and 5-hydroxymethylcytosine with nanopore MspA. *Proc Natl Acad Sci U S A* 2013, 110:18904–18909.
  20. Langecker M, Arnaut V, Martin TG, List J, Renner S, Mayer M, Dietz H, Simmel FC. Synthetic lipid membrane channels formed by designed DNA nanostructures. *Science* 2012, 338:932–936.
  21. Graham M. The Coulter principle: foundation of an industry. *J Assoc Lab Autom* 2003, 8:72–81.
  22. Wanunu M. A journey towards DNA sequencing. *Phys Life Rev* 2012, 9:125–158.
  23. Hall JE. Access resistance of a small circular pore. *J Gen Physiol* 1975, 66:531–532.
  24. Nef C, Frederix PLTM, Brunner J, Schönenberger C, Calame M. Force-conductance correlation in individual molecular junctions. *Nanotechnology* 2012, 23:365201.
  25. Xu B, Tao NJ. Measurement of single-molecule resistance by repeated formation of molecular junctions. *Science* 2003, 301:1221–1223.
  26. Lindsay S, He J, Sankey O, Hapala P, Jelinek P, Zhang P, Chang S, Huang S. Biochemistry and semiconductor electronics—the next big hit for silicon? *Nanotechnology* 2010, 21:262001.
  27. Branton D, Deamer DW, Marziali A, Bayley H, Benner SA, Butler T, Di Ventra M, Garaj S, Hibbs A, Huang X, et al. The potential and challenges of nanopore sequencing. *Nat Biotechnol* 2008, 26:1146–1153.
  28. Derrington IM, Butler TZ, Collins MD, Manrao E, Pavlenok M, Niederweis M, Gundlach JH. Nanopore DNA sequencing with MspA. *Proc Natl Acad Sci U S A* 2010, 107:16060.
  29. Wendell D, Jing P, Geng J, Subramaniam V, Lee TJ, Montemagno C, Guo P. Translocation of double-stranded DNA through membrane-adapted phi29 motor protein nanopores. *Nat Nanotechnol* 2009, 4:765–772.
  30. Healy K, Schiedt B, Morrison AP. Solid-state nanopore technologies for nanopore-based DNA analysis. *Nanomedicine (Lond)* 2007, 2:875–897.
  31. Dekker C. Solid-state nanopores. *Nat Nanotechnol* 2007, 2:209–215.
  32. Howorka S, Siwy Z. Nanopore analytics: sensing of single molecules. *Chem Soc Rev* 2009, 38:2360–2384.
  33. Bell N, Engst C, Ablay M, Divitini G, Ducati C, Liedl T, Keyser UF. DNA origami nanopores. *Nano Lett* 2011, 12:512.
  34. Meller A, Nivon L, Brandin E, Golovchenko J, Branton D. Rapid nanopore discrimination between single polynucleotide molecules. *Proc Natl Acad Sci U S A* 2000, 97:1079–1084.
  35. Wanunu M, Dadosh T, Ray V, Jin J, McReynolds L, Drndić M. Rapid electronic detection of probe-specific microRNAs using thin nanopore sensors. *Nat Nanotechnol* 2010, 5:807–814.
  36. Halverson KM, Panchal RG, Nguyen TL, Gussio R, Little SF, Misakian M, Bavari S, Kasianowicz JJ. Anthrax biosensor, protective antigen ion channel asymmetric blockade. *J Biol Chem* 2005, 280:34056–34062.
  37. Rodriguez-Larrea D, Bayley H. Multistep protein unfolding during nanopore translocation. *Nat Nanotechnol* 2013, 8:288.
  38. Boersma AJ, Brain KL, Bayley H. Real-time stochastic detection of multiple neurotransmitters with a protein nanopore. *ACS Nano* 2012, 6:5304–5308.
  39. Thompson JF, Oliver JS. Mapping and sequencing DNA using nanopores and nanodetectors. *Electrophoresis* 2012, 33:3429–3436.
  40. Howorka S, Cheley S, Bayley H. Sequence-specific detection of individual DNA strands using engineered nanopores. *Nat Biotechnol* 2001, 19:636–639.
  41. Wanunu M, Meller A. Chemically modified solid-state nanopores. *Nano Lett* 2007, 7:1580–1585.
  42. Wanunu M, Sutin J, McNally B, Chow A, Meller A. DNA translocation governed by interactions with solid-state nanopores. *Biophys J* 2008, 95:4716–4725.
  43. Venta K, Shemer G, Puster M, Rodriguez-Manzo JA, Balan A, Rosenstein JK, Shepard KL, Drndić M. Differentiation of short, single-stranded dna homopolymers in solid-state nanopores. *ACS Nano* 2013, 7:4629–4636.
  44. Kowalczyk SW, Wells DB, Aksimentiev A, Dekker C. Slowing down DNA translocation through a nanopore in lithium chloride. *Nano Lett* 2012, 12:1038–1044.
  45. Polonsky S, Rossnagel S, Stolovitzky G. Nanopore in metal-dielectric sandwich for DNA position control. *Appl Phys Lett* 2007, 91:153103.
  46. Lieberman KR, Cherf GM, Doody MJ, Olasagasti F, Kolodji Y, Akeson M. Processive replication of single DNA molecules in a nanopore catalyzed by phi29 DNA polymerase. *J Am Chem Soc* 2010, 132:17961–17972.
  47. Beamish E, Kwok H, Tabard-Cossa V, Godin M. Precise control of the size and noise of solid-state nanopores using high electric fields. *Nanotechnology* 2012, 23:405301.
  48. Hoogerheide D, Garaj S, Golovchenko J. Probing surface charge fluctuations with solid-state nanopores. *Phys Rev Lett* 2009, 102:256804.
  49. Bezrukov SM, Kasianowicz JJ. Current noise reveals protonation kinetics and number of ionizable sites in an open protein ion channel. *Phys Rev Lett* 1993, 70:2352–2355.
  50. Tabard-Cossa V, Trivedi D, Wiggin M, Jetha NN, Marziali A. Noise analysis and reduction in solid-state nanopores. *Nanotechnology* 2007, 18:305505.



51. Rosenstein JK, Wanunu M, Merchant CA, Drndic M, Shepard KL. Integrated nanopore sensing platform with sub-microsecond temporal resolution. *Nat Methods* 2012, 9:487.
52. Kim D, Goldstein B, Tang W, Sigworth FJ, Culurciello E. Noise analysis and performance comparison of low current measurement systems for biomedical applications. *IEEE Trans Biomed Circuits Syst* 2012, 2011:1–11.
53. Chen C, Yemencioğlu S, Uddin A, Corgliano E, Theogarajan L. A CMOS enhanced solid-state nanopore based single molecule detection platform. *IEEE Eng Med Biol Conf* 2013, 2013:164–167.
54. Besteman K, Lee JO, Wiertz FGM, Heering HA, Dekker C. Enzyme-coated carbon nanotubes as single-molecule biosensors. *Nano Lett* 2003, 3:727–730.
55. Lee C-S, Kim SK, Kim M. Ion-sensitive field-effect transistor for biological sensing. *Sensors (Basel)* 2009, 9:7111–7131.
56. Cui Y, Wei Q, Park H, Lieber CM. Nanowire nanosensors for highly sensitive and selective detection of biological and chemical species. *Science* 2001, 293:1289–1292.
57. Cui Y, Lauhon LJ, Gudiksen MS, Wang J, Lieber CM. Diameter-controlled synthesis of single-crystal silicon nanowires. *Appl Phys Lett* 2001, 78:2214.
58. Stern E, Klemic JF, Routenberg DA, Wyrembak PN, Turner-Evans DB, Hamilton AD, LaVan DA, Fahmy TM, Reed MA. Label-free immunodetection with CMOS-compatible semiconducting nanowires. *Nature* 2007, 445:519–522.
59. Zheng G, Gao XPA, Lieber CM. Frequency domain detection of biomolecules using silicon nanowire biosensors. *Nano Lett* 2010, 10:3179–3183.
60. Xie P, Xiong Q, Fang Y, Qing Q, Lieber CM. Local electrical potential detection of DNA by nanowire-nanopore sensors. *Nat Nanotechnol* 2011, 7:119.
61. McEuen PL, Fuhrer MS. Single-walled carbon nanotube electronics. *IEEE Trans Nanotechnol* 2002, 1:78–85.
62. Dénervaud N, Becker J, Delgado-gonzalo R, Damay P, Rajkumar AS, Unser M. A chemostat array enables the spatio-temporal analysis of the yeast proteome. *Proc Natl Acad Sci U S A* 2013, 110:15842.
63. Cai H, Cao X, Jiang Y, He P, Fang Y. Carbon nanotube-enhanced electrochemical DNA biosensor for DNA hybridization detection. *Anal Bioanal Chem* 2003, 375:287–293.
64. Zhang X, Chenet D, Kim B, Yu J, Tang J, Nuckolls C, Hone J. Fabrication of hundreds of field effect transistors on a single carbon nanotube for basic studies and molecular devices. *J Vac Sci Technol B* 2013, 31:06FI01.
65. Zevenbergen MAG, Singh PS, Goluch ED, Wolfrum B, Lemay SG. Stochastic sensing of single molecules in a nanofluidic electrochemical device. *Nano Lett* 2011, 11:2881–2886.
66. Kang S, Nieuwenhuis AF, Mathwig K, Mampallil D, Lemay SG. Electrochemical single-molecule detection in aqueous solution using self-aligned nanogap transducers. *ACS Nano* 2013, 7:10931–10937.
67. Kleijn SEF, Lai SCS, Miller TS, Yanson AI, Koper MTM, Unwin PR. Landing and catalytic characterization of individual nanoparticles on electrode surfaces. *J Am Chem Soc* 2012, 134:18558–18561.
68. Rosenblatt S, Yaish Y, Park J, Gore J, Sazonova V, McEuen PL. High performance electrolyte gated carbon nanotube transistors. *Nano Lett* 2002, 2:869–872.
69. Goldsmith BR, Coroneus JG, Khalap VR, Kane AA, Weiss GA, Collins PG. Conductance-controlled point functionalization of single-walled carbon nanotubes. *Science* 2007, 315:77–81.
70. Choi Y, Moody IS, Sims PC, Hunt SR, Corso BL, Perez I, Weiss GA, Collins PG. Single-molecule lysozyme dynamics monitored by an electronic circuit. *Science* 2012, 335:319–324.
71. Sorgenfrei S, Chiu CY, Gonzalez RL, Yu YJ, Kim P, Nuckolls C, Shepard KL. Label-free single-molecule detection of DNA-hybridization kinetics with a carbon nanotube field-effect transistor. *Nat Nanotechnol* 2011, 6:125–131.
72. Fei J, Bronson JE, Hofman JM, Srinivas RL, Wiggins CH, Gonzalez RL. Allosteric collaboration between elongation factor G and the ribosomal L1 stalk directs tRNA movements during translation. *Proc Natl Acad Sci U S A* 2009, 106:15702–15707.
73. Realov S, Shepard KL. Random telegraph noise in 45-nm CMOS: Analysis using an on-chip test and measurement system. In: *Electron Devices Meeting (IEDM)*. 2010 IEEE International; 2010, 28.2.1.
74. Bronson JE, Fei J, Hofman JM, Gonzalez RL Jr, Wiggins CH. Learning rates and states from biophysical time series: a Bayesian approach to model selection and single-molecule FRET data. *Biophys J* 2009, 97:3196–3205.
75. Männik J, Heller I, Janssens AM, Lemay SG, Dekker C. Charge noise in liquid-gated single-wall carbon nanotube transistors. *Nano Lett* 2008, 8:685–688.
76. Ishigami M, Chen JH, Williams ED, Tobias D, Chen YF, Fuhrer MS. Hooge's constant for carbon nanotube field effect transistors. *Appl Phys Lett* 2006, 88:203116.
77. Carminati M, Vergani M, Ferrari G, Caranzi L, Caironi M, Sampietro M. Accuracy and resolution limits in quartz and silicon substrates with microelectrodes for electrochemical biosensors. *Sens Actuators B* 2012, 174:168–175.

78. Aviram A, Ratner M. Molecular rectifiers. *Chem Phys Lett* 1974, 29:277.
79. Tao NJ. Electron transport in molecular junctions. *Nat Nanotechnol* 2006, 1:173–181.
80. Aradhya SV, Venkataraman L. Single-molecule junctions beyond electronic transport. *Nat Nanotechnol* 2013, 8:399–410.
81. McCreery RL, Yan H, Bergren AJ. A critical perspective on molecular electronic junctions: there is plenty of room in the middle. *Phys Chem Chem Phys* 2013, 15:1065–1081.
82. Venkataraman L, Klare JE, Tam IW, Nuckolls C, Hybertsen MS, Steigerwald ML. Single-molecule circuits with well-defined molecular conductance. *Nano Lett* 2006, 6:458.
83. Venkataraman L, Klare JE, Nuckolls C, Hybertsen MS, Steigerwald ML. Dependence of single-molecule junction conductance on molecular conformation. *Nature* 2006, 442:1–4.
84. Tsutsui M, Taniguchi M, Yokota K, Kawai T. Identifying single nucleotides by tunnelling current. *Nat Nanotechnol* 2010, 5:286.
85. Stava E, Yu M, Shin HC, Shin H, Kreft DJ, Blick RH. Rapid fabrication and piezoelectric tuning of micro- and nanopores in single crystal quartz. *Lab Chip* 2013, 13:156–160.
86. Tsutsui M, Rahong S, Iizumi Y, Okazaki T, Taniguchi M, Kawai T. Single-molecule sensing electrode embedded in-plane nanopore. *Sci Rep* 2011, 1:46.
87. Ilyas A, Asghar W, Billo JA, Syed EAQ, Iqbal SM. IEEE/NIH Life Science Systems and Applications Workshop (LiSSA), 2011, 79–82.
88. Mészáros G, Li C, Pobelov I, Wandlowski T. Current measurements in a wide dynamic range-applications in electrochemical nanotechnology. *Nanotechnology* 2007, 18:424004.
89. Cheng Z, Skouta R, Vazquez H, Widawsky JR, Schneebeli S, Chen W, Hybertsen MS. In situ formation of highly conducting covalent Au–C contacts for single-molecule junctions. *Nat Nanotechnol* 2011, 6:353–357.
90. Chang S, Huang S, He J, Liang F, Zhang P, Li S, Chen X, Sankey O, Lindsay S. Electronic signatures of all four DNA nucleosides in a tunneling gap. *Nano Lett* 2010, 10:1070–1075.
91. Zhao Y, Ashcroft B, Zhang P, Liu H, Sen S, Song W, Im J, Gyarfás B, Manna S, Biswas S. Single-molecule spectroscopy of amino acids and peptides by recognition tunnelling. *Nat Nanotechnol* 2014, 9:466.
92. Tsutsui M, Taniguchi M, Kawai T. Single-molecule identification via electric current noise. *Nat Commun* 2010, 1:138.
93. Zwolak M, Di Ventra M. Electronic signature of DNA nucleotides via transverse transport. *Nano Lett* 2005, 5:421–424.
94. Albrecht T. Electrochemical tunnelling sensors and their potential applications. *Nat Commun* 2012, 3:810–829.
95. Huang S, He J, Chang S, Zhang P, Liang F, Li S, Tuchband M, Fuhrmann A, Ros R, Lindsay S. Identifying single bases in a DNA oligomer with electron tunnelling. *Nat Nanotechnol* 2010, 5:868.
96. Tsutsui M, Matsubara K, Ohshiro T, Furuhashi M, Taniguchi M, Kawai T. Electrical detection of single methylcytosines in a DNA oligomer. *J Am Chem Soc* 2011, 133:9124–9128.
97. Fan FRF, Bard AJ. Electrochemical detection of single molecules. *Science* 1995, 267:871–874.
98. Fan FRF, Kwak J, Bard AJ. Single molecule electrochemistry. *J Am Chem Soc* 1996, 118:9669–9675.
99. Bard AJ, Fan FRF. Electrochemical detection of single molecules. *Acc Chem Res* 1996, 29:572–578.
100. Sun P, Mirkin MV. Electrochemistry of individual molecules in zeptoliter volumes. *J Am Chem Soc* 2008, 130:8241–8250.
101. Zevenbergen MAG, Krapf D, Zuiddam MR, Lemay SG. Mesoscopic concentration fluctuations in a fluidic nanocavity detected by redox cycling. *Nano Lett* 2007, 7:384–388.
102. Zevenbergen MAG, Singh PS, Goluch ED, Wolfrum BL, Lemay SG. Electrochemical correlation spectroscopy in nanofluidic cavities. *Anal Chem* 2009, 81:8203–8212.
103. Singh PS, Katelhon E, Mathwig K, Wolfrum B, Lemay SG. Stochasticity in single-molecule nanoelectrochemistry: origins, consequences, and solutions. *ACS Nano* 2012, 6:9662–9671.
104. Xiao X, Bard AJ. Observing single nanoparticle collisions at an ultramicroelectrode by electrocatalytic amplification. *J Am Chem Soc* 2007, 129:9610.
105. Xiao X, Fan F-RF, Zhou J, Bard AJ. Current transients in single nanoparticle collision events. *J Am Chem Soc* 2008, 130:16669–16677.
106. Kwon SJ, Zhou H, Fan F-RF, Vorobyev V, Zhang B, Bard AJ. Stochastic electrochemistry with electrocatalytic nanoparticles at inert ultramicroelectrodes-theory and experiments. *Phys Chem Chem Phys* 2011, 13:5394–5402.
107. Bernhardt PV. Enzyme electrochemistry—biocatalysis on an electrode. *Aust J Chem* 2006, 59:233–256.
108. Armstrong FA, Heering HA, Hirst J. Reactions of complex metalloproteins studied by protein-film voltammetry. *Chem Soc Rev* 1997, 26:169–179.
109. Bard AJ. Toward single enzyme molecule electrochemistry. *ACS Nano* 2008, 2:2437–2440.
110. Hoeben FJM, Meijer FS, Dekker C, Albracht SPJ, Heering HA, Lemay SG. Toward single-enzyme molecule electrochemistry: NiFe–hydrogenase protein film voltammetry at nanoelectrodes. *ACS Nano* 2008, 2:2497–2504.

111. Rosenstein J, Sorgenfrei S, Shepard KL. Noise and bandwidth performance of single-molecule biosensors. In: *Custom Integrated Circuits Conference (CICC)*, San Jose, CA, 2011.
112. Crescentini M, Bennati M, Carminati M, Tartagni M. Noise limits of CMOS current interfaces for biosensors: a review. *IEEE Trans Biomed Circuits Syst* 2013, 2013:1–15.
113. Nazari M, Mazhab-Jafari H. CMOS neurotransmitter microarray: 96-channel integrated potentiostat with on-die microsenors. *IEEE Trans Biomed Circuits Syst* 2012, 7:338.
114. Farre C, Fertig N. HTS techniques for patch clamp-based ion channel screening—advances and economy. *Expert Opin Drug Discovery* 2012, 7:515–524.
115. Sigworth FJ, Klemic KG. Microchip technology in ion-channel research. *IEEE Trans Nanobioscience* 2005, 4:121–127.
116. Lindsay S. Biochemistry and semiconductor electronics—the next big hit for silicon? *J Phys Condens Matter* 2012, 24:164201.
117. Rosenstein JK, Ramakrishnan S, Roseman J, Shepard KL. Single ion channel recordings with CMOS-anchored lipid membranes. *Nano Lett* 2013, 13:2682.
118. Kim J, Pedrotti K, Dunbar WB. An area-efficient low-noise CMOS DNA detection sensor for multichannel nanopore applications. *Sens Actuators B* 2012, 176:1051.
119. Goldstein B, Member GS, Kim D, Xu J, Vanderlick TK, Culurciello E, Member S. CMOS low current measurement system for biomedical applications. *IEEE Trans Biomed Circuits Syst* 2012, 6:111–119.
120. Ferrari G, Gozzini F, Molari A, Sampietro M. Transimpedance amplifier for high sensitivity current measurements on nanodevices. *IEEE J Solid-State Circuits* 2009, 44:1609–1616.
121. Vanisri T, Toumazou C. Integrated high frequency low-noise current-mode optical transimpedance preamplifiers: theory and practice. *IEEE J Solid-State Circuits* 1995, 30:677–685.
122. Radeka V. Low-noise techniques in detectors. *Annu Rev Nucl Part Sci* 1988, 38:217.
123. Denison TA. The development of a nanoscale coulter counter for rapid genetic sequence recognition. Appendix D: circuits schematics. PhD Thesis, *Massachusetts Institute of Technology*, 2000.
124. English AC, Richards S, Han Y, Wang M, Vee V, Qu J, Qin X, Muzny DM, Reid JG, Worley KC, et al. Mind the gap: upgrading genomes with Pacific Biosciences RS long-read sequencing technology. *PLoS ONE* 2012, 7:e47768.
125. Maitra RD, Kim J, Dunbar WB. Recent advances in nanopore sequencing. *Electrophoresis* 2012, 33:3418.
126. Pennisi E. DNA sequencers still waiting for the nanopore revolution. *Science* 2014, 343:829–830.
127. Rothberg JM, Hinz W, Rearick TM, Schultz J, Mileski W, Davey M, Leamon JH, Johnson K, Milgrew MJ, Edwards M, et al. An integrated semiconductor device enabling non-optical genome sequencing. *Nature* 2011, 475:348–352.
128. Ivanov AP, Instuli E, McGilvery CM, Baldwin G, McComb DW, Albrecht T, Edel JB. DNA tunneling detector embedded in a nanopore. *Nano Lett* 2011, 11:279–285.
129. Liang X, Chou S. Nanogap detector inside nanofluidic channel for fast real-time label-free DNA analysis. *Nano Lett* 2008, 8:1472–1476.
130. Park YS, Whalley AC, Kamenetska M, Steigerwald ML, Hybertsen MS, Nuckolls C, Venkataraman L. Contact chemistry and single-molecule conductance: a comparison of phosphines, methyl sulfides, and amines. *J Am Chem Soc* 2007, 129:15768–15769.