

## The Glove3

***Citation for published version (APA):***

Ungureanu, C., Aarts, R. M., & Arends, J. B. A. M. (2015). The Glove3: a new device for measurement of vital signs. In *Proceedings of the 5th Dutch Bio-Medical Engineering Conference, Egmond aan Zee, The Netherlands, 22-23 January 2015*

***Document status and date:***

Published: 01/01/2015

***Document Version:***

Accepted manuscript including changes made at the peer-review stage

***Please check the document version of this publication:***

- A submitted manuscript is the version of the article upon submission and before peer-review. There can be important differences between the submitted version and the official published version of record. People interested in the research are advised to contact the author for the final version of the publication, or visit the DOI to the publisher's website.
- The final author version and the galley proof are versions of the publication after peer review.
- The final published version features the final layout of the paper including the volume, issue and page numbers.

[Link to publication](#)

***General rights***

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal.

If the publication is distributed under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license above, please follow below link for the End User Agreement:

[www.tue.nl/taverne](http://www.tue.nl/taverne)

***Take down policy***

If you believe that this document breaches copyright please contact us at:

[openaccess@tue.nl](mailto:openaccess@tue.nl)

providing details and we will investigate your claim.

# **THE GLOVE3, A NEW DEVICE FOR MEASUREMENT OF VITAL SIGNS**

**Constantin Ungureanu\*<sup>‡</sup>, Ronald M. Aarts\*<sup>‡</sup> and Johan B. A. M. Arends\*<sup>‡</sup>**

\* Dept. of Electrical Engineering, Eindhoven University of Technology, The Netherlands;

<sup>†</sup> Epilepsiecentrum Kempenhaeghe, Heeze, The Netherlands;

<sup>‡</sup>HOBO Heeze, The Netherlands

<sup>‡</sup>Philips Research Eindhoven, The Netherlands

Email: c.ungureanu@tue.nl

## **ABSTRACT**

Heart rate, respiratory rate, body temperature, oxygen saturation, and blood pressure are the main vital signs of the human body [1]. Monitoring of these parameters can provide information about the general health status of patients. In general, each vital parameter is measured with a different device.

In this paper we present the details of a preclinical prototype for measurement of heart and respiratory rate plus oxygen saturation. This device is embedded in a glove that has an optical sensor to record the photoplethysmographic (PPG) signal from the patient. The optical module situated on the palmar side of the glove uses a reflection mode pulse oximeter at 25 degrees incidence. Further, the PPG signal is recorded and transmitted wirelessly by a Shimmer sensor node situated on the dorsal side of the glove. On the same side a solar cell is located which is used to charge the battery of the glove system. The PPG signals are received by a laptop and processed in real time to extract the vital signs.

To measure the vital signs, the nurse or the doctor simply touches the patient skin while using the glove. We evaluated the glove on 10 healthy volunteers. The respiratory and heart rate errors were under 4 breaths respective beats per minute when compared with reference. This device can be used in rural areas by doctors to evaluate the general health status of patients. In addition, the glove can communicate with smart phones where databases can be generated.

## **REFERENCES**

- [1] M. Elliot and A. Coventry, Critical care: the eight vital signs of patient monitoring, Vol. 21(10), pp. 621-625, (2012)