

# Enabling Spectrally Resolved Single-Molecule Localization Microscopy at High Emitter Densities

**Citation for published version (APA):**

Martens, K. J. A., Gobes, M., Archontakis, E., Brillas, R. R., Zijlstra, N., Albertazzi, L., & Hohlbein, J. (2022). Enabling Spectrally Resolved Single-Molecule Localization Microscopy at High Emitter Densities. *Nano Letters*, 22(21), 8618-8625. <https://doi.org/10.1021/acs.nanolett.2c03140>

**Document license:**

CC BY

**DOI:**

[10.1021/acs.nanolett.2c03140](https://doi.org/10.1021/acs.nanolett.2c03140)

**Document status and date:**

Published: 09/11/2022

**Document Version:**

Publisher's PDF, also known as Version of Record (includes final page, issue and volume numbers)

**Please check the document version of this publication:**

- A submitted manuscript is the version of the article upon submission and before peer-review. There can be important differences between the submitted version and the official published version of record. People interested in the research are advised to contact the author for the final version of the publication, or visit the DOI to the publisher's website.
- The final author version and the galley proof are versions of the publication after peer review.
- The final published version features the final layout of the paper including the volume, issue and page numbers.

[Link to publication](#)

**General rights**

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal.

If the publication is distributed under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license above, please follow below link for the End User Agreement:

[www.tue.nl/taverne](http://www.tue.nl/taverne)

**Take down policy**

If you believe that this document breaches copyright please contact us at:

[openaccess@tue.nl](mailto:openaccess@tue.nl)

providing details and we will investigate your claim.

# Enabling Spectrally Resolved Single-Molecule Localization Microscopy at High Emitter Densities

Koen J. A. Martens,\* Martijn Gobes, Emmanouil Archontakis, Roger R. Brillas, Niels Zijlstra, Lorenzo Albertazzi, and Johannes Hohlbein\*



Cite This: *Nano Lett.* 2022, 22, 8618–8625



Read Online

ACCESS |


Metrics & More

Article Recommendations

Supporting Information

**ABSTRACT:** Single-molecule localization microscopy (SMLM) is a powerful super-resolution technique for elucidating structure and dynamics in the life- and material sciences. Simultaneously acquiring spectral information (spectrally resolved SMLM, sSMLM) has been hampered by several challenges: an increased complexity of the optical detection pathway, lower accessible emitter densities, and compromised spatio-spectral resolution. Here we present a single-component, low-cost implementation of sSMLM that addresses these challenges. Using a low-dispersion transmission grating positioned close to the image plane, the +1<sup>st</sup> diffraction order is minimally elongated and is analyzed using existing single-molecule localization algorithms. The distance between the 0<sup>th</sup> and 1<sup>st</sup> order provides accurate information on the spectral properties of individual emitters. This method enables a 5-fold higher emitter density while discriminating between fluorophores whose peak emissions are less than 15 nm apart. Our approach can find widespread use in single-molecule applications that rely on distinguishing spectrally different fluorophores under low photon conditions.

**KEYWORDS:** Single-molecule spectroscopy, multicolor imaging, single-molecule Förster resonance energy transfer (smFRET), stochastic optical reconstruction microscopy (STORM), point accumulation for imaging in nanoscale topography (PAINT)



Super-resolution microscopy, or nanoscopy, has revolutionized the life and material sciences as it allows surpassing the optical diffraction limit by more than an order of magnitude.<sup>1–4</sup> One frequently used implementation is single-molecule localization microscopy (SMLM), in which the stochastic activation of single fluorescent emitters leads to spatially separated point spread functions (PSFs) that are used to determine the position of each emitter with sub-50 nm accuracy. Localizations obtained via (direct) stochastic optical reconstruction microscopy ((d)STORM),<sup>1,4</sup> point accumulation for imaging in nanoscale topography (PAINT),<sup>5–8</sup> or photoactivated localization microscopy (PALM)<sup>3,9,10</sup> provide access to detailed structural images or can quantify dynamics and mobilities via single-particle tracking (spt).<sup>10,11</sup> In this capacity, SMLM has led to breakthroughs<sup>12,13</sup> in fields such as DNA–protein interactions,<sup>14–17</sup> cell biology,<sup>18–20</sup> and soft matter.<sup>8,21–23</sup>

Improving throughput via multiplexing of different fluorophores in SMLM,<sup>5</sup> enabling microenvironmental characterization,<sup>24,25</sup> or studying fluorophore-to-fluorophore distance (via single-molecule Förster resonance energy transfer (smFRET)),<sup>26,27</sup> can be accomplished by combining SMLM with the additional spectral characterization of emitters. Spectral information on single emitters can be acquired and


analyzed using various implementations that all rely on placing additional components into the optical detection pathway.


The first implementation of spectrally resolved SMLM uses the ratiometric distinction of spectral emission profiles and is based on placing one (or more) suitable dichroic mirror(s) in the emission pathway.<sup>28–31</sup> Photons emitted from the sample are separated based on their wavelength and directed toward two different detection channels. This entails either two separate detectors or using two areas on the same camera chip after using additional lenses or mirrors to guide the beams. Then the PSFs and their integrated intensities obtained in the two channels are matched, and the intensity ratio of photons is used to discriminate between the emission spectra of different fluorophores. Importantly, this method requires photons to be directed toward each channel, implicating that only a defined spectral range around the cutoff wavelength of the dichroic mirror can be accessed.

**Received:** August 8, 2022

**Revised:** October 17, 2022

**Published:** October 21, 2022





**Figure 1.** Implementation of low-dispersion spectrally resolved single-molecule localization microscopy. (a) A low-dispersion blazed transmission grating is placed in the emission path of a typical SMLM capable microscope such that the distance of the grating to the image plane is minimized. Around 50% of the light passing the grating will not have any dispersion, causing a 0<sup>th</sup> order point spread function (PSF) to appear. The other 50% of the light is dispersed based on its wavelength and will create a second, slightly elongated 1<sup>st</sup> order diffraction pattern. Image not to scale. (b) Simulation of low-dispersion sSMLM data of two spectrally different emitters with  $\lambda_1$  (corresponding to  $d_1$ ) <  $\lambda_2$  (corresponding to  $d_2$ ). Six emitters create in total five 0<sup>th</sup> and 1<sup>st</sup> order diffraction pairs on the detector, which can be linked together (two 1<sup>st</sup> order diffraction patterns are not captured in the field of view). The obtained distances between the 0<sup>th</sup> and 1<sup>st</sup> order diffraction patterns ( $d_1$  and  $d_2$ ) are a measure for the average emission wavelength. (c) Single frame of raw data of a DNA-PAINT nanoruler sample showing 85 spatio-spectrally resolvable emitters in a  $31 \times 31 \mu\text{m}^2$  field of view. The red outline is enlarged in the inset, in which the 0<sup>th</sup> and 1<sup>st</sup> order diffraction patterns are encircled in magenta and green, respectively. Scale bar represents  $5 \mu\text{m}$ . (d) Comparison of achievable emitter density in standard (non-spectrally resolved) SMLM (black dotted line), our approach (red line), and sSMLM with 20–30 pixels spectral pattern elongation, taken from Bongiovanni et al.<sup>37</sup> The shaded background indicates the standard deviation as determined from repeating the simulations.


The second implementation uses point spread function engineering to obtain spectral information on emitters. Here, a spatial light modulator (SLM) or a phase mask (PM) is employed in the Fourier plane of the emission path.<sup>32,33</sup> The introduced phase offset by these elements depends on the incoming wavelength, which can be exploited to design a pattern so that different PSF shapes are realized when photons of different wavelength arrive at the detector. However, small spectral emission differences in the order of tens of nanometers in the peak emission cannot create sufficiently distinct PSF shapes, hindering discrimination of spectrally close fluorophores. Moreover, the voltage (for SLM) or phase (for PM) has to be specifically tuned for certain emission wavelengths, complicating this method when different fluorophores are used.

In the third implementation, spectral dispersion, a spectrally dispersive optical element is added, after which the spatial and spectral profiles are guided to different regions on a single camera chip or to completely separate detectors.<sup>34–41</sup> Generally, the spatial profile is then analyzed with regular single-molecule localization algorithms,<sup>42–46</sup> while the spectral profile is spread out over tens of pixels and is used to determine the corresponding emission profile. While this implementation allows a large spectral range to be used and allows the discrimination of fluorophores with similar emission spectra, it has various downsides. First, the entire emission pathway needs to be modified to separate the spatial from the spectral channel.<sup>37,47</sup> Second, as the spectral information is widened over tens of pixels, the signal-to-noise ratio obtainable in this channel is compromised, leading to a loss in spectral accuracy.<sup>48</sup> The wide spreading of emission further directly limits the usable density of emitters in the sample, as overlapping spectral profiles cannot be resolved.

In comparing these implementations, it is obvious that there is a need to combine an easy implementation with a broad spectral range and good specificity. Here, we demonstrate spectrally resolved single-molecule localization microscopy (sSMLM) using an inexpensive (blazed) transmission grating

that can be easily implemented in most microscope configurations allowing for widespread adoption while maximizing the achievable signal-to-noise ratio. By using a grating with a large line pitch and placing it close to the image plane to minimize the dispersion of the 1<sup>st</sup> order, we created a low-dispersion sSMLM implementation in which the 0<sup>th</sup> and 1<sup>st</sup> diffraction order of every emitter is imaged and analyzed by existing single-molecule localization algorithms in a single field of view without the need for any additional optical elements. This concept can be straightforwardly combined with SMLM methods employing structured excitation profiles<sup>49–54</sup> and with 3-dimensional PSF engineering approaches,<sup>43,44,55–59</sup> further increasing the potential use of our implementation. Our novel implementation is capable of accurately determining spectral properties of single molecules at much higher emitter densities than other spectral-dispersing sSMLM implementations. With our sSMLM approach, we show the technical feasibility of spectral multiplexing and ability to distinguish between 0 and 15% FRET efficiency in single-molecule FRET experiments.

To maximize the signal-to-noise ratio and the achievable molecular density in spectrally resolved single-molecule localization microscopy (sSMLM), the available photon budget should be distributed over as few camera pixels as possible.<sup>48</sup> To fulfill this criterion, we placed a transmission grating with low dispersion (70 lines/mm) as close as possible to a camera chip representing the image plane (<1 cm, Figure 1a) by placing it inside the camera housing. This arrangement minimizes the separation of the 0<sup>th</sup> and 1<sup>st</sup> order diffraction patterns and thus results in the highest achievable fluorophore density and signal-to-noise ratio for the 1<sup>st</sup> order diffraction pattern (Figure 1b,c). Notably, and distinctly different from earlier implementations,<sup>34–37</sup> our arrangement allows imaging of both spatial and spectral information in the same field of view, thereby maximizing the usable detection area on the sensor. Furthermore, our approach does not require any additional optical components such as mirrors, beam splitters, or secondary detectors.



**Figure 2.** Multiplexing dSTORM of fixed Cos7 cells with CF660-labeled clathrin and CF680-labeled tubules. (a) Obtained dSTORM image, color-coded on 0<sup>th</sup>-to-1<sup>st</sup>-order distance. Separation between tubule and clathrin can be observed without further data analysis. Scale bar represents 1  $\mu\text{m}$ . (b) Emission spectra of CF660 (bright red) and CF680 (dark red). Dotted lines represent full spectra, while the solid lines represent emission spectra corrected for the transmission characteristics of the optical components in the detection path of the microscope. (c) Histograms representing 0<sup>th</sup>-to-1<sup>st</sup>-order distances of fluorophores belonging to areas indicated by dotted outlines in panel a. These populations are fitted with Gaussian curves (see main text) and attributed to CF660 (green) or to CF680 (magenta).

As the PSFs from the obtained 1<sup>st</sup> order diffraction pattern show only minor excess width compared to the PSFs in the 0<sup>th</sup> order, we were able to employ existing super-resolution algorithms to independently obtain subpixel localizations of the 0<sup>th</sup> and 1<sup>st</sup> order diffraction patterns. Next, we linked the localizations with each other in the dispersion direction, with the distance between the 0<sup>th</sup> and 1<sup>st</sup> order diffraction patterns ( $d_1$  and  $d_2$  in Figure 1b; further called “0<sup>th</sup>-to-1<sup>st</sup>-order distance”) being a direct measure for the average emission wavelength of the emitter ( $\lambda_1 < \lambda_2$  with  $d_1 < d_2$ ). Moreover, the excess width of the 1<sup>st</sup> order diffraction pattern compared to the width of the PSF in the 0<sup>th</sup> order is a measure for the width of the emission spectrum. This directly results in a spectral accuracy being limited primarily by spatial single-molecule localization accuracy, an area of research that is progressing very rapidly via both software and hardware developments.<sup>42,55,60</sup> Using minimal dispersion, the achievable density in our implementation without the need for specialized high-density fitting algorithms is around 5 times higher than earlier implementations of sSMLM, where the spectral information is spread out over 20–30 pixels (Figure 1d; Methods, Supporting Information).<sup>37</sup>


We determined the distance between the dispersion-inducing optics of the grating to the camera chip to be  $6.9 \pm 0.1$  mm in our system (Methods, Supporting Information,

Supplementary Figure 1a,b). The spectral dispersion (SD) was determined by calculating the 0<sup>th</sup>-to-1<sup>st</sup>-order distance of a sample labeled with ATTO542 and ATTO655 (Supplementary Figure 1c). From the median distance of these obtained distances and the mean emission profile of the fluorophores, a spectral dispersion of  $\sim 0.21$  nm/nm (spatial/spectral; equivalent here to  $\sim 27$  nm/px (spectral/spatial)) was determined. We did not observe a wavelength dependency on the angle between the spatial and spectral profiles (Supplementary Figure 1d).

The signal-to-noise ratio of the obtained 1<sup>st</sup> order diffraction pattern, combined with high resolution of subpixel localization algorithms, indicates that small spectral differences can be elucidated. We imaged double-labeled fixed Cos7 cells, in which clathrin was labeled with the fluorophore CF660, and tubulin with CF680. Pseudocolor coding a super-resolved image based on 0<sup>th</sup>-to-1<sup>st</sup>-order distance reveals good separation of the labeled structures without further analysis (Figure 2a), even though these fluorophores only have a  $\sim 10$  nm intensity-weighted spectral separation in our microscope (CF660: 692 nm, CF680: 702 nm, Figure 2b; 12 nm difference in peak wavelength (CF660: 686 nm, CF680: 698 nm)).

Selecting image regions with mostly CF660- or CF680-labeled structures (dotted outlines in Figure 2a) and fitting the corresponding 0<sup>th</sup>-to-1<sup>st</sup>-order distances with a Gaussian profile





**Figure 3.** Low-dispersion sSMLM is capable of distinguishing ATTO647N from ATTO655 in DNA-PAINT. (a) Two different nanoparticles have associated DNA-PAINT imager strands containing either ATTO647N or ATTO655. Scheme not to scale. (b) Emission spectra of ATTO647N (bright red) and ATTO655 (dark red). Dotted lines represent full spectra, while the solid lines represent emission spectra corrected for the optical components present in the microscope. (c) Histograms representing observed 0<sup>th</sup>-to-1<sup>st</sup>-order distances of fluorophores belonging to individual NPs. These populations are fitted with Gaussian curve(s) (see main text) and attributed to NPs accepting ATTO647N-DNA (green) or to NPs accepting ATTO655-DNA (magenta). (d) Visualization of six individual NPs, with individual localizations color-coded based on the dotted line shown in panel b. Scale bar represents 500 nm.

reveals that CF660 has a 0<sup>th</sup>-to-1<sup>st</sup>-order distance of  $3077 \pm 2$  nm ( $\sigma = 56 \pm 2$  nm; mean  $\pm 95\%$  confidence interval (CI)) and  $3128 \pm 2$  nm for CF680 ( $\sigma = 62 \pm 2$  nm; Figure 2c). This is a difference of  $51 \pm 2$  nm in the raw data, which corresponds to a spectral distance of  $10.9 \pm 0.4$  nm, in agreement with the weighted average and peak position difference of the microscope-corrected emission profiles.

Next, we performed a DNA-PAINT experiment on polystyrene nanoparticles (NPs) that have DNA-PAINT docking strands for either ATTO647N or for ATTO655 (Figure 3a). These fluorophores have a weighted average emission wavelength separated by only  $\sim 9$  nm (685 nm and 693 nm, respectively, after correcting for optical components in our microscope, Figure 3b) and a peak emission wavelength separation of  $\sim 14$  nm (665 and 679 nm, respectively, after correcting for optical components in our microscope, Figure 3b). After isolating localizations belonging to individual beads and analyzing the 0<sup>th</sup>-to-1<sup>st</sup>-order distances of these emitters, two populations can be observed (Figure 3c).

The population with the lowest 0<sup>th</sup>-to-1<sup>st</sup>-order distance (green; Gaussian fit peak position:  $3014 \pm 1$  nm,  $\sigma = 34 \pm 1$  nm, mean  $\pm 95\%$  CI) was attributed to ATTO647N fluorophores. The population with the larger distances (magenta) was fitted with a combination of two Gaussian curves: one restricted to the fit of the first observed population; along with a unique Gaussian curve (Gaussian fit peak position:  $3090 \pm 2$  nm,  $\sigma = 41 \pm 3$  nm). This population was attributed to ATTO655 fluorophores. The larger standard deviation of the ATTO655 population compared to the ATTO647N population can be attributed to a lower median localization accuracy (42 nm vs 50 nm), possibly caused by a difference in quantum yield (65% vs 30%). The spectral distance between these fitted peak positions ( $76 \pm 2$  nm distance; corresponding to  $16.2 \pm 0.4$  nm spectral separation) is close to the difference between the emission peaks of both fluorophores but higher than the weighted average wavelength. This is possibly caused by deviations of the described wavelength-dependent efficiency of optical elements compared to our hardware implementation, which could lead to a shifted weighted mean emission wavelength of ATTO647N, as its


emission maximum is close to the lower spectral cutoff ( $\sim 660$  nm) of the filters and beam splitters used.

Next, all linked localizations were color-coded according to their distance (cutoff at the black dotted line in Figure 3c at 3054 nm). Visualization of the individual NPs (Figure 3d) then reveals their fluorophore distribution. This shows that the NPs are populated by either one DNA-PAINT docking strand or the other, with minimal cross-talk between the used fluorophores, which can be attributed to unspecific DNA–DNA interactions.

Taken together with the dSTORM data of the fixated cells, the obtained order of the mean emission wavelength of these four fluorophores (ATTO647N, CF660, ATTO655, CF680) coincided with that of the 0<sup>th</sup>-to-1<sup>st</sup>-order distance (calculated mean emission wavelength from spectra provided by the manufacturers and corrected for the optical properties of our setup): 685, 692, 693, 701 nm; mean 0<sup>th</sup>-to-1<sup>st</sup>-order distance: 3014, 3077, 3090, 3128 nm). The separation of these 0<sup>th</sup>-to-1<sup>st</sup>-order distances suggests that simultaneous multiplexing of at least three fluorophores with single-wavelength excitation is technically possible under realistic experimental conditions (Supplementary Figure 2).

Next, we were interested whether we could expand the low-dispersion sSMLM to assess single-molecule Förster resonance energy transfer (smFRET). In a typical surface-based smFRET experiment, probes labeled with a donor and an acceptor fluorophore are immobilized and monitored over time. Depending on the experiment, changes in FRET and/or changes in acceptor/donor activity (i.e., blinking and bleaching) can be expected. Conventionally, a ratiometric spectral determination method is applied to separate the donor emission from the acceptor emission on different positions on a camera chip, and the intensity ratio between these is a measure for the relative FRET efficiency  $E^*$ . In our implementation, we can use the full field of view of our camera and determine  $E^*$  via the 0<sup>th</sup>-to-1<sup>st</sup>-order distance. We further can utilize the width of the 1<sup>st</sup> order diffraction pattern as an additional way of discriminating FRET.

We performed smFRET measurements on well-characterized samples of immobilized double-stranded DNA that is dual-labeled with ATTO550 and ATTO647N.<sup>27</sup> Two samples



**Figure 4.** Single-molecule FRET analysis with low-dispersion sSMLM. (a) Predicted emission spectra of donor-only (top, ATTO550), 15% FRET (middle), and 55% FRET (bottom). Dotted lines represent full spectra, while the solid lines represent emission spectra corrected for the transmission characteristics of the optical components present in the microscope. Schemes represent donor (green; ATTO550) and acceptor (red; ATTO647N) fluorophore placements on a DNA strand. (b) Simulated raw data obtained in our low-dispersion sSMLM implementation based on the emission profiles determined in panel a. Vertical dotted white lines are a guide for the eye. (c, d) Two-dimensional histograms of experimental data of (c) 15% FRET and (d) 55% FRET. The histograms were globally fitted with multiple Gaussian functions (shown here centered around black crosses, with solid ellipses representing 1 sigma, and dotted ellipses representing 2 sigma; Methods (Supporting Information), main text). (e–h) Single emitter time trace analysis of (e, f) a bleaching acceptor fluorophore in a 15% FRET pair, and of (g, h) a blinking acceptor fluorophore in a 55% FRET pair. Horizontal gray lines with red and blue shading represent (e) 15% or (g) 55% FRET populations (red) and donor-only populations (blue), determined from the fits in panels c and d. The vertical green, yellow, and orange shading represent current FRET pair state, with green representing donor-only, yellow representing 15% FRET, and orange representing 55% FRET. The raw data corresponding to these FRET pairs throughout the observation time are shown in panels f and h. Scale bars in panels f and h represent 500 nm.

with different distances between the labeling sites were used: 23-bp separation ( $\sim 8.4$  nm) and 15-bp separation ( $\sim 5.9$  nm), leading to apparent FRET efficiencies of  $\sim 0.15$  and  $\sim 0.55$ , respectively. First, we performed simulations using the known emission profiles corrected for the fluorophore's quantum yield and the optical elements in our microscope (Methods, Supporting Information; Figure 4a,b). These simulations of donor-only, 15% FRET, and 55% FRET samples show that the 0<sup>th</sup>-to-1<sup>st</sup>-order distance follows  $d_{\text{donor}} < d_{15\%} < d_{55\%}$ , and the width of the 1<sup>st</sup> order diffraction pattern follows  $\sigma_{\text{donor}} < \sigma_{15\%} < \sigma_{55\%}$ .

Experimentally, the immobilized DNA strands were imaged separately. Contrary to the multiplexing before, both the 0<sup>th</sup>-to-1<sup>st</sup>-order distance and the width of 1<sup>st</sup> order diffraction pattern are measures for the FRET efficiency and were therefore visualized (Figure 4c,d). The experimental data agree with the simulations showing  $d_{\text{donor}} < d_{15\%} < d_{55\%}$  (2620 nm, 2653 nm, 2781 nm, respectively) and  $\sigma_{\text{donor}} < \sigma_{15\%} < \sigma_{55\%}$  (172 nm, 204 nm, 238 nm, respectively).

Next, we explored to which extent we can monitor dynamic behavior using spectrally resolved smFRET. While no direct state transitions are expected for this sample, there is occasional acceptor fluorophore blinking or bleaching, leading to a transition of FRET emission to donor-only emission. For this, we fitted the combined  $[d, \sigma]$  2-dimensional histogram with four 2-dimensional Gaussian profiles (Figure 4c,d black crosses and ellipses). These profiles comprise donor-only, 15% FRET efficiency, 55% FRET efficiency, and “background” states (background state not shown). The “background” state is attributed to nonsense linkages occurring from sparse localizations unrelated to the FRET sample.

Time traces of individual emitters were further assessed (Figure 4e–h). The likelihood of an emitter belonging to the predetermined states was calculated (Methods, Supporting Information), and the most likely state dictated the background color of the graphs in Figure 4e and g. With this methodology, we were able to determine acceptor bleaching (Figure 4e) and acceptor blinking (Figure 4g) in 15% and 55%

FRET experimental data. Accurate state determination of the 15% FRET measurement proved to be difficult due to the overlapping Gaussian profiles representing the FRET and donor-only states (Figure 4c), whereas this was better discriminable for 55% FRET.

Here we have demonstrated minimal-dispersion spectrally resolved SMLM (sSMLM), which fundamentally maximizes signal-to-noise and emitter density due to lowest possible photon spread on the detector. In our implementation, we used a single optical component add-on to the detection path leading to a spectral dispersion of just  $\sim 0.2$  nm/nm (spatial/spectral), orders of magnitude lower than typical grating-based sSMLM implementations. With this implementation, we realized a five-times increased emitter density compared to similar approaches, achieved good separation of emitters with a  $\sim 10$  nm intensity-weighted spectral difference in STORM, and were able to observe changes between 0%, 15%, and 55% FRET efficiency in smFRET.

The low spectral dispersion allows us to use subpixel localization algorithms, a field that is advancing rapidly,<sup>42</sup> and all future developments are directly applicable to our sSMLM implementation, potentially benefiting from custom deep-learning routines addressing both spectral orders.<sup>56,61</sup> This could open up avenues for better spatial and spectral precision, more information obtained from the 1<sup>st</sup> order pattern shape, and sSMLM at even higher emitter density. We additionally note that our implementation can readily be combined with SMLM techniques that modulate excitation patterns to increase localization precision.<sup>49–54</sup> Finally, low-dispersion sSMLM as presented here can be combined with three-dimensional SMLM by engineering the PSF in the emission path.<sup>57–59</sup>

Work by Song et al. employed subpixel localization algorithms for sSMLM, based on data obtained via the  $-1^{\text{st}}$  and  $+1^{\text{st}}$  orders of a nonblazed transmission grating.<sup>38</sup> While this solution is elegant and uses all photons that arrive on the camera chip for both spatial and spectral localization, blocking out the 0<sup>th</sup> order leads to a significant loss of photons. Moreover, the implementation requires additional optical components (mirrors and lenses) to direct only the  $-1^{\text{st}}$  and  $+1^{\text{st}}$  orders on the camera chip. We additionally note that minimizing the spectral dispersion maximizes spectral precision by minimization of effects by shot-, background-, and read-out noise.<sup>48</sup>

Further minimization of spectral dispersion in our implementation is possible by using a lower dispersion blazed grating by decreasing the effective distance of the grating to the camera chip, if the camera housing permits, or by placing the grating close to an intermediate image plane. While this could result in better spectral resolution and higher achievable sSMLM density, a trade-off of this minimization is decreased information content about the shape of the emission spectrum and the risk of imaging the grating itself on the camera. In this study, we used the shape of the emission spectrum only to discriminate FRET states from the donor-only state.

Taken together, we believe that our implementation of low-dispersion sSMLM will find widespread use due to its inherent simplicity and photon efficiency providing access to maximized spatiotemporal and spectral resolution. We further envision applications in which the photon efficient separation of spectrally different entities is desired such as in low-signal flow cytometry. Here, the ideas taken from super-resolution microscopy such as subpixel localization and spectral peak

determination can be equally applied even for low-magnification configurations.

## ■ ASSOCIATED CONTENT

### Data Availability Statement

The code used in this manuscript can be found on Github, <https://github.com/HohlbeinLab/sSMLMAnalyzer>. Data underlying the manuscript are available at [10.5281/zenodo.6778964](https://doi.org/10.5281/zenodo.6778964).

### Supporting Information

The Supporting Information is available free of charge at <https://pubs.acs.org/doi/10.1021/acs.nanolett.2c03140>.

Detailed description of used materials and methods; calibration of low-dispersion sSMLM; technical showcase of triple-fluorophore multiplexing with single-wavelength excitation; ssDNA sequences used for DNA-PAINT on nanoparticles; algorithmic determination of pairing distance and rotation (PDF)

## ■ AUTHOR INFORMATION

### Corresponding Authors

**Koen J. A. Martens** – *Laboratory of Biophysics, Wageningen University and Research, 6708 WE Wageningen, The Netherlands*; Present Address: Institute for Microbiology and Biotechnology, Rheinische-Friedrich-Wilhelms-Universität Bonn, Bonn, Germany; [orcid.org/0000-0002-9447-8579](https://orcid.org/0000-0002-9447-8579); Email: [koenjamartens@gmail.com](mailto:koenjamartens@gmail.com)

**Johannes Hohlbein** – *Laboratory of Biophysics, Wageningen University and Research, 6708 WE Wageningen, The Netherlands*; *Microspectroscopy Research Facility, Wageningen University and Research, 6708 WE Wageningen, The Netherlands*; [orcid.org/0000-0001-7436-2221](https://orcid.org/0000-0001-7436-2221); Email: [Johannes.Hohlbein@wur.nl](mailto:Johannes.Hohlbein@wur.nl)

### Authors

**Martijn Gobes** – *Laboratory of Biophysics, Wageningen University and Research, 6708 WE Wageningen, The Netherlands*

**Emmanouil Archontakis** – *Department of Biomedical Engineering, Institute for Complex Molecular Systems (ICMS), Eindhoven University of Technology, 5600 MB Eindhoven, Netherlands*

**Roger R. Brillas** – *Department of Biomedical Engineering, Institute for Complex Molecular Systems (ICMS), Eindhoven University of Technology, 5600 MB Eindhoven, Netherlands*

**Niels Zijlstra** – *Laboratory of Biophysics, Wageningen University and Research, 6708 WE Wageningen, The Netherlands*

**Lorenzo Albertazzi** – *Department of Biomedical Engineering, Institute for Complex Molecular Systems (ICMS), Eindhoven University of Technology, 5600 MB Eindhoven, Netherlands*; *Nanoscopy for Nanomedicine, Institute for Bioengineering of Catalonia, 08028 Barcelona, Spain*; [orcid.org/0000-0002-6837-0812](https://orcid.org/0000-0002-6837-0812)

Complete contact information is available at:

<https://pubs.acs.org/doi/10.1021/acs.nanolett.2c03140>

### Author Contributions

Conceptualization: K.J.A.M., J.H. Data curation: K.J.A.M. Formal analysis: K.J.A.M., M.G. Funding acquisition: J.H., L.A. Investigation: K.J.A.M., M.G., N.Z., E.A., R.R.B., L.A. Methodology: K.J.A.M., M.G., E.A., R.R.B. Project admin-



istration: J.H. Software: K.J.A.M., M.G. Supervision: N.Z., J.H. Visualization: K.J.A.M. Writing—original draft: K.J.A.M. Writing—review and editing: All authors.

## Notes

The authors declare the following competing financial interest(s): Wageningen University filed a patent application describing this method (WO2022053649).

## ACKNOWLEDGMENTS

The authors thank current and previous members of our laboratories for stimulating discussions and ongoing support. We thank Dr. Mattia Fontana for help on the smFRET samples. This manuscript is part of several research projects (#KIEM.K20.01.054 of the research programme NWO KIEM 2020, #18854 of the research programme NWO Take-off phase I), which are (partly) financed by the Dutch Research Council (NWO). K.J.A.M. is funded by a VLAG PhD-fellowship grant awarded to J.H. R.R.B. is funded by the European Research Council/Horizon 2020 (ERC-StG-757397) and E.A. is funded by the MSCA ITN project THERACAT (765497), both awarded to L.A. We further acknowledge support from a Road to Innovation grant from the Value Creation Office at Wageningen University & Research.

## REFERENCES

- (1) Rust, M. J.; Bates, M.; Zhuang, X. Sub-diffraction-limit imaging by stochastic optical reconstruction microscopy (STORM). *Nat. Methods* **2006**, *3*, 793.
- (2) Hell, S. W.; Wichmann, J. Breaking the diffraction resolution limit by stimulated emission: stimulated-emission-depletion fluorescence microscopy. *Optics Letters* **1994**, *19*, 780–782.
- (3) Betzig, E.; et al. Imaging intracellular fluorescent proteins at nanometer resolution. *Science* **2006**, *313*, 1642–1645.
- (4) Heilemann, M.; et al. Subdiffraction-Resolution Fluorescence Imaging with Conventional Fluorescent Probes. *Angew. Chem., Int. Ed.* **2008**, *47*, 6172–6176.
- (5) Jungmann, R.; et al. Multiplexed 3D cellular super-resolution imaging with DNA-PAINT and Exchange-PAINT. *Nat. Methods* **2014**, *11*, 313.
- (6) Sharonov, A.; Hochstrasser, R. M. Wide-field subdiffraction imaging by accumulated binding of diffusing probes. *Proc. Natl. Acad. Sci. U.S.A.* **2006**, *103*, 18911–18916.
- (7) Delcanale, P.; Miret-Ontiveros, B.; Arista-Romero, M.; Pujals, S.; Albertazzi, L. Nanoscale Mapping Functional Sites on Nanoparticles by Points Accumulation for Imaging in Nanoscale Topography (PAINT). *ACS Nano* **2018**, *12*, 7629–7637.
- (8) Fuentes, E.; et al. PAINT-ing Fluorenylmethoxycarbonyl (Fmoc)-Diphenylalanine Hydrogels. *Chem. Eur. J.* **2020**, *26*, 9869–9873.
- (9) Hess, S. T.; Girirajan, T. P. K.; Mason, M. D. Ultra-High Resolution Imaging by Fluorescence Photoactivation Localization Microscopy. *Biophys. J.* **2006**, *91*, 4258–4272.
- (10) Manley, S.; et al. High-density mapping of single-molecule trajectories with photoactivated localization microscopy. *Nat. Methods* **2008**, *5*, 155–157.
- (11) Shen, H.; et al. Single Particle Tracking: From Theory to Biophysical Applications. *Chem. Rev.* **2017**, *117*, 7331–7376.
- (12) Lelek, M.; et al. Single-molecule localization microscopy. *Nat. Rev. Methods Primers* **2021**, *1*, 1–27.
- (13) Liu, S.; Hoess, P.; Ries, J. Super-Resolution Microscopy for Structural Cell Biology. *Annu. Rev. Biophys.* **2022**, *51*, 301.
- (14) Uphoff, S.; Reyes-Lamothe, R.; Garza de Leon, F.; Sherratt, D. J.; Kapanidis, A. N. Single-molecule DNA repair in live bacteria. *Proc. Natl. Acad. Sci. U.S.A.* **2013**, *110*, 8063–8068.
- (15) Elf, J.; Barkefors, I. Single-Molecule Kinetics in Living Cells. *Annu. Rev. Biochem.* **2019**, *88*, 635–659.
- (16) Vink, J. N. A.; et al. Direct Visualization of Native CRISPR Target Search in Live Bacteria Reveals Cascade DNA Surveillance Mechanism. *Mol. Cell* **2020**, *77*, 39–50.e10.
- (17) Martens, K. J. A.; et al. Visualisation of dCas9 target search in vivo using an open-microscopy framework. *Nat. Commun.* **2019**, *10*, 3552.
- (18) Sauer, M.; Heilemann, M. Single-Molecule Localization Microscopy in Eukaryotes. *Chem. Rev.* **2017**, *117*, 7478–7509.
- (19) Leterrier, C.; et al. Nanoscale Architecture of the Axon Initial Segment Reveals an Organized and Robust Scaffold. *Cell Rep.* **2015**, *13*, 2781–2793.
- (20) Turkowyd, B.; Virant, D.; Endesfelder, U. From single molecules to life: microscopy at the nanoscale. *Anal. Bioanal. Chem.* **2016**, *408*, 6885–6911.
- (21) Wöll, D.; Flors, C. Super-resolution Fluorescence Imaging for Materials Science. *Small Methods* **2017**, *1*, 1700191.
- (22) Martens, K.; van Duynhoven, J.; Hohlbein, J. Spatiotemporal heterogeneity of  $\kappa$ -carrageenan gels investigated via single-particle-tracking fluorescence microscopy. *Langmuir* **2020**, *20*, 5502–5509.
- (23) Pujals, S.; Feiner-Gracia, N.; Delcanale, P.; Voets, I.; Albertazzi, L. Super-resolution microscopy as a powerful tool to study complex synthetic materials. *Nat. Rev. Chem.* **2019**, *3*, 68–84.
- (24) Greenspan, P.; Fowler, S. D. Spectrofluorometric studies of the lipid probe, Nile red. *J. Lipid Res.* **1985**, *26*, 781–789.
- (25) Halder, A.; et al. Lipid chain saturation and the cholesterol in the phospholipid membrane affect the spectroscopic properties of lipophilic dye Nile red. *Spectrochimica Acta Part A: Molecular and Biomolecular Spectroscopy* **2018**, *191*, 104–110.
- (26) Hohlbein, J.; Craggs, T. D.; Cordes, T. Alternating-laser excitation: single-molecule FRET and beyond. *Chem. Soc. Rev.* **2014**, *43*, 1156–1171.
- (27) Hellenkamp, B.; et al. Precision and accuracy of single-molecule FRET measurements—a multi-laboratory benchmark study. *Nat. Methods* **2018**, *15*, 669–676.
- (28) Purohit, A.; Benigno, S. P. C.; Richtering, W.; Wypyssek, S. K.; Wöll, D. Microgel PAINT—Nanoscale Polarity Imaging of Adaptive Microgels without Covalent Labelling. *Chem. Sci.* **2019**, *10*, 10336–10342.
- (29) Baddeley, D.; et al. 4D super-resolution microscopy with conventional fluorophores and single wavelength excitation in optically thick cells and tissues. *PLoS one* **2011**, *6*, No. e20645.
- (30) Testa, I.; et al. Multicolor fluorescence nanoscopy in fixed and living cells by exciting conventional fluorophores with a single wavelength. *Biophys. J.* **2010**, *99*, 2686–2694.
- (31) Gimber, N.; Strauss, S.; Jungmann, R.; Schmoranzler, J. Simultaneous Multicolor DNA-PAINT without Sequential Fluid Exchange Using Spectral Demixing. *Nano Lett.* **2022**, *22*, 2682–2690.
- (32) Shechtman, Y.; Weiss, L. E.; Backer, A. S.; Lee, M. Y.; Moerner, W. E. Multicolour localization microscopy by point-spread-function engineering. *Nat. Photonics* **2016**, *10*, 590.
- (33) Broeken, J.; Rieger, B.; Stallinga, S. Simultaneous measurement of position and color of single fluorescent emitters using diffractive optics. *Opt. Lett.* **2014**, *39*, 3352–3355.
- (34) Dong, B.; et al. Super-resolution spectroscopic microscopy via photon localization. *Nat. Commun.* **2016**, *7*, 1–8.
- (35) Mlodzianoski, M. J.; Curthoys, N. M.; Gunewardene, M. S.; Carter, S.; Hess, S. T. Super-resolution imaging of molecular emission spectra and single molecule spectral fluctuations. *PLoS one* **2016**, *11*, No. e0147506.
- (36) Zhang, Z.; Kenny, S. J.; Hauser, M.; Li, W.; Xu, K. Ultrahigh-throughput single-molecule spectroscopy and spectrally resolved super-resolution microscopy. *Nat. Methods* **2015**, *12*, 935.
- (37) Bongiovanni, M. N.; et al. Multi-dimensional super-resolution imaging enables surface hydrophobicity mapping. *Nat. Commun.* **2016**, *7*, 13544.



(38) Song, K.-H.; Zhang, Y.; Brenner, B.; Sun, C.; Zhang, H. F. Symmetrically dispersed spectroscopic single-molecule localization microscopy. *Light Sci. Appl.* **2020**, *9*, 92.

(39) Jeffet, J.; et al. Multimodal single-molecule microscopy with continuously controlled spectral resolution (CoCoS). *Biophys. J.* **2022**, *121*, 301a.

(40) Kumar Gaire, S.; et al. Accelerating multicolor spectroscopic single-molecule localization microscopy using deep learning. *Biomed. Opt. Express* **2020**, *11*, 2705–2721.

(41) Song, K.-H.; et al. Monolithic dual-wedge prism-based spectroscopic single-molecule localization microscopy. *Nanophotonics* **2022**, *11*, 1527–1535.

(42) Sage, D.; et al. Super-resolution fight club: assessment of 2D and 3D single-molecule localization microscopy software. *Nat. Methods* **2019**, *16*, 387.

(43) Li, Y.; et al. Real-time 3D single-molecule localization using experimental point spread functions. *Nat. Methods* **2018**, *15*, 367–369.

(44) Martens, K. J. A.; Bader, A. N.; Baas, S.; Rieger, B.; Hohlbein, J. Phasor based single-molecule localization microscopy in 3D (pSMLM-3D): An algorithm for MHz localization rates using standard CPUs. *J. Chem. Phys.* **2018**, *148*, 123311.

(45) Smith, C. S.; Joseph, N.; Rieger, B.; Lidke, K. A. Fast, single-molecule localization that achieves theoretically minimum uncertainty. *Nat. Methods* **2010**, *7*, 373–375.

(46) Martens, K. J. A.; Jabermoradi, A.; Yang, S.; Hohlbein, J. Integrating engineered point spread functions into the phasor-based single-molecule localization microscopy framework. *Methods* **2021**, *193*, 107–115.

(47) Butler, C.; et al. Multi-Dimensional Spectral Single Molecule Localization Microscopy. *Front. Bioinform.* **2022**, *2*, 813494.

(48) Song, K.-H.; Dong, B.; Sun, C.; Zhang, H. F. Theoretical analysis of spectral precision in spectroscopic single-molecule localization microscopy. *Rev. Sci. Instrum.* **2018**, *89*, 123703.

(49) Balzarotti, F.; et al. Nanometer resolution imaging and tracking of fluorescent molecules with minimal photon fluxes. *Science* **2017**, *355*, 606–612.

(50) Reymond, L.; et al. SIMPLE: Structured illumination based point localization estimator with enhanced precision. *Opt. Express*, **2019**, *27*, 24578–24590.

(51) Gu, L.; et al. Molecular resolution imaging by repetitive optical selective exposure. *Nat. Methods* **2019**, *16*, 1114–1118.

(52) Cnossen, J.; et al. Localization microscopy at doubled precision with patterned illumination. *Nat. Methods* **2020**, *17*, 59–63.

(53) Jouchet, P.; et al. Nanometric axial localization of single fluorescent molecules with modulated excitation. *Nat. Photonics* **2021**, *15*, 297–304.

(54) Gwosch, K. C.; et al. MINFLUX nanoscopy delivers 3D multicolor nanometer resolution in cells. *Nat. Methods* **2020**, *17*, 217–224.

(55) Xu, F.; et al. Three-dimensional nanoscopy of whole cells and tissues with in situ point spread function retrieval. *Nat. Methods* **2020**, *17*, 531–540.

(56) Nehme, E.; et al. DeepSTORM3D: dense 3D localization microscopy and PSF design by deep learning. *Nat. Methods* **2020**, *17*, 734–740.

(57) Huang, B.; Wang, W.; Bates, M.; Zhuang, X. Three-Dimensional Super-Resolution Imaging by Stochastic Optical Reconstruction Microscopy. *Science* **2008**, *319*, 810–813.

(58) Shechtman, Y.; Weiss, L. E.; Backer, A. S.; Sahl, S. J.; Moerner, W. E. Precise three-dimensional scan-free multiple-particle tracking over large axial ranges with tetrapod point spread functions. *Nano Lett.* **2015**, *15*, 4194–4199.

(59) Pavani, S. R. P.; et al. Three-dimensional, single-molecule fluorescence imaging beyond the diffraction limit by using a double-helix point spread function. *Proc. Natl. Acad. Sci. U.S.A.* **2009**, *106*, 2995–2999.

(60) Reymond, L.; Huser, T.; Ruprecht, V.; Wieser, S. Modulation-enhanced localization microscopy. *J. Phys. Photonics* **2020**, *2*, 041001.

(61) Speiser, A.; et al. Deep learning enables fast and dense single-molecule localization with high accuracy. *Nat. Methods* **2021**, *18*, 1082–1090.

## Recommended by ACS

### Single-Shot High-Throughput Phase Imaging with Multibeam Array Interferometric Microscopy

Jie Dong, Alexander W. Koch, et al.

NOVEMBER 18, 2021  
ACS PHOTONICS

READ 

### Visualizing Quantum Coherence Based on Single-Molecule Coherent Modulation Microscopy

Haitao Zhou, Suotang Jia, et al.

JANUARY 28, 2021  
NANO LETTERS

READ 

### Calibration-Free Single-Molecule Absolute Quantification Using Super-resolution Microscopy

Guang Li and Qiqing Zhang

APRIL 07, 2021  
ANALYTICAL CHEMISTRY

READ 

### Cryo-nanoimaging of Single Human Macrophage Cells: 3D Structural and Chemical Quantification

Chiara Gramaccioni, Peter Cloetens, et al.

MARCH 12, 2020  
ANALYTICAL CHEMISTRY

READ 

Get More Suggestions >