



Enfermedades Infecciosas y Microbiología Clínica

www.elsevier.es/eimc



Original

Evaluation and optimization of the Sysmex UF1000i system for the screening of urinary tract infection in primary health care elderly patients



Guillermo Martín-Gutiérrez^{a,*}, Ana Porras-González^a, Carlos Martín-Pérez^b, Jose Antonio Lepe^{a,c}, Javier Aznar^{a,c}

^a Infectious Diseases, Clinical Microbiology and Preventive Medicine Unit, Hospital Universitario Virgen del Rocío, Seville, Spain

^b UGC Marquesado, AGS Nordeste de Granada; SAMSERAP group, Spain

^c Institute of Biomedicine of Seville (IBiS), University Hospital Virgen del Rocío/CSIC/University of Seville, Seville, Spain

ARTICLE INFO

Article history:

Received 9 April 2014

Accepted 25 July 2014

Available online 18 October 2014

Keywords:

UTI

Elderly

Flow cytometry

ABSTRACT

Objective: Urinary tract infections (UTIs) are a common problem in the elderly population. Urine culture is still considered the “gold standard” to diagnose infection in this population. However, urine cultures are laborious and costly, and most samples will yield no growth.

Methods: An evaluation was made of the Sysmex UF-1000i flow cytometer as a screening tool for UTI in an elderly population older than 65 years who lived in the community, using 346 urine samples submitted for culture.

Results: The Receiver Operating Characteristic (ROC) analysis showed a significant difference ($P < 0.01$) between 0.98 bacteria area under the curve value and 0.82 of white blood cells (WBC). The combination of both counts for screening did not show any improvement in specificity or sensitivity. According to our data, the use of a single cut-off point of 200 bacteria/ μL is suggested, in which the sensitivity and specificity were 99.11% and 91.59%, respectively, with a NPV of 99.49%. Moreover, this cut-off value could avoid 60.24% of the samples to be cultured, with a minimal false negative results rate of 0.87%.

Conclusions: The stratification of age groups stratification helps in selecting a more adjusted Sysmex UF1000i cut-off limit, leading to an improvement in the screening parameters that would imply a better management of these infections, as well as a high reduction in the workload and cost savings.

© 2014 Elsevier España, S.L.U. and Sociedad Española de Enfermedades Infecciosas y Microbiología Clínica. All rights reserved.

Evaluación y optimización del sistema Sysmex UF-1000i como método de cribaje para el estudio de infecciones urinarias en pacientes ancianos de Atención Primaria

RESUMEN

Objetivo: Evaluar y optimizar el uso del citómetro de flujo (Sysmex UF1000i®) como cribado para las infecciones urinarias (ITUs) en pacientes ≥ 65 años procedentes de Atención Primaria.

Métodos: Se estudiaron 346 orinas de pacientes ≥ 65 años con sospecha de infección urinaria, enviadas al Hospital Universitario Virgen del Rocío, durante el periodo enero-mayo 2013. Las muestras se estudiaron mediante citometría de flujo y cultivo cuantitativo en medio cromogénico.

Resultados: Se incluyeron 346 pacientes, cuya edad media fue de $76,70 \pm 0,75$ años. De las 346 muestras 113 (32,65%) fueron positivas, 214 (61,84%) negativas y 19 (5,49%) contaminadas. El área bajo la curva ROC utilizando el número de bacterias (0,98) fue mayor que para los leucocitos (0,82), existiendo diferencias significativas entre ellas ($P < 0.01$). El estudio conjunto de bacterias y leucocitos no supuso ninguna mejora,

Palabras clave:

UTI

Mayores

Citómetro de flujo

* Corresponding author.

E-mail address: guiller_mg86@hotmail.es (G. Martín-Gutiérrez).

por lo que se utilizaron distintos umbrales basados en el número de bacterias. De acuerdo con nuestros datos, proponemos un punto de corte de 200 bacterias/ μL con el cual obtenemos una sensibilidad del 99,11% y especificidad del 91,59%, con un *valour* predictivo negativo del 99,49%. Además este punto de corte nos permitiría evitar el 60,24% de los cultivos, con una mínima tasa de falsos negativos (0,87%).

Conclusiones: La categorización de la población según criterios de positividad para el diagnóstico de la ITU, así como en grupos etarios con las mismas condiciones clínicas, permite incrementar la exactitud de los resultados obtenidos, reduciendo la carga de trabajo sustancialmente así como los costes asociados.

© 2014 Elsevier España, S.L.U. y Sociedad Española de Enfermedades Infecciosas y Microbiología Clínica.
Todos los derechos reservados.

Introduction

In the elderly population urinary tract infections (UTIs) are common, representing the second most frequent infection in elderly women living in the community.^{1,2} The diagnosis of UTI requires the presence of significant bacteriuria ($\geq 10^5$ Colony Forming Unit (CFU) per mL) and genitourinary symptoms. However, these patients may not refer typical urinary complaints; thus, the diagnosis may be more challenging than in other age groups. Most of these cases are self-limited and have no long-term sequelae, but underlying structural or functional abnormalities of the genitourinary tract are not rare,³ and serious complications may develop, such as pyelonephritis or sepsis.^{4,5}

Microbiologic cultures are still considered the diagnostic “gold standard”, allowing identification and quantification of the causal agents.⁶ Nevertheless, culturing of the samples is both time and labour consuming, with most of the samples yielding no growth.^{3,7,8} Screening methods can improve the laboratory efficiency by ruling out UTI-negative samples, thereby reducing the workload. The Sysmex UF-1000i (TOA Medical Electronics, Kobe, Japan) is an automated urine particle analyzer using laser-based fluorescent flow cytometry. Some data have shown that it is valid as a screening test for UTIs, suggesting that cut-off values should vary depending on both age and gender.^{9–11} Further studies are required to establish the value of this system, interpret the results and adapt the criteria of a positive result to the characteristics of a given population.

In 2012, 34,128 urine samples from Primary Care Units were received at the Microbiology Service of the Virgen del Rocío University Hospital, out of which 10,223 (29.95%) were from elderly patients, 70% being negative cultures (data not shown). This high number of negative specimens emphasizes the need for a urine samples screening method prior to be cultured, and therefore to improve the laboratory efficiency by reducing not only the economical and workload costs but also by shortening the laboratory response time.

The aim of this study was to evaluate and to optimize the use of the Sysmex UF-1000i as a screening method for urine samples obtained from an in community-dwelling elderly population older than 65 year.

Materials and methods

Patients and urine samples

From January 2013 to June 2013, 346 patients attending at the Primary Care Units of Virgen del Rocío University Hospital were selected using a systematic random sampling. Four to five urine samples per day submitted for culture from elderly outpatients (≥ 65 years old) were randomly selected. Sample size was determined by the Carley et al. method,¹² using a 95% of sensitivity and a precision of 5% for the expected UTI prevalence of elderly

patients. Midstream catch urine was collected in sterile preservative tubes SRO-1-25B with boric acid (Soria Melguizo S.A., Madrid, Spain), transported at cold temperature and processed within 4–8 h after the sampling.

Culture and urinalysis

Ten microlitres of the urine specimen were quantitatively cultured onto Brilliance UTI Clarity Agar (Oxoid, Basingstoke, UK) plates. All plates were aerobically incubated for 18–24 h at 37 °C, and the results were expressed as the number of colony-forming units (CFUs) per millilitre. A threshold of $\geq 10^5$ CFUs/mL for women and $\geq 10^4$ CFUs/mL for men was established for positive cultures. The presence of two or more different isolates as well as the growth of one or more non-pathogens was defined as contamination of the specimen. Identification of the isolates was performed by conventional biochemical tests (biochemical testing, pigment production, growth, and colony characteristics) and MicroScan WalkAway® plus System (Siemens Healthcare Diagnostics, WestSacramento, CA). When the identification was uncertain, it was confirmed by Bruker Biotyper MALDI-TOF MS system (Bruker Daltonik GmbH, Leipzig, Germany).

All of the urine specimens were also analyzed by the Sysmex UF-1000i, which allows the discrimination and quantification of bacteria, erythrocytes, WBC, epithelial cells, casts, crystals, fungi, and sperm. The results obtained with Sysmex UF-1000i and those from the urine cultures were compared.

Statistics

A logistic regression model was performed, in which age, gender, bacterial count and WBC count were included as independent variables, to predict the probability of a positive culture. The cut-off values for bacteria and leucocytes were evaluated in the Sysmex UF-1000i according to the area under the ROC curve, which was estimated by using the Hanley and McNeil's nonparametric method.¹³ The bacteria and leucocytes ROC curves were compared by the De Long et al. method.¹⁴ Sensitivity, specificity, positive predictive value (PPV) and negative predictive value (NPV) were calculated for those cut-off points. A *P* value ≤ 0.05 was considered as significant. The data analysis was carried out using the software STATA Release 10.1 statistical package (StataCorp LP, Lakeway Drive, TX, USA).

Results

In total, 346 elderly patients, 261 women (75, 43%) and 85 men (24, 56%) were included in this study. The mean age was 76.70 ± 0.75 years. One hundred and thirteen (32.65%) of the urine samples yielded positive cultures, 214 (61.84%) were negative and 19 (5.49%) were considered as contaminated and not further tested. The clinical isolates obtained were: *Escherichia coli* (72), *Klebsiella pneumoniae*,¹³ *Proteus mirabilis*,⁹ *Enterococcus faecalis*,⁵




Fig. 1. Receiver operating characteristic curve of bacterial and WBC counts on the UF-1000i flow cytometer.

Pseudomonas aeruginosa,⁴ ESBL-*E. coli*,³ ESBL-*K. pneumoniae*,¹ *Klebsiella oxytoca*,¹ *Serratia liquefaciens*,¹ *Morganella morgani*,¹ *Citrobacter freundii*,¹ *Citrobacter koseri*¹ and *Citrobacter amalonaticus*.¹

The ROC curves for WBC and bacterial counts from the UF-1000i are given in Fig. 1. The area under the curve (AUC) for bacterial counts was 0.98, a significantly higher value than the 0.82 AUC for WBC count ($P < 0.01$). WBC count was not as effective as the former as screening method for UTI, and the use of both bacterial and WBC counts did not show any advantage in terms of sensitivity or specificity compared with bacterial counts alone. Therefore, only different cut-off values for bacterial counts have been applied to predict community-acquired UTI in the elderly. Parameters associated with the different cut-off values for bacterial counts generated by Sysmex UF-1000i are shown in Table 1.

The most balanced cut-off value was 338 bacteria/ μL , with sensitivity, specificity and negative predictive (NP) values of 93.81%, 93.93% and 96.39% respectively. Nevertheless, screening for negative urine samples requires both a high sensitivity and high NP values in order to minimize the number of false-negative results. Sensitivity and specificity values of 99.11% and 91.59% respectively, with a NPV of 99.49% (Table 2) were obtained with a cut-off of 200 bacteria/ μL .

Discussion

Several studies, with a considerable heterogeneity in variables such as gender, age or clinical condition (in or outpatients, comorbidities, use of medical devices, etc) have shown the usefulness of

the UF-1000i screening for UTI.^{4,6,9} The use of a unique standard cut-off value for a heterogeneous population results in over- or under-culturing specimens among different groups, thereby affecting both predictive values and diminishing the reliability of the screening. In addition, stratification according to different established UFC/mL criteria for the diagnosis of UTI is likely to increase the accuracy of the studies' results in different groups of populations.¹⁵ These and other factors should be also considered when deciding the proper cut-off value in a precise group of population. Regarding elderly people, the criteria of 10^5 UFC/mL for women and 10^4 UFC/mL for men meet the standard one, but the higher prevalence of UTI increases the PPV and decreases the NPV. Moreover, misclassification of patients with or without UTI may have serious implications due to frequent presence of multimorbidity, polytreatment and impaired renal function in this age group. All these factors emphasize the convenience to lower the ROC curve most balanced cut-off value in order to raise both sensitivity and NPV, and therefore to achieve a lower false negative rate. In addition, fast and reliable negative results may allow an antibiotic treatment discontinuation, bringing along a decrease in the risk of drug interactions and side effects in an often-multimorbid patient.

In our study, the sensitivity, specificity, and AUC of bacterial count in the Sysmex UF-1000i analyzer system were higher than those of WBC count, and the combination of both counts for screening did not show any specificity or sensitivity improvements to bacterial counts alone, as previously reported.⁷ Nevertheless, in contrast with these results, some articles showed the effectiveness of screening with WBC plus bacterial counts,^{6,10} with an increase

Table 1
Parameters depending on the cut-off values for the bacterial count.

Definition	S ^a	95% CI	SP ^a	95% CI	PPV ^a	95% CI	NPV ^a	95% CI
≥ 50 bacteria/ μL	100	96.71–100	73.83	67.55–79.26	66.86	59.46–73.51	100	97.62–100
≥ 100 bacteria/ μL	100	96.71–100	83.18	77.59–87.59	75.84	68.37–82.00	100	97.89–100
≥ 150 bacteria/ μL	99.11	95.16–99.84	87.38	82.26–91.18	80.57	73.21–86.29	99.47	97.05–99.91
≥ 200 bacteria/ μL	99.11	95.16–99.84	91.59	87.09–94.61	86.15	79.17–91.06	99.49	97.18–99.91
≥ 250 bacteria/ μL	96.46	91.25–98.61	92.06	87.64–94.98	86.50	79.45–91.40	98.00	94.99–99.22
≥ 300 bacteria/ μL	93.80	87.76–96.97	92.52	88.20–95.34	86.88	79.75–91.76	96.58	93.12–98.33
≥ 350 bacteria/ μL	92.92	86.65–96.37	94.39	90.45–96.76	89.74	82.92–94.03	96.19	92.66–98.05

^a S, sensitivity; SP, specificity; PPV, positive predictive value; NPV, negative predictive value.

Table 2
Number and percentage of savings with different cut-off values.

Definition	No. necessary cultures	No. unnecessary cultures	% unnecessary cultures	% false negatives	95% CI
≥50 bacteria/μL	169	158	48.32	0	0–3.29
≥100 bacteria/μL	149	178	54.43	0	0–3.29
≥150 bacteria/μL	139	188	57.49	0.87	0.17–4.84
≥200 bacteria/μL	130	197	60.24	0.87	0.17–4.84
≥250 bacteria/μL	126	201	61.47	3.54	1.38–8.75
≥300 bacteria/μL	121	206	63.00	6.19	3.03–12.28
≥350 bacteria/μL	117	210	64.22	7.08	3.63–13.35

in sensitivity but a decrease in specificity. This difference could be attributable to the two different populations studied, outpatients only⁷ or both in- and outpatients.^{6,10} Moreover, some studies reveal that pyuria is not a good marker of UTI. In the study by Kishore et al., the relationship between pyuria and culture positivity did not reach statistical significance in both males and females in the community elderly.¹⁶ Another study performed by Kupelian et al.¹⁷ reported that pyuria demonstrates poor sensitivity in patients with UTI. A systematic review and meta-analysis performed by Shang et al. conclude that UF-1000i may be used as an effective screening method for UTI by measuring WBC and bacterial counts of urine samples.¹⁸ Nevertheless, in this study the overall estimates of sensitivity and specificity show a very high heterogeneity ($i^2 > 90\%$), limiting the validity of these results. Broeren et al.³ hypothesized that the lack of improvement shown in some studies was due to the lack of use of boric acid, which acts as a stabilizing compound. Kupelian et al.¹⁷ describes that cell destruction appears to be retarded by boric acid, although significant cell loss appears inevitable. Our data do not support this hypothesis, as we used boric acid containers and no improvement was observed. It has been suggested that gender-specific cut-off points could improve the screening,¹⁰ but again we did not find any significant differences between men and women for bacterial counts.

According to our data, the recommended cut-off value is 200 bacteria/μL, higher than those mentioned in previous studies,^{6,7,19–22} and with which, we obtain a 60, 24% reduction of the samples to be cultured, with a very low 0.87% false negative rate. The only false negative culture result found was $a > 10^5$ CFU/mL *Proteus mirabilis* culture. False negative results mainly with Gram-positive pathogens,^{14,17} but also with Gram-negatives have been documented with the use of Sysmex UF-1000i.^{23,24} Further studies using larger sample sizes would probably permit to address this issue. A limitation in the design of our study is that the cohort was selected from urine samples sent to the laboratory instead of from a group of patients eligible for UTI.

Conclusions

In conclusion, age-groups stratification allows selecting a more adjusted Sysmex UF1000i cut-off, bringing along an improvement in the screening parameters that would imply a high reduction in the workload and cost savings. In the community-dwelling elderly and according to our data, a 200 bacteria/μL cut-off value should be used. This would mean reducing waiting times in the case of negative cultures, thus eliminating possible antibiotic adverse effects and interactions, with a minimal false negatives rate. Further prospective studies are needed in order to contrast these data with larger populations, and to establish the more convenient cut-off for each group of patients.

Conflict of interest

The authors declare no conflict of interest.

References

- Matthews SJ, Lancaster J. Urinary tract infections in the elderly population. *Am J of Geriatr Pharmacother.* 2011;9:286–309.
- Schmiemann G, Kniehl E, Gebhardt K, Matejczyk MM, Hummers-Pradier E. The diagnosis of urinary tract infection. A systematic review. *Dtsch Arztebl Int.* 2010;107:361–7.
- Beveridge LA, Davey PG, Phillips G, McMurdo ME. Optimal management of urinary tract infections in older people. *Clin Interv Aging.* 2011;6:173–80.
- Broeren MA, Bahceci S, Vader HL, Arents NL. Screening for urinary tract infection with the Sysmex UF-1000i urine flow cytometer. *J Clin Microbiol.* 2011;49:1025–9.
- Yates M. Urinary tract infection in the elderly. *Aust J Hosp Pharmacy.* 1999;29:166–70.
- Gutiérrez-Fernández J, Lara A, Bautista MF, de Dios Luna J, Polo P, Miranda C, et al. Performance of the Sysmex UF1000i system in screening for significant bacteriuria before quantitative culture of aerobic/facultative fast-growth bacteria in a reference hospital. *J Appl Microbiol.* 2012;113:609–14.
- Hu X, Zhang J, Zhang X. Evaluation of the Sysmex UF-1000i urine analyzer as a screening test to reduce the need for urine cultures for urinary tract infection. *Lab Med.* 2010;41:349–52.
- Kim SY, Kim YJ, Lee SM, Hwang SH, Kim HH, Son HC, et al. Evaluation of the Sysmex UF-100 urine cell analyzer as a screening test to reduce the need for urine cultures for community-acquired urinary tract infection. *Am J Clin Pathol.* 2007;128:922–5.
- Pieretti B, Brunati P, Pini B, Colzani C, Congedo P, Rocchi M, et al. Diagnosis of bacteriuria and leukocyturia by automated flow cytometry compared with urine culture. *J Clin Microbiol.* 2010;48:3990–6.
- Giesen CD, Greeno AM, Thompson KA, Patel R, Jenkins SM, Lieske JC. Performance of flow cytometry to screen urine for bacteria and white blood cells prior to urine culture. *Clin Biochem.* 2013;46:810–3.
- Jolkkonen S, Paattiniemi EL, Kärpänoja P, Sarkkinen H. Screening of urine samples by flow cytometry reduces the need for culture. *J Clin Microbiol.* 2010;3117–22.
- Carley S, Dosman S, Jones SR, Harrison M. Simple nomograms to calculate sample size in diagnostic studies. *Emerg Med J.* 2005;22:180–1.
- Hanley JA, McNeil BJ. The meaning and use of the area under a receiver operating characteristic (ROC) curve. *Radiology.* 1982;143:29–36.
- De Long ER, De Long DM, Clarke-Pearson DL. Comparing the areas under two or more correlated receiver operating characteristic curves: a nonparametric approach. *Biometrics.* 1988;44:837–45.
- Kubik MJ, McCarter YS. Controversies in the diagnosis of urinary tract infections. *Clin Microbiol Newslett.* 2012;34:185–91.
- Kishore D, Gambhir IS, Diwaker A, Khurana V, Dant R, Anupurba S. Study of urinary infection in community based indian elderly. *Indian J Res.* 2013;7:32–9.
- Kupelian AS, Horsley H, Khasriya R, Amussah RT, Badiani R, Courtney AM, et al. Discrediting microscopic pyuria and leucocyte esterase as diagnostic surrogates for infection in patients with lower urinary tract symptoms: results from a clinical and laboratory evaluation. *BJU Int.* 2013;112:231–8.
- Shang Y, Wang Q, Zhang J, Xu Y, Zhang W, Chen Y, et al. Systematic review and meta-analysis of flow cytometry in urinary tract infection screening. *Clin Chim Acta.* 2013;424:90–5.
- Krongvorakul J, Phundhusuwannakul S, Santanirand P, Kunakorn M. A flow cytometric urine analyzer for bacteria and white blood cell counts plus urine dipstick test for rapid screening of bacterial urinary tract infection. *Asian Biomed.* 2012;6:601–8.
- Muñoz-Algarra M, Martínez-Ruiz R, Orden-Martínez B. Evaluación del sistema automatizado UF-1000i® en el diagnóstico de infección urinaria. *Enferm Infecc Microbiol Clin.* 2013;31:29–31.
- Wang J, Zhang Y, Xu D, Shao W, Lu Y. Evaluation of the Sysmex UF-1000i for the diagnosis of urinary tract infection. *Am J Clin Pathol.* 2010;133:577–82.
- Kadkhoda K, Manickam K, Degagne P, Sokolowski P, Pang P, Kontzie N, et al. UF-1000i flow cytometry is an effective screening method for urine specimens. *Diagn Microbiol Infect Dis.* 2011;69:130–6.
- De Rosa R, Grosso S, Bruchetta G, Avolio M, Stano P, Modolo ML. Evaluation of the Sysmex UF 1000i flow cytometer for ruling out bacterial urinary tract infection. *Clin Chim Acta.* 2010;411:1137–42.
- Manoni F, Fornasiero L, Ercolin M, Tinello A, Ferrian M, Hoffer P, et al. Cutoff values for bacteria and leukocytes for urine flow cytometer Sysmex UF-1000i in urinary tract infections. *Diagn Microbiol Infect Dis.* 2009;65:103–7.