

TOPICAL REVIEW • **OPEN ACCESS**

Review and recommendations on deformable image registration uncertainties for radiotherapy applications

To cite this article: Lena Nenoff *et al* 2023 *Phys. Med. Biol.* **68** 24TR01

View the [article online](#) for updates and enhancements.

You may also like

- [Detailed dosimetric evaluation of inter-fraction and respiratory motion in lung stereotactic body radiation therapy based on daily 4D cone beam CT images](#)
Carlos Huesa-Berral, Celia Juan-Cruz, Simon van Kranen *et al.*
- [The distance discordance metric—a novel approach to quantifying spatial uncertainties in intra- and inter-patient deformable image registration](#)
Ziad H Saleh, Aditya P Apte, Gregory C Sharp *et al.*
- [An automated, quantitative, and case-specific evaluation of deformable image registration in computed tomography images](#)
R G J Kierkels, L A den Otter, E W Korevaar *et al.*



TOPICAL REVIEW

OPEN ACCESS

Review and recommendations on deformable image registration uncertainties for radiotherapy applications

RECEIVED
11 April 2023REVISED
30 October 2023ACCEPTED FOR PUBLICATION
15 November 2023PUBLISHED
13 December 2023

Original content from this work may be used under the terms of the [Creative Commons Attribution 4.0 licence](#).

Any further distribution of this work must maintain attribution to the author(s) and the title of the work, journal citation and DOI.



Lena Nenoff^{1,2,3,4,*} , Florian Amstutz^{5,6,7} , Martina Murr⁸, Ben Archibald-Heeren⁹ , Marco Fusella¹⁰ ,
Mohammad Hussein¹¹, Wolfgang Lechner¹² , Ye Zhang⁶ , Greg Sharp^{1,2} and Eliana Vasquez Osorio¹³ 

¹ Department of Radiation Oncology, Massachusetts General Hospital, Boston, MA, United States of America

² Harvard Medical School, Boston, MA, United States of America

³ OncoRay—National Center for Radiation Research in Oncology, Faculty of Medicine and University Hospital Carl Gustav Carus, Technische Universität Dresden, Helmholtz-Zentrum Dresden—Rossendorf, Dresden Germany

⁴ Helmholtz-Zentrum Dresden—Rossendorf, Institute of Radiooncology—OncoRay, Dresden, Germany

⁵ Department of Physics, ETH Zurich, Switzerland

⁶ Center for Proton Therapy, Paul Scherrer Institute, Villigen PSI, Switzerland

⁷ Division of Medical Radiation Physics and Department of Radiation Oncology, Inselspital, Bern University Hospital, and University of Bern, Bern, Switzerland

⁸ Section for Biomedical Physics, Department of Radiation Oncology, University of Tübingen, Germany

⁹ Icon Cancer Centres, Sydney, NSW, Australia

¹⁰ Department of Radiation Oncology, Abano Terme Hospital, Italy

¹¹ Metrology for Medical Physics, National Physical Laboratory, Teddington, United Kingdom

¹² Department of Radiation Oncology, Medical University of Vienna, Austria

¹³ Division of Cancer Sciences, The University of Manchester, Manchester, United Kingdom

* Authors to whom any correspondence should be addressed.

E-mail: lena.nenoff@oncoray.de

Keywords: deformable image registration, DIR uncertainty, radiotherapy, structure propagation, dose accumulation, mapping, image deformation


Supplementary material for this article is available [online](#)

Abstract

Deformable image registration (DIR) is a versatile tool used in many applications in radiotherapy (RT). DIR algorithms have been implemented in many commercial treatment planning systems providing accessible and easy-to-use solutions. However, the geometric uncertainty of DIR can be large and difficult to quantify, resulting in barriers to clinical practice. Currently, there is no agreement in the RT community on how to quantify these uncertainties and determine thresholds that distinguish a good DIR result from a poor one. This review summarises the current literature on sources of DIR uncertainties and their impact on RT applications. Recommendations are provided on how to handle these uncertainties for patient-specific use, commissioning, and research. Recommendations are also provided for developers and vendors to help users to understand DIR uncertainties and make the application of DIR in RT safer and more reliable.

1. Introduction

Deformable image registration (DIR) is used in multiple applications in radiotherapy (RT), including image fusion, contour propagation, dose mapping, and dose accumulation. Many improvements in patient quality of care may be facilitated by DIR, including clinical delineations using multiple images (Brock *et al* 2017, Barber *et al* 2020), organ sparing with adaptive techniques (Albertini *et al* 2020, Glide-Hurst *et al* 2021), and better understanding of patient morbidity and mortality risks incorporating adaptive RT (ART) with accumulated dose (Murr *et al* 2023, Smolders *et al* 2023b). The efficacy of these techniques relies on the accuracy and reproducibility of the results of DIR. Incorporation of DIR-facilitated processes without an understanding of the impact of uncertainties may affect RT patient treatments.



The potential and risks of DIR in RT are well covered in current literature (Brock *et al* 2017, Paganelli *et al* 2018, Lowther *et al* 2022, Murr *et al* 2023). The American Association of Physicists in Medicine Task Group 132 (AAPM TG-132) report (Brock *et al* 2017) provided early guidance for work on qualification and commissioning of DIR algorithms and processes. AAPM TG-132 remains an excellent review of DIR and quality assurance (QA), but the report does suffer from some limitations. Latifi *et al* noted difficulties in applying the AAPM TG-132 recommendations in clinical practice (Latifi *et al* 2018). Hussein *et al* and Rigaud *et al* report barriers to DIR clinical implementation with a lack of suitable evaluation tools and consensus on their implementation (Rigaud *et al* 2019, Hussein *et al* 2021). Barber *et al* and Paganelli *et al* addressed the requirements of patient-specific DIR QA and commissioning, and discussed the difficulties of consensus DIR QA metrics (Paganelli *et al* 2018, Barber *et al* 2020). Recent position papers out of the Australasian College of Physical Scientists and Engineers in Medicine (ACPSEM) (Barber *et al* 2020) and the Medical Image Registration Special Interest Group (MIRSIG) (Lowther *et al* 2022) have proposed consensus evaluation strategies for local geometric accuracy and vector grid suitability.

Despite recommendations on geometric tolerances present in the literature, the reporting of uncertainty quantification in clinically implemented DIR is not well standardised for RT applications in today's literature, particularly with respect to dosimetric measures. This review aims to summarise the current understanding of uncertainties in DIR-facilitated processes and their clinical impact. The authors analysed the current literature about uncertainties in multiple DIR-facilitated applications, and summarised and extended recommendations with the general aim of raising awareness.


This review is structured as follows: We first summarise DIR algorithms used in RT (Chapter 2), and give a short explanation about the sources of DIR uncertainties (Chapter 3). Next, we review methods to quantify DIR uncertainties geometrically and dosimetrically (Chapter 4), and describe the effects and severity of these uncertainties for different RT applications (Chapter 5). Finally, we discuss uncertainty tolerances (Chapter 6) and summarise and expand current recommendations and recommend future research avenues (Chapter 7).

2. DIR algorithms

DIR is applied between two images, aiming at aligning corresponding anatomic regions in both images. The result of a DIR is a transformation, which is often represented as a displacement vector field (DVF), which can be applied to images, structures, or dose distributions (figure 1). The earliest DIR algorithms were based on optical flow (Horn and Schunck 1981) or thin plate splines (Bookstein 1989). Classical algorithms, such as intensity-based matching or biomechanical models remain popular, but recently research in deep-learning (DL) methods is increasing. For a comprehensive overview of DIR algorithms, we refer the reader to review articles (Maintz and Viergever 1998, Holden 2008, Haskins *et al* 2020, Chen *et al* 2021, Teuwen *et al* 2022, Zou *et al* 2022).

2.1. Classical image registration

Classical methods, in their simplest form, follow a process illustrated in figure 2. There are two input images, a moving image and a fixed image, where the goal is to deform the moving image into the coordinate system of the fixed image. The algorithm proceeds by iteratively optimising transformation parameters to find a registration



that minimises a similarity metric. The transformation parameters represent a displacement field, a velocity field, spline parameters, or other deformable transform representations. The similarity metric typically includes a regularisation term, which limits permissible transformations to those considered desirable or physically plausible, in addition to a similarity metric that matches intensity, such as mutual information or correlation coefficient.

Intensity-based DIR matching criteria are developed to use image intensity to optimise metrics such as mutual information (MI), sum of the squared difference (SSD) of image intensity, or cross-correlation (CC) (Oh and Kim 2017, Li *et al* 2021). Intensity-based DIR can achieve high accuracy for image areas with clear image features and high contrast. In poor contrast regions, intensity-based DIR accuracy may be less robust (Elmahdy *et al* 2019, Li *et al* 2021, Tascón-Vidarte *et al* 2022). To improve DIR accuracy, hybrid DIR algorithms consider point landmarks or structures defined on both image sets to improve registration results (Zhong *et al* 2012, Weistrand and Svensson 2015, Qin *et al* 2018, Motegi *et al* 2019, Shah *et al* 2021). Some algorithms rely on distance criteria to determine correspondence and transformations (Xiong *et al* 2006, Vásquez Osorio *et al* 2009, Zakariaee *et al* 2016) others use biomechanical properties.

Biomechanical algorithms are influenced by modelled physical properties of the tissues (Sotiras *et al* 2013, Polan *et al* 2017, Velec *et al* 2015, 2017). Finite element methods (FEM) model the properties of the tissues under mechanical force. Although the use of FEM requires the challenging definition of material properties, geometry, and boundary conditions, its robustness and plausibility are well demonstrated (Sotiras *et al* 2013). Compared to intensity-based DIR, it can improve multi-modal registration and registration in low-contrast regions (Velec *et al* 2015).

2.2. Deep learning-based DIR

In the past decade, machine learning algorithms in radiotherapy have increased dramatically, and DL has likewise made advances in the field of medical DIR (Teuwen *et al* 2022, Zou *et al* 2022). Topical reviews of the literature present extensive summaries of the current state of DL algorithms within DIR (Boveiri *et al* 2020, Xiao *et al* 2021, Zou *et al* 2022). DL in image registration is implemented through two approaches: deep similarity metrics in classical image registration algorithms, and deep neural networks that directly estimate the DVF.

2.2.1. Deep similarity metrics (DSMs)

As described in section 2.1, classical algorithms approach the problem of image alignment through a process of iterative optimization. These algorithms search for a global minimum of the solution space, but the choice of similarity metric remains problematic. DSMs aim to improve classical iterative image registration by improving the similarity term. This approach is particularly useful in multi-modal imaging where it has been shown to outperform mutual information (Wu *et al* 2013, Simonovsky *et al* 2016). Improvements in difficult monomodal registration problems, low contrast regions and large transformations, have been reported in the literature (Zhao and Jia 2015).

2.2.2. Direct determination of DVFs by machine learning algorithms

Direct DVF DL algorithms use historic DVFs or artificial DVFs as training data to determine registrations. The optimization phase happens in the training phase, where model parameters are determined. The vast majority of DL models aim for a direct regression of DVF transforms in a supervised approach. Variation between models is primarily a result of algorithm design and methodology.

Reviews (Boveiri *et al* 2020) cover a range of algorithm architectures. DL architectures include stacked auto-encoders (SAEs) (Wang *et al* 2017, Krebs *et al* 2018), bayesian frameworks (Deshpande and Bhatt 2019, Khawaled and Freiman 2020, 2022a), implicit neural representations (Wolterink *et al* 2022) and convolution

neural networks (CNNs) (Cao *et al* 2018, Ferrante *et al* 2018, Hu *et al* 2018, Balakrishnan *et al* 2018, 2019, Kim *et al* 2019, Kuang and Schmah 2019, Liu *et al* 2019, Jian *et al* 2022, Wolterink *et al* 2022, Xi *et al* 2022, Liang *et al* 2023). CNNs have been researched for direct DVF regression, with reported improvements in DVF when coupled with spatial transformer networks (Jaderberg *et al* 2015). CNN architecture use encoder-decoder networks, rather than a fully connected layer. Such approaches are currently implemented in well-cited solutions (VoxelMorph (Balakrishnan *et al* 2018, 2019) and U-NET (Liang *et al* 2023)). Despite the growth of multimodal foundational models in image creation, these reviews do not find application in image registration.

In general, DL training is divided between supervised and unsupervised learning methods (Chen *et al* 2021). For supervised registration methods, ground truth is either a DVF or a segmentation. The DVF may be created by a conventional DIR algorithm or from synthetic deformations, and the segmentations may be created by manual contouring or other methods. Unsupervised registration methods are further split into training by similarity metrics or generative adversarial networks (GANs) (Mahapatra *et al* 2018, Elmahdy *et al* 2019). If similarity metrics are used no ground truth is needed for the learning process but, as in traditional image registration, these models are limited by the same issues as similarity metrics in classical DIR optimization. If GAN is used, a discriminator judges if the warped moving image can be discriminated from the fixed image. When the warped image cannot be distinguished from the fixed image, the registration is deemed to be optimal (Goodfellow *et al* 2014). GANs show promise for multi-modality DIR problems as they do not require image similarity terms.

One advantage of DL algorithms is improvements in multi-modal registration, which is challenging for classical similarity metrics. Additionally, DL-based algorithms are more computationally efficient (Rohé *et al* 2017, Cao *et al* 2018, Balakrishnan *et al* 2018, 2019).

3. Source of uncertainties

The uncertainties of DIR can arise from a variety of sources. Many are image-based uncertainties, caused by anatomical changes, artifacts and different image modalities, as well as algorithm-based uncertainties, caused by intrinsic mathematical limitations and similarity metrics.

3.1. Image-based

3.1.1. Anatomical changes

Non-rigid variations in patient anatomy, such as weight gain or loss, neck flexion and tumour changes can be poorly mapped by rigid and affine registrations. DIR can improve the locally accurate alignment of anatomy (Hill *et al* 2001). While regularisation is useful to reduce the likelihood of physically unrealistic deformations, the magnitude of anatomical changes may exceed those allowed by an algorithm's settings. This can result in large DIR errors in areas near significant shape changes, particularly in low contrast image regions (Kashani *et al* 2008) or due to forced anatomical changes such as between external beam RT and brachytherapy (Vásquez Osorio *et al* 2015) (figure 3(a)).

Anatomical changes can be elastic, where the surrounding tissue follows the change and occupies the previous space (e.g. movement, position changes, or displacement) or inelastic, where the surrounding tissue stays in place (e.g. tissue growth, regression or emptying/filling cavities) (Amugongo *et al* 2022) figures 3(b) and (c). Modelling these changes is challenging (Sonke and Belderbos 2010, Mencarelli *et al* 2014, Sonke *et al* 2019). Certain implementations of regularisation can result in significant registration inaccuracies in sites in which naturally sliding boundaries occur, such as a rib bone and its adjacent lung (Sonke *et al* 2019). Some solutions were proposed to incorporate missing tissue during the DIR (Nithianathan *et al* 2012, Vishnevskiy *et al* 2017, Eiben *et al* 2018).

3.1.2. Artifacts/Image quality

The anatomical changes caused by natural patient motion, such as respiration, muscle contraction, and blood flow can lead to image artifacts (Nehmeh and Erdi 2008, Zhang *et al* 2012, Spin-Neto and Wenzel 2016, Giganti *et al* 2022). For example, motion artifacts during the image acquisition can result in implausible anatomy (Yamamoto *et al* 2008, Persson *et al* 2010) and implants such as prostheses in the imaging area can lead to streaking or voids (Ritter *et al* 2009, Fontenele *et al* 2018, Lee *et al* 2021). As these artifacts disrupt the true image intensity gradients of the patient tissue several papers have demonstrated decreased intensity-based DIR quality in their presence (Serban *et al* 2008, Sonke and Belderbos 2010, Fusella *et al* 2016) (figure 4).

Sensitivity of DIR algorithms to image noise, resolution (Constable and Henkelman 1991, Verdun *et al* 2015, Zhao *et al* 2016, Sarrut *et al* 2017), field of view (Barber *et al* 2020) and image contrast (Mencarelli *et al* 2014, Barber *et al* 2020, Dowling and O'Connor 2020) has been demonstrated in the literature. However, other studies




Figure 3. Examples of large anatomical changes. (a) Large changes during combined treatments with external beam radiotherapy (EBRT) and brachytherapy (BT) of the uterus. Image from (Vásquez Osorio *et al* 2015) with permission. (b) Large anatomical changes in the lung. (c) Weight loss for a head and neck patient during the course of treatment.




Figure 4. Examples of (a) dental artifacts (image from the United States National Cancer Institute (NCI) ‘The cancer imaging archive’ (TCIA) (Clark *et al* 2013, Ang *et al* 2014, Bosch *et al* 2015)), (b) 4 D artefacts in lung (image from TCIA (Roman *et al* 2012, Balik *et al* 2013, Clark *et al* 2013, Hugo *et al* 2017)) and (c) metal artifact in liver.

find that the effect of image noise has only minor effects on DIR results for computed tomography (CT) to CT registrations (Nesteruk *et al* 2022).

Research on the implementation of iterative image reconstruction algorithms has shown reduced noise and improved image quality for both CT and cone-beam CT (CBCT) (Held *et al* 2016, Giacometti *et al* 2019, Jarema and Aland 2019, Greffier *et al* 2020, Loi *et al* 2020), which may allow for improved quality intensity-based DIR.




Figure 5. Example of gradient effects in MRI that may increase DIR uncertainties. Figure from (Vásquez Osorio *et al* 2012), with permission. MRI: magnetic resonance imaging, DIR: Deformable image registration.

3.1.3. Multimodal registration

Multimodal DIR offers considerable clinical benefit in contour propagation (Söhn *et al* 2008, Vásquez Osorio *et al* 2012, Barber *et al* 2020, Zachiu *et al* 2020). However, multimodal DIR remains challenging, and similarity metrics must be selected with care.

For example, magnetic resonance imaging (MRI) to CT registration in the lung is difficult because of low contrast and resolution in MRI (Yang *et al* 2015) and in the prostate, lack of a clear boundary of the prostate gland in CT may lead to failures in MR-CT DIR (Zhong *et al* 2015). In the HN, limited soft tissue contrast and dental artifacts in CT images compared to MR influence the DIR uncertainty (Nix *et al* 2017, Kiser *et al* 2019). Additionally, gradient contrast artifacts in MRI may impair the DIR quality between different image modalities (figure 5) (Vásquez Osorio *et al* 2012). McKenzie *et al.* found monomodal registration from synthetic CT (generated from the MRI) to CT to be more accurate than the multimodal registration from the original MRI to CT for large deformations of HN patients (McKenzie *et al* 2020). Of course, the synthetic CT generation also faces uncertainties, for example the resulting Hounsfield units (HUs) differ between CT and synthetic CT. Boulanger *et al* report a mean absolute error of 76 HU in head and liver, and 42 HU in the pelvic area in average over multiple methods generating synthetic CTs (Boulanger *et al* 2021). Geometric differences between structures can also appear (Palmér *et al* 2021).

3.2. Algorithm-based

The choice of the DIR algorithm and parameter settings influences the DVF obtained when registering the same image pair. Several studies investigate the performance of different DIR algorithms, for example in HN (Hardcastle *et al* 2012, Močnik *et al* 2018, Qin *et al* 2018, Lee *et al* 2020, Kubli *et al* 2021), lung (Kadoya *et al* 2014, Scaggion *et al* 2020a), liver (Zhang *et al* 2012, Sen *et al* 2020) or pelvis (Hammers *et al* 2020). Some commercial DIR algorithms offer the possibility of parameter adjustments, such as registration metrics, guiding structures, regularisation levels, regularisation weights, or contrast level sensitivity, which causes uncertainty of the algorithm to vary (Ziegler *et al* 2019). However, most commercial algorithms are closed systems and not adjustable. Some studies find that even a single commercial DIR software can show variability in the results (Kadoya *et al* 2016, Miura *et al* 2017), depending on the specific workflows used. The performance of the same DIR algorithm might also vary between anatomical sites. For example, in a series of three separate studies comparing the Velocity and MIM algorithms (Kadoya *et al* 2016, Pukala *et al* 2016, Fukumitsu *et al* 2017) on different patient anatomies, the published results come to different conclusions regarding the performance.

4. Quantification of uncertainties

The quantification and evaluation of uncertainties in the applications of DIR are difficult due to multiple aspects. Firstly, a true ground truth is lacking and secondly, there are a wide range of DIR-facilitated applications which have differing requirements for accuracy. For dose monitoring, a low point-to-point error is necessary in steep dose gradients, while in low gradient or homogeneous dose regions, even larger point-to-point errors will not impact the mapped dose. For contour propagation, a high correspondence between organ boundaries is of importance (Rigaud *et al* 2019). Quantifying DIR uncertainties is crucial, as the DIR results are used for consecutive steps (Brock *et al* 2017, Paganelli *et al* 2018). So far, there is no standard procedure for uncertainty quantification of DIRs. Indeed, most commercial and research systems omit uncertainties entirely.

4.1. Using a digital or physical phantom as ground truth

The validation of DIR results is challenging due to the lack of ground truth. Therefore, evaluation strategies have been developed, questioned, and improved over the past decades. To create a ground-truth surrogate, digital phantoms and physical phantoms have been proposed. Digital phantoms (Castillo *et al* 2009, Vandemeulebroucke *et al* 2011, Brock *et al* 2017) are created using voxel-based reference deformations, which DIR algorithms are expected to recover. This can cause bias in results. For example, a digital phantom deformed with displacements generated by a B-spline might result in better agreement when testing algorithms that use B-spline transformations (Fatyga *et al* 2015, Loi *et al* 2018, Balakrishnan *et al* 2019, Boyd *et al* 2021). Digital phantoms allow for the comparison of contour-based evaluation methods *and* direct evaluation of DVF errors. In contrast, physical phantoms (Graves *et al* 2015, Niebuhr *et al* 2019, Kadoya *et al* 2021) provide geometrical verification through landmarks or structures. Therefore, physical phantoms suffer from a similar lack of ground truth as patient images. Intrinsic errors due to inter- and intra-observer variability due to the manual identification (Machiels *et al* 2019, Roach *et al* 2019) remain present in phantoms. The use of markers (Machiels *et al* 2019), guidelines (Lin *et al* 2020), auto segmentation (Rey *et al* 2002, Yang *et al* 2018, Cardenas *et al* 2019, Schipaanboord *et al* 2019, Vrtovec *et al* 2020, Harrison *et al* 2022) and automated landmark extraction (Paganelli *et al* 2018) can reduce observer uncertainties, but are not necessarily more accurate. Also, just as with patient images, these methods quantify DIR performance only near the points or structures under consideration (Shi *et al* 2021) and do not provide a holistic assessment of the DIR performance. The deformations of physical phantoms might not always be anatomically realistic. While both, digital and physical phantoms, are useful for commissioning and QA of applications involving DIR, it is important to keep their weaknesses in mind.

4.2. Geometric and dosimetric uncertainty quantification

With the lack of ground truth, alternative measures have to be used to quantify the effects of DIR uncertainty. Most commonly geometric measures are used, comparing warped points of interest or structures to reference points and structures. These reference-based geometric measures are however not always available and have their own uncertainties, such as intra- and inter-observer variability. Reference-free measures have also been proposed, they can be applied without reference data. A short summary of various geometric uncertainty measures is given in table 1. For a more detailed overview about geometric measures and which methods are proposed for specific applications please refer to the AAPM TG 132 (Brock *et al* 2017) and MIRSIG (Lowther *et al* 2022). In addition to geometric measures multiple methods to visualise and quantify dosimetric uncertainties have been proposed (table 1).

In this review, we refer to dose mapping as the process of warping/projecting/transferring a dose distribution, defined in one image, to a second image of the same patient. We refer to dose accumulation as the summation of the mapped dose distribution and a secondary dose distribution defined in the second image. Quantifying the correctness of dose mapping is challenging but essential in RT (Murr *et al* 2023). Some authors suggest using TG-132 thresholds (Xiao *et al* 2020), but the TG-132 report explicitly states '[t]he use of deformable registration for dose accumulation . . . is outside of the scope of this task group.' (Brock *et al* 2017). For this reason, we feel that the metrics and thresholds proposed by TG-132 are not sufficient to evaluate image registration for dose mapping/accumulation. Instead, dosimetry uncertainty measures for clinical practice are needed.

4.2.1. Correlation within measures

Geometric measures are not independent and self-correlate. Loi *et al* found a linear relationship between mean distance to agreement (MDA) and dice similarity coefficient (DSC) (Loi *et al* 2018). Also, a correlation between distance discordance metric (DDM) and Harmonic energy (HE) has been found (Kierkels *et al* 2018). Reporting multiple measures is still useful despite being redundant. For example, the DSC limitations can be critically analysed in conjunction with other metrics, such as MDA for different structures and volumes (Jena *et al* 2010, Brock *et al* 2017, Loi *et al* 2018). Combining different geometrical metrics can improve the understanding of the overall quality of the DIR for a specific application.

Different implementations and specific ways to use the same measure can lead to vastly different results. For example variations of up to 50% in DSC, 50% in Hausdorff distance (HD) and 200% in MDA were found between the same structure sets, evaluated by different institutions (Gooding *et al* 2022). Comparing results from different studies and centres should therefore be taken with care. The correlation between geometric and dosimetric measures was found to be low (Hvid *et al* 2016, Pukala *et al* 2016, Poel *et al* 2021, Nash *et al* 2022, Kamath *et al* 2023).

Table 1. Description, strengths and limitations of commonly used geometric and dosimetric uncertainty quantification metrics. DVH: Dose-volume-histogram.

Metric	Description	Strengths (+) / Limitations (-)
Reference-based	<p>Target registration error (TRE)</p> <ul style="list-style-type: none"> Distance between anatomical landmarks defined by different methods, e.g. warped with DIR versus physician-drawn reference (Fitzpatrick <i>et al</i> 1998, Datteri and Dawant 2012, Brock <i>et al</i> 2017) $TRE = T(p_f) - p_m$ $T(p_f)$: estimated transformation of point from fixed image, p_m position of point on moving image 	<ul style="list-style-type: none"> + Distance, in mm + Spatially resolved - Reference points required (manual or automatic definition), additional inherent uncertainties, and time-consuming definition - Validity depends point quantity and relevance - Limited to areas with sufficient image contrast <ul style="list-style-type: none"> - Requires reference/ground truth + Widely used, useful to compare to literature
	<p>Dice similarity coefficient (DSC)</p> <ul style="list-style-type: none"> Measure of the overlap between two contours (Dice 1945, Brock <i>et al</i> 2017) $DSC = \frac{2 X \cap Y }{ X + Y }$: volume covered by both structures, $X + Y$: volume covered by at least one of the structures 	<ul style="list-style-type: none"> - Strongly volume dependent, lacks sensitivity for large structures - Special care needed for tubular structures - Hard to interpret/no meaningful unit - Requires reference/ground truth
	<p>Hausdorff distance (HD)</p> <ul style="list-style-type: none"> Maximum distance of the closest approach of each point on one contour to all points of the other contour (Hausdorff 1920, Huttenlocher <i>et al</i> 1993) $HD(X, Y) = \max(d(X, Y), d(Y, X))$ with $d(X, Y) = \max_{x \in X} \min_{y \in Y} \ x - y\$ $d(X, Y)$ distance between two pointsets 	<ul style="list-style-type: none"> + Distance, in mm - Sensitive to outliers
	<p>Mean distance to agreement (MDA)</p> <ul style="list-style-type: none"> Mean distance of the closest approach of each point on one contour to all points of the other contour (Vrtovec <i>et al</i> 2020) $MDA(X, Y) = \text{mean}(d(X, Y), d(Y, X))$ 	<ul style="list-style-type: none"> - Requires reference/ground truth + Distance, in mm + Less sensitive to outliers than HD - Misses local uncertainties - Requires reference/ground truth
	<p>Centre of mass displacement (COM)</p> <ul style="list-style-type: none"> Shift in center of mass between two structures (Choi <i>et al</i> 2011, Takayama <i>et al</i> 2017) $COM = \sqrt{\Delta x^2 + \Delta y^2 + \Delta z^2}$ with $\Delta x^2 = \vec{R}_{1,x} - \vec{R}_{2,x}$, $\Delta y^2 = \vec{R}_{1,y} - \vec{R}_{2,y}$, $\Delta z^2 = \vec{R}_{1,z} - \vec{R}_{2,z}$ and $\vec{R} = \frac{1}{M} \int \int \int \rho(\vec{r}) \vec{r} dV$ M: mass of the structure, $\rho(\vec{r})$ density distribution structure 	<ul style="list-style-type: none"> + Distance, in mm - Lacks sensitivity to variations in contour boundary - Requires reference/ground truth + Useful in contrast-poor areas

Table 1. (Continued.)

	Metric	Description	Strengths (+) / Limitations (-)
Reference- free measures	Distance discordance metric (DDM)	<ul style="list-style-type: none"> • Mean distance of points from moving images which are registered to the same point in the a fixed reference image (Saleh <i>et al</i> 2014) • For mathematical description please refer to the original publication (Saleh <i>et al</i> 2014) 	<ul style="list-style-type: none"> + Spatially resolved, reference-free - Needs at least four registered images + Spatially resolved, reference-free
	Local uncertainty metric (LU)	<ul style="list-style-type: none"> • Uncertainties within a uniformly-dense structures can be calculated based on points defined on the organ edges (Takemura <i>et al</i> 2018) • For mathematical description please refer to the original publication (Takemura <i>et al</i> 2018) 	<ul style="list-style-type: none"> + Works in uniformly-dense regions - Requires contours + Local volume gain/loss detection
	Jacobian determinant	<ul style="list-style-type: none"> • The first derivative of the DVF, distinguish between regions which are locally expanding in volume $J > 1$ and those shrinking with volume $J < 1$ (Chung <i>et al</i> 2001) • $J = \det\left(\frac{dT}{dx}\right) = \det\left(\begin{matrix} \frac{dT_x}{dx} & \frac{dT_x}{dy} & \frac{dT_x}{dz} & \frac{dT_y}{dx} & \frac{dT_y}{dy} & \frac{dT_y}{dz} & \frac{dT_z}{dx} & \frac{dT_z}{dy} & \frac{dT_z}{dz} \end{matrix}\right)$ with T the transformation 	<ul style="list-style-type: none"> + Spatially resolved, reference-free - Misleading for actual mass change - Necessary but not sufficient + Measure for smoothness
	Harmonic energy (HE)	<ul style="list-style-type: none"> • A measure of the nonlinearity of the transformation, inversely proportional to the smoothness of the deformation (Forsberg <i>et al</i> 2012, Varadhan <i>et al</i> 2013) • $HE = \ Jac\ _F = \sqrt{\sum_{i=1}^3 \sum_{j=1}^3 t_{ij} ^2}$ beeing the Frobenius norm of the Jacobian 	<ul style="list-style-type: none"> + Spatially resolved, reference-free - Hard to interpret - Fails with sliding surfaces + Related to algorithm repeatability
	Inverse consistency error (ICE) / Transitivity error (TE)	<ul style="list-style-type: none"> • Applying a registration from image A to image B and then back to image A, it is assumed that all points will be mapped on their original position. ICE is defined as the difference between the original point and the transformed point mapped back to the fixed image grid (Bender and Tomé 2009), TE extends this idea to more than two images (Bender <i>et al</i> 2012) 	<ul style="list-style-type: none"> + Spatially resolved, reference-free - No indication of accuracy in the result - Necessary but not sufficient
	Dosimetric measures	Dose parameter variations and DVH bands	<ul style="list-style-type: none"> • Report of relevant dosimetric point variations (e.g. V95%, D2%, V10Gy, mean dose) and DVH bands caused by uncertainties in propagated structures or dose mapping/accumulation (Nassef <i>et al</i> 2016, Lowther <i>et al</i> 2020a, 2020b, García-Alvarez <i>et al</i> 2022) • A known or estimated DIR uncertainty is necessary, either simulated (Wang <i>et al</i> 2018, Smolders <i>et al</i> 2022b), DIR variations (Nenoff <i>et al</i> 2020, Amstutz <i>et al</i> 2021b) or with known reference deformations (Kirby <i>et al</i> 2016, Covele <i>et al</i> 2021)

Table 1. (Continued.)

Metric	Description	Strengths (+) / Limitations (-)
Local uncertainty maps	<ul style="list-style-type: none"> • Highlights regions with anticipated discrepancies due to voxel-wise uncertainties • Voxel-wise uncertainties can be based on geometric factors (Salguero <i>et al</i> 2011), principal component analysis (Murphy <i>et al</i> 2012) or stochastic methods (Hub <i>et al</i> 2012) 	+ Spatially resolved dosimetric uncertainty information + No reference required
Energy-conservation-based criterion	<ul style="list-style-type: none"> • Structure-wise comparison of delivered energy with the energy of the warped representation of the dose (Zhong and Chetty 2017, Wu <i>et al</i> 2023) 	<ul style="list-style-type: none"> – Previous DIR uncertainty measure required + Reliability measure for regions with mass/volume change – References required – Only structure-wise information

4.3. AI/DL-based uncertainty quantification

Further to its implementation as a DVF generator for the registration process, DL can also be used for the quantification or prediction of uncertainties in DIR (Smolders *et al* 2022b, 2022a, 2023a). DSM that are not used in the optimization of output DVFs, provide further uncertainty quantification metrics that can be used to determine the quality of the overall registration and highlight regions of poor accuracy (Galib *et al* 2020). The implementation of algorithms for automated image segmentation allows for the potential use of reference-based DIR evaluations (table 1) with limited or no user interaction. In this case the segmented structures must be consistent between the datasets used in the image registration. Additionally, DL-based DIR showed the potential of having inherent uncertainty assessments within the DL framework (Grigorescu *et al* 2021, Gong *et al* 2022, Khawaled and Freiman 2022b).

4.4. Treatment margins

Uncertainties in any DIR-facilitated process that is used to generate contours (e.g., image registration for standard treatment planning or atlas-based segmentation) should be quantified and included in the treatment margins. To achieve this, population-based studies would be required where the calculated uncertainties can be used in the margin formula (van Herk *et al* 2000). However, guidelines detailing the quantification and inclusion of these uncertainties are missing.

5. Application-specific DIR uncertainty

In this chapter, studies investigating the effect of DIR uncertainties for the deformation of images, structures and doses used in RT are reviewed (figure 1).

5.1. Deformed images

5.1.1. Applications at planning

The TG-132 report and other recommendations suggest imaging the patient in the treatment position whenever possible to minimise the magnitude of the required deformation during registration (Brock *et al* 2017, Barber *et al* 2020).

5.1.2. Intrafraction applications

DIR has been used to derive motion-corrected images from 4D CT scans (Wolthaus *et al* 2008) with average landmark-position differences of 0.5 mm for all directions in the tumour region. DIR is also used to reconstruct time-resolved 4D MRI (Nie *et al* 2020), with reported centre of mass differences of 2.9 ± 0.6 mm. We expect the geometrical uncertainties of propagated images to be similar to those of structure propagation, considering both utilise the same input data.

5.1.3. Interfraction applications

With MRI linac or CBCT-based online adaptation becoming more commonly available, the interest in deforming images between fractions for dose calculation and optimization is increasing (Kraus *et al* 2017, Tenhunen *et al* 2018, Irmak *et al* 2020, Byrne *et al* 2021). In these workflows, the calculated dose distribution is unlikely to be accurate considering the spatial uncertainties in the deformed CT, especially in areas with large density changes.

To correct for density changes that are not represented by the deformed image such as moving air in the gastro-intestinal organs, the density in these areas is often overwritten with the density of air or water (van Timmeren *et al* 2020). Research investigating the impacts of these overwrites on photon RT has found these impacts to be not clinically relevant (Pham *et al* 2022), except for very large air cavities (Thapa *et al* 2019). For protons, these density corrections are likely more relevant.

To avoid the use of DIR and manual density overwrites, direct dose calculation on the MRI or CBCT images has been investigated. The generation of synthetic CT images from MRI is reviewed elsewhere (Owraangi *et al* 2018, Hoffmann *et al* 2020, Boulanger *et al* 2021). Methods of scatter correction to make CBCT usable for dose calculation are widely explored (Kurz *et al* 2016, Giacometti *et al* 2019, Jarema and Aland 2019, Lalonde *et al* 2020, Trapp *et al* 2022).

5.1.4. Intervention follow-up

Follow-up images after intervention can be registered to a planning CT to understand the relation and location of local failure such as recurrence or necrosis with a planned dose distribution and planning structures (Chang *et al* 2018, Kamal *et al* 2020, Abdel-Aty *et al* 2022). In these cases, dramatic changes are observed caused by the time between images, surgical intervention, or other medical issues. Systematic studies quantifying the impact of

DIR uncertainties for intervention follow-up are rare and further work needs to be done to quantify and account for them in patterns of failure.

5.2. Propagated structures

5.2.1. At planning

For treatment planning, structures are commonly defined on the planning CT. Structure definition can be challenging on CT due to low contrast compared to other imaging modalities such as MRI. Including multiple imaging modalities for contouring can lead to a reduction in inter-observer variability (Caldwell *et al* 2001, Farina *et al* 2017, Hall *et al* 2018). Though it is common to merge CT with PET, MRI or other images for contouring, the effect of deformable registration errors is not well investigated. (Barber *et al* 2020) therefore suggested using rigid registration wherever possible.

DIR is also used in atlas-based auto-segmentation, which is increasingly used in clinics to assist contouring. In this case, DIR is applied between images from different patients (Vrtovec *et al* 2020). Research showing a time benefit in using atlas-based contours also show the necessity of manual corrections (Gooding *et al* 2013, Cardenas *et al* 2019, Welgemoed *et al* 2023). To our knowledge, there are no systematic studies on the impact of DIR implementation and DIR uncertainty for atlas-based segmentation. Studies do however investigate the impact of atlas selection (Schipaanboord *et al* 2019) or institution-specific implementation (Gooding *et al* 2013). As an alternative to atlas-based auto-segmentation, DL-based auto-segmentation was developed. Different auto-segmentation methods are reviewed elsewhere (Cardenas *et al* 2019, Schipaanboord *et al* 2019, Vrtovec *et al* 2020, Harrison *et al* 2022). The details of auto segmentation methods are out of the scope of this study.

5.2.2. Intrafraction applications

DIR is used to map contours between different breathing phases, or intrafraction changes in patient treatment positions, to reduce the time needed for contouring. In clinical practice, the propagated structures are visually checked and corrected if necessary (Gaede *et al* 2011, Peroni *et al* 2013, Liu *et al* 2016, Ma *et al* 2017, Willigenburg *et al* 2022).

5.2.3. Interfraction applications

Structure propagation can speed up recontouring for repeated imaging of a patient (Sonke *et al* 2019). This can be used for evaluation of recalculated doses or adaptive planning, but it is especially important for online adaptive workflows. The clinical availability of regular or daily imaging, such as scheduled repeated CT or daily CBCT, and the implementation of online adaptive workflows has led to multiple studies on the quality of deformed structures for adaptive planning.

Table 2 summarises recent studies investigating geometric DIR uncertainties for structure propagation for different anatomical sites and imaging modalities. The structures have been evaluated using geometrical measures introduced in table 1. Commercial and open access algorithms show similar performance (Scaggion *et al* 2020b). The majority of studies incorporate reference structures of a single expert physician. Variations between structures defined by different physicians are observed in many studies, and these inter-expert structure variations are currently de-facto the clinically accepted variability. Research comparing DIR propagated structure uncertainties to physician-to-physician uncertainties, has demonstrated results approaching inter-expert contour variation (Riegel *et al* 2016, Woerner *et al* 2017, Rigaud *et al* 2019, Nash *et al* 2022).

Currently, there is no consensus on the use of DIR propagated structures for plan adaptation in the literature. Some authors conclude that propagated structures can be used for reoptimization and/or dose evaluation (Beasley *et al* 2016, Hvid *et al* 2016, Qiao *et al* 2019, Nenoff *et al* 2021b, Nash *et al* 2022), while others found that manual corrections are still necessary (Li *et al* 2017, Christiansen *et al* 2020, 2021). Generally, the literature agrees that a visual inspection of the DIR propagated structures remains necessary for dose evaluation and optimization. Furthermore, it has been observed that for most organs at risk (OARs) geometric uncertainties correlated only weakly to dosimetric errors (Hvid *et al* 2016, Pukala *et al* 2016, Nash *et al* 2022).

There are a small number of studies evaluating the dosimetric effect of uncertainties in propagated structures for dosimetric evaluation or plan optimization during adaptive RT (table 3). For pancreas stereotactic body radiotherapy (SBRT), physician-drawn structures were compared to propagated structures by MIM and Precision DIR algorithms (Magallon-Baro *et al* 2022). They compared uncorrected propagated structures with physician-drawn structures in 0.5, 1 and 3 cm distance rings from the target. They found that replanning with uncorrected propagated structures improves the target coverage and OAR sparing compared to no adaptation. For the majority of fractions, manual correction of propagated structures could be avoided or be limited to the region closest to the target. Ray *et al* evaluated the use of automatic deformed CTVs compared to physician defined CTVs and proposed a framework to determine PTV margins based on automatic deformed CTVs for adaptive planning (Ray *et al* 2020). Nash *et al* showed that even large geometrical structure differences rarely had

Table 2. Literature review of quantified geometric uncertainties in DIR-facilitated processes for different anatomical regions and imaging modalities.

Indication	Image modality	DIR algorithm and/or vendor	Assessment method	Study type	DICE	HD	others (TRE, MDA, COM, ...)	Reference
Brain	MRI to MRI	Demons, HAMMER, and state-of-the-art registration methods with integrated learned features from unsupervised deep learning. ICA: Independent Component Analysis	Compare to segmented structures in the datasets	IXI dataset and ADNI dataset	XI dataset: Demons: 0.752 M+PCA: 0.790 M+ISA: 0.789 HAMMER: 0.789 H+PCA: 0.754 H+ISA: 0.801 ADNI dataset: Demons: 0.869 M+PCA: 0.789 M+ISA: 0.844 HAMMER: 0.821 H+PCA: 0.820 H+ISA: 0.873			(Wu <i>et al</i> 2013)
Brain	MRI to MRI	ANTs, VoxelMorph-1 (DL-based), VoxelMorph-2 (DL-based)	Compare to segmentations performed by FreeSurfer checked by visual inspection	7829 T1 weighted brain MRI scans from eight publicly available datasets	Average DICE: Affine only: 0.567 ANTs: 0.749 VoxelMorph-1: 0.724 VoxelMorph-2: 0.750			(Balakrishnan <i>et al</i> 2018)
Brain	MRI to MRI	Cue-Aware Deep Regression Network (DL-based)	Compare to segmented structures in the dataset	Three databases, i.e. LONI LPBA40, IXI, and ADNI	Average overall DICE: 0.7526		Average surface distance (ASD) in mm: Overall \approx 0.6–0.7 (25th–75th percentile)	(Cao <i>et al</i> 2018)
Brain	MRI to MRI (2D)	Unsupervised DL-based (Bayesian Framework)	Compare to 4 largest anatomical structures in the reference dataset	MGH10 dataset, 10 subjects, 10 slices each	Overall DICE			(Khawaled and Freiman 2020)
Brain	Inter-patient MRI	TransMorph: Transformer for unsupervised image	Inter-patient MRI: compare to 30 anatomical structures labeled by FreeSurfer	Inter-patient MRI: 260 T1-weighted brain MRI images from John Hopkins University	VoxelMorph: 0.7109 Proposed: 0.736 Please consult (Chen <i>et al</i> 2022) for extensive comparison of DICE values	Please consult (Chen <i>et al</i> 2022) for extensive comparison of HD values	Please consult (Chen <i>et al</i> 2022) for extensive comparison of SDlogJ and SSIM values	(Chen <i>et al</i> 2022)

Table 2. (Continued.)

Indication	Image modality	DIR algorithm and/or vendor	Assessment method	Study type	DICE	HD	others (TRE, MDA, COM, ...)	Reference
Head and neck	Atlas-to-patient MRXCAT-to-CT	registration	Atlas-to-patient MRI: compare to 30 anatomical structures labeled by FreeSurfer	Atlas-to-patient MRI: 576 T1-weighted brain MRI images from the IXI database XCAT-to-CT: XCAT phantom and 50 non-contrast chest-abdomen-pelvis CT scans				
Head and neck	CT to CT	MIM, Velocity, Raystation, Pinnacle, Eclipse	Compared to physician drawn reference	10 virtual head and neck phantoms (DIREP)			Mean TRE: 0.5 mm – 3 mm	(Pukala <i>et al</i> 2016)
Head and neck	CT to CT	MIM, Velocity, Eclipse	Compared to physician drawn reference	35 institutions, 10 virtual head and neck phantoms (DIREP)			Maximum TRE: 22 mm Mean TRE:	(Kubli <i>et al</i> 2021)
Head and neck	CT to CT	Raystation (simple Anaconda, detailed Anaconda, simple Morfeus, detailed Morfeus)	Compared to physician drawn reference	10 head and neck cancer patients	GTV DSC: Simple Anaconda 0.78 ± 0.11 ; Detailed Anaconda 0.96 ± 0.02 ; Simple Morfeus 0.64 ± 0.15 ; Detailed Morfeus 0.91 ± 0.03 ; Larger DSC for OARs larger than the eye compared to smaller OARs		Velocity 2.04 ± 0.35 mm; MIM 1.10 ± 0.29 mm; Eclipse 2.35 ± 0.15 mm All mean TRE < 3 mm Maximum errors > 2 cm	(Zhang <i>et al</i> 2018)
Head and neck	CT to CT	10 DIR combinations using demons and free form deformations (FFD)	Compared against each other and 2 experts using landmarks	15 patients, 6 weekly CTs			Landmark Registration Error: inter-observer distance 2.01 mm (1.29 mm), most effective DIRs 2.44 mm (and 1.30 mm)	(Rigaud <i>et al</i> 2019)
Head and neck	CT to CBCT	NiftyReg	Compared to physician drawn reference	5 head and neck patients	Mean DSC: 0.850 External contour: 0.986 Bony anatomy: 0.846 Soft tissue: 0.790 (DIR better than rigid registration)			(Veiga <i>et al</i> 2014)

Table 2. (Continued.)

Indication	Image modality	DIR algorithm and/or vendor	Assessment method	Study type	DICE	HD	others (TRE, MDA, COM, ...)	Reference
Head and neck	CT to CBCT	Five commercially available DIRs (RayStation, ADMIRE, Mirada, ProSoma, Pinnacle)	Compared to physician drawn and STAPLE reference	10 head and neck patients: 5 oropharyngeal, 2 oral cavity, 1 hypopharynx, 1 supraglottic and 1 of unknown primary (target below nasal region)	clinician drawn reference: Brainstem 0.68(0.09), Spinal Cord, 0.62(0.14), Larynx 0.75(0.1), Left Parotid 0.72(0.08), Right Parotid 0.76(0.06) STAPLE reference: Brainstem 0.93(0.04), Spinal Cord 0.87(0.04), Larynx 0.93(0.04), Left Parotid 0.93(0.06), Right Parotid 0.92(0.03)	clinician drawn reference: Brainstem 10.8(3.5), Spinal Cord, 7.1(2.8), Larynx 10.2(4.5), Left Parotid 12.9(4.8), Right Parotid 12.2(3.9) STAPLE reference: Brainstem 4.4(2.7), Spinal Cord 4.3(2.7), Larynx 3.5(1.1), Left Parotid 3.5(1.1), Right Parotid 3.4(1.1)	MDA: clinician drawn reference: Brainstem 2.9(0.1), Spinal Cord, 1.5(0.5), Larynx 2.2(1.1), Left Parotid 2.2(0.5), Right Parotid 2.0(0.5) STAPLE reference: Brainstem 0.8(0.5), Spinal Cord, 0.5(0.2), Larynx 0.5(0.3), Left Parotid 0.5(0.2), Right Parotid 0.5(0.2) Centroid separation in mm: clinician drawn reference: Brainstem 5.7(2.9), Spinal Cord, 9.7(5.8), Larynx 3.2(2.7), Left Parotid 3.6(1.6), Right Parotid 3.1(1.4) STAPLE reference: Brainstem 1.6(1.3), Spinal Cord, 2.8(2.1), Larynx 0.9(1.0), Left Parotid 0.9(0.6), Right Parotid 0.9(0.6)	(Nash et al 2022)
Head and neck	CT to CBCT	MIM DIR	Compared to physician structures	30 HN patients, squamous cell carcinoma of the oral cavity, pharynx or larynx, DIR to first and last CBCT	First CBCT Parotid L 0.95 Parotid R 0.95 Submandibular L 0.91 Submandibular R 0.93 Esophagus 0.85 Spinal cord 0.89 Last CBCT Parotid L 0.95 Parotid R 0.95 Submandibular L 0.85 Submandibular R 0.87 Esophagus 0.84 Spinal cord 0.87	First CBCT Parotid L 0.7 cm Parotid R 0.7 cm Submandibular L 0.6 cm Submandibular R 0.6 cm Esophagus 0.5 cm Spinal cord 0.3 cm Last CBCT Parotid L 0.7 cm Parotid 0.7 cm Submandibular 0.7 cm Esophagus 0.8 cm Spinal cord 0.3 cm		(Hvid et al 2016)
Head and neck	CT to CBCT	10 DIRs (optical flow, Demons, Level set, Spline)	Compared to physician reference	21 HN patients	data not shown in tables, please refer to the plots in the paper.	data not shown in tables, please refer to the plots in the paper.		(Li et al 2017)
Head and neck	MRI to MRI	Monaco DIR	Compared to manual defined structures and intra observer variability	17 patients, larynx (3), oropharynx (10), oral cavity (1) and hypopharynx (3), planning MRO + 3 repeated MRI	MRI to MRIGTV-T 0.55 GTV-N 0.58 Brain Stem 0.89 Spinal cord 0.86 R parotid 0.81 L parotid 0.82 R submand 0.77 L submand 0.78 Thyroid 0.74 IOVGTV-T 0.68 GTV-N 0.72 Brain Stem 0.96 Spinal cord 0.89 R parotid 0.93 L parotid 0.88 R submand 0.89 L submand 0.88 Thyroid 0.81	IOVGTV-T 9.8 mm GTV-N 5.0 mm Brain Stem 3.0 mm Spinal cord 2.8 mm R parotid 3.7 mm L parotid 4.4 mm R submand 3.1 mm L submand 3.3 mm Thyroid 4.3 mm MRI to MRIGTV-T 7.6 mm GTV-N 5.7 mm Brain Stem 4.3 mm Spinal cord 5.0 mm R parotid 7.7 mm L parotid 7.1 mm R submand 5.0 mm L submand 4.6 mm Thyroid 7.2 mm	mean surface distance, IOVGTV-T 2.2 mm GTV-N 1.1 mm Brain Stem 0.2 mm Spinal cord 0.5 mm R parotid 0.4 mm L parotid 0.8 mm R submand 0.5 mm L submand 0.6 mm Thyroid 0.8 mm MRI to MRIGTV-T 2.0 mm GTV-N 1.6 mm Brain Stem 1.0 mm Spinal cord 0.6 mm R parotid 1.2 mm L parotid 1.1 mm R submand 1.1 mm L submand 0.9 mm Thyroid 1.4 mm	(Christiansen et al 2021)

Table 2. (Continued.)

Indication	Image modality	DIR algorithm and/or vendor	Assessment method	Study type	DICE	HD	others (TRE, MDA, COM, ...)	Reference
Head and neck, thorax, pelvis	CT to CT	Velocity	Compared to two observers	30 head and neck and 20 prostate cancer patients		mean HD, structure dependence H- N Intraobserver variation 0.7mm-2.3 mm, interobserver variation 1.0mm-5.0 mm, DIR error 1.1mm-3.0 mm Pelvis Intraobserver variation 1.3mm-2.5 mm, interobserver variation 1.6mm-3.1 mm, DIR error 1.9mm-3.1 mm		(Riegel et al 2016)
Head and neck, thorax, pelvis	CT to CT	RayStation, MIM, Velocity AI and Smart Adapt, Mirada XD, ABAS	Compared to reference contours generated with a ground truth DVF	synthetic CT images (simQA), thirteen institutions	HN 0.84–0.93 Thorax 0.52–0.97 Pelvis 0.45–0.87		Mean Distance to Conformity (MDC) in mm HN 2.26–3.36 Thorax 2.38–4.57 Pelvis 3.69–6.03	(Loi et al 2018)
Head and neck, pelvis	CT to CT	MIM-Maestro, Raystation, Velocity	Compared to reference contours generated with a ground truth DVF	9 pairs of synthetic CTs (simQA)	trachea, esophagus, spinal cord, and spinal canal 0.95–0.98 pituitary 0.34–0.92		MDA (mm):trachea, esophagus, spinal cord, and spinal canal 2.10–2.70 pituitary 3.02–3.81	(Shi et al 2021)
Head and neck, Prostate, Pancreas	CBCT to CT	Physician-to-physician, Velocity	Compared to physician drawn reference	HN 6 patients, prostate 5 patients, pancreas 5 patients	HN Mean DSC:Physician-to-physician 0.87 DIR 0.77 Prostate Mean DSC:Physician-to-physician 0.9 DIR 0.74 Pancreas: Mean DSC: Physician-to-physician 0.93 DIR 0.84	All: Mean HD: Physician-to-physician 11.32 mm Rigid 12.1 mm DIR 12.0 mm		(Woerner et al 2017)
virtual phantoms and brain, HN, cervix, prostate	CT to CT	Smart Adapt (Eclipse)	Compared to physician structures	10 virtual phantoms, and brain (n = 5), HN (n = 9), cervix (n = 18) and prostate (n = 23) patients	Brain 0.91 (0.04) HN 0.84 (0.03) Prostate 0.81 (0.05) Cervix 0.77 (0.05) per-structure DSCs in paper	Brain 1.37 (0.97) HN 1.06 (0.22) Prostate 2.70 (0.24) Cervix 3.23 (0.78) per-structure HD in paper	Center of mass, Brain 1.69 (0.84) HN 1.63 (0.30) Prostate 5.19 (1.34) Cervix 5.79 (1.42) per-structure COM in paper	(Jamema et al 2018)
Abdominal, Head and neck, Thoracic	4DCT	Mirada	Compared to physician drawn reference	3 abdominal patients, 7 thoracic patients, two images from extreme respiratory phases	Abdominal: Nearly all OARs DSC > 0.90, pancreas 0.74–0.88 HN: Lower DSC, lowest for pharyngeal constrictor low contrast in this region, small size of structure and proximity to air cavities, Thorax: Nearly all OARs DSC > 0.90, esophagus 0.79–0.85		Thoracic: Mean TRE: 3.4–8.9 mm (above AAPM report recommendation) Maximum TRE: 10.1–29.0 mm	(Latifi et al 2018)
Retina and Heart	Retina: Colour fundus images to	GAN (DL-based)	Retina: Compare to registration ground-	Retina: 26 image pairs Heart:	Average DICE: Retina GAN: 0.946 DIRNet: 0.911 Elastix:	HD95 (95th percentile HD): Retina GAN: 4.2 DIRNet: 5.9 Elastix:	Mean absolute surface distance (MAD): Retina GAN: 3.1 DIRNet:	(Mahapatra et al 2018)

Table 2. (Continued.)

Indication	Image modality	DIR algorithm and/or vendor	Assessment method	Study type	DICE	HD	others (TRE, MDA, COM, ...)	Reference
	fluorescein angiography Heart: MRI to MRI		truth derived with ITKHeart: Compare to manual segmented structures	Sunybrook cardiac dataset, 45 cardiac cine MRI scans (short-axis cardiac image slices each containing 20 timepoints)	0.874Before registration: 0.843Heart:GAN: 0.85DIRNet: 0.80Elastix: 0.77Before registration: 0.62	9.7Before registration: 11.4Heart: GAN: 3.9DIRNet: 5.03Elastix: 5.21Before registration: 7.79	5.0Elastix: 8.7Before registration: 9.1Heart:GAN: 1.3DIRNet: 1.83Elastix: 2.12Before registration: 2.12	
abdomen, thorax, pelvis	4DCT, MR-MR, CT-MR	Morpheus (Raystation)	Compared to manually defined contours and landmarks	74 patients, thoracic and abdominal 4DCT and MR,, liver CT-MR, prostate MR-Mr			mean DTA <1 mm for controlling structures and 1.0–3.5 mm for implicitly deformed structuresTRE: 2.0 mm – 5.1 mm	(Velec et al 2017)
Thorax/Esophagus	4DCT	Bspline (Velocity), free form deform (FDD), Horn-Schunk optical flow (OF), Demons	Compared to manual landmarks	5 esophagus patients from DIR lab dataset			3D registration errors B-spline 1.84 (0.97)–3.72 (3.17) mmFDD 2.49 (1.21)–4.52 (3.45) OF 1.42 (0.92) –3.40 (2.93) Demons 1.40 (0.96) –4.39 (4.23)	(Kadoya et al 2014)
Thorax	CT to CT	4 RayStation (RaySearch5 MIM Software (Cleveland, OH),3 used Velocity	Compared to expert defined anatomical landmarks (DIR-Lab references)	10 patients with esophageal or lung cancer			3D registration errorRayStation 1.26–3.91 mm,MIM 2.17–3.61 mmVelocity 4.02–6.20 mm	(Kadoya et al 2016)
Lung	CT to CT	Demons, Salient Feature BAsed registration (PInnacle), Morphons	Compared to physician structures	17 NSCLC patients, 4D CTs (50% exhale was used)	data not shown in tables, please refer to the plots in the paper.	data not shown in tables, please refer to the plots in the paper.	COMGTV-tumor 0.27–0.29 cmnodal-GTVs 0.31–0.37 cm	(Hardcastle et al 2013)
Lung	4DCT-4DCBCT	Demons, SICLE	Compared to physician drawn reference	10 locally advanced non-small cell lung cancer patients, one 4D fan-beam CT and 7 weekly cone-beam CT;Day-to-day and phase-to-phase registrations	Day-to-dayMean DSC:SICLE 0.75Demons 0.70Rigid-tumor registration 0.66Rigid-bone registration 0.6Phase-to-phase (4D CBCT):SICLE 0.8Demons: 0.79			(Balik et al 2013)
Lung	4DCT	In-house Bspline, MIM freeform	Compared to physician drawn reference	4D-CTs of 12 lung cancer patients acquired in prone and supine positions	Mean DSC:In-house Bspline 0.8MIM 0.8	Mean HD:In-house Bspline 22.5 mmMIM 22.6 mm	Mean MDA:In-house Bspline 2.3 mmMIM 2.1 mm	(Guy et al 2019)
Lung	4DCT							

Table 2. (Continued.)

Indication	Image modality	DIR algorithm and/or vendor	Assessment method	Study type	DICE	HD	others (TRE, MDA, COM, ...)	Reference
		10 DIR algorithms (optical flow, demons)	Compared to physician defined reference/ fiducials (FM)	5 patients implanted fiducial markers (FM) as ground truth			TREFM positions 1.82–1.98 mm tumor position TREs 1.29–1.78 mm	(Han et al 2022)
Heart	MRI to MR	SVF-Net (DL-based)	Compare to segmented structures in the dataset	187 3D MRI cardiac images	No numbers reported, only plots, box-plot (25%-75%): Left ventricle myocardium ≈ 0.75 – 0.8 Right ventricle myocardium ≈ 0.45 – 0.55 Left ventricle blood pool ≈ 0.85 – 0.9 Right ventricle myocardium ≈ 0.75 – 0.85	No numbers reported, only plots, box-plot (25%-75%): Left ventricle myocardium ≈ 4 – 5.5 mm Right ventricle myocardium ≈ 5 – 6 mm Left ventricle blood pool ≈ 4 – 5 mm Right ventricle myocardium ≈ 4.5 – 6 mm		(Rohé et al 2017)
Cervical cancer	CT to CT	Velocity, Elastix	Compared to physician drawn reference	5 cervical brachytherapy patients	Mean DSC: Bladder Velocity 0.85 Rectum Velocity 0.72 Rectosigmoid Velocity 0.47 Bladder Elastix 0.76 Rectum Elastix 0.68 Rectosigmoid Elastix 0.50	Mean HDR Rectosigmoid Velocity 35.94 mm Rectosigmoid Elastix 40.76 mm		(Belon et al 2015)
Intrahepatic cholangiocarcinoma (IHCC)	CT to CT	Five commercially available DIRs (Demons, B-splines, salient feature-based, anatomically constrained, finite element-based algorithm)	Compared to physician drawn reference	29 IHCC patients			Mean TRE: Demons 4.6 ± 2.0 mm; B-splines 7.4 ± 2.7 mm; salient feature-based 7.2 ± 2.6 mm; anatomically constrained 6.3 ± 2.3 mm; finite element-based 7.5 ± 4.0 mm; Maximum errors > 1 cm for all techniques	(Sen et al 2020)
Liver	CT to CT	MIM, Velocity, .	Compared to fiducial markers (FM) as ground truth	24 Patients with liver tumor, pre and post treatment images (median 10 months)			FM error MIM: 0.4 – 32.9 (9.3 ± 9.9) mm Velocity 0.5 – 38.6 (11.0 ± 10.0) mm	(Fukumitsu et al 2017)
Liver	CT to CT	Unsupervised Cycle-Consistent CNN (DL-based)	Compare to 20 anatomical points in the liver and adjacent organs marked by radiologists	Liver cancer (HCC) patients at Asan Medical Center, Seoul, South Korea: 555 scans for training, 50 scans for testing			TRE Arterial to Portal Elastix: 3.26 VoxelMorph: 6.67 CNN: 4.91 Delayed to Portal Elastix: 2.96 VoxelMorph: 5.35 CNN: 3.76% of Jacobian determinant ≤ 0 Arterial to Portal VoxelMorph: 0.0327 CNN: 0.0175 Delayed to Portal VoxelMorph: 0.0311 CNN: 0.0181 NMSE (normalized mean	(Kim et al 2019)

Table 2. (Continued.)

Indication	Image modality	DIR algorithm and/or vendor	Assessment method	Study type	DICE	HD	others (TRE, MDA, COM, ...)	Reference
pancreatic	CT to CBCT	B-spline registration mutual-information (MI), mattes mutual-information (mattes) and gradient magnitude (GM) and also different regularization levels $\lambda \in \{0.05; 0.005; 0.00025\}$, GM($\lambda = 0.05$), Rigid, MIM	Compared to physician drawn reference	Fifteen pancreatic cancer patients			square error) Arterial to PortalVoxelMorph: 0.0278 CNN: 0.0277 Delayed to PortalVoxelMorph: 0.0213 CNN: 0.0199 best registration outcome for the visual comparison, the lowest median deviation was obtained with GM($\lambda = 0.005$) and GM($\lambda = 0.05$), whereas the variation over the patient collective was much smaller for GM($\lambda = 0.05$).	(Ziegler et al 2019)
Prostate	CT to ultrasound	Rigid, MIM		10 prostate patients, HDR-brachytherapy	Mean DSC: Rigid 0.78 ± 0.06 DIR 0.93 ± 0.01	Mean HD: Rigid 11.64 ± 2.38 mm DIR 5.19 ± 1.47 mm	Mean MDA: Rigid 2.50 ± 0.70 mm DIR 0.69 ± 0.06 mm	(Vozzo et al 2021)
Prostate	CT to CT	intensity based Elastix	Compared to manual delineation	18 prostate cancer patients, 7–10 repeat CT	prostate 0.87 ± 0.05 , seminal vesicles 0.63 ± 0.18 , lymph nodes 0.89 ± 0.03 , Rectum 0.76 ± 0.06 , Bladder 0.86 ± 0.09	95 percentile HD prostate 3.35 ± 1.19 mm, seminal vesicles 4.76 ± 2.77 mm, lymph nodes 3.57 ± 0.99 mm, Rectum 10.83 ± 5.93 mm, Bladder 8.91 ± 6.76 mm	mean surface distance (MSD) prostate 1.42 ± 0.48 mm, seminal vesicles 1.97 ± 1.22 mm, lymph nodes 1.46 ± 0.44 mm, Rectum 3.29 ± 1.31 mm, Bladder 2.92 ± 1.90 mm	(Qiao et al 2019)
Prostate	CT to CT	improved AI DIR in Elastix	Compared to manual delineation	evaluation on 2 datasets 14+18 patients, follow up on Quiao et al, improved adaptive dose constraints with this one	results on two datasets Prostate $0.87 \pm 0.08/0.87 \pm 0.12$ seminal vesicles $0.70 \pm 0.13/0.75 \pm 0.18$ Lymph nodes $0.87 \pm 0.07/-$ Rectum $0.82 \pm 0.12/0.78 \pm 0.15$ Bladder $0.89 \pm 0.12/0.83 \pm 0.17$	results on two datasets in mm Prostate $3.07 \pm 1.30/3.93 \pm 2.24$ seminal vesicles $3.82 \pm 3.19/4.92 \pm 5.13$ Lymph nodes $3.74 \pm 1.02/-$ Rectum $8.66 \pm 6.92/10.4 \pm 7.77$ Bladder $5.11 \pm 4.38/11.5 \pm 12.5$	mean surface distance (MSD) results on two datasets in mm Prostate $1.29 \pm 0.39/1.54 \pm 0.67$ seminal vesicles $1.48 \pm 1.16/1.67 \pm 1.38$ Lymph nodes $1.49 \pm 0.44/-$ Rectum $2.39 \pm 1.92/2.67 \pm 1.76$ Bladder $1.72 \pm 1.17/3.89 \pm 4.00$	(Elmahdy et al 2019)
prostate	CT-CBCT	anaconda	Compared to physician drawn reference	10 prostate patients	body ROI controlling: prostate 0.84 ± 0.05 rectum 0.75 ± 0.05 bladder 0.69 ± 0.07 seminal vesicles 0.65 ± 0.11 all ROIs controlling: prostate 0.98 ± 0.00 rectum 0.97 ± 0.01 bladder 0.98 ± 0.00 seminal vesicles 0.94 ± 0.03		COM body ROI controlling (mm): prostate 2.0 ± 1.5 rectum 3.7 ± 1.4 bladder 7.8 ± 2.2 seminal vesicles 3.6 ± 1.2 all ROIs controlling (mm): prostate 0.1 ± 0.0 rectum 0.3 ± 0.2 bladder 0.2 ± 0.1 seminal vesicles 0.6 ± 0.6	(Takayama et al 2017)
Prostate	CT to CT and CT to CBCT	3 DIR algorithms implemented in MIM (DIR)	Compared to manually drawn reference	20 patients (453 fractions)	CT to CT bladder: $0.729-0.943$ rectum: $0.737-0.913$ CT	CT to CT bladder: 7.26–18.40 mm rectum: 9.63–16.37 mm CT to	MDA in mm CT to CT bladder: $0.86-4.47$ rectum:	(Hammers et al 2020)

Table 2. (Continued.)

Indication	Image modality	DIR algorithm and/or vendor	Assessment method	Study type	DICE	HD	others (TRE, MDA, COM, ...)	Reference
Prostate	CT to MRI, MRI to MRI	Profile, normalized intensity-based (NIB) and shadowed NIB DIR algorithms Monaco DIR	Compared to manual defined structures and intra observer variability	12 high-risk prostate cancer patients, prostate and pelvic lymph nodes treated on MRI linac	to CBCTbladder: 0.713–0.906 rectum: 0.710–0.879 CT to MRI Prostate 0.84, Seminal Vesicles 0.68, Rectum 0.77, Bladder 0.87, R fem. Head 0.93, L fem. Head 0.91 MRI to MRI Prostate 0.90, Seminal Vesicles 0.76, Rectum 0.87, Bladder 0.92, R fem. Head 0.95, L fem. Head 0.94 Inter observer Prostate 0.92, Seminal Vesicles 0.81, Rectum 0.95, Bladder 0.97, R fem. Head 0.95, L fem. Head 0.94	CBCTbladder: 12.24–22.57 mm rectum: 11.25–18.49 mm CT to MRI Prostate 7.16 mm, Seminal Vesicles 6.55 mm, Rectum 12.36 mm, Bladder 10.88 mm, R fem. Head 4.96 mm, L fem. Head 4.98 mm MRI to MRI Prostate 5.10 mm, Seminal Vesicles 5.54 mm, Rectum 8.89 mm, Bladder 5.71 mm, R fem. Head 4.77 mm, L fem. Head 4.75 mm Inter observer Prostate 4.89 mm, Seminal Vesicles 5.31 mm, Rectum 07.65 mm, Bladder 4.05 mm, R fem. Head 4.41 mm, L fem. Head 5.21	0.89–2.96 CT to CBCTbladder: 1.51–4.68 rectum: 1.31–3.29 mean surface distance, CT to MRI-Prostate 1.6 mm, Seminal Vesicles 1.48 mm, Rectum 2.41 mm, Bladder 1.96 mm, R fem. Head 1.09 mm, L fem. Head 1.37 mm MRI to MRI Prostate 1.00 mm, Seminal Vesicles 1.17 mm, Rectum 1.25 mm, Bladder 1.11 mm, R fem. Head 0.81 mm, L fem. Head 0.81 mm Inter observer Prostate 0.88 mm, Seminal Vesicles 0.86 mm, Rectum 0.65 mm, Bladder 0.55 mm, R fem. Head 0.75 mm, L fem. Head 1.05 mm	(Christiansen et al 2020)
Prostate	MRI to transrectal ultrasound	Weakly-supervised CNN (DL-based)	Compare to manually segmented structures	108 pairs of T2-weighted MR and TRUS images	Composite-Net Median: 0.82 Percentiles [25th, 75th]: [0.78, 0.86]		TRE (mm): Composite-Net Median: 4.7 Percentiles [25th, 75th]: [3.3, 7.5]	(Hu et al 2018, p 218)
Phantoms		Elastix, BRAINS, Plasticmatch, Raystation	Compared to results from synthetic image datasets from applying synthetic DVFs	4 computational anthropomorphic phantoms	Mostly DSC > 0.85 Only smallest structures mild failure DSC < 0.75		In case of severe deformations MDC > 3 mm	(Scaggion et al 2020a)

DIR: deformable image registration, HN: Hausdorff distance, TRE: target registration error, MDA: mean distance to agreement, COM: center of mass, HD: Hausdorff distance

Table 3. Literature review of quantified dosimetric uncertainties in DIR-facilitated processes for different anatomical regions and imaging modalities. DIR: deformable image registration, SBRT: stereotactic body radiotherapy, VMAT: volumetric modulated arc radiotherapy, IMPT: intensity modulated proton therapy, DDM distance discordance metric.

Application	Indication	Image modality	DIR algorithm and/or vendor	Assessment method	Study type	Dosimetric uncertainty	Reference	
Structure propagation	Photons interfraction dose recalculation	Head and neck	CT to CBCT	MIM	compared to physician reference	30 head and neck patients, squamous cell carcinoma of the oral cavity, pharynx or larynx, DIR to first and last CBCT	Dose difference when dose is evaluated on propagated versus reference structures <u>First CBCT</u> Parotid L 0.1 Gy Parotid R -0.1 Gy Submandibular L 0.1 Gy Submandibular R 0.1 Gy Esophagus 0.0 Gy Spinal cord 0.1 Gy <u>Last CBCT</u> Parotid L 0.1 Gy Parotid R -0.1 Gy Submandibular L -0.3 Gy Submandibular R -0.5 Gy Esophagus 0.3 Gy Spinal cord 0.0 Gy	(Hvid <i>et al</i> 2016)
	Photons interfraction dose recalculation	Head and neck	CT to CBCT	Five commercially available DIRs (RayStation, ADMIRE, Mirada, ProSoma, Pinnacle)	compared to physician drawn and STAPLE reference	10 head and neck patients: 5 oropharyngeal, 2 oral cavity, 1 hypopharynx, 1 supra-glottic and 1 of unknown primary (target below nasal region)	Spinal cord D1cc occasionally exceeds planning tolerance (44 Gy) by 7–250 cGy Brainstem D1cc occasionally exceeds planning tolerance (54 Gy) by (29–199 cGy) Despite poor geometric agreement, the DVH parameters of propagated contours gave a reliable estimate of the organ dose	(Nash <i>et al</i> 2022)
Photon adaptive planning (cyberknife)	Pancreas	CT to CT	Precision, MIM	compared to physician reference	35 pancreas patients, 98 fx CTs, breathhold	Plans optimized on propagated and reference contours, evaluated on reference contours Dose difference between no adaptation and <u>a) Physician reference</u> PTV -2.0% GTV -0.1% Stomach V35 Gy -0.2 cc <u>b) Precision</u> PTV -2.7% GTV -0.4% Stomach V35 Gy -0.1 cc <u>c) MIM</u> PTV -5.1% GTV -1.6% Stomach V35 Gy -0.1 cc	(Magallon-Baro <i>et al</i> 2022)	

Table 3. (Continued.)

Application	Indication	Image modality	DIR algorithm and/or vendor	Assessment method	Study type	Dosimetric uncertainty				Reference	
						Duodenum V35 Gy -0.4 cc	Duodenum V35 Gy -0.2 cc	Duodenum V35 Gy -0.2 cc			
Proton adaptive planning	Prostate	CT to CT	Elastix	compared to manual delineation	18 prostate cancer patients, 7-10 repeat CT	Plans optimized on propagated and reference contours, evaluated on reference contours Propagated contours could be directly used for reoptimization (V95% ≥ 98% for each target volume) in 89% of cases				(Qiao et al 2019)	
Proton adaptive planning	Lung	CT to CT	Plastimatch (B-splines, demons, Velocity, Mirada, Raystation (Anaconda, Morfeus)	compared to physician reference	5 NSCLC patients with 9 repeated DIBH CTs	Plans optimized on propagated and reference contours, evaluated on reference contours 0.04% average difference in CTV V95 with DIR versus 0.06% with rigid propagation and 9.7% without adaptation				(Nenoff et al 2021b)	
Proton adaptive planning	Lung & Head and neck	CT to CT	Rigid registration, Plastimatch B-splines, Commercial CNN, patient-specific CNN	autocontouring techniques compared to manual delineation	5 NCSLC patients 9 repeated CTs and 5 head and neck cancer patients with 4-7 repeated CTs	Plans optimized on automatic OARs contours showed small dependence on the contouring method (<5%). For automatic target contours the dosimetric effect can be larger than 5%. Compared to non-adaptive approaches the automatic contouring showed improved target coverage.				(Smolders et al 2023a)	
Dose accumulation	Photon adaptive planning	Head and neck	CT to CT	Raystation Anaconda (simple & detailed) Raystation Morfeus (simple & detailed)	Not applicable	10 head and neck patients with weekly offline replanning	Deformed weekly doses accumulated and compared to primary planning dose- Difference to primary planned dose:				(Zhang et al 2018)
							Simple Anaconda	Detailed Anaconda	Simple Morfeus	Detailed Morfeus	
							Homogeneity index	0.137 ± 0.115 ± 0.032	0.006 ± 0.096	0.006 ± 0.033	
							Main difference between simple and detailed algorithms.				
							Simple presetting: 344.6 cGy, 109.9 cGy, 329.0 cGy for D95, Dmean, Dmin in average				
							Detailed presetting: less than 20 cGy				
	Photon 4D dose calculation	Lung, liver	4DCT	6 open source algorithms from EMPIRE challenge (ANTS,	Not applicable	5 patients with multiple lung metastasis, 5 patients with multiple	GTV D95% difference between plan on average CT and 4D dose simulation Lung metastasis: Variations mostly negligible (<0.5 Gy), but up to 7.85 Gy Liver metastasis: Larger variations more diverging, higher negative, up to -29.09 Gy				(Mogadas et al 2018)

Table 3. (Continued.)

Application	Indication	Image modality	DIR algorithm and/or vendor	Assessment method	Study type	Dosimetric uncertainty	Reference																												
Photon 4D dose calculation	Lung	4DCT	VarReg, DIR-ART, NiftyReg, Elastix, Plastimatch)	Not applicable	liver meta-satsis, VMAT 6 lung SBRT patients, VMAT	If results are limited to visually acceptable deformed images: Maximum difference in the evaluated DVH parameters was $\leq 3.0\%$ for GTV D98, spinal cord D2%, heart D2% and $\leq 3.6\%$ of the total structure volume for the ipsilateral lung	(Sarudis <i>et al</i> 2019)																												
Proton 4D dose calculation	Liver	4DCT (generated from 4D MRI)	SmartAdapt, Velocity, Anaconda (Raystation)	Not applicable	9 liver cancer patients with generated 4DCTs, applying motion from 4DMRI, IMPT	CTV V95 differences up to $11.34 \pm 12.57\%$ for single fields without rescanning, large motion CTV V95 differences up to $3.46 \pm 1.40\%$ for three-field plans with rescanning, large motion CTV V95 differences up to $0.37 \pm 0.38\%$ for three-field plans with rescanning, small motion	(Ribeiro <i>et al</i> 2018)																												
Photon dose calculation inter and intra-fraction	Lung	CT to CBCT	Admire (Eleta)	Not applicable	20 lung SBRT patients, comparison if inter- and intrafractional differences	95-percentile of DDM (in mm) and dosimetric errors (in Gy)	(Huesa-Berral <i>et al</i> 2022)																												
						<table border="1"> <thead> <tr> <th>Structure</th> <th>DDM Intrafraction in mm</th> <th>DDM Interfraction in mm</th> <th>DDM Interfraction dosimetric in Gy</th> </tr> </thead> <tbody> <tr> <td>GTV</td> <td>0.93</td> <td>1.54</td> <td>1.67</td> </tr> <tr> <td>Lung</td> <td>1.86</td> <td>2.16</td> <td>0.86</td> </tr> <tr> <td>Ribs</td> <td>1.66</td> <td>5.13</td> <td>1.05</td> </tr> <tr> <td>Heart</td> <td>6.26</td> <td>2.34</td> <td>0.57</td> </tr> <tr> <td>Esophagus</td> <td>1.38</td> <td>2.55</td> <td>0.29</td> </tr> <tr> <td>Spinal cord</td> <td>0.16</td> <td>8.00</td> <td>1.28</td> </tr> </tbody> </table>	Structure	DDM Intrafraction in mm	DDM Interfraction in mm	DDM Interfraction dosimetric in Gy	GTV	0.93	1.54	1.67	Lung	1.86	2.16	0.86	Ribs	1.66	5.13	1.05	Heart	6.26	2.34	0.57	Esophagus	1.38	2.55	0.29	Spinal cord	0.16	8.00	1.28	
Structure	DDM Intrafraction in mm	DDM Interfraction in mm	DDM Interfraction dosimetric in Gy																																
GTV	0.93	1.54	1.67																																
Lung	1.86	2.16	0.86																																
Ribs	1.66	5.13	1.05																																
Heart	6.26	2.34	0.57																																
Esophagus	1.38	2.55	0.29																																
Spinal cord	0.16	8.00	1.28																																
Photon dose calculation	Abdomen	CT to CT	Thin Plate Spline—Robust Point Matching	Not applicable	16 liver SBRT patients, DIR uncertainty modeled by systematic	The dosimetric impact of Interfraction changes is larger than intrafraction motion After selection of ‘realistic’ deformations, average difference between the 1st and 99th percentile of the cumulative maximum doses: 1.4 Gy for esophagus 0.7 Gy for stomach 0.9 Gy for duodenum (maximum difference for one patient: 3.3 Gy)	(Wang <i>et al</i> 2018)																												

Table 3. (Continued.)

Application	Indication	Image modality	DIR algorithm and/or vendor	Assessment method	Study type	Dosimetric uncertainty	Reference
Tomotherapy	Head and neck	CT to megavoltage CT	algorithum with variable settings PreciseART (Accuray)	Not applicable	variation of registration parameters 20 Head and neck patients with daily MVCTs	Doses from daily MVCTs reconstructed and accumulated on the planning CT and compared to planned dose with warped contours on the daily MVCTs. Average dose uncertainty bounds (and confidence interval) for the cumulative treatment were: Parotids mean dose: 3.5% (97.1%–107.0%) Parotids D50%: 6.6% (98.2%–110.4%) Parotids V20Gy: 4.6% (95.6%–111.1%) PTV D95%: 0.4% (98.2%–100.2%)	(García-Alvarez <i>et al</i> 2022)
Photon adaptive planning	Head and neck	CT to CBCT	4 different NiftyReg approaches	Not applicable	5 Head and neck cancer patients with weekly CBCTs	The four DIR methods resulted in similar geometrical matching, but smoothness and inverse consistency differed. The root mean squared dose difference of the different warped doses was $1.9\% \pm 0.8\%$. $9\% \pm 4\%$ of voxels within the treated volume failed a 2% dose difference test, this value was larger in high dose gradient regions ($21\% \pm 6\%$) and for poor CBCT quality regions ($28\% \pm 9\%$).	(Veiga <i>et al</i> 2015)
Photon VMAT	Head and neck	CT to CBCT	Bspline DIR, Varian's demons DIR	Not applicable	12 Head and neck patients with 4 CBCTs	In-silico reference created with a B-spline algorithm. Inverse consistency was assessed by forward and backward deformation. Dose was reconstructed by the demons algorithm and compared to the in-silico ground truth. 98.5% of all voxels were inverse consistent with the following confidence interval for the dose reconstruction of a single fraction relative to planned dose: Target structures: [2.3%; +2.1%] Critical OARs: [10.2%; +15.2%] Non-critical OARs: [9.5%; +12.5%] Inverse inconsistent voxels were associated with higher uncertainties.	(Lowther 2020a, 2020b)
Photon dose calculation	Prostate	CT to CT	Demons algorithm	Not applicable	1 prostate patient with 9 CTs	Quantification of errors with unbalanced energy (UE) and compared to standard displacement error (SDE). High Pearson correlation above 70% between UE and SDE. Mean dose reconstruction error in target over nine fractions 1.68%.	(Zhong <i>et al</i> 2008)
Photon dose calculation inter-fraction	Prostate	CT to CBCT	Demons algorithm	Not applicable	24 prostate patients with 8 weeklc CBCTs for 21 patients and daily CBCTs for 3 patients	Quantification of differences between planned and cumulated doses using DIR-based dose accumulation and quantifying the dose accumulation uncertainties with a numerical pelvis phantom. Standard deviation of the dose difference between planned and accumulated dose Mean bladder dose: 6.9 Gy Mean rectum wall dose: 2.0 Gy Dose accumulation uncertainty: Mean bladder dose: 2.7 Gy Mean rectum wall dose: 1.2 Gy	(Nassef <i>et al</i> 2016)
Proton adaptive planning	Lung	CT to CT	Plastimatch (B-splines, demons),	Not applicable	5 NSCLC patients with 9 repeated DIBH CTs	PTV-V95 decrease without adaptation by 14% (range: 1.5% – 40.5%) DIR-caused variations in PTV-V95 of accumulated doses on average 8.7% (range 1.0% – 26.3%)	(Nenoff <i>et al</i> 2021b)

Table 3. (Continued.)

Application	Indication	Image modality	DIR algorithm and/or vendor	Assessment method	Study type	Dosimetric uncertainty	Reference
Proton adaptive planning	Head and neck	CT to CT	Velocity, Mirada, Raystation (Anaconda, Morfeus)	Not applicable	1 Head and neck patient with 8 repeated CTs	After individually warping the dose with the different DIR algorithms, the volume for which the dose uncertainty in the accumulated dose was larger than 10% was ($V_{\text{dosediff}>10\%}$): Contralateral parotid: 28.1% Ipsilateral parotid: 13.9% Contralateral Retina: 9.4% Contralateral Macula: 8.9%	(Amstutz <i>et al</i> 2021a)
Proton, Photon and Combined proton-photon adaptive planning	Lung	CT to CT	Plastimatch (B-splines, Velocity, Mirada, Raystation (Anaconda, Morfeus)	Not applicable	5 NSCLC patients with 3 repeated DIBH CTs	Difference between the deposited fractional energy and the energy in the representation of the warped dose on the planning CT: Energy conservation violation in the accumulated energy averaged over treatment modalities and DIR algorithms compared to fractional deposited energy: GTV: 40.9% PTV: 32.1% OARs: randomly distributed within $\pm 10\%$ Energy conservation violation in traditional intensity-based DIR is linearly correlated to mass/volume variations.	(Wu <i>et al</i> 2023)

a statistically significant impact on OAR dose-volume-histograms (DVH) parameters and concluded that DIR propagated structures are suitable for dose evaluation (Nash *et al* 2022). Similar conclusions were found by Hvid *et al* (2016).

Also for proton therapy the dosimetric impact of using propagated structures for proton dose evaluation and optimization has been investigated: Qiao *et al* and Elmahdy *et al* investigated prostate structures propagated from CT to CT with the open source DIRs in Elastix (Elmahdy *et al* 2019, Qiao *et al* 2019). They gave an extensive geometrical evaluation (included in table 2) and dosimetric evaluation (included in table 3) that showed that DIR propagated structures can be used for optimization in online-adaptive intensity-modulated proton therapy (IMPT). Similar conclusions were found for lung cancer patients by Nenoff *et al*, showing that daily IMPT optimization on CT based on propagated, uncorrected structures was better than no adaptation (Nenoff *et al* 2021b). Daily manual recontouring on each CT gives a small additional benefit for some patients and OARs. They also investigated if including the inter-algorithm variation between structures propagated with DIR in the adaptive IMPT optimization could improve the adapted plan against structure uncertainties (Nenoff *et al* 2022). They found that adaptation on propagated, uncorrected structures showed a benefit over no adaptation for MRI-to-MRI registrations for pancreas and liver patients and CT-to-CT registrations for HN patients. Only for the HN patients including structure propagation uncertainties in the optimization significantly improved the adapted plan. Recently, Smolders *et al* compared the effect of different auto-segmentation methods, among those DIR based structure propagation, for the dosimetric quality of online adaptive proton therapy plans. They found the dosimetric influence of using automatic contours for the optimization to be small for OARs and larger for targets, with DIR propagated structures performing best for both OARs and targets (Smolders *et al* 2023a).

5.3. Mapped/Accumulated doses

In this section we outline the influence of DIR uncertainty on dose mapping and accumulation. For more details about dose mapping and accumulation, including direct dose mapping versus energy/mass mapping, biological considerations and (dis)appearing tissue please refer to the recent review of (Murr *et al* 2023).

5.3.1. Intrafraction applications

A commonly proposed use of dose accumulation is for 4D treatment planning or the dose reconstruction of the dose in a moving area. Both, 4D optimisation (Graeff *et al* 2013, Engwall *et al* 2018, Spautz *et al* 2023) and 4D dose evaluation (Zhang *et al* 2019, Meijers *et al* 2020) require DIR and therefore show DIR-related uncertainties. Both are mostly applied in anatomical areas affected by breathing motion, registering all phases of a 4D CT or 4D MRI scan into a reference phase or average image (Rosu and Hugo 2012, Engwall *et al* 2018, Meijers *et al* 2020). Most 4D dose optimisation and dose calculation studies do not investigate DIR-facilitated dosimetric uncertainties (table 3). Those who do, report different metrics between different studies, to quantify these uncertainties. For example, (Ribeiro *et al* 2018) found differences in the target V95% of up to 11.34% for 4D dose accumulation of liver cancer proton therapy. In contrast, (Sarudis *et al* 2019) found only dose deviations of $\leq 3.0\%$ between different visually acceptable DIRs in 4D lung volumetric modulated arc therapy (VMAT) dose accumulations. Mogadas *et al* tested five open-source registration algorithms on lung and liver SBRT, using the delta D95% of the target using 4D dose reconstruction compared to the static plan. For lung metastases, accumulated dose distributions were similar regardless of the DIR algorithm. In contrast, for liver metastases, accumulated dose distributions strongly varied, due to large DIR uncertainties in low contrast regions (Mogadas *et al* 2018).

5.3.2. Interfraction applications

Another DIR-facilitated application has been to map doses re-calculated (or re-optimized) on 3D images from different fractions on the planning CT, to get an estimation of the total delivered treatment dose (Chetty and Rosu-Bubulac 2019, Ziegler *et al* 2019, Nenoff *et al* 2020). This technique has been extended to evaluate the validity of treatment plans with reduced margins (Wu *et al* 2009, van Kranen *et al* 2016, Lowther *et al* 2020b, van der Bijl *et al* 2022). The reporting of the uncertainties is very application-dependent and not standardised, which makes direct comparisons challenging (table 3). Examples of the magnitude of dose mapping uncertainties for interfractional changes include (Nenoff *et al* 2020), reporting differences caused by DIR uncertainties of 8.7% in the CTV of accumulated proton doses and (Wang *et al* 2018) reporting a maximum dose variation of 3.3 Gy for hollow organs in the abdomen for interfraction dose mapping. More recently, (Huesa-Berral *et al* 2022) reported a dosimetric uncertainty between fractions below 2 Gy in tumour and OAR in lung SBRT. This study also concluded that inter-fraction variations dominated and that dose accumulation for these patients should prioritise day-to-day changes over respiratory motion.

There have been several propositions to also predict uncertainties on the geometrical and the dosimetric level. The inter-algorithm variability was proposed to be used for geometric as well as dosimetric DIR uncertainties (Nenoff *et al* 2020, Amstutz *et al* 2021b). Probabilistic unsupervised DL methods have also been proposed to predict the variance of DVFs in interfraction datasets (Gong *et al* 2022, Smolders *et al* 2022b, 2022a).

5.3.3. Intertreatment applications

Dose mapping and accumulation have been used in work on treatment method combinations and patient re-irradiation. Application of the technology presents the possibility of greater outcome modelling in combined methodologies, and long term outcomes in re-irradiation. Research regarding the combination of external beam RT and brachytherapy was done for cervical cancer patients (Vásquez Osorio *et al* 2015, Swamidás *et al* 2020, Zeng *et al* 2020). Van Heerden did not find clinically relevant improvements when using DIR for dose accumulation compared to adding uniform external beam RT doses or overlapping high dose volumes (van Heerden *et al* 2017).

In recent years, improved survival has led to an increase in the numbers of re-irradiations (Nieder *et al* 2013, Andratschke *et al* 2022) with particular focus made on cancers of the brain (glioma), lung, HN, abdomen, pelvis and spine (Abusaris *et al* 2012, Mantel *et al* 2013, De Ruyscher *et al* 2014, Nieder *et al* 2016). Dose from previous treatments can be deformed to the current anatomy to evaluate potential dose overlap (Meijneke *et al* 2013, Nix *et al* 2022). Thereby, being used to define safe dose tolerances in those previously treated regions (Embring *et al* 2021, Andratschke *et al* 2022, Brooks *et al* 2022, Nix *et al* 2022). In addition, DIR-facilitated dose warping can be used to correlate places of local failure with previously planned and/or delivered dose distributions (Boman *et al* 2017, McVicar *et al* 2018, Skjøtskift *et al* 2018, Embring *et al* 2021, Nix *et al* 2022). Registration algorithms are challenged by dramatic anatomical changes caused by the time between treatments, often months or years, not to mention sequels of treatments such as fibrosis resulting from radiation or surgery (Nix *et al* 2022, Vasquez Osorio *et al* 2023b). Systematic studies about the DIR uncertainties in the re-irradiation setting are rare, but some reports indicate that DIR uncertainty increases with the magnitude of anatomical changes, in particular for lung radiographic changes after SBRT (Mahon *et al* 2020). DIR uncertainty is only one of multiple uncertainty factors which makes the definition of organ constraints for re-irradiation challenging. The lack of standardised toxicity scoring or cumulative DVHs over multiple treatments, partially influenced by DIR uncertainty remain reasons why the recovery of organs over time is not well quantified. The calculation of biologically effective dose can improve the understanding of normal tissue responses over time (Brooks *et al* 2022, Nix *et al* 2022) and allow a better estimation of safe dose constraints during re-irradiation.

5.4. Other DIR-facilitated applications

DIR uncertainties can affect other medical physics and imaging applications.


5.4.1. TCP and NTCP calculation

Currently, tumour control probability (TCP) and normal tissue complication probability (NTCP) models are built on planned doses. They are however designated to correlate to delivered doses which can differ from the planned dose. Dose accumulation of reconstructed doses on repeated images, requiring DIR in most anatomical areas, is the closest surrogate to the delivered dose that is available. The impact of DIR uncertainty on the accumulated doses directly affects the outcome calculation (Nenoff *et al* 2021a, Smolders *et al* 2023b). Deformation-free methods (Niemierko 1997, Niebuhr *et al* 2021) have their own (not well quantified) uncertainties. Niebuhr *et al* found larger differences when assuming a registration error of 3 mm, compared to changing alpha-beta values for prostate RT. (Niebuhr *et al* 2021) There is more research needed to fully understand and quantify the impact of DIR uncertainty for outcome calculation.

5.4.2. Outcome modelling based on spatial/voxel-based analyses

Conventional outcome modelling simplifies the planned dose distribution to a single value, often using DVH statistics. Voxel-based analysis techniques that maintain the spatial distribution of doses have been used to explore local correlations between dose and treatment outcomes. Voxel-based analysis (figure 6) relies on DIR to 'spatially normalise' dose distributions into a common reference anatomy (Palma *et al* 2020, Shortall *et al* 2021). In summary, DIR is first performed between the planning CTs of each patient and an arbitrarily selected reference CT scan. The DIR result is then used to map the dose distributions to the reference anatomy allowing the local dose to be correlated with the studied outcome. The region is evaluated with statistical modelling, often quantifying the improvement in model discrimination when the dose to the identified region is included in a multivariable predictive model (including other demographic and clinical variables). The region is then used to generate hypotheses which are then tested and validated in external cohorts aiming at generating dose constraints to ultimately improve treatment outcomes.

With voxel-based techniques, doses to anatomical subregions have been linked to outcomes, such as the dose to the base of the heart to overall survival in lung RT (McWilliam *et al* 2017, Green *et al* 2020), the



inferior–anterior hemi-anorectum dose to rectal bleeding in prostate RT (Dréan *et al* 2016) and the cricopharyngeus muscle, cervical oesophagus and the base of the brainstem dose to dysphagia in HN RT (Monti *et al* 2017).




Table 4. Voxel-wise dosimetric uncertainty as a function of the dose gradient and the uncertainty of the DIR. DIR: deformable image registration.

	Dose gradient		
	Low	Medium	High
DIR uncertainty	1 %/mm	10 %/mm	25 %/mm
Low 1 mm	1%	10%	25%
Medium 5 mm	5%	50%	125%
High 10 mm	10%	100%	250%

Several measures to evaluate DIR uncertainty for voxel-based analysis have been proposed (Palma *et al* 2020, Shortall *et al* 2021, McWilliam *et al* 2023). Quantified DIR uncertainties are often incorporated in the analysis by treating them as random errors and blurring the mapped dose distributions (McWilliam *et al* 2017, Beasley *et al* 2018, Green *et al* 2020, Vasquez Osorio *et al* 2023a). Therefore, DIR uncertainties can result in a decrease of significance for small radiosensitive regions and local changes in their shapes.

6. Uncertainty tolerances of DIR-facilitated dosimetric procedures

Specifying tolerances for the uncertainty in DIR-facilitated procedures is a challenging task and these should be based on clinical needs rather than achievable results. The demands on the accuracy of DIR vary by application. In a retrospective analysis, larger tolerances might be sufficient, while for interventional applications tighter tolerances might be indicated. For example, visualising a voxel-wise dose uncertainty map might be sufficient for a crude estimation of the dose in a re-irradiation case while precise DVH metrics along with their uncertainty estimation are necessary for correlating the dose to organs with outcome and toxicity data in clinical trials. In contrast to tolerances for geometric uncertainties, there is a scarcity of literature describing these for dose mapping or accumulation. There is no generally accepted approach on how to analyse and report DIR-related dosimetric uncertainties. Publications evaluating DIR-facilitated dosimetric differences are summarised in chapter 5 and table 3. A common finding in dose accumulation studies is that areas with steep dose gradients are more sensitive to DIR-facilitated uncertainties (Saleh-Sayah *et al* 2011, Swamidas *et al* 2020, Amstutz *et al* 2021b). Therefore, in areas with steep dose gradients DIR uncertainties are more relevant than in homogeneous dose areas or areas with low doses. Table 4 shows clinically relevant examples of dose gradients as well as geometric DIR uncertainties. Multiplying the dose gradient with the geometric DIR uncertainty gives an assessment of the dosimetric uncertainties expected in these situations. Low dose gradients are typically found in the central region of the target. Medium dose gradients are found in OARs in the beam path and high dose gradients are found close to the target boundary. As both the dose gradient and DIR uncertainty typically vary within an organ, voxel-wise dose uncertainty maps can visualise dose distribution uncertainties (figure 7).

	Task	Source of information	Availability / development status	
Step 1	Select appropriate algorithm for the application	See description of algorithms in section 2	<ul style="list-style-type: none"> • Open source available • Commercially available 	Level of knowledge
Step 2	Commissioning the algorithm	See recommendation section and Appendix 1	<ul style="list-style-type: none"> • Few guidelines available • Recommendation by groups of experts 	
Step 3	Geometric assessment of the uncertainty of the algorithm in different regions	Select appropriate method/metric in dimension of length (see Table 1 - geometric measures).	<ul style="list-style-type: none"> • Open source • Rarely commercially available • Ongoing research 	
Step 4	Generate a voxel-wise model of geometric uncertainties based on the results of Step 3	Define voxel-wise DVF uncertainty map (see section 6)	<ul style="list-style-type: none"> • Ongoing research • Not (widely) available 	
Step 5	Combine voxel-wise geometric uncertainty with dose gradient to visualise dose uncertainty map, subsequently, calculate DVH metrics	Table 1 - dosimetric metrics	<ul style="list-style-type: none"> • Ongoing research • Not (widely) available 	
Step 6	Use results of step 5 to derive tolerances while considering clinical needs	-	<ul style="list-style-type: none"> • Future research 	
				Low

Figure 8. An example approach on how to assess dosimetric uncertainties of accumulated dose caused by DIR-uncertainties. The shading indicates the level of knowledge/confidence of the individual steps.

Since there is no standard agreed upon in the literature on how to quantify dosimetric DIR uncertainties or tolerances, we propose a short ‘recipe’ (figure 8). The first step is the selection of the DIR algorithm. Second, the algorithm must be commissioned for the specified application (recommendations in chapter 7 and commissioning document in the supplement). Third, the DIR uncertainty is evaluated using geometric measures. We consider geometric measures in dimension of distance (e.g. target registration error (TRE), MHD) necessary to define tolerances. The quantification of geometric measures needs to be done for different structures and points of interest such as targets, OARs, anatomical landmarks close to the target or in the beam path.

Steps 1–3 are described in multiple recommendations (Brock *et al* 2017, Barber *et al* 2020, Lowther *et al* 2022). In step 4 a voxel-wise geometric uncertainty map of geometrical measures is created (Amstutz *et al* 2021b). The simplest method is using the worst-case or average difference distance in all directions for all voxels of a given region or structure. More individualised methods have been investigated (Amstutz *et al* 2021b, Smolders *et al* 2022b, 2023c) and provide patient specific voxel-wise uncertainty maps. We recommend using such voxel-wise uncertainty maps whenever possible. However, due to the lack of commercial implementations, simpler global geometrical metrics are easy-to-implement alternatives. Using these metrics may lead to locally over- or underestimated geometric uncertainties, but its use is an improvement over no geometric uncertainty estimation, and will help pave the way to include such concepts in clinics.

In step 5 the geometric uncertainty map is applied to the dose by calculating the scalar product of the dose gradient and the geometric uncertainty of the DIR transformation on a voxel-wise level. This uncertainty map can be used to calculate DIR-facilitated variations of DVH parameters or DVH bands. To define geometric

tolerances of DVFs, steps 6 to 3 can be propagated backwards: starting with a maximum allowed DVH variation or uncertainty in a given voxel resulting in a maximum allowed DVF uncertainty. Since multiple relevant methods are not yet widely available or still require future research the definition of tolerances is not trivial.

Another method to calculate the required accuracy of a registration to achieve a given tolerance is the distance-to-dose difference (DTD), proposed by Saleh-Sayah *et al*. The DTD indicates how large local registration errors can be before they introduce mapping errors breaching the given tolerance. For example accurate DVFs (1 mm) are required in high dose gradient regions while large DVF errors (>20 mm) are acceptable in low dose gradient regions. Another approach is to divide the acceptable dosimetric tolerance by the dose gradient (TDG). Compared to the TDG, the DTD gives a more conservative assessment (Saleh-Sayah *et al* 2011, Saleh *et al* 2014).

7. Recommendations

Several publications have offered recommendations for methods and action thresholds for assessing registration quality. We endorse these efforts. This section summarises these recommendations and extends recommendations for the community.

7.1. Recommendations for patient-specific use

TG-132 recommends visual inspection for patient-specific use, using split-screen, fusion, contour overlay, or other tools (Brock *et al* 2017). Visualisation should focus on alignment of anatomic landmarks, organ or tissue boundaries, vessels, and other distinct features. When software allows, the displacement field should be inspected to identify implausible deformations. Qualitative assessment can optionally be verified using quantitative metrics such as those summarised in table 2. TG-132 also recommends a threshold of 2–3 mm accuracy in TRE and MDA, although this is not achievable in practice (Rong *et al* 2021). Vector field smoothness should be tested for locations with negative Jacobian determinant. We suggest this threshold might lie between 0.2 and 2.0. MIRSIG recommends additional tests on the displacement field using DVF histograms, transitivity errors, and harmonic energy, but no thresholds are provided. TG-132 recommends a 2–3 mm threshold for inverse consistency, and a 0.8–0.9 threshold for DSC, with the caveat that DSC varies widely by structure volume.

Applications using dose deformation or dose accumulation should focus on the important regions of interest. Usually these are the volumes with meaningful dose levels, relevant structures and high dose gradients. The recipe proposed in chapter 6 can provide guidance how to calculate dosimetric uncertainties on a voxel-wise level.

7.2. Recommendations for commissioning

System commissioning requires testing software interchange, and TG-132 recommends using a physical phantom for this purpose. It also recommends testing on digital phantoms to recover known, artificial deformations. Best practices prospectively evaluate registration software on treatment sites of interest, but there are few guidelines on this. Glide-Hurst *et al* recommend centralised review of each fraction for at least the first three cases in adaptive therapy clinical trials (Glide-Hurst *et al* 2021). We recommend five representative patient cases to assess with quantitative metrics. These metrics should be compared to typical values from the literature (tables 2 and 3, commissioning document in the supplement) and with inter-observer variability.

7.3. Recommendations for developers, vendors, and the community

TG-132 recommends that vendors provide a basic description of the registration algorithm, vector field export, and basic quantitative tools (DSC, MDA, TRE). Unfortunately software providers still fail to apply these quantitative assessment tools (Rong *et al* 2021). More recently, Murr *et al* evaluated contour distance metrics and DVF analysis tools, such as DVF visualisation, transitivity analysis, and Jacobian determinant (Murr *et al* 2023). They recommend vendors to implement dose uncertainty tools, a region of interest (ROI) tool to limit registration domain, multiple algorithms for sensitivity analysis, and a greater selection of state-of-the-art algorithms.

In addition to these recommendations, we add:

- Tools for generating artificial warps
- ROI tools for quantitative metrics within a contour or dose level
- DIR correction tools, such as a smudge tool to locally push the registration, vector field smoothing tool, landmark-based correction, and contour-based correction

- Open access resources of reference images, structures, landmarks, and vector fields
- Tools to restrict DIR to be locally rigid or locally mass-preserving
- Tools to import and export DVFs in a consistent dicom format
- Voxel-wise uncertainty quantification and visualisation

7.4. Recommendations for future research

Finally, we propose areas where research is still needed.

TCP and NTCP metrics. It is unclear how DIR-generated dose distributions are related to clinical outcomes, considering uncertainties. Uncertainties in DIR-generated doses should be quantified with the metrics described in table 1 and utilised with the aim of generating more accurate TCP and NTCP models.

DIR failure modes. While it is possible to obtain typical uncertainty estimates during commissioning, many DIR algorithms have unexpected failure modes which are hard to enumerate. It is desirable to better understand the causes of these failures so that automated tests can be performed.

Uncertainty estimation methodology. There are multiple methods in use for estimating the uncertainty of DIR, and they are difficult to compare as they measure different aspects. Efforts should be made to find consensus on which methods should be preferred for each application.

Avoiding DIR. For online ART, improvements in imaging and dose calculation could eliminate the need to deform images with DIR for daily dose calculation and plan optimisation, and thereby eliminate it as a source of overall uncertainty. To evaluate the total accumulated treatment dose, DIR will remain necessary.

8. Summary

DIR is a powerful and versatile tool for RT. It has many applications, but is also associated with considerable uncertainties. Many clinical DIR solutions have been implemented, but they generally lack tools for uncertainty quantification. In the community, there are no agreed thresholds to distinguish between a good or bad DIR result when using a combination of geometric and dosimetric measures. Multiple quantification metrics, mostly using geometrical measures, and tolerances have been proposed. The reporting of dosimetric measures and uncertainties caused by DIR uncertainty is less standardised and highly application dependent. It is important to reach an agreement and standardisation in the evaluation of DIR uncertainties for different RT applications. In this review we summarised DIR-facilitated uncertainties for different applications and gave recommendations on the quantification of DIR uncertainties. We then outlined a potential path towards definition of tolerances. It should be emphasised that the presented recommendations are only a starting point, they should be challenged and refined by the community.

Acknowledgments

We thank the ESTRO physics section for supporting us with organising the physics workshop 2021 on deformable image registration. We thank the reviewers for their helpful and thorough reviews, which helped improve the manuscript substantially.

Data availability statement

The paper is a review paper, no new data was acquired. The data that support the findings of this study are available upon reasonable request from the authors.

Funding

LN: reports funding from the Swiss National Science Foundation (SNSF): grant 199943 and from the NIH R01 CA229178. FA: reports funding from the Swiss Cancer Research Foundation (KFS-4528-08-2018) and Krebsliga Schweiz Research Grant (KFS-4517-08-2018). MM: reports funding from DFG ZI 736/2-1 and MU 6403/1-1 (PAK 997/1-1). MH: supported by the National Measurement System of the UK's Department for Business, Energy and Industrial Strategy. GCS: reports funding from the NIH R01 CA229178. EVO: reports funding by Cancer Research UK RadNet Manchester [C1994/A28701]. BAH, MF, WL, YZ: No funding to disclose.

Conflicts of interest

LN: No conflict of interest. FA: No conflict of interest. MM: reports institutional research agreements with Elekta, Philips, TheraPanacea, Kaiku, PTW Freiburg. BAH: No conflict of interest. MF: No conflict of interest. MH: reports institutional collaboration agreement with RaySearch. WL: reports institutional research agreements with Brainlab, Elekta, Philips and RaySearch. YZ: No conflict of interest. GCS: No conflict of interest. EVO: No conflict of interest.

ORCID iDs

Lena Nenoff  <https://orcid.org/0000-0002-7468-835X>
Florian Amstutz  <https://orcid.org/0000-0003-1416-8093>
Ben Archibald-Heeren  <https://orcid.org/0000-0001-9013-6048>
Marco Fusella  <https://orcid.org/0000-0002-5361-8387>
Wolfgang Lechner  <https://orcid.org/0000-0001-9211-7510>
Ye Zhang  <https://orcid.org/0000-0003-1608-4467>
Eliana Vasquez Osorio  <https://orcid.org/0000-0003-0741-994X>

References

- Abdel-Aty H et al 2022 Mapping Local Failure following bladder radiotherapy according to dose *Clin. Oncol.* **34** e421–9
- Abusaris H, Hoogeman M and Nuytens JJ 2012 Re-irradiation: outcome, cumulative dose and toxicity in patients retreated with stereotactic radiotherapy in the abdominal or pelvic region *Technol. Cancer Res. Treat.* **11** 591–7
- Albertini F, Matter M, Nenoff L, Zhang Y and Lomax A 2020 Online daily adaptive proton therapy *Br. J. Radiol.* **93** 20190594
- Amstutz F, D'Almeida P G, Albertini F, Nenoff L, Weber D C, Lomax A J and Zhang Y 2021a PO-1535 Deformable image registration uncertainty for dose accumulation of proton therapy for H&N tumors *Radiother. Oncol.* **161** S1259–61
- Amstutz F, Nenoff L, Albertini F, Ribeiro C O, Knopf A C, Unkelbach J, Weber D C, Lomax A J and Zhang Y 2021b An approach for estimating dosimetric uncertainties in deformable dose accumulation in pencil beam scanning proton therapy for lung cancer *Phys. Med. Biol.* **66**
- Amugongo L M, Green A, Cobben D, van Herk M, McWilliam A and Osorio E V 2022 Identification of modes of tumor regression in non-small cell lung cancer patients during radiotherapy *Med. Phys.* **49** 370–81
- Andratschke N et al 2022 European society for radiotherapy and oncology and european organisation for research and treatment of Cancer consensus on re-irradiation: definition, reporting, and clinical decision making *Lancet Oncol.* **23** e469–78
- Ang K K et al 2014 Randomized Phase III trial of concurrent accelerated radiation plus cisplatin with or without cetuximab for stage III to IV head and neck carcinoma: RTOG 0522 *J. Clin. Oncol.* **32** 2940–50
- Balakrishnan G, Zhao A, Sabuncu M R, Dalca A V and Gutttag J 2018 An unsupervised learning model for deformable medical image registration 2018 IEEE/CVF Conference on Computer Vision and Pattern Recognition 2018 IEEE/CVF Conference on Computer Vision and Pattern Recognition pp 9252–60
- Balakrishnan G, Zhao A, Sabuncu M R, Gutttag J and Dalca A V 2019 VoxelMorph: a learning framework for deformable medical image registration *IEEE Trans. Med. Imaging* **38** 1788–800
- Balik S et al 2013 Evaluation of four-dimensional computed tomography to four-dimensional cone-beam computed tomography deformable image registration for lung cancer adaptive radiation therapy *Int. J. Radiat. Oncol. Biol. Phys.* **86** 372–9
- Barber J et al 2020 Deforming to Best Practice: key considerations for deformable image registration in radiotherapy *J. Med. Radiat. Sci.* **67** 318–32
- Beasley W et al 2018 Image-based data mining to probe dosimetric correlates of radiation-induced trismus *Int. J. Radiat. Oncol. Biol. Phys.* **102** 1330–8
- Beasley W J, McWilliam A, Aitkenhead A, Mackay R I and Rowbottom C G 2016 The suitability of common metrics for assessing parotid and larynx autosegmentation accuracy *J. Appl. Clin. Med. Phys.* **17** 41–9
- Belon M R, Liu C, Elshaikh M A and Chetty I J 2015 Uncertainties in contour propagation for cervical brachytherapy due to the selection of deformable image registration algorithm *Int. J. Radiat. Oncol. Biol. Phys.* **93** E619–20
- Bender E T, Hardcastle N and Tomé W A 2012 On the dosimetric effect and reduction of inverse consistency and transitivity errors in deformable image registration for dose accumulation *Med. Phys.* **39** 272–80
- Bender E T and Tomé W A 2009 The utilization of consistency metrics for error analysis in deformable image registration *Phys. Med. Biol.* **54** 5561–77
- Boman E, Kapanen M, Pickup L and Lahtela S-L 2017 Importance of deformable image registration and biological dose summation in planning of radiotherapy retreatments *Med. Dosim.* **42** 296–303
- Bookstein F L 1989 Principal warps: thin-plate splines and the decomposition of deformations *IEEE Trans. Pattern Anal. Mach. Intell.* **11** 567–85
- Bosch W R, Straube W L, Matthews J W and Purdy J A 2015 Data from head-Neck_Cetuximab *The Cancer Imaging Archive*
- Boulangier M, Nunes J-C, Chourak H, Largent A, Tahri S, Acosta O, De Crevoisier R, Lafond C and Barateau A 2021 Deep learning methods to generate synthetic CT from MRI in radiotherapy: a literature review *Phys. Med.* **89** 265–81
- Boveiri H R, Khayami R, Javidan R and Mehdizadeh A 2020 Medical image registration using deep neural networks: a comprehensive review *Comput. Electr. Eng.* **87** 106767
- Boyd R, Basavatia A and Tomé W A 2021 Validation of accuracy deformable image registration contour propagation using a benchmark virtual HN phantom dataset *J. Appl. Clin. Med. Phys.* **22** 58–68
- Brock K K, Mutic S, McNutt T R, Li H and Kessler M L 2017 Use of image registration and fusion algorithms and techniques in radiotherapy: report of the AAPM Radiation Therapy Committee Task Group No. 132 *Med. Phys.* **44** e43–76

- Brooks E D, Wang X, De B, Verma V, Williamson T D, Hunter R, Mohamed A S R, Ning M S, Zhang X and Chang J Y 2022 An algorithm for thoracic re-irradiation using biologically effective dose: a common language on how to treat in a 'no-treat zone' *Radiat. Oncol.* **17** 4
- Byrne M et al 2021 Varian ethos online adaptive radiotherapy for prostate cancer: early results of contouring accuracy, treatment plan quality, and treatment time *J. Appl. Clin. Med. Phys.* **23** e13479
- Caldwell C B, Mah K, Ung Y C, Danjoux C E, Balogh J M, Ganguli S N and Ehrlich L E 2001 Observer variation in contouring gross tumor volume in patients with poorly defined non-small-cell lung tumors on CT: the impact of 18FDG-hybrid PET fusion *Int. J. Radiat. Oncol.* **51** 923–31
- Cao X, Yang J, Wang L, Xue Z, Wang Q and Shen D 2018 Deep learning based inter-modality image registration supervised by intra-modality similarity *Machine Learning in Medical Imaging Lecture Notes in Computer Science* ed Y Shi, H-I Suk and M Liu (Cham: Springer International Publishing) pp 55–63
- Cardenas C E, Yang J, Anderson B M, Court L E and Brock K B 2019 Advances in auto-segmentation *Semin. Radiat. Oncol.* **29** 185–97
- Castillo R, Castillo E, Guerra R, Johnson V E, McPhail T, Garg A K and Guerrero T 2009 A framework for evaluation of deformable image registration spatial accuracy using large landmark point sets *Phys. Med. Biol.* **54** 1849
- Chang J S et al 2018 Mapping patterns of locoregional recurrence following contemporary treatment with radiation therapy for breast cancer: a multi-institutional validation study of the ESTRO consensus guideline on clinical target volume *Radiother. Oncol. J. Eur. Soc. Ther. Radiol. Oncol.* **126** 139–47
- Chen J, Frey E C, He Y, Segars W P, Li Y and Du Y 2022 TransMorph: transformer for unsupervised medical image registration *Med. Image Anal.* **82** 102615
- Chen X, Diaz-Pinto A, Ravikumar N and Frangi A F 2021 Deep learning in medical image registration *Prog. Biomed. Eng.* **3** 012003
- Chetty I J and Rosu-Bubulac M 2019 Deformable registration for dose accumulation *Semin. Radiat. Oncol.* **29** 198–208
- Choi H J et al 2011 Inter- and intra-observer variability in contouring of the prostate gland on planning computed tomography and cone beam computed tomography *Acta Oncol.* **50** 539–46
- Christiansen R L, Dysager L, Bertelsen A S, Hansen O, Brink C and Bernchou U 2020 Accuracy of automatic deformable structure propagation for high-field MRI guided prostate radiotherapy *Radiat. Oncol.* **15** 32
- Christiansen R L, Johansen J, Zukauskaitė R, Hansen C R, Bertelsen A S, Hansen O, Mahmood F, Brink C and Bernchou U 2021 Accuracy of automatic structure propagation for daily magnetic resonance image-guided head and neck radiotherapy *Acta Oncologica* **60** 589–97
- Chung M K, Worsley K J, Paus T, Cherif C, Collins D L, Giedd J N, Rapoport J L and Evans A C 2001 A unified statistical approach to deformation-based morphometry *NeuroImage* **14** 595–606
- Clark K et al 2013 The cancer imaging archive (TCIA): maintaining and operating a public information repository *J. Digit. Imaging* **26** 1045–57
- Constable R T and Henkelman R M 1991 Contrast, resolution, and detectability in MR imaging *J. Comput. Assist. Tomogr.* **15** 297–303
- Covele B M, Carroll C J and Moore K L 2021 A practical method to quantify knowledge-based DVH prediction accuracy and uncertainty with reference cohorts *J. Appl. Clin. Med. Phys.* **22** 279–84
- Datterli R D and Dawant B M 2012 Estimation and reduction of target registration error *Med. Image Comput. Comput.-Assist. Interv. MICCAI Int. Conf. Med. Image Comput. Comput.-Assist. Interv.* **15** 139–46
- De Ruysscher D, Faivre-Finn C, Le Pechoux C, Peeters S and Belderbos J 2014 High-dose re-irradiation following radical radiotherapy for non-small-cell lung cancer *Lancet Oncol.* **15** e620–4
- Deshpande V S and Bhatt J S 2019 Bayesian deep learning for deformable medical image registration *Pattern Recognition and Machine Intelligence Lecture Notes in Computer Science* ed B Deka et al (Cham: Springer International Publishing) pp 41–9
- Dice L R 1945 Measures of the amount of ecologic association between species *Ecology* **26** 297–302
- Dowling J A and O'Connor L M 2020 Deformable image registration in radiation therapy *J. Med. Radiat. Sci.* **67** 257–9
- Dréan G et al 2016 Identification of a rectal subregion highly predictive of rectal bleeding in prostate cancer IMRT *Radiother. Oncol.* **119** 388–97
- Eiben B, Tran E H, Menten M J, Oelfke U, Hawkes D J and McClelland J R 2018 Statistical motion mask and sliding registration *Lect. Notes Comput. Sci. Subser. Lect. Notes Artif. Intell. Lect. Notes Bioinforma.* **10883** 13–23 Online
- Elmahdy M S et al 2019 Robust contour propagation using deep learning and image registration for online adaptive proton therapy of prostate cancer *Med. Phys.* **46** 3329–43
- Embring A, Onjukka E, Mercke C, Lax I, Berglund A, Boredal S, Wennberg B, Dalqvist E and Friesland S 2021 Re-irradiation for head and neck cancer: cumulative dose to organs at risk and late side effects *Cancers* **13** 3173
- Engwall E, Fredriksson A and Glimelius L 2018 4D robust optimization including uncertainties in time structures can reduce the interplay effect in proton pencil beam scanning radiation therapy *Med. Phys.* **45** 4020–9
- Farina E, Ferioli M, Castellucci P, Farina A, Rambaldi G Z, Cilla S, Cammelli S, Fanti S and Morganti A G 2017 18F-Fdg-PET-guided planning and re-planning (Adaptive) radiotherapy in head and neck cancer: current state of art *Anticancer Res.* **37** 6523–32
- Fatya M, Dogan N, Weiss E, Sleeman W C, Zhang B, Lehman W J, Williamson J F, Wijesooriya K and Christensen G E 2015 A Voxel-by-voxel comparison of deformable vector fields obtained by three deformable image registration algorithms applied to 4DCT lung studies *Front. Oncol.* **5** 17
- Ferrante E, Oktay O, Glocker B and Milone D H 2018 On the adaptability of unsupervised CNN-based deformable image registration to unseen image domains *Machine Learning in Medical Imaging Lecture Notes in Computer Science* ed Y Shi, H-I Suk and M Liu (Cham: Springer International Publishing) pp 294–302
- Fitzpatrick J M, West J B and Maurer C R 1998 Predicting error in rigid-body point-based registration *IEEE Trans. Med. Imaging* **17** 694–702
- Fontenele R C, Nascimento E H, Vasconcelos T V, Noujeim M and Freitas D Q 2018 Magnitude of cone beam CT image artifacts related to zirconium and titanium implants: impact on image quality *Dentomaxillofacial Radiol.* **47** 20180021
- Forsberg D, Andersson M and Knutsson H 2012 Extending image registration using polynomial expansion to diffeomorphic deformations Fukumitsu N et al 2017 Registration error of the liver CT using deformable image registration of MIM Maestro and Velocity AI *BMC Med. Imaging* **17** 30
- Fusella M, Loi G, Fiandra C and Lanzi E 2016 Impact of image quality on deformable image registration performances in pelvis patients *Phys. Medica Eur. J. Med. Phys.* **32** 24
- Gaede S, Olsthoorn J, Louie A V, Palma D, Yu E, Yaremko B, Ahmad B, Chen J, Bzdusek K and Rodrigues G 2011 An evaluation of an automated 4D-CT contour propagation tool to define an internal gross tumour volume for lung cancer radiotherapy *Radiother. Oncol.* **101** 322–8
- Galib S M, Lee H K, Guy C L, Riblett M J and Hugo G D 2020 A fast and scalable method for quality assurance of deformable image registration on lung CT scans using convolutional neural networks *Med. Phys.* **47** 99–109

- García-Alvarez J A, Zhong H, Schultz C J, Li X A and Kainz K 2022 Incorporating uncertainty bounds in daily deformable dose accumulation for adaptive radiation therapy of head-and-neck cancer *Med. Phys.*
- Giacometti V, King R B, Agnew C E, Irvine D M, Jain S, Hounsell A R and McGarry C K 2019 An evaluation of techniques for dose calculation on cone beam computed tomography *Br. J. Radiol.* **92** 20180383
- Giganti F, Kasivisvanathan V, Kirkham A, Punwani S, Emberton M, Moore C M and Allen C 2022 Prostate MRI quality: a critical review of the last 5 years and the role of the PI-QUAL score *Br. J. Radiol.* **95** 20210415
- Glide-Hurst C K et al 2021 Adaptive radiation therapy (ART) strategies and technical considerations: a state of the ART review from NRG Oncology *Int. J. Radiat. Oncol. Biol. Phys.* **109** 1054–75
- Gong X, Khaïdem L, Zhu W, Zhang B and Doermann D 2022 Uncertainty learning towards unsupervised deformable medical image registration 2022 *IEEE/CVF Winter Conference on Applications of Computer Vision (WACV) 2022 IEEE/CVF Winter Conference on Applications of Computer Vision (WACV) (Waikoloa, HI, USA: IEEE)* pp 1555–64 Online <https://ieeexplore.ieee.org/document/9707086/>
- Goodfellow I, Pouget-Abadie J, Mirza M, Xu B, Warde-Farley D, Ozair S, Courville A and Bengio Y 2014 Generative adversarial nets *Advances in Neural Information Processing Systems* vol 27 (Curran Associates, Inc.) Online https://proceedings.neurips.cc/paper_files/paper/2014/hash/5ca3e9b122f61f8f06494c97b1afcf3-Abstract.html
- Gooding M J, Boukerroui D, Vasquez Osorio E, Monshouwer R and Brunenberg E 2022 Multicenter comparison of measures for quantitative evaluation of contouring in radiotherapy *Phys. Imaging Radiat. Oncol.* **24** 152–8
- Gooding M J, Chu K, Conibear J, Dilling T, Durrant L, Fuss M, Gujral D, Ove R, Stevens C and Kadir T 2013 Multicenter clinical assessment of DIR atlas-based autocontouring *Int. J. Radiat. Oncol. Biol. Phys.* **87** S714–5
- Graeff C, Lüchtenborg R, Eley J G, Durante M and Bert C 2013 A 4D-optimization concept for scanned ion beam therapy *Radiother. Oncol.* **109** 419–24
- Graves Y J, Smith A-A, McIlvena D, Manilay Z, Lai Y K, Rice R, Mell L, Jia X, Jiang S B and Cerviño L 2015 A deformable head and neck phantom with in-vivo dosimetry for adaptive radiotherapy quality assurance *Med. Phys.* **42** 1490–7
- Green A, Vasquez Osorio E, Aznar M C, McWilliam A and van Herk M 2020 Image based data mining using per-voxel cox regression *Front. Oncol.* **10** 1178
- Greffier J, Frandon J, Larbi A, Beregi J P and Pereira F 2020 CT iterative reconstruction algorithms: a task-based image quality assessment *Eur. Radiol.* **30** 487–500
- Grigorescu I, Uus A, Christiaens D, Cordero-Grande L, Hutter J, Batalle D, Edwards A D, Hajnal J V, Modat M and Deprez M 2021 Uncertainty-aware deep learning based deformable registration *Uncertainty for Safe Utilization of Machine Learning in Medical Imaging, and Perinatal Imaging, Placental and Preterm Image Analysis Lecture Notes in Computer Science* vol 12959 ed C H Sudre, R Licandro, C Baumgartner, A Melbourne, A Dalca, J Hutter, R Tanno, E Abaci Turk, K Van Leemput, J Torrents Barrena, W M Wells and C Macgowan (Cham: Springer International Publishing) pp 54–63 Online https://link.springer.com/10.1007/978-3-030-87735-4_6
- Guy C L, Weiss E, Che S, Jan N, Zhao S and Rosu-Bubulac M 2019 Evaluation of image registration accuracy for tumor and organs at risk in the thorax for compliance With TG 132 recommendations *Adv. Radiat. Oncol.* **4** 177–85
- Hall W A et al 2018 Pancreatic gross tumor volume contouring on computed tomography (CT) compared with magnetic resonance imaging (MRI): results of an international contouring conference *Pract. Radiat. Oncol.* **8** 107–15
- Hammers J E, Pirozzi S, Lindsay D, Kaidar-Person O, Tan X, Chen R C, Das S K and Mavroidis P 2020 Evaluation of a commercial DIR platform for contour propagation in prostate cancer patients treated with IMRT/VMAT *J. Appl. Clin. Med. Phys.* **21** 14–25
- Han M C et al 2022 Performance evaluation of deformable image registration algorithms using computed tomography of multiple lung metastases *Technol. Cancer Res. Treat.* **21** 15330338221078464
- Hardcastle N et al 2012 A multi-institution evaluation of deformable image registration algorithms for automatic organ delineation in adaptive head and neck radiotherapy *Radiat. Oncol.* **7** 90
- Hardcastle N, van Elmp W, De Ruyscher D, Bzdusek K and Tomé W A 2013 Accuracy of deformable image registration for contour propagation in adaptive lung radiotherapy *Radiat. Oncol. Lond. Engl.* **8** 243
- Harrison K, Pullen H, Welsh C, Oktay O, Alvarez-Valle J and Jena R 2022 Machine learning for auto-segmentation in radiotherapy planning *Clin. Oncol.* **34** 74–88
- Haskins G, Kruger U and Yan P 2020 Deep learning in medical image registration: a survey *Mach. Vis. Appl.* **31** 8
- Hausdorff F 1920 Hausdorffs Grundzuege der Mengenlehre *Bull. Amer. Math. Soc.* **27** 116–29
- Held M, Cremers F, Sneed P K, Braunstein S, Fogh S E, Nakamura J, Barani I, Perez-Andujar A, Pouliot J and Morin O 2016 Assessment of image quality and dose calculation accuracy on kV CBCT, MV CBCT, and MV CT images for urgent palliative radiotherapy treatments *J. Appl. Clin. Med. Phys.* **17** 279–90
- Hill D L G, Batchelor P G, Holden M and Hawkes D J 2001 Medical image registration *Phys. Med. Biol.* **46** R1–45
- Hoffmann A et al 2020 MR-guided proton therapy: a review and a preview *Radiat. Oncol.* **15** 129
- Holden M 2008 A review of geometric transformations for nonrigid body registration *IEEE Trans. Med. Imaging* **27** 111–28
- Horn B K P and Schunck B G 1981 Determining optical flow *Artif. Intell.* **17** 185–203
- Hu Y et al 2018 Weakly-supervised convolutional neural networks for multimodal image registration *Med. Image Anal.* **49** 1–13
- Hub M, Thieke C, Kessler M L and Karger C P 2012 A stochastic approach to estimate the uncertainty of dose mapping caused by uncertainties in b-spline registration *Med. Phys.* **39** 2186–92
- Huesa-Berral C, Juan-Cruz C, van Kranen S, Rossi M, Belderbos J S A, Azcona J-D, Burguete J and Sonke J-J 2022 Detailed dosimetric evaluation of inter-fraction and respiratory motion in lung stereotactic body radiation therapy based on daily 4D cone beam CT images *Phys. Med. Biol.* Online <http://iopscience.iop.org/article/10.1088/1361-6560/aca94d>
- Hugo G D, Weiss E, Sleeman W C, Balik S, Keall P J, Lu J and Williamson J F 2017 A longitudinal four-dimensional computed tomography and cone beam computed tomography dataset for image-guided radiation therapy research in lung cancer *Med. Phys.* **44** 762–71
- Hussein M, Akintonde A, McClelland J, Speight R and Clark C H 2021 Clinical use, challenges, and barriers to implementation of deformable image registration in radiotherapy—the need for guidance and QA tools *Br. J. Radiol.* **94** 20210001
- Huttenlocher D P, Klanderman G A and Rucklidge W J 1993 Comparing images using the Hausdorff distance *IEEE Trans. Pattern Anal. Mach. Intell.* **15** 850–63
- Hvid C A, Elström U V, Jensen K, Alber M and Grau C 2016 Accuracy of software-assisted contour propagation from planning CT to cone beam CT in head and neck radiotherapy *Acta Oncol.* **55** 1324–30
- Irmak S, Georg D and Lechner W 2020 Comparison of CBCT conversion methods for dose calculation in the head and neck region *Z. Med. Phys.* **30** 289–99

- Jaderberg M, Simonyan K, Zisserman A and Kavukcuoglu K 2015 *Spatial Transformer Networks Advances in Neural Information Processing Systems* vol 28 (Curran Associates, Inc.) Online https://proceedings.neurips.c.c./paper_files/paper/2015/hash/33ceb07bf4eeb3da587e268d663aba1a-Abstract.html
- Jamema S V, Phurailatpam R, Paul S N, Joshi K and Deshpande D D 2018 Commissioning and validation of commercial deformable image registration software for adaptive contouring *Phys. Med.* **47** 1–8
- Jarema T and Aland T 2019 Using the iterative kV CBCT reconstruction on the Varian Halcyon linear accelerator for radiation therapy planning for pelvis patients *Phys. Medica PM Int. J. Devoted Appl. Phys. Med. Biol. Off. J. Ital. Assoc. Biomed. Phys. AIFB* **68** 112–6
- Jena R, Kirkby N F, Burton K E, Hoole A C F, Tan L T and Burnet N G 2010 A novel algorithm for the morphometric assessment of radiotherapy treatment planning volumes *Br. J. Radiol.* **83** 44–51
- Jian B et al 2022 Weakly-supervised Biomechanically-constrained CT/MRI Registration of the Spine Online <http://arxiv.org/abs/2205.07568>
- Johnson H J, McCormick M M, Ibáñez L and The Insight Software Consortium 2019 *The ITK Software Guide* Online <https://itk.org/ITKSoftwareGuide/html/>
- Kadoya N et al 2014 Evaluation of various deformable image registration algorithms for thoracic images *J. Radiat. Res. (Tokyo)* **55** 175–82
- Kadoya N et al 2016 Multi-institutional validation study of commercially available deformable image registration software for thoracic images *Int. J. Radiat. Oncol. Biol. Phys.* **96** 422–31
- Kadoya N et al 2021 Development of a physical geometric phantom for deformable image registration credentialing of radiotherapy centers for a clinical trial *J. Appl. Clin. Med. Phys.* **22** 255–65
- Kamal M et al 2020 Patterns of failure after intensity modulated radiation therapy in head and neck squamous cell carcinoma of unknown primary: implication of elective nodal and mucosal dose coverage *Adv. Radiat. Oncol.* **5** 929–35
- Kamath A, Poel R, Willmann J, Andratschke N and Reyes M 2023 How sensitive are deep learning based radiotherapy dose prediction models to variability in organs at risk segmentation? *2023 IEEE 20th International Symposium on Biomedical Imaging (ISBI) 2023 IEEE 20th International Symposium on Biomedical Imaging (ISBI)* (Cartagena, Colombia: IEEE) pp 1–4 Online <https://ieeexplore.ieee.org/document/10230559/>
- Kashani R et al 2008 Objective assessment of deformable image registration in radiotherapy: a multi-institution study *Med. Phys.* **35** 5944–53
- Khawaled S and Freiman M 2020 Unsupervised deep-learning based deformable image registration: a bayesian framework *ArXiv Comput. Vis. Pattern Recognit.* Online <https://semanticscholar.org/paper/Unsupervised-Deep-Learning-Based-Deformable-Image-A-Khawaled-Freiman/c6e859a8e8a0849c7759fc5edef938b38b345fcd>
- Khawaled S and Freiman M 2022a NPBDREG: uncertainty assessment in diffeomorphic brain MRI registration using a non-parametric Bayesian deep-learning based approach *Comput. Med. Imaging Graph.* **99** 102087
- Khawaled S and Freiman M 2022b NPBDREG: uncertainty assessment in diffeomorphic brain MRI registration using a non-parametric Bayesian deep-learning based approach *Comput. Med. Imaging Graph.* **99** 102087
- Kierkels R G J, Otter L A den, Korevaar E W, Langendijk J A, Schaaf A van der, Knopf A C and Sijtsma N M 2018 An automated, quantitative, and case-specific evaluation of deformable image registration in computed tomography images *Phys. Med. Ampmathsemicolon Biol.* **63** 045026
- Kim B, Kim J, Lee J-G, Kim D H, Park S H and Ye J C 2019 Unsupervised deformable image registration using cycle-consistent CNN *Medical Image Computing and Computer Assisted Intervention—MICCAI 2019 Lecture Notes in Computer Science* ed D Shen, T Liu, T M Peters, L H Staib, C Essert, S Zhou, P-T Yap and A Khan (Cham: Springer International Publishing) pp 166–74
- Kirby N, Chen J, Kim H, Morin O, Nie K and Pouliot J 2016 An automated deformable image registration evaluation of confidence tool *Phys. Med. Biol.* **61** N203–14
- Kiser K et al 2019 Prospective quantitative quality assurance and deformation estimation of MRI-CT image registration in simulation of head and neck radiotherapy patients *Clin. Transl. Radiat. Oncol.* **18** 120–7
- Kraus K M, Jäkel O, Niebuhr N I and Pfaffenberger A 2017 Generation of synthetic CT data using patient specific daily MR image data and image registration *Phys. Med. Biol.* **62** 1358–77
- Krebs J, Mansi T, Mailhé B, Ayache N and Delingette H 2018 Unsupervised Probabilistic deformation modeling for robust diffeomorphic registration *Deep Learning in Medical Image Analysis and Multimodal Learning for Clinical Decision Support Lecture Notes in Computer Science* ed D Stoyanov, Z Taylor, G Carneiro, T Syeda-Mahmood, A Martel, L Maier-Hein, J M R S Tavares, A Bradley, J P Papa, V Belagiannis, J C Nascimento, Z Lu, S Conjeti, M Moradi, H Greenspan and A Madabhushi (Cham: Springer International Publishing) pp 101–9
- Kuang D and Schmah T 2019 FAIM—A ConvNet Method for Unsupervised 3D Medical Image Registration *Machine Learning in Medical Imaging Lecture Notes in Computer Science* ed H-I Suk, M Liu, P Yan and C Lian (Cham: Springer International Publishing) pp 646–54
- Kubli A, Pukala J, Shah A P, Kelly P, Langen K M, Bova F J, Mañon R R and Meeks S L 2021 Variability in commercially available deformable image registration: a multi-institution analysis using virtual head and neck phantoms *J. Appl. Clin. Med. Phys.* **22** 89–96
- Kurz C et al 2016 Investigating deformable image registration and scatter correction for CBCT-based dose calculation in adaptive IMPT *Med. Phys.* **43** 5635–46
- Lalonde A, Winey B, Verburg J, Paganetti H and Sharp G C 2020 Evaluation of CBCT scatter correction using deep convolutional neural networks for head and neck adaptive proton therapy *Phys. Med. Ampmathsemicolon Biol.* **65** 245022
- Latifi K, Caudell J, Zhang G, Hunt D, Moros E G and Feygelman V 2018 Practical quantification of image registration accuracy following the AAPM TG-132 report framework *J. Appl. Clin. Med. Phys.* **19** 125–33
- Lee E M, Ibrahim E-S H, Dudek N, Lu J C, Kalia V, Runge M, Srinivasan A, Stojanovska J and Agarwal P P 2021 Improving MR image quality in patients with metallic implants *RadioGraphics* **41** E126–37
- Lee S S, Choi S H, Kim K B and Paik E K 2020 Validation of deformable image registration by using a B-spline and optical-flow algorithm in head and neck cancer cases *J. Korean Phys. Soc.* **76** 194–201
- Li X, Zhang Y, Shi Y, Wu S, Xiao Y, Gu X, Zhen X and Zhou L 2017 Comprehensive evaluation of ten deformable image registration algorithms for contour propagation between CT and cone-beam CT images in adaptive head & neck radiotherapy *PLoS One* **12** e0175906
- Li Y, Kubota Y, Okamoto M, Shiba S, Okazaki S, Matsui T, Komatsu S and Ohno T 2021 Determination of deformable image registration algorithms for accumulating dose in carbon-ion radiotherapy for pancreatic cancer *Anticancer Res.* **41** 835–43
- Liang X, Morgan H, Bai T, Dohopolski M, Nguyen D and Jiang S 2023 Deep learning based direct segmentation assisted by deformable image registration for cone-beam CT based auto-segmentation for adaptive radiotherapy *Phys. Med. Biol.* **68** 045012
- Lin D, Lapen K, Sherer M V, Kantor J, Zhang Z, Boyce L M, Bosch W, Korenstein D and Gillespie E F 2020 A systematic review of contouring guidelines in radiation oncology: analysis of frequency, methodology, and delivery of consensus recommendations *Int. J. Radiat. Oncol. Biol. Phys.* **107** 827–35

- Liu X, Jiang D, Wang M and Song Z 2019 Image synthesis-based multi-modal image registration framework by using deep fully convolutional networks *Med. Biol. Eng. Comput.* **57** 1037–48
- Liu Y, Jin R, Chen M, Song E, Xu X, Zhang S and Hung C-C 2016 Contour propagation using non-uniform cubic B-splines for lung tumor delineation in 4D-CT *Int. J. Comput. Assist. Radiol. Surg.* **11** 2139–51
- Loi G et al 2018 Performance of commercially available deformable image registration platforms for contour propagation using patient-based computational phantoms: a multi-institutional study *Med. Phys.* **45** 748–57
- Loi G et al 2020 Computed tomography to cone beam computed tomography deformable image registration for contour propagation using head and neck, patient-based computational phantoms: a multicenter study *Pract. Radiat. Oncol.* **10** 125–32
- Lowther N, Louwe R, Yuen J, Hardcastle N, Yeo A, Jameson M and the Medical Image and Registration Special Interest Group (MIRSI) of the ACPSEM 2022 MIRSI position paper: the use of image registration and fusion algorithms in radiotherapy *Phys. Eng. Sci. Med.* **45** 421–8
- Lowther N J, Marsh S H and Louwe R J W 2020a Quantifying the dose accumulation uncertainty after deformable image registration in head-and-neck radiotherapy *Radiother. Oncol.* **143** 117–25
- Lowther N J, Marsh S H and Louwe R J W 2020b Dose accumulation to assess the validity of treatment plans with reduced margins in radiotherapy of head and neck cancer *Physics and Imaging in Radiation Oncology* **14** 53–60
- Ma Y, Liu X, Dai Z, He P, Yan Y, Shen G, Yuan P and Li Q 2017 Evaluation of mesh- and binary-based contour propagation methods in 4D thoracic radiotherapy treatments using patient 4D CT images *Phys. Med.* **36** 46–53
- Machiels M et al 2019 Reduced inter-observer and intra-observer delineation variation in esophageal cancer radiotherapy by use of fiducial markers *Acta Oncol.* **58** 943–50
- Magallon-Baro A, Milder M T W, Granton P V, den Toom W, Nuytens J J and Hoogeman M S 2022 Impact of using unedited CT-Based DIR-propagated autocontours on online ART for pancreatic SBRT *Front. Oncol.* **12** 910792
- Mahapatra D, Antony B, Sedai S and Garnavi R 2018 Deformable medical image registration using generative adversarial networks 2018 *IEEE 15th International Symposium on Biomedical Imaging (ISBI 2018) 2018 IEEE 15th International Symposium on Biomedical Imaging (ISBI) 2018* pp 1449–53
- Mahon R N, Jan N, Hugo G D, Muscu S and Weiss E 2020 Accuracy of deformable image registration for reirradiation of lung cancer following stereotactic body radiotherapy *Int. J. Radiat. Oncol. Biol. Phys.* **108** e304–5
- Maintz J B and Viergever M A 1998 A survey of medical image registration *Med. Image Anal.* **2** 1–36
- Mantel F, Flentje M and Guckenberger M 2013 Stereotactic body radiation therapy in the re-irradiation situation—a review *Radiat. Oncol.* **8** 7
- McKenzie E M, Santhanam A, Ruan D, O'Connor D, Cao M and Sheng K 2020 Multimodality image registration in the head-and-neck using a deep learning-derived synthetic CT as a bridge *Med. Phys.* **47** 1094–104
- McVicar N, Thomas S, Liu M, Carolan H and Bergman A 2018 Re-irradiation volumetric modulated arc therapy optimization based on cumulative biologically effective dose objectives *J. Appl. Clin. Med. Phys.* **19** 341–5
- McWilliam A et al 2023 Voxel-based analysis: roadmap for clinical translation *Radiother. Oncol. J. Eur. Soc. Ther. Radiol. Oncol.* **188** 109868
- McWilliam A, Kennedy J, Hodgson C, Vasquez Osorio E, Faivre-Finn C and van Herk M 2017 Radiation dose to heart base linked with poorer survival in lung cancer patients *Eur. J. Cancer Oxf. Engl.* **1990** 85 106–13
- Meijers A, Knopf A-C, Crijs A P G, Ubbels J F, Niezink A G H, Langendijk J A, Wijsman R and Both S 2020 Evaluation of interplay and organ motion effects by means of 4D dose reconstruction and accumulation *Radiother. Oncol.* **150** 268–74
- Meijneke T R, Petit S F, Wentzler D, Hoogeman M and Nuytens J J 2013 Reirradiation and stereotactic radiotherapy for tumors in the lung: dose summation and toxicity *Radiother. Oncol.* **107** 423–7
- Mencarelli A, van Kranen S R, Hamming-Vrieze O, van Beek S, Nico Rasch C R, van Herk M and Sonke J-J 2014 Deformable image registration for adaptive radiation therapy of head and neck cancer: accuracy and precision in the presence of tumor changes *Int. J. Radiat. Oncol.* **90** 680–7
- Miura H, Ozawa S, Nakao M, Furukawa K, Doi Y, Kawabata H, Kenjou M and Nagata Y 2017 Impact of deformable image registration accuracy on thoracic images with different regularization weight parameter settings *Phys. Med.* **42** 108–11
- Mogadas N, Sothmann T, Knopp T, Gauer T, Petersen C and Werner R 2018 Influence of deformable image registration on 4D dose simulation for extracranial SBRT: a multi-registration framework study *Radiother. Oncol.* **127** 225–32
- Monti S, Palma G, D'Avino V, Gerardi M, Marvaso G, Ciardo D, Pacelli R, Jereczek-Fossa B A, Alterio D and Cella L 2017 Voxel-based analysis unveils regional dose differences associated with radiation-induced morbidity in head and neck cancer patients *Sci. Rep.* **7** 7220
- Motegi K, Tachibana H, Motegi A, Hotta K, Baba H and Akimoto T 2019 Usefulness of hybrid deformable image registration algorithms in prostate radiation therapy *J. Appl. Clin. Med. Phys.* **20** 229–36
- Močnik D, Ibragimov B, Xing L, Strojjan P, Likar B, Pernuš F and Vrtovec T 2018 Segmentation of parotid glands from registered CT and MR images *Phys. Medica PM Int. J. Devoted Appl. Phys. Med. Biol. Off. J. Ital. Assoc. Biomed. Phys. AIFB* **52** 33–41
- Murphy M J, Salguero F J, Siebers J V, Staub D and Vaman C 2012 A method to estimate the effect of deformable image registration uncertainties on daily dose mapping *Med. Phys.* **39** 573–80
- Murr M, Brock K K, Fusella M, Hardcastle N, Hussein M, Jameson M G, Wahlstedt I, Yuen J, McClelland J R and Vasquez Osorio E 2023 Applicability and usage of dose mapping/accumulation in radiotherapy *Radiother. Oncol.* 109527
- Nash D, Juneja S, Palmer A L, van Herk M, McWilliam A and Osorio E V 2022 The geometric and dosimetric effect of algorithm choice on propagated contours from CT to cone beam CTs *Phys. Med.* **100** 112–9
- Nassef M, Simon A, Cazoulat G, Duménil A, Blay C, Lafond C, Acosta O, Balosso J, Haigron P and de Crevoisier R 2016 Quantification of dose uncertainties in cumulated dose estimation compared to planned dose in prostate IMRT *Radiother. Oncol.* **119** 129–36
- Nehmeh S A and Erdi Y E 2008 Respiratory motion in positron emission tomography/computed tomography: a review *Semin. Nucl. Med.* **38** 167–76
- Nenoff L et al 2020 Deformable image registration uncertainty for inter-fractional dose accumulation of lung cancer proton therapy *Radiother. Oncol.* **147** 178–85
- Nenoff L et al 2021a TCP and NTCP calculations based on treatment doses instead of planned doses for daily adaptive proton therapy of lung cancer *Int. J. Radiat. Oncol.* **111** e133
- Nenoff L et al 2021b Dosimetric influence of deformable image registration uncertainties on propagated structures for online daily adaptive proton therapy of lung cancer patients *Radiother. Oncol.* **159** 136–43
- Nenoff L, Buti G, Bobić M, Lalonde A, Nesteruk K P, Winey B, Sharp G C, Sudhyadhom A and Paganetti H 2022 Integrating structure propagation uncertainties in the optimization of online adaptive proton therapy plans *Cancers* **14** 3926

- Nesteruk K P, Bobić M, Sharp G C, Lalonde A, Winey B A, Nenoff L, Lomax A J and Paganetti H 2022 Low-dose computed tomography scanning protocols for online adaptive proton therapy of head-and-neck cancers *Cancers* **14** 5155
- Nie X, Saleh Z, Kadbi M, Zakian K, Deasy J, Rimmer A and Li G 2020 A super-resolution framework for the reconstruction of T2-weighted (T2w) time-resolved (TR) 4DMRI using T1w TR-4DMRI as the guidance *Med. Phys.* **47** 3091–102
- Niebuhr N I et al 2021 Biologically consistent dose accumulation using daily patient imaging *Radiat. Oncol.* **16** 65
- Niebuhr N I, Johnen W, Echner G, Runz A, Bach M, Stoll M, Giske K, Greilich S and Pfaffenberger A 2019 The ADAM-pelvis phantom—an anthropomorphic, deformable and multimodal phantom for MRgRT *Phys. Med. Biol.* **64** 04NT05
- Nieder C, Andratschke N H and Grosu A L 2013 Increasing frequency of reirradiation studies in radiation oncology: systematic review of highly cited articles *Am. J. Cancer Res.* **3** 152–8
- Nieder C, Andratschke N H and Grosu A L 2016 Re-irradiation for recurrent primary brain tumors *Anticancer Res.* **36** 4985–96
- Niemierko A 1997 Reporting and analyzing dose distributions: a concept of equivalent uniform dose *Med. Phys.* **24** 103–10
- Nithiananthan S, Schafer S, Mirota D J, Stayman J W, Zbijewski W, Reh D D, Gallia G L and Siewerdsen J H 2012 Extra-dimensional Demons: a method for incorporating missing tissue in deformable image registration *Med. Phys.* **39** 5718–31
- Nix M et al 2022 Dose summation and image registration strategies for radiobiologically and anatomically corrected dose accumulation in pelvic re-irradiation *Acta Oncol.* **61** 64–72
- Nix M G, Prestwich R J D and Speight R 2017 Automated, reference-free local error assessment of multimodal deformable image registration for radiotherapy in the head and neck *Radiother. Oncol.* **125** 478–84
- Oh S and Kim S 2017 Deformable image registration in radiation therapy *Radiat. Oncol. J.* **35** 101–11
- Owringi A M, Greer P B and Glide-Hurst C K 2018 MRI-only treatment planning: benefits and challenges *Phys. Med. Biol.* **63** 05TR01
- Paganelli C, Meschini G, Molinelli S, Riboldi M and Baroni G 2018 Patient-specific validation of deformable image registration in radiation therapy: overview and caveats *Med. Phys.* **45** e908–22
- Palma G, Monti S and Cella L 2020 Voxel-based analysis in radiation oncology: a methodological cookbook *Phys. Medica PM Int. J. Devoted Appl. Phys. Med. Biol. Off. J. Ital. Assoc. Biomed. Phys. AIFB* **69** 192–204
- Palmér E, Karlsson A, Nordström F, Petruson K, Siverson C, Ljungberg M and Sohlin M 2021 Synthetic computed tomography data allows for accurate absorbed dose calculations in a magnetic resonance imaging only workflow for head and neck radiotherapy *Phys. Imaging Radiat. Oncol.* **17** 36–42
- Peroni M, Spadea M F, Riboldi M, Falcone S, Vaccaro C, Sharp G C and Baroni G 2013 Validation of automatic contour propagation for 4D treatment planning using multiple metrics *Technol. Cancer Res. Treat.* **12** 501–10
- Persson G F, Nygaard D E, Brink C, Jahn J W, Munck af Rosenschöld P, Specht L and Korreman S S 2010 Deviations in delineated GTV caused by artefacts in 4DCT *Radiother. Oncol.* **96** 61–6
- Pham J, Cao M, Yoon S M, Gao Y, Kishan A U and Yang Y 2022 Dosimetric effects of air cavities for MRI-guided online adaptive radiation therapy (MRgART) of prostate bed after radical prostatectomy *J. Clin. Med.* **11** 364
- Poel R, Rüfenacht E, Hermann E, Scheib S, Manser P, Aebersold D M and Reyes M 2021 The predictive value of segmentation metrics on dosimetry in organs at risk of the brain *Med. Image Anal.* **73** 102161
- Polan D F, Feng M, Lawrence T S, Ten Haken R K and Brock K K 2017 Implementing radiation dose-volume liver response in biomechanical deformable image registration *Int. J. Radiat. Oncol.* **99** 1004–12
- Pukala J, Johnson P B, Shah A P, Langen K M, Bova F J, Staton R J, Mañon R R, Kelly P and Meeks S L 2016 Benchmarking of five commercial deformable image registration algorithms for head and neck patients *J. Appl. Clin. Med. Phys.* **17** 25–40
- Qiao Y, Jagt T, Hoogeman M, Lelieveldt B P F and Staring M 2019 Evaluation of an open source registration package for automatic contour propagation in online adaptive intensity-modulated proton therapy of prostate cancer *Front. Oncol.* **9** Online <https://frontiersin.org/article/10.3389/fonc.2019.01297>
- Qin A, Ionascu D, Liang J, Han X, O'Connell N and Yan D 2018 The evaluation of a hybrid biomechanical deformable registration method on a multistage physical phantom with reproducible deformation *Radiat. Oncol.* **13** 240
- Ray X, Moazzezi M, Bojchko C and Moore K L 2020 Data-driven margin determination for online adaptive radiotherapy using batch automated planning *Int. J. Radiat. Oncol. Biol. Phys.* **108** e370
- Rey D, Subsol G, Delingette H and Ayache N 2002 Automatic detection and segmentation of evolving processes in 3D medical images: application to multiple sclerosis *Med. Image Anal.* **6** 163–79
- Ribeiro C O, Knopf A, Langendijk J A, Weber D C, Lomax A J and Zhang Y 2018 Assessment of dosimetric errors induced by deformable image registration methods in 4D pencil beam scanned proton treatment planning for liver tumours *Radiother. Oncol.* **128** 174–81
- Riegel A C, Antone J G, Zhang H, Jain P, Raince J, Rea A, Bergamo A M, Kapur A and Potters L 2016 Deformable image registration and interobserver variation in contour propagation for radiation therapy planning *J. Appl. Clin. Med. Phys.* **17** 347–57
- Rigaud B, Simon A, Castelli J, Lafond C, Acosta O, Haigron P, Cazoulat G and de Crevoisier R 2019 Deformable image registration for radiation therapy: principle, methods, applications and evaluation *Acta Oncol.* **58** 1225–37
- Ritter L, Mischkowski R A, Neugebauer J, Dreiseidler T, Scheer M, Keeve E and Zöller J E 2009 The influence of body mass index, age, implants, and dental restorations on image quality of cone beam computed tomography *Oral Surg. Oral Med. Oral Pathol. Oral Radiol. Endodontology* **108** e108–16
- Roach D et al 2019 Multi-observer contouring of male pelvic anatomy: highly variable agreement across conventional and emerging structures of interest *J. Med. Imaging Radiat. Oncol.* **63** 264–71
- Rohé M-M, Datar M, Heimann T, Sermesant M and Pennec X 2017 SVF-Net: learning deformable image registration using shape matching *Medical Image Computing and Computer Assisted Intervention – MICCAI 2017 Lecture Notes in Computer Science* ed M Descoteaux, L Maier-Hein, A Franz, P Jannin, D L Collins and S Duchesne (Cham: Springer International Publishing) pp 266–74
- Roman N O, Shepherd W, Mukhopadhyay N, Hugo G D and Weiss E 2012 Interfractional positional variability of fiducial markers and primary tumors in locally advanced non-small-cell lung cancer during audiovisual biofeedback radiotherapy *Int. J. Radiat. Oncol.* **83** 1566–72
- Rong Y et al 2021 Rigid and deformable image registration for radiation therapy: a self-study evaluation guide for NRG oncology clinical trial participation *Pract. Radiat. Oncol.* **11** 282–98
- Rosu M and Hugo G D 2012 Advances in 4D radiation therapy for managing respiration: II. 4D treatment planning *Z. Für Med. Phys.* **22** 272–80
- Saleh Z H et al 2014 The distance discordance metric—a novel approach to quantifying spatial uncertainties in intra- and inter-patient deformable image registration *Phys. Med. Biol.* **59** 733–46
- Saleh-Sayah N K, Weiss E, Salguero F J and Siebers J V 2011 A distance to dose difference tool for estimating the required spatial accuracy of a displacement vector field *Med. Phys.* **38** 2318–23

- Salguero F J, Saleh-Sayah N K, Yan C and Siebers J V 2011 Estimation of three-dimensional intrinsic dosimetric uncertainties resulting from using deformable image registration for dose mapping *Med. Phys.* **38** 343–53
- Sarrut D, Baudier T, Ayadi M, Tanguy R and Rit S 2017 Deformable image registration applied to lung SBRT: usefulness and limitations *Phys. Med.* **44** 108–12
- Sarudis S, Karlsson A, Bibac D, Nyman J and Bäck A 2019 Evaluation of deformable image registration accuracy for CT images of the thorax region *Phys. Med.* **57** 191–9
- Scaggion A, Fiandra C, Loi G, Vecchi C and Fusella M 2020a Free-to-use DIR solutions in radiotherapy: benchmark against commercial platforms through a contour-propagation study *Phys. Med.* **74** 110–7
- Scaggion A, Fiandra C, Loi G, Vecchi C and Fusella M 2020b Free-to-use DIR solutions in radiotherapy: benchmark against commercial platforms through a contour-propagation study *Phys. Medica PM Int. J. Devoted Appl. Phys. Med. Biol. Off. J. Ital. Assoc. Biomed. Phys. AIFB* **74** 110–7
- Schipsaanboord B, Boukerroui D, Peressutti D, van Soest J, Lustberg T, Kadir T, Dekker A, van Elmpt W and Gooding M 2019 Can atlas-based auto-segmentation ever be perfect? insights from extreme value theory *IEEE Trans. Med. Imaging* **38** 99–106
- Sen A et al 2020 Accuracy of deformable image registration techniques for alignment of longitudinal cholangiocarcinoma CT images *Med. Phys.* **47** 1670–9
- Serban M, Heath E, Stroian G, Collins D L and Seuntjens J 2008 A deformable phantom for 4D radiotherapy verification: design and image registration evaluation *Med. Phys.* **35** 1094–102
- Shah K, Shackelford J A, Kandasamy N and Sharp G C 2021 Improving deformable image registration accuracy using a hybrid similarity metric for adaptive radiation therapy *Medical Imaging 2021: Image Processing Image Processing* ed B A Landman and I Išgum Online Only (United States: SPIE) p 122 Online <https://spiedigitallibrary.org/conference-proceedings-of-spie/11596/2582164/Improving-deformable-image-registration-accuracy-using-a-hybrid-similarity-metric/10.1117/12.2582164.full>
- Shi L, Chen Q, Barley S, Cui Y, Shang L, Qiu J and Rong Y 2021 Benchmarking of deformable image registration for multiple anatomic sites using digital data sets with ground-truth deformation vector fields *Pract. Radiat. Oncol.* **11** 404–14
- Shortall J, Palma G, Mistry H, Osorio E V, McWilliam A, Choudhury A, Aznar M, Herk M van and Green A 2021 Flogging a dead salmon? Reduced dose posterior to prostate correlates with increased PSA progression in voxel-based analysis of 3 randomized phase 3 trials *Int. J. Radiat. Oncol. Biol. Phys.* **110** 696–9
- Simonovsky M, Gutiérrez-Becker B, Mateus D, Navab N and Komodakis N 2016 A deep metric for multimodal registration *Medical Image Computing and Computer-assisted Intervention—MICCAI 2016 Lecture Notes in Computer Science* ed S Ourselin, L Joskowicz, M R Sabuncu, G Unal and W Wells (Cham: Springer International Publishing) pp 10–8
- Skjøtskift T, Evensen M E, Furre T, Moan J M, Amdal C D, Bogsrud T V, Malinen E and Dale E 2018 Dose painting for re-irradiation of head and neck cancer *Acta Oncol.* **57** 1693–9
- Smolders A, Amstutz F, Zhang Y, Weber D C, Lomax T and Albertini F 2022a Fast deformable image registration uncertainty estimation for contour propagation in daily adaptive proton therapy *Medical Imaging with Deep Learning* Online <https://openreview.net/forum?id=6b60oHnnST4>
- Smolders A, Choulilitsa E, Czerska K, Bizzocchi N, Krcek R, Lomax A, Weber D C and Albertini F 2023a Dosimetric comparison of autocontouring techniques for online adaptive proton therapy *Phys. Med. Biol.* **68** 175006
- Smolders A, Hengeveld A C, Both S, Wijsman R, Langendijk J A, Weber D C, Lomax A J, Albertini F and Guterres Marmitt G 2023b Inter- and intrafractional 4D dose accumulation for evaluating Δ NTCP robustness in lung cancer *Radiother. Oncol.* **182** 109488
- Smolders A, Lomax A, Weber D C and Albertini F 2022b Deformable Image Registration uncertainty quantification using deep learning for dose accumulation in adaptive proton therapy 10th International Workshop on Biomedical Image Registration Online <https://openreview.net/forum?id=B0MxIXCh50Y>
- Smolders A J, Lomax A J, Weber D C and Albertini F 2023c Deep learning based uncertainty prediction of deformable image registration for contour propagation and dose accumulation in online adaptive radiotherapy *Phys. Med. Biol.* Online <https://iopscience.iop.org/article/10.1088/1361-6560/ad0282>
- Söhn M, Birkner M, Chi Y, Wang J, Yan D, Berger B and Alber M 2008 Model-independent, multimodality deformable image registration by local matching of anatomical features and minimization of elastic energy *Med. Phys.* **35** 866–78
- Sonke J-J, Aznar M and Rasch C 2019 Adaptive radiotherapy for anatomical changes *Semin. Radiat. Oncol.* **29** 245–57
- Sonke J-J and Belderbos J 2010 Adaptive radiotherapy for lung cancer *Semin. Radiat. Oncol.* **20** 94–106
- Sotiras A, Davatzikos C and Paragios N 2013 Deformable medical image registration: a survey *IEEE Trans. Med. Imaging* **32** 1153–90
- Spautz S, Haase L, Tschiche M, Makocki S, Richter C, Troost E G C and Stützer K 2023 Comparison of 3D and 4D robustly optimized proton treatment plans for non-small cell lung cancer patients with tumour motion amplitudes larger than 5 mm *Phys. Imaging Radiat. Oncol.* **27** 100465
- Spin-Neto R and Wenzel A 2016 Patient movement and motion artefacts in cone beam computed tomography of the dentomaxillofacial region: a systematic literature review *Oral Surg. Oral Med. Oral Pathol. Oral Radiol.* **121** 425–33
- Swamidas J, Kirisits C, De Brabandere M, Hellebust T P, Siebert F-A and Tanderup K 2020 Image registration, contour propagation and dose accumulation of external beam and brachytherapy in gynecological radiotherapy *Radiother. Oncol.* **143** 1–11
- Takayama Y et al 2017 Evaluation of the performance of deformable image registration between planning CT and CBCT images for the pelvic region: comparison between hybrid and intensity-based DIR *J. Radiat. Res. (Tokyo)* **58** 567–71
- Takemura A, Nagano A, Kojima H, Ikeda T, Yokoyama N, Tsukamoto K, Noto K, Isomura N, Ueda S and Kawashima H 2018 An uncertainty metric to evaluate deformation vector fields for dose accumulation in radiotherapy *Phys. Imaging Radiat. Oncol.* **6** 77–82
- Tascón-Vidarte J D, Stick L B, Josipovic M, Risum S, Jomier J, Erleben K, Vogelius I R and Darkner S 2022 Accuracy and consistency of intensity-based deformable image registration in 4DCT for tumor motion estimation in liver radiotherapy planning *PLoS One* **17** e0271064
- Tenhunen M, Korhonen J, Kapanen M, Seppälä T, Koivula L, Collan J, Saarilahti K and Visapää H 2018 MRI-only based radiation therapy of prostate cancer: workflow and early clinical experience *Acta Oncol.* **57** 902–7
- Teuwen J, Gouw Z A R and Sonke J-J 2022 Artificial intelligence for image registration in radiation oncology *Semin. Radiat. Oncol.* **32** 330–42
- Thapa R, Ahunbay E E, Zhang J and Li A 2019 Management of Daily Variation of Air cavities during MRI-guided online adaptive radiation therapy *Int. J. Radiat. Oncol. Biol. Phys.* **105** E247–8
- Trapp P, Maier J, Susenburger M, Sawall S and Kachelrieß M 2022 Empirical scatter correction: CBCT scatter artifact reduction without prior information *Med. Phys.* **49** 4566–84
- van Heerden L E, Houweling A C, Koedooder K, van Kesteren Z, van Wieringen N, Rasch C R N, Pieters B R and Bel A 2017 Structure-based deformable image registration: added value for dose accumulation of external beam radiotherapy and brachytherapy in cervical cancer *Radiother. Oncol.* **123** 319–24

- van Herk M, Remeijer P, Rasch C and Lebesque J V 2000 The probability of correct target dosage: dose-population histograms for deriving treatment margins in radiotherapy *Int. J. Radiat. Oncol. Biol. Phys.* **47** 1121–35
- van Kranen S, Hamming-Vrieze O, Wolf A, Damen E, van Herk M and Sonke J-J 2016 Head and neck margin reduction with adaptive radiation therapy: robustness of treatment plans against anatomy changes *Int. J. Radiat. Oncol. Biol. Phys.* **96** 653–60
- van Timmeren J E et al 2020 Treatment plan quality during online adaptive re-planning *Radiat. Oncol.* **15** 203
- van der Bijl E, Remeijer P, Sonke J-J, van der Heide U A and Janssen T 2022 Adaptive margins for online adaptive radiotherapy *Phys. Med. Biol.* **67**
- Vandemeulebroucke J, Rit S, Kybic J, Clarysse P and Sarrut D 2011 Spatiotemporal motion estimation for respiratory-correlated imaging of the lungs *Med. Phys.* **38** 166–78
- Varadhan R, Karangelis G, Krishnan K and Hui S 2013 A framework for deformable image registration validation in radiotherapy clinical applications *J. Appl. Clin. Med. Phys.* **14** 4066
- Vasquez Osorio E, Abravan A, Green A, van Herk M, Lee L W, Ganderton D and McPartlin A 2023a Dysphagia at 1 year is associated with mean dose to the inferior section of the brain stem *Int. J. Radiat. Oncol. Biol. Phys.* **117** 903–13
- Vasquez Osorio E, Mayo C, Jackson A and Appelt A 2023b Challenges of re-irradiation: a call to arms for physicists - and radiotherapy vendors *Radiother. Oncol.* **182** 109585
- Vásquez Osorio E M, Hoogeman M S, Bondar L, Levendag P C and Heijmen B J M 2009 A novel flexible framework with automatic feature correspondence optimization for nonrigid registration in radiotherapy *Med. Phys.* **36** 2848–59
- Vásquez Osorio E M, Hoogeman M S, Méndez Romero A, Wielopolski P, Zolnay A and Heijmen B J M 2012 Accurate CT/MR vessel-guided nonrigid registration of largely deformed livers *Med. Phys.* **39** 2463–77
- Vásquez Osorio E M, Kolkman-Deurloo I-K K, Schuring-Pereira M, Zolnay A, Heijmen B J M and Hoogeman M S 2015 Improving anatomical mapping of complexly deformed anatomy for external beam radiotherapy and brachytherapy dose accumulation in cervical cancer *Med. Phys.* **42** 206–20
- Veiga C, Lourenço A M, Mouinuddin S, van Herk M, Modat M, Ourselin S, Royle G and McClelland J R 2015 Toward adaptive radiotherapy for head and neck patients: uncertainties in dose warping due to the choice of deformable registration algorithm *Med. Phys.* **42** 760–9
- Veiga C, McClelland J, Mouinuddin S, Lourenço A, Ricketts K, Annkah J, Modat M, Ourselin S, D'Souza D and Royle G 2014 Toward adaptive radiotherapy for head and neck patients: feasibility study on using CT-to-CBCT deformable registration for 'dose of the day' calculations *Med. Phys.* **41** 031703
- Veleg M, Juang T, Moseley J L, Oldham M and Brock K K 2015 Utility and validation of biomechanical deformable image registration in low-contrast images *Pract. Radiat. Oncol.* **5** e401–8
- Veleg M, Moseley J L, Svensson S, Hårdemark B, Jaffray D A and Brock K K 2017 Validation of biomechanical deformable image registration in the abdomen, thorax, and pelvis in a commercial radiotherapy treatment planning system *Med. Phys.* **44** 3407–17
- Verdun F R et al 2015 Image quality in CT: from physical measurements to model observers *Phys. Med.* **31** 823–43
- Vishnevskiy V, Gass T, Szekely G, Tanner C and Goksel O 2017 Isotropic total variation regularization of displacements in parametric image registration *IEEE Trans. Med. Imaging* **36** 385–95
- Vozzo M, Poder J, Yuen J, Buccì J and Haworth A 2021 Use of deformable image registration techniques to estimate dose to organs at risk following prostate external beam radiation therapy and high-dose-rate brachytherapy *J. Contemp. Brachytherapy* **13** 72–9
- Vrtovec T, Močnik D, Strojan P, Pernuš F and Ibragimov B 2020 Auto-segmentation of organs at risk for head and neck radiotherapy planning: from atlas-based to deep learning methods *Med. Phys.* **47** e929–50
- Wang S, Kim M, Wu G and Shen D 2017 Chapter 11 - scalable high performance image registration framework by unsupervised deep feature representations learning *Deep Learning for Medical Image Analysis* ed S K Zhou, H Greenspan and D Shen (New York: Academic) pp 245–69 Online <https://sciedirect.com/science/article/pii/B9780128104088000158>
- Wang Y, Petit S F, Osorio E V, Gupta V, Romero A M and Heijmen B 2018 An individualized strategy to estimate the effect of deformable registration uncertainty on accumulated dose in the upper abdomen *Phys. Med. Biol.* **63** 125005
- Weistrand O and Svensson S 2015 The ANACONDA algorithm for deformable image registration in radiotherapy *Med. Phys.* **42** 40–53
- Welgemoed C, Spezi E, Riddle P, Gooding M J, Gujral D, McLauchlan R and Aboagye E O 2023 Clinical evaluation of atlas-based auto-segmentation in breast and nodal radiotherapy *Br. J. Radiol.* **96** 20230040
- Willigenburg T, Zachiu C, Lagendijk J, van der Voort van Zyp J R N, de Boer H C J and Raaymakers B W 2022 Fast and accurate deformable contour propagation for intra-fraction adaptive magnetic resonance-guided prostate radiotherapy *Phys. Imaging Radiat. Oncol.* **21** 62–5
- Woerner A J, Choi M, Harkenrider M M, Roeske J C and Surucu M 2017 Evaluation of Deformable Image Registration-Based Contour Propagation From Planning CT to Cone-Beam CT *Technol. Cancer Res. Treat.* **16** 801–10
- Wolterink J M, Zwienenberg J C and Brune C 2022 Implicit neural representations for deformable image registration *Proc. of The 5th Int. Conf. on Medical Imaging with Deep Learning Int. Conf. on Medical Imaging with Deep Learning (PMLR)* pp 1349–59 Online <https://proceedings.mlr.press/v172/wolterink22a.html>
- Wolthaus J W H, Sonke J J, van Herk M and Damen E M F 2008 Reconstruction of a time-averaged midposition CT scan for radiotherapy planning of lung cancer patients using deformable registration *Med. Phys.* **35** 3998–4011
- Wu G, Kim M, Wang Q, Gao Y, Liao S and Shen D 2013 Unsupervised deep feature learning for deformable registration of MR brain images *Med. Image Comput. Comput.-Assist. Interv. MICCAI Int. Conf. Med. Image Comput. Comput.-Assist. Interv.* **16** 649–56
- Wu Q, Chi Y, Chen P Y, Krauss D J, Yan D and Martinez A 2009 Adaptive replanning strategies accounting for shrinkage in head and neck IMRT *Int. J. Radiat. Oncol. Biol. Phys.* **75** 924–32
- Wu X, Amstutz F, Weber D C, Unkelbach J, Lomax A J and Zhang Y 2023 Patient-specific quality assurance for deformable IMRT/IMPT dose accumulation: proposition and validation of energy conservation based validation criterion *Med. Phys.* **n/a** Online <https://onlinelibrary.wiley.com/doi/abs/10.1002/mp.16564>
- Xi J, Luo L, Li J, Zhang Y, Wang Y and Jiang D 2022 An attention-based residual neural network for deformable image registration *Thirteenth Int. Conf. on Information Optics and Photonics (CIOP 2022) Thirteenth Int. Conf. on Information Optics and Photonics (CIOP 2022) 12478 (SPIE)* pp 373–83
- Xiao H, Ren G and Cai J 2020 A review on 3D deformable image registration and its application in dose warping *Radiat. Med. Prot.* **1** 171–8
- Xiao H, Teng X, Liu C, Li T, Ren G, Yang R, Shen D and Cai J 2021 A review of deep learning-based three-dimensional medical image registration methods *Quant. Imaging Med. Surg.* **11** 4895–916
- Xiong L, Viswanathan A, Stewart A J, Haker S, Tempny C M, Chin L M and Cormack R A 2006 Deformable structure registration of bladder through surface mapping *Med. Phys.* **33** 1848–56
- Yamamoto T, Langner U, Loo B W, Shen J and Keall P J 2008 Retrospective analysis of artifacts in four-dimensional CT images of 50 abdominal and thoracic radiotherapy patients *Int. J. Radiat. Oncol. Biol. Phys.* **72** 1250–8

- Yang J *et al* 2018 Auto-segmentation for thoracic radiation treatment planning: a grand challenge at AAPM 2017 *Med. Phys.* **45** 4568–81
- Yang Y X, Teo S-K, Van Reeth E, Tan C H, Tham I W K and Poh C L 2015 A hybrid approach for fusing 4D-MRI temporal information with 3D-CT for the study of lung and lung tumor motion *Med. Phys.* **42** 4484–96
- Zachiu C, Denis de Senneville B, Willigenburg T, Voort van Zyp J R N, de Boer J C J, Raaymakers B W and Ries M 2020 Anatomically-adaptive multi-modal image registration for image-guided external-beam radiotherapy *Phys. Med. Biol.* **65** 215028
- Zakariaee R, Hamarneh G, Brown C J and Spadinger I 2016 Validation of non-rigid point-set registration methods using a porcine bladder pelvic phantom *Phys. Med. Biol.* **61** 825–54
- Zeng J, Chen J, Zhang D, Meng M, Zhang B, Qu P, Pang Q and Wang P 2020 Assessing cumulative dose distributions in combined external beam radiotherapy and intracavitary brachytherapy for cervical cancer by treatment planning based on deformable image registration *Transl. Cancer Res.* **9** 6107–15
- Zhang L, Wang Z, Shi C, Long T and Xu X G 2018 The impact of robustness of deformable image registration on contour propagation and dose accumulation for head and neck adaptive radiotherapy *J. Appl. Clin. Med. Phys.* **19** 185–94
- Zhang Y, Boye D, Tanner C, Lomax A J and Knopf A 2012 Respiratory liver motion estimation and its effect on scanned proton beam therapy *Phys. Med. Biol.* **57** 1779–95
- Zhang Y, Huth I, Weber D C and Lomax A J 2019 Dosimetric uncertainties as a result of temporal resolution in 4D dose calculations for PBS proton therapy *Phys. Med. Biol.* **64** 125005
- Zhao C, Carass A, Jog A and Prince J 2016 *Effects of Spatial Resolution on Image Registration Proceedings of SPIE—the International Society for Optical Engineering* vol **9784** p97840Y
- Zhao L and Jia K 2015 Deep adaptive log-demons: diffeomorphic image registration with very large deformations *Comput. Math. Methods Med.* **2015** 836202
- Zhong H and Chetty I J 2017 Caution must be exercised when performing deformable dose accumulation for tumors undergoing mass changes during fractionated radiation therapy *Int. J. Radiat. Oncol. Biol. Phys.* **97** 182–3
- Zhong H, Kim J, Li H, Nurusev T, Movsas B and Chetty I J 2012 A finite element method to correct deformable image registration errors in low-contrast regions *Phys. Med. Biol.* **57** 3499
- Zhong H, Weiss E and Siebers J V 2008 Assessment of dose reconstruction errors in image-guided radiation therapy *Phys. Med. Biol.* **53** 719–36
- Zhong H, Wen N, Gordon J, Elshaikh M A, Movsas B and Chetty I J 2015 An adaptive MR-CT registration method for MRI-guided prostate cancer radiotherapy *Phys. Med. Biol.* **60** 2837–51
- Ziegler M, Nakamura M, Hirashima H, Ashida R, Yoshimura M, Bert C and Mizowaki T 2019 Accumulation of the delivered treatment dose in volumetric modulated arc therapy with breath-hold for pancreatic cancer patients based on daily cone beam computed tomography images with limited field-of-view *Med. Phys.* **46** 2969–77
- Zou J, Gao B, Song Y and Qin J 2022 A review of deep learning-based deformable medical image registration *Front. Oncol.* **12** Online <https://frontiersin.org/articles/10.3389/fonc.2022.1047215>