

External Quality Assessment on Molecular Tumor Profiling with Circulating Tumor DNA-Based Methodologies Routinely Used in Clinical Pathology within the COIN Consortium

Paul van der Leest,^{a,b,†} Pim Rozendal,^{a,†} John Hinrichs,^c Carel J.M. van Noesel,^d Karen Zwaenepoel,^e Birgit Deiman,^{f,g,h,i} Cornelis J.J. Huijsmans,^j Ronald van Eijk,^k Ernst Jan M. Speel,^l Rick J. van Haastert,^m Marjolijn J.L. Ligtenberg,^{n,o} Ron H.N. van Schaik,^p Maurice P.H.M. Jansen,^q Hendrikus J. Dubbink,^r Wendy W. de Leng,^s Mathie P.G. Leers,^t Menno Tamminga,^u Daan van den Broek,^b Léon C. van Kempen,^{a,e} and Ed Schuurin^{a,*}

BACKGROUND: Identification of tumor-derived variants in circulating tumor DNA (ctDNA) has potential as a sensitive and reliable surrogate for tumor tissue-based routine diagnostic testing. However, variations in pre(analytical) procedures affect the efficiency of ctDNA recovery. Here, an external quality assessment (EQA) was performed to determine the performance of ctDNA mutation detection work flows that are used in current diagnostic settings across laboratories within the Dutch COIN consortium (ctDNA on the road to implementation in The Netherlands).

METHODS: Aliquots of 3 high-volume diagnostic leukapheresis (DLA) plasma samples and 3 artificial reference plasma samples with predetermined mutations were distributed among 16 Dutch laboratories. Participating laboratories were requested to perform ctDNA analysis for *BRAF* exon 15, *EGFR* exon 18–21, and *KRAS* exon 2–3 using their regular circulating cell-free DNA (ccfDNA) analysis work flow. Laboratories were assessed based on adherence

to the study protocol, overall detection rate, and overall genotyping performance.

RESULTS: A broad range of preanalytical conditions (e.g., plasma volume, elution volume, and extraction methods) and analytical methodologies (e.g., droplet digital PCR [ddPCR], small-panel PCR assays, and next-generation sequencing [NGS]) were used. Six laboratories (38%) had a performance score of >0.90; all other laboratories scored between 0.26 and 0.80. Although 13 laboratories (81%) reached a 100% overall detection rate, the therapeutically relevant *EGFR* p.(S752_I759del) (69%), *EGFR* p.(N771_H773dup) (50%), and *KRAS* p.(G12C) (48%) mutations were frequently not genotyped accurately.

CONCLUSIONS: Divergent (pre)analytical protocols could lead to discrepant clinical outcomes when using the same plasma samples. Standardization of (pre)analytical work flows can facilitate the implementation of reproducible liquid biopsy testing in the clinical routine.

^aDepartment of Pathology, University Medical Center Groningen, University of Groningen, Groningen, the Netherlands; ^bDepartment of Laboratory Medicine, Netherlands Cancer Institute, Amsterdam, the Netherlands; ^cDepartment of Pathology, Symbiant B.V., Alkmaar, the Netherlands; ^dDepartment of Pathology, Amsterdam University Medical Centre, University of Amsterdam, Amsterdam, the Netherlands; ^eDepartment of Pathology, Antwerp University Hospital, University of Antwerp, Edegem, Belgium; ^fClinical Laboratory, Catharina Hospital Eindhoven, Eindhoven, the Netherlands; ^gInstitute for Complex Molecular Systems, Laboratory of Chemical Biology, Eindhoven University of Technology, Eindhoven, the Netherlands; ^hDepartment of Biomedical Engineering, Laboratory of Chemical Biology, Eindhoven University of Technology, Eindhoven, the Netherlands; ⁱExpert Center Clinical Chemistry Eindhoven, Eindhoven, the Netherlands; ^jPathologie-DNA, Laboratory for Molecular Diagnostics, Location Jeroen Bosch Hospital, 's-Hertogenbosch, the Netherlands; ^kDepartment of Pathology, Leiden University Medical Centre, Leiden, the Netherlands; ^lDepartment of Pathology, GROW-School for Oncology and Reproduction, Maastricht University Medical Center, Maastricht, the Netherlands; ^mDepartment of Clinical Chemistry, St. Antonius Hospital, Nieuwegein, the Netherlands; ⁿDepartment of

Human Genetics, Radboud Institute for Medical Innovation, Radboud University Medical Center, Nijmegen, the Netherlands; ^oDepartment of Pathology, Radboud Institute for Medical Innovation, Radboud University Medical Center, Nijmegen, the Netherlands; ^pDepartment of Clinical Chemistry, Erasmus MC University Medical Center, Rotterdam, the Netherlands; ^qDepartment of Medical Oncology, Laboratory of Translational Genomics, Erasmus MC University Medical Center, Rotterdam, the Netherlands; ^rDepartment of Pathology, Erasmus MC University Medical Center, Rotterdam, the Netherlands; ^sDepartment of Pathology, University Medical Center Utrecht, Utrecht, the Netherlands; ^tDepartment of Clinical Chemistry & Hematology, Zuyderland Medical Center, Heerlen, the Netherlands; ^uDepartment of Pulmonary Medicine, University Medical Center Groningen, University of Groningen, Groningen, the Netherlands.

*Address correspondence to this author at: Department of Pathology, University Medical Center Groningen, University of Groningen, Hanzeplein 1, 9713GZ Groningen, the Netherlands, tel +31-50-36-0020; E-mail e.schuuring@umcg.nl.

†Authors contributed equally to this work.

Received October 2, 2023; accepted December 21, 2023.
<https://doi.org/10.1093/clinchem/hvae014>

Introduction

Liquid biopsy approaches, particularly the detection of tumor-derived variants in circulating tumor DNA (ctDNA) in the plasma of cancer patients, have gained increasing interest in oncological diagnostics (1–4). Circulating cell-free DNA (ccfDNA)—which includes ctDNA—can easily be collected using a minimally invasive blood withdrawal. Liquid biopsy applications include early screening, molecular target detection, treatment response prediction, early detection of therapy resistance, and minimal residual disease (MRD) and disease monitoring (1, 5). However, clinical application of liquid biopsy remains limited due to the absence of harmonized (pre) analytical work flows, the lack of data showing clinical utility, regulation, and standardization (6, 7).

The COIN consortium (ctDNA on the road to implementation in The Netherlands; www.cfdna.nl/coin) is an endeavor to implement ctDNA as a biomarker in the Netherlands, which requires harmonization of methods. Current preanalytical protocols are very diverse concerning blood collection tubes and the processing thereof, storage conditions, ccfDNA extraction methods, and elution volumes (8–10). While ccfDNA quantity, quality, and mutation detection sensitivity can be dramatically impacted by preanalytical processes (11), standardization of these processes is lacking. A previous round-robin trial within the framework of the COIN consortium demonstrated that, indeed, preanalytical conditions vary significantly across laboratories in the Netherlands, and that the choice of ccfDNA extraction method affects the overall ccfDNA yield and ctDNA abundance (6). Other studies confirm that plasma volume, extraction method, and automation have an impact on the ccfDNA fraction that can be extracted from (artificial) plasma (8, 12–14), which was reviewed recently (15).

Because only a few plasma-based ctDNA detection assays have received FDA approval and/or CE marking required for in vitro diagnostic (IVD) devices for use in the clinical setting, many analyses are performed with research-use-only tests (5, 16). In addition, methodologies vary, from single-target PCR-based approaches through small, targeted panels, to broad next-generation sequencing (NGS) strategies. Each method has particular value for specific applications in clinical diagnostics: single-target approaches are suitable for tumor-informed monitoring or ctDNA dynamics, while expanded panel analyses are applicable to identify tumor-derived variants at primary detection or resistance mechanisms at progression to assist in treatment decision-making. However, other considerations (e.g.,

turnaround time, costs, reimbursement, and demand for expertise) frequently affect the choice of testing. Studies to compare the effect of different (pre)analytical ctDNA testing work flows on analyte detection are required to demonstrate a possible impact of the method of liquid biopsy testing in clinical practice.

The aim of this study was to gain insight into the performance of ctDNA mutation detection work flows that are used in current diagnostic settings across laboratories within the Dutch COIN consortium. An external quality assessment (EQA) was performed for the detection of tumor-derived driver mutations in ccfDNA involving 16 laboratories that are presently using ctDNA testing in clinical or research settings. The participating laboratories received aliquots of 3 commercial spiked-in reference plasma samples and 3 patient-derived high-volume diagnostic leukapheresis (DLA) plasma samples with a pathogenic mutation in either *EGFR* or *KRAS* for their analysis. Molecular reports were collected and assessed by a central reference laboratory.

Materials and Methods

RECRUITMENT OF PARTICIPANTS

A survey was conducted among laboratories specialized in pathology, medical oncology, and clinical chemistry within the Dutch COIN consortium to identify sites performing plasma-derived ctDNA analysis. Laboratories were invited to participate in this interlaboratory EQA. Seventeen laboratories indicated their willingness to participate in the EQA. All laboratories were requested to track their proceedings and report the results according to a standardized form (online [Supplemental Appendix 1](#)). One laboratory failed to report its results before the deadline and was excluded; hence, 16 laboratories were evaluated in this EQA.

PLASMA SAMPLE SELECTION AND CHARACTERIZATION

Aliquots (4 mL each) of 6 high-volume plasma samples were distributed among the participating laboratories. Three samples constituted of artificial ccfDNA (25 ng/mL) in plasma either spiked-in or not spiked-in with several clinically relevant mutations (purchased from LGC Clinical Diagnostics; online [Supplemental Table 1](#)). DLA plasma samples from 3 patients with metastatic non-small cell lung cancer (NSCLC) who were treated at the University Medical Center Groningen (UMCG, Groningen, the Netherlands) were retrieved from the UMCG DLA-Biobank. Patients provided written informed consent for the use of their samples for research and validation purposes. DLA plasma samples were selected based on (a) the availability of at least 100 mL of plasma, (b) identification of a common pathogenic mutation in *BRAF* exon 15, *EGFR* exon 18–21 or *KRAS* exon

2–3 during tumor tissue NGS analysis, and (c) that the identified common pathogenic mutation is also detectable in the DLA plasma. On-treatment cell-free DLA plasma samples were collected in citrate as described previously (17, 18) and stored as 50 mL fractions at -80°C within 30 min after withdrawal. After thawing, 2 DLA fractions of 50 mL were pooled and centrifuged at $1600g$ for 10 min to separate the plasma from the debris. Plasma was aliquoted to 1 mL volumes and stored at -80°C until their shipment on dry ice to the participating laboratories. Prior to distribution, plasma samples were analyzed for the presence of all mutations in *BRAF* exon 15, *EGFR* exon 18–21, and *KRAS* exon 2–3 in an independent analysis using the NGS-based Avenio ctDNA Expanded Kit (Roche). ccfDNA was extracted from 2 mL of DLA plasma with the silica membrane-based QIAamp Circulating Nucleic Acid Kit and quantified using the Qubit dsDNA HS assay kit (Thermo Fischer Scientific) as reported previously (11). ccfDNA concentration and ctDNA NGS-based variant allele frequencies (VAFs) detected in the plasma samples are described in Table 1.

ctDNA ANALYSIS BY THE PARTICIPANTS

All laboratories received 4 mL aliquots of each sample and were requested to perform ctDNA analysis for *BRAF* exon 15, *EGFR* exon 18–21, and *KRAS* exon 2–3 using their regular ccfDNA analysis work flow. Procedures from ccfDNA extraction to molecular profiling were documented and reported to the central reference laboratory (UMCG). Laboratories were assessed based on the execution of the study protocol (analysis of the requested loci), overall sensitivity of mutation detection, and overall performance according to custom performance score criteria (online Supplemental Table 2).

STATISTICAL ANALYSES

Descriptive statistics were used to characterize data. A performance score was calculated using custom marking criteria (Supplemental Table 2) by addition of all points and deductions awarded and division by the total number of variants tested for. Figures were generated with Prism version 9.1.0. (GraphPad Software).

Results

(PRE)ANALYTICAL WORK FLOW PROCEDURES

Table 2 describes the different procedures for ccfDNA extraction and quantification by the 16 participants. A broad range of plasma input and elution volume was used for the ccfDNA extraction; 8 different extraction methods were used. Eleven laboratories quantified the ccfDNA prior to analysis, most frequently using

fluorescence-based methods (i.e., Qubit, Nanodrop, TapeStation).

Mutation analyses were performed with a great variety of methodologies (online Supplemental Appendix 2). Laboratories used either single-target analysis for a set of mutations using droplet digital PCR (ddPCR), small-panel PCR assays (i.e., Cobas), in-house-developed or commercially available NGS approaches (i.e., Oncomine, Avenio), or a combination of techniques to determine ctDNA mutations (Fig. 1). Some laboratories did not analyze all requested exons, which was considered during the overall detection rate and performance score analysis.

OVERALL DETECTION RATE AND PERFORMANCE SCORE

The reference samples (Samples 1 and 3) each contained 9 mutations and the clinical samples each had one actionable variant within the loci of interest (i.e., *BRAF* exon 15, *EGFR* exon 18–21, and *KRAS* exon 2–3; Table 1). The detection rate for the 21 variants across 5 samples was analyzed (Fig. 2; Supplemental Appendix 2). Several laboratories did not apply an assay that can detect every mutation; these mutations were removed from the total number of analyzed mutations for these laboratories to determine the overall detection rate, defined as the number of mutations identified (either specifically or nonspecifically genotyped) divided by the number of analyzed mutations (see “overall detection rate” in online Supplemental Table 3). Although 13 laboratories reported all mutations tested for, some clinically relevant mutations were not accurately genotyped (Fig. 2; Supplemental Table 3). The therapeutically relevant *EGFR* p.(S752_I759del) and p.(N771_H773dup) mutations were not identified by 11 (69%) and 8 (50%) of the participating laboratories, respectively (Fig. 2). Furthermore, only 33 out of a possible 64 calls (52%) specific for *KRAS* p.(G12C) were reported, which, in contrast to other codon 12/13 mutations, could render a patient eligible for targeted therapy. Lack of detection of these actionable mutations was primarily due to the inability of the applied assay to correctly genotype the expected variant.

The performance score was determined for each laboratory, considering the accuracy of detection and false-negative and false-positive rates (online Supplemental Tables 3 and 4). Six laboratories had a performance score of >0.90 (Fig. 3; Supplemental Table 3), of which 5 used an NGS approach and one used ddPCR (Fig. 1). The 9 laboratories with a performance score between 0.56 and 0.80 mostly applied ddPCR and Cobas approaches to determine the requested variants. Laboratory 4, which used an NGS approach, had a low performance score due to many nonspecific and false-negative calls (Supplemental Table 3). For each participating laboratory, the coverage

Table 1. Sample characteristics.

Sample type	Sample ID	Mutations in loci of interest ^a	ccfDNA, ng/μL	Input, ng	VAF	Mutant molecules ^b
Reference plasma samples	Sample 1 ^c	<i>BRAF</i> c.1799T > A; p.(V600E)	1.34	25.0	0.54%	53
		<i>EGFR</i> c.2235_2249del; p.(E746_A750del)			0.38%	38
		<i>EGFR</i> c.2240_2257del; p.(L747_P753delinsS)			0.21%	21
		<i>EGFR</i> c.2254_2277del; p.(S752_I759del)			0.17%	16
		<i>EGFR</i> c.2369C > T; p.(T790M)			0.42%	41
		<i>EGFR</i> c.2573T > G; p.(L858R)			0.40%	40
		<i>KRAS</i> c.34G > T; p.(G12C)			0.46%	45
		<i>KRAS</i> c.35G > A; p.(G12D)			0.51%	50
		<i>KRAS</i> c.183A > C; p.(Q61H)			0.59%	58
		Sample 2 ^d	—		1.22	25.0
Sample 3 ^e	Sample 3 ^e	<i>BRAF</i> c.1799T > A; p.(V600E)	1.51	25.0	1.14%	132
		<i>EGFR</i> c.2235_2249del; p.(E746_A750del)			1.10%	127
		<i>EGFR</i> c.2240_2257del; p.(L747_P753delinsS)			1.43%	166
		<i>EGFR</i> c.2254_2277del; p.(S752_I759del)			1.15%	133
		<i>EGFR</i> c.2369C > T; p.(T790M)			0.89%	103
		<i>EGFR</i> c.2573T > G; p.(L858R)			0.85%	99
		<i>KRAS</i> c.34G > T; p.(G12C)			0.95%	110
		<i>KRAS</i> c.35G > A; p.(G12D)			0.72%	84
		<i>KRAS</i> c.183A > C; p.(Q61H)			0.72%	83
Clinical plasma samples	Sample 4	<i>EGFR</i> c.2311_2319dup; p.(N771_H773dup)	0.392	15.2	9.3%	278
	Sample 5	<i>KRAS</i> c.34G > T; p.(G12C)	0.545	11.8	4.5%	100
	Sample 6	<i>KRAS</i> c.34G > T; p.(G12C)	0.878	36.3	1.6%	109

^aLoci of interest encompass *BRAF* exon 15, *EGFR* exon 18–21, and *KRAS* exon 2–3. Mutations were determined in an independent run using the Avenio ctDNA Expanded Kit at the central analysis laboratory (University Medical Center Groningen [UMCG]).

^bDepicted as mutant copies/mL of plasma.

^cSeraCare ctDNA Complete™ Reference Material VAF0.5% (see Supplemental Table 1).

^dSeraCare ctDNA Complete™ Reference Material WT (see Supplemental Table 1).

^eSeraCare ctDNA Complete™ Reference Material VAF1% (see Supplemental Table 1).

of requested exons, overall detection rate, and performance score are depicted in Fig. 3.

Discussion

Previous studies have determined the effect of various facets in the (pre)analytical work flow on the quantity and composition of ccfDNA, which resulted in overviews of the critical factors in liquid biopsy testing (15, 20, 21). However, whether these variables eventually affect the overall detection rate of tumor-derived mutations in ctDNA remained largely unclear. Here, the impact of divergent (pre)analytical protocols on the detection of clinically relevant mutations in cell-free plasma was evaluated across 16 laboratories within the Dutch COIN consortium.

Since (inter)national harmonization of (pre)analytics has not yet been achieved (22, 23), most laboratories develop their local liquid biopsy work flows and implement them for research and clinical purposes. Particularly in the preanalytical phase, variations that affect the quantity and quality of ccfDNA could be induced due to many different commercially available products (15, 21). To gain insight into the use of (pre)analytical products in the Netherlands, a general survey was conducted among laboratories contributing to the COIN consortium. Laboratories willing to participate in an interlaboratory EQA reported a wide variety of (pre)analytical work flows for liquid biopsy applications. As expected, differences in plasma use, ccfDNA extraction methods and conditions, and mutation detection assays were applied by the laboratories participating in this EQA. Variations in preanalytical conditions can considerably affect the quantity and quality

Table 2. Laboratory preanalytical conditions.

Lab	Plasma volume, mL	Elution volume, μ L	Extraction method ^a	ccfDNA quantification method	Remarks
1	4	100	CNA	None	None
2	4	40	CNA	Qubit	None
3	4 (2 \times 2)	120 (2 \times 60)	DSP	Qubit	None
4	3	80	ME	None	None
5	4 (2 \times 2)	200 (2 \times 100)	COB	Qubit	None
6	4 (2 \times 2)	160 (2 \times 80)	COB	None	None
7	3	50	CNA	None	None
8	4	75	RSC	Nanodrop	None
9	4 (2 \times 2)	200 (2 \times 100)	COB	None	None
10	4	55	Custom ^b	ddPCR	None
11	2	52	CNA	Qubit	None
12	4	60	AVE	TapeStation	Yes ^c
13	4	25	MMA	Qubit	None
14	3	50	CNA	Qubit	None
15	4	50	CNA	Qubit	None
16	4	70	RSC	Qubit	None

^aAbbreviations: CNA, QIAamp Circulating Nucleic Acid Kit; DSP, QIASymphony DSP Circulating DNA Kit; ME, QIAamp MinElute ccfDNA Mini Kit; COB, Cobas ccfDNA Sample Preparation Kit; RSC, Maxwell RSC LV ccfDNA Kit; AVE, AVENIO cfDNA Isolation Kit; MMA, MagMAX Cell-Free Total Nucleic Acid Kit.


^bApplied a modified protocol (see Supplemental Appendix 1).

^cPlasma Samples 4 to 6 were nearly thawed upon arrival at the site.

of the ccfDNA, which ultimately can hamper the sensitivity and accuracy of mutation detection (6, 15). Moreover, the choice of an analytical method to detect tumor-derived mutations affects the detection rate as well (24, 25); 10 different assays were used in this EQA. Previous EQAs have shown highly discrepant detection rates among laboratories for plasma samples with mutations spiked-in at 1% VAF (1, 26). To this end, 3 commercially available reference plasma samples spiked-in with clinically relevant mutations at 1%, 0.5%, and 0% VAF were distributed. Biological features of clinical patient-derived plasma samples, however, might affect the detection of ctDNA in the plasma (20). Therefore, 3 patient-derived plasma samples with proven actionable mutations were included as well. Since the VAFs of these variants were quite high (>1%), which should be above the analytical sensitivity of the applied molecular profiling methods, the patient samples were also included in the performance analysis.


Laboratories were scored based on whether the requested loci of interest were tested, and on the overall detection rate of the expected variants. Molecular profiling was performed by 11/16 laboratories (69%) for at least one variant in all requested exons, of which 5 laboratories (31%) applied a test that could detect all variants present across the samples. Considering only variants for which analysis was performed, 13/16 laboratories (81%) scored

an overall detection rate of 100% for the expected variants with or without accurate genotyping (i.e., specific amino acid change), implicating a high sensitivity of the (screening) assay used. However, reporting of a variant without the specific amino acid change (e.g., *KRAS* p.(G12/G13)) cannot support clinical decision-making if an actionable mutation (i.e., *KRAS* p.(G12C)) is present. In current routine diagnostics, accurate characterization of the molecular profile of tumors is essential to treat NSCLC patients appropriately. Identification of the precise nucleic acid change could therefore determine whether a patient is considered eligible for targeted treatment or will receive nonspecific therapy. To this end, the participating laboratories were graded on performance based on custom marking criteria as well, considering the therapeutic implications of the molecular report and false-positive calls in the exons analyzed. In general, laboratories using screening assays (i.e., ddPCR screening assays, Cobas mutations tests) had a lower performance score as the methods were unable to distinguish the specific variant from other alterations at that codon. In contrast, NGS-based methodologies can define specific variants more accurately, but are more prone to false-positive calls since a wide range of genomic loci are analyzed simultaneously. Laboratory 16 had a performance score of 1 while using ddPCR but only tested for 57% of the variants, highlighting that—



although single-target testing can be sensitive and specific—a large number of sequential or parallel single-target tests will be required to analyze all loci of interest. Most laboratories that did not apply an NGS-based methodology were particularly unable to appropriately characterize the *KRAS* p.(G12C) and *EGFR* p.(N771_H773dup) mutations, which have clinical significance regarding treatment decision-making. *KRAS* p.(G12C)-mutant NSCLC patients with progressive disease following first-line (chemo-)immunotherapy are considered eligible for subsequent treatment with sotorasib or adagrasib according to international guidelines (https://www.nccn.org/professionals/physician_gls/PDF/nscl.pdf, accessed on November 6, 2023), while no targeted therapies are available for other *KRAS* mutations. Upon detection of most common *EGFR* mutations, patients are treated with osimertinib in first-line. However, *EGFR* exon 20 insertions are negative predictors for osimertinib treatment and, therefore, these patients generally receive amivantamab-targeted antibodies subsequent to (chemo-)immunotherapy. NSCLC patients with missense mutations at codon 600 of *BRAF* are preferably treated with first-line dabrafenib/trametinib combination therapy (27).

For this study, gathering plasma through the DLA procedure enabled collection and distribution of aliquots of identical clinical samples to the participating laboratories which would not have been feasible with regular blood draws. However, the VAFs in the patient-derived DLA samples were excessively high (>1%), while plasma ctDNA generally represents a low fraction of the total ccfDNA, particularly in early stages of disease (28). Detection with high sensitivity and specificity of ctDNA variants with low abundance in plasma can be challenging because of a possible low signal-to-noise ratio. Therefore, sequencing artifacts and clonal hematopoiesis of indeterminate potential (CHIP) will be more difficult to discriminate from true tumor-derived mutations (29, 30), especially for clinical applications in which low quantities of ctDNA are expected (e.g., screening, MRD monitoring). Novel technologies have been developed to resolve these issues (e.g., unique molecular identifiers [UMIs], patient-specific tumor-informed sequencing panels, and bioinformatics approaches), however remain to be validated on clinical samples with low plasma tumor fractions (31). Here, only Sample 1 contained variants with <0.5% (Supplemental Table 1). These were detected by all participants who performed






Fig. 3. Performance score, exon coverage, and overall detection rate. Performance was assessed using custom marking criteria (Supplemental Table 2). The percentage of analyzed exons is depicted based on the coverage of the applied assays. The percentage of identified variants detectable by the applied assays is expressed as the overall detection rate (also see Supplemental Table 2).

a test that should identify the particular variants, with Laboratory 4 as an exception. However, this was a commercial reference plasma sample containing high-quality spiked-in DNA and, therefore, does not reflect clinical samples. As such, the current study did not address in detail the accuracy of the detection of variants with low VAF in clinical samples.

This interlaboratory EQA exemplifies how divergent (pre)analytical protocols could lead to discrepant clinical outcomes when using the same plasma samples. Preanalytical work flows affect the quality of the material analyzed, whereas the choice of an analytical assay determines which variants could be identified. Although the sensitivity and specificity of single-target approaches or small panels are sufficient (8, 24), they are inadequate to cover all required genomic loci according to guidelines, rendering them inappropriate for primary screening purposes. National and international consortia should pursue the standardization of (pre)analytical work flows to facilitate the implementation of liquid biopsy testing in the clinical routine.

Supplemental Material

Supplemental material is available at *Clinical Chemistry* online.

Nonstandard Abbreviations: ctDNA, circulating tumor DNA; EQA, external quality assessment; COIN, ctDNA on the road to implementation in The Netherlands; DLA, diagnostic leukapheresis; ccfDNA, circulating cell-free DNA; ddPCR, droplet digital PCR; NGS, next-generation sequencing; IVD, in vitro diagnostic; NSCLC, non-small cell lung cancer; VAF, variant allele frequency.

Human Genes: *BRAF*, B-Raf proto-oncogene, serine/threonine kinase; *EGFR*, epidermal growth factor receptor; *KRAS*, KRAS proto-oncogene, GTPase.

Author Contributions: The corresponding author takes full responsibility that all authors on this publication have met the following required criteria of eligibility for authorship: (a) significant contributions to the conception and design, acquisition of data, or analysis and interpretation of data; (b) drafting or revising the article for intellectual content; (c) final approval of the published article; and (d) agreement to be accountable for all aspects of the article thus ensuring that questions related to the accuracy or integrity of any part of the article are appropriately investigated and resolved. Nobody who qualifies for authorship has been omitted from the list.

Authors' Disclosures or Potential Conflicts of Interest: Upon manuscript submission, all authors completed the author disclosure form.

Research Funding: The COIN consortium is funded by a ZonMW research grant (project number: 848101011) to D. van den Broek, paid to the Netherlands Cancer Institute. M. Tamminga has received a CANCER-ID IMI grant to support the collection of the DLA samples (project number: 115749).

Disclosures: C.J.J. Huijsmans is a consultant on the advisory board for Janssen-Cilag. E.J.M. Speel is a consultant on advisory boards for AstraZeneca, GlaxoSmithKline, Janssen-Cilag, and Merck Serono, and has received grants from Bayer and Pfizer. M.J.L. Ligtenberg is a consultant on advisory boards for AstraZeneca, GlaxoSmithKline, Illumina, and Janssen-Cilag, and has performed lectures for Medtalks and Rorij congressen. M.P.H.M. Jansen has received grants from GlaxoSmithKline. H.J. Dubbink is a consultant on the advisory board and has received grants from AstraZeneca. W.W. de Leng has a leadership role in the section Clinical and Experimental Molecular Pathology of the Dutch Pathology Society, is a consultant on advisory boards for Janssen-Cilag and Novartis, and has received grants from Bristol-Myers Squibb and Roche. L.C. van Kempen has a leadership role in the EORTC Melanoma Group and the Commission Personalized Medicine—Belgium, is a consultant on advisory boards for Cyclomics, Janssen-Cilag, LOGEX, Merck Serono, Protyon, and Roche, has received grants and non-financial support from Amgen, AstraZeneca, Bayer, Bristol-Myers Squibb, Eli Lilly, Janssen-Cilag, LOGEX, Lynxcare, Merck Serono, nanoString, Novartis, Pfizer, Roche, and ThermoFisher, and has stocks in Cyclomics. E. Schuurin has a leadership role in the Dutch Society of Pathology, the European Society of Pathology, the European Liquid Biopsy Society, the CieBOD, and the national guideline committee, is a consultant on advisory boards for Agena Bioscience, Amgen, Astellas Pharma, AstraZeneca, Bayer, Bristol-Myers Squibb, CC Diagnostics, Eli Lilly, GlaxoSmithKline, Illumina, Janssen-Cilag, Merck Serono, Novartis, Roche, SinnoVision Lab, and Sysmex, has performed lectures for Agena Bioscience, Biocartis, Bio-Rad, Eli Lilly, Illumina, Roche, and SeraCare Life Sciences, and has received grants and non-financial support from Abbott, Agena Bioscience, ArcherDX, AstraZeneca, Bayer, Biocartis, Bio-Rad, Boehringer Ingelheim, CC Diagnostics, Eli Lilly, Illumina, Merck Serono, Roche, and SeraCare Life Sciences.

Role of Sponsor: The funding organizations played no role in the design of study, choice of enrolled patients, review and interpretation of data, preparation of manuscript, or final approval of manuscript.

Acknowledgments: We thank all contributing members of the COIN consortium, in particular F.J.G. Scherpen (Department of Pathology, University Medical Center Groningen), D. van Egmond and L. Steeghs (both Leiden University Medical Center), L.I. Kroeze and M.J. Geerlings (both Radboud University Medical Center), E. de

Jonge (Department of Clinical Chemistry, Erasmus MC University Medical Center), C. Beaufort and J. Helmijr (both Department of Medical Oncology, Erasmus MC University Medical Center), B. Jongen (Zuyderland Medical Center), and C. Meeues (Netherlands Cancer Institute) for their contributions to this study.

References

- Heitzer E, Haque IS, Roberts CES, Speicher MR. Current and future perspectives of liquid biopsies in genomics-driven oncology. *Nat Rev Genet* 2019;20:71–88.
- Moding EJ, Diehn M, Wakelee HA. Circulating tumor DNA testing in advanced non-small cell lung cancer. *Lung Cancer* 2018;119:42–7.
- Ossandon MR, Agrawal L, Bernhard EJ, Conley BA, Dey SM, Divi RL, et al. Circulating tumor DNA assays in clinical cancer research. *J Natl Cancer Inst* 2018;110:929–34.
- Aggarwal C, Thompson JC, Black TA, Katz SI, Fan R, Yee SS, et al. Clinical implications of plasma-based genotyping with the delivery of personalized therapy in metastatic non-small cell lung cancer. *JAMA Oncol* 2019;5:173–80.
- Heitzer E, van den Broek D, Denis MG, Hofman P, Hubank M, Moulriere F, et al. Recommendations for a practical implementation of circulating tumor DNA mutation testing in metastatic non-small-cell lung cancer. *ESMO Open* 2022;7:100399.
- van der Leest P, Ketelaar EM, van Noesel CJM, van den Broek D, van Boerdonk RAA, Deiman B, et al. Dutch National Round Robin Trial on plasma-derived circulating cell-free DNA extraction methods routinely used in clinical pathology for molecular tumor profiling. *Clin Chem* 2022;68:963–72.
- Alix-Panabières C, Pantel K. Liquid biopsy: from discovery to clinical application. *Cancer Discov* 2021;11:858–73.
- Lampignano R, Neumann MHD, Weber S, Klotten V, Herdean A, Voss T, et al. Multicenter evaluation of circulating cell-free DNA extraction and downstream analyses for the development of standardized (pre)analytical work flows. *Clin Chem* 2020;66:149–60.
- Deans ZC, Williams H, Dequeker EMC, Keppens C, Normanno N, Schuurin E, et al. Review of the implementation of plasma ctDNA testing on behalf of IQN path ASBL: a perspective from an EQA providers' survey. *Virchows Arch* 2017;471:809–13.
- Godsey JH, Silvestro A, Barrett JC, Bramlett K, Chudova D, Deras I, et al. Generic protocols for the analytical validation of next-generation sequencing-based ctDNA assays: a joint consensus recommendation of the BloodPAC's analytical variables working group. *Clin Chem* 2020;66:1156–66.
- van der Leest P, Boonstra PA, ter Elst A, van Kempen LC, Tibbesma M, Koopmans J, et al. Comparison of circulating cell-free DNA extraction methods for downstream analysis in cancer patients. *Cancers (Basel)* 2020;12:1222.
- van Dessel LF, Beije N, Helmijr JCA, Vitale SR, Kraan J, Look MP, et al. Application of circulating tumor DNA in prospective clinical oncology trials—standardization of preanalytical conditions. *Mol Oncol* 2017;11:295–304.
- Warton K, Graham L-J, Yuwono N, Samimi G. Comparison of 4 commercial kits for the extraction of circulating DNA from plasma. *Cancer Genet* 2018;228–229:143–50.
- Klotten V, Röchel N, Bröchle NO, Gasthaus J, Freudenmacher N, Steib F, et al. Liquid biopsy in colon cancer: comparison of different circulating DNA extraction systems following absolute quantification of KRAS mutations using Intplex allele-specific PCR. *Oncotarget* 2017;8:86253–63.
- van der Leest P, Schuurin E. Critical factors in the analytical workflow of circulating tumor DNA-based molecular profiling. *Clin Chem* 2024;70:220–33.
- Food and Drug Administration. List of cleared or approved companion diagnostic devices (in vitro and imaging tools). <https://www.fda.gov/medical-devices/in-vitro-diagnostics/list-cleared-or-approved-companion-diagnostic-devices-in-vitro-and-imaging-tools> (Accessed October 2023).
- Tamminga M, de Wit S, Hiltermann TJN, Timens W, Schuurin E, Terstappen LWMM, Groen HJM. Circulating tumor cells in advanced non-small cell lung cancer patients are associated with worse tumor response to checkpoint inhibitors. *J Immunother Cancer* 2019;7:173.
- Tamminga M, Oomens L, Hiltermann TJN, Andree KC, Tibbe A, Broekmaat J, et al. Microsieves for the detection of circulating tumor cells in leukapheresis product in non-small cell lung cancer patients. *Transl Lung Cancer Res* 2020;9:1093–100.
- Hofste LSM, Geerlings MJ, von Rhein D, Tolmeijer SH, Weiss MM, Gillissen C, et al. Circulating tumor DNA-based disease monitoring of patients with locally advanced esophageal cancer. *Cancers (Basel)* 2022;14:4417.
- Moser T, Kühberger S, Lazzeri I, Vlachos G, Heitzer E. Bridging biological cfDNA features and machine learning approaches. *Trends Genet* 2023;39:285–307.
- Ungerer V, Bronkhorst AJ, Holdenrieder S. Preanalytical variables that affect the outcome of cell-free DNA measurements. *Crit Rev Clin Lab Sci* 2020;57:484–507.
- Rolfo C, Mack P, Scagliotti GV, Aggarwal C, Arcila ME, Barlesi F, et al. Liquid biopsy for advanced NSCLC: a consensus statement from the International Association for the study of lung cancer. *J Thorac Oncol* 2021;16:1647–62.
- Pascual J, Attard G, Bidard F-C, Curigliano G, De Mattos-Arruda L, Diehn M, et al. ESMO recommendations on the use of circulating tumour DNA assays for patients with cancer: a report from the ESMO Precision Medicine working group. *Ann Oncol* 2022;33:750–68.
- Elazezy M, Joosse SA. Techniques of using circulating tumor DNA as a liquid biopsy component in cancer management. *Comput Struct Biotechnol J* 2018;16:370–8.
- Weber S, Spiegl B, Perakis SO, Ulz CM, Abuja PM, Kashofer K, et al. Technical evaluation of commercial mutation analysis platforms and reference materials for liquid biopsy profiling. *Cancers (Basel)* 2020;12:1588.
- Van Casteren K, Keppens C, Schuurin E, Deans ZC, Normanno N, Patton SJ, et al. External quality assessment schemes for biomarker testing in oncology. *J Mol Diagn* 2020;22:736–47.
- de Jager VD, Timens W, Bayle A, Botling J, Brcic L, Büttner R, et al. Developments in predictive biomarker testing and targeted therapy in advanced stage non-small cell lung cancer and their application across European countries. *Lancet Reg Health Eur* 2023;36:100787.
- Fairley JA, Cheetham MH, Patton SJ, Rouleau E, Denis M, Dequeker EMC, et al. Results of a worldwide external quality assessment of cfDNA testing in lung cancer. *BMC Cancer* 2022;22:759.
- Stetson D, Ahmed A, Xu X, Nuttall BRB, Lubinski TJ, Johnson JH, et al. Orthogonal comparison of four plasma NGS tests with tumor suggests technical factors are a major source of assay discordance. *JCO Precis Oncol* 2019;3:1–9.
- Weber S, van der Leest P, Donker HC, Schlange T, Timens W, Tamminga M, et al. Dynamic changes of circulating tumor DNA predict clinical outcome in patients with advanced non-small-cell lung cancer treated with immune checkpoint inhibitors. *JCO Precis Oncol* 2021;5:1540–53.
- Dang DK, Park BH. Circulating tumor DNA: current challenges for clinical utility. *J Clin Invest* 2022;132:e154941.