

2024

## Synopsis on the Role of Diode Laser in Root Canal Disinfection

Ashraf M. Abu-Seida

*1Faculty of Dentistry, Galala University. Suez, Egypt 2Department of Surgery, Anesthesiology & Radiology, Faculty of Veterinary Medicine, Cairo University. PO: 12211, Giza - Egypt.*

Mohamed M.N. El-Tayeb

*Department of Endodontics, Faculty of Oral and Dental Medicine, Badr University in Cairo*

Elhassan E. Hassanein

*HHU Heinrich Heine Universität Düsseldorf, Germany*

Follow this and additional works at: <https://digitalcommons.aaru.edu.jo/iajd>



Part of the [Endodontics and Endodontology Commons](#)

---

### Recommended Citation

Abu-Seida, Ashraf M.; El-Tayeb, Mohamed M.N.; and Hassanein, Elhassan E. (2024) "Synopsis on the Role of Diode Laser in Root Canal Disinfection," *International Arab Journal of Dentistry*. Vol. 15: Iss. 1, Article 18.

DOI: <https://doi.org/10.70174/iajd.v15i1.971>

Available at: <https://digitalcommons.aaru.edu.jo/iajd/vol15/iss1/18>

This Review Article is brought to you for free and open access by Arab Journals Platform. It has been accepted for inclusion in International Arab Journal of Dentistry by an authorized editor. The journal is hosted on [Digital Commons](#), an Elsevier platform. For more information, please contact [rakan@aarj.edu.jo](mailto:rakan@aarj.edu.jo), [marah@aarj.edu.jo](mailto:marah@aarj.edu.jo), [u.murad@aarj.edu.jo](mailto:u.murad@aarj.edu.jo).

## SYNOPSIS ON THE ROLE OF DIODE LASER IN ROOT CANAL DISINFECTION

Ashraf M. Abu-Seida<sup>1,2</sup> | Mohamed M.N. El-Tayeb<sup>3</sup> | Elhassan Hassanein<sup>4</sup>

**Abstract:** Effective root canal disinfection is necessary for success of endodontic treatment. Traditional chemo-mechanical techniques such as sodium hypochlorite, mechanical preparation and calcium hydroxide used for intracanal disinfection usually fail in cases of biofilm-related infections. Therefore, several new strategies are applied for root canal disinfection. One of these strategies is using of diode laser that has recently gained significant attention in the endodontic practice due to its antibacterial action, bio-stimulation, decrease of pain and improvement of success. Moreover, it has several advantages like its low cost, easily application, penetration power into dentine and smaller size device compared with other types of lasers. Diode laser has been applied for root canal disinfection with various degrees of success in both *in vitro* and *in vivo* studies. Nevertheless, small number of published randomized clinical trials in humans with small sample sizes and high heterogeneity provide insufficient support for its efficacy. Furthermore, the results of diode laser in root canal disinfection are controversial. Future high quality studies are recommended to assess the effectiveness of diode laser on various microbial endodontic species and to confirm its role in disinfection of the root canal system in human teeth. This synopsis describes the current scientific understanding and the status of diode laser efficacy in root canal disinfection.

**Keywords:** Apical periodontitis, *Enterococcus faecalis*, Necrotic pulp, Regenerative endodontic therapy, Root canal

---

**Corresponding author:**

Prof. Dr. Ashraf M. Abu-Seida, E-mails: ashrafseida@cu.edu.eg

**Conflicts of interest:**

The authors declare no conflicts of interest.

1. Faculty of Dentistry, Galala University, Suez, Egypt.

2. Department of Surgery, Anesthesiology & Radiology, Faculty of Veterinary Medicine, Cairo University. PO: 12211, Giza - Egypt.

E-mails: ashrafseida@cu.edu.eg, ashrafseida@gu.edu.eg

3. Department of Endodontics, Faculty of Oral and Dental Medicine, Badr University in Cairo, Cairo, Egypt.

E-mail: eltayeb65@hotmail.com

4. HHU Heinrich Heine University Düsseldorf, Germany.

## **SYNOPSIS SUR LE RÔLE DU LASER À DIODE DANS LA DÉSINFECTION DU CANAL RADICULAIRE**

**Résumé:** Une désinfection efficace du canal radiculaire est nécessaire pour un bon succès du traitement endodontique. Les techniques chimio-mécaniques traditionnelles telles que l'hypochlorite de sodium, la préparation mécanique et l'hydroxyde de calcium utilisées pour la désinfection intracanalaires échouent généralement en cas d'infections liées au biofilm. Par conséquent, plusieurs nouvelles stratégies sont appliquées pour la désinfection des canaux radiculaires. L'une de ces stratégies est l'utilisation du laser à diode qui a récemment attiré une attention considérable dans la pratique endodontique en raison de son action antibactérienne, de sa biostimulation, de sa diminution de la douleur et de son haut taux de succès. De plus, le laser à diode présente plusieurs avantages tels que son faible coût, sa facilité d'application, son pouvoir de pénétration dans la dentine et son dispositif de plus petite taille par rapport aux autres types de lasers. Il a été appliqué pour la désinfection des canaux radiculaires avec divers degrés de succès dans des études *in vitro* et *in vivo*. Néanmoins, un petit nombre d'essais cliniques randomisés publiés chez l'homme avec des échantillons de petite taille et une forte hétérogénéité ne soutiennent pas suffisamment son efficacité. De plus, les résultats du laser à diode dans la désinfection canalaire sont controversés. De futures études de haute qualité sont recommandées pour évaluer l'efficacité du laser à diode sur diverses espèces endodontiques microbiennes et pour confirmer son rôle dans la désinfection du système canalaire des dents humaines. Cet article décrit les connaissances scientifiques actuelles et l'efficacité du laser à diode dans la désinfection canalaire.

**Mots clés :** Parodontite apicale, *Enterococcus faecalis*, Pulpe nécrotique, Thérapie endodontique régénérative, Canal radiculaire

## Introduction

A fundamental prerequisite for endodontic treatment is effective root canal disinfection. Traditionally, mechanical preparation, sodium hypochlorite irrigant, and calcium hydroxide have been recommended for root canal disinfection [1-5]. However, they are ineffectual in cases of biofilm-related chronic infection, therefore other intracanal drugs, such as triple antibiotic paste (TAP) and double antibiotic paste (DAP), are proposed as substitutes or adjuncts in root canal disinfection [6-8]. Moreover, some natural products like propolis, chitosan and Aloe vera have been applied for intracanal disinfection [9-13].

Although sodium hypochlorite is the gold standard for root canal disinfection [14-16], it has some drawbacks, including substantivity and smear layer removal ability [17]. Furthermore, chemical disinfection is dependent on the extent to which chemical agents penetrate the root canal system [18]. As a result, the search for an optimal new intracanal disinfectant is ongoing due to the current rise in antibiotic-resistant bacteria and negative effects caused by the chemicals utilized [19, 20].

Several research on the possible benefits of laser in endodontics have been published since the introduction of the ruby laser by Maiman in 1960 and the application of laser for endodontic therapy by Weichman in 1971 [21]. The use of laser in endodontics has expanded in recent years due to the development of thinner, more flexible, and durable laser fibers [22]. Lasers, such as diode lasers, can be sent into the root canal system via fiber optics or hollow tubes, like Er-YAG lasers. These various go delivery strategies can have an impact on the performance of similar lasers in canal disinfection [23].

Due to its antibacterial action, bio-stimulation, boosting success rate and lifespan, and exploiting the thermal effect of laser on surrounding tissues, the use of laser power

has recently attracted significant interest in the endodontic profession [24, 25]. This may save patients from invasive surgical intervention, saving both the patient and the dentist time.

The interaction of laser light with tissues occurs in four ways: reflection, scattering, transmission, and absorption [26]. The wavelength of the laser, as well as the optical features of the target tissues such as pigmentation (chromophores) and water content, determine laser absorption. Melanin, hemoglobin, and water are the primary chromophores in intraoral soft tissue [27]. Water and hydroxyapatite are the primary chromophores in hard tissues. Because different wavelengths are required for different uses, different types of laser power are necessary in dental treatment [26, 27].

The laser wavelengths used for disinfecting the root canal include: erbium: yttrium aluminium garnet (Er:YAG), 2940 nm; erbium, chromium: yttrium scandium gallium garnet (Er,Cr:YSGG), 2780 nm; neodimium: yttrium aluminium garnet (Nd:YAG), 1064 nm; diode, 635 to 980 nm; potassium titanyl phosphate (KTP), 532 nm and carbon dioxide (CO<sub>2</sub>), 9600 and 10600 nm [28].

Diode laser is a solid-state semiconductor laser with wavelengths for dental applications ranging from 800 to 1,064nm. Because of their high absorption in pigments such as melanin and hemoglobin as well as transmission in water and hydroxyapatite, diode lasers are increasingly being used for root canal therapy. It promotes cell proliferation while inhibiting inflammation-causing enzymes. The efficacy of root canal disinfection utilizing diode laser wavelengths 940 nm and 980 nm was reported to have a good bactericidal impact on *Enterococcus faecalis* (*E. faecalis*) [29, 30].

Another important issue with endodontic treatment is the post-operative endodontic pain that many patients endure. When patients experience discomfort following endodontic treatment, they may

mistrust the clinician's abilities. As a result, effective pain management is crucial and diode laser is able to inhibit the inflammation-inducing enzymes [31].

This synopsis describes the current scientific understanding and the status of diode laser efficacy in root canal disinfection.

### Antibacterial activity of diode laser

Although several studies have been conducted to assess the role of diode laser in disinfection of the root canal system, the results of these studies are controversial. Some authors recommended using of diode laser as an adjunctive treatment to the chemo-mechanical methods [32-36]. On contrary, others concluded that diode laser, as a monotherapy, has a beneficial antibacterial action during root canal disinfection [37-40].

Several recent studies concluded that diode laser therapy has a limited effect on bacterial cell eradication and use of a laser as an adjuvant to regular sodium hypochlorite irrigation is a beneficial strategy [32-36]. Moreover, diode laser in continuous mode is more effective than both diode laser in pulse mode and 5.2% sodium hypochlorite solution [37].

Several studies have been concluded that diode laser can be used for complementary disinfection of root canals due to its antibacterial action. When diode laser treatment was used for root canal disinfection, a statistically significant reduction in microorganisms was observed [38-40].

The principles underlying a high-power diode laser's antibacterial properties are based on the thermal and photodisruptive effects of the produced light. High-energy radiation characterizes diode lasers with wavelengths of 980 nm. The energy is transformed to heat when it is absorbed by the water molecules in the tissues. As a result, water evaporation causes cell damage and microorganism mortality. However, there are additional published theories about bacterial cell death by laser light [41]. One plau-

sible assumption is that following irradiation, the temperature rises to an exceedingly high level for a brief period of time, and the intense heat results in the instantaneous killing of the bacteria [41].

Another theory assumes that immediate cell death may not occur during laser irradiation, but sublethal damage. Sublethal damage is explained as a disruption of cell wall integrity and the accumulation of denaturated proteins. Those changes cease the cell growth and result in cell lysis [10]. This last assumption was also described by Moritz *et al.* [42]. They concluded that laser irradiation had a cumulative effect on *E. faecalis* lethality. Heat from laser application is a stress component that causes non-lethal reversible damage that may be changed into lethal damage once the stress is repeated [43].

Diode laser is applied in many investigations due to its efficacy in lowering intracanal bacterial counts and penetrating 500 microns into dentin, low cost and smaller size than other types of lasers [44]. Morsy *et al.* evaluated antibacterial effect of 980 nm diode laser versus conventional endodontic treatment in necrotic teeth with chronic periapical lesions. They concluded that diode laser has efficient antibacterial effect [25]. An *in vitro* study concluded that diode laser irradiation (970 nm) has efficient antibacterial effect in comparison to silver nanoparticles on *E. faecalis* bacterial strain [24]. Moreover, a statistically significant reduction in the number of intracanal *E. faecalis* colonies was observed following a single and repeated 980 nm diode laser application [45].

Similarly, when compared to NaOCl 5.25% solution, ER, Cr: YSGG laser 2780nm wavelength and diode laser 940nm wavelength show approximately identical antibacterial activity on *E. faecalis* [46]. Bago *et al.* found that using a diode laser (985 nm, 2 W, 3x20 s) against *E. faecalis* biofilm was comparable to us-

ing 2.5% NaOCl for 60 seconds [47].

Unfortunately, there haven't been enough randomized clinical trials evaluating endodontic treatment outcomes after laser use [28, 44]. Future high-quality studies are highly recommended on this topic.

#### **Effect of diode laser on the post-operative endodontic pain**

Endodontic disorders such as irreversible pulpitis, apical periodontitis, and acute apical abscess are among the most common causes of dental pain [48, 49]. Pre-operative and chronic pain is a significant risk factor for endodontic treatment failure [50]. Meanwhile, as a common complication following endodontic therapy, post-operative endodontic pain has been found to have a prevalence of up to 40%, and pain control has been identified as a critical aim of endodontic therapies [51].

Standard chemomechanical-preparation may pose an ongoing risk of inflammation and post-operative pain. All the while, diode laser may reach microbes in deeper dentin layers without causing damage to the tooth structure [52]. Lasers have been shown to decrease pain-inducing chemicals (such as substance P, histamine, dopamine, and prostaglandins) while increasing synthesis of lymphokines, immunoglobulins, anti-inflammatory prostaglandins, and beta-endorphins [53].

Multiple studies have been carried out to study the effect of diode laser on pain after endodontic therapy [54-56]. All of these studies confirmed the beneficial role of diode laser in decreasing the pain after endodontic treatment. According to the findings of Kaplan *et al.*, diode laser can reduce post-operative pain after endodontic therapy in teeth with apical periodontitis [30]. Moreover, Morsy *et al.* found that diode laser 980 nm induces efficient decrease in the post-operative pain versus conventional endodontic treatment in necrotic teeth with chronic periapical lesions [25].

#### **Effect of diode laser on biomodulation**

Photobiomodulation therapy through the application of photonic energy at specific wavelengths within the optical window of 650-1350 nm, works on the principle of inducing a biological response through energy transfer [57]. Cellular metabolism can be modulated, leading to secondary effects which modify cellular behavior. The benefits of this approach can be described as anti-inflammatory, analgesic and therapeutic and with a correct incident dose applied; photobiomodulation therapy has no appreciable thermal effects in irradiated tissue [58, 60]. *In vivo* studies have shown that photobiomodulation can inhibit nerve function. Other alterations include local conduction blockage that give rise to pain relief and are reversible without side effects [61].

A recent study compared the effect of low-level laser therapy on post-operative pain after single-visit root canal retreatment on mandibular molars. It concluded that low-level laser therapy reduces post-operative pain after single-visit root canal retreatment of mandibular molars only four hours following the procedure [62]. Using biostimulation technique as low-level laser treatment enhances regeneration processes without significantly elevating body temperature. Biostimulation accelerates cellular metabolism, protein expression, cell migration, proliferation, and differentiation [38, 63].

An *in vitro* study evaluated the effect of 970 nm diode laser potentiality on proliferation of dental pulp stem cells with fiber optic tip 320nm in continuous wave and noncontact mode. The authors concluded that the diode laser power enhances proliferation of dental pulp stem cells [63]. In another study, the efficacy of diode laser in maturogenesis of immature teeth with necrotic pulps was evaluated. The study concluded that the diode laser has efficient antibacterial effect and can be used instead of TAP [64].

Recently, effect of biostimulation on the regenerative response of immature teeth with necrotic pulp and apical periodontitis was assessed in dogs. The results revealed that using of biostimulation by seven sessions at 808 nm diode laser and output power of 300 mW for 90 seconds showed marked increase in root length, thickness and decrease in apical diameter. Furthermore, it showed significant higher score of vital tissue infiltration and low inflammatory scores than non-biostimulated group. The authors concluded that biostimulation enhances the response of immature teeth with necrotic pulp and apical periodontitis to regenerative endodontic therapy (RET) improving root maturation [65]. Similarly, diode laser 980 nm can alternate DAP as a disinfection method of the root canal during RET for mature necrotic teeth [39]. Therefore diode laser can accelerate RET for both the patient and dentist in a single visit.

#### **Limitations of diode laser in root canal disinfection**

Power density, irradiation frequency, wavelength, cycle duration, and dentine thickness all have a direct impact on the photothermal effect of high power laser light [66]. Proper laser parameter selection not only has a therapeutic impact, but also aids in the prevention of damage to periodontal ligament

cells and alveolar bone. The photothermal effect, which generates a temperature increase of 10°C for more than one minute, has the potential to cause irreversible changes to the root dentine and surrounding tissues [67].

Several constraints must be considered while using lasers inside the root canal. To begin, laser light with a divergence angle of only 18 to 20 degrees is emitted in a straight line from the tip of an optical plain-ended fiber or a laser guide [68]. With such a unidirectional laser beam, it is difficult to achieve uniform irradiation of the whole root canal dentine surface [68, 69]. Furthermore, root canal preparation and retreatment with laser and plain fibers are risky in curved root canals due to the potential of generating ledges and holes [70, 71]. When irradiating root canal dentine, a helicoidal withdrawal motion from apical to coronal section is proposed to increase the surface area [72].

To ensure that the laser covered the whole interior wall of the canal, the laser fiber is moved in a circular motion from the apex to the coronal end and back again. The internal canal wall is not touched with the fiber tip to avoid melting the dentin and extending the heat effect to the region of the periodontal ligaments [39]. In order to provide a powerful laser beam, the fiber tip was also kept clean [73].

Intracanal irradiation is typically done in a pulsed fashion to reduce the danger of thermal harm to the external root surface, periodontal tissues, or underlying bone, which reduces post-operative pain and improves periapical healing [31, 39, 74]. Furthermore, new conical side-firing fiber tips with 80% lateral and 20% forward radiation cover the whole intracanal wall [75].

Another constraint is the safe use of lasers in the root canal, particularly when utilizing erbium lasers at ablative settings. Lasers may cause thermal injury to periradicular tissues through the open apical foramen [76].

## **Conclusion**

When compared to other types of lasers, diode laser has various advantages such as efficient antibacterial activity, low cost, ease of administration, high penetration power into dentine, and a smaller size device. Nonetheless, a few numbers of human randomized clinical trials with small sample sizes and substantial heterogeneity give insufficient support for its efficacy. Future research should look at the antibacterial effects of diode laser on different microbial endodontic species, as well as the role of diode laser in root canal disinfection in human teeth.

## References

- Hassanien EE, Abu-Seida AM, Hashem AA, Khanbash SS. Histologic evaluation of furcation perforation treated with mineral trioxide aggregate and bioaggregate. *Asian J Anim Sci.* 2015; 9: 148-56.
- Alazrag MA, Abu-Seida AM, El-Batouty KM, El Ashry SH. Adaptability, solubility and biocompatibility of TheraCal LC compared with MTA-Angelus and Biodentine as a furcation perforation repair material. *BMC Oral Health.* 2020; 20: 298. doi: 10.1186/s12903-020-01289-y.
- Okasha H, Abu-Seida AM, Hashem AA, El Ashry SH, Nagy MM. Inflammatory response and immunohistochemical characterization of experimental calcium silicate-based perforation repair material. *Int J Exp Pathol.* 2022; 103(4): 156-63. doi: 10.1111/iep.12439.
- Abboud KM, Abu-Seida AM, Hassanien EE, Tawfik HM. Biocompatibility of NeoMTA plus® versus MTA angelus as delayed furcation perforation repair materials in a dog model. *BMC Oral Health.* 2021, 21: 192. doi: 10.1186/s12903-021-01552-w.
- Badr M, Abu-Seid AM, Hashem AA, El Ashry SH. Immunohistochemical analysis of Biodentine versus MTA in repair of furcation perforation: an animal study. *G Ital Endod.* 2023; 37(1): 85-93. doi: 10.32067/GIE.2023.37.01.09
- Abbas KF, Tawfik H, Hashem AAR, Ahmed HMA, Abu-Seida AM, Refai HM. Histopathological evaluation of different regenerative protocols using chitosan-based formulations for management of immature non-vital teeth with apical periodontitis: In vivo study. *Aust endodo J.* 2020; 46:405-14. doi: 10.1111/aej.12426.
- El Ashry S, Abu-Seida AM, Bayoumi AA, Hashem AA. Regenerative potential of immature permanent non-vital teeth following different dentin surface treatments. *Exp Toxicol Pathol.* 2016; 68: 181-90. doi: 10.1016/j.etp.2015.12.001.
- El-Ashry SH, Abu-Seida AM, Al-Boghdady H, El-Batouty K, Abdel-Fattah M. The effect of different formulations of calcium hydroxide on healing of intentionally induced periapical lesions in dogs. *Pak Vet J.* 2013; 33: 48-52.
- El-Tayeb MM, Abu-Seida AM, El Ashry SH, El-Hady SA. Evaluation of antibacterial activity of propolis on regenerative potential of necrotic immature permanent teeth in dogs. *BMC Oral Health.* 2019; 19: 174. doi: 10.1186/s12903-019-0835-0.
- Abdelsalam N, Abu Seida AM, Fayyad D, Tawfik H. Radiographic and histopathologic outcomes of immature dog teeth with apical periodontitis after revascularization using propolis. *Saudi Endod J.* 2020; 10 (3): 199-207. DOI:10.4103/sej.sej\_178\_19
- Priyadarshini P, Gurunathan D. Comparative evaluation of quality of obturation using endoflas and endoflas powder with Aloe vera gel as obturating materials in primary mandibular molars: A double blinded randomized controlled trial. *J Complement Med Res.* 2020; 11: 203-13. doi: 10.5455/jcmr.2020.11.03.27
- Abu-Seida AM, Seif H. Aloe vera in dentistry: Current status and future prospects. *Int Arab J Dent.* 2023; 11 (2): 188-199.
- Goel P, Galhotra V, Makkar S, Mohan J, Bala N, Kaur T. An *in vitro* study comparing the antimicrobial efficacy of 0.2% chitosan, 3% sodium hypochlorite, 2% chlorhexidine against *Enterococcus faecalis*, alone and in conjunction with diode laser. *Int J Clin Pediatr Dent.* 2022; 15(1):109-15. doi: 10.5005/jp-journals-10005-2351.
- Abada HM, Hashem AAR, Abu-Seida AM, Nagy MM. The effect of changing apical foramen diameter on regenerative potential of mature teeth with necrotic pulp and apical periodontitis. *Clin Oral Investig.* 2022; 26(2): 1843-53. doi: 10.1007/s00784-021-04159-1.
- El Halaby HM, Abu-Seida AM, Fawzy MI, Farid MH, Bastawy HA. Evaluation of regenerative potential of dentin conditioning and naturally derived scaffold for necrotic immature permanent teeth in a dog model. *Int J Exp Pathol.* 2020, 101(6): 264-76. doi: 10.1111/iep.12372.
- Elkhashab RA, Mahran AH, Badr M, Abu-Seida AM. Histopathology and immunohistochemical reactions of Nano-White MTA versus NeoMTATM Plus® and MTA Angelus® as immediate furcation perforation repair materials in a dog model. *G Ital Endod.* 2024; 38(1): 16-26. https://doi.org/10.32067/GIE.2023.37.01.27
- Prabhakar J, Senthilkumar M, Priya MS, Mahalakshmi K, Sehgal PK, Sukumaran VG. Evaluation of antimicrobial efficacy of herbal alternatives (Triphala and green tea polyphenols), MTAD, and 5% sodium hypochlorite against *Enterococcus faecalis* biofilm formed on tooth substrate: an in vitro study. *J Endod.* 2010; 36(1): 83-6. doi: 10.1016/j.joen.2009.09.040.
- Jones DS, Gorman SP, Mc Cafferty DF, Woolfson AD. The effects of three non-antibiotic, antimicrobial agents on the surface hydrophobicity of certain micro-organisms evaluated by different methods. *J Applied Bacteriol.* 1991; 71: 218-27. doi: 10.1111/j.1365-2672.1991.tb04451.x.

19. Tyagi SP, Sinha DJ, Garg P, Singh UP, Mishra CC, Nagpal R. Comparison of antimicrobial efficacy of propolis, *Morinda citrifolia*, *Azadirachta indica* (Neem) and 5% sodium hypochlorite on *Candida albicans* biofilm formed on tooth substrate: An in-vitro study. *J Conserv Dent*. 2013; 16(6):532-5. doi: 10.4103/0972-0707.120973.
20. Yadav P, Chaudhary S, Saxena RK, Talwar S, Yadav S. Evaluation of antimicrobial and antifungal efficacy of Chitosan as endodontic irrigant against *Enterococcus faecalis* and *Candida albicans* biofilm formed on tooth substrate. *J Clin Exp Dent*. 2017 9(3):e361-e7. doi: 10.4317/jced.53210.
21. Kimura Y, Wilder-Smith P, Matsumoto K. Laser in endodontics: A review. *Int Endod J*. 2000; 33: 173-85.
22. Sen BH, Piskin B, Demirci T. Observation of bacteria and fungi in infected root canals and dentinal tubules by SEM. *Endod Dent Traumatol*. 1995; 11: 6-9. doi: 10.1111/j.1600-9657.1995.tb00671.x.
23. Mohammadi Z. Laser applications in endodontics: an update review. *Int Dent J*. 2009; 59: 35-46.
24. Sadony DM, Montasser K. Evaluation and comparison between the bactericidal effect of diode laser irradiation (970 nm) and silver nanoparticles on *Enterococcus faecalis* bacterial strain (an in vitro study). *Bull Nat Res Cent*. 2019; 43:155. DOI: 10.1186/s42269-019-0188-5
25. Morsy DA, Negm M, Diab A, Ahmed G. Postoperative pain and antibacterial effect of 980 nm diode laser versus conventional endodontic treatment in necrotic teeth with chronic periapical lesions: A randomized control trial. *F1000Res*. 2018; 7: 1795. doi:10.12688/f1000research.16794.1.
26. Maiti N, Benedicenti S, Henna, Ambesh S, Gudapati S, Mustafa M, Agarwal K. Assessment of efficiency of diode laser in root canal disinfection: An original research. *J Pharm Bioallied Sci*. 2022; 14(Suppl 1): S248-S250. doi: 10.4103/jpbs.jpbs\_710\_21.
27. Hyder T. Diode lasers in dentistry: Current and emerging applications. *J Pakistan Dent*. 2022; 31:100-5.
28. Jurić IB, Anić I. The use of lasers in disinfection and cleanliness of root canals: A review. *Acta Stomatol Croat*. 2014; 48(1): 6-15
29. Stanusi AS, Camen A, Diaconu OA, Albulescu DM, Turcu A, Nicola A, Petcu C, Gheorghită LM, Georgescu RV, Tuculină MJ. Statistical study on the use of dental laser in endodontics. *Rom J Oral Rehab*. 2022; 14:86-96.
30. Kaplan T, Sezgin GP, Kaplan SS. Effect of a 980-nm diode laser on post-operative pain after endodontic treatment in teeth with apical periodontitis: a randomized clinical trial. *BMC Oral Health*. 2021; 21(1): Article number: 41. doi: 10.1186/s12903-021-01401-w.
31. Anagnostaki E, Mylona V, Parker S, Lynch E, Grootveld M. Systematic review on the role of lasers in endodontic therapy: Valuable adjunct treatment? *Dent J (Basel)*. 2020; 8: 63. doi: 10.3390/dj8030063.
32. Bytyqi A, Aliu X, Barani M, Stubljarić D, Jukić T, Starc A, Krasniqi S. Disinfection of infected artificial dental periapical lesions with diode laser: An in vitro study. *Med Sci Monit Basic Res*. 2021; 27: e932492-7. doi: 10.12659/MSMBR.932492.
33. Shanshan Dai, Gang Xiao, Ning Dong, Fei Liu, Shuyang He, Qingyu Guo. Bactericidal effect of a diode laser on *Enterococcus faecalis* in human primary teeth- An in vitro study. *BMC Oral Health*. 2018; 18: 154. doi: 10.1186/s12903-018-0611-6.
34. Asnaashari M, Sadeghian A, Hazrati P. The effect of high-power lasers on root canal disinfection: A systematic review. *J Lasers Med Sci*. 2022; 13: e66. doi: 10.34172/jlms.2022.66.
35. Pelozo LL, Silva-Neto RD, Salvador SL, Sousa-Neto MD, Souza-Gabriel AE. Adjuvant therapy with a 980-nm diode laser in root canal retreatment: A randomized clinical trial with 1-year follow-up. *Lasers Med Sci*. 2023; 38(1): 77. doi: 10.1007/s10103-022-03659-0.
36. Alves DRS, Decucio DA, de Alencar AHG, Estrela CR, de Souza JB, Pinheiro ALB, Estrela C. Effect of low-power diode laser on infected root canals. *Braz Dent J*. 2022; 33(3): 8-17. doi: 10.1590/01036440202204999
37. Mishra A, Koul M, Abdullah A, Khan N, Dhawan P, Bhat A. Comparative evaluation of antimicrobial efficacy of diode laser (Continuous Mode), diode laser (Pulse Mode), and 5.25% of Sodium hypochlorite in disinfection of root canal: A short study. *Int J Clin Pediatr Dent*. 2022; 15(5): 579-83. doi: 10.5005/jp-journals-10005-2444.
38. Sohrabi K, Sooratgar A, Zolfagharnasab K, Kharazifard MJ, Afkhami F. Antibacterial activity of diode laser and sodium hypochlorite in *Enterococcus faecalis*-contaminated root canals. *Iran Endodo J*. 2016 11(1): 8-12. doi: 10.7508/iej.2016.01.002.
39. Eldessoky AE, Khalefa MM, Abu-Seida AM. Regenerative endodontic therapy in mature teeth with necrotic pulp and apical periodontitis using two disinfection protocols. *BMC Oral Health*, 2023, 23(1): 163. doi: 10.1186/s12903-023-02863-w.
40. Eldessoky AE, Khalefa MM, Abu-Seida, AM. Comparison of antibacterial activity of diode laser 980 nm and double antibiotic paste during regenerative



- endodontic therapy of mature necrotic teeth. *G Ital Endod.* 2022; 36(2): 10-18. <https://doi.org/10.32067/GIE.2022.36.02.10>
41. Ando Y, Aoki A, Watanabe H, Ishikawa I. Bactericidal effect of erbium: YAG laser on periodontopathic bacteria. *Lasers Surg Med.* 1996; 19:190-200. doi: 10.1002/(SICI)1096-9101(1996)19:2<190::AID-LSM11>3.0.CO;2-B.
  42. Moritz A, Jakolitsch S, Goharkhay K, Schoop U, Kluger W, Mallinger R, Sperr W, Georgopoulos A. Morphologic changes correlating to different sensitivities of *Escherichia coli* and *Enterococcus faecalis* to Nd:YAG laser irradiation through dentin. *Lasers Surg Med.* 2000; 26: 250-61. doi: 10.1002/(sici)1096-9101(2000)26:3<250::aid-lsm2>3.0.co;2-h.
  43. Jung H. A Generalized concept for cell killing by heat. *Radiat Res.* 1986; 106: 56-72.
  44. Dawasaz AA. In vivo efficacy of diode laser as a monotherapy in root canal disinfection: a systematic review and meta-analysis. *Photobiomodul Photomed Laser Surg.* 2022; 40(1):59-70. <https://doi.org/10.1089/photob.2021.0073>.
  45. Prażmo EJ, Godlewska AR, Sałkiewicz M, Mielczarek AB. Effects of 980 nm diode laser application protocols on the reduction of *Enterococcus faecalis* intracanal biofilm: An in vitro study. *Dent Med Probl.* 2017; 54:333-8. DOI 10.17219/dmp/77898
  46. El-Tayeb MM, Nabeel M. Effect of two laser systems on root canal disinfection: An in vitro study *Ain Shams Dent J.* 2021; XXIV (1): 73-9. doi:10.15644/asc48/1/1.
  47. Bago I, Plečko V, Gabrić Pandurić D, Schauerperl Z, Baraba A, Anić I. Antimicrobial efficacy of a high-power diode laser, photo-activated disinfection, conventional and sonic activated irrigation during root canal treatment. *Int Endod J.* 2013; 46(4): 339-47. 10.1111/j.1365-2591.2012.02120.x.
  48. Huang Q, Li Z, Lyu P, Zhou X, Fan Y. Current applications and future directions of lasers in endodontics: A narrative review. *Bioengineering (Basel).* 2023; 10(3): 296. doi: 10.3390/bioengineering10030296.
  49. Cope AL, Francis N, Wood F, Chestnutt IG. Systemic antibiotics for symptomatic apical periodontitis and acute apical abscess in adults. *Cochrane Database Syst Rev.* 2018; 9:CD010136. doi: 10.1002/14651858.CD010136.pub3.
  50. Van Nieuwenhuysen JP, D'Hoore W, Leprince JG. What ultimately matters in root canal treatment success and tooth preservation: A 25-year cohort study. *Int Endod J.* 2023; 56(5):544-57. doi: 10.1111/iej.13895.
  51. Pak JG, White SN. Pain prevalence and severity before, during, and after root canal treatment: A systematic review. *J Endod.* 2011; 37: 429-38. doi: 10.1016/j.joen.2010.12.016.
  52. He WL, Li CJ, Liu ZP, Sun JF, Hu ZA, Yin X, Zou SJ. Efficacy of low-level laser therapy in the management of orthodontic pain: A systematic review and meta-analysis. *Lasers Med Sci.* 2013; 28:1581-9. doi: 10.1007/s10103-012-1196-y.
  53. Pires D, Xavier M, Araújo T, Silva JA, Aimbire F, Albertini R. Low-level laser therapy (LLLT; 780 Nm) acts differently on mRNA expression of anti- and pro-inflammatory mediators in an experimental model of collagenase-induced tendinitis in rat. *Lasers Med Sci.* 2011; 26: 85-94. doi: 10.1007/s10103-010-0811-z.
  54. Genc Sen O, Kaya M. Effect of root canal disinfection with a diode laser on postoperative pain after endodontic retreatment. *Photobiomodul Photomed Laser Surg.* 2019; 37(2): 85-90. doi: 10.1089/photob.2018.4539.
  55. Kaplan T, Kaplan SS, Sezgin GP. The effect of different irrigation and disinfection methods on post-operative pain in mandibular molars: a randomised clinical trial. *BMC Oral Health.* 2022; 22(1): 601. doi: 10.1186/s12903-022-02651-y.
  56. Metin R, Tatli U, Evlice B. Effects of low-level laser therapy on soft and hard tissue healing after endodontic surgery. *Lasers Med Sci.* 2018; 33(8):1699-706. doi: 10.1007/s10103-018-2523-8.
  57. Rubio F, Wienecke F, Arnabat-Domínguez J, Betancourt P. Photobiomodulation therapy and endodontic treatment of teeth with apical periodontitis using 940-nm diode laser. Report of two cases. *J Clin Exp Dent.* 2022; 14 (3): 298-302. doi: 10.4317/jced.59058.
  58. Abdelgawad LM, Salah N, Sabry D, Abdelgwad M. Efficacy of photobiomodulation and vitamin d on odontogenic activity of human dental pulp stem cells. *J Lasers Med Sci.* 2021; 12: e30. doi: 10.34172/jlms.2021.30.
  59. Özverel CS, Islam A, Yılmaz HG. The biostimulative effectiveness of photobiomodulation therapy application on thawed dental pulp stem cells. *J Innov Opt Health Sci.* 2021; 14 (3): 2150006. <https://doi.org/10.1142/S1793545821500061>.
  60. Malthiery E, Chouaib B, Hernandez-Lopez AM, Martin M, Gergely C, Torres JH, Cuisinier FJ, Collart-Dutilleul PY. Effects of green light photobiomodulation on dental pulp stem cells enhanced proliferation and improved wound healing by cytoskeleton reorganization and cell softening. *Lasers Med Sci.* 2021; 36 (2): 437-45. doi: 10.1007/s10103-020-03092-1.

61. Chow RT, Armati PJ. Photobiomodulation: Implications for anesthesia and pain relief. *Photomed Laser Surg.* 2016; 34 (12): 599-609. doi: 10.1089/pho.2015.4048.
62. Fazlyab M, Shahmirzadi S, Esnaashari E, Azizi A, Moshari AA. Effect of low-level laser therapy on postoperative pain after single-visit root canal re-treatment of mandibular molars: A randomized controlled clinical trial. *Int Endod J.* 2021; 54 (11): 2006-15. doi: 10.1111/iej.13608.
63. Abo El-Dahab MM, Gheith M, Soliman NL, Aly RM. Effect of diode laser potentiality on proliferation of dental pulp stem cells (*in vitro study*). *Bull Nati Res Cent.* 2020; 44:170. <https://doi.org/10.1186/s42269-020-00414-9>
64. Mekkawi O, Kataia A, Ali MM. Evaluation of the efficacy of diode laser in maturogenesis of immature teeth with necrotic pulps: an *in vivo* study “part one” *Ind J Pub Health Res Dev.* 2020; 11:1387. <https://doi.org/10.37506/ijphrd.v11i3.1721>.
65. Fouad E, Fawzy M, Elhousiny M, Saafan A, Abdel Ghani S. Effect of biostimulation on response of immature teeth with necrotic pulp and apical periodontitis to regenerative endodontic therapy in immature dogs teeth. *Al-Azhar Dent J Girl.* 2022; 9 (2): 265-76. DOI: 10.21608/ADJG.2022.103757.1445.
66. Bergmans L, Moisiadis P, Teughels W, Van Meerbeek B, Quirynen M, Lambrechts P. Bactericidal effect of Nd:YAG laser irradiation on some endodontic pathogens *ex vivo*. *Int Endod J.* 2006; 39: 547-57. DOI: 10.1111/j.1365-2591.2006.01115.x .
67. Eriksson AR, Albrektsson T. Temperature threshold levels for heatinduced bone tissue injury: A vital-microscopic study in the rabbit. *J Prosthet Dent.* 1983; 50: 101-7.
68. Stabholz A, Sahar-Helft S, Moshonov J. Lasers in endodontics. *Dent Clin North Am.* 2004; 48(4):809-32. 10.1016/j.cden.2004.05.012 .
69. Matsuoka E, Yonaga K, Kinoshita J, Kimura Y, Matsumoto K. Morphological study on the capability of Er:YAG laser irradiation for root canal preparation. *J Clin Laser Med Surg.* 2000; 18(4): 215-9. 10.1089/10445470050144074.
70. Jahan KM, Hossain M, Nakamura Y, Yoshishige Y, Kinoshita J, Matsumoto K. An assessment following root canal preparation by Er,Cr: YSGG laser irradiation in straight and curved roots, *in vitro*. *Lasers Med Sci.* 2006; 21(4): 229-34. 10.1007/s10103-006-0401-2.
71. Gutknecht N, van Gogswaardt D, Conrads G, Apel C, Schubert C, Lampert F. Diode laser radiation and its bactericidal effect in root canal wall dentin. *J Clin Laser Med Surg.* 2000; 18 (2): 57-60.
72. George R, Walsh JL. Laser fiber-optic modifications and their role in endodontics. *J Laser Dent.* 2012; 20(1): 24-30.
73. Nageh M, Ahmed GM, El-Baz AA. Assessment of regaining pulp sensibility in mature necrotic teeth using a modified revascularization technique with platelet-rich fibrin: a clinical study. *J Endod.* 2018; 44:1526-33. <https://doi.org/10.1016/j.joen.2018.06.014>.
74. Kapasi A, Kapasi F. When less is more - an update on low level laser therapy applications in endodontics for predictable outcomes - A review article. *Acta Sci Dent Sci.* 2019; 3: 83-93.
75. Komori T, Yokoyama K, Matsumoto Y, Matsumoto K. Erbium:YAG and holmium:YAG laser root resection of extracted human teeth. *J Clin Laser Med Surg.* 1997; 15(1): 9-13.
76. Hellingwerf KJ, Hoff WD, Crielaard W. Photobiology of microorganisms: how photosensors catch a photon to initialize signalling. *Mol Microbiol.* 1996; 21(4): 683-93. 10.1046/j.1365-2958.1996.411402.x