

Journal section: *Biomaterials and Bioengineering in Dentistry*

Publication Types: *Research*

doi:10.4317/medoral.18671

<http://dx.doi.org/doi:10.4317/medoral.18671>

## A new 3D method for measuring cranio-facial relationships with cone beam computed tomography (CBCT)

Natalia Zamora <sup>1</sup>, Rosa Cibrián <sup>2</sup>, Jose-Luis Gandia <sup>3</sup>, Vanessa Paredes <sup>4</sup>

<sup>1</sup> Orthodontist, Research Assistant Professor, Department of Orthodontics, Faculty of Medicine and Dentistry, University of Valencia, Valencia, Spain

<sup>2</sup> Associate Professor, Department of Physiology, Faculty of Medicine and Dentistry, University Valencia, Valencia, Spain

<sup>3</sup> Orthodontist, Professor and Department Chair of Postgraduate Orthodontics Masters Course, Department of Orthodontics, Faculty of Medicine and Dentistry, University of Valencia, Valencia, Spain

<sup>4</sup> Orthodontist and PhD Associate Professor, Department of Orthodontics, Faculty of Medicine and Dentistry, University of Valencia, Valencia, Spain

### Correspondence:

*Clinica Odontológica*

*Departamento Ortodoncia 5º piso*

*c/ Gasco Oliag nº1*

*46010, Valencia, Spain*

*natalia.zamora@uv.es*

Zamora N, Cibrián R, Gandia JL, Paredes V. A new 3D method for measuring cranio-facial relationships with cone beam computed tomography (CBCT). *Med Oral Patol Oral Cir Bucal*. 2013 Jul 1;18 (4):e706-13. <http://www.medicinaoral.com/medoralfree01/v18i4/medoralv18i4p706.pdf>

Received: 19/07/2012

Accepted: 17/12/2012

Article Number: 18671 <http://www.medicinaoral.com/>  
© Medicina Oral S. L. C.I.F. B 96689336 - pISSN 1698-4447 - eISSN: 1698-6946  
eMail: [medicina@medicinaoral.com](mailto:medicina@medicinaoral.com)

### Indexed in:

Science Citation Index Expanded  
Journal Citation Reports  
Index Medicus, MEDLINE, PubMed  
Scopus, Embase and Emcare  
Indice Médico Español

### Abstract

**Objectives:** CBCT systems, with their high precision 3D reconstructions, 1:1 images and accuracy in locating cephalometric landmarks, allows us to evaluate measurements from craniofacial structures, so enabling us to replace the anthropometric methods or bidimensional methods used until now. The aims are to analyse cranio-facial relationships in a sample of patients who had previously undergone a CBCT and create a new 3D cephalometric method for assessing and measuring patients. **Study Design:** 90 patients who had a CBCT (i-Cat®) as a diagnostic register were selected. 12 cephalometric landmarks on the three spatial planes (X,Y,Z) were defined and 21 linear measurements were established. Using these measurements, 7 triangles were described and analysed. With the sides of the triangles: (CdR-Me-CdL); (FzR-Me-FzL); (GoR-N-GoL); and the Gl-Me distance, the ratios between them were analysed. In addition, 4 triangles in the mandible were measured (body: GoR-DB-Me and GoL-DB-Me and ramus: KrR-CdR-GoR and KrL-CdL-GoL). **Results:** When analyzing the sides of the CdR-Me-CdL triangle, it was found that the 69.33% of the patients could be considered symmetric.

Regarding the ratios between the sides of the following triangles: CdR-Me-CdL, FzR-Me-FzL, GoR-N-GoL and the Gl-Me distance, it was found that almost all ratios were close to 1:1 except between the CdR-CdL side with respect the rest of the sides. With regard to the ratios of the 4 triangles of the mandible, it was found that the most symmetrical relationships were those corresponding to the sides of the body of the mandible and the most asymmetrical ones were those corresponding to the base of such triangles. **Conclusions:** A new method for assessing cranio-facial relationships using CBCT has been established. It could be used for diverse purposes including diagnosis and treatment planning.

**Key words:** *Craniofacial relationship, CBCT, 3D cephalometry.*

## Introduction

Facial symmetry, is one element of bodily symmetry, including fluctuating asymmetry. Along with traits such as averageness and youthfulness it influences judgements of aesthetic traits of physical attractiveness and beauty, and is associated with fitness-linked traits including health(1). It is also hypothesized as a factor in both interpersonal attraction and interpersonal chemistry.

British orthodontist R.J. Edler (2) cited research supporting the claim that bilateral symmetry is an important indicator of freedom from disease, and worthiness for mating. Random differences between the two sides, known in biological terms as fluctuating asymmetry, and not deliberate asymmetrical structures found in some animals, develops throughout the lifespan of the individual and is a sign of the phenotype being subjected to some levels of stress.

The ability to cope with these pressures is partly reflected in the levels of symmetry. "Perfect beauty" exists for each person when all the exact relationships of beauty are met. A higher degree of symmetry indicates a better coping system for environmental factors. While the visible signs of this may not be particularly apparent, it is thought that they have at least an unconscious effect on people's perception of their beauty.

Ricketts (3) in 1982, described "the divine proportion" (analysis of the divine rectangle, pentagon and triangle) by undertaking a facial analysis based on both, facial width and height ratios in soft tissues. Likewise, he found those divine proportions on the skeletal level, both sagittal and frontal. On the sagittal level, he found eight divine proportions and on the frontal level three divine proportions. In this way, he found "the divine proportion" by locating various frontal and lateral measurements on the cranium and in the soft tissues of the human face.

Perrett et al.(4) showed that beauty consists of a more complex concept, as they discovered in a study that the average number of faces of all the women studied was less attractive than the average of the faces that separately were considered attractive.

Jefferson (5) revised divine proportions both in soft and hard tissues claiming that the normal face was the most attractive given that it was made up of "divine proportions" and this provided it with a more healthy, strong and fertile look.

Hönn and Göz (6) determined to what extent the beauty of a face could be measured and how symmetry provided a tool for doing so. In undertaking these measurements of facial proportions, they found that various authors stated that anthropometric methods were preferable to cephalometric measurements since they involved non-invasive methods and provided a three dimensional assessment of the structures under study.

Despite these claims of Hönn and Göz (6), with the appearance of Cone Beam Computed Tomography (CBCT)

with its high precision (1:1) and accurately locating the cephalometric points bringing them over more to the 3D reality. (7-16) it is now possible to assess different measurements and facial relationships in patients.

There have been several studies that have attempted bringing them over more to the 3D reality to design cephalometric analysis in 3D, some of them adapting conventional 2D cephalometric analysis to 3D, using the visualization of craniofacial structures in three dimensions, while others have designed new measurements for the study of cleftpalate patients, patients with asymmetries or orthognatic surgical candidates, or as part of pilot studies for other applications.

The development and implement of this 3D cephalometry will mean progression in daily practice since measurements will be done directly in three-dimensional reconstruction of the patient with the many advantages that this implies. It is a matter of fact that a 3D structure always gives us a more realistic view from what we face. The aims of this study were to describe and analyse the relationships between different measurements corresponding to craniofacial structures in a sample of patients who had undergone CBCT.

## Material and Methods

A study approved by the ethical committee of the Clinical University Hospital of the University of Valencia was undertaken. An initial sample consisted of 122 CBCTs of patients who had previously undergone a full cranial scan at the University of Valencia, Spain. The CBCTs used were obtained from the data base of those patients who had previously undergone such diagnostic tool because different reasons: included teeth like canines or third molars, agenesis or supernumerary teeth, all without moderate to severe skeletal asymmetries. No patient was scanned because of the purpose of the present study.

From this initial sample, 32 CBCTs were excluded for the following reasons:

9 Patients with open mouth.

11 Patients with multiple missing teeth.

12 Patients with moderate to severe asymmetries.

The final sample was composed of 90 CBCTs of patients between the ages of 8 and 50. The mean age was (18.05 years  $\pm$  8.69).

Each of the patients had undergone a scan using the i-CAT® (Imaging Sciences International, Hatfield, Pa) equipment. This CBCT device uses an amorphous silicon flat panel sensor to capture the fields of view (FOV). The field of view (FOV) employed was the portrait mode that captures data in extended FOV mode and includes the full head of 170 mm in height x 230 mm in diameter with a scanning time of 8.9 seconds. It was set at medium quality and high resolution mode. It generates a total of 326 slices, with an image matrix size of 400x400. The voxel size is of 0.4 x 0.4 x 0.4 mm. The focal size is

0.5mm y and the size of its base is 119x142 cm. Tube voltage is 120 kV and its intensity 23.87 mAs. The size of the data files generated is in the order of 35 megabytes.

The raw data and the slices obtained from the CBCTs were imported to the InVivo5® software (Anatomage, San Jose, CA) which was used to visualize the slices and three dimensional images that are obtained from a CBCT. This is where the three-dimensional reconstruction of the DICOM images (Digital Imaging and Communications in Medicine) is made.

In order to undertake this particular study, 12 cephalometric points were defined on each of the three spatial planes (X, Y, Z) (Table 1). The procedure for locating each point requires the selection of the most appropriate view or plane (sagittal, coronal, axial) and then adjusting that point on the other planes for better accuracy. Employing this location method, all the spatial positions of each point have been pinpointed on each of these axes as numerical values.

In a previous study undertaken by the same authors (17), the reproducibility and reliability in locating ce-

phalometric points was analyzed. To assess the reproducibility on landmark location and the differences in the measurements, two observers with the same background and six years of experience in the field of orthodontics located 41 landmarks at three separate times. A total of 11,070 data were processed using the 15.0 SPSS statistical package®. To discover the reproducibility of the method on landmark location, an ANOVA analysis was undertaken using two factors of variation: time (t1, t2 and t3) and observers (Ob1 and Ob2) for each axis (X, Y and Z) and landmark. The order of the CBCT scans submitted to the observers (Ob1, Ob2) at t1, t2, t3 were different and randomly allocated. The intraclass correlation coefficient (ICC) was calculated. The results of the study showed that both intra-and inter-observer reliability were high, both being ICC ≥ 0.99. It was also found that the best ICC frequency was on axis Z.

Once the cephalometric points have been adapted to the 3D reality, that is, have been defined on each of the spa-

**Table 1.** Definition of the three spatial planes of the 12 points used in this study. Anterior point (AP), midpoint (MP), posterior point (PP), lowest point (LP), upper point (UP), highest point (HP).

Landmark	Anatomical definition	Sagittal or lateral view	Coronal or frontal view	Axial view
<b>Nasion (N)</b>	Most AP of the frontonasal suture	Most AP	MP	Most AP and MP of the anterior contour
<b>Menton (Me)</b>	LP of the mandibular symphysis	LP	LP	LP and MP
<b>Right and left Condylion (CdR and CdL)</b>	UP of the head of the condyle	UP and most PP	Most UP and MP	Most PP
<b>Right and left Gonion (GoR and GoL)</b>	Most PP of the posterior edge of the ramus. Bisection of two tangents: posterior edge of the ramus and lower part of the body	Most PP	Most PP and MP	Most PP determined superoinferiorly by the other two slices
<b>Right and left frontozygomatic suture (FzR and FzL)</b>	Most antero-lower point of the frontozygomatic suture	Most AP	MP	MP determined supero-inferiorly by the other two slices
<b>Glabella (Gl)</b>	Most prominent point of the frontal bone measured within the Sagittal plane. Located between the supraciliar arches.	Most AP	MP	AP and MP of the anterior contour
<b>Right and left coronoid process point (KrR and KrL)</b>	HP of the Coronoid process	Most HP	Most HP and MP	MP
<b>Bonwill dental Point (DB)</b>	Intersection point between upper and lower central incisors	Most MP and AP	MP	MP

tial planes (X, Y, Z) and have been duly located in their correct position, the next step to design a 3D cephalometric analysis of the craniofacial relationships, was to define 21 linear measurements (Table 2).

After that, using those 21 linear measurements 7 triangles were created (Fig. 1).

2 Triangles of the mandibular body: formed by the DB-Me-GoR points on the right side and the DB-Me-GoL points on the left side (Fig. 1).


An Excel® sheet, version 12.0 for Windows (Microsoft Corporation), was created to introduce all the variables and measurements obtained in the 3D cephalometric

**Table 2.** Description, global mean (mm) and standard deviation of the mean (mm) of the 21 linear measurements used in this study.

Linear Measurements		Description	Global Mean (mm)	Standard Deviation (mm)
1	Gl-Me	Distance between point Gl and point Me	118,67	7,93
2	CdR-CdL	Distance between point CdR and point CdL	99,74	9,71
3	CdR-Me	Distance between point CdR and point Me	114,65	9,61
4	CdL-Me	Distance between point CdL and point Me	113,18	9,71
5	FzR-Me	Distance between point FzR and point Me	117,12	8,86
6	FzL-Me	Distance between point FzL and point Me	116,75	8,53
7	FzR-FzL	Distance between point FzR and point FzL	93.40	3.56
8	Na-GoR	Distance between point Na and point GoR	116,27	7,96
9	Na-GoL	Distance between point Na and point GoL	115,81	7,89
10	GoR-GoL	Distance between point GoR and point GoL	88.70	5.87
11	DB-Me	Distance between point DB and point Me	39,82	4,52
12	DB-GoR	Distance between point DB and point GoR	81,63	6,96
13	DB-GoL	Distance between point DB and point GoL	81,58	7,14
14	Me-GoR	Distance between point Me and point GoR	85,40	9,58
15	Me-GoL	Distance between point Me and point GoL	85,30	10,10
16	KrR-CdR	Distance between point KrR and point CdR	35,78	4,70
17	KrL-CdL	Distance between point KrL and point CdL	36,94	4,68
18	KrR-GoR	Distance between point KrR and point GoR	49,00	6,44
19	KrL-GoL	Distance between point KrL and point GoL	49,86	5,76
20	CdR-GoR	Distance between point CdR and point GoR	53.28	6.05
21	CdL-GoL	Distance between point CdL and point GoL	51.34	5.87

1 Triangle formed by the CoR-Me-CoL points (Fig. 1).  
 1 Triangle formed by the FzR- Me-FzL points (Fig. 1).  
 1 Triangle formed by the GoR- N-GoL points (Fig. 1).  
 2 Triangles of the mandibular ramus: formed by the KrR-CoR-GoR points on the right side and the KrL-CoL-GoL points on the left side (Fig. 1).

analysis of the 90 patients After that all data were introduced in version 17.0 of the statistical package SPSS® for Windows (SPSS Inc., Chicago, IL). The means and standard deviations for each of the measurements were obtained. Finally, the correlations between variables were found using the Bonferroni and Scheffe analyses.



**Fig. 1.** a) CdR-Me-CdL triangle, b) FzR-Me-FzL triangle, GoR-N-GoL triangle and GI-Me distance c) Triangle of the mandibular ramus: CdR-CrR-GoR d) Triangles of the mandibular body: DB-Me-GoR; DB-Me-GoL.

**Results**

Once the 3D analysis was created, all the values of the different linear measurements and triangles were analysed. Table 2 shows the means and overall typical deviations for the 21 linear measurements used in this study.

To eliminate bias in the study, and to be able to compare between subjects of different ages we only analysed angles and ratios between distances.

Firstly, we studied the relationships between the sides of the CdR-Me-CdL triangle (Fig. 1); this triangle gives information about the shape and position of the mandible. To consider the mandible symmetric, we have analyzed the value of the lower angle of the triangle, the CdR-Me-CdL angle.

By doing this, we were searching for the symmetry of the left and right sides of the mandible.

When dividing this particular angle into two semiangles (Fig. 2), to be considered symmetrical, they should measure the same.


A midpoint in between the CdR-CdL line (CdM point) was determined. Then, a line starting on this point and finishing in Me was used to calculate the values of the two semiangles: CdR-Me-CdM and CdL-Me-CdM (Fig. 2).

Figure 2 shows the relationship between the semiangles. The diagonal line represents perfect symmetry between sides. Dotted lines included those patients with  $\leq 4^\circ$  of discrepancy between right and left side.

In our study, we observed that with this approach the 69.33% of the patients could be considered symmetric.

Secondly, we wanted to study if there was any type of relationship between the sides of the triangles formed by the CdR-Me-CdL; GoR-N-GoL; FzR-Me-FzL points and the GI-Me distance. Some authors (3,6) in their studies and reviews stated that, in order to reach symmetry, a ratio of 1:1 between all sides must be achieved.

In our study, to measure this characteristic, the minimum, maximum and mean value of the different ratios between sides was assessed (Table 3).



**Fig. 2.** a) Distribution of semiangles (CdR-Me-CdM; CdL-Me-CdM) of the CdR-Me-CdL triangle. b) relationship between the semiangles. Diagonal line: represents perfect symmetry between sides. Dotted line: included those patients with  $\leq 4^\circ$  of discrepancy between right and left side.

We observed that, in our sample, all the ratios studied between the different sides that form these triangles, presented values close to 1:1, indicating, thus, that those distances were similar to each other. Only the CdR-CdL distance did not approach a 1:1 ratio when compared with the values of the other distances.

We have also compared the mean difference (in mm) between right and left sides of the following distances: FzR-



**Table 3.** Minimum, maximum and mean value of the ratios between distances. The most asymmetrical ratios are highlighted in grey.

Distance Ratio	Minimum	Maximum	Mean ratio between sides
CdR-CdL_FzR-Me	0.52	1.04	0.844
CdR-CdL_FzL-Me	0.51	1	0.845
CdR-CdL_GoR-N	0.54	1.02	0.848
CdR-CdL_GoL-N	0.53	1.02	0.849
FzR-Me_GoL-N	0.93	1.11	1.007
FzL-Me_GoL-N	0.91	1.1	1.006
FzR-Me_GoR-N	0.91	1.11	1.006
FzL-Me_GoR-N	0.89	1.1	1.005
Gl-Me_CdR-CdL	1.01	1.93	1.212
Gl-Me_CdR-Me	0.93	1.11	1.032
Gl-Me_CdL-Me	0.93	1.16	1.036
Gl-Me_FzR-Me	0.93	1.06	1.014
Gl-Me_FzL-Me	0.94	1.07	1.015
Gl-Me_GoR-N	0.91	1.11	1.019
Gl-Me_GoL-N	0.93	1.09	1.02
CdL-Me_CdR-CdL	0.96	1.8	1.175
CoR-Me_CdR-CdL	0.99	1.83	1.171
CdR-Me_CdL-Me	0.96	1.04	1.004

Me; FzL-Me; GoR-N; GoL-N; CdR-Me y CdL-Me. According to this, 66.66% of the cases studied present a maximum discrepancy of 5mm between right and left sides.

Thirdly, we analysed the ratios between the ramus and body triangles of the mandible (KrR-CdR-GoR and KrL-CdL-GoL) and (DB-Me-GoR and DB-Me-GoL). (Figs. 1).

A total of 9 relationships between these last four triangles were recorded. Table 4 shows the values of all ratios. In dark grey the most symmetrical relationships have been represented. Those corresponding to the sides of the body of the mandible (DB-GoR with DB-GoL and Me-GoR with Me-GoL) are considered the most symmetrical ones. In light grey, the most asymmetrical relationships, those corresponding to the bases of such triangles, are represented (relationship between KrL-CdL and DB-Me and relationship between KrR-CdR and DB-Me).

## Discussion

To carry out this study, records of patients who had previously undergone a CBCT as a diagnostic tool because some additional alteration (agenesia, supernumerary teeth, included teeth...) were used; so undertaking a CBCT was justified. Patients with moderate and severe asymmetries were not included in this study. Despite the fact that the sample was large, including 90 patients,

the drawback of carrying out a study of these characteristics is that irradiating actual patients only for research purposes is not justified.

In the present study, only angles and ratios have been assessed. Consequently, no age groups have been made as there is no difference between ages when measuring ratios and angles (3.4.6).

According to some authors who studied the symmetry of the skull, (3,4,6) it was found that in skulls with perfect symmetry the mandible must respond to geometric triangulation; in our study we examined the mandible as a triangle (CdR-Me-CdL) of an equilateral type whose sides must be equal.

However, it is difficult for people to have such perfect symmetry, because of multiple influencing factors (1-2,4,6). We considered that measuring the angles of the mandible is a good method of comparing sides and symmetry. It is clear that, the more equal the angles between sides are, the more symmetry we will find in a mandible. With respect to the first triangle analysed (CdR-Me-CdL), we have found that the major part of our sample can be considered symmetric. Particularly, having the two semiangles of the lower angle (CdR-Me-CdM/ CdL-Me-CdM) equal or almost equal means that the other two upper angles that form that triangle must be equal as well, and this means symmetry.

**Table 4.** Analysis of the 9 relationships of all sides that form the four triangles of the body and ramus of the mandible.. R\_Kr\_Cd= relationship between KrL-CdL and KrR-CdR; R\_Kr\_Go= relationship between KrL-GoL and KrR-GoR; R\_Cd\_Go= relationship between CdL-GoL and CdR-GoR; R\_DB\_Go= relationship between DB-GoL and DB-GoR; R\_Go\_Me= relationship between GoL-Me and GoR-Me; R\_DB\_GoL\_Me= relationship between DB-GoL and GoL-Me; R\_DB\_GoR\_Me= relationship between DB-GoR and GoR-Me; R\_KrL\_CdL\_DB\_Me= relationship between KrL-CdL and DB-Me; R\_KrR\_CdR\_DB\_Me= relationship between KrR-CdR and DB-Me.

	Mean Ratio	Typical De- viation	Mnimum	Maximum	Confidence Interval 95%	
					Inferior	Superior
R_Kr_Cd	1.03505	0.06578	0.86	1.22	1.0199	1.0502
R_Kr_Go	1.02054	0.0511	0.91	1.16	1.0088	1.0323
R_Cd_Go	0.96506	0.05472	0.78	1.14	0.9525	0.9776
R_DB_Go	0.99987	0.03608	0.92	1.11	0.9916	1.0082
R_Go_Me	1.00037	0.06759	0.87	1.17	0.9848	1.0159
R_DB_GoL_Me	0.96692	0.12328	0.74	1.26	0.9386	0.9953
R_DB_GoR_Me	0.96575	0.12388	0.73	1.27	0.9372	0.9943
R_KrL_CdL_DB_Me	0.93543	0.1269	0.71	1.27	0.9062	0.9646
R_KrR_CdR_DB_Me	0.90576	0.12368	0.66	1.19	0.8773	0.9342

Secondly, on analysing the ratios existing between each of the sides of the following triangles: CdR-Me-CdL; FzR-Me-FzL; GoR-N-GoL and the Gl-Me distance, in order to reach symmetry, each of the sides must measure the same. We found, in our study, that almost all ratios between distances were close to 1:1. Only the relationship between the CdR-CdL distance with respect the rest of the distances, wasn't close to 1:1.

If we interpret these values, we can find that, in the majority of the patients, a good relationship between different structures is present, which as some authors stated, means symmetry and aesthetics (3,5,6). This could be a good method, thus, to evaluate the characteristics of a particular patient, when looking for a complete diagnosis.

Finally, we have analysed the four isosceles triangles obtained from the mandibular body and ramus and the relationships that existed between sides. In this study, all the ratios existing between the sides that form these 4 triangles of the body and ramus of the mandible have been established.

On analysing the ratios between the base of the mandibular body and ramus, it was found that the most symmetrical relationships were those corresponding to the sides of the body of the mandible and the most asymmetrical relationships were those corresponding to the base of such triangles.

With the introduction of CBCT in orthodontic diagnosis, new possibilities have opened up in the study of cranio-facial relationships (8,13). We considered it of interest to create and study a sample of patients using these new technologies. This type of 3D analysis are go-

ing to be implemented in the future, for all the patients who need a CBCT. It is, thus, important to create norms and measurements in order to diagnose them following also these criteria.

The conclusions of our study are:

We observed that regarding to the CdR-Me-CdL triangle, the 69.33% of the patients could be considered symmetric.

As regards the relationship between the sides of the CdR-Me-CdL, FzR-Me-FzL, GoR-N-GoL triangles and the Gl-Me distance, it was found that almost all ratios were close to 1:1. Only the relationship between the CdR-CdL distance with respect the rest of the distances, wasn't close to 1:1.

With regard to the ratios of the 4 triangles of the mandible, it was found that the most symmetrical relationships were those corresponding to the sides of the body of the mandible and the most asymmetrical relationships were those corresponding to the base of such triangles.

To conclude, it must be said that we did not find articles regarding the topic of the present study. Only few authors had studied the same topic in the past, but using measurements directly made on skulls, or using cephalometry with 2D technologies (3-6), which means that comparisons of our results could not be made as extensively as we would like.

**References**

1. Rhodes G, Zebrowitz LA, Clark A, Kalick SM, Hightower A, McKay R. Do facial averageness and symmetry signal health? *Evol Hum Behav.* 2001;22:31-46.

2. Edler RJ. Background considerations to facial aesthetics. *J Orthod.* 2001;28:159-68.
3. Ricketts RM. The biologic significance of the divine proportion and Fibonacci series. *Am J Orthod.* 1982;81:351-70.
4. Perrett DI, May KA, Yoshikawa S. Facial shape and judgements of female attractiveness. *Nature.* 1994; 368:239-42.
5. Jefferson Y. Facial beauty--establishing a universal standard. *Int J Orthod Milwaukee.* 2004;15:9-22.
6. Hönn M, Göz G. The ideal of facial beauty: a review. *J Orofac Orthop.* 2007;68:6-16.
7. Kitaura H, Yonetsu K, Kitamori H, Kobayashi K, Nakamura T. Standardization of 3-D CT measurements for length and angles by matrix transformation in the 3-D coordinate system. *Cleft Palate Craniofac J.* 2000;37:349-56.
8. Pinsky HM, Dyda S, Pinsky RW, Misch KA, Sarment DP. Accuracy of three-dimensional measurements using cone-beam CT. *Dentomaxillofac Radiol.* 2006;35:410-6.
9. Lascala CA, Panella J, Marques MM. Analysis of the accuracy of linear measurements obtained by cone beam computed tomography (CBCT-NewTom). *Dentomaxillofac Radiol.* 2004;33:291-4.
10. Marmulla R, Wörtche R, Mühling J, Hassfeld S. Geometric accuracy of the NewTom9000 Cone Beam CT. *Dentomaxillofac Radiol.* 2005;34:28-31.
11. Mischkowski RA, Pulsfort R, Ritter L, Neugebauer J, Brochhagen HG, Keeve E, Zöllner JE. Geometric accuracy of a newly developed cone-beam device for maxillofacial imaging. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2007;104:551-9.
12. Moerenhout BA, Gelaude F, Swennen GR, Casselman JW, Van Der Sloten J, Mommaerts MY. Accuracy and repeatability of cone-beam computed tomography (CBCT) measurements used in the determination of facial indices in the laboratory setup. *J Craniomaxillofac Surg.* 2009;37:18-23.
13. Lagravère MO, Carey J, Toogood RW, Major PW. Three-dimensional accuracy of measurements made with software on cone-beam computed tomography images. *Am J Orthod Dentofacial Orthop.* 2008;134:112-6.
14. Stratemann SA, Huang JC, Maki K, Miller AJ, Hatcher DC. Comparison of cone beam computed tomography imaging with physical measures. *Dentomaxillofac Radiol.* 2008;37:80-93.
15. Periago DR, Scarfe WC, Moshiri M, Scheetz JP, Silveira AM, Farman AG. Linear Accuracy and Reliability of Cone Beam CT Derived 3-Dimensional Images Constructed Using an Orthodontic Volumetric Rendering Program. *Angle Orthod.* 2008;78:387-95.
16. Kobayashi K, Shimoda S, Nakagawa Y, Yamamoto A. Accuracy in measurement of distance using limited cone-beam computerized tomography. *Int J Oral Maxillofac Implants.* 2004;19:228-31.
17. Zamora N, Llamas JM, Cibrián R, Gandia JL, Paredes V. A study on the reproducibility of cephalometric landmarks when undertaking a three-dimensional (3D) cephalometric analysis. *Med Oral Patol Oral Cir Bucal.* 2012;17:e678-88.

#### *Acknowledgments*

The present work has been granted by a Precompetitive Project (UV-INV-PRECOMP 12-80391) of the University of Valencia.