


C-A/AP/6
October 1999

Pre-conceptual Design of a Rapid Cycling Medical Synchrotron

The AES/BNL Collaboration
October 27, 1999

AES: A. Favale, T. Myers, J. Rathke, J. Sredniawski, A. Todd
BNL: D. Barton, M. Brennan, A.Z. Diaz, D. Lowenstein, A. Luccio,
S. Peggs, A. Pendzick, D. Phillips, P. Pile, T. Roser, J. Sandberg,
S. Tepikian, J. Tuozzolo, W. T. Weng, S.Y. Zhang
LBL: B. Ludewigt

Editor: S. Peggs




**Collider-Accelerator Department
Brookhaven National Laboratory
Upton, NY 11973**

Pre-conceptual Design of a Rapid Cycling Medical Synchrotron

The AES/BNL collaboration
October 27, 1999

AES: A. Favale, T. Myers, J. Rathke, J. Sredniawski, A. Todd
BNL: D. Barton, M. Brennan, A.Z. Diaz, D. Lowenstein, A. Luccio, S. Peggs,
A. Pendzick, D. Phillips, P. Pile, T. Roser, J. Sandberg, S. Tepikian,
J. Tuozzolo, B. Weng, S.Y. Zhang
LBL: B. Ludewigt

Editor: S. Peggs



Contents

1	Introduction and Overview	5
1.1	Key features	6
1.2	Performance upgrades	7
2	Injector	9
2.1	Proton Source & Low Energy Beam Transport (LEBT)	10
2.2	RF Accelerator	11
2.3	High Energy Beam Transport (HEBT)	14
2.4	Control System	14
2.5	Support Subsystems	15
3	Synchrotron	17
3.1	Optics	18
3.2	Magnets	22
3.3	Main power supplies and ramping	23
3.4	Radio Frequency	28
3.5	Longitudinal dynamics	30
3.6	Accelerator controls	36
4	Beam delivery	37
4.1	Switch yard and transport line	37
4.2	Planar gantry	40
4.3	Variable focusing nozzle	41
4.4	Treatment planning and patient safety	43

Abstract

The Rapid Cycling Medical Synchrotron (RCMS) delivers more than 3×10^{12} protons per minute at a maximum extraction energy of 270 MeV, with single turn “fast extraction” at a repetition rate of 15 Hz. Upgrade paths are available to deliver substantially more protons per minute, to increase the extraction energy to 320 MeV, and to enable “slow extraction”.

The main advantages for proton therapy are:

- **easy pulse-by-pulse adjustment of energy and intensity**
- **less beam on more frequent cycles in digital dose delivery**
- **upgradability: wobbling, scanning, proton radiography**
- **identical simple light magnets in both synchrotron and gantry**
- **easy installation and maintenance with small modular components**

Three key design choices account for these advantages – **fast extraction, rapid cycling, and strong focussing**. At all levels the underlying design goals are:

- reliability, simplicity, and modularity
- early and economic patient treatment

In its initial configuration the RCMS delivers fast extracted beam into a double scattering system at the end of a fixed beamline, or at the end of a gantry. The energy of the extracted beam is easily adjusted to any desired value between 70 MeV and 270 MeV.

Three dimensional voxel scanning is also best performed using fast extraction. Nonetheless, the RCMS can easily be upgraded to include slow extraction capabilities, if this is deemed necessary for scanning or for other purposes.

Less beam is accelerated per cycle in a rapid cycling synchrotron, making the accelerator performance more reliable, the design simpler, and the commissioning more rapid. Fast extraction and rapid cycling together make possible dose delivery with digital accuracy.

Strong focusing is achieved by building the synchrotron from “FODO cell” modules containing dipole (bending) magnets, and quadrupole (focusing) magnets.

Each dipole is short, permitting a rectangular design with a small cross section that is easy to construct. The dipole is light in weight, and economical in operation.

The same dipole and quadrupole magnets are used in the fixed beamlines and gantries. The planar gantry design is simple and light, with flexible FODO cell matched achromatic optics.




Figure 1: The racetrack footprint of the RCMS, with two straight sections for injection, extraction, and RF systems, and two 180 degree arcs. The geometry of the planar gantry is also shown, at four rotational angles.

1 Introduction and Overview

A synchrotron is the machine of choice for proton therapy. The most significant advantage over a linac or a cyclotron is that an appropriately designed synchrotron can easily produce a beam whose output energy is continuously variable within wide limits, a capability necessary for contemporary beam delivery modalities. Linacs and cyclotrons can produce more intense beams, but a well-designed synchrotron – especially a rapid cycling synchrotron – can easily provide enough beam intensity for a therapy facility.

The performance parameters of the RCMS design, displayed in Table 1 and elsewhere below, are realistic. They do not approach the performance limits that have been achieved in other accelerators. The RCMS is also conservative in its adoption of technologies that are well-established by successful use elsewhere. The design emphasizes simplicity and modularity, for example by minimizing the number of different component types, so that many parts are interchangeable.

Property [units]	Value
Injection kinetic energy [MeV]	7
Maximum extraction kinetic energy [MeV]	270
Repetition rate, f_{rep} [Hz]	15
Protons per bunch, N	3.3×10^9
Accelerated flux, R [protons/min]	3.0×10^{12}
Circumference, C [m]	28.6

Table 1: Primary parameters of the Rapid Cycling Medical Synchrotron.

The space requirements for the racetrack synchrotron shown in Figure 1 are modest – the circumference is only 28.6 meters. It is likely that the beam transport system will occupy much more floor space, as at other facilities. In principle the circumference could be reduced by a few percent, despite the increased installation and maintenance difficulties which would ensue. The cost saving achieved by reducing the circumference would be small or negligible, since the total length of all the dipole magnets – the most costly magnets – is unaltered.

In contrast, a careful design of the beam delivery system, including the gantries, can result in large cost savings. The fixed beam lines and the gantries of the RCMS, described below, are simple, cheap, and elegant. Their optics are

matched and achromatic, using the same small and light magnets that are used in the synchrotron. Unlike the synchrotron, there is a considerable trade-off between cost and space in the design of the gantry. A “corkscrew” gantry saves building space at the expense of needing almost twice as many dipoles as the “planar” gantry described below. It is assumed in what follows that, while space constraints are important, they are not of the same paramount (and sometimes competing) importance as reliability, simplicity, and modularity.

1.1 Key features

Three features set the RCMS apart in its ability to achieve its design goals: fast extraction, rapid cycling, and strong focusing.

Fast extraction and rapid cycling – a digital approach to dose delivery.

In nominal operation, the rapid cycling RCMS uses single turn fast extraction to deliver beam into a double scattering system. (Nonetheless, an available upgrade path allows beam to be extracted from the RCMS in a resonant “slow extraction” mode, on a time scale of order one second.)

Fast extraction and rapid cycling are also optimal for scanning large tumor volumes by delivering one beam bunch after another to neighboring voxels. This beam delivery method is a very promising digital approach to realizing the full potential of proton therapy. Fast extraction is simple, and has the great advantage that a single shot of beam can be fast extracted at arbitrary energy simply by changing the firing time of the extraction kicker.

Unlike slow spill, fast extraction imposes no special requirements on the good-field aperture and on field stability at extraction time. Fast extraction is very insensitive to noise and fluctuations. Painting the beam over a tumor volume during slow spill requires a very uniform spill and/or a feed-back system; it is hard in practice to avoid slow spill fluctuations of the order of ten per cent or more.

A slow cycle also demands that a lot of protons occupy the accelerator simultaneously, in order to achieve acceptably short treatment times. Slow extraction therefore carries the threat of suddenly dumping a lot of beam into the patient. Rapid cycling allows the average rate requirements to be met with modest intensities per cycle, also eliminating intensity-dependent problems such as coherent instabilities and space-charge effects.

The high level of repeatability that comes with rapid cycling provides a natural way to ensure delivery of the intended number of protons on each cycle.

A rare abnormal cycle with a beam intensity that falls outside tight intensity specifications is easily expelled from the synchrotron – and not delivered to the patient.

Strong focusing.

The magnet layout includes regularly spaced quadrupoles that alternately focus and defocus the beam. This results in “strong focusing” in both transverse planes. The space between quadrupoles is either magnetically empty, or contains a single bending dipole. Straight and bending “half-cells” are assembled to make the racetrack layout shown in Figure 1.

The horizontal and vertical beam sizes are naturally small, allowing the use of small aperture magnets that are small, light, and cheap. Only one kind of dipole magnet and one kind of quadrupole magnet are necessary for the synchrotron, and for the beam transport and gantry systems.

1.2 Performance upgrades

A maximum kinetic energy of 270 MeV is sufficient for proton therapy because it allows the beam to penetrate most of the way through most human bodies in any direction perpendicular to the long axis. However, the RCMS design includes provision for a higher kinetic energy of 320 MeV to permit proton computed tomography, which requires penetration all the way through the body. Proton CT is like X-ray CT except that the images are based not on X-ray attenuation measurements but on measurements of proton energy loss or residual range for multiple trajectories through the body. The 19% change in kinetic energy increases the range by 31% while raising the maximum beam momentum by 10% and the size of the synchrotron by only about 6%; the cost impact is modest.

Proton CT can produce medical images of quality comparable to conventional X-ray CT machines while delivering significantly lower doses. The ability to acquire high-quality images quickly using the same beam and with the patient positioned in the same way as for treatment could be a significant addition to the capability of the proton therapy facility. Such images could be very useful for verifying correct patient positioning; the ability to monitor changes that may occur during the course of treatment could also prove valuable. The proton CT capabilities provided in an RCMS upgrade represent a potentially interesting research opportunity.

The RCMS achieves an extracted beam intensity of more than 3×10^{12} protons per minute, sufficient to supply the needs of a large therapy facility. Although this is small compared to the intensity achieved by high-energy physics

synchrotrons like the BNL AGS, it is not trivial to achieve in a slow cycling low-energy medical synchrotron with a low injection energy.

Higher intensities may prove valuable for shortening treatment times and/or for supplying beam to additional treatment rooms in the future. The high injection energy of 7 MeV and the rapid cycling rate of 15 Hz ensure that the RCMS will not constrain either the size of a tumor that can be treated, or the scope of the facility.

2 Injector

The first beam line component of the RCMS system is the injector subsystem (hereafter called the "injector"). This subsystem serves as the proton source for injection into the synchrotron subsystem. The baseline configuration selected for this report is a linear accelerator (linac) comprised of an multi-cusp plasma type proton source, and a series of RF powered accelerating cavities that serve to group the protons into bunches and accelerate them to a high enough energy for injection into the synchrotron. In addition to these linac components, the injector subsystem contains major support subsystems that provide vacuum, cooling, operational control, and structural support and alignment of the linac components.

Property [units]	Value
Injector Subsystem final energy [MeV]	7.0
Injector Subsystem Delivered Current [mA]	2.1
Pulse Rate [Hz]	15
Pulse Length [nsec]	250
Energy Spread [keV]	± 32
Transverse emittance [μm]	0.3

Table 2: Top level requirements of the injector system.

Initial system optimization studies of the RCMS system indicate that the top-level operational requirements of the injector system are those listed in Table 2. A comparison of these requirements with established operational linac capabilities worldwide indicates that the RCMS injection subsystem can be easily designed from existing operational systems which supply similar, or more complex, beams. Further, the RCMS injector subsystem should be able to achieve high availability based on the excellent operational experience with existing linacs in this power class.

The selected final energy of 7 MeV offers several benefits over lower injection energies. The protons are more dynamically rigid, ameliorating the disruptive effects of the space charge forces. In conjunction the high injection energy dictates a smaller frequency swing of the RF subsystem in the synchrotron, simplifying the design of the ring magnets and minimizing the overall cost of the entire synchrotron system.




Figure 2: Baseline injector subsystem side elevation sketch identifying major beam line components.

The injector subsystem configuration selected to provide the required parameters is depicted in Figure 2. It is comprised of a RF driven multi-cusp plasma source and low energy beam transport section, a radio-frequency quadrupole (RFQ), a single tank Alvarez type drift tube linac (DTL), an output transport section, a three magnet transport section, and a bending magnet. Final insertion into the ring will be accomplished by an injection kicker subsystem of the ring subsystem that will effectively select 250 nano-second sections from the longer linac pulses. The un-used portion of the linac beam will be dumped in the kicker system.

2.1 Proton Source & Low Energy Beam Transport (LEBT)

The baseline proton source is a multi-cusp magnetically confined configuration employing a hydrogen plasma. The hydrogen plasma provides the ions that are extracted from the plasma chamber by a set of electrodes at an energy of 30 keV. The low energy proton beam emerging from the ion source is divergent in nature. It is then magnetically focussed by a pair of solenoid electromagnets for acceptance into the next section, the RFQ.

The plasma of the ion source requires a source of energy to break down the hydrogen gas into positive and negative ions. This is provided by a low frequency (2 MHz) RF source rated at 40 kW pulsed duty. In addition, there are power supplies required to provide the accelerating voltage across the electrodes, as well as a source of hydrogen gas for the plasma, and associated controls. The major properties of the ion source and LEBT are listed in Table 3.

Property [units]	Value
Injector Subsystem Delivered Current [mA]	2.1
Pulse Rate [Hz]	15
Pulse Length [μ s]	5
Transverse emittance [μ m]	< 0.1
Overall Length Power Supplies, Source and LEBT [m]	3.22

Table 3: Major properties of the ion source and Low Energy Beam Transport.

2.2 RF Accelerator

The RF accelerator consists of two accelerating structures: an RFQ and a DTL. Also included as part of the RF accelerator WBS is the RF power to drive these structures.




Figure 3: Radio Frequency Quadrupole developed for the Beam Experiment Aboard Rocket.

The first structure, the RFQ, is envisioned as a standard single physical segment vane-type configuration. The RFQ will accelerate the ion source output beam of 30 keV to a final energy of 1.66 MeV. The protons are initially bunched in the front section of the RFQ, and thereafter accelerated up to final energy.

High availability is expected from this device based on an extensive operational database of very similar RFQs. One such system, shown in Figure 3, has been operated for many years in the AES laboratory and has proven to be extremely reliable. It requires approximately 180 kW of pulsed RF power at 425 MHz. The major properties of the RCMS RFQ are listed in Table 4.

Property [units]	Value
RFQ Input Energy [keV]	30
Nominal Input Current [mA]	0-10
Transmission	90%
Accelerating Frequency [Hz]	425
Kilpatrick Factor	1.664
Output Energy [MeV]	1.66
RF Power Required (Beam & Cu) [kW]	154
Overall Length [cm]	127.0

Table 4: Major properties of the Radio Frequency Quadrupole (RFQ).

The next accelerating element is a single tank classic Alvarez type DTL. Figure 4 depicts a view inside a small DTL tank. The DTL will accelerate the beam from the RFQ output energy of 1.66 MeV to the final injector subsystem energy of 7.0 MeV. The DTL will require 51 $\beta\lambda$ accelerating cells employing a FODO focusing lattice. Individual drift tube magnets will be two centimeter long permanent magnets with a strength of 215 T/m. The DTL tank will be a ramped gradient design (from 1.8 to 3.0 MV/m) with a total length of 3.160 meters, requiring 450 kW of RF power. The major properties of the DTL are listed in Table 5.

Existing DTLs, like the one depicted in Fig. 4, provide emittances far lower than the RCMS requirements. The DTL in Fig. 4 routinely produced an output transverse rms emittance of 0.1 μm for a 20 mA beam (at 1.76 MeV). This is significantly better than what is required for the RCMS. At the lower current for RCMS, the emittance will be even better.




Figure 4: Internal DTL Details (425 MHz).

Property [units]	Value
DTL Input Energy [MeV]	1.66
Nominal Input Current [mA]	2.2
Transmission	99.9%
Accelerating Frequency [MHz]	425
Focusing Lattice	$\beta\lambda$ - FODO
Magnet Strength [T/m]	215
Magnet Length [m]	.02
No. of Drift Tubes	50
Output Energy [MeV]	7.02
Transverse emittance (rms) [μm]	0.3
RF Power Required (Beam & Cu) [kW]	450
Overall Length [m]	3.16

Table 5: Major properties of the Drift Tube Linac (DTL).

2.3 High Energy Beam Transport (HEBT)

The HEBT subsystem includes the beam transport tube with beam focusing quadrupole magnets, a bending magnet, and a beam dump. The quadrupoles maintain the beam size and shape during transport, while the bending magnet serves to direct the proton beam into the synchrotron injection septum. The final angle of the magnet is still to be determined but it is expected to be in the range of 5 to 10 degrees. The magnet will be powered by a DC supply during injection. For checkout and tuning of the injection system the magnet will not be powered and the proton beam will be delivered to the downstream beam dump. The beam dump will be capable of handling the total beam power for extended time.

2.4 Control System

The control system includes both the hardware and the software necessary to operate the injector. The control system architecture consists of a set of distributed processors, each performing independent functions. Each of the subsystems has a common set of hardware, which influences the writing of a common set of software modules. This approach is cost effective by limiting the amount of software that must be created. The processors that will be selected are off-the-shelf PCs

and PLCs that offer reliability and are cost effective. The subsystem processors are connected to the injector master control processor via a high-speed data highway network. In addition, provisions for discreet signals routed to each of the processors will be made. The main processors for the system are:

- Injector master control – provides overall control and synchronization with the RCMS master control
- Ion source high voltage processor – located in the high voltage equipment rack for the ion source, it provides local control for the ion source
- Beam transport processor – is responsible for control of the LEBT components and the HEBT components which are magnets, steerers, and diagnostics
- RF system processor – controls RF amplifiers for all of the accelerating cavities

2.5 Support Subsystems

The support subsystems for the injector includes cooling water, vacuum, prime power distribution, support structure, and personnel safety.

Cooling Water Subsystem – provides de-ionized cooling water for the ion source, the RFQ, and the DTL, the dipole bending magnet, the RF amplifiers and the vacuum pumps. The subsystem consists of a non-refrigerated recirculating cooler with heat exchange to facility water. Shut-off valves and flow meters are provided for each equipment tap-off. Flow switches are provided for critical components to de-energize them in the event of loss of cooling water.

Vacuum Subsystem – provides the high vacuum pumping for the ion source, LEBT, RFQ, and DTL. The use of cryopumps is incorporated throughout because of their cost effectiveness. A mechanical roughing pump will be provided for initial pump out from atmosphere.

Prime Power Distribution – will receive facility electric power from a single source and distribute to each component requiring it. A fused disconnect will be provided for each powered component. Power to the ion source high voltage equipment rack will be fed through a magnet contactor wired to disconnect power to the rack in the event of loss of cooling water.

Support Structure – this structure provides the physical support and alignment provisions for the entire accelerator and beam transport subsystem. It is made up from aluminum extrusion pieces, which are bolted and or welded together as required.

Personnel Safety Subsystem – provides personnel safety against hazards in the injector system. Other than facility hazards such as radiation, which must be protected by shielding which is not part of this WBS, the only personnel hazard is the high voltage applied to the ion source rack. To assure personnel protection, a cage at ground potential and an entry door will enclose all items that operate at high voltage. Opening the entry door will result in de-energizing the high voltage rack.

3 Synchrotron

Figure 1 shows the racetrack footprint of the synchrotron, consisting of two straight sections and two 180 degree bend sections or arcs. Each straight section contains five FODO half-cells without dipoles and each arc consists of eight half-cells with dipoles. With a grand total of 26 half-cells, the total circumference of the RCMS is 28.6 meters. Figure 5 shows the optical functions – the beta functions β_x and β_y and the horizontal dispersion η – for the synchrotron.

Rapid Cycling Medical Synchrotron




Figure 5: Optical functions for the synchrotron.

The straight sections on each side accommodate the functions of injection, extraction, and acceleration. They are made dispersion free by setting the horizontal betatron phase advance per cell to 90 degrees, so that the net horizontal phase advance per arc is 360 degrees. Two adjacent half-cells on one side are

occupied by radio frequency cavities. Each side has a fast kicker and a septum magnet separated by one empty half-cell. One straight is used for injection, and the other for extraction. The nominal parameters of the beam injected into the synchrotron are listed in Table 6.


Property [units]	Value
Injection pulse length, Δt [nsec]	250
Injection pulse current [mA]	2.11
Normalized rms emittance, ϵ [μm]	0.3
Momentum width at injection (rms), σ_p/p	0.001
Total momentum width, $\Delta p/p$	± 0.0023
Total kinetic energy width, ΔK [keV]	± 32

Table 6: Synchrotron beam input parameters.



Variation of the extraction energy is achieved simply by adjusting the extraction time. This potentially avoids the use of energy degraders, delivering higher quality beam with better resolution, but it does not avoid the need to change the excitation of the transport line magnets in proportion to the extraction momentum. However, the transport lines which are described below are in general insensitive to momentum matching errors, since they are achromatic and (mostly) dispersionless.

3.1 Optics

Figure 6 illustrates the layout of a bending half-cell, including a quadrupole, a dipole, a Beam Position Monitor (BPM), and a dipole corrector. The RCMS has only one type of each of these elements, minimizing the effort required for component design and development, and reducing the required number of kinds of spares. The dipole is placed as far upstream as possible to maximize the free space available between its downstream end and the next quadrupole. The BPM may be integrated into the vacuum pipe in the center of the RCMS quadrupoles. Table 7 lists the key geometric parameters of the half-cell, and a selection of parameters which summarize the optical functions shown in Figure 5.



HALF CELL



MAGNET	RING DIPOLE
APERTURE	4 cm x 8 cm (HxV)
CORE LENGTH	63 cm
CURRENT	2000 A (Max)
COIL	12 Turns/Coil
WEIGHT	1100 lbs

MAGNET	RING QUADRUPOLE
APERTURE	Ø4 cm
CORE LENGTH	12 cm
CURRENT	600 A (Max)
COIL	6 Turns/Coil
WEIGHT	100 lbs

Figure 6: The FODO half cell and its principal magnets.

Property [units]	Value
Half-cell length [m]	1.1
Phase advance per cell, H and V [deg]	90
Number of FODO cells	13
Max. horizontal beta function, β_{xmax} [m]	3.39
Max. vertical beta function, β_{ymax} [m]	4.16
Max. dispersion function, η_{max} [m]	2.18
Horizontal tune, Q_x	3.25
Vertical tune, Q_y	2.85
Natural horizontal chromaticity, ξ_x	-0.88
Natural vertical chromaticity, ξ_y	-0.69
Transition gamma, γ_T	2.39

Table 7: Optical parameters of the synchrotron.

The horizontal and vertical root mean square beam sizes σ_x and σ_y at a given location in the synchrotron are related to these quantities as follows:

$$\sigma_x = \left(\frac{\beta_x \epsilon}{(\beta\gamma)} + \left(\eta \frac{\sigma_p}{p} \right)^2 \right)^{1/2} \quad (1)$$

$$\sigma_y = \left(\frac{\beta_y \epsilon}{(\beta\gamma)} \right)^{1/2} \quad (2)$$

where the normalized emittance ϵ is assumed to be the same in both planes, σ_p is the rms momentum spread, and $(\beta\gamma)$ is a relativistic Lorentz factor. The second term in the brackets is absent in the vertical case because the vertical dispersion is zero. Table 8 lists the maximum expected beam sizes (and other parameters) at the bottom and at the top of the ramp, using the beam parameters from Table 6 and the optical parameters of Table 7. The contribution of the transverse emittance to the beam size dominates that of the momentum spread.

Since the dipoles are rectangular, their pole face rotations focus the beam only in the vertical plane. This means that the horizontal phase advance in the straight cells is the same as in the bending cells, so that the total horizontal tune is $Q_x \approx 3.25$. It also means that the vertical tune is $Q_y \approx 4.05$ if the focusing (F) and defocusing (D) quadruples were of the same integrated strength.

Property [units]	injection	top
Kinetic energy, K [MeV]	7.0	270.0
Momentum, p [MeV/c]	114.8	761.3
Lorentz γ	1.0075	1.2878
Lorentz β	0.121	0.630
Revolution frequency, F_{rev} [MHz]	1.273	6.605
Revolution period, T_{rev} [μ s]	0.785	0.151
Rigidity, $B\rho$ [Tm]	0.383	2.539
Dipole field, B [T]	0.226	1.500
Quad field gradient, G [T/m]	3.52	23.32
Quad pole tip field, B_{pole} [T]	0.070	0.466
Vertical rms beam size, σ_y [mm]	4.13	1.60
Horizontal rms betatron beam size [mm]	3.73	1.44
Horizontal rms momentum beam size [mm]	2.18	2.18
Total horizontal rms size, σ_x [mm]	4.32	2.61

Table 8: Kinematic and beam parameters at the bottom (injection) and top of the acceleration ramp.

The vertical tune is adjusted by varying the F and D quadrupole strengths independently. This can be achieved in practice, with identical quadrupoles, by allowing for two different current taps.

3.2 Magnets

There is one rectangular dipole with a magnetic length of 0.665 meters between the quadrupoles in each bending half-cell. Each of the 16 dipoles bends the beam through 22.5 degrees, with a maximum field of 1.5 T. Rectangular dipoles – with parallel end faces and straight sides – are used instead of sector dipoles because they are more easily constructed. More information about the magnets and other aspects of the synchrotron lattice can be found in Table 9 and Figure 6.

Property [units]	Value
Number of dipoles	16
Magnetic length of dipole [m]	0.665
Max. dipole field (top) [T]	1.50
Max. dipole dB/dt [T/s]	60.0
Dipole bend angle [deg]	22.5
Dipole bend radius [m]	1.693
Dipole sagitta [mm]	32.5
Dipole gap width [mm]	80
Dipole gap height [mm]	40
Dipole weight (each) [lbs]	1100
Number of quadrupoles	26
Magnetic length of quadrupole [m]	0.14
F quadrupole gradient (top) [T/m]	23.8
D quadrupole gradient (top) [T/m]	19.7
Quadrupole pole tip diameter [mm]	40.0
Quadrupole weight (each) [lbs]	100

Table 9: Dipole and quadrupole magnet parameters.

The rapid cycling nature of the synchrotron leads to particular, although not unprecedented, challenges in three areas. Two of these areas – the main power supply system and the radio frequency system – are addressed at some length, directly below.

The third area is in the design of the vacuum chamber pipe through which the beam passes. A conventional solid metal beam pipe cannot be used inside the dipoles, because the eddy currents induced during the ramp would be far too strong. A ceramic beam pipe is a reasonable solution for the quadrupoles and perhaps also for the dipoles. A second option that is used in practice at Cornell and at Fermilab is to put the vacuum chamber around the outside of the dipole. A third option is to use a hybrid vacuum chamber that allows beam image currents to flow along the direction of the beam, but does not allow eddy current loops across the beam pipe. Any one of these three schemes could be made to work. The optimal choice has not been made at the time of writing this pre-conceptual design report.

3.3 Main power supplies and ramping

In the initial configuration of the RCMS, a single resonant power supply drives all the dipoles in series. A direct current bias I_{DC} is added to the sinusoidal alternating current I_{AC} to ensure that the minimum current matches the required injection field. The total dipole current I is therefore

$$I(t) = I_{DC} - I_{AC} \cos(2\pi f_{rep}t) \quad (3)$$

Injection occurs at $t = 0$ when $I = I_{DC} - I_{AC} = 300A$. Beam is extracted from the ring sometime before $I = I_{max}$ at $f_{rep}t = 0.5$, when $I = I_{DC} + I_{AC} = 2000$ A. This waveform is illustrated in Figure 7. Except for iron saturation effects the beam momentum is directly proportional to the dipole current.

The dipoles comprise the inductive part of a series LC circuit, driven at a resonant frequency that is the same as the repetition rate f_{rep} of the synchrotron. This is illustrated in Figure 8. Similar resonant power supplies have been used successfully for more than two decades in rapid cycling combined function synchrotrons, such as the Cornell electron synchrotron (60 Hz) and the Fermilab proton booster (15 Hz). Such systems are both reliable and economical.

The resonant circuit is driven by two programmable excitation power supplies. In a series resonant topology the excitation power supplies deliver the full magnet current, but at a significantly reduced voltage, when compared to a non-resonant system. Chokes must be placed across each capacitor bank to provide a path for the DC bias current to flow. The chokes are designed with secondary windings which will all be connected in parallel to provide coupling between the individual resonant circuits.

Besides their simplicity, resonant power supplies have the major advantage of “recycling” most of the stored energy between inductors and capacitors. The




Figure 7: The sinusoidal waveform in RCMS dipoles.

maximum energy stored in the dipoles, when the current and the field are at their maximum values, is approximately

$$U_{max} = \frac{B_{max}^2}{2\mu_0}(LWH)N_{dip} \quad (4)$$

where B_{max} is the dipole field, $\mu_0 = 4\pi \cdot 10^{-7}$ [H/m] is the permeability of free space, L, W, H are the length, width, and height of the dipole field aperture, and N_{dip} is the total number of dipoles.

In the approximation that $I_{inj} \ll I_{max}$, the peak *reactive* dipole power




Figure 8: The resonant programmable main dipole power supply system.

P_{max} is related to U_{max} through

$$P_{max} \approx U_{max} \frac{3\sqrt{3}\pi}{4} f_{rep} = 4.08 U_{max} f_{rep} = 1.86 \text{ [MW]} \quad (5)$$

Fortunately the resistive power losses, which need to be continuously replaced by the excitation supplies and the power line, are much less than the reactive power. This makes a resonant system very economical from both a capital equipment and an operational point of view. Table 10 shows that the total of the resistive losses to the dipoles, the choke, and the capacitors is about 182 kW.

The main power supply driving the quadrupoles is much less demanding in power and performance. In order to compensate for field saturation effects in the dipoles (and, to a lesser extent, in the quadrupoles), the quadrupole power supply must be independently programmable. Such a non-resonant power supply is illustrated in Fig. 9. Although non-resonant power supplies are less efficient, the wall plug power necessary to drive the quadrupoles is only 15 kW, almost negligible in comparison to the 182 kW consumption of the dipole power supply.

Property [units]	Value
Topology	Series Resonant
Repetition rate [Hz]	15
Number of excitation power supplies	2
Excitation power supply current [A]	2000
Excitation power supply voltage [V]	100
Injection current [A]	300
Peak current	2000
Required magnet current, I_{DC} [A]	1150
Required magnet current, I_{AC} [A]	850
Number of capacitance banks	16
Number of bypass chokes	16
Number of dipoles	16
Capacitance per bank [F]	0.175
Inductance of choke [mH]	1.95
Inductance of dipole [mH]	0.97
Resistance of choke [$m\Omega$]	4.6
DC resistance of dipole [$m\Omega$]	2
Quality factor	26
Stored energy [kJ]	30.5
Maximum reactive power [MW]	1.865
Capacitor losses [kW]	13
Choke losses [kW]	105
Magnet losses [kW]	64
TOTAL losses [kW]	182

Table 10: Main dipole power supply parameters.

The quadrupole power supply is a standard switch mode type of unit. Switch mode supplies have the advantage of operating at a high frequency, typically 40 kHz, which allows very good regulation and economical filtering. The power supply has a thyristor controlled pre-regulator, which reduces the amount of reactive power the supply draws from the line. This supply is readily available commercially and has a proven reliability.




Figure 9: The main quadrupole power supply.

The upgrade path to slow extraction has an important impact on the main dipole power supply, for it, too, must become non-resonant in this scenario. A major advantage of the topology chosen for the resonant dipole power supply is that the excitation power supplies, which are programmable, are rated for the full magnet current. With the reduced ramp rate requirements needed for slow extracted beam, these power supplies can be reconfigured to drive the dipole magnets directly. High power programmable power supplies are in routine operation at the BNL AGS Booster.

The power line swing caused by a resonant power supply is much lower than that due to a comparable rapid cycling programmable power supply; this is a significant advantage in choosing a resonant system. Large variations in

Property [units]	Value
Topology	Switch Mode
Repetition rate [Hz]	15
Number of power supplies	1
Power supply current [A]	650
Power supply voltage [V]	75
Peak Current [A]	595
Number of magnets	26
Inductance of magnet [μH]	51
DC resistance of magnet [$\text{m}\Omega$]	1.7
Magnet losses total [kW]	15

Table 11: Main quadrupole power supply parameters.

reactive power flow cause voltage flicker problems, which can be very costly to solve. This advantage of the resonant system disappears as the ramp rate is decreased.

3.4 Radio Frequency

Two ferrite loaded RF cavities provide the voltage needed for acceleration and bunch stability. The unique challenge for these cavities is the very high rate of change of the RF frequency. Typically the resonance frequency of a cavity is changed dynamically to match the changing beam frequency by applying a variable DC bias to the ferrite. However, ferrites have a natural inertia of their permeability which limits the maximum rate of change of frequency. The rate of change that would be required to keep the RCMS cavities on resonance exceeds the value achieved in any previous synchrotron application by a factor of ten.

Fortunately, RCMS cavities do not need to be operated on-resonance, because the voltage required is modest (~ 3 kV each) and the beam intensity presents negligible beam loading. To demonstrate the validity of this assertion a laboratory measurement was performed with a spare cavity from the AGS RF system. By reconfiguring the cavity appropriately and using an available RF amplifier it was shown that the necessary RF voltage and frequency change rate can be achieved without dynamically biasing the cavity. Figure 10 shows the test results for the achieved voltage and the required current. A peak voltage of




Figure 10: Lab measurement of AGS rf cavity from 1 to 7 MHz with no dynamic tuning. Top is gap voltage, bottom is current from each of two amplifiers, linear scale.

0.6 kV was attained with acceptable variation of frequency from only 200 Watts of drive power.

These results were used to establish the parameters of circuit model for the cavity/amplifier system, which yielded the results of Fig. 11. From this model a cavity optimized for the RCMS application was developed, by careful choice of the ferrite material and of the geometry of the acceleration gap. The calculated performance of the model RCMS cavity is shown in Figure 12. The calculation shows that the required voltage may be achieved by driving such a cavity with an off-the-shelf solid-state amplifier of less than 5 kW.

The significance of the solid-state amplifier for this project is great. Such an amplifier is completely turn-key in implementation and operation. That is, no RF specialist is required for tune up and maintenance. In contrast, vacuum




Figure 11: Circuit model calculation of lab measurement, establishing values for parameters of the model.

tube amplifiers need a great deal of ancillary equipment for high voltages and water cooling. They also need periodic maintenance to replace vacuum tubes which have inherently finite life times due to filament failure.

3.5 Longitudinal dynamics

In its initial configuration the RCMS ramps the dipole magnetic field from 0.23 T to 1.50 T in 33 ms. This leads to a relatively large value for the maximum rate of change of field, $dB/dt = 60.0$ T/s. The RF voltage V_{RF} roughly tracks the dB/dt ramp in the nominal RF cycle, keeping the synchronous phase approximately constant throughout acceleration. The relatively low RF voltage which




Figure 12: Using the circuit model and optimizing the parameters of the cavity (ferrite material and gap capacitance) the required voltage is achieved with two 2 kW solid state amplifiers.

is applied at injection implies a low bucket height. The bucket profile at injection matches the given injection beam pulse length and the beam momentum spread, which is naturally small, thanks to the use of an RFQ injector. (The small momentum spread and the small dispersion of the synchrotron lattice also help to reduce the transverse size requirements of the beam pipe and magnets.)

The linac beam profile is an important factor in determining the response of the beam to the RF program. The Drift Tube Linac section of the injector does not include a debuncher section, so that the full kinetic energy spread of the beam is $\Delta K \approx \pm 32 kV$, corresponding to a full momentum spread of $\Delta p/p \approx \pm 0.0023$. Simple calculations show that the moving RF bucket is amply large enough to contain this beam at the single turn injection. Simulations of multi-

particle motion over many synchrotron periods must be applied to accurately determine the evolution of the bunch area and the longitudinal emittance, et cetera.

The simulation results presented in graphical form below correspond to the model parameters shown in Table 12. This table also includes summary results from the simulation.

Figure 13a shows the RF voltage V_{RF} , dB/dt , and B for the entire cycle. The RF voltage ramps from 300 V at injection to a maximum value of 4.5 KV after 15 ms. Figure 14 shows the simulated particle distribution in phase space for the first turn, and then for every 500 turns, up to 5,500 turns, or about 4 ms. The bunch length very quickly increases from its injection value of 250 ns to to about 350 ns at 2 ms. This situation is optimized by carefully tuning the injection parameters.




Figure 13: The evolution of longitudinal dynamics parameters during the RCMS acceleration ramp.

Property [units]	Value
Harmonic number, h	1
Number of cavities	2
Maximum total voltage, V_{rf} [kV]	4.5
Maximum accelerating voltage [kv]	4.2
Longitudinal emittance (total) [eVs]	0.03
Normalized rms emittance [μm]	0.3
Revolution period (inj), T_{rev} [μs]	.785
Revolution period (top), T_{rev} [μs]	.151
Revolution frequency (inj), F_{rev} [MHz]	1.274
Revolution frequency (top), F_{rev} [MHz]	6.605
Maximum beam momentum spread	± 0.0042
Maximum synchronous phase [degrees]	41
Maximum synchrotron frequency [kHz]	6.1
Maximum space charge tune shift, ΔQ	0.044
Maximum peak current [mA]	23
Maximum df_{RF}/dt [MHz/s]	270
Minimum bucket length [degrees]	149
Bunching factor at injection	0.24
Bunch length at injection [ns]	400

Table 12: RCMS Radio Frequency system parameters.




Figure 14: Simulated evolution of the injected beam in longitudinal phase space during the first 4 ms of acceleration.

The length of the smooth bunch is assumed to be 400 ns at injection for an analytical calculation whose results show that there is a comfortable margin in the bucket to bunch area ratio, length, and height. Figures 13b, c, and d record these results. Figure 13 shows that there is ample space available in the latter part of the cycle, allowing the RF voltage to be reduced.

Figure 15a shows the space charge tune spread for a normalized rms transverse emittance of $\epsilon = 0.3 [\mu\text{m}]$, the value at the output of the DTL. Figure 15b shows the maximum beam radius, horizontally and vertically. The largest beam size, at injection, is 4.3 mm. Finally, Figures 15c and d show the RF frequency and synchrotron frequency as a function of time.




Figure 15: The evolution of more longitudinal dynamics parameters during the RCMS acceleration ramp.

3.6 Accelerator controls

The principal hardware items which appear to be unique in controls required for a synchrotron-based system are probably only two, both related to the fast-changing time dependence of magnet and other ramped equipment. A distributed, microsecond resolution timing link, preferably using encoded signals would be required for scheduling synchrotron cycle actuations. In addition, fast multi-setpoint waveform generators and digitizers would be required for programmed magnet and rf control. BNL has considerable experience with both items and could contribute specifications and/or designs for such items.

The main area in which BNL could be expected to contribute to the controls software for running the facility would be with high level control algorithms for automation of synchrotron setup and operation and expert diagnostics.

4 Beam delivery

This section describes three optical components of the RCMS design, through which beam passes before reaching the patient:

1. A variable geometry “switch yard and transport line” section, which takes the beam from the synchrotron and directs it towards one of the patient treatment rooms. The optical design of the simplest and most minimal delivery system – a fixed horizontal beam line with a conventional double scattering nozzle – differs little from the transport line that feeds it, except in the immediate vicinity of the nozzle.
2. A “planar gantry” system that enables beam delivery to the patient from any direction around the gantry axis. The solution which is presented implicitly assumes that space constraints do not force the adoption of a corkscrew geometry. The distance from the end of the gantry to the patient is adjustable (in the design stage).
3. A “variable focusing nozzle” magnetically shapes the round spot size of the beam at the patient over a wide dynamic range of almost two orders of magnitude, always with a converging beam. This nozzle can be used either at the end of a fixed beam line, or at the end of a gantry.

All transport lines and gantries discussed here are constructed from the same dipoles and quadrupoles used in the synchrotron itself.

4.1 Switch yard and transport line

Figure 16 shows the schematic layout of a horizontal beam transport line – or fixed horizontal beam line – which extracts protons from the synchrotron and delivers them over approximately 34 meters (a distance that is easily adjusted) to the patient. The figure shows the behavior of the optical functions in one configuration of the entire line. Table 13 lists various parameters of the transport line.

The first two half-cells of the transport line shown here are the same two half-cells that begin the extraction straight in the synchrotron, starting with a D quadrupole. Immediately after the D quadrupole is a fast kicker that deflects the beam vertically by a small angle of approximately 6.0 milliradians. This causes the beam to be displaced by about 24 millimeters at the location of a horizontally bending Lambertson septum magnet one cell downstream. A full-energy kicked beam experiences a field of about 1.2 T in the 0.4 m long Lambertson, giving it a horizontal angle of approximately 0.178 radians, and displacing the beam horizontally by 170 millimeters at the next quadrupole




Figure 16: Optical functions through extraction, switch yard, transport line, and variable focusing nozzle, in a sample horizontal fixed beam line.

downstream. The extracted beam is physically separate from the RCMS at this stage but is still traveling in a FODO cell structure like that of the ring.

Both the horizontal and the vertical extraction angles launch dispersion waves in the transport line, but the vertical dispersion is negligible small. The horizontal dispersion wave can be canceled by placing a dipole with the same bend angle as the Lambertson $4N + 2$ full cells downstream, where N is any integer, or by placing an equal and opposite bend $4N$ full cells downstream. In the case presented here, a bend of the same polarity is placed 2 full cells downstream, about 5.2 meters downstream from the extraction Lambertson. A switch yard is easily constructed to serve as many dispersion free transport lines as desired. The ability to add treatment lines in the future is guaranteed, so long as floor space remains available.

Property [units]	Value
Kicker bend angle [radians]	0.006
Lambertson bend angle [radians]	0.178
Magnetic length of Lambertson [m]	0.4
Maximum Lambertson field [T]	1.2
Length of transport line half-cells [m]	2.6

Table 13: General parameters of the horizontal delivery line.

Short half-cells are desirable in the synchrotron since the beta functions are proportionally small, leading to tighter focusing and a smaller beam. This allows the use of magnets with small apertures and small transverse dimensions in general. The tight focusing of the synchrotron can be relaxed in the transport lines (by spacing the same quadrupoles further apart) because the beam size at even the lowest extraction energy is much smaller than the beam size at injection.

If the quadrupoles in the transport line are placed twice as far apart as normal, the beta functions are twice as large, and the beam size is $\sqrt{2}$ times as large (at the same momentum). The quadrupole excitation current is also halved, so the power consumption per magnet is one quarter of nominal. Since there are half as many quadrupoles per meter of transport, the power consumption per meter is only one eighth of the amount that would be required if the nominal half-cell length were used.

The “lengthener” part of the transport line smoothly matches the optical functions of nominal length synchrotron FODO cells to the double length transport FODO cells. The height of a box representing a quadrupole in Figure 16 is proportional to the strength of that quadrupole, showing that the three “lengthener” quadrupoles each have different strengths. These three quadrupoles must have separate power supplies, or shunts.

Again it should be emphasized that the design presented here is pre-conceptual, and has not undergone parameter optimization. For example, the optimum length of the long half-cells is probably not exactly twice the RCMS half-cell length.

4.2 Planar gantry

The planar gantry is a simple structure using standard synchrotron magnets. Its optics are shown in Figure 17, and its geometry in Figure 18. Thus, the gantry is light and economical in construction and operation.

The optics are fully matched, achromatic, and flexible. The distance between the end of the gantry and the isocenter is a variable which is set at design time. Values in the range from 1.2 m to 3.7 m are trivially achievable. In the design presented here a nominal value of 3 m is assumed to be sufficient to insert a conventional double scattering system, and for a scanning upgrade.




Figure 17: Fully matched achromatic strongly focused optical functions through the planar gantry.

Gantry Geometry




Figure 18: Geometry of the planar gantry, showing a radius of just under 6 meters.

4.3 Variable focusing nozzle

Figure 19 shows the optical performance of the variable focusing nozzle as the beta function at the patient, $\beta_{patient}$, is adjusted over three orders of magnitude, from 0.1 to 100 meters. In the absence of multiple scattering in the patient, this is equivalent to beam spot sizes varying from 0.35 to 11.1 millimeters at 270 MeV. When $\beta_{patient}$ is very small, the beam is strongly convergent, significantly reducing the entrance radiation dose. However, the smallest practical beam size is limited by multiple scattering, so that such sharp focusing is not advantageous except for shallow problems such as ocular tumors. When $\beta_{patient}$ is very large, the practical limit to the maximum size of the beam is the aperture of the triplet quadrupoles. In this configuration the beam is barely convergent – it

is essentially parallel. Note that in no case can the beam be considered to be coming from a point source; hence there is no inverse-square augmentation of entrance dose, and the SAD is always negative.




Figure 19: Optical functions of the variable focus nozzle at four settings, showing a dynamic range of 1000 in β_{patient} , with values from 0.1 m to 100 m.

Table 14 shows the excitations of the four nozzle quadrupoles, as measured by the slope of the B-field, in the four configurations shown in Figure 19. The table shows that quadrupoles Q1 and Q4, which undergo the largest variations, must have separate power supplies. It may be possible to power quadrupoles Q2 and Q3 using a single power supply and a shunt. If the maximum acceptable quadrupole pole tip field is 1.5 T when the field gradient is 13.4 Tesla per meter, then the pole tip diameter of the triplet quadrupoles must be less than 224 millimeters. This is much larger than the largest beam size of 16 millimeters

that occurs when the 70 MeV beams are essentially parallel, so that the beam has the same size in the triplet as in the patient.

$\beta_{patient}$ [m]	0.1	1.0	10.0	100.0
Focal spot size @ 70 MeV [mm]	0.50	1.59	5.04	15.9
Focal spot size @ 270 MeV [mm]	0.35	1.11	3.51	11.1
Gradient of Q1 [T/m]	10.3	3.5	1.7	5.0
Gradient of Q2 [T/m]	10.5	8.4	5.6	6.7
Gradient of Q3 [T/m]	13.4	12.2	9.8	9.9
Gradient of Q4 [T/m]	17.4	4.6	7.4	28.1

Table 14: Four focusing configurations of the nozzle. The spot sizes quoted do not include the effects of multiple scattering. Except for the first row of the table, all values are quoted for a kinetic energy of 270 MeV, and an rms emittance of $\epsilon = 1.0$ [μm].

4.4 Treatment planning and patient safety

The goals of patient treatment planning are conceptually simple: to deliver the intended dose to the treatment volume while minimizing damage to normal tissue and avoiding damage to critical anatomical structures. However, achieving these goals in a realistic three-dimensional situation of an irregularly shaped tumor surrounded by anatomical structures of variable density and composition is a challenge to computing and control systems. Fully realizing the potential of proton therapy requires an integrated approach to medical imaging, patient treatment planning in three dimensions, patient localization, and control of accelerator and beam delivery systems using state-of-the-art computing techniques. Digital dose delivery holds the most promise for achieving these goals.

A rapid cycling synchrotron lends itself to a simple, reliable approach to patient safety; this is one of its main virtues. Conditions will be monitored on each cycle; if dangerous abnormal conditions are detected, any beam currently in the synchrotron will be aborted and subsequent cycles will be inhibited at the source. An abort system internal to the synchrotron is also appropriate. Two dipole correctors separated by 180 degrees of phase advance in one of the arcs can be pulsed to sweep the beam into an absorber block halfway between them. The trigger to the extraction kicker would simultaneously be inhibited.

Examples of situations that would produce this response include an unexpectedly large proton bunch intensity, out of tolerance orbit fluctuations, errors in gantry or patient alignment, or a hardware failure in the final beam steering system. Detection of alarm conditions will require continuous monitoring of beam position and circulating-beam intensity monitors. The final intensity monitor will also serve to log beam delivery automatically for the patients records. The record will contain pulse-by-pulse information of exposure and orientation so that dosages can be reconstructed in complete detail.

One can imagine rare but perhaps not impossible failure modes that conspire to circumvent such monitoring systems. A great safety advantage of using a rapid cycling synchrotron is that the maximum single bunch intensity is too low to cause serious damage even if a single bunch goes astray and strikes the patient in an unexpected way. As a last line of defense in a multi-layered approach to patient safety, monitors just upstream of the patient will detect the first such out-of-tolerance bunch, and the beam will be inhibited at the source on subsequent cycles until the situation is understood. Our goal is to provide safe, precise, effective, thoroughly monitored treatment using highly reliable, easily maintained (or replaced) hardware.