

Commissioning RHIC'S electron lens

**X. Gu, Z. Altinbas, M. Anerella, D. Bruno, M. Costanzo,
W.C. Dawson, A.K. Drees, W. Fischer,
B.M. Frak, D.M. Gassner, K. Hamdi, J. Hock, L.T. Hoff,
A.K. Jain, J. Jamilkowski, R. Lambiase,
Y. Luo, M. Mapes, A. Marone, C. Mi, R. Michnoff, T. Miller,
M. Minty, C. Montag, S. Nemesure,
W. Ng, D. Phillips, A.I. Pikin, S.R. Plate, P. J. Rosas, J. Sandberg,
P. Sampson, L. Snyderstrup, Y. Tan,
R. Than, C.W. Theisen, P. Thieberger, J. Tuozzolo,
P. Wanderer, W. Zhang**

*Presented at the North American Particle Accelerator Conference (NA-PAC 13)
Pasadena, CA
September 29 – October 4, 2013*

**Collider-Accelerator Department
Brookhaven National Laboratory**

**U.S. Department of Energy
DOE Office of Science**

Notice: This manuscript has been authored by employees of Brookhaven Science Associates, LLC under Contract No. DE-AC02-98CH10886 with the U.S. Department of Energy. The publisher by accepting the manuscript for publication acknowledges that the United States Government retains a non-exclusive, paid-up, irrevocable, world-wide license to publish or reproduce the published form of this manuscript, or allow others to do so, for United States Government purposes.

This preprint is intended for publication in a journal or proceedings. Since changes may be made before publication, it may not be cited or reproduced without the author's permission.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

COMMISSIONING RHIC'S ELECTRON LENS *

X. Gu[#], Z. Altinbas, M. Anerella, D. Bruno, M. Costanzo, W.C. Dawson, A.K. Drees, W. Fischer, B. M. Frak, D.M. Gassner, K. Hamdi, J. Hock, L.T. Hoff, A.K. Jain, J. Jamilkowski, R. Lambiase, Y. Luo, M. Mapes, A. Marone, C. Mi, R. Michnoff, T. Miller, M. Minty, C. Montag, S. Nemesure, W. Ng, D. Phillips, A.I. Pikin, S.R. Plate, P. J. Rosas, J. Sandberg, P. Sampson, L. Snyderstrup, Y. Tan, R. Than, C.W. Theisen, P. Thieberger, J. Tuozzolo, P. Wanderer, and W. Zhang

Brookhaven National Laboratory, Upton, NY 11973

Abstract

In the 2013 RHIC polarized-proton run, it was found that the intensity of the RHIC bunch had reached a limit due to the head-on beam-beam interaction at intensity of 2×10^{11} , as we expected from our simulations [1]. To overcome this limitation, we have planned to implement two electron lenses for beam-beam compensation. During and after the 2013 RHIC run, some e-lens systems were commissioned. The effect of the e-lens warm solenoids on the protons orbit was observed and corrected by orbit feedback. The blue electron-lens system was fully tested, except for the superconducting magnet; the electron beam was propagated from the gun to the collector, and most of the instrumentation for the blue e-lens was commissioned. The straightness of the superconducting solenoid #2 field was measured for the first time. The installation of the yellow e-lens system and two superconducting magnets are underway.

INTRODUCTION

Figure 1 schematically depicts the layout of the electron lens that was used during the 2013 run (A), and will be implemented for the upcoming 2014 run (B).

During the 2013 run, all e-lens warm magnets [2, 3] were ready and tested with proton beam; we placed the superconducting solenoid #1 on the yellow e-lens side and replaced the blue superconducting solenoid with a spare EBIS superconducting solenoid, which was operated at 3T. We tested the blue e-lens system with the electron beam.

For the 2014 run (Figure 1 B), the yellow e-lens system will also be available and the two e-lens systems will be ready for commissioning.




Figure 1: Schematic depictions of the layout of the electron lens during the 2013 run (upper) and the 2014 run (lower).

For commissioning the e-lens during the 2013 255 GeV polarized proton run, we first commissioned a new e-lens lattice [4]; then, we studied the effects of e-lens magnetic

field on the proton beam with the e-lens warm solenoids. We found that when the currents from the warm solenoids were ramped up, there were some effects on the beam

* Work supported by Brookhaven Science Associates, LLC under Contract No. DE-AC02-98CH10886 with the U.S. Department of Energy
[#]xgu@bnl.gov

vertical orbit; we can correct this distortion by beam feedback. There were no apparent effects on beam loss, or on the beam intensity, emittance, or luminosity.

We also tested some instrumentation and applications for single-bunch beam-beam compensation during the 2013 polarized-proton run. Two BPMs were commissioned for the proton beam and electron beam.

For commissioning the blue e-lens, we propagated the electron beam through a system with a 3T EBIS spare superconducting-magnet. The properties of the cathode and the profile of the beam were measured with both a pinhole scanner and a YAG screen. The beam profile has Gaussian distribution. The modulator was commissioned with the 80 kHz and the DC beam for 14 and 9.5 hours respectively.

To further investigate the e-lens system reliability and make it more robust, several scenarios of system failure also were undertaken, such as Machine Protection System failure.

MAGNET EFFECTS DURING PROTON STORE

During the 2013 polarized proton run, we tested the e-lens warm solenoids with a proton beam. According to our simulation [5], the orbit of the proton beam is affected by the e-lens warm solenoids, especially in the vertical plane.

Figure 2 shows the distortion of the proton beam during the ramp-up of the e-lens warm solenoids with the orbit feedback on.




Figure 2: The vertical orbit rms with orbit feedback (top, in mm) and the warm magnet currents (bottom, in A)

The green (Yrms) trace in Figure 2, top, shows that the vertical beam rms position changes (~0.15mm) when the e-lens warm solenoid is ramped up even with the orbital feedback on, but recovered subsequently. A change in the orbit feedback parameters later reduced this effect. The effects on the horizontal plane were less.




Figure 3: E-lens magnets current (bottom, in A) and luminosity (top).

Although e-lens warm magnets have some effects on the proton orbit, Figure 3, it is hard to discern any visible effects on RHIC luminosity.

ELECTRON BEAM COMMISSIONING

The blue e-lens system was fully commissioned with the help of the EBIS spare superconducting magnet, including the electron gun [6] and the collector. Figure 4 shows the perveance measurement of electron gun, which is $1.0 \times 10^{-6} \text{ AV}^{-3/2}$.




Figure 4: Perveance measurement of the electron gun.

We fully tested the electron beam running modes, and also demonstrated the system reliability. We ran the DC beam (1.256A) and the 80 kHz (parasitic) beam, respectively, for 9.5 hours and 14 hours; they were stopped intentionally for another test.

Figure 5 illustrates the 14 hour test of the 80 kHz electron beam. The bottom plot is the vacuum in the collector, which indicates that electron current was available in the collector; the upper plot, the anode voltage, denotes that the beam current was constant.

We also measured the electron beam profile with a YAG screen and a pinhole scanner [7]. Figure 6A shows the YAG profile for a 70 mA beam; Figure 6B shows the profile measured by the pinhole scanner for 1150 mA beam. The profile is Gaussian, with a small flat top, similar to the profile from the test bench [8, 9]. The total

beam radius is 2.76 rms widths; the model yielded a value of 2.8.




Figure 5: 80 kHz modulator test with more than a 1A beam




Figure 6: YAG screen (A) and pinhole (B) profile measurements of the electron beam.

E-LENS SUPERCONDUCTING MAGNETS

Two e-lens superconducting magnets were also tested. A horizontal test was done for both solenoid #1 and #2. With only 2 quenches, solenoid 1 reached 5 T. Solenoid #2 reached about 5.5 T. Both magnets have reached 6 T in a vertical test, and are expected to reach 6 T in the cryostat with further training quenches.

Figure 7 shows our measurements for the field straightness test at 2.83 T for solenoid #2, which was obtained at Brookhaven National Laboratory's Superconducting Magnet Department via magnetic needle [10]. The measurement was taken only with main coil and without fringe coil, anti-fringe coil, and correctors.

The green line in Figure 7 is the field measurement and the red line is the measured deviation from a straight line. The blue line is the computed deviation after the application of five correctors.




Figure 7: Measurement of the superconducting magnet field straightness in the horizontal plane.

As the blue line in figure 7 shows, the straightness specification of $\pm 50 \mu\text{m}$ is satisfied in X plane from about -900 mm to 900 mm without correctors. The uncorrected straightness deviation is even less in the Y (vertical) plane.

ACKNOWLEDGMENTS

The authors would like to acknowledge the help of the instrumentation group, vacuum group and controls group. The authors are also grateful to the valuable discussions with the FNAL TEL staff, in particular with V. Shiltsev, A. Valishev, and G. Stancari.

REFERENCES

- [1] Y. Luo, et al., "6-D weak-strong simulation of head-on beam-beam compensation in the Relativistic Heavy Ion Collider". Phys. Rev. ST Accel. Beams 15, 051004 (2012).
- [2] W. Fischer, et al., "Status of head-on beam-beam compensation in RHIC", proceedings of the ICFA Mini-Workshop on Beam-Beam Effects in Hadron Colliders (BB3013), CERN (2013).
- [3] X. Gu, et al., "Designing a Beam Transport System for RHIC's Electron Lens", PAC'11, New York, USA TUP208.
- [4] V.H. Ranjbar, et al. "RHIC Polarized Proton Operation for 2013". IPAC2013, Shanghai, China, May 12-17, 2013, pp. 1544-1546 (2013).
- [5] X. Gu, et al., "The Effects of The RHIC e-lenses Magnetic Structure Layout on The Proton Beam Trajectory", PAC'11, New York, USA TUP207.
- [6] A. Pikin, et al., "Structure and Design of the Electron Lens for RHIC", PAC'11, New York, USA THP100.
- [7] D. M. Gassner, et al., "RHIC Electron Lens Test Bench Diagnostics". DIPAC'11, Hamburg, Germany, MOPD04.
- [8] X. Gu, et al., this proceedings.
- [9] T. Miller, et al, ELECTRON-LENS TEST STAND INSTRUMENTATION PROGRESS, IBIC12, Tsukuba, Japan, THCB01
- [10] C. Crawford, et al, "Magnetic Field Alignment in the Beam-Beam compensation Device", PAC'99, New York, USA, pp. 3321-3323.