

UCRL-CONF-215311



LAWRENCE  
LIVERMORE  
NATIONAL  
LABORATORY

# High resolution trace element and isotopic imaging of microbial systems by NanoSIMS

P. K. Weber, S. J. Fallon, J. Pett-Ridge, S. Ghosal, C. E. Ramon, I. D. Hutcheon

September 13, 2005

AGU Fall Meeting 2005  
San Francisco, CA, United States  
December 5, 2005 through December 9, 2005

## **Disclaimer**

---

This document was prepared as an account of work sponsored by an agency of the United States government. Neither the United States government nor Lawrence Livermore National Security, LLC, nor any of their employees makes any warranty, expressed or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States government or Lawrence Livermore National Security, LLC. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States government or Lawrence Livermore National Security, LLC, and shall not be used for advertising or product endorsement purposes.

## High resolution trace element and isotopic imaging of microbial systems by NanoSIMS

Peter K. Weber, Stewart J. Fallon, Jennifer Pett-Ridge, Sutapa Ghosal, Christina E. Ramon, and Ian D. Hutcheon

Lawrence Livermore National Laboratory, Livermore, CA, USA

The NanoSIMS 50 is the state of the art in high spatial resolution secondary ion mass spectrometry (SIMS), combining unprecedented spatial resolution (as good as 50 nm) with ultra-high sensitivity (minimum detection limit of ~200 atoms). The NanoSIMS has an array of detectors, enabling simultaneous collection of 5 species originating from the same sputtered volume of a sample. The primary ion beam ( $\text{Cs}^+$  or  $\text{O}^-$ ) can be scanned across the sample to produce quantitative secondary ion images. This capability provides a novel new approach to the study of microbial systems. We have applied our NanoSIMS to various microbial systems. We have analyzed sub-regions of bacterial cells, biofilms, and other associated materials to map trace element and isotopic ratios on a submicron scale. Growth and metabolism have been tracked using stable isotope labels. High resolution SIMS is particularly powerful when used in combination with other high resolution techniques, such as FIB and TEM. Examples will be presented to demonstrate the range of capabilities of this technique for microbial systems.

This work was performed under the auspices of the U.S. Department of Energy by the University of California, Lawrence Livermore National Laboratory under Contract W-7405-Eng-48.