

# Focusing nucleic acid-based molecular diagnostics and xenomonitoring approaches for human helminthiasis amenable to preventive chemotherapy

CORRADO MINETTI, E. JAMES LACOURSE\*, LISA REIMER and J. RUSSELL STOTHARD

*Liverpool School of Tropical Medicine, Pembroke Place, Liverpool L3 5QA, UK*

*(Received 19 June 2016; revised 4 September 2016; accepted 12 September 2016)*

## SUMMARY

The current mainstay for control of the four major helminth diseases in humans (lymphatic filariasis, onchocerciasis, soil-transmitted helminthiasis and schistosomiasis) is with preventive chemotherapy by mass administration of key anthelmintics. Following the London Declaration on Neglected Tropical Diseases in 2012, a roadmap for the elimination and control of these helminthiasis by 2020 has been devised. With expected declines in prevalence and intensity of these infections, there is urgent need for implementing more sensitive, high-throughput and cost-effective diagnostic tools. Currently available diagnostic approaches for surveying, monitoring and evaluating helminth control programmes are based on microscopical observation of eggs/larvae, and/or detection of antibodies or parasite antigens in stool, urine or blood; all relatively low-throughput and of limited sensitivity and specificity. Newly proposed approaches for helminthiasis diagnosis include the nucleic acid-based methods of (multiplex) real-time polymerase chain reaction assays, loop-mediated isothermal amplification and recombinase polymerase amplification. However, as well as sensitivity/specificity evaluation, their comparison to current 'gold standard' diagnostics and future application in individual-/community-based diagnosis, or in xenomonitoring requires consideration of relative costs, agreement of standard methods and strategic interpretation of resulting data before control/elimination programmes might best utilize molecular diagnostics to inform decision making. We review current nucleic-acid-based molecular diagnostic methods and highlight the needs and future research required to refine these tools for monitoring and evaluation of control and elimination programmes for four major human helminthiasis.

Key words: neglected tropical diseases, molecular diagnosis, xenomonitoring, LAMP, RPA, real-time PCR, qPCR.

## INTRODUCTION

Lymphatic filariasis (LF), onchocerciasis, soil-transmitted helminthiasis (STHs) and schistosomiasis are four major helminth diseases in humans targeted for control with preventive chemotherapy (PC) by mass drug administration (MDA). Following the London Declaration on Neglected Tropical Diseases (NTDs) in 2012, the World Health Organization (WHO) developed a roadmap to prioritize and guide the scaling-up of interventions for control and elimination of NTDs by 2020 (World Health Organization, 2015). Current goals include elimination of LF and control of onchocerciasis, soil-transmitted helminths and schistosomiasis by sustaining, expanding and extending MDA alongside condition-specific interventions including vector control and water, sanitation and hygiene measures to abate transmission.

Access to and deployment of appropriate diagnostic tools is crucial within disease control and elimination programmes throughout their inherent four key time points: (i) initial mapping to establish disease prevalence, (ii) monitoring the impact of

interventions after they are started, (iii) deciding when to stop interventions and (iv) post-elimination surveillance (Solomon *et al.* 2012). Diagnostic tools which are both highly sensitive and cost-effective, particularly when scaling-up MDA and control interventions and within 'endgame' final stages of elimination scenarios and post-MDA surveillance, are key to the impact and success of control and elimination programmes (McCarthy *et al.* 2012). The expected reduction in both infection prevalence and intensity in communities will result in the need to detect very low levels of infection and screen larger numbers of people or vectors, often returning a much greater number of negative tests to provide confidence in excluding presence of infections. The reliability of diagnostics under such scenarios is vital when the decision to stop MDA in an area has to be made, as well as in the post-MDA monitoring stages: residual low levels of infection, if overlooked, can cause resurgence of transmission and subsequent endemic infection. More sensitive diagnostics can also aid in the early detection of drug resistance potentially emerging in parasite populations as a result of increased exposure to drugs due to scaling-up of MDA. Additionally, given the nature of integrated MDA, further unique challenges exist in devising strategies,

\* Corresponding author: Liverpool School of Tropical Medicine, Pembroke Place, Liverpool L3 5QA, UK. E-mail: [james.lacourse@lstmed.ac.uk](mailto:james.lacourse@lstmed.ac.uk)

protocols and tests for each disease in parallel when appropriate.

Currently, for MDA-targeted helminths the pragmatic 'gold standards' for human diagnosis as endorsed by the WHO for implementation in resource-poor settings include the microscopic identification of parasite larvae in blood (LF) or skin snips (onchocerciasis), eggs or larvae in stools or urine (STHs and schistosomiasis) or the detection of antibodies or worm antigens circulating in blood (LF, onchocerciasis and schistosomiasis) or urine (schistosomiasis) (McCarthy *et al.* 2012). All these 'gold standard' tools, although appropriate for mapping disease in moderate-to-high intensity and prevalence areas, show their limitations for impact monitoring and post-elimination stages: in particular for the latter stage, it is proposed that the ideal test format would be a high-throughput laboratory assay (Solomon *et al.* 2012).

Research requirements to advance current diagnostic tools for human helminthiasis in the context of control and elimination programmes have been extensively reviewed (McCarthy *et al.* 2012; World Health Organization, 2012). However, there remains lack of funding commitment despite recognition of the fundamental role of diagnostics in meeting the 2020 goals for NTDs (Hotez *et al.* 2016). With more countries scaling up MDA and associated interventions and approaching elimination of at least some of these infections, an important question is whether efforts should be focused on developing novel sensitive diagnostics for resource-poor and/or field settings, or validating throughput, cost-effectiveness and building required capacity in implementing existing tools.

Future solutions to the need for more sensitive, high-throughput diagnostics could include development of antibody, antigen and nucleic acid detection-based methods as all these approaches could lend themselves to levels of automation, handling of larger sample numbers, and sensitivity that helminth elimination programmes will need. These approaches, however, would require investment in capacity building with equipment, training and general resources as well as the necessary initial research to develop, validate, standardize and quality assure assays. Whilst antibody and antigen detection assays are of great potential value and most certainly warrant further investment, it is the nucleic-acid-based methods that arguably may currently offer the most rapid route to multiplexing and implementation within helminth control and elimination programmes. Furthermore, nucleic-acid-based methods may also offer ancillary benefit through their potential to monitor population genetic variations in helminths through possible drug pressure selection changes and thus inform strategies for judicious use of the few efficacious anthelmintics upon which we rely heavily.

Proposed nucleic-acid-based molecular approaches as candidates for sensitive, high-throughput

diagnostics, include (multiplex) real-time quantitative polymerase chain reaction (qPCR) assays or isothermal amplification methods, such as loop-mediated isothermal amplification (LAMP), nucleic acid sequence-based amplification (NASBA) and recombinase polymerase amplification (RPA). However, although these molecular methods do offer the required sensitivity and high-throughput, their acceptance and uptake is hampered by three main 'obstacles', that may be categorized as; 'Cost', through the expense of equipment, reagents and training; 'Standard Method', in terms of there being as yet no agreement on a specific 'best approach/method and quality standards'; and 'Result Interpretation' with regard to how control/elimination programmes might best use molecular data to inform decision making (summarized in Table 1).

PCR is one of the most promising approaches to detect parasite DNA or RNA in human or vector samples due to its higher diagnostic sensitivity and specificity compared to microscopy or immunodiagnostics through antibody or antigen detection. In particular real-time qPCR has become the mainstay of molecular diagnostics in many clinical microbiology laboratories across the world for detecting bacteria, viruses, fungi and parasites in clinical samples (Espy *et al.* 2006; Verweij and Stensvold, 2014), with a comprehensive number of published qPCR assays for pathogen detection.

However, operational issues such as the high cost of the instrumentation required for thermal cycling and signal detection and the need for well-equipped laboratories, have so far hindered use of qPCR in resource-poor settings. In the past few years, isothermal methods allowing the amplification of nucleic acids at constant temperature and with minimal equipment have been described and presented as more 'field-friendly' and cost-effective alternatives to qPCR, particularly for their potential use as point-of-care tests (Craw and Balachandran, 2012). Amongst these methods, the LAMP (Notomi *et al.* 2000) is the most researched and characterized for parasite diagnostics. For example, a variety of LAMP assays have been published for the detection of malaria (Han, 2013) and trypanosomes (Deborggraeve and Büscher, 2012). Another isothermal amplification technique, known as NASBA is a method, which amplifies RNA templates, mimicking the retroviral replication mechanism, and is particularly suited for the diagnosis of RNA viruses and in detecting viable targets (Craw and Balachandran, 2012). Although NASBA has been successfully applied to the diagnosis of parasitic infections such as leishmaniasis (de Ruyter *et al.* 2014) and trypanosomiasis (Mugasa *et al.* 2014), it has yet to be applied to human helminths. More recently a further isothermal method, the RPA (Piepenburg *et al.* 2006) has come to the attention of the scientific community. In RPA, target amplification is rapidly

Table 1. Current obstacles, potential solutions and research priorities for the large-scale implementation of molecular diagnostics in MDA-targeted helminths control and elimination programmes

Obstacle	Potential solutions	Research priorities
'Cost' of molecular diagnostics is prohibitive for current helminth NTDs programmes	Reduce sample collection and processing costs by using existing NTDs diagnostic pathways as a source for suitable human and vector samples	<ul style="list-style-type: none"> <li>• Foster a greater collaboration across groups working on different diseases</li> <li>• Optimize sample collection, storage and extraction methods for helminth NTDs screening</li> </ul>
	Reduce the cost and increase the throughput capacity of molecular screening by devising strategies for pooling samples (increasing sample size) and detecting multiple pathogens from the same samples	<ul style="list-style-type: none"> <li>• Optimize sample pooling methods (alongside extraction and amplification) to increase sample size without sacrificing sensitivity of detection</li> <li>• Further development of multiplex assays</li> </ul>
	Increase the investments from key stakeholders to support post-MDA surveillance implementation of molecular diagnostics	<ul style="list-style-type: none"> <li>• Modelling the costs of programme failure compared with the costs of surveillance programme</li> <li>• Modelling the actual costs of molecular diagnostics compared with the 'gold standards', including factors such as sample volumes, frequency of sampling and equipment, supplies, personnel and training costs</li> </ul>
'Standard Method' – No clear consensus on which molecular diagnostic platform(s) would best serve the needs of control and elimination programmes	Ensure the platform that is most appropriate for a particular setting and that responds to the needs and capabilities of control programmes is used, to balance costs and diagnostic effectiveness	<ul style="list-style-type: none"> <li>• Complete costing and thorough comparison of the available molecular diagnostic platforms as applied to different helminth diseases and scenarios</li> <li>• Identify (with inputs from key stakeholders) new molecular diagnostic platforms and approaches to address the changing needs of control and elimination programmes</li> </ul>
	Use of standard quality assessment schemes and of quantified reference materials to ensure robust and comparable standards between programmes and countries Promote non-invasive sampling strategies to maintain community support and to simplify sampling logistics	<ul style="list-style-type: none"> <li>• Establishment of globally agreed standard quality assessment schemes and of a central a resource of quantified reference materials</li> <li>• Further development of molecular diagnostic assays using non-conventional biological samples (urine, saliva, insect excreta)</li> </ul>
'Result Interpretation' – Little or no guidance on how to use and interpret molecular diagnostics to confirm residual transmission and/or local elimination	Develop suitable guidelines for the determination of the level of infection intensity (rather than prevalence) needed for sustained transmission	<ul style="list-style-type: none"> <li>• Evaluation of the potential for molecular diagnostic assays to precisely quantify the worm burden in the human host and vectors, also as indicators of morbidity</li> </ul>
	Develop a focal sampling strategy aimed towards reservoirs of infection	<ul style="list-style-type: none"> <li>• Develop risk maps (including factors such as human and vector behaviour) which can be used to direct sampling for molecular diagnostics</li> </ul>

accomplished (20–40 min) using a combination of target-specific primers and a recombinase enzyme (which form a recombinase filament inserting the primers when the specific target is encountered via a strand exchange mechanism) and single-strand DNA-binding proteins, which stabilizes and prevents re-annealing of the DNA double-strand

(Craw and Balachandran, 2012). The RPA approach has been applied for the detection of a variety of viral pathogens, including Ebola (Faye *et al.* 2015) and also of urinary schistosomiasis (Rosser *et al.* 2015). With increasing focus towards in-country laboratory capacity building and the necessity for integrating interventions and monitoring activities for different

NTDs, molecular diagnostic methods are now suitable to explore in terms of value and effectiveness in helminth control and elimination programmes.

#### AIMS OF THE REVIEW

The aims of this review are threefold. First, to compare the features of qPCR and the two isothermal tools for DNA amplification (LAMP and RPA) for which examples in human helminths exist, and so are relevant for their potential use in control and elimination programmes of the four major human helminthiases. Second, to review the qPCR, LAMP and RPA assays currently available for patient and/or community diagnosis or xenomonitoring of LF, onchocerciasis, STHs and schistosomiasis. Third, to identify the research needs and implementation issues which should be considered for the large-scale adoption of such molecular diagnostic tools to assist the control and elimination programmes, particularly during the post-MDA surveillance stages.

#### WHAT FEATURES REALLY MATTER: qPCR VS LAMP VS RPA

When comparing the features of different molecular detection formats or assays, it is important to consider the characteristics most relevant to the purpose for which the tools are meant to be used. Analytical or diagnostic sensitivity and specificity are influenced by how the assay is designed and optimized and how samples are processed before testing (Verweij and Stensvold, 2014). For example, choosing a molecular target with multi-copy genes and using an appropriate method to extract nucleic acids efficiently (in terms of both yield and purity) from samples will allow higher sensitivity of detection. Following careful design and extensive testing of extraction methods and reaction set up, similar and very low analytical sensitivities can be accomplished using either qPCR or isothermal assays (Han, 2013). However, the lack of standardization in protocols between laboratories and published papers makes a direct comparison of assays based on the above characteristics difficult: variability between laboratories is commonly observed in terms of sample volumes, storage conditions and extraction methods (de Ruiter *et al.* 2014). Also, when comparing the sensitivity of detection between a new molecular assay and one that is already available, the choice of the comparison assay is crucial. For example, the sensitivity of LAMP assays is commonly reported against a PCR performed using the same two primers employed in the LAMP reaction. In general and whenever possible, comparison should be made against the most sensitive assay available.

Cost-effectiveness and scalability of molecular diagnostics will depend on equipment and reaction

costs, throughput capacity and the ability for tests to be multiplexed to amplify multiple targets from the same sample. It will be crucial to optimize these features for use in control and elimination campaigns of MDA-targeted helminths. A general comparison of qPCR, LAMP and RPA based on some of these features is shown in Table 2.

#### *Cost of reagents and equipment and the issue of target quantification*

Real-time qPCR has a high cost associated with reagents and equipment needed for thermal cycling and signal detection, although the cost can vary between models and according to the machine multiplex capacity. During qPCR the amplified target detection happens ‘real-time’ using fluorescent intercalating dyes binding to the nucleic acids (such as SYBR-Green) or fluorescence-labelled probes (such as TaqMan probes) (Verweij and Stensvold, 2014): this feature makes the detection of amplification quantitative, allowing the determination of the pathogen load in samples. Quantifying parasite load is a key aspect in the diagnostics of helminths and assessment of morbidity and control, where determining infection intensity is more important than calculating prevalence (McCarthy *et al.* 2012).

Isothermal methods can amplify nucleic acids at a constant (and relatively lower, if compared with qPCR) temperature using a simple water bath or heat block: in the case of RPA, successful amplification using simply the body heat of the operator has been proven possible (Crannell *et al.* 2014). A wide range of approaches are available for the qualitative detection of LAMP and RPA products, including gel electrophoresis, in-tube fluorescence, lateral flow detection with immunochromatographic strips (Craw and Balachandran, 2012) and also by naked eye using pH-sensitive dyes (Tanner *et al.* 2015). The ability to perform nucleic acid amplification in the absence of expensive instrumentation, under field conditions in resource-poor setting, has been the major selling point for LAMP and similar methods as a ‘new generation’ of molecular diagnostic tools. However, if quantification of the target is needed, the advantage of isothermal methods over qPCR in terms of equipment requirements vanishes since detection must be accomplished by turbidimetric or fluorometric approaches requiring similar or identical equipment to qPCR (Zhang *et al.* 2014).

#### *One or more than one pathogen(s)? The issue of multiplexing*

Another important consideration to be made when estimating the cost per reaction for each platform is whether the aim is to detect one (singleplex) or more than one (multiplex) targets (parasite species

Table 2. Qualitative comparison of qPCR, LAMP and RPA based on features relevant to implementation of molecular diagnostics for MDA-targeted helminths control and elimination

Method	Equipment cost QL & QT	Cost/reaction SX & MX <sup>a</sup>	Throughput capacity	Multiplexing capacity	Target quantification	Sensitivity to inhibitors	Genotyping capacity	Essential references
qPCR	– & –	(±) & +	+	+	+	(±)	+	(Espy <i>et al.</i> 2006; Verweij and Stensvold, 2014)
LAMP	– & +	+ & unknown	–	(±)	(±)	+	(±)	(Notomi <i>et al.</i> 2000; Zhang <i>et al.</i> 2014; Craw and Balachandran, 2012)
RPA	– & +	– & unknown	–	(±)	(±)	+	(±)	(Piepenburg <i>et al.</i> 2006; Craw and Balachandran, 2012)

The ‘positive’/‘negative’ symbols (+, ±, –) indicate qualitatively assigned comparative suitability, performance, benefit and/or advantage of methods for the indicated features as follows: ‘+’ indicates relative comparable superior suitability performance, increased benefit and/or advantage; ‘±’ indicates relative comparable/no difference in suitability, performance, similar benefit and/or equitable vantage; ‘–’ indicates relative comparable inferior suitability, performance, reduced benefit and/or disadvantage.

<sup>a</sup> Excluding sample collection, extraction, consumables and primers costs: qPCR SX \$0.55 (intercalating dye detection) – \$0.81 (probe detection), MX (two targets, probe detection) \$0.87 (Source: <http://blog.biosearchtech.com/how-much-is-your-qpcr-assay-really-costing-you>); LAMP SX \$0.26 (estimated using the *Bst* 2·0 DNA polymerase kit from New England Biolabs: \$264 for 1 mL of 8000 units mL<sup>-1</sup> of enzyme for 1000 reactions at 1 µL of enzyme in 25 µL total reaction volume per manufacturer’s standard protocol); RPA SX \$4.27 (estimated using the TwistAmp<sup>®</sup> Basic kit from TwistDX: \$410 for 96 reactions). qPCR, quantitative (real-time) polymerase chain reaction; LAMP, loop-mediated isothermal amplification; RPA, recombinase polymerase amplification; QL, qualitative detection; QT, quantitative detection; SX, singleplex reaction (one target); MX, multiplex reaction (more than one target); mL, millilitre; µL, microlitre; \$, US dollar



in this case) simultaneously. Currently, RPA has the highest cost per sample and LAMP the lowest for a single target reaction (Table 2). The cost-effectiveness of qPCR at less than a dollar for a singleplex reaction increases as more targets are amplified and detected in the same tube: recent estimates showed that the cost per reaction tube increases only slightly from a single target (\$ 0.82) up to, for example, four targets amplified (\$ 1.01). As a consequence, the multiplexing capacity of qPCR implies a significant decrease in the hands-on time and increase in throughput capacity. The upper limit of targets that can be amplified in a multiplex qPCR assay using, for example, TaqMan probes depends on the number of different dyes the machine is able to detect, but using microfluidic technologies such as TaqMan Array Cards (TACs) can allow the detection of up to 19 different targets (Liu *et al.* 2013). Whether the same multiplexing dynamics can in time be applied to the isothermal tools remain to be seen. Although the possibility of amplifying multiple targets has been shown in both LAMP (Aonuma *et al.* 2010; Zhang *et al.* 2014) and RPA (Crannell *et al.* 2016), both the multiplexing and sample throughput capacity of these approaches are still reduced if compared with qPCR and are not well established at present.

#### *Parasite genotyping: an added value*

The possibility of obtaining additional information on genetic makeup of parasites alongside detection should be considered an added value too. For example, molecular diagnostics can be applied to the detection of single nucleotide polymorphisms (SNPs) in the beta-tubulin gene sequence, which has been shown to be associated with resistance to benzimidazole drugs in a variety of animal nematode parasites (Kotze *et al.* 2014), and potentially STHs (Diawara *et al.* 2013) and *Wuchereria bancrofti* (Schwab *et al.* 2005). The detection of SNPs by qPCR using TaqMan probes is well established (Espy *et al.* 2006), whereas it is still in its infancy for isothermal tools such as LAMP and RPA (Craw and Balachandran, 2012). Where and when drug resistance genes and alleles are known, inclusion of such molecular tests within standard diagnostic monitoring surveys could ensure early detection of emergence of potential drug resistance and thus inform appropriate responsible strategies in judicious use of existing precious drugs on which MDA programmes rely significantly.

#### *Are we really considering all costs?*

It is important to stress that the real cost-effectiveness of molecular diagnostics must also include costs, which are usually overlooked, such as those associated with sample collection, storage and extraction. Particularly in relation to assay sensitivity

and throughput capacity, the extraction of DNA and/or RNA from the samples is a crucial step (Verweij and Stensvold, 2014). Depending on the nature and number of samples processed, extractions can be expensive and time-consuming. The possibility of running molecular assays using unprocessed or minimally processed samples is highly attractive in terms of reducing the cost and time of extractions. It is a common belief that isothermal methods are more robust than PCR-based methods are to amplification inhibition due to sample contaminants, as shown in publications reporting pathogen detection by RPA directly from urine (Krölov *et al.* 2014) or by LAMP from crude or boiled insects (Alhassan *et al.* 2014; Nzelu *et al.* 2014) or freshwater samples (Koloren *et al.* 2011). Although these results are promising, the effect of amplification inhibition on qPCR, LAMP and RPA must be thoroughly tested in different samples to avoid compromising the sensitivity of detection in seeking to reduce extraction costs.

#### OVERVIEW OF AVAILABLE qPCR, LAMP AND RPA ASSAYS FOR EACH HELMINTHIASIS

The major qPCR, LAMP and RPA assays described for the human diagnosis and xenomonitoring of LF, onchocerciasis, STHs and schistosomiasis are summarized in Table 3.

#### *Lymphatic filariasis*

PCR detection of the causative agents of LF *W. bancrofti* and *Brugia malayi* in blood samples or mosquitoes has been widely explored through a variety of formats and gene targets (Alhassan *et al.* 2015). Real-time PCR assays have been developed using highly repeated sequences in parasites' genomes, namely SspI (Lulitanond *et al.* 2004) and LDR (Rao *et al.* 2006a) in *W. bancrofti* and HhaI (Rao *et al.* 2006b) in *Brugia*. The *W. bancrofti* LDR qPCR has been successfully used to amplify *W. bancrofti* DNA from dried nucleopore membranes stored from 6 months up to 4 years, as well as from dried blood on ICT sample application pads and individual mosquitoes (Rao *et al.* 2006a). The same assay later allowed a quantification of microfilarial load from dried blood spots, and proved to be suitable for screening pooled blood spots with a detection limit of one positive spot detected in a pool of 12 negative aliquots (Plichart and Lemoine, 2013). The detection of *B. malayi* in blood by qPCR has also been successfully achieved (Rao *et al.* 2006b). In this study, 17 out of 31 (54.8%) amicrofilaremic blood samples collected during the day tested positive by PCR. A PCR-positivity rate in day-collected blood samples similar to night-collected samples has been also observed in areas with nocturnally periodic *W. bancrofti* (Lucena *et al.* 1998).

Table 3. Overview of qPCR, LAMP and RPA assays published for diagnosis and xenomonitoring of the MDA-targeted helminths

Parasite species	Assay type	Target gene(s)	Sample type(s)	Detection limit (comparison assay)	Performance on clinical or field samples (comparison assays, if any) [gold standard test, if any]	Reference(s)
Filarial worms						
<i>W. bancrofti</i>	qPCR	LDR	Human blood (dried blood spots, nucleopore membranes and ICT sample pads), mosquitoes	1 mf/50 µL dried blood diluted in 12 negative blood spots	Blood: PR 24·3% (microscopy 17·1%; ICT 53·4%) Mosquitoes: PR 31% (PCR 22%)	Rao <i>et al.</i> (2006a); Plichart and Lemoine (2013)
<i>W. bancrofti</i>	qPCR	SspI	Mosquitoes	1 L3 larva/mosquito	SE 100%, SP 100% [experimental infections]	Lulitanond <i>et al.</i> (2004)
<i>W. bancrofti</i> , <i>P. falciparum</i>	Multiplex qPCR	LDR, 18S rDNA ( <i>Plasmodium</i> )	Pooled mosquitoes (~8/pool)	0·1 pg gDNA	PR: <i>W. bancrofti</i> 31·8%, <i>P. falciparum</i> 19·2%, <i>P. vivax</i> 14·8%	Rao <i>et al.</i> (2009))
<i>B. malayi</i>	qPCR	HhaI	Human blood	0·1–100 fg gDNA	PR 66·6% (microscopy 63·1%)	Rao <i>et al.</i> (2006b)
<i>W. bancrofti</i> , <i>B. malayi</i>	Multiplex qPCR	SspI, HhaI	Mosquitoes (experimentally infected)	1 L3/30 pooled mosquitoes (both species)	ND	Intapan <i>et al.</i> (2009)
<i>W. bancrofti</i>	RT-qPCR	Wb cut-1·2	Mosquitoes (experimentally infected)	One infected mosquito/30 pooled	ND	Laney <i>et al.</i> (2010)
<i>W. bancrofti</i>	LAMP	Nuclear scaffold/matrix attachment region	Human blood, mosquitoes spiked with mf, field-collected mosquitoes	0·1 pg gDNA (1/10th mf); 1 mf mL <sup>-1</sup> blood; 1 mf/60 pooled mosquitoes (=SspI PCR)	Individual mosquitoes PR 1·8% pooled mosquitoes PR 0·31%	Takagi <i>et al.</i> (2011); Kouassi <i>et al.</i> (2015)
<i>B. malayi</i> , <i>B. timori</i>	LAMP	HhaI	Cat blood	0·5–1 pg gDNA (1/200th–1/100th mf) (=Hha I PCR); 1 mf/40 µL blood	ND	Poole <i>et al.</i> (2012)
<i>O. volvulus</i>	qPCR	O-150	Skin snips	10 fg gDNA µL <sup>-1</sup> (1 pg PCR)	PR: qPCR 56·9% (PCR 33·9%; skin snip microscopy 26%; nodule palpation 37·9%)	Lloyd <i>et al.</i> (2015)
<i>O. volvulus</i>	LAMP	O-150	Skin snips	0·1 pg gDNA	PR: 65·7% (PCR F3/R3 37·1%, microscopy 31·4%)	Alhassan <i>et al.</i> (2016)
<i>O. volvulus</i>	LAMP	OvGST1a	Black flies spiked with parasite DNA	0·01 ng gDNA (1/10th mf) (=OvGST1a PCR F3/R3); 1 mf/200 pooled flies	ND	Alhassan <i>et al.</i> (2014)
Soil-transmitted helminthes						
<i>A. lumbricoides</i> , <i>A. duodenale</i> , <i>N. americanus</i> , <i>S. stercoralis</i>	Multiplex qPCR	ITS1,2 18S rDNA	Human feces	10–10 <sup>3</sup> copies	<i>Ascaris</i> : PR 0·03% (0·04) Hookworms: PR 0·18% (0·01) <i>Strongyloides</i> : PR 0·19% (0·03) (KK coproscopy)	Basuni <i>et al.</i> (2011)

Table 3. (Cont.)

Parasite species	Assay type	Target gene(s)	Sample type(s)	Detection limit (comparison assay)	Performance on clinical or field samples (comparison assays, if any) [gold standard test, if any]	Reference(s)
<i>A. lumbricoides</i> , <i>A. duodenale</i> , <i>N. americanus</i> , <i>S. stercoralis</i> , <i>T. trichiura</i>	qPCR	ITS1,2 SSU rDNA	Human feces	0.001 fg gDNA $\mu\text{L}^{-1}$	<i>Ascaris</i> : PR 7% (5.5) Hookworms: PR 0.5% (0) <i>Strongyloides</i> : PR 0.7% (0) <i>Trichuris</i> : PR 3% (0.7) (direct coproscopy)	Mejia <i>et al.</i> (2013)
<i>A. lumbricoides</i> , <i>T. trichiura</i> and 17 other enteropathogens	Multiplex qPCR	ITS1 SSU rDNA	Human feces	$10^6$ copies $\text{g}^{-1}$ feces	SE 100%, SP 98.7% [coproscopy]	Liu <i>et al.</i> (2013)
<i>A. lumbricoides</i> , <i>A. duodenale</i> , <i>N. americanus</i> , <i>S. stercoralis</i> , <i>T. trichiura</i>	Multiplex qPCR	Species-specific, non-coding, high copy-number repeat sequences	Human feces	2 fg gDNA $\mu\text{L}^{-1}$	<i>Ascaris</i> : PR 59.4% (50.6) <i>Necator</i> : PR 75.9% (60.7) <i>Trichuris</i> : PR 22.8% (17.7) (multiplex qPCR)	Pilotte <i>et al.</i> (2016b)
<i>S. stercoralis</i>	LAMP	28S rDNA	Human feces	<10 copies; $10^{-2}/10^{-1}$ larvae /250 $\mu\text{L}$ feces	SE 100%, SP 100% [coproscopy and real-time PCR]	Watts <i>et al.</i> (2014)
<i>N. americanus</i>	LAMP	ITS2	Human feces	0.4 fg gDNA	SE 97%, SP 100% [KK coproscopy]	Mugambi <i>et al.</i> (2015)
Schistosomes <i>S. haematobium</i> , <i>S. mansoni</i>	Multiplex qPCR	Cox1	Human stools	100 fg gDNA (both species)	<i>S. haematobium</i> : PR 55% (0%) <i>S. mansoni</i> : PR 73% (80%) (duplicate KK)	ten Hove <i>et al.</i> (2008)
<i>S. haematobium</i> , <i>S. mansoni</i>	Singleplex qPCR	IGS	Snails	0.05 pg gDNA ( <i>haematobium</i> ) 0.1–1 pg gDNA ( <i>mansoni</i> )	<i>S. haematobium</i> : PR 79–82% (82%) <i>S. mansoni</i> : PR 26.3–47.4% (21%) (PCR)	Kane <i>et al.</i> (2013)
<i>S. haematobium</i>	LAMP	IGS	Human urine	100 fg gDNA (1–10 fg spiked in urine)	SE 100%, SP 86.67% [urine filtration microscopy]	Gandasegui <i>et al.</i> (2015)
<i>S. haematobium</i>	RPA	DraI	Parasite DNA spiked with urine	100 fg gDNA	ND	Rosser <i>et al.</i> (2015)
<i>S. haematobium</i> , <i>S. mansoni</i>	LAMP	DraI Sm1-7	Snails	0.1 fg gDNA (10 fgPCR)	ND	Abbasi <i>et al.</i> (2010); Hamburger <i>et al.</i> (2013)
<i>S. mansoni</i>	LAMP	Mitochondrial minisatellite region	Mouse feces	1 fg gDNA	ND	Fernández-Soto <i>et al.</i> (2014)



<i>S. japonicum</i>	LAMP	SjR2	Human serum	0.08 fg gDNA (0.8 pg PCR)	SE 95.5–96.7%, SP 100% (PCR SE 60%, SP 100%; ELISA SE 84.6%, SP 85.7%; IHA SE 91.8%, SP 88.1%) [KKK]	Xu <i>et al.</i> (2010, 2015)
<i>S. japonicum</i>	LAMP	28S rDNA	Snails	100 fg gDNA (1 miracidium) (1 pg PCR); 1 infected/100 uninfected snails	PR: 25% (PCR 14%)	Kumagai <i>et al.</i> (2010); Tong <i>et al.</i> (2014)

ND, not determined; gDNA, parasite genomic DNA; SE, sensitivity; SP, specificity; PR, prevalence; PCR, polymerase chain reaction; PCR F3/R3, PCR using the LAMP outer primers F3 and R3; qPCR, real-time quantitative polymerase chain reaction; RT-qPCR, reverse transcriptase quantitative PCR; LAMP, loop-mediated isothermal amplification; RPA, recombinase polymerase amplification; ELISA, enzyme-linked immunosorbent assay; IHA, indirect haemagglutination assay; KK, Kato-Katz; LDR, SspI, *W. bancrofti* repeat sequences; Hha I, *Brugia* repeat sequence; Wb cut-1-2, *W. bancrofti* cuticular collagen gene; O-150, *O. volvulus* tandem repeat; Ov-GST1a, *O. volvulus* glutathione S-transferase 1a; SSU rDNA, small subunit ribosomal DNA (nuclear); 12S rDNA, 12S ribosomal DNA (mitochondrial); 16S rDNA, 16S ribosomal DNA (mitochondrial); 28S rDNA, 28S ribosomal DNA (nuclear); IGS, intergenic spacer region of the ribosomal DNA; ITS1,2, internal transcribed spacer region 1 or 2 of the ribosomal DNA; Cox1, cytochrome *c* oxidase subunit 1; SjR2, *S. japonicum* repetitive retrotransposon; DraI, *S. haematobium* tandem repeat; Sml1,7, *S. mansoni* tandem repeat; g, gram; ng, nanogram; pg, picogram; fg, femtogram; mL, millilitre; mf, microfilariae; L3, L3 larva.

Xenomonitoring will play a major role in the ‘endgame’ final stages of elimination programmes and post-MDA surveillance of LF, since it gives proxy estimates of disease burden and existing potential for transmission while also being less invasive for the community (Okorie and de Souza, 2016). Furthermore, a reduction in infection rates in vectors offers a valuable demonstration of interruption of transmission (Owusu *et al.* 2015). The detection of filarial DNA or RNA in mosquitoes has several advantages over the traditional method of dissections. For example, it is relatively faster since it can be used on pooled mosquitoes and is less reliant on labour intensive and highly skilled dissection methods; it is more sensitive and allows the species-specific identification of filarial worms (Okorie and de Souza, 2016). During an entomological survey in American Samoa, the estimated infection rate in pooled *Aedes polynesiensis* was 0.68% compared with the overall 0.11% rate observed by dissection of individual mosquitoes (Chambers *et al.* 2009). Real-time qPCR has recently been used for detection of *W. bancrofti* in mosquito populations across the world, including Sri Lanka (Rao *et al.* 2014), Samoa (Hapairai *et al.* 2015), Ghana (Owusu *et al.* 2015) and Tanzania (Irish *et al.* 2015). Under the perspective of a programmatic use in detection of filarial DNA from both blood and mosquitoes as part of post-MDA surveillance and verification of LF elimination, the possibility of applying the same molecular test to both sample types could potentially constitute a cost and operational advantage since laboratories in developing countries will be required to be proficient in only one assay.

Multiplex assays are also available, including the simultaneous detection of both *W. bancrofti* and *B. malayi* with a sensitivity of a single L3 larva of either species in up to 30 pooled mosquitoes (Intapan *et al.* 2009). Of note is the duplex qPCR assay co-amplifying DNA from *W. bancrofti* and two species of *Plasmodium* (*Plasmodium falciparum* or *P. vivax*) in mosquito pools (Rao *et al.* 2009), which also showed a better sensitivity compared to the respective singleplex assays. This method has yet to be tested for its applicability in diagnosing filariasis and malaria coinfection in human blood, but it may prove useful in determining infection rates in mosquitoes from areas of filariasis and malaria co-endemicity. The possibility of diagnosing coinfections has been explored before using a ligation detection reaction-fluorescent microsphere assay (LDR-FMA) multiplex PCR on blood samples of 517 people from Papua New Guinea (Mehlotra *et al.* 2010). The assay, capable of detecting simultaneously *W. bancrofti* and four species of *Plasmodium*, revealed a frequent co-occurrence of bancroftian filariasis and any of the four malaria parasites in 29% of samples.

The ability to discriminate infective mosquitoes (harbouring the L3 infective stage of the parasite and thus contributing to transmission) from mosquitoes which, although having taken an infected blood meal do not support the development of the parasite and thus remain unable to infect the human host, is essential to obtain more precise estimates of actual disease transmission potential from xenomonitoring data (Okorie and de Souza, 2016). For this purpose, reverse transcriptase (RT) PCR and qPCR assays have been developed for detection of L3 larvae-specific RNA transcripts in mosquitoes (Laney *et al.* 2008, 2010; Vasuki *et al.* 2008). The assays showed a detection limit of a single L3 larva in pools of up to 25 mosquitoes (Laney *et al.* 2008; Vasuki *et al.* 2008) or one infected mosquito in a group of 30 uninfected insects (Laney *et al.* 2010). The qPCR assay for *W. bancrofti* has also been multiplexed to detect both one L3 larvae-specific transcript and one further transcript present in all parasite stages (Laney *et al.* 2010) thus allowing simultaneous detection of infected and infective insects in one reaction. Neither of the RT-PCR assays has been tested to date on field-collected mosquitoes from entomological surveys.

In terms of isothermal diagnostic tools, only one LAMP assay has been developed for bancroftian filariasis (Takagi *et al.* 2011). Targeting a portion of the parasite nuclear scaffold/matrix attachment region, the method successfully amplified the DNA extracted from a single *W. bancrofti* microfilaria spiked in one millilitre of human blood or in a pool of 60 *Culex* mosquitoes. This LAMP assay has recently been used in Ghana screening for *W. bancrofti* infection in 112 individual *Anopheles* and 3635 *Culex* mosquitoes (arranged in 184 pools, with a range of 7–20 insects per pool) (Kouassi *et al.* 2015). However, the very low number of positive mosquitoes (1.8%) and pools (0.31%) and the absence of a reference test prevented sufficient evaluation of the assay performance. Nevertheless, the two single mosquitoes and seven out of 11 (64%) of pools testing positive by LAMP, subsequently tested negative by PCR.

A LAMP assay has also been developed for the detection of *B. malayi* and *B. timori*, based on the tandem repeat HhaI (Poole *et al.* 2012). The assay showed the same limit of detection (corresponding to approximately 1/100th of a microfilaria) as the original HhaI PCR assay (Lizotte *et al.* 1994) when pure parasite genomic DNA was used, but a lower limit (1/200th of a microfilaria) was observed when diluted cat blood was used as template due perhaps to dead parasite DNA circulating in blood (Poole *et al.* 2012).

Despite the potential applications and the possibility of developing a less invasive testing, the amplification of filarial DNA from samples such as sputum or urine has been attempted only in a few instances. Using PCR assays targeting the *W. bancrofti* repeated

elements SspI and AccI, DNA of the parasite was amplified from the sputum of 32 (94%) out of 34 patients with confirmed bancroftian filariasis with a detection limit of one infected sputum sample mixed with 14 uninfected samples (Abbasi *et al.* 1999). In a latter study involving 304 people from Kenya, the sputum SspI PCR detected the presence of the parasite in more patients (42.8%) compared with ICT (39.1%) and blood microscopy (22%) (Kagai *et al.* 2008). Thus, taking ICT as ‘gold standard’, the calculated sensitivity and specificity for sputum PCR was 97.5 and 92.4%, respectively. The DNA of *W. bancrofti* has also been successfully detected in urine from 39 people from Brazil using the same PCR assay (Lucena *et al.* 1998) for an overall detection rate of 48.7–59% (in the first and second urine samples collected in the morning, respectively).

The possibility of detecting filarial DNA shed in the insect excreta and/or feces as a form of passive xenomonitoring tool (allowing then to bypass the need for pooling and extracting large numbers of mosquitoes) has been recently explored (Pilotte *et al.* 2016a). In this study, the authors showed that using the HhaI repeat qPCR assay the DNA of *B. malayi* can be detected in the excreta/feces of a single mosquito (fed with microfilaraemic blood) in pooled excreta from up to 500 uninfected insects. The same approach was also used to detect the presence of DNA from the malaria parasite *Plasmodium vivax* (Pilotte *et al.* 2016a). The high throughput and level of adaptability of excreta/feces collection and DNA testing to a variety of mosquito trapping systems makes this novel approach to xenomonitoring of LF very attractive and worth investigating in the near future, bearing also in mind its potential applicability for the surveillance of other blood-borne and mosquito-borne pathogens such as malaria or arboviruses.

### *Onchocerciasis*

Molecular detection tools for *Onchocerca volvulus* have largely been based on the PCR amplification of the parasite-specific O-150 tandem repeat from skin snips, particularly in the assessment of skin microfilarial load reduction following PC (Alhassan *et al.* 2015), and also from individual and pooled black flies in the Americas (Rodríguez-Pérez *et al.* 1999). The O-150 PCR has been adapted into a qPCR format, showing increased sensitivity of detection of infection in skin snips compared with either microscopy (Fink *et al.* 2011; Lloyd *et al.* 2015) and also nodule palpation (Lloyd *et al.* 2015). The TaqMan assay developed allowed a semi-quantitative estimation of the microfilarial load from skin snips in only 30 min (Lloyd *et al.* 2015).

The applicability of the O-150 assay for onchocerciasis xenomonitoring in Africa is partially hindered

by the need for a DNA probe hybridization step to distinguish *O. volvulus* from the often co-endemic cattle parasite *Onchocerca ochengi*, which is transmitted by the same vectors (Alhassan *et al.* 2015). However, the recently developed TaqMan assay did not cross-amplify DNA from *O. ochengi* (Lloyd *et al.* 2015) and so its future evaluation for the detection of *O. volvulus* in African black flies appears promising. The detection of *Onchocerca* DNA in the urine of 38 people from an endemic area in Cameroon has also been reported, but with a significantly lower sensitivity (14%) compared with either skin snip PCR (90%) or serum antigen detection (100%) (Vincent *et al.* 2000).

To date, only one LAMP assay has been developed for specific detection of *O. volvulus* DNA in black flies (Alhassan *et al.* 2014). The assay was capable of amplifying an amount of DNA equivalent to a single microfilaria spiked in a pool of up to 200 flies, irrespective of whether purified DNA or simply boiled flies were used as template. In the latter case using boiled flies, a reduction in analytical sensitivity was observed for PCR (one mf in 150 flies) (Alhassan *et al.* 2014).

The applicability of LAMP to detect *Onchocerca* microfilariae in skin snips has also been tested using the O-150 tandem repeat as a target (Alhassan *et al.* 2016). When applied to skin snips from 70 individuals living in endemic areas of Ghana, the LAMP assay showed a significantly higher sensitivity (67.5% prevalence) compared with both PCR using the external LAMP primers (37.1%) and microscopy (31.4%). Additionally, the detection of amplified parasite DNA was accomplished visually using pH-sensitive dyes (Alhassan *et al.* 2016). To date, the amplification of L3 larvae-specific transcripts by RT-PCR for the detection of infective black flies (as it has been for LF in mosquitoes) has yet to be reported.

#### Soil-transmitted helminths (STHs)

Infection with the soil-transmitted helminths *Ascaris lumbricoides*, the hookworms (*Ancylostoma duodenale*, *A. ceylanicum*, *Necator americanus*), *Trichuris trichiura* and *Strongyloides stercoralis* is usually determined by finding the parasite eggs or larvae in stools by microscopy or culture (for hookworms and *Strongyloides*). However, coproscopy (either by concentrated or direct wet smear, egg flotation or Kato-Katz method) is relatively insensitive and time consuming, being most suitable for only moderate-to-heavy intensity egg loads, and thus light infections may typically be missed.

Molecular and biochemical assays that may confirm the presence of fecally-diagnosed helminths (STHs and schistosomes under our consideration in this review) through detection of antigens or DNA could therefore offer greater sensitivity. It is of

course important to acknowledge detection limitations, as is the case in microscopic detection of eggs, when considering parasite antigens and DNA in low parasite burdens. As host parasite numbers fall, eggs become more difficult to detect and so DNA levels are reduced also. In the case of STHs, antigens and DNA would be present in feces from both eggs and the bodies of parasites themselves (through excretion/secretion, regurgitation, sloughing off from cuticle turnover, etc.), whilst schistosome antigens and DNA in feces would largely derive from eggs. Coproantigen detection methods are currently not commercially available or adopted for STHs or *S. stercoralis* diagnosis. Serological assays to detect antibodies against exposure to parasite antigens are also not available for STHs although several have been demonstrated in use for *Strongyloides* infections where the clinical importance of this infection may require greater attention [for recent reviews see (Buonfrate *et al.* 2015) and (Krolewiecki *et al.* 2013)]. Since co-infection with multiple STH species is common in endemic areas and the intensity of infection (worm burden) is the most important parameter to assess the efficacy of STH control programmes, the development, testing and refinement of multiplex and quantitative molecular assays should be considered a priority for these parasites.

In recent years a variety of PCR assays, both conventional and real-time, have been developed for the molecular detection of STHs in human stools (Verweij and Stensvold, 2014). Singleplex qPCR assays have been described and tested on the stools of 400 asymptomatic children from Ecuador (Mejia *et al.* 2013) and 99 people from Argentina (Cimino *et al.* 2015). In both studies, qPCR detected additional infections for all STHs, which were missed by microscopy, and there was significant correlation between the egg counts of *Ascaris*, *Trichuris* and hookworms in samples and concentration of parasite DNA determined by qPCR. In the Ecuador study, qPCR was also used to confirm complete clearance of *Ascaris* infection in a cohort of 125 children treated with albendazole and ivermectin (Mejia *et al.* 2013).

Multiplex PCR and qPCR assays have also been described and applied successfully in epidemiological studies globally (Basuni *et al.* 2011; Taniuchi *et al.* 2011; Liu *et al.* 2013; Gordon *et al.* 2015; Llewellyn *et al.* 2016). One particular assay (targeting *Ascaris*, *Ancylostoma*, *Necator* and *Strongyloides*) showed an overall sensitivity of detection eight times higher than using a combination of three wet smears and one Kato-Katz slide on 77 stool samples collected in Malaysia (Basuni *et al.* 2011). In another study on 680 stool samples from Timor-Leste and Cambodia, the prevalence of *Ascaris* and hookworm infections determined by multiplex PCR was respectively 1.2 and 2.9 higher than by microscopy

(Llewellyn *et al.* 2016). A multiplex PCR assay with Luminex beads was developed for detection of *Ascaris*, *Ancylostoma*, *Necator*, *Strongyloides* and three intestinal protozoa, showing high sensitivity (90–100%) and specificity (95–100%) on 319 stool samples previously tested by multiplex qPCR (Taniuchi *et al.* 2011). Two qPCR assays amplifying *Ascaris* and *Trichuris* DNA have also been incorporated into a TAC allowing the simultaneous detection of other 17 other pathogens (including amongst others rotavirus, *Salmonella*, *Vibrio cholerae*, *Entamoeba histolytica*, *Cryptosporidium* and *Giardia*) (Liu *et al.* 2013). In the Philippines, the simultaneous detection in stool samples of the two cestode parasites *Taenia saginata* and *T. solium* alongside *Ascaris* and hookworms has also been successful using a multiplex approach (Gordon *et al.* 2015). In all the above studies, qPCR showed an increased capacity in detecting polyparasitism compared with coproscopy.

In a study on 796 stool samples from Kenya, qPCR was significantly more sensitive than two Kato-Katz slides from a single stool sample in detecting *Ascaris* and hookworm infections (Easton *et al.* 2016). Furthermore significant correlation was also shown between these egg counts by Kato-Katz, the DNA concentration determined by qPCR and the number of expelled *Ascaris* worms, showing that qPCR can be a good predictor for worm burden. The possibility of using multiplex qPCR to monitor changes in STH infection prevalence and intensity over time and in particular following MDA should be considered a priority (Easton *et al.* 2016). A recent study clearly showed the importance of using novel technologies to refine and improve the diagnostic potential of multiplex qPCR as applied to STHs (Pilotte *et al.* 2016b), where next generation sequencing data were used to identify non-coding, high copy-number repeated DNA sequences in *Ascaris*, *Ancylostoma*, *Necator*, *Strongyloides* and *Trichuris*, which were then included in a novel multiplex qPCR assay. When compared with a previous pentaplex assay (Llewellyn *et al.* 2016) using field collected stool samples, the new assay showed improved sensitivity detecting consistently down to 2 femtograms (fg) of parasite DNA (corresponding to a small fraction of the estimated DNA content of a fertilized egg or single L1 larva of the species considered) (Pilotte *et al.* 2016b).

In contrast to qPCR the use of isothermal amplification methods for STH stool diagnosis has been overlooked so far, with the exception of two LAMP assays targeting *S. stercoralis* (Watts *et al.* 2014) and *N. americanus* (Mugambi *et al.* 2015), respectively. The former assay, amplifying a region of the *S. stercoralis* 28S ribosomal DNA sequence, showed excellent performance (100% sensitivity and specificity) when tested on 28 stool specimens

from patients with *S. stercoralis* infection confirmed by a combination of microscopy and real-time PCR (Watts *et al.* 2014). The latter assay, targeting the internal transcribed spacer region 2 of the *N. americanus* ribosomal DNA region, was used on 86 positive and 20 negative specimens (confirmed by Kato-Katz microscopy) detecting the parasite DNA with a 97% sensitivity and 100% specificity (Mugambi *et al.* 2015). The *N. americanus* ITS2 LAMP was not tested on DNA from the other two hookworms infecting humans (*A. duodenale* and *A. ceylanicum*), so its applicability as a general hookworm diagnostic tool from feces has yet to be verified.

### Schistosomiasis

Conventional and real-time PCR assays have been developed for the detection of *Schistosoma* spp. DNA in numerous types of clinical sample (feces, urine, plasma, serum, vaginal lavages) over the years, showing overall increased sensitivity of detection compared with either stool/urine microscopy or serology and antigen detection (Verweij and Stensvold, 2014). In only one case the amplification of different schistosomes, specifically *S. haematobium* and *S. mansoni*, has been attempted with a multiplex qPCR approach (ten Hove *et al.* 2008), using 176 stool samples collected over 2 weeks from 88 people in Senegal. Although the method failed to detect a small number of *S. mansoni* infections (confirmed by duplicate Kato-Katz), the DNA of *S. haematobium* was frequently detected in stools in the absence of eggs from this species (including people in which eggs were found in the urine and also a few in which the urine was egg-negative) (ten Hove *et al.* 2008).

Few LAMP assays have been published for schistosomal DNA detection in clinical samples. The first described assay was developed to target *S. japonicum* DNA in human serum (Xu *et al.* 2010). The assay was used on sera from 20 infected and 30 uninfected patients diagnosed by Kato-Katz microscopy with only one false negative detected by LAMP, which otherwise displayed high sensitivity and specificity (96.7 and 100%, respectively) compared to conventional PCR. In a successive study using sera from 110 infected and 42 uninfected patients (Xu *et al.* 2015), the same LAMP assay had a diagnostic performance (95.5% sensitivity and 100% specificity) which surpassed both enzyme-linked immunosorbent assay (ELISA) (84.6% sensitivity and 85.7% specificity) and the indirect haemagglutination assay (91.8% sensitivity and 88.1% specificity). In the same study, the LAMP assay also detected *S. japonicum* DNA in 10 out of 60 (16.7%) serum samples from people that tested negative by triplicate Kato-Katz, ELISA and IHA. The minimum time required to detect infection with the three methods



was also assessed using experimentally infected rabbits (Xu *et al.* 2015). While ELISA and IHA were positive only at 5 and 4 weeks post-infection respectively, LAMP was positive already at three days. The isothermal detection by LAMP of *S. haematobium* DNA in human urine has also been described (Gandasegui *et al.* 2015). With the ability to detect as little as 1 fg of parasite DNA using just the heated urine supernatant as template, the LAMP amplified all 18 *S. haematobium*-confirmed patient samples tested and eight out of 53 (15%) specimens negative for eggs by microscopy (Gandasegui *et al.* 2015).

The RPA method has also been tested for its potential to diagnose urinary schistosomiasis using oligochromatographic lateral flow strips (Rosser *et al.* 2015). Although the sensitivity and specificity of the method was not evaluated on patients' samples, the amplification efficiency for parasite DNA was tested adding variable volume percentages of crude urine (from 1.25 to 10% of total reaction volume) at two parasite DNA concentrations [100 fgs and 10 picograms (pg) L<sup>-1</sup>]. Significant amplification was observed even in the presence of 10% of added volume of crude urine where 10 pg of DNA was used, but not at lower DNA concentrations (Rosser *et al.* 2015). These preliminary results suggest further optimization of this RPA assay could provide a rapid diagnostic test requiring only minimal processing of urine samples. To date, only one LAMP assay is described for detection of *S. mansoni* in stool samples, and is described being tested using mice feces (Fernández-Soto *et al.* 2014). Providing a detection limit of 1 fg of parasite DNA the LAMP diagnostic gave positive results in mice at 1-week post-infection in comparison with the 4 weeks required for ELISA and 6 weeks for Kato-Katz microscopy (Fernández-Soto *et al.* 2014).

Xenomonitoring of schistosomes in the snail intermediate host has been advocated as a key component of monitoring and surveillance when emphasis shifts from disease control to elimination (Rollinson *et al.* 2013). Molecular tools also allow detection of prepatent infections in snails that do not shed parasite cercariae (Stothard *et al.* 2014). Furthermore, they also allow the precise identification of the infecting schistosome species which is crucial in areas where the snails act also as intermediate hosts for other animal schistosomes (Abbasi *et al.* 2012). PCR and multiplex PCR methods are published for schistosome DNA detection in snails, with some allowing concomitant snail species identification (Stothard *et al.* 2014). Particularly interesting are two qPCR assays allowing detection of *S. haematobium* and *S. mansoni* DNA from *Bulinus* spp. and *Biomphalaria* spp. snails by either fluorescence or oligochromatographic lateral flow strips (Kane *et al.* 2013). Both assays showed a comparable or even higher sensitivity of detection of parasite DNA than traditional PCR approaches.

Schistosome DNA detection in snail vectors by LAMP has been attempted too. One assay was capable of detecting DNA from either a single *S. japonicum* miracidium or one infected snail added to a pool of up to 100 uninfected snails (Kumagai *et al.* 2010). The assay has also been tested alongside a nested PCR assay using field-collected snails from different sites to determine the prevalence of *S. japonicum* (Tong *et al.* 2014). Overall in pools of 50 snails from each of the 28 sampled sites, parasite DNA was detected in seven out of 28 sites (25% prevalence) compared with only 4 (14.3% prevalence) detected by PCR. Two LAMP assays targeting *S. haematobium* and *S. mansoni* are also published and were compared for detection in snails (Abbasi *et al.* 2010). Both methods showed a detection limit 10 times lower (at 1 fg) than the standard PCR assays (at 10 fg). These two methods have also been used to test individual field-collected snails in Kenya (Hamburger *et al.* 2013) where the LAMP assay returned infection prevalence estimates comparable with those obtained through PCR and qPCR. Interestingly, authors reported the possibility of pre-mixing all reaction components and stabilizing them with the addition of 17.5% sucrose to attain prolonged storage at room temperature (Hamburger *et al.* 2013).

#### Concluding remarks and future directions

It is clear that a series of obstacles in terms of costs, agreement on methods/standards, interpretation and strategic application of result data must be overcome to pave the way for nucleic acid-based molecular diagnostics to become valuable tools for the control and elimination programmes of MDA-targeted helminths in general (as summarized in Table 1, with individual parasite-specific research priorities summarized in Box 1). In general, besides technical feasibility and laboratory capacity, the major issue related to the large-scale and routine implementation of molecular diagnostics is its cost-effectiveness. Robust assessment of cost-effectiveness will require careful consideration of several factors and may need to be viewed on a case-by-case basis depending upon the situation, stage and needs of the programmes, and stakeholder requirements. Aside from costs associated with equipment and reagents for sample extraction and amplification, other factors to consider include the time involved for the predicted number of samples to be screened and the sampling frequency, whether the samples are used to screen for multiple pathogens and all costs associated with planning and executing the sampling (supplies, transport, personnel and training). Since these aspects could be overlooked, applying both mathematical modelling and full cost-analysis should be considered pivotal in generating evidence about actual cost-effectiveness of molecular diagnostics and set against a numerical sampling



**Box 1. Proposed research priorities for the molecular diagnostics of specific MDA-targeted helminths as applied to control and elimination programmes**

- **LF:** optimization of mosquito pooling strategies and pool extractions, thorough field evaluation of the L3 larvae-specific RT-qPCR to assess mosquito infectivity as compared with the simple detection of DNA in vectors, explore the possibility of molecular xenomonitoring via indirect sampling strategies (insect excreta), exploration of less invasive sampling methods to determine infection prevalence in the communities (use of urine and saliva), using blood samples for the simultaneous screening of LF, malaria and schistosomiasis.
- **Onchocerciasis:** testing the existing O-150 qPCR assay for the xenomonitoring of onchocerciasis in African blackflies, developing an *O. volvulus* L3 larvae-specific RT-qPCR assay to detect infective black flies.
- **STHs:** Optimization of extant multiplex assays and development of new platforms for the simultaneous detection of all STHs (and schistosomes) from stools.
- **Schistosomiasis:** development of highly specific molecular assays for the xenomonitoring of schistosomes in snail populations, and assessment of their value for interruption of transmission evaluation, test snail pooling strategies for molecular detection using a similar approach as for LF.
- **All helminths:** further evaluation of the use of molecular detection as a quantitative indicator of worm burden and potentially morbidity.

framework to judge affordability when used in discrete applications, for example, to certify elimination. We propose that with a greater focus on laboratory capacity building, cost-effectiveness analysis, agreements on standards in methods (including in particular the formation of quality assessment schemes and the use of quantified reference material) and their strategic application, the full potential of current and future nucleic acid-based molecular diagnostics for MDA-targeted helminths could now realize the more rapid situational assessment required in achieving press timeframes stipulated in the WHO 2020 Roadmap.

ACKNOWLEDGEMENTS

C. M., L. R. and J. R. S. received funding from DFID, UK as part of the COUNTDOWN implementation research consortium on neglected tropical diseases (NTDs).

FINANCIAL SUPPORT

This work was supported by the Liverpool School of Tropical Medicine Research Development Fund (grant number RDF150303Jla02 to J. L., R. S. and C. M.).

CONFLICT OF INTEREST

None.

ETHICAL STANDARDS

Not applicable.

REFERENCES

- Abbasi, I., Githure, J., Ochola, J. J., Agure, R., Koech, D. K., Ramzy, R. M., Williams, S. A. and Hamburger, J. (1999). Diagnosis of *Wuchereria bancrofti* infection by the polymerase chain reaction employing patients' sputum. *Parasitology Research* **85**, 844–849.
- Abbasi, I., King, C. H., Muchiri, E. M. and Hamburger, J. (2010). Detection of *Schistosoma mansoni* and *Schistosoma haematobium* DNA by loop-mediated isothermal amplification: identification of infected snails from early prepatency. *American Journal of Tropical Medicine and Hygiene* **83**, 427–432.
- Abbasi, I., Hamburger, J., Kariuki, C., Mungai, P. L., Muchiri, E. M. and King, C. H. (2012). Differentiating *Schistosoma haematobium* from related animal schistosomes by PCR amplifying inter-repeat sequences flanking newly selected repeated sequences. *American Journal of Tropical Medicine and Hygiene* **87**, 1059–1064.
- Alhassan, A., Makepeace, B. L., LaCourse, E. J., Osei-Atweneboana, M. Y. and Carlow, C. K. S. (2014). A simple isothermal DNA amplification method to screen black flies for *Onchocerca volvulus* infection. *PLoS ONE* **9**, e108927.
- Alhassan, A., Li, Z., Poole, C. B. and Carlow, C. K. S. (2015). Expanding the MDx toolbox for filarial diagnosis and surveillance. *Trends in Parasitology* **31**, 391–400.
- Alhassan, A., Osei-Atweneboana, M. Y., Kyeremeh, K. F., Poole, C. B., Li, Z., Tettevi, E., Tanner, N. A. and Carlow, C. K. S. (2016). Comparison of a new visual isothermal nucleic acid amplification test with PCR and skin snip analysis for diagnosis of *Onchocerciasis* in humans. *Molecular and Biochemical Parasitology* 6–8 (in press). doi: 10.1016/j.molbiopara.2016.07.006.
- Aonuma, H., Yoshimura, A., Kobayashi, T., Okado, K., Badolo, A., Nelson, B., Kanuka, H. and Fukumoto, S. (2010). A single fluorescence-based LAMP reaction for identifying multiple parasites in mosquitoes. *Experimental Parasitology* **125**, 179–183.
- Basuni, M., Muhi, J., Othman, N., Verweij, J. J., Ahmad, M., Miswan, N., Rahumatullah, A., Aziz, F. A., Zainudin, N. S. and Noordin, R. (2011). A pentaplex real-time polymerase chain reaction assay for detection of four species of soil-transmitted helminths. *The American Journal of Tropical Medicine and Hygiene* **84**, 338–343.
- Buonfrate, D., Formenti, F., Perandin, F. and Bisoffi, Z. (2015). Novel approaches to the diagnosis of *Strongyloides stercoralis* infection. *Clinical Microbiology and Infection* **21**, 543–552.
- Chambers, E. W., McClintock, S. K., Avery, M. F., King, J. D., Bradley, M. A., Schmaedick, M. A., Lammie, P. J. and Burkot, T. R. (2009). Xenomonitoring of *Wuchereria bancrofti* and *Dirofilaria immitis* infections in mosquitoes from American Samoa: trapping considerations and a comparison of polymerase chain reaction assays with dissection. *The American Journal of Tropical Medicine and Hygiene* **80**, 774–781.
- Cimino, R. O., Jeun, R., Juarez, M., Cajal, P. S., Vargas, P., Echazú, A., Bryan, P. E., Nasser, J., Krolewiecki, A. and Mejia, R. (2015). Identification of human intestinal parasites affecting an asymptomatic peri-urban Argentinian population using multi-parallel quantitative real-time polymerase chain reaction. *Parasites & Vectors* **8**, 380.

- Crannell, Z. A., Rohrman, B. and Richards-Kortum, R. (2014). Equipment-free incubation of recombinase polymerase amplification reactions using body heat. *PLoS ONE* 9, 1–7.
- Crannell, Z., Castellanos-Gonzalez, A., Nair, G., Mejia, R., White, A. C. and Richards-Kortum, R. (2016). Multiplexed Recombinase Polymerase Amplification Assay to detect intestinal protozoa. *Analytical Chemistry* 88, 1610–1616.
- Craw, P. and Balachandran, W. (2012). Isothermal nucleic acid amplification technologies for point-of-care diagnostics: a critical review. *Lab on a Chip* 12, 2469.
- de Ruiter, C. M., Van Der Veer, C., Leeflang, M. M. G., Deborggraeve, S., Lucas, C. and Adams, E. R. (2014). Molecular tools for diagnosis of visceral leishmaniasis: systematic review and meta-analysis of diagnostic test accuracy. *Journal of Clinical Microbiology* 9, 3147–3155.
- Deborggraeve, S. and Büscher, P. (2012). Recent progress in molecular diagnosis of sleeping sickness. *Expert Review of Molecular Diagnostics* 12, 719–730.
- Diawara, A., Halpenny, C. M., Churcher, T. S., Mwandawiro, C., Kihara, J., Kaplan, R. M., Streit, T. G., Idaghdour, Y., Scott, M. E., Basáñez, M.-G. and Prichard, R. K. (2013). Association between response to albendazole treatment and  $\beta$ -tubulin genotype frequencies in soil-transmitted helminths. *PLoS Neglected Tropical Diseases* 7, e2247.
- Easton, A. V., Oliveira, R. G., O'Connell, E. M., Kepha, S., Mwandawiro, C. S., Njenga, S. M., Kihara, J. H., Mwatele, C., Odiero, M. R., Brooker, S. J., Webster, J. P., Anderson, R. M. and Nutman, T. B. (2016). Multi-parallel qPCR provides increased sensitivity and diagnostic breadth for gastrointestinal parasites of humans: field-based inferences on the impact of mass deworming. *Parasites & Vectors* 9, 38.
- Espy, M. J., Uhl, J. R., Sloan, L. M., Buckwalter, S. P., Jones, M. F., Vetter, E. A., Yao, J. D. C., Wengenack, N. L., Rosenblatt, J. E., Cockerill, F. R. and Smith, T. F. (2006). Real-time PCR in clinical microbiology: applications for routine laboratory testing. *Clinical Microbiology Reviews* 19, 165–256.
- Faye, O., Faye, O., Soropogui, B., Patel, P., El Wahed, A. A., Loucoubar, C., Fall, G., Kiory, D., Magassouba, N., Keita, S., Kondé, M. K., Diallo, A. A., Koivogui, L., Karlberg, H., Mirazimi, A., Nentwich, O., Piepenburg, O., Niedrig, M., Weidmann, M. and Sall, A. A. (2015). Development and deployment of a rapid recombinase polymerase amplification Ebola virus detection assay in Guinea in 2015. *Euro Surveillance* 20, 10–18.
- Fernández-Soto, P., Gandasegui Arahetes, J., Sánchez Hernández, A., López Abán, J., Vicente Santiago, B. and Muro, A. (2014). A loop-mediated isothermal amplification (LAMP) assay for early detection of *Schistosoma mansoni* in stool samples: a diagnostic approach in a murine model. *PLoS Neglected Tropical Diseases* 8, e3126.
- Fink, D. L., Fahle, G. A., Fischer, S., Fedorko, D. F. and Nutman, T. B. (2011). Toward molecular parasitologic diagnosis: enhanced diagnostic sensitivity for filarial infections in mobile populations. *Journal of Clinical Microbiology* 49, 42–47.
- Gandasegui, J., Fernández-Soto, P., Carranza-Rodríguez, C., Pérez-Arellano, J. L., Vicente, B., López-Abán, J. and Muro, A. (2015). The rapid-heat LAMPellet method: a potential diagnostic method for human urogenital Schistosomiasis. *PLoS Neglected Tropical Diseases* 9, e0003963.
- Gordon, C. A., McManus, D. P., Acosta, L. P., Olveda, R. M., Williams, G. M., Ross, A. G., Gray, D. J. and Gobert, G. N. (2015). Multiplex real-time PCR monitoring of intestinal helminths in humans reveals widespread polyparasitism in Northern Samar, the Philippines. *International Journal for Parasitology* 45, 477–483.
- Hamburger, J., Abbasi, I., Kariuki, C., Wanjala, A., Mzungu, E., Mungai, P., Muchiri, E. and King, C. H. (2013). Evaluation of loop-mediated isothermal amplification suitable for molecular monitoring of schistosome-infected snails in field laboratories. *American Journal of Tropical Medicine and Hygiene* 88, 344–351.
- Han, E.-T. (2013). Loop-mediated isothermal amplification test for the molecular diagnosis of malaria. *Expert Review of Molecular Diagnostics* 13, 205–218.
- Hapairai, L. K., Plichart, C., Naseri, T., Silva, U., Tesimale, L., Pemita, P., Bossin, H. C., Burkot, T. R., Ritchie, S. A., Graves, P. M., Melrose, W. and Joseph, H. (2015). Evaluation of traps and lures for mosquito vectors and xenomonitoring of *Wuchereria bancrofti* infection in a high prevalence Samoan Village. *Parasites & Vectors* 8, 287.
- Hotez, P. J., Pecoul, B., Rijal, S., Boehme, C., Aksoy, S., Malecela, M., Tapia-Conyer, R. and Reeder, J. C. (2016). Eliminating the neglected tropical diseases: translational science and new technologies. *PLoS Neglected Tropical Diseases* 10, e0003895.
- Intapan, P. M., Thanchomngang, T., Lulitanond, V. and Maleewong, W. (2009). Rapid detection of *Wuchereria bancrofti* and *Brugia malayi* in mosquito vectors (Diptera: Culicidae) using a real-time fluorescence resonance energy transfer multiplex PCR and melting curve analysis. *Journal of Medical Entomology* 46, 158–164.
- Irish, S. R., Stevens, W. M. B., Derua, Y. A., Walker, T. and Cameron, M. M. (2015). Comparison of methods for xenomonitoring in vectors of lymphatic filariasis in northeastern Tanzania. *American Journal of Tropical Medicine and Hygiene* 93, 983–989.
- Kagai, J. M., Mpoke, S., Muli, F., Hamburger, J. and Kenya, E. U. (2008). Molecular technique utilising sputum for detecting *Wuchereria bancrofti* infections in Malindi, Kenya. *East African Medical Journal* 85, 118–122.
- Kane, R. A., Stothard, J. R., Rollinson, D., Leclipteux, T., Evaerts, J., Standley, C. J., Allan, F., Betson, M., Kaba, R., Mertens, P. and Laurent, T. (2013). Detection and quantification of schistosome DNA in freshwater snails using either fluorescent probes in real-time PCR or oligochromatographic dipstick assays targeting the ribosomal intergenic spacer. *Acta Tropica* 128, 241–249.
- Koloren, Z., Sotiriadou, I. and Karanis, P. (2011). Investigations and comparative detection of *Cryptosporidium* species by microscopy, nested PCR and LAMP in water supplies of Ordu, Middle Black Sea, Turkey. *Annals of Tropical Medicine & Parasitology* 105, 607–615.
- Kotze, A. C., Hunt, P. W., Suce, P., von Samson-Himmelstjerna, G., Martin, R. J., Sager, H., Krücken, J., Hodgkinson, J., Lespine, A., Jex, A. R., Gilleard, J. S., Beech, R. N., Wolstenholme, A. J., Demeler, J., Robertson, A. P., Abraham, C. L., Neveu, C., Kaminsky, R., Rufener, L., Alberich, M., Menez, C. and Prichard, R. K. (2014). Recent advances in candidate-gene and whole-genome approaches to the discovery of anthelmintic resistance markers and the description of drug/receptor interactions. *International Journal for Parasitology: Drugs and Drug Resistance* 4, 164–184.
- Kouassi, B. L., de Souza, D. K., Goepogui, A., Narh, C. A., King, S. A., Mamadou, B. S., Diakité, L., Dadzie, S. K., Boakye, D. A., Utzinger, J., Bockarie, M. J. and Koudou, B. G. (2015). Assessing the presence of *Wuchereria bancrofti* in vector and human populations from urban communities in Conakry, Guinea. *Parasites & Vectors* 8, 492.
- Krolewiecki, A. J., Lammie, P., Jacobson, J., Gabrielli, A.-F., Levecke, B., Socias, E., Arias, L. M., Sosa, N., Abraham, D., Cimino, R., Echazú, A., Crudo, F., Vercruyse, J. and Albonico, M. (2013). A public health response against *Strongyloides stercoralis*: time to look at soil-transmitted helminthiasis in full. *PLOS Neglected Tropical Diseases* 7, e2165.
- Krölov, K., Frolova, J., Tudoran, O., Suhorutsenko, J., Lehto, T., Sibul, H., Mäger, I., Laanpere, M., Tulp, I. and Langel, Ü. (2014). Sensitive and rapid detection of chlamydia trachomatis by recombinase polymerase amplification directly from urine samples. *Journal of Molecular Diagnostics* 16, 127–135.
- Kumagai, T., Furushima-Shimogawara, R., Ohmae, H., Wang, T. P., Lu, S., Chen, R., Wen, L. and Ohta, N. (2010). Detection of early and single infections of *Schistosoma japonicum* in the intermediate host snail, *Oncomelania hupensis*, by PCR and loop-mediated isothermal amplification (LAMP) assay. *American Journal of Tropical Medicine and Hygiene* 83, 542–548.
- Laney, S. J., Buttaro, C. J., Visconti, S., Pilotte, N., Ramzy, R. M. R., Weil, G. J. and Williams, S. A. (2008). A reverse transcriptase-PCR assay for detecting filarial infective larvae in mosquitoes. *PLoS Neglected Tropical Diseases* 2, 1–9.
- Laney, S. J., Ramzy, R. M. R., Helmy, H. H., Farid, H. A., Ashour, A. A., Weil, G. J. and Williams, S. A. (2010). Detection of *Wuchereria bancrofti* L3 larvae in mosquitoes: a reverse transcriptase PCR assay evaluating infection and infectivity. *PLoS Neglected Tropical Diseases* 4, e602.
- Liu, J., Gratz, J., Amour, C., Kibiki, G., Becker, S., Janaki, L., Verweij, J. J., Taniuchi, M., Sobuz, S. U., Haque, R., Haverstick, D. M. and Houpt, E. R. (2013). A laboratory-developed taqman array card for simultaneous detection of 19 enteropathogens. *Journal of Clinical Microbiology* 51, 472–480.
- Lizotte, M. R., Supali, T., Partono, F. and Williams, S. A. (1994). A polymerase chain reaction assay for the detection of *Brugia malayi* in blood. *The American Journal of Tropical Medicine and Hygiene* 51, 314–321.
- Llewellyn, S., Inpankaew, T., Nery, S. V., Gray, D. J., Verweij, J. J., Clements, A. C. A., Gomes, S. J., Traub, R. and McCarthy, J. S. (2016). Application of a multiplex quantitative PCR to assess prevalence and intensity of intestinal parasite infections in a controlled clinical trial. *PLoS Neglected Tropical Diseases* 10, e0004380.
- Lloyd, M. M., Gilbert, R., Taha, N. T., Weil, G. J., Meite, A., Kouakou, I. M. M. and Fischer, P. U. (2015). Conventional parasitology

- and DNA-based diagnostic methods for onchocerciasis elimination programmes. *Acta Tropica* **146**, 114–118.
- Lucena, W. A., Dhalia, R., Abath, F. G. C., Nicolas, L., Regis, L. N. and Furtado, A. F.** (1998). Diagnosis of *Wuchereria bancrofti* infection by the polymerase chain reaction using urine and day blood samples from microfilaraemic patients. *Transactions of the Royal Society of Tropical Medicine and Hygiene* **92**, 290–293.
- Lulitanond, V., Intapan, P. M., Pipitgool, V., Choochote, W. and Maleewong, W.** (2004). Rapid detection of *Wuchereria bancrofti* in mosquitoes by LightCycler polymerase chain reaction and melting curve analysis. *Parasitology Research* **94**, 337–341.
- McCarthy, J. S., Lustigman, S., Yang, G.-J., Barakat, R. M., García, H. H., Sripan, B., Willingham, A. L., Prichard, R. K. and Basañez, M.-G.** (2012). A research agenda for Helminth diseases of humans: diagnostics for control and elimination programmes. *PLoS Neglected Tropical Diseases* **6**, e1601.
- Mehlotra, R. K., Gray, L. R., Blood-Zikursh, M. J., Kloos, Z., Henry-Halldin, C. N., Tisch, D. J., Thomsen, E., Reimer, L., Kastens, W., Baea, M., Baea, K., Baisor, M., Tarongka, N., Kazura, J. W. and Zimmerman, P. A.** (2010). Molecular-based assay for simultaneous detection of four *Plasmodium* spp. and *Wuchereria bancrofti* infections. *The American Journal of Tropical Medicine and Hygiene* **82**, 1030–1033.
- Mejia, R., Vicuña, Y., Broncano, N., Sandoval, C., Vaca, M., Chico, M., Cooper, P. J. and Nutman, T. B.** (2013). A novel, multi-parallel, real-time polymerase chain reaction approach for eight gastrointestinal parasites provides improved diagnostic capabilities to resource-limited at-risk populations. *American Journal of Tropical Medicine and Hygiene* **88**, 1041–1047.
- Mugambi, R. M., Agola, E. L., Mwangi, I. N., Kinyua, J., Shiraho, E. A. and Mkoji, G. M.** (2015). Development and evaluation of a loop mediated isothermal amplification (LAMP) technique for the detection of hookworm (*Necator americanus*) infection in fecal samples. *Parasites & Vectors* **8**, 574.
- Mugasa, C. M., Katiti, D., Boobo, A., Lubega, G. W., Schallig, H. D. F. H. and Matovu, E.** (2014). Comparison of nucleic acid sequence-based amplification and loop-mediated isothermal amplification for diagnosis of human African trypanosomiasis. *Diagnostic Microbiology and Infectious Disease* **78**, 144–148.
- Notomi, T., Okayama, H., Masubuchi, H., Yonekawa, T., Watanabe, K., Amino, N. and Hase, T.** (2000). Loop-mediated isothermal amplification of DNA. *Nucleic Acids Research* **28**, E63.
- Nzeli, C. O., Gomez, E. a., Cáceres, A. G., Sakurai, T., Martini-Robles, L., Uezato, H., Mimori, T., Katakura, K., Hashiguchi, Y. and Kato, H.** (2014). Development of a loop-mediated isothermal amplification method for rapid mass-screening of sand flies for *Leishmania* infection. *Acta Tropica* **132**, 1–6.
- Okorie, P. N. and de Souza, D. K.** (2016). Prospects, drawbacks and future needs of xenomonitoring for the endpoint evaluation of lymphatic filariasis elimination programs in Africa. *Transactions of the Royal Society of Tropical Medicine and Hygiene* **110**, 90–97.
- Owusu, I. O., de Souza, D. K., Anto, F., Wilson, M. D., Boakye, D. A., Bockarie, M. J. and Gyapong, J. O.** (2015). Evaluation of human and mosquito based diagnostic tools for defining endpoints for elimination of *Anopheles* transmitted lymphatic filariasis in Ghana. *Transactions of the Royal Society of Tropical Medicine and Hygiene* **109**, 628–635.
- Piepenburg, O., Williams, C. H., Stemple, D. L. and Armes, N. a.** (2006). DNA detection using recombination proteins. *PLoS Biology* **4**, 1115–1121.
- Pilotte, N., Zaky, W. I., Abrams, B. P., Chadee, D. D. and Williams, S. A.** (2016a). A novel Xenomonitoring Technique using mosquito excreta/feces for the detection of filarial parasites and malaria. *PLoS Neglected Tropical Diseases* **10**, e0004641.
- Pilotte, N., Papaikakovou, M., Grant, J. R., Bierwert, L. A., Llewellyn, S., McCarthy, J. S. and Williams, S. A.** (2016b). Improved PCR-based detection of soil transmitted Helminth infections using a next-generation sequencing approach to assay design. *PLoS Neglected Tropical Diseases* **10**, e0004578.
- Plichart, C. and Lemoine, A.** (2013). Monitoring and evaluation of lymphatic filariasis interventions: an improved PCR-based pool screening method for high throughput *Wuchereria bancrofti* detection using dried blood spots. *Parasites & Vectors* **6**, 110.
- Poole, C. B., Tanner, N. a, Zhang, Y., Evans, T. C. and Carlow, C. K. S.** (2012). Diagnosis of brugian filariasis by loop-mediated isothermal amplification. *PLoS Neglected Tropical Diseases* **6**, e1948.
- Rao, R. U., Atkinson, L. J., Ramzy, R. M. R., Helmy, H., Farid, H. A., Bockarie, M. J., Susapu, M., Laney, S. J., Williams, S. A. and Weil, G. J.** (2006a). A real-time PCR-based assay for detection of *Wuchereria bancrofti* DNA in blood and mosquitoes. *The American Journal of Tropical Medicine and Hygiene* **74**, 826–832.
- Rao, R. U., Weil, G. J., Fischer, K., Supali, T. and Fischer, P.** (2006b). Detection of *Brugia* parasite DNA in human blood by real-time PCR. *Journal of Clinical Microbiology* **44**, 3887–3893.
- Rao, R. U., Huang, Y., Bockarie, M. J., Susapu, M., Laney, S. J. and Weil, G. J.** (2009). A qPCR-based multiplex assay for the detection of *Wuchereria bancrofti*, *Plasmodium falciparum* and *Plasmodium vivax* DNA. *Transactions of the Royal Society of Tropical Medicine and Hygiene* **103**, 365–370.
- Rao, R. U., Nagodavithana, K. C., Samarasekera, S. D., Wijegunawardana, A. D., Premakumara, W. D. Y., Perera, S. N., Settinayake, S., Miller, J. P. and Weil, G. J.** (2014). A comprehensive assessment of lymphatic filariasis in Sri Lanka six years after cessation of mass drug administration. *PLoS Neglected Tropical Diseases* **8**, e3281.
- Rodriguez-Pérez, M. A., Danis-Lozano, R., Rodriguez, M. H., Unnasch, T. R. and Bradley, J. E.** (1999). Detection of *Onchocerca volvulus* infection in *Simulium ochraceum sensu lato*: comparison of a PCR assay and fly dissection in a Mexican hypoendemic community. *Parasitology* **119**, 613–619.
- Rollinson, D., Knopp, S., Levitz, S., Stothard, J. R., Tchuem Tchuenté, L.-A., Garba, A., Mohammed, K. A., Schur, N., Person, B., Colley, D. G. and Utzinger, J.** (2013). Time to set the agenda for schistosomiasis elimination. *Acta Tropica* **128**, 423–440.
- Rosser, A., Rollinson, D., Forrest, M. and Webster, B. L.** (2015). Isothermal Recombinase Polymerase amplification (RPA) of *Schistosoma haematobium* DNA and oligochromatographic lateral flow detection. *Parasites & Vectors* **8**, 446.
- Schwab, A. E., Boakye, D. a, Kyelem, D. and Prichard, R. K.** (2005). Detection of benzimidazole resistance-associated mutations in the filarial nematode *Wuchereria bancrofti* and evidence for selection by albendazole and ivermectin combination treatment. *The American Journal of Tropical Medicine and Hygiene* **73**, 234–238.
- Solomon, A. W., Engels, D., Bailey, R. L., Blake, I. M., Brooker, S., Chen, J. X., Chen, J. H., Churcher, T. S., Drakeley, C. J., Edwards, T., Fenwick, A., French, M., Gabrielli, A. F., Grassly, N. C., Harding-Esch, E. M., Holland, M. J., Koukounari, A., Lammie, P. J., Leslie, J., Mabey, D. C., Rhajaoui, M., Secor, W. E., Stothard, J. R., Wei, H., Willingham, A. L., Zhou, X. N. and Peeling, R. W.** (2012). A diagnostics platform for the integrated mapping, monitoring, and surveillance of neglected tropical diseases: rationale and target product profiles. *PLoS Neglected Tropical Diseases* **6**, 1–5.
- Stothard, J. R., Stanton, M. C., Bustinduy, A. L., Sousa-Figueiredo, J. C., Van Dam, G. J., Betson, M., Waterhouse, D., Ward, S., Allan, F., Hassan, A. A., Al-Helal, M. A., Memish, Z. A. and Rollinson, D.** (2014). Diagnostics for schistosomiasis in Africa and Arabia: a review of present options in control and future needs for elimination. *Parasitology* **141**, 1947–1961.
- Tagaki, H., Itoh, M., Kasai, S., Yahathugoda, T. C., Weerasooriya, M. V. and Kimura, E.** (2011). Development of loop-mediated isothermal amplification method for detecting *Wuchereria bancrofti* DNA in human blood and vector mosquitoes. *Parasitology International* **60**, 493–497.
- Taniuchi, M., Verweij, J. J., Noor, Z., Sobuz, S. U., Van Lieshout, L., Petri, W. a., Haque, R. and Houpt, E. R.** (2011). High throughput multiplex PCR and probe-based detection with luminex beads for seven intestinal parasites. *American Journal of Tropical Medicine and Hygiene* **84**, 332–337.
- Tanner, N. A., Zhang, Y. and Evans, T. C.** (2015). Visual detection of isothermal nucleic acid amplification using pH-sensitive dyes. *BioTechniques* **58**, 59–68.
- ten Hove, R. J., Verweij, J. J., Vereecken, K., Polman, K., Dieye, L. and van Lieshout, L.** (2008). Multiplex real-time PCR for the detection and quantification of *Schistosoma mansoni* and *S. haematobium* infection in stool samples collected in northern Senegal. *Transactions of the Royal Society of Tropical Medicine and Hygiene* **102**, 179–185.
- Tong, Q. B., Chen, R., Zhang, Y., Yang, G. J., Kumagai, T., Furushima-Shimogawara, R., Lou, D., Yang, K., Wen, L. Y., Lu, S. H., Ohta, N. and Zhou, X. N.** (2014). A new surveillance and response tool: risk map of infected *Oncomelania hupensis* detected by loop-mediated isothermal amplification (LAMP) from pooled samples. *Acta Tropica* **141**, 170–177.
- Vasuki, V., Hoti, S. L. and Patra, K. P.** (2008). RT-PCR assay for the detection of infective (L3) larvae of lymphatic filarial parasite, *Wuchereria bancrofti*, in vector mosquito *Culex quinquefasciatus*. *Journal of Vector Borne Diseases* **45**, 207–216.

- Verweij, J. J. and Stensvold, R. C.** (2014). Molecular testing for clinical diagnosis and epidemiological investigations of intestinal parasitic infections. *Clinical Microbiology Reviews* **27**, 371–418.
- Vincent, J. a, Lustigman, S., Zhang, S. and Weil, G. J.** (2000). A comparison of newer tests for the diagnosis of onchocerciasis. *Annals of Tropical Medicine & Parasitology* **94**, 253–258.
- Watts, M. R., James, G., Sultana, Y., Ginn, A. N., Outhred, A. C., Kong, F., Verweij, J. J., Iredell, J. R., Chen, S. C. A. and Lee, R.** (2014). A loop-mediated isothermal amplification (LAMP) assay for *Strongyloides stercoralis* in stool that uses a visual detection method with SYTO-82 fluorescent dye. *American Journal of Tropical Medicine and Hygiene* **90**, 306–311.
- World Health Organization** (2012). *Research Priorities for helminth infections: Technical Report of the TDR Disease Reference Group on Helminth Infections*. World Health Organization, Geneva.
- World Health Organization** (2015). *Investing to Overcome the Global Impact of Neglected Tropical Diseases - Third WHO Report on Neglected Tropical Diseases*. World Health Organization, Geneva.
- Xu, J., Rong, R., Zhang, H. Q., Shi, C. J., Zhu, X. Q. and Xia, C. M.** (2010). Sensitive and rapid detection of *Schistosoma japonicum* DNA by loop-mediated isothermal amplification (LAMP). *International Journal for Parasitology* **40**, 327–331.
- Xu, J., Guan, Z.-X., Zhao, B., Wang, Y.-Y., Cao, Y., Zhang, H.-Q., Zhu, X.-Q., He, Y.-K. and Xia, C.-M.** (2015). DNA detection of *Schistosoma japonicum*: diagnostic validity of a LAMP assay for low-intensity infection and effects of chemotherapy in humans. *PLoS Neglected Tropical Diseases* **9**, e0003668.
- Zhang, X., Lowe, S. B. and Gooding, J. J.** (2014). Brief review of monitoring methods for loop-mediated isothermal amplification (LAMP). *Biosensors & Bioelectronics* **61C**, 491–499.