



University of Pennsylvania
ScholarlyCommons

Departmental Papers (BE)

Department of Bioengineering

11-12-2012

Scanner-Free and Wide-Field Endoscopic Imaging by Using a Single Multimode Optical Fiber

Youngwoon Choi

Korea University

Changhyeong Yoon

Korea University

Moonseok Kim

Korea University

Taeseok Daniel Yang

Korea University

Christopher Fang-Yen

University of Pennsylvania, fangyen@seas.upenn.edu

See next page for additional authors

Follow this and additional works at: http://repository.upenn.edu/be_papers

 Part of the [Biomedical Engineering and Bioengineering Commons](#)

Recommended Citation

Choi, Y., Yoon, C., Kim, M., Yang, T., Fang-Yen, C., Dasari, R. R., Lee, K., & Choi, W. (2012). Scanner-Free and Wide-Field Endoscopic Imaging by Using a Single Multimode Optical Fiber. Retrieved from http://repository.upenn.edu/be_papers/186

Choi, Y., Yoon, C., Kim, M., Yang, T. D., Fang-Yen, C., Dasara, R. R., Lee, K. J., Choi, W. (2012). Scanner-Free and Wide-Field Endoscopic Imaging by Using a Single Multimode Optical Fiber. *Physical Review Letters*, 109(20), 203901. doi: 10.1103/PhysRevLett.109.203901

© 2012 American Physical Society

This paper is posted at ScholarlyCommons. http://repository.upenn.edu/be_papers/186

For more information, please contact libraryrepository@pobox.upenn.edu.

Scanner-Free and Wide-Field Endoscopic Imaging by Using a Single Multimode Optical Fiber

Abstract

A single multimode fiber is considered an ideal optical element for endoscopic imaging due to the possibility of direct image transmission via multiple spatial modes. However, the wave distortion induced by the mode dispersion has been a fundamental limitation. In this Letter, we propose a method for eliminating the effect of mode dispersion and therefore realize wide-field endoscopic imaging by using only a single multimode fiber with no scanner attached to the fiber. Our method will potentially revolutionize endoscopy in various fields encompassing medicine and industry.

Disciplines

Biomedical Engineering and Bioengineering

Comments

Choi, Y., Yoon, C., Kim, M., Yang, T. D., Fang-Yen, C., Dasara, R. R., Lee, K. J., Choi, W. (2012). Scanner-Free and Wide-Field Endoscopic Imaging by Using a Single Multimode Optical Fiber. *Physical Review Letters*, 109(20), 203901. doi: [10.1103/PhysRevLett.109.203901](https://doi.org/10.1103/PhysRevLett.109.203901)

© 2012 American Physical Society

Author(s)

Youngwoon Choi, Changhyeong Yoon, Moonseok Kim, Taeseok Daniel Yang, Christopher Fang-Yen, Ramchandra R. Dasari, Kyoung Jin Lee, and Wonshik Choi



Scanner-Free and Wide-Field Endoscopic Imaging by Using a Single Multimode Optical Fiber

Youngwoon Choi,¹ Changhyeong Yoon,¹ Moonseok Kim,¹ Taeseok Daniel Yang,¹ Christopher Fang-Yen,²
Ramachandra R. Dasari,³ Kyoung Jin Lee,¹ and Wonshik Choi^{1,*}

¹*Department of Physics, Korea University, Seoul 136-701, Korea*

²*Department of Bioengineering, University of Pennsylvania, Philadelphia, Pennsylvania 19104, USA*

³*G. R. Harrison Spectroscopy Laboratory, Massachusetts Institute of Technology, Cambridge, Massachusetts 02139, USA*
(Received 1 August 2012; revised manuscript received 24 September 2012; published 12 November 2012)

A single multimode fiber is considered an ideal optical element for endoscopic imaging due to the possibility of direct image transmission via multiple spatial modes. However, the wave distortion induced by the mode dispersion has been a fundamental limitation. In this Letter, we propose a method for eliminating the effect of mode dispersion and therefore realize wide-field endoscopic imaging by using only a single multimode fiber with no scanner attached to the fiber. Our method will potentially revolutionize endoscopy in various fields encompassing medicine and industry.

DOI: [10.1103/PhysRevLett.109.203901](https://doi.org/10.1103/PhysRevLett.109.203901)

PACS numbers: 42.25.Dd, 07.60.Vg, 42.30.-d

Since the invention of a rigid type endoscope, endoscopy has become a widely used tool for industrial, military, and medical purposes because of its ability to visualize otherwise inaccessible structures. In the history of technological developments, the advent of the optical fiber has been an important turning point. Especially, a multimode optical fiber has drawn interest because numerous independent spatial modes can be used for parallel information transport. However, the single multimode fiber could not be used in itself for the imaging purpose. When a light wave couples to and propagates through the fiber, the wave is distorted into a complex pattern because of mode dispersion [1]. Moreover, it has been much more challenging to use a single multimode fiber for practical endoscopic imaging in which the imaging operation should be performed in the reflection mode. In such a case, the light wave injected into the fiber is distorted twice, i.e., on the way in for the illumination and on the way out for the detection. For this reason, a fiber bundle [2–4] has been widely used and thus became a standard for commercial endoscopes. In this so-called fiber scope, each fiber constituting the bundle acts as a single pixel of an image, and the number of fibers in the bundle determines the pixel resolution. Therefore, the requirement for a large number of fibers for high resolution imaging has posed constraints on the diameter of the endoscope, thereby causing considerable limitation on the accessibility of the device.

The use of a multimode fiber for the endoscopic purpose can have a clear advantage over the fiber bundle in terms of the diameter of the imaging unit. If the numerous individual optical fibers in the bundle are replaced with the multiple propagating modes of a single fiber, the diameter of the imaging unit can be significantly reduced. The pixel density (which is the number of modes per unit area in this case) of a multimode fiber is given approximately by $\rho \sim 16/\pi(\text{NA}/\lambda)^2$, where NA is the numerical aperture of

the fiber and λ the wavelength of the light source [1]. For $\text{NA} = 0.5$, each pixel occupies an area of about λ^2 . By contrast, each fiber in a fiber bundle has a typical area of $\sim 50\text{--}100\lambda^2$. Therefore the pixel density can be 1–2 orders of magnitude greater than that possible with a fiber bundle.

In order to exploit this benefit of a multimode fiber, a number of ways were proposed to handle the image distortion. For example, the object images were encoded in spectral, spatial, and wavelength-time content and the transmitted waves were decoded to recover the original image information [5,6]. Optical phase conjugation was used to deliver image information in a double-path configuration [7,8] and to generate a focused spot [9]. However, both approaches are not suitable for endoscopic purposes in which an object image should be retrieved at the reflection mode of detection. Indeed, none of the previous studies to date could achieve the wide-field endoscopic imaging by a single multimode fiber.

We note that the concept of the recent studies enabling a disordered medium to be used as imaging optics can be extended to achieve the reflection mode of imaging. In these studies, the so-called transmission matrix (which describes the input-output response of an arbitrary optical system by connecting the free modes at the input plane to those at the output plane) of an unknown complex medium was used for the delivery of an image through the medium [10–12] and for the control of the light field at the opposite side of the disordered medium [13,14]. In fact, a multimode optical fiber can also be considered as a complex medium because bending and twisting of the fiber complicate wave propagation through the fiber. Čižmár *et al.* demonstrated that a wave-front shaping technique in conjunction with the knowledge of a transmission matrix can shape the intensity pattern at the far side of the fiber [15]. But so far, the transmission matrix approach itself was used only for the transmission mode of imaging.

In this Letter, we make use of the measured transmission matrix of a multimode optical fiber to reverse the on-the-way-out distortion of the detected light. By employing the speckle imaging method, the on-the-way-in distortion was also eliminated. In doing so, we converted the multimode fiber into a self-contained 3D imaging device that does not require a scanner or a lens. We have demonstrated wide-field endoscopic imaging by using a single multimode fiber at relatively high speed (1 frame/s for 3D imaging) with a high spatial resolution ($1.8 \mu\text{m}$). The diameter of the endoscope is then reduced to the fundamental limit—the diameter of a single fiber ($230 \mu\text{m}$ in diameter). We call this method lensless microendoscopy by a single fiber (LMSF). As a proof-of-concept study, we performed 3D imaging of a biological tissue from a rat intestine using LMSF.

Our experimental setup is shown in Fig. 1 (for more details, see the Supplemental Material [16]). A He-Ne laser with a wavelength of 633 nm is split into two by a beam splitter (BS1) and the transmitted beam is reflected off by a two-axis galvanometer scanning mirror (GM). After being reflected at a second beam splitter (BS2), the beam illuminates at the input plane (IP) of the fiber, couples to the fiber, and subsequently propagates toward the object plane (OP) located at the exit of the fiber to illuminate a target object. The OP is determined to be the conjugate plane to the pivot point of the GM used for recording the transmission matrix [see Fig. S1(b) in the Supplemental Material [16]]. Here we use a 1 m-long multimode optical fiber with $200 \mu\text{m}$ core diameter, $15 \mu\text{m}$ clad thickness ($230 \mu\text{m}$ in total diameter), and 0.48 NA. For a test object, we used either a United States Air Force (USAF) target or a biological tissue. The light reflected from the object is collected by the same optical fiber. The light collected by the fiber is guided back through the fiber to the IP, and the output image is then delivered through the beam splitters (BS2 and BS3) into the camera. The laser beam reflected by BS1




FIG. 1 (color online). Experimental scheme of LMSF. The setup is based on an interferometric phase microscope. The interference between the reflected light from the object located at the opposite side of a single multimode fiber and the reference light is recorded by a camera. A two-axis galvanometer scanning mirror (GM) controls the incident angle $(\theta_x, \theta_y)_S$ to the fiber. BS1, BS2, and BS3: beam splitters. OL: objective lens. IP: input plane of a multimode optical fiber (0.48 NA, $200 \mu\text{m}$ core diameter). OP: object plane.

is combined with the beam from the fiber to form an interference image at the camera. Using an off-axis digital holography algorithm, both amplitude and phase of the image from the fiber are retrieved [17,18].

The recorded images of the USAF target at the camera are shown in Fig. 2(c). Two processes mask the object information: (1) distortion of illumination light on the way in (IP \rightarrow OP) and (2) the scrambling of the light reflected by the object on the way out (OP \rightarrow IP). In order to overcome these two distortions, we employ methods of speckle imaging [19–21] and turbid lens imaging (TLI) [11], respectively.

First we describe the application of TLI for unscrambling the distortion for the object image. As shown in Fig. 2(b), we first characterize the input-output response of the optical fiber from the object side to the camera side (OP \rightarrow IP). We scan the incident wave in its angle $(\theta_\xi, \theta_\eta)_T$ to the fiber and record its transmitted image at the camera. A set of these angle-dependent transmission images is called the transmission matrix T [10,11,14]. We measure the transmission matrix elements up to 0.22 NA by recording 15 000 images at 500 frames/s. This characterization for the fiber transmission is performed just in 30 s. Because this characterization is performed prior to recording the reflected images of the sample, it is not included in the actual image acquisition process. Figure 2(a) shows a few representative measured transmission matrix elements of the multimode fiber. In this Letter, only the amplitude images are shown due to the




FIG. 2 (color online). Image reconstruction process. (a) Representative images of a measured transmission matrix following the scheme shown in (b). (b) Scheme of a separate setup from Fig. 1 for measuring a transmission matrix of the single fiber from OP to IP. The incident angle $(\theta_\xi, \theta_\eta)_T$ is scanned and the transmitted images are recorded. (c) Reflected object images taken by the setup shown in Fig. 1 at different illumination angles $(\theta_x, \theta_y)_S$. (d) Reconstructed object images after applying TLI method on the images in (c). Image scale is the same as that of (c). (e) Averaging of all reconstructed images in (d). Scale bars indicate $100 \mu\text{m}$ in (a),(c) and $50 \mu\text{m}$ in (e).

space limit, but the corresponding phase information is simultaneously recorded (see Fig. S2 in the Supplemental Material [16]). The transmitted patterns are random speckles and the average size of the speckle decreases as we increase the angle of illumination. This is because a high angle of illumination mostly couples to the high-order modes of the fiber.

Once the transmission matrix is recorded, we can reconstruct an object image at the OP from the distorted image recorded at the IP [each image in Fig. 2(c)] by using the TLI method (for details of the algorithm, see the Supplemental Material [16]). The process is an inversion of the transmission matrix given by the relation

$$E_{OP}(\xi, \eta) = T^{-1}E_{IP}(x, y). \quad (1)$$

Here E_{IP} and T^{-1} represent the recorded image at the IP and the inversion of the measured transmission matrix, respectively, and E_{OP} is the image at the OP. Using Eq. (1), the distortion from OP to IP is reversed. The recovered images at the OP are shown in Fig. 2(d). The reconstructed images remain devoid of object structure because the illumination light is distorted due to the propagation of the illumination from IP to OP.

Next, we deal with this distortion of the illumination light at the OP by using the speckle imaging method. We scan the illumination angle, $(\theta_x, \theta_y)_S$, at the IP as shown in Fig. 1. The change in the illumination angle causes the variation of the illumination light at the target object; in other words, a different speckle field is generated as we vary the angle $(\theta_x, \theta_y)_S$. According to the speckle imaging method, a clean object image can be acquired if we average sufficient numbers of images recorded at different speckle illuminations [20]. Then the complex speckle illumination patterns are averaged out, leaving a clean object image. In order to implement this, we record multiple object images, typically at 500 different illumination angles, $(\theta_x, \theta_y)_S$. This procedure takes 1 s, which is the actual image acquisition time of our LMSF. Since the frame rate of the camera is the major factor for determining the image acquisition time, the use of a higher speed camera will significantly reduce the data acquisition time. The representative multiple recorded images are shown in Fig. 2(c). Then the TLI method is applied to each of them in order to obtain corresponding object images at the OP [each in Fig. 2(d)]. Here each image contains object information even though they all look dissimilar due to the different speckle pattern of illumination. By adding all the intensity images in Fig. 2(d), we obtain a clear image of the object as shown in Fig. 2(e). The image processing for 500 raw images takes about 30 s when a personal computer with an Intel Core i5 CPU is used. Therefore the combination of speckle imaging and TLI methods enables single fiber microendoscopic imaging.

LMSF presents spatial resolution equivalent to incoherent imaging due to the speckle illumination, which is two times better than coherent imaging. The effective NA is

extended up to twice that covered by the transmission matrix. In our LMSF, transmission matrix is measured up to 0.22 NA, so that the diffraction limit of the imaging is about $1.8 \mu\text{m}$. In Fig. 2(e), the structures in group 8 of the USAF target (whose smallest element has a peak-to-peak distance of $2.2 \mu\text{m}$) are resolved. The field of view is about the same as the $200 \mu\text{m}$ diameter of the fiber core.

In conventional diagnostic endoscopy, medical practitioners control the object end of an endoscope to image tissues within the body. Although our method does not support this fully flexible endoscopic operation due to the variation of transmission matrix induced by the bending of the fiber, we found that the endoscopic operation is attainable up to a partially flexible degree. In order to demonstrate the flexible endoscopic operation in LMSF, we translated the end of the fiber facing a sample [Fig. 3(a)]. We measured the transmission matrix of the fiber at an initial position of the fiber end and then used the same matrix to reconstruct object images taken while moving the fiber end. For each position, we acquired 500 distorted object images and used the recorded transmission matrix to reconstruct the original object image. Figure 3(b) shows the composite image of the USAF target in which multiple images are stitched together. Each constituting image has a view field of about $200 \mu\text{m}$ in diameter and the translation ranges of the fiber end are about $800 \mu\text{m}$ in horizontal and about $1000 \mu\text{m}$ in vertical directions. Structures ranging in a wide area are all clearly visible. Although bending induced by the movement of the fiber causes a change of the transmission matrix, many of the matrix elements stay intact and contribute to the image reconstruction. According to our experiment, the LMSF is working well up to a centimeter-travel of the fiber end, confirming that our LMSF has partial flexibility for searching for the view field (see Fig. S3 in the Supplemental Material [16]).

To demonstrate the potential of LMSF as an endoscope to image biological tissue, we imaged a sample of rat




FIG. 3 (color online). Scanning of the fiber end to enlarge the view field. (a) The fiber end at the object side is translated to take images at different parts of the target object. The object images taken at different sites are reconstructed with the same transmission matrix measured at the initial position of the fiber. (b) Reconstructed images are stitched to enlarge the field of view. Scale bar indicates $100 \mu\text{m}$.




FIG. 4 (color online). Image of villus in a rat intestine tissue taken by LMSF. (a) A composite image made by several bright field images taken by conventional microscope in the transmission geometry. (b) A composite image made by several LMSF images at the same site. Scale bar, $100\ \mu\text{m}$. (c) The same image as (b), but after numerical refocusing by $40\ \mu\text{m}$ toward the fiber end. The arrows indicated by point A refer to the villus in focus at (c) while those indicated by point B refer to the villus in focus at (b).

intestine tissue. We loaded the tissue at the OP and scanned the fiber end to extend the view field as done in Fig. 3(a). Figure 4(b) is the composite image taken by the LMSF for the site where multiple villi are seen. The villi are clearly visible with their external features in excellent agreement with the conventional bright field image taken by transmission geometry [Fig. 4(a)]. In fact, the contrast of LMSF image is better than that of the bright field image. This is because the absorption by the villus is weak for transmission mode imaging while the reflection at the rough surface provides relatively strong contrast for the LMSF imaging.

LMSF can capture the 3D map of biological tissue without depth scanning of the fiber end. As shown in Eq. (1), we acquire complex field maps, $E_{\text{OP}}(\xi, \eta)$, of a specimen which contain both amplitude and phase of the reflected waves. Once the complex field is known, one can numerically solve wave propagation to a different depth [22] (see the Supplemental Material [16]). This enables numerical refocusing of an acquired image to an arbitrary depth. Figure 4(c) is the result of numerical refocusing from the original image [Fig. 4(b)] by $40\ \mu\text{m}$ toward the fiber end. The villus indicated by point A [that used to be blurred in Fig. 4(b)] comes into focus and its boundary becomes sharp. On the other hand, the villus at point B gets blurred due to the defocusing. By comparing Figs. 4(b) and 4(c), we can clearly distinguish relative depths of multiple villi in the view field. This 3D imaging ability of LMSF will eliminate the need for physical depth scanning, such that it will dramatically enhance the speed of volumetric imaging.

In conclusion, we have demonstrated partially flexible wide-field endoscopic imaging with a single multimode optical fiber without a lens or any scanning element attached to the fiber. After characterizing the transmission property of the optical fiber, unique imaging methods of TLI and speckle imaging were used to see an object through a curved optical fiber in the reflection configuration. The

image acquisition time was 1 s for the 3D imaging with 12 300 effective image pixels, and the view field was given by the core diameter of the fiber. The pixel density of LMSF is about 30 times greater than that of a typical fiber bundle ($10\ \mu\text{m}$ diameter for each pixel), and can be increased further if a full NA transmission matrix is measured.

The proposed method is applicable only for partially flexible endoscopic operation because significant bending of the fiber modifies the transmission matrix. However, LMSF will still find many new possibilities in various fields, especially in the biomedical field. In particular, it can replace the rigid type endoscopes which are widely used in the medical practices to explore and biopsy the abdomen, urinary bladder, nose, ears, and thorax. Since LMSF can miniaturize the rigid type endoscope to its fundamental limit—the diameter of the optical fiber, the accessibility of the endoscope to the target object can be significantly improved. In addition, our technique can be implemented in a needle biopsy [23,24] in such a way that a single multimode fiber is inserted through a fine needle. By imaging a tissue faced in front of the fiber, an *in situ* diagnosis will be made possible.

Our approach is fundamentally different from other single fiber endoscopy techniques [25] such as confocal endoscopy [26,27], optical coherence tomography based endoscopy [24,28], and spectrally encoded endoscopy [29]. In these techniques, a single-mode fiber is used to generate a spot at the object and the spot is scanned for image acquisition. Therefore, minimum achievable diameter of the endoscope is mainly limited by the scanning unit, which is typically larger than the fiber diameter. In comparison with a recent study in which a single multimode optical fiber is used for the image acquisition by a point scanning method [30,31], our method of wide-field imaging is advantageous in the image acquisition speed and the pixel resolution. As such, the proposed method will find its unique applications in the biomedical field and contribute to greatly improving the ability to find fatal diseases, in particular.

The authors thank Dr. Euiheon Chung for helpful discussions. This research was supported by Basic Science Research Program through National Research Foundation of Korea (NRF) funded by the Ministry of Education, Science and Technology (MEST) (Grants No. 2010-0011286, 2011-0016568, and 2012R1A2A1A01008021), and also by the National Institutes of Biomedical Imaging and Bioengineering (9P41-EB015871-26A1).

*wonshik@korea.ac.kr

- [1] B. E. A. Saleh and M. C. Teich, *Fundamentals of Photonics* (John Wiley & Sons, New York, 2007), 2nd ed.
- [2] H. H. Hopkins and N. S. Kapany, *Nature (London)* **173**, 39 (1954).

- [3] F. Koenig, J. Knittel, and H. Stepp, *Science* **292**, 1401 (2001).
- [4] B. A. Flusberg, E. D. Cocker, W. Piyawattanametha, J. C. Jung, E. L. M. Cheung, and M. J. Schnitzer, *Nat. Methods* **2**, 941 (2005).
- [5] A. A. Friesem and U. Levy, *Opt. Lett.* **2**, 133 (1978).
- [6] A. M. Tai and A. A. Friesem, *Opt. Lett.* **8**, 57 (1983).
- [7] G. J. Dunning and R. C. Lind, *Opt. Lett.* **7**, 558 (1982).
- [8] J.-Y. Son, V. I. Bobrinev, H.-W. Jeon, Y.-H. Cho, and Y.-S. Eom, *Appl. Opt.* **35**, 273 (1996).
- [9] I. N. Papadopoulos, S. Farahi, C. Moser, and D. Psaltis, *Opt. Express* **20**, 10583 (2012).
- [10] S. Popoff, G. Lerosey, M. Fink, A. C. Boccara, and S. Gigan, *Nature Commun.* **1**, 1 (2010).
- [11] Y. Choi, T. D. Yang, C. Fang-Yen, P. Kang, K. J. Lee, R. R. Dasari, M. S. Feld, and W. Choi, *Phys. Rev. Lett.* **107**, 023902 (2011).
- [12] Y. Choi, M. Kim, C. Yoon, T. D. Yang, K. J. Lee, and W. Choi, *Opt. Lett.* **36**, 4263 (2011).
- [13] I. M. Vellekoop, A. Lagendijk, and A. P. Mosk, *Nature Photon.* **4**, 320 (2010).
- [14] M. Kim, Y. Choi, C. Yoon, W. Choi, J. Kim, Q.-H. Park, and W. Choi, *Nature Photon.* **6**, 583 (2012).
- [15] T. Čižmár and K. Dholakia, *Opt. Express* **19**, 18871 (2011).
- [16] See Supplemental Material at <http://link.aps.org/supplemental/10.1103/PhysRevLett.109.203901> for more details on the experimental setup, image processing algorithms, and additional data.
- [17] T. Ikeda, G. Popescu, R. R. Dasari, and M. S. Feld, *Opt. Lett.* **30**, 1165 (2005).
- [18] W. Choi, C. Fang-Yen, K. Badizadegan, S. Oh, N. Lue, R. R. Dasari, and M. S. Feld, *Nat. Methods* **4**, 717 (2007).
- [19] M. C. Pitter, C. W. See, and M. G. Somekh, *Opt. Lett.* **29**, 1200 (2004).
- [20] Y. Park, W. Choi, Z. Yaqoob, R. R. Dasari, K. Badizadegan, and M. S. Feld, *Opt. Express* **17**, 12285 (2009).
- [21] Y. Choi, T. D. Yang, K. J. Lee, and W. Choi, *Opt. Lett.* **36**, 2465 (2011).
- [22] J. W. Goodman, *Introduction to Fourier Optics* (Roberts, Greenwood Village, CO, 2005), 3rd ed.
- [23] J. G. Fujimoto, M. E. Brezinski, G. J. Tearney, S. A. Boppart, B. Bouma, M. R. Hee, J. F. Southern, and E. A. Swanson, *Nature Med.* **1**, 970 (1995).
- [24] G. J. Tearney, M. E. Brezinski, B. E. Bourma, S. A. Boppart, C. Pitris, J. F. Southern, and J. G. Fujimoto, *Science* **276**, 2037 (1997).
- [25] D. R. Rivera, C. M. Brown, D. G. Ouzounov, I. Pavlova, D. Kobat, W. W. Webb, and C. Xu, *Proc. Natl. Acad. Sci. U.S.A.* **108**, 17598 (2011).
- [26] G. J. Tearney, R. H. Webb, and B. E. Bouma, *Opt. Lett.* **23**, 1152 (1998).
- [27] R. Kiesslich, M. Goetz, M. Vieth, P. R. Galle, and M. F. Neurath, *Nat. Clin. Pract. Oncol.* **4**, 480 (2007).
- [28] X. Li, C. Chudoba, T. Ko, C. Pitris, and J. G. Fujimoto, *Opt. Lett.* **25**, 1520 (2000).
- [29] D. Yelin, I. Rizvi, W. M. White, J. T. Motz, T. Hasan, B. E. Bouma, and G. J. Tearney, *Nature (London)* **443**, 765 (2006).
- [30] S. Bianchi and R. Di Leonardo, *Lab Chip* **12**, 635 (2012).
- [31] T. Čižmár, and K. Dholakia, *Nature Commun.* **3**, 1027 (2012).