

# Survivorship and clinical outcome of the minimally invasive Uniglide medial fixed bearing, all-polyethylene tibia, unicompartmental knee arthroplasty at a mean follow-up of 7.3 years

Forster-Horváth Cs<sup>a1</sup>, Artz N<sup>a2</sup>, Hassaballa MA<sup>a</sup>, Robinson JR<sup>a</sup>, Porteous AJ<sup>a</sup>, Murray JR<sup>a</sup>, Newman JH<sup>a3</sup>

<sup>a</sup>The Bristol Knee Group, Avon Orthopaedic Centre, Southmead Hospital, Southmead Way, Avon, Bristol, BS10 5NB, United Kingdom

Present address:

<sup>1</sup> Orthopaedic Department, University Hospital Basel, Spitalstrasse 21, 4031 Basel, Switzerland

<sup>2</sup> Institute of Sport and Exercise Science, University of Worcester, Henwick Grove, Worcester, WR2 6AJ, United Kingdom

<sup>3</sup>Litfield House Medical Center, 1 Litfield Place, Clifton Down, Bristol BS8 3LS, United Kingdom

Corresponding author:

Csaba Forster-Horvath, email: [dr\\_forster@icloud.com](mailto:dr_forster@icloud.com), telephone: +41414928222, fax: +41414928199

## Abstract

**Background:** Medial UKA performed in England and Wales represents 7 to 11% of all knee arthroplasty procedures, and is most commonly performed using mobile-bearing designs.

Fixed bearing eliminates the risk of bearing dislocation, however some studies have shown higher revision rates for all-polyethylene tibial components compared to those that utilize metal-backed implants. The aim of the study is to analyse survivorship and maximum 8-year clinical outcome of medial fixed bearing, Uniglide unicompartmental knee arthroplasty performed using an all-polyethylene tibial component with a minimal invasive approach.

**Methods:** Between 2002 and 2009, 270 medial fixed UKAs were performed in our unit. Patients were reviewed pre-operatively, 5 and 8 years post-operatively. Clinical and radiographic reviews were carried out. Patients' outcome scores (Oxford, WOMAC and American Knee Score) were documented in our database and analysed.

**Results:** Survival and clinical outcome data of 236 knees with a mean 7.3 years follow-up are reported. Every patient with less than 4.93 years follow-up underwent a revision. The patients' average age at the time of surgery was 69.5 years. The American Knee Society Pain and Function scores, the Oxford Knee Score and the WOMAC score all improved significantly. The 5 years survival rate was 94.1% with implant revision surgery as an end point. The estimated 10 years survival rate is 91.3%. 14 patients were revised before the 5 year follow-up.

Conclusion: Fixed bearing Uniglides UKA with an all-polyethylene tibial component is a valuable tool in the management of a medial compartment osteoarthritis, affording good short term survivorship.

Level of evidence IV

Keywords:

unicompartmental knee arthroplasty, unicompartmental knee arthroplasty, survival rate, revision, all-polyethylene tibial component, patient reported outcome

## 1. Introduction

TKA, UKA and HTO are accepted alternative surgical treatments for medial compartment osteoarthritis.

A recent meta-analysis comparing HTO versus UKA indicated that UKA is a more favourable technique for improving clinical outcome and relief of pain up to 10-years following surgery [1]. Survivorship did not differ significantly but there was a trend towards UKA beyond 12 years postoperatively. UKA was also associated with a lower rate of post-operative infection [2, 3]. Studies comparing UKA and TKA for treatment of medial joint OA have shown that patients with UKA achieve higher levels of post-operative function [4], range of motion [5, 6] and task specific activities such as kneeling [7] up to 10, 15 and 2 years after surgery respectively. In addition, lower mortality rates, reduced post-operative infection rates and fewer perioperative complications [8] have all been shown with UKA [9, 10].

Medial UKA performed in England and Wales represents 7 to 11% of all knee arthroplasty procedures, and is most commonly performed using mobile-bearing designs. These may have advantages in reducing linear polyethylene wear and have been shown, in some studies, to be capable of producing good long term survivorship [11]. However bearing dislocation may occur in 1-5.3% of medial UKAs [12, 13] and has been identified as the fourth most frequent mode of failure for mobile-bearing implants [14]. Fixed bearing designs have been shown in several studies to have equivalent clinical and radiographic outcomes compared to mobile-bearing implant designs at mid and long term follow up [15]. A fixed bearing eliminates the risk of bearing dislocation, however some studies have shown higher revision rates for all-polyethylene tibial components compared to those that utilize metal-backed implants [16-18].

Clear advantage of this implant is low cost. Also the non inferior performance of all-poly TKR may support its use [19, 20]. Disadvantages of all-polyethylene implant are the lack of modularity, thus care must be taken not to overstuff the joint as one simply can't downsize the bearing;

the lack of potential to change an isolated bearing during future reoperations (e.g. bicompartamental, patellofemoral OA) and the lack of uncemented option.

The aim of this study was to analyse survival and up to 8-year clinical data of fixed bearing all-polyethylene tibia Uniglides UKA and compare to literature data.

## **2. Methods**

### **2.1 Patients**

Between 2002 and 2009, 270 medial fixed bearing all-polyethylene tibia UKAs (Uniglide, Corin Group PLC, Cirencester, England, UK) were implanted in 236 patients (112 female and 124 male) were performed at our unit.

Patients were offered a UKA if they had typical anteromedial pattern osteoarthritis with radiographic evidence of full thickness loss of articular cartilage confirmed on either an AP or Rosenberg weight bearing view. All patients had a minimum of 90° knee flexion, a maximum 15° of passively correctable varus deformity, maximum 10° of fixed flexion deformity and the presence of a functioning anterior cruciate ligament. This was determined by clinical examination. In some cases varus/valgus stress X-rays were performed to confirm cartilage thickness in the lateral compartment, although this was not routinely performed.

Patients with less than 90° of flexion, severe symptomatic patellofemoral arthritis or evidence of lateral tibio-femoral osteoarthritis (more than Ahlbäck grade 1 [21]) were not offered a UKA. Fibrillation or minor circumscribed cartilage lesions of the medial aspect of the lateral femoral condyle or the patellofemoral joint were not seen as contraindications.

### **2.2 Prosthesis design**

The Uniglide (Corin Group PLC, Cirencester, UK) femoral component has a triple-radius femoral geometry and is made of titanium nitride coated cobalt chrome. It is available in cemented or uncemented form. The tibia has both fixed and mobile-bearing options. The ultra-high molecular-weight all-polyethylene tibial fixed bearing component is flat, with a central keel, which is cemented to the prepared surface of the medial tibial plateau. The tibiofemoral articulation formed is unconstrained and non-congruous (Figure 1.).

Prosthesis design

Figure 1. The Uniglide™ fixed bearing unicompartamental knee replacement

### **2.3 Surgical technique**

Depending on individual surgeon preference, the patient was either positioned as for total knee arthroplasty, with a foot rest and lateral side support or, alternatively, using a leg holding device with the lower leg hanging. All medial UKAs were performed using MIS technique with a skin incision of approximately 8 centimeters and a mini mid-vastus or a subvastus approach. The lateral compartment was inspected for evidence of arthrosis not determined radiographically. A Langenbeck retractor was placed under the patellar ligament in slight flexion. This gave a limited view, however enough to judge the distal joint surface of the lateral femoral and tibial condyle. An extra-medullary tibial jig was used to set the valgus/varus alignment and the posterior slope of the axial tibial cut. The tibial sagittal cut was made referencing from the tibial jig, aligned with the second metatarsal. A stylus was used to determine the tibial resection depth. Tibial resection was adjusted to allow easy insertion of a 7 mm spacer feeler gauge, taking into account the thinnest fixed bearing all-polyethylene tibial insert (7 mm).

An extra-medullary jig was used to set the femoral component valgus/varus and internal/external rotation. A guide rod was placed through the jig to ensure that flexion/extension of the femoral component was set parallel with the femoral shaft. In the coronal plane, the rod was set to point at a marker dot attached to the patient showing the position of the femoral head midway between the anterior superior iliac spine and pubic symphysis. The posterior femoral cut was made first and then the distal femoral condyle was reamed with the aim to carefully balance the flexion and extension gaps and to ensure that the mechanical axis was not over corrected.

To reduce the risk of cement extrusion posteriorly cement is pressed into tibia with a wet osteotome or gloved finger. Minimal cement is then applied to the all-polyethylene component. During implantation the all-polyethylene tibia is inserted at an angle so that the posterior part of the prosthesis is compressed first allowing excess cement to extrude anteriorly. Any cement that does extrude posteriorly is scraped away prior to implantation of the femur.

### **2.4 Outcome measures**

Pre-and post-operative data were collected prospectively. Either a research nurse or physiotherapist carried out follow-up in a research clinic. Patients underwent physical and radiographic examination of the knee and completed a questionnaire consisting of the Oxford Knee Score (OKS, 0 worst and 48 best), the American Knee Society Score pain and function domains (AKSS Pain, 0 worst and 50 best, AKSS Function, 0 worst, 100 best) and Western Ontario and McMaster Universities Arthritis Index (WOMAC, 60 worst and 12 best; pain domain worst 25 and best 5;

function domain worst 35 and best 7) [22-24] at 5 and 8 years postoperatively. Revision of the prosthesis was used to define survivorship.

## **2.5 Statistical Analysis**

Kaplan-Meier survival analysis was used to determine the survivorship. Only patients with known outcome were included, thus patients who died (even with unrevised implants) and were lost to FU were excluded. Outcome measures were compared using student t-test for parametric data and the Mann-Whitney U-test for non parametric data. Normal distribution was assessed using the Kolgomorov-Smirnov-test. Significance was set at a  $p < 0.05$ . SPSS version 21 and MedCalc version 14.12.0 were used for statistical analysis.

## **2.6 Ethics and registration**

The procedures followed were in accordance with the ethical standards of the responsible committee on human experimentation and with the Helsinki Declaration of 1975, as revised in 2000.

The Avon Orthopaedic Centre Knee Database South West Regional Ethics Committee number is 09/H0206/72.

## **3. Results**

### **3.1 Patient demographics**

236 of the original 270 knees could be followed-up. Seven patients (eight knees) could not be contacted and were considered lost to follow-up giving a follow-up rate of 87.4%. Twenty-three patients (26 knees) have since died from unrelated medical conditions. Eleven patients (14 knees) died with the prosthesis in situ. This was verified by contacting the patient's general practitioner or their relatives. In the other deceased patients no data regarding prosthetic revision could be identified (Table 1).

#### Overview of excluded cases

| Reason for exclusion         | Number of knees (number of patients) |
|------------------------------|--------------------------------------|
| Total                        | 270 (236)                            |
| Lost to follow-up            | 8 (7)                                |
| Died / died without revision | 26 (23) / 14 (11)                    |
| Included                     | 236 (206)                            |

Table 1. Breakdown of excluded cases

The mean follow-up was 7.33 years (SD: 1.94). Every patient with less than 4.93 years FU was revised. The average age at operation and weight at time of surgery was 68.2 +-9.9 years (range 41-87) and 81.4+-17.6 kg (range 51.2-161) respectively. Of the 236 knees followed up, 98 were right-sided, 82 were left-sided and 28 were bilateral UKAs. 233 knees had a diagnosis of primary medial compartment osteoarthritis, two were post-traumatic, and one suffered from crystal arthropathy.

### 3.2 Survivorship analysis

Figure 2 shows the survival curve for the 236 medial fixed UKA knees. The survival probability at 5-years following surgery were 94.1% and at 10-years was 91.3%. Table 2 shows life-table for the 236 implants.

Survival curve of included implants

Figure 2. Kaplan-Meier survivorship curve for 236 medial Uniglide knees. with revision as the endpoint (red lines indicate 95% upper and lower CI). Mean days = 2744.52 (95% CI = 2655.3 – 2833.7) SD = 695.8.

| <i>Year</i> | <i>Number at risk</i> | <i>Revised</i> | <i>Censored</i> | <i>Survivorship</i> | <i>Standard Error</i> |
|-------------|-----------------------|----------------|-----------------|---------------------|-----------------------|
| 0-1         | 236                   | 1              | 0               | 0.996               | 0.004                 |
| 1-2         | 235                   | 4              | 0               | 0.979               | 0.009                 |
| 2-3         | 231                   | 3              | 0               | 0.966               | 0.012                 |
| 3-4         | 228                   | 4              | 0               | 0.949               | 0.014                 |
| 4-5         | 224                   | 2              | 3               | 0.941               | 0.015                 |
| 5-6         | 219                   | 4              | 25              | 0.923               | 0.017                 |
| 6-7         | 190                   | 2              | 29              | 0.913               | 0.019                 |
| 7-8         | 159                   | 0              | 49              | 0.913               | 0.019                 |
| 8-9         | 110                   | 0              | 59              | 0.913               | 0.019                 |
| 9-10        | 51                    | 0              | 44              | 0.913               | 0.019                 |
| 10-11       | 7                     | 0              | 3               | 0.913               | 0.019                 |
| 11-12       | 4                     | 0              | 4               | 0.913               | 0.019                 |

Table 2. Life-table of the medial Uniglide prosthesis in 236 knees

### 3.3 Revisions

Altogether twenty of the 236 knees were revised. The mean time to revision was 3.45 years (SD 1.78 ). Eighteen knees were revised to TKA, fifteen of them to standard primary implants (Genesis II, Smith and Nephew, Memphis, TN, USA; Triathlon, Kalamazoo, MI, USA). The revision cases, with cause and timing of revision and revision prosthesis used are shown in Table 3.

One patient was revised to a TKA at 5.6 years in the private sector and was subsequently lost to follow-up with no data being available as to the reason for revision.

### Summary of revised Uniglide fixed bearing cases

| <b>Patient / Age (years) / Gender</b> | <b>Time to revision/ failure (years)</b> | <b>New implant</b>                              | <b>Cause of revision</b>   |
|---------------------------------------|--|---|--|
| 1 / 78 / F                            | 0.5                                      | Genesis II                                      | Periprosthetic fracture  |
| 2 / 51 / F                            | 1.1                                      | Triathlon                                       | After 3.5 months revised to one size smaller femoral component for malalignment. At 13.5 months revised to Triathlon due to constant pain  |
| 3 / 63 / F                            | 1.2                                      | Triathlon                                       | Progression of OA lateral and patellofemoral   |
| 4 / 58 / F                            | 1.5                                      | Uniglide fixed bearing                          | Femoral rotational malalignment, pain  |
| 5 / 72 / M                            | 1.5                                      | Genesis II                                      | Pain, implant in excellent condition   |
| 6 / 67 / M                            | 2.2                                      | Triathlon                                       | Persistent pain  |
| 7 / 54 / F                            | 2.5                                      | Triathlon                                       | Aseptic loosening of the tibial component  |
| 8 / 82 / M                            | 2.9                                      | Triathlon with patella                          | Progressive OA lateral and patellofemoral and tibial aseptic loosening   |
| 9 / 64 / M                            | 3.1                                      | Triathlon                                       | Aseptic loosening of both components   |
| 10 / 55 / F                           | 3.2                                      | Genesis II                                      | Aseptic loosening of both components   |
| 11 / 55 / F                           | 3.5                                      | Genesis II                                      | Rotational malalignment and pain   |
| 12 / 72 / M                           | 3.6                                      | Uniglide Fixed bearing                          | At 16 months arthroscopic synovial biopsy, removal of osteophytes/ anterior scar tissue and loose body. At 43.5 months revised to fixed bearing Uniglide for synovitis and femoral component wear. |
| 13 / 72 / F                           | 3.6                                      | Genesis II with Legion stemmed tibial baseplate | Progression of OA lateral  |
| 14 / 61 / F                           | 4.5                                      | Genesis II                                      | Progression of OA patellofemoral   |
| 15 / 53 / M                           | 5.0                                      | Triathlon                                       | Aseptic loosening of the tibial component  |
| 16 / 66 / M                           | 5.1                                      | Legion  | Aseptic loosening  |
| 17 / 52 / M                           | 5.2                                      | Triathlon                                       | Progression of OA lateral  |

|             |     |                              |  |
|-------------|-----|------------------------------|--|
| 18 / 66 / M | 5.6 | Revised to<br>TKR            | No data  |
| 19 / 71 / F | 6.4 | Genesis II                   | Progression of OA                                    |
| 20 / 80 / F | 6.7 | RT plus<br>rotating<br>hinge | Progression of OA, incompetent MCL<br>preoperatively |

Table 3. Summary of revised Uniglides fixed bearing cases

### 3.4 Patient reported and clinical outcome measures

Table 4 shows pre-operative and 5-year post-operative OKS, WOMAC and AKSS pain and function scores. Samples were tested for normality by the Kolmogorov-Smirnov test. None of the samples showed normality except for the preoperative OKS and WOMAC scores. Differences between pre- and postoperative scores were found to be significant for each of the scores calculated by Wilcoxon non-parametric test ( $p < 0.05$ ).

#### Pre- and postoperative patient reported outcome measures

| Mean (CI95%)                            | AKSS Pain                 | AKSS Function              | OKS                        | WOMAC Pain                 | WOMAC Function             | WOMAC Total               |
|---|---------------------------|----------------------------|----------------------------|----------------------------|----------------------------|---------------------------|
| PREOPERATIVE<br>(270 KNEES)             | 8.5<br>(7.1-<br>9.9)      | 55.2 (53.0-<br>57.3)       | 19.9 (18.94-<br>20.86)     | 15.52<br>(12.34-<br>18.71) | 21.47<br>(20.81-<br>22.12) | 37 (36.0-<br>38.0)        |
| 5 YEARS<br>POSTOPERATIVE<br>(228 KNEES) | 40.1<br>(37.91-<br>42.35) | 76.68<br>(73.23-<br>80.15) | 37.4 (35.8-<br>39)         | 8.19 (7.56-<br>8.82)       | 11.96<br>(11.14-<br>12.87) | 20.2<br>(18.68-<br>21.62) |
| 8 YEARS<br>POSTOPERATIVE<br>(106 KNEES) | 42.5<br>(38.41-<br>46.59) | 70.6 (61.1-<br>80.1)       | 34.22<br>(30.22-<br>38.22) | 9.1 (7.32-<br>10.88)       | 14.1<br>(11.43-<br>16.77)  | 23.1<br>(18.64-<br>27.56) |

Table 4. Pre- and 5-year postoperative patient reported outcome measures differed significantly (all  $p < 0.0001$ )

## 4. Discussion

This study reports the short-term survivorship and outcome of 236 patients with a fixed medial unicompartmental knee arthroplasty.

In our series a survival estimate of 97.9% at 2-years, 94.1% at 5-years, and 91.3% at 10-years was calculated following medial fixed bearing UKA. Our follow up rate of 97.1% is similar to other studies of this nature [25, 26]. Recent studies with all-polyethylene tibial components showed worse or similar survival rates to ours [27, 28]. Studies reporting on the survivorship of other fixed bearing implants (Zimmer I and II, Marmor, St Georg, Brigham), showed that 10-year survival rates ranged between 80 and 93.7% [29]. A study on the St. Georg sled showed 85.9% survival at 18-20 years with revision as the endpoint in patients with a mean age at operation of 67 years [30]. Pennington et al. reported on patients younger than 60 years showing 92% survival at 10 years with Miller-Galante



implants [31]. 94% of the same implants done by a single surgeon survived with a mean age at surgery of 66.54 years [32]. 10-year cumulative revision rates of approximately 12-13.5 % and 11.78 % (CI 95% 11.80-13.34) were reported for fixed bearing UKR in the 2014 Swedish Knee Arthroplasty Register (SKAR) and the 11<sup>th</sup> annual report of the National Joint Registry England and Wales (NJR) respectively. Medial and lateral UKAs show similar survival rates based on recent studies [26, 33, 34].

Although several studies have reported similar survival data for UKA and TKA [4, 35, 36], NJR data show significantly worse survival for UKAs. The lower survivorship of UKA compared with TKA has been a cause for concern for some surgeons leading them to avoid UKA. However of the 20 cases in our study cohort that have been revised, 90% were revised to either a standard primary TKR prosthesis or further UKR. Only 2 cases required a formal revision prosthesis system (one a Legion and another an RT plus (Smith and Nephew, Memphis, TN, USA).

The lower survival rates for UKA may relate to surgeons having a lower threshold to offer revision to patients with a problematic UKA. Revisions for unexplained knee pain may be partly responsible for the increased revisions of UKA compared to TKA [37]. Goodfellow et al. investigated the management of patients with poor Oxford Knee Scores following UKA and TKA. In patients with OKS <20 12% of TKAs were revised, whereas 63% of the UKRs with similar scores were revised [38]. The post-operative improvements in patient reported outcomes measured in our study appeared to be consistent with other UKA studies in the literature [39]. In our study, the mean 5-year postoperative OKS and improvement compared to the preoperative level were higher than those reported in the National Joint Registry for England and Wales 8<sup>th</sup> Annual Report, [37]. 27 of our cases reported a post-operative OKS <20. Of the 20 patients revised 12 had an OKS of <20. Mean OKS in this cohort was 19.2 (CI 95% 12.3-26.1).

The reasons for revision in our study are consistent with published mechanisms of UKA failure [14, 25, 37, 40, 41], which include aseptic loosening, arthritis progression and unexplained knee pain. The most frequent reason for revision in our study was progression of OA. This occurred in 6 cases and most frequently the lateral compartment was affected. Aseptic loosening was the second most common reason for revision in 5 of the 20 revised cases. This occurred most commonly on the tibial side. One patient was affected by both progression of OA and aseptic loosening of the tibial component. Five patients in our series were revised for knee pain. In our experience the majority of unexplained knee pain following UKA will settle with time. Important causes of pain not related to the prosthesis should be excluded, such as neuromas of the infra-patellar branch of the saphenous nerve [42]. We apply and propose

an oblique minimal invasive skin incision avoiding the course of the infra-patellar branch of the saphenous nerve for this procedure. In some cases MRI can be helpful in exploring the source of pain [43]. Indicating the operation with full thickness cartilage loss has been shown to reduce revision rate [44, 45].

It has been suggested that increased strain on the antero-medial tibial cortical bone is a cause of discomfort, which settles with gradual osseous remodeling [46]. The increased tibial strain associated with all-polyethylene tibial designs might be a cause of pain and aseptic loosening [47] and some studies have reported inferior clinical outcome and survival [16-18, 48, 49]. These findings are also supported by biomechanical studies [47]. Also inferior alignment with these implants has been described compared to metal-backed ones [50]. Yet even other studies have shown comparable survival rates [51, 52]. The results for the Uniglides fixed bearing all-polyethylene tibia, presented in this study are good, suggesting that good results with an all-polyethylene tibia may be design specific.

Older series of fixed bearing UKAs reported polyethylene wear as a typical mode of failure. Linear wear rates have been estimated at 0.15 mm/year for fixed bearing versus 0.04 mm/year in mobile bearing implants [11, 53]. However, recent NJR reports describe similar revision rates for fixed and mobile bearing UKAs, where the cause for revision was described as polyethylene wear [34]. This may be attributed to the introduction of ultra-high molecular-weight polyethylene and improved sterilization methods and shelf life with improved wear resistance [54] counteracting higher compression and tensile stress in non-congruous fixed bearing designs. Wear rates may be further reduced with improved surface coating of the femoral component. The cobalt chrome femoral component of the Uniglides UKA used in the study has a titanium nitride coating which has been shown to reduce wear of both polyethylene and metal counterparts in vitro [55].

There is some contrast in orthopedic centers 10 years survival rate reports and those of joint registry data according to Labek et al. [56]. Registry data for the Uniglides has to be carefully assessed as the Uniglides has both fixed and mobile bearing options and the registry reports do not generally distinguish between the two designs.

The reasons explaining our superior results may be the somewhat higher mean age, standardized surgical technique in a unit with 20 years experience with fixed bearing UKR including attention not to (over)correct the mechanical axis. The operation was only performed by high volume surgeons in a tertiary referral center unit.

Another possible explanation is the long learning curve. Surgeons with less than 23 cases per year produce significantly lower survival rates (Swedish Knee Arthroplasty Register 2004) [57]. Rees et al. showed that the average American Knee Society Score of the first 10 cases was significantly lower than that of the subsequent ones Rees 2004. Appropriate patient selection and correct clinical indication for UKA are

important factors that could affect survivorship. Pre-operative cartilage thickness in the medial compartment appears to affect survival since re-revision rate was found to be 6-fold for knees where there was more than 2 mm of joint space preserved preoperatively [45].

Limitations of the study are the relatively short follow up, and the multiple surgeon design, however 89,3 % of the cases were performed by the senior author (JHN).

To our knowledge this is the first study reporting on survivorship of the all-polyethylene fixed bearing Uniglide UKA prosthesis.

## **5. Conclusion**

The Uniglide UKA with all-polyethylene tibial component provides a relatively low cost option for UKA and our results demonstrate satisfactory patients outcomes and survivorship rates comparable to other bearing designs.

## **6. Acknowledgement**

The authors would like to thank Suzanne Miller and Rachel Bray for their contribution to this work.

## References

1. Spahn G, Hofmann GO, von Engelhardt LV, Li M, Neubauer H, Klinger HM. The impact of a high tibial valgus osteotomy and unicondylar medial arthroplasty on the treatment for knee osteoarthritis: a meta-analysis. *Knee Surg Sports Traumatol Arthrosc* 2013;21:96-112.
2. Osti M, Gohm A, Schlick B, Benedetto KP. Complication rate following high tibial open-wedge osteotomy with spacer plates for incipient osteoarthritis of the knee with varus malalignment. *Knee Surg Sports Traumatol Arthrosc* 2015;23:1943-1948.
3. Morris MJ, Molli RG, Berend KR, Lombardi AV, Jr. Mortality and perioperative complications after unicompartmental knee arthroplasty. *Knee* 2013;20:218-220.
4. Newman J, Pydisetty RV, Ackroyd C. Unicompartmental or total knee replacement: the 15-year results of a prospective randomised controlled trial. *J Bone Joint Surg Br* 2009;91:52-57.
5. Ackroyd CE, Whitehouse SL, Newman JH, Joslin CC. A comparative study of the medial St Georg sled and kinematic total knee arthroplasties. Ten-year survivorship. *J Bone Joint Surg Br* 2002;84:667-672.
6. Longo UG, Loppini M, Trovato U, Rizzello G, Maffulli N, Denaro V. No difference between unicompartmental versus total knee arthroplasty for the management of medial osteoarthritis of the knee in the same patient: a systematic review and pooling data analysis. *Br Med Bull* 2015;114:65-73.
7. Hassaballa MA, Porteous AJ, Newman JH, Rogers CA. Can knees kneel? Kneeling ability after total, unicompartmental and patellofemoral knee arthroplasty. *Knee* 2003;10:155-160.
8. Brown NM, Sheth NP, Davis K, Berend ME, Lombardi AV, Berend KR, et al. Total knee arthroplasty has higher postoperative morbidity than unicompartmental knee arthroplasty: a multicenter analysis. *J Arthroplasty* 2012;27:86-90.
9. Hunt LP, Ben-Shlomo Y, Clark EM, Dieppe P, Judge A, MacGregor AJ, et al. 45-day mortality after 467,779 knee replacements for osteoarthritis from the National Joint Registry for England and Wales: an observational study. *Lancet* 2014;384:1429-1436.
10. Liddle AD, Judge A, Pandit H, Murray DW. Adverse outcomes after total and unicompartmental knee replacement in 101,330 matched patients: a study of data from the National Joint Registry for England and Wales. *Lancet* 2014;384:1437-1445.
11. Psychoyios V, Crawford RW, O'Connor JJ, Murray DW. Wear of congruent meniscal bearings in unicompartmental knee arthroplasty: a retrieval study of 16 specimens. *J Bone Joint Surg Br* 1998;80:976-982.
12. Saxler G, Temmen D, Bontemps G. Medium-term results of the AMC-unicompartmental knee arthroplasty. *Knee* 2004;11:349-355.
13. Choy WS, Kim KJ, Lee SK, Yang DS, Lee NK. Mid-term results of oxford medial unicompartmental knee arthroplasty. *Clin Orthop Surg* 2011;3:178-183.
14. National Joint Registry for England WaNI. 11th annual report. 2014.

15. Parratte S, Pauly V, Aubaniac JM, Argenson JN. No long-term difference between fixed and mobile medial unicompartmental arthroplasty. *Clin Orthop Relat Res* 2012;470:61-68.
16. Saenz CL, McGrath MS, Marker DR, Seyler TM, Mont MA, Bonutti PM. Early failure of a unicompartmental knee arthroplasty design with an all-polyethylene tibial component. *Knee* 2010;17:53-56.
17. Furnes O, Espehaug B, Lie SA, Vollset SE, Engesaeter LB, Havelin LI. Failure mechanisms after unicompartmental and tricompartmental primary knee replacement with cement. *J Bone Joint Surg Am* 2007;89:519-525.
18. Gladnick BP, Nam D, Khamaisy S, Paul S, Pearle AD. Onlay tibial implants appear to provide superior clinical results in robotic unicompartmental knee arthroplasty. *HSS J* 2015;11:43-49.
19. Dojcinovic S, Ait Si Selmi T, Servien E, Verdonk PC, Neyret P. [A comparison of all-polyethylene and metal-backed tibial components in total knee arthroplasty]. *Rev Chir Orthop Reparatrice Appar Mot* 2007;93:364-372.
20. Norgren B, Dalen T, Nilsson KG. All-poly tibial component better than metal-backed: a randomized RSA study. *Knee* 2004;11:189-196.
21. Ahlbäck S. Osteoarthritis of the knee. A radiographic investigation. *Acta Radiol Diagn (Stockh)* 1968;Suppl 277:7-72.
22. Dawson J, Fitzpatrick R, Murray D, Carr A. Questionnaire on the perceptions of patients about total knee replacement. *J Bone Joint Surg Br* 1998;80:63-69.
23. Insall JN, Dorr LD, Scott RD, Scott WN. Rationale of the Knee Society clinical rating system. *Clin Orthop Relat Res* 1989:13-14.
24. Bellamy N, Buchanan WW, Goldsmith CH, Campbell J, Stitt LW. Validation study of WOMAC: a health status instrument for measuring clinically important patient relevant outcomes to antirheumatic drug therapy in patients with osteoarthritis of the hip or knee. *J Rheumatol* 1988;15:1833-1840.
25. Foran JR, Brown NM, Della Valle CJ, Berger RA, Galante JO. Long-term survivorship and failure modes of unicompartmental knee arthroplasty. *Clin Orthop Relat Res* 2013;471:102-108.
26. Smith JR, Robinson JR, Porteous AJ, Murray JR, Hassaballa MA, Artz N, et al. Fixed bearing lateral unicompartmental knee arthroplasty--short to midterm survivorship and knee scores for 101 prostheses. *Knee* 2014;21:843-847.
27. Bruni D, Gagliardi M, Akkawi I, Raspugli GF, Bignozzi S, Marko T, et al. Good survivorship of all-polyethylene tibial component UKA at long-term follow-up. *Knee Surg Sports Traumatol Arthrosc* 2014.
28. Manzotti A, Cerveri P, Pullen C, Confalonieri N. A flat all-polyethylene tibial component in medial unicompartmental knee arthroplasty: a long-term study. *Knee* 2014;21 Suppl 1:S20-25.
29. Murray DW, Goodfellow JW, O'Connor JJ. The Oxford medial unicompartmental arthroplasty: a ten-year survival study. *J Bone Joint Surg Br* 1998;80:983-989.
30. Steele RG, Hutabarat S, Evans RL, Ackroyd CE, Newman JH. Survivorship of the St Georg Sled medial unicompartmental knee replacement beyond ten years. *J Bone Joint Surg Br* 2006;88:1164-1168.

31. Pennington DW, Swienckowski JJ, Lutes WB, Drake GN. Unicompartmental knee arthroplasty in patients sixty years of age or younger. *J Bone Joint Surg Am* 2003;85-A:1968-1973.
32. John J, Mauffrey C, May P. Unicompartmental knee replacements with Miller-Galante prosthesis: two to 16-year follow-up of a single surgeon series. *Int Orthop* 2011;35:507-513.
33. Ollivier M, Abdel MP, Parratte S, Argenson JN. Lateral unicondylar knee arthroplasty (UKA): contemporary indications, surgical technique, and results. *Int Orthop* 2014;38:449-455.
34. Baker PN, Jameson SS, Deehan DJ, Gregg PJ, Porter M, Tucker K. Mid-term equivalent survival of medial and lateral unicondylar knee replacement: an analysis of data from a National Joint Registry. *J Bone Joint Surg Br* 2012;94:1641-1648.
35. Price AJ, Dodd CA, Svard UG, Murray DW. Oxford medial unicompartmental knee arthroplasty in patients younger and older than 60 years of age. *J Bone Joint Surg Br* 2005;87:1488-1492.
36. Svard UC, Price AJ. Oxford medial unicompartmental knee arthroplasty. A survival analysis of an independent series. *J Bone Joint Surg Br* 2001;83:191-194.
37. Baker PN, Petheram T, Jameson SS, Avery PJ, Reed MR, Gregg PJ, et al. Comparison of patient-reported outcome measures following total and unicondylar knee replacement. *J Bone Joint Surg Br* 2012;94:919-927.
38. Goodfellow JW, O'Connor JJ, Murray DW. A critique of revision rate as an outcome measure: re-interpretation of knee joint registry data. *J Bone Joint Surg Br* 2010;92:1628-1631.
39. Pandit H, Jenkins C, Gill HS, Barker K, Dodd CA, Murray DW. Minimally invasive Oxford phase 3 unicompartmental knee replacement: results of 1000 cases. *J Bone Joint Surg Br* 2011;93:198-204.
40. Citak M, Cross MB, Gehrke T, Dersch K, Kendoff D. Modes of failure and revision of failed lateral unicompartmental knee arthroplasties. *Knee* 2015;22:338-340.
41. Saragaglia D, Bonnin M, Dejour D, Deschamps G, Chol C, Chabert B, et al. Results of a French multicentre retrospective experience with four hundred and eighteen failed unicondylar knee arthroplasties. *Int Orthop* 2013;37:1273-1278.
42. Harris JD, Fazalare JJ, Griesser MJ, Flanigan DC. Infrapatellar branch of saphenous neurectomy for painful neuroma: a case report. *Am J Orthop (Belle Mead NJ)* 2012;41:37-40.
43. Park CN, Zuiderbaan HA, Chang A, Khamaisy S, Pearle AD, Ranawat AS. Role of magnetic resonance imaging in the diagnosis of the painful unicompartmental knee arthroplasty. *Knee* 2015;22:341-346.
44. Maier MW, Kuhs F, Streit MR, Schuhmacher P, Walker T, Ewerbeck V, et al. Unicompartmental knee arthroplasty in patients with full versus partial thickness cartilage loss (PTCL): equal in clinical outcome but with higher reoperation rate for patients with PTCL. *Arch Orthop Trauma Surg* 2015.
45. Niinimäki TT, Murray DW, Partanen J, Pajala A, Leppilähti JI. Unicompartmental knee arthroplasties implanted for osteoarthritis with partial loss of joint space have high re-operation rates. *Knee* 2011;18:432-435.

46. Simpson DJ, Price AJ, Gulati A, Murray DW, Gill HS. Elevated proximal tibial strains following unicompartmental knee replacement--a possible cause of pain. *Med Eng Phys* 2009;31:752-757.
47. Scott CE, Eaton MJ, Nutton RW, Wade FA, Pankaj P, Evans SL. Proximal tibial strain in medial unicompartmental knee replacements: A biomechanical study of implant design. *Bone Joint J* 2013;95-B:1339-1347.
48. Murray DW, MacLennan GS, Breeman S, Dakin HA, Johnston L, Campbell MK, et al. A randomised controlled trial of the clinical effectiveness and cost-effectiveness of different knee prostheses: the Knee Arthroplasty Trial (KAT). *Health Technol Assess* 2014;18:1-235, vii-viii.
49. Hutt JR, Farhadnia P, Masse V, Lavigne M, Vendittoli PA. A randomised trial of all-polyethylene and metal-backed tibial components in unicompartmental arthroplasty of the knee. *Bone Joint J* 2015;97-B:786-792.
50. Suero EM, Citak M, Njoku IU, Pearle AD. Does the type of tibial component affect mechanical alignment in unicompartmental knee replacement? *Technol Health Care* 2013;21:81-85.
51. Lustig S, Paillot JL, Servien E, Henry J, Ait Si Selmi T, Neyret P. Cemented all polyethylene tibial insert unicompartmental knee arthroplasty: a long term follow-up study. *Orthop Traumatol Surg Res* 2009;95:12-21.
52. O'Rourke MR, Gardner JJ, Callaghan JJ, Liu SS, Goetz DD, Vittetoe DA, et al. The John Insall Award: unicompartmental knee replacement: a minimum twenty-one-year followup, end-result study. *Clin Orthop Relat Res* 2005;440:27-37.
53. Ashraf T, Newman JH, Desai VV, Beard D, Nevelos JE. Polyethylene wear in a non-congruous unicompartmental knee replacement: a retrieval analysis. *Knee* 2004;11:177-181.
54. McGovern TF, Ammeen DJ, Collier JP, Currier BH, Engh GA. Rapid polyethylene failure of unicondylar tibial components sterilized with gamma irradiation in air and implanted after a long shelf life. *J Bone Joint Surg Am* 2002;84-A:901-906.
55. Pappas MJ, Makris G, Buechel FF. Titanium nitride ceramic film against polyethylene. A 48 million cycle wear test. *Clin Orthop Relat Res* 1995:64-70.
56. Labek G, Sekyra K, Pawelka W, Janda W, Stockl B. Outcome and reproducibility of data concerning the Oxford unicompartmental knee arthroplasty: a structured literature review including arthroplasty registry data. *Acta Orthop* 2011;82:131-135.
57. Rees JL, Price AJ, Beard DJ, Dodd CA, Murray DW. Minimally invasive Oxford unicompartmental knee arthroplasty: functional results at 1 year and the effect of surgical inexperience. *Knee* 2004;11:363-367.